

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043204.D
 Acq On : 18 Nov 2025 10:02
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:49:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	105095	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	283830	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	234772	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	172524	10.025	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.300%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.496	85	140789	9.498	ppbv	100
3) Chlorodifluoromethane	1.473	51	138443	9.497	ppbv	100
4) Chloromethane	1.531	50	52372	9.450	ppbv	100
5) Vinyl Chloride	1.577	62	51045	8.422	ppbv	100
6) Bromomethane	1.664	94	26542	9.034	ppbv	100
7) Chloroethane	1.700	64	20375	8.632	ppbv	100
8) Dichlorotetrafluoroethane	1.551	85	114670	9.300	ppbv	100
9) Propene	1.480	41	52671	7.789	ppbv #	1
10) Heptane	4.965	43	168611	9.090	ppbv	100
11) Trichlorofluoromethane	1.875	101	130362	9.077	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.137	101	103679	9.247	ppbv	100
13) Ethanol	1.716	45	17464	8.023	ppbv	100
14) Bromoethene	1.777	108	38037	8.849	ppbv	100
15) Acetone	1.832	43	108422	7.426	ppbv	100
16) 1,3-Butadiene	1.606	39	57871	8.890	ppbv	100
17) tert-Butyl alcohol	2.036	59	179201	9.350	ppbv	100
18) 1,1-Dichloroethene	2.030	96	46313	8.871	ppbv	100
19) Isopropyl Alcohol	1.881	45	71862	8.710	ppbv	100
20) Methylene Chloride	2.056	84	41868	9.708	ppbv	100
21) Allyl Chloride	2.091	41	89266	9.087	ppbv	100
22) trans-1,2-Dichloroethene	2.334	96	52470	8.793	ppbv	100
23) Vinyl Acetate	2.457	43	170459	9.670	ppbv	100
24) 1,1-Dichloroethane	2.405	63	106970	9.287	ppbv	100
25) Ethyl Acetate	2.829	43	289921	9.190	ppbv	100
26) Hexane	2.820	57	129809	9.402	ppbv	100
27) Carbon Disulfide	2.146	76	132086	9.371	ppbv	100
28) Methyl tert-Butyl Ether	2.434	73	62115	9.165	ppbv	100
29) Chloroform	2.842	83	187620	9.561	ppbv	100
30) Cyclohexane	3.846	84	107621	9.297	ppbv	100
31) cis-1,2-Dichloroethene	2.716	61	130357	9.499	ppbv	100
32) 1,1,1-Trichloroethane	3.360	97	198036	9.381	ppbv	100
34) 2-Butanone	2.551	43	182749	9.121	ppbv	100
35) Carbon Tetrachloride	3.755	117	200831	9.322	ppbv	100
36) Benzene	3.648	78	256954	9.453	ppbv	100
37) 1,2-Dichloroethane	3.211	62	150769	9.483	ppbv	100
38) Trichloroethene	4.535	130	100136	9.398	ppbv	100
39) 1,2-Dichloropropane	4.302	63	95097	9.559	ppbv	100
40) 1,4-Dioxane	4.567	88	40441	9.043	ppbv #	100
41) Tetrahydrofuran	3.046	42	104294	9.426	ppbv	100
42) Bromodichloromethane	4.477	83	211504	9.643	ppbv	100
43) Methyl Methacrylate	4.865	69	131003	9.080	ppbv	100
44) 2,2,4-Trimethylpentane	4.626	57	441234	9.451	ppbv	100
45) t-1,3-Dichloropropene	6.399	75	129122	9.601	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043204.D
 Acq On : 18 Nov 2025 10:02
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:49:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.584	75	151655	9.603	ppbv	100
47) 1,1,2-Trichloroethane	6.564	97	97512	9.193	ppbv	100
48) Dibromochloromethane	7.377	129	175529	9.670	ppbv	100
49) Bromoform	9.474	173	157563	9.800	ppbv	100
50) 4-Methyl-2-Pentanone	5.733	43	268369	9.509	ppbv	100
51) 2-Hexanone	7.419	43	208268	9.503	ppbv #	100
52) Tetrachloroethene	8.212	164	93518	9.249	ppbv	100
53) Toluene	6.920	91	305169	9.422	ppbv	100
54) 1,2-Dibromoethane	7.639	107	148973	9.588	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.914	131	158856	9.207	ppbv	100
57) Chlorobenzene	8.911	112	227490	9.438	ppbv	100
58) Ethyl Benzene	9.345	91	427530	9.461	ppbv	100
59) m/p-Xylene	9.542	91	673371m	18.519	ppbv	
60) o-Xylene	9.956	91	333044	9.202	ppbv	100
61) Styrene	9.863	104	152172	9.607	ppbv	100
62) Isopropylbenzene	10.533	105	600155	9.269	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.953	83	209272	9.109	ppbv	100
64) n-propylbenzene	10.982	120	154239	9.271	ppbv	100
65) tert-Butylbenzene	11.523	119	522106	9.029	ppbv	100
66) Benzyl Chloride	11.607	91	56620	9.649	ppbv	100
67) sec-Butylbenzene	11.762	105	729573	9.039	ppbv	100
69) p-Isopropyltoluene	11.918	119	624586	9.135	ppbv	100
70) n-Butylbenzene	12.274	91	619610	9.341	ppbv	100
71) 2-Chlorotoluene	10.895	91	460118	9.245	ppbv	100
72) 4-Ethyltoluene	11.118	105	410587	9.316	ppbv	100
73) 1,3,5-Trimethylbenzene	11.196	105	341900	9.244	ppbv	100
74) 1,2,4-Trimethylbenzene	11.529	105	363476	9.005	ppbv	100
75) 1,3-Dichlorobenzene	11.601	146	219894	9.335	ppbv	100
76) 1,4-Dichlorobenzene	11.665	146	220576	9.363	ppbv	100
77) 1,2-Dichlorobenzene	11.940	146	204826	9.044	ppbv	100
78) Hexachloro-1,3-Butadiene	13.876	225	161034	8.520	ppbv	100
79) Naphthalene	13.507	128	351148	9.795	ppbv	100
80) Naphthalene,2-methyl-	14.452	142	93113	9.895	ppbv	100
81) 1,2,4-Trichlorobenzene	13.436	180	159091	9.923	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

