



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043202.D
Acq On : 18 Nov 2025 08:46
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Wed Nov 19 01:11:25 2025



AutoFind: Scans 3175, 3176, 3177; Background Corrected with Scan 3156

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	22.1	30285	PASS
75	95	30	66	55.6	76296	PASS
95	95	100	100	100.0	137266	PASS
96	95	5	9	6.7	9227	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	63.6	87269	PASS
175	174	4	9	7.8	6847	PASS
176	174	93	101	95.5	83301	PASS
177	176	5	9	6.6	5509	PASS

Average of 10.364 to 10.371 min.: VL043202.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
33.15	1017	45.05	1373	56.05	2485	67.10	650
33.80	90	45.95	190	57.05	3511	68.05	1515
36.05	1728	47.00	799	58.00	209	69.05	1587
37.00	7775	47.20	474	60.00	1493	70.00	1182
38.05	6647	47.95	600	61.00	7430	70.90	18
39.00	2637	48.20	316	62.05	6404	71.20	48
40.05	176	49.10	6876	63.10	5348	72.05	772
41.00	228	50.05	30285	63.95	548	73.05	6279
42.20	99	51.05	9432	65.10	18	74.10	25637
43.00	63	52.00	422	65.75	54	75.10	76296
44.05	598	55.05	162	66.10	23	76.10	6470

Average of 10.364 to 10.371 min.: VL043202.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
77.05	853	88.00	3931	102.80	24	109.85	96
77.95	601	88.70	128	103.10	20	110.85	101
79.00	3536	90.95	202	103.90	564	111.80	44
79.95	911	92.05	3581	104.90	156	112.20	28
81.00	3819	93.05	5411	105.75	294	114.70	70
81.90	690	94.10	15989	106.00	240	114.95	67
82.90	77	95.05	137266	106.95	132	115.95	360
83.10	136	96.05	9227	107.65	48	116.80	153
84.15	40	97.00	116	108.10	27	116.95	414
86.00	131	97.90	24	109.00	61	117.90	356
87.05	4162	102.10	22	109.70	38	118.95	710

Average of 10.364 to 10.371 min.: VL043202.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
123.10	19	135.05	222	145.80	142	157.00	250
123.90	99	135.90	30	146.10	100	158.60	28
125.20	19	136.95	182	146.80	34	159.00	120
125.70	56	138.90	65	147.10	127	161.00	78
126.10	58	139.90	157	147.90	406	170.90	87
127.85	394	141.00	1242	148.90	165	172.20	73
128.10	59	141.90	141	149.95	156	174.00	87269
129.00	325	142.95	1524	151.00	22	175.05	6847
129.95	344	143.70	75	151.80	76	176.00	83301
130.95	175	144.00	78	152.90	62	177.00	5509
134.60	21	145.00	87	154.80	296	177.95	166

Average of 10.364 to 10.371 min.: VL043202.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
207.00	28						
268.30	20						
294.80	17						