



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_L\Data\VL120825\  
Data File : VL043304.D  
Acq On : 08 Dec 2025 07:56  
Operator : SY/MD  
Sample : BFB  
Misc : 400mL/MSVOA\_L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_L\methods\VL111825AIR.M  
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Fri Aug 26 06:05:16 2022  
Last Update : Wed Nov 19 01:11:25 2025



AutoFind: Scans 3176, 3177, 3178; Background Corrected with Scan 3162

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 8               | 40              | 22.8         | 30677      | PASS                |
| 75             | 95              | 30              | 66              | 55.8         | 74976      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 134320     | PASS                |
| 96             | 95              | 5               | 9               | 6.7          | 8968       | PASS                |
| 173            | 174             | 0.00            | 2               | 0.9          | 774        | PASS                |
| 174            | 95              | 50              | 120             | 67.2         | 90299      | PASS                |
| 175            | 174             | 4               | 9               | 7.8          | 7010       | PASS                |
| 176            | 174             | 93              | 101             | 93.8         | 84736      | PASS                |
| 177            | 176             | 5               | 9               | 6.8          | 5800       | PASS                |

Average of 10.367 to 10.374 min.: VL043304.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 33.20 | 412    | 47.05 | 1211   | 56.10 | 2151   | 65.85 | 71     |
| 36.05 | 1537   | 47.85 | 883    | 57.05 | 3791   | 66.75 | 29     |
| 37.10 | 8108   | 49.05 | 6574   | 58.00 | 87     | 67.05 | 22     |
| 38.00 | 7724   | 50.05 | 30677  | 58.15 | 144    | 68.05 | 1458   |
| 39.00 | 2675   | 51.00 | 9026   | 60.00 | 1280   | 69.05 | 14873  |
| 41.20 | 58     | 51.95 | 352    | 61.00 | 6682   | 70.05 | 1219   |
| 42.70 | 142    | 52.70 | 80     | 62.10 | 6686   | 70.90 | 41     |
| 43.20 | 18     | 53.15 | 172    | 63.05 | 5301   | 71.15 | 39     |
| 44.05 | 1520   | 54.50 | 37     | 63.80 | 190    | 71.60 | 36     |
| 45.00 | 1337   | 54.80 | 241    | 64.10 | 735    | 71.95 | 673    |
| 46.05 | 161    | 55.10 | 407    | 65.05 | 876    | 72.20 | 234    |

Average of 10.367 to 10.374 min.: VL043304.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z    | abund. | m/z    | abund. |
|-------|--------|-------|--------|--------|--------|--------|--------|
| 73.00 | 5829   | 81.85 | 925    | 93.05  | 6657   | 110.20 | 17     |
| 74.05 | 25208  | 82.90 | 215    | 94.10  | 15627  | 111.80 | 208    |
| 75.10 | 74976  | 83.25 | 61     | 95.10  | 134320 | 112.80 | 82     |
| 76.10 | 6025   | 86.00 | 142    | 96.10  | 8968   | 113.10 | 18     |
| 76.90 | 491    | 86.90 | 1029   | 97.05  | 342    | 114.70 | 38     |
| 77.30 | 59     | 87.10 | 2128   | 102.95 | 92     | 115.00 | 94     |
| 77.80 | 177    | 88.05 | 3657   | 103.95 | 578    | 116.00 | 619    |
| 78.05 | 291    | 90.00 | 128    | 104.95 | 165    | 116.95 | 795    |
| 78.95 | 3704   | 90.70 | 92     | 105.95 | 748    | 117.80 | 107    |
| 79.90 | 1113   | 91.05 | 427    | 106.90 | 143    | 118.00 | 233    |
| 81.00 | 3576   | 92.05 | 4014   | 109.70 | 22     | 118.95 | 739    |

Average of 10.367 to 10.374 min.: VL043304.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|--------|--------|--------|--------|--------|--------|--------|--------|
| 123.90 | 119    | 141.00 | 1621   | 149.95 | 311    | 171.10 | 184    |
| 125.95 | 44     | 141.95 | 220    | 152.00 | 33     | 172.00 | 1272   |
| 127.10 | 22     | 142.95 | 1671   | 152.95 | 137    | 172.90 | 774    |
| 127.95 | 650    | 143.90 | 18     | 155.00 | 333    | 173.10 | 296    |
| 129.10 | 355    | 144.10 | 40     | 155.90 | 33     | 174.00 | 90299  |
| 129.90 | 284    | 144.95 | 216    | 156.40 | 21     | 175.05 | 7010   |
| 130.10 | 164    | 145.60 | 49     | 156.80 | 36     | 176.00 | 84736  |
| 130.90 | 198    | 146.05 | 272    | 157.00 | 194    | 177.05 | 5800   |
| 131.80 | 35     | 146.95 | 146    | 158.90 | 191    | 178.00 | 190    |
| 135.00 | 227    | 147.75 | 244    | 160.85 | 110    | 208.80 | 33     |
| 137.00 | 260    | 148.90 | 86     | 170.80 | 55     | 233.70 | 17     |

Average of 10.367 to 10.374 min.: VL043304.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z | abund. | m/z | abund. | m/z | abund. |
|--------|--------|-----|--------|-----|--------|-----|--------|
| 235.45 | 47     |     |        |     |        |     |        |
| 250.20 | 42     |     |        |     |        |     |        |
| 268.10 | 26     |     |        |     |        |     |        |

Instrument :

MSVOA\_L

ClientSampleId :

BFB