

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GFEL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3798</u>
Instrument ID:	<u>MSVOA_L</u>	Calibration Date/Time:	<u>12/08/2025 08:44</u>
Lab File ID:	<u>VL043305.D</u>	Init. Calib. Date(s):	<u>11/18/2025 11/18/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:29 13:24</u>
GC Column:	<u>RTX-1</u>	ID:	<u>0.32</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.577	0.526		-8.84	30
1,1-Dichloroethane	1.096	1.133		3.38	30
cis-1,2-Dichloroethene	1.306	1.287		-1.46	30
1,1,1-Trichloroethane	1.990	2.167		8.89	30
Trichloroethene	0.375	0.424		13.07	30
Tetrachloroethene	0.356	0.385		8.15	30
1-Bromo-4-Fluorobenzene	0.733	0.753		2.73	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.