

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL111825

Review By	Semsettin Yesilyurt	Review On	11/19/2025 8:33:42 AM		
Supervise By	Mahesh Dadoda	Supervise On	11/19/2025 9:44:32 AM		
SubDirectory	VL111825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2691
Initial Calibration Stds	AP2694,AP2695,AP2696
CCC	AP2694
Internal Standard/PEM	AP2691
ICV/I.BLK	AP2697
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL043202.D	18 Nov 2025 08:46		SY/MD	Ok
2	VSTDIC0.5	VSTDIC0.5	VL043203.D	18 Nov 2025 09:29		SY/MD	Ok,M
3	VSTDICCC010	VSTDICCC010	VL043204.D	18 Nov 2025 10:02	LR-20	SY/MD	Ok,M
4	VSTDIC002	VSTDIC002	VL043205.D	18 Nov 2025 10:37		SY/MD	Ok,M
5	VSTDIC001	VSTDIC001	VL043206.D	18 Nov 2025 11:10		SY/MD	Ok,M
6	VSTDIC0.5	VSTDIC0.5	VL043207.D	18 Nov 2025 11:43	Not used	SY/MD	Not Ok
7	VSTDIC0.1	VSTDIC0.1	VL043208.D	18 Nov 2025 12:16		SY/MD	Ok,M
8	VSTDIC0.03	VSTDIC0.03	VL043209.D	18 Nov 2025 12:49		SY/MD	Ok,M
9	VSTDIC015	VSTDIC015	VL043210.D	18 Nov 2025 13:24		SY/MD	Ok,M
10	VSTDICV010	ICVVL111825	VL043211.D	18 Nov 2025 13:56	ICV failed for com.#66	SY/MD	Ok,M
11	VL1118ABL01	VL1118ABL01	VL043212.D	18 Nov 2025 14:41		SY/MD	Ok
12	VL1118ABS01	VL1118ABS01	VL043213.D	18 Nov 2025 15:24		SY/MD	Ok,M
13	Q3651-03	IA1	VL043214.D	18 Nov 2025 15:57	Need 20X	SY/MD	Dilution
14	Q3651-03DUP	IA1DUP	VL043215.D	18 Nov 2025 16:31		SY/MD	Ok,M
15	Q3651-04	IA2	VL043216.D	18 Nov 2025 17:06	Need 10X	SY/MD	Dilution
16	Q3651-06	IA3	VL043217.D	18 Nov 2025 17:40	Need 10X	SY/MD	Dilution
17	Q3650-03	IA1	VL043218.D	18 Nov 2025 18:15		SY/MD	Ok,M
18	Q3650-04	IA2	VL043219.D	18 Nov 2025 18:49		SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL111825

Review By	Semsettin Yesilyurt	Review On	11/19/2025 8:33:42 AM		
Supervise By	Mahesh Dadoda	Supervise On	11/19/2025 9:44:32 AM		
SubDirectory	VL111825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME	STD REF.#				
Tune/Reschk	AP2691				
Initial Calibration Stds	AP2694,AP2695,AP2696				
CCC	AP2694				
Internal Standard/PEM	AP2691				
ICV/I.BLK	AP2697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3650-05	IA3	VL043220.D	18 Nov 2025 19:23		SY/MD	Ok,M
20	Q3650-06	OA1	VL043221.D	18 Nov 2025 19:58		SY/MD	Ok,M
21	Q3651-01	SV1	VL043222.D	18 Nov 2025 20:29		SY/MD	Ok,M
22	Q3651-02	SV2	VL043223.D	18 Nov 2025 21:01		SY/MD	Ok,M
23	Q3651-05	SV3	VL043224.D	18 Nov 2025 21:33		SY/MD	Ok,M
24	Q3650-01	SV1	VL043225.D	18 Nov 2025 22:05		SY/MD	Ok,M
25	Q3650-02	SV2	VL043226.D	18 Nov 2025 22:37		SY/MD	Ok,M
26	Q3651-03DL	IA1DL	VL043227.D	18 Nov 2025 23:09		SY/MD	Ok,M
27	Q3651-03DL	IA1DL	VL043228.D	18 Nov 2025 23:41	Not Required	SY/MD	Not Ok
28	Q3651-04DL	IA2DL	VL043229.D	19 Nov 2025 00:13		SY/MD	Ok,M
29	Q3651-06DL	IA3DL	VL043230.D	19 Nov 2025 00:44		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL120825

Review By	Semsettin Yesilyurt	Review On	12/9/2025 8:08:46 AM		
Supervise By	John Carlone	Supervise On	12/9/2025 1:27:26 PM		
SubDirectory	VL120825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2692
Initial Calibration Stds	AP2694,AP2695,AP2696
CCC	AP2694
Internal Standard/PEM	AP2692
ICV/I.BLK	AP2697
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL043304.D	08 Dec 2025 07:56		SY/MD	Ok
2	VSTDCCC010	VSTDCCC010	VL043305.D	08 Dec 2025 08:44		SY/MD	Ok,NS
3	VL1208ABL01	VL1208ABL01	VL043306.D	08 Dec 2025 09:36		SY/MD	Ok
4	VL1208ABS01	VL1208ABS01	VL043307.D	08 Dec 2025 10:24		SY/MD	Ok,NS
5	QC120225 CAN 10297	QC120225 CAN 10297	VL043308.D	08 Dec 2025 11:11		SY/MD	Ok
6	QC120425 CAN 10606	QC120425 CAN 10606	VL043309.D	08 Dec 2025 11:44		SY/MD	Ok
7	Q3799-01	AA-1	VL043310.D	08 Dec 2025 12:40		SY/MD	Ok,NS
8	Q3799-01DUP	AA-1DUP	VL043311.D	08 Dec 2025 13:16		SY/MD	Ok,NS
9	Q3799-02	9B-IA-1	VL043312.D	08 Dec 2025 13:52		SY/MD	Ok,NS
10	Q3799-03	10-IA-1	VL043313.D	08 Dec 2025 14:29	Need 10X	SY/MD	Dilution
11	Q3799-04	11-IA-1	VL043314.D	08 Dec 2025 15:37		SY/MD	Ok,NS
12	Q3799-05	13-IA-1	VL043315.D	08 Dec 2025 16:25		SY/MD	Ok,NS
13	Q3799-06	16-IA-1	VL043316.D	08 Dec 2025 17:00	Need 10X	SY/MD	Dilution
14	Q3797-02	IA1	VL043317.D	08 Dec 2025 17:36		SY/MD	Ok,NS
15	Q3797-04	IA2	VL043318.D	08 Dec 2025 18:11		SY/MD	Ok,NS
16	Q3797-05	OA1	VL043319.D	08 Dec 2025 18:46		SY/MD	Ok,NS
17	Q3798-02	IA1	VL043320.D	08 Dec 2025 19:21		SY/MD	Ok,NS
18	Q3798-03	OA1	VL043321.D	08 Dec 2025 19:56		SY/MD	Ok,NS

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL120825

Review By	Semsettin Yesilyurt	Review On	12/9/2025 8:08:46 AM		
Supervise By	John Carlone	Supervise On	12/9/2025 1:27:26 PM		
SubDirectory	VL120825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME	STD REF.#				
Tune/Reschk	AP2692				
Initial Calibration Stds	AP2694,AP2695,AP2696				
CCC	AP2694				
Internal Standard/PEM	AP2692				
ICV/I.BLK	AP2697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3799-03DL	10-IA-1DL	VL043322.D	08 Dec 2025 20:50	Need 40X;Not Match exactly please confirm	SY/MD	Dilution
20	Q3799-03DL2	10-IA-1DL2	VL043323.D	08 Dec 2025 21:23	Need further dilution, rerun 400X	SY/MD	Not Ok
21	Q3799-06DL	16-IA-1DL	VL043324.D	08 Dec 2025 21:56		SY/MD	Ok
22	Q3797-01	SV1	VL043325.D	08 Dec 2025 22:28		SY/MD	Ok,NS
23	Q3797-01DL	SV1DL	VL043326.D	08 Dec 2025 23:01	Not Required	SY/MD	Not Ok
24	Q3797-03	SV2	VL043327.D	08 Dec 2025 23:34		SY/MD	Ok,NS
25	Q3797-03DL	SV2DL	VL043328.D	09 Dec 2025 00:06	Not Required	SY/MD	Not Ok
26	Q3798-01	SV1	VL043329.D	09 Dec 2025 00:38		SY/MD	Ok,NS
27	Q3798-01DL	SV1DL	VL043330.D	09 Dec 2025 01:11	Not Required	SY/MD	Not Ok
28	Q3799-03DL2	10-IA-1DL2	VL043331.D	09 Dec 2025 07:30	Not match exactly please check	SY/MD	Ok,NR

M : Manual Integration