



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3798
Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	SV1							
Q3798-01	SV1	Air	Tetrachloroethene	1.49		0.41	0.81	ug/m3
			Total Voc :	1.49				
			Total Concentration:	1.49				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client: GFE LLC
Project: 339 Flatbush Ave Brooklyn NY
Client Sample ID: SV1
Lab Sample ID: Q3798-01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/04/25
Date Received: 12/04/25
SDG No.: Q3798
Matrix: Air
Test: VOCMS Group2

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-01-4	Vinyl Chloride	0.10	0.26	U	4	0.26	0.31	12/09/25 00:38	VL120825
75-34-3	1,1-Dichloroethane	0.52	2.10	U	4	2.10	8.09	12/09/25 00:38	VL120825
156-59-2	cis-1,2-Dichloroethene	0.40	1.59	U	4	1.59	7.93	12/09/25 00:38	VL120825
71-55-6	1,1,1-Trichloroethane	0.060	0.33	U	4	0.33	0.65	12/09/25 00:38	VL120825
79-01-6	Trichloroethene	0.10	0.54	U	4	0.54	0.64	12/09/25 00:38	VL120825
127-18-4	Tetrachloroethene	0.22	1.49		4	0.41	0.81	12/09/25 00:38	VL120825
SURROGATES									
460-00-4	1-Bromo-4-Fluorobenzene	10.3				65 - 135	103% SPK: 10		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	136000							
540-36-3	1,4-Difluorobenzene	344000							
3114-55-4	Chlorobenzene-d5	270000							

U = Not Detected
RL = Reporting Limit
MDL = Method Detection Limit
E = Value Exceeds Calibration Range
D = Dilution

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client: GFE LLC
Project: 339 Flatbush Ave Brooklyn NY
Client Sample ID: IA1
Lab Sample ID: Q3798-02
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/04/25
Date Received: 12/04/25
SDG No.: Q3798
Matrix: Air
Test: VOCMS Group2

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
------------	-----------	---------------	----------------	------	----	--------------	---------------------	-----------	---------

TARGETS

75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 19:21	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 19:21	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 19:21	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 19:21	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 19:21	VL120825
127-18-4	Tetrachloroethene	0.020	0.14	U	1	0.14	0.20	12/08/25 19:21	VL120825

SURROGATES

460-00-4	1-Bromo-4-Fluorobenzene	10.1				65 - 135	101% SPK: 10		
----------	-------------------------	------	--	--	--	----------	--------------	--	--

INTERNAL STANDARDS

INTERNAL STANDARDS		Area Count
74-97-5	Bromochloromethane	132000
540-36-3	1,4-Difluorobenzene	343000
3114-55-4	Chlorobenzene-d5	273000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client: GFE LLC
Project: 339 Flatbush Ave Brooklyn NY
Client Sample ID: OA1
Lab Sample ID: Q3798-03
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/04/25
Date Received: 12/04/25
SDG No.: Q3798
Matrix: Air
Test: VOCMS Group2

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
------------	-----------	---------------	----------------	------	----	--------------	---------------------	-----------	---------

TARGETS

75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 19:56	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 19:56	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 19:56	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 19:56	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 19:56	VL120825
127-18-4	Tetrachloroethene	0.020	0.14	U	1	0.14	0.20	12/08/25 19:56	VL120825

SURROGATES

460-00-4	1-Bromo-4-Fluorobenzene	9.90				65 - 135	99% SPK: 10		
----------	-------------------------	------	--	--	--	----------	-------------	--	--

INTERNAL STANDARDS

		Area Count
74-97-5	Bromochloromethane	134000
540-36-3	1,4-Difluorobenzene	341000
3114-55-4	Chlorobenzene-d5	271000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client: GFE LLC
Project: 339 Flatbush Ave Brooklyn NY
Client Sample ID: AA-1DUP
Lab Sample ID: Q3799-01DUP
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/08/25
Date Received: 12/08/25
SDG No.: Q3798
Matrix: Air
Test: VOCMS Group2

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 13:16	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 13:16	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 13:16	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 13:16	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 13:16	VL120825
127-18-4	Tetrachloroethene	0.020	0.14	U	1	0.14	0.20	12/08/25 13:16	VL120825
SURROGATES									
460-00-4	1-Bromo-4-Fluorobenzene	9.80				65 - 135	98% SPK: 10		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	155000							
540-36-3	1,4-Difluorobenzene	398000							
3114-55-4	Chlorobenzene-d5	319000							

U = Not Detected
RL = Reporting Limit
MDL = Method Detection Limit
E = Value Exceeds Calibration Range
D = Dilution

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
Q = indicates LCS control criteria did not meet requirements

LAB CHRONICLE

OrderID:	Q3798	OrderDate:	12/5/2025 12:44:00 PM
Client:	GFE LLC	Project:	339 Flatbush Ave Brooklyn NY
Contact:	Frank Galdun	Location:	Air Lab,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3798-01	SV1	Air	VOCMS Group2	TO-15	12/04/25		12/09/25	12/04/25
Q3798-02	IA1	Air	VOCMS Group2	TO-15	12/04/25		12/08/25	12/04/25
Q3798-03	OA1	Air	VOCMS Group2	TO-15	12/04/25		12/08/25	12/04/25