

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3798</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>GFE LLC</u>	Datafile :	<u>VL043307.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL1208ABS01	Vinyl Chloride	10	9.20	ppbv	92			70	130	
	1,1-Dichloroethane	10	10.5	ppbv	105			70	130	
	cis-1,2-Dichloroethene	10	9.90	ppbv	99			70	130	
	1,1,1-Trichloroethane	10	10.9	ppbv	109			70	130	
	Trichloroethene	10	11.5	ppbv	115			70	130	
	Tetrachloroethene	10	11.1	ppbv	111			70	130	