

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GFEL01

Lab Code: ACE

SDG NO.: Q3798

Lab File ID: VL043202.D

BFB Injection Date: 11/18/2025

Instrument ID: MSVOA_L

BFB Injection Time: 08:46

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.1
75	30.0 - 66.0% of mass 95	55.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 120.0% of mass 95	63.6
175	4.0 - 9.0% of mass 174	5 (7.8) 1
176	93.0 - 101.0% of mass 174	60.7 (95.5) 1
177	5.0 - 9.0% of mass 176	4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC0.5	VSTDICC0.5	VL043203.D	11/18/2025	09:29
VSTDICCC010	VSTDICCC010	VL043204.D	11/18/2025	10:02
VSTDICC002	VSTDICC002	VL043205.D	11/18/2025	10:37
VSTDICC001	VSTDICC001	VL043206.D	11/18/2025	11:10
VSTDICC0.1	VSTDICC0.1	VL043208.D	11/18/2025	12:16
VSTDICC0.03	VSTDICC0.03	VL043209.D	11/18/2025	12:49
VSTDICC015	VSTDICC015	VL043210.D	11/18/2025	13:24