

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GFEL01

Lab Code: ACE

SDG NO.: Q3798

Lab File ID: VL043304.D

BFB Injection Date: 12/08/2025

Instrument ID: MSVOA_L

BFB Injection Time: 07:56

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.8
75	30.0 - 66.0% of mass 95	55.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.6 (0.9) 1
174	50.0 - 120.0% of mass 95	67.2
175	4.0 - 9.0% of mass 174	5.2 (7.8) 1
176	93.0 - 101.0% of mass 174	63.1 (93.8) 1
177	5.0 - 9.0% of mass 176	4.3 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL043305.D	12/08/2025	08:44
VL1208ABL01	VL1208ABL01	VL043306.D	12/08/2025	09:36
VL1208ABS01	VL1208ABS01	VL043307.D	12/08/2025	10:24
AA-1DUP	Q3799-01DUP	VL043311.D	12/08/2025	13:16
IA1	Q3798-02	VL043320.D	12/08/2025	19:21
OA1	Q3798-03	VL043321.D	12/08/2025	19:56
SV1	Q3798-01	VL043329.D	12/09/2025	00:38