

**DATA PACKAGE
VOLATILE ORGANICS**

PROJECT NAME : CHADWICK SQUARE

**RTP ENVIRONMENTAL
239 US Highway 22 East**

**Green Brook, NJ - 08812
Phone No: 732-968-9600**

**ORDER ID : Q3799
ATTENTION : David Gabel**



Laboratory Certification ID # 20012



1) VOLATILES DATA	2
2) Signature Page	4
3) Shipping Document	5
3.1) Chain Of Custody	6
3.2) Lab Certificate	12
3.3) Internal COC	13
3.4) Air sample Pressure And Dilution Log Book	14
3.5) Storage Location	16
4) Case Narrative	19
5) Qualifier Page	21
6) QA Checklist	22
7) Chronicle	23
8) Hit Summary	24
9) Air Conversion Table	27
9.1) MDL	45
10) QC Data Summary For TO-15	50
10.1) Deuterated Monitoring Compound Summary	51
10.2) LCS/LCSD Summary	52
10.3) Duplicate Sample Summary	54
10.4) Method Blank Summary	56
10.5) GS/MS Tune Summary	57
10.6) Internal Standard Area and RT Summary	59
11) Sample Data	60
11.1) AA-1	61
11.2) 9B-IA-1	75
11.3) 10-IA-1	89
11.4) 10-IA-1DL	105
11.5) 10-IA-1DL2	114
11.6) 11-IA-1	122
11.7) 13-IA-1	137
11.8) 16-IA-1	151
11.9) 16-IA-1DL	168
12) Calibration Data Summary	177
12.1) 12.1) Initial Calibration Data	178
12.2) 12.1.1) VL111825	178

Table Of Contents for Q3799

12.3) Continued Calibration Data	464
12.3.1) VL043305.D	464
13) QC Sample Data	514
13.1) Tune Raw Data	515
13.2) Method Blank Data	517
13.3) LCS Data	523
13.4) DUP Data	569
14) Manual Integration	583
15) Analytical Runlogs	590
16) Standard Prep Logs	598
17) Canister Data	610
18) Not Reviewed Data	676

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : Q3799

Project ID : CHADWICK SQUARE

Client : RTP Environmental

Lab Sample Number

Q3799-01
Q3799-02
Q3799-03
Q3799-04
Q3799-05
Q3799-06

Client Sample Number

AA-1
9B-IA-1
10-IA-1
11-IA-1
13-IA-1
16-IA-1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:30 pm, Dec 15, 2025

Date: 12/12/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Client Contact Information						Bottle Order ID : B2512003				Courier :				1 of 6 COCs															
Client ID : RTPE01 Project ID : Air						Sampler Name(s) :				Analysis				Matrix															
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :						Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																			
						Phone Number : 732-968-9600																							
						Fax Number :																							
						Site Details: <i>Chalnick</i>																							
Analysis Turnaround Time						Data Package Type : <i>NJ Regulatory</i>																							
Standard : 10 business days OR						EDD Type : <i>NJ Haz/Excel w/ conversion</i>																							
Rush (Specify): _____ Days						table in excel.																							
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas												
AA-1	12/3-4/2025	1140	1140	-30	-7	35	35	 	-3.9	10520	10280	6 L	4.16	VL043290.D	X		X												
Temperature (Fahrenheit) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				GC/MS Analyst Signature (TO-15) <i>[Signature]</i>							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Pressure (Inches of Hg) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Special Instructions/QC Requirements & Comments :																													
Suspected Contamination: High Medium Low PID Readings:																													
Sampling site (State):																													
Quick Connector required : <i>NO</i>																													
Canisters Shipped by: <i>Sam</i>				Date/Time: <i>12/2/25</i>				Canisters Received by: <i>[Signature]</i>				Date/Time: <i>12/2/25</i>				B2512003 - 2													
Samples Relinquished by: <i>CJ</i>				Date/Time:				Received by:				Date/Time: <i>12-4-25</i> <i>500</i>																	
Relinquished by:				Date/Time:				Received by:				Date/Time:																	

Client Contact Information					Bottle Order ID : B2512003					Courier :					<u>2</u> of <u>6</u> COCs																
Client ID : RTPE01 Project ID : Air					Sampler Name(s) :					Analysis					Matrix																
Customer Name : RTP Environmental Address : 239 US Highway 22 East					Project Manager : David gabel					AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																					
					Phone Number : 732-968-9600																										
					Fax Number :																										
					Site Details: <i>Chadwick</i>																										
City : Green Brook					Analysis Turnaround Time					TO-15 Indoor/Ambient Air Soil Gas																					
State : NJ					Standard : 10 business days OR															Data Package Type :											
Zip Code : 08812					Rush (Specify): Days															EDD Type :											
Country :																															
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID																	
9B-IA-1	12/3-4/25	1115	1115	-30	-5	60	60	 	-5.5	10771	10608	6 L	4.16	VL043290.D	X			X													
Temperature (Fahrenheit) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th></th> <th>Ambient</th> <th>Maximum</th> <th>Minimum</th> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table> GC/MS Analyst Signature (TO-15)											Ambient	Maximum	Minimum	Start				Stop				*** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.									
	Ambient	Maximum	Minimum																												
Start																															
Stop																															
Pressure (Inches of Hg) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th></th> <th>Ambient</th> <th>Maximum</th> <th>Minimum</th> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop													
	Ambient	Maximum	Minimum																												
Start																															
Stop																															
Special Instructions/QC Requirements & Comments : Suspected Contamination: High Medium Low PID Readings: Sampling site (State):																															
Quick Connector required : <i>NO</i>																															
Canisters Shipped by: <i>Sam</i>					Date/Time: <i>12/2/25</i>					Canisters Received by: <i>Sam</i>					Date/Time: <i>12-4-25 1500</i>					B2512003 - 6											
Samples Relinquished by:					Date/Time:					Received by:					Date/Time:																
Relinquished by:					Date/Time:					Received by:					Date/Time:																

Client Contact Information				Bottle Order ID : B2512003				Courier :				3 of 6 COCs							
Client ID : RTPE01 Project ID : Air				Sampler Name(s) :				Analysis				Matrix							
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified											
				Phone Number : 732-968-9600															
				Fax Number :															
				Site Details: <i>Chadwick</i>															
Analysis Turnaround Time				Data Package Type :															
Standard : 10 business days OR				EDD Type :															
Rush (Specify): Days																			
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15			Indoor/Ambient Air	Soil Gas
10-IA1	12/5/11	1120	1120	-30	-6	70	70	-30	-7.5	10207	10283	6 L	4.16	VL043259.D	X			X	
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) <i>[Signature]</i>									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Special Instructions/QC Requirements & Comments :																			
Suspected Contamination: High Medium Low PID Readings:																			
Sampling site (State):																			
Quick Connector required : <i>No</i>																			
Canisters Shipped by: <i>Sam</i>				Date/Time: <i>12/4/11</i>				Canisters Received by: <i>[Signature]</i>				Date/Time: <i>12-4-25 1500</i>				B2512003 - 3			
Samples Relinquished by: <i>/</i>				Date/Time:				Received by:				Date/Time:							
Relinquished by:				Date/Time:				Received by:				Date/Time:							

Client Contact Information				Bottle Order ID : B2512003				Courier :				4 of 6 COCs																	
Client ID : RTPE01				Project ID : Air				Sampler Name(s) :				Analysis		Matrix															
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :				Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Individual Certified																					
				Phone Number : 732-968-9600																									
				Fax Number :																									
				Site Details: <i>Chedrick</i>																									
Analysis Turnaround Time				Standard : 10 business days OR				Data Package Type :																					
Rush (Specify): Days				EDD Type :																									
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambinet Air	Soil Gas												
11-IA-1	1/3-4/25	11:25	11:25	-29	-7	65	65	 	-63	10240	10315	6 L	4.16	VL043290.D	X														
Temperature (Fahrenheit) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table> GC/MS Analyst Signature (TO-15) <i>[Signature]</i>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Pressure (Inches of Hg) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop											
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Special Instructions/QC Requirements & Comments :																													
Suspected Contamination: High Medium Low PID Readings: Sampling site (State):																													
Quick Connector required : <i>NO</i>																													
Canisters Shipped by: <i>San</i>				Date/Time: <i>12/2/25</i>				Canisters Received by: <i>[Signature]</i>				Date/Time: <i>12-4-25 1500</i>				B2512003 - 4													
Samples Relinquished by:				Date/Time:				Received by:				Date/Time:																	
Relinquished by:				Date/Time:				Received by:				Date/Time:																	

Client Contact Information						Bottle Order ID : B2512003				Courier :				5 of 6 COCs															
Client ID : RTPE01 Project ID : Air						Sampler Name(s) :				Analysis				Matrix															
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :						Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																			
						Phone Number : 732-968-9600																							
						Fax Number :																							
						Site Details: <i>Chadwick</i>																							
Analysis Turnaround Time						Data Package Type :																							
Standard : 10 business days OR						EDD Type :																							
Rush (Specify): Days																													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambinet Air	Soil Gas												
13-1A-1	12/3-4/25	11:30	11:30	-30	-15	70	70	 	14.1	10626	10320	6 L	4.16	VL043290.D	X	X													
Temperature (Fahrenheit) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				GC/MS Analyst Signature (TO-15) <i>[Signature]</i>							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Pressure (Inches of Hg) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Special Instructions/QC Requirements & Comments :																													
Suspected Contamination: High Medium Low PID Readings:																													
Sampling site (State):																													
Quick Connector required : <i>No</i>																													
Canisters Shipped by: <i>Seun</i>				Date/Time: <i>12/1/15</i>				Canisters Received by: <i>[Signature]</i>				Date/Time: <i>12-4-25 1500</i>				B2512003 - 5													
Samples Relinquished by:				Date/Time:				Received by:				Date/Time:																	
Relinquished by:				Date/Time:				Received by:				Date/Time:																	

Client Contact Information						Bottle Order ID : B2512003				Courier :				6 of 6 COCs																	
Client ID : RTPE01 Project ID : Air						Sampler Name(s) :				Analysis				Matrix																	
Customer Name : RTP Environmental Address : 239 US Highway 22 East City : Green Brook State : NJ Zip Code : 08812 Country :						Project Manager : David gabel				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																					
						Phone Number : 732-968-9600																									
						Fax Number :																									
						Site Details: <i>Chadwick</i>																									
Analysis Turnaround Time						Standard : 10 business days OR				Data Package Type :																					
Rush (Specify): Days						EDD Type :																									
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15			Indoor/Ambinet Air	Soil Gas												
16-IA-1	12/3-4/25	1135	1135	-30	-7	70	70	-30	-5.9	10482	10266	6 L	4.16	VL043259.D	X			X													
Temperature (Fahrenheit) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table> GC/MS Analyst Signature (TO-15) <i>[Signature]</i>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.									
	Ambient	Maximum	Minimum																												
Start																															
Stop																															
Pressure (Inches of Hg) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop													
	Ambient	Maximum	Minimum																												
Start																															
Stop																															
Special Instructions/QC Requirements & Comments : Suspected Contamination: High Medium Low PID Readings: Sampling site (State):																															
Quick Connector required : <i>N</i>																															
Canisters Shipped by: <i>Sam</i>					Date/Time: <i>12/6/25</i>					Canisters Received by: <i>[Signature]</i>					Date/Time: <i>12-4-25 1500</i>																
Samples Relinquished by:					Date/Time:					Received by:					Date/Time:																
Relinquished by:					Date/Time:					Received by:					Date/Time:																

B2512003 - 1

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

Internal Chain of Custody

Instructions: Use 1 form for each 20 samples of aliquot

Laboratory Person Breaking Field Seal on Sample Shuttle & Accepting Responsibility for Sample

Laboratory: Chemtech

Location: 284 Sheffield Street, Mountainside, NJ 7092

QA68:

Title: Sample Custodian

Field Sample Seal No. Q3799

Date Broken 12/4/2025

Military Time Seal Broken: 15:00:00

Case No.: CHADWICK SQUARE

Analytical Parameter/Fraction TO-15

Sample No.	Aliquot/Extract No.	Sample No.	Aliquot/Extract No.
Q3799-01	AA-1		
Q3799-02	9B-IA-1		
Q3799-03	10-IA-1		
Q3799-04	11-IA-1		
Q3799-05	13-IA-1		
Q3799-06	16-IA-1		

Date	Time	Relinquished By	Received By	Purpose of Change of Custody
12/5/25	14:00	Signature	Signature	
		Printed Name <u>Chadwick Square</u>	Printed Name <u>Sample Custodian</u>	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	

Distribution: White - Original (Sent With Report)

Yellow - Contractor Archive

Pink - Sample Custodian - Interim Copy

AIR SAMPLE PRESSURE & DILUTION LOGBOOKAnalyst Signature: Supervisor Signature: 

METHOD: TO-15

Pressure Gauge ID: A255971

Date	Sample Number	Canister #	Initial Pressure psia	Initial Pressure Hg	Final Pressure psia	Final Pressure Hg	Dilution Factor	Comment
11/25/25	Q3696-06	10053	13.6	-2.2				sy
	Q3696-07	10602	13.6	-2.2				
	Q3698-01	10158	12.0	-5.5				
	Q3698-02	10282	12.8	-3.9				
	Q3698-03	10267	12.2	-5.1				
	Q3698-04	10599	12.2	-5.1				
11/26/25	Q3744-01	10311	12.2	-5.1				
	Q3744-02	10285	13.0	-3.5				
	Q3744-03	10604	12.4	-4.7				
	Q3744-04	10601	12.6	-4.3				
	Q3744-05	10443	13.4	-2.7				
12/04/25	Q3797-01	10153	12.6	-4.3				sy
	Q3797-02	10060	12.6	-4.3				
	Q3797-03	10296	12.4	-4.7				
	Q3797-04	10332	12.6	-4.3				
	Q3797-05	10300	13.6	-2.2				
12/07/25	Q3798-01	10492	12.4	-4.7				sy
	Q3798-02	10588	12.0	-5.5				sy
	Q3798-03	10609	13.0	-3.5				sy
12/04/25	Q3799-01	10280	12.8	-3.9				
	Q3799-02	10608	12.0	-5.5				

Analyst Signature:

Supervisor Signature: _____

METHOD: TO-15

Pressure Gauge ID:Document Control # A3041228

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

Client Sample ID #: 9B-IA-1
Client Name: RTP
Project Name: Chadwick
Date: 12/3/25 Time: 11:15
Analysis: TO-15
Comment: _____

CHEMTECH
Date Received

Storage Location: Air Lab
Sample: Q3799-02
Cust #: 9B-IA-1

of Disposal: _____

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

Client Sample ID #: AA-1
Client Name: RTP
Project Name: Chadwick
Date: 12/3/25 - 12/4/25 Time: 11:40
Analysis: TO-15
Comment: _____

CHEMTECH
Date Received

Storage Location: Air Lab
Sample: Q3799-01
Cust #: AA-1

of Disposal: _____

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

Client Sample ID #: 10-IA-1
Client Name: RTP
Project Name: Chadwick
Date: 12/3/25 - 12/4/25 Time: 11:20
Analysis: TO-15
Comment: _____

Storage Location: Air Lab
Sample: Q3799-03
Cust #: 10-IA-1

CHEMTECH
Date Received

of Disposal: _____

10329-114

CHEMTECH
284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

Client Sample ID #: 13-IA-1
Client Name: RTP
Project Name: Chadwick
Date: 12/3-4/25 Time: 1135
Analysis: TO-15
Comments: _____

CHEMTECH'S
Date Received: _____

Storage Location: Air Lab
Sample: Q3799-05
Cust #: 13-IA-1

Proposal: _____

82512003

10315-63

CHEMTECH
284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

Client Sample ID #: 11-IA-1
Client Name: RTP
Project Name: Chadwick
Date: 12/3/25 Time: 1128
Analysis: TO-15
Comments: _____

CHEMTECH'S
Date Received: _____

Storage Location: Air Lab
Sample: Q3799-04
Cust #: 11-IA-1

Proposal: _____

82512003

15-1
Q260

CHEMTECH
284 Sheffield Street, Mountainside, NJ 07092 P: (908) 789-8900 F: (908) 789-8922

Client Sample ID #: 16-IA-1
Client Name: RTP
Project Name: Chedwick
Date: 12/3-4/25 Time: 1135
Analysis: TD-15
Comments: _____
CHEMTECH: _____
Date Received: _____

Storage Location: Air Lab
Sample: Q3799-06
Cust #: 16-IA-1

Disposal: _____

12003

CASE NARRATIVE

RTP Environmental

Project Name: CHADWICK SQUARE

Project # N/A

Order ID # Q3799

Test Name: TO-15

A. Number of Samples and Date of Receipt:

6 Air samples were received on 12/04/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: TO-15. This data package contains results for TO-15.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of TO-15 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD for {Q3799-01DUP} with File ID: VL043311.D met criteria except for Methylene Chloride [31.6%] due to difference in results of original and DUP.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 30% in the Initial Calibration method (VL111825AIR.M) for Methylene chloride passing on Linear regression.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Samples 10-IA-1, 10-IA-1DL and 16-IA-1 were diluted due to high concentrations.

E. Additional Comments:

The Sample #AA-1, 9B-IA-1, 10-IA-1, 11-IA-1 and 16-IA-1 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:37 pm, Dec 15, 2025

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3799

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 12/12/2025

LAB CHRONICLE

OrderID:	Q3799	OrderDate:	12/5/2025 1:04:00 PM
Client:	RTP Environmental	Project:	CHADWICK SQUARE
Contact:	David Gabel	Location:	Air Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3799-01	AA-1	Air	TO-15	TO-15	12/03/25		12/08/25	12/04/25
Q3799-02	9B-IA-1	Air	TO-15	TO-15	12/03/25		12/08/25	12/04/25
Q3799-03	10-IA-1	Air	TO-15	TO-15	12/03/25		12/08/25	12/04/25
Q3799-03DL	10-IA-1DL	Air	TO-15	TO-15	12/03/25		12/08/25	12/04/25
Q3799-03DL 2	10-IA-1DL2	Air	TO-15	TO-15	12/03/25		12/09/25	12/04/25
Q3799-04	11-IA-1	Air	TO-15	TO-15	12/03/25		12/08/25	12/04/25
Q3799-05	13-IA-1	Air	TO-15	TO-15	12/03/25		12/08/25	12/04/25
Q3799-06	16-IA-1	Air	TO-15	TO-15	12/03/25		12/08/25	12/04/25
Q3799-06DL	16-IA-1DL	Air	TO-15	TO-15	12/03/25		12/08/25	12/04/25

Hit Summary Sheet SW-846

SDG No.: Q3799
Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	AA-1							
Q3799-01	AA-1	Air	Dichlorodifluoromethane	1.93	J	0.54	2.47	ug/m3
Q3799-01	AA-1	Air	Chloromethane	4.54		0.21	1.03	ug/m3
Q3799-01	AA-1	Air	Trichlorofluoromethane	1.18	J	0.96	2.81	ug/m3
Q3799-01	AA-1	Air	Acetone	11.4		0.43	1.19	ug/m3
Q3799-01	AA-1	Air	Methylene Chloride	8.34		0.80	1.74	ug/m3
Q3799-01	AA-1	Air	2-Butanone	1.47		0.18	1.47	ug/m3
Q3799-01	AA-1	Air	Carbon Tetrachloride	0.57		0.19	0.19	ug/m3
Q3799-01	AA-1	Air	Chloroform	0.73	J	0.49	2.44	ug/m3
Q3799-01	AA-1	Air	Benzene	0.58	J	0.26	1.60	ug/m3
Q3799-01	AA-1	Air	Hexane	4.23		0.56	1.76	ug/m3
Total Voc :				35.0				
Total Concentration:				35.0				
Client ID:	9B-IA-1							
Q3799-02	9B-IA-1	Air	Dichlorodifluoromethane	1.98	J	0.54	2.47	ug/m3
Q3799-02	9B-IA-1	Air	Chloromethane	1.14		0.21	1.03	ug/m3
Q3799-02	9B-IA-1	Air	Trichlorofluoromethane	1.12	J	0.96	2.81	ug/m3
Q3799-02	9B-IA-1	Air	Acetone	20.9		0.43	1.19	ug/m3
Q3799-02	9B-IA-1	Air	Methylene Chloride	1.67	J	0.80	1.74	ug/m3
Q3799-02	9B-IA-1	Air	2-Butanone	0.88	J	0.18	1.47	ug/m3
Q3799-02	9B-IA-1	Air	Carbon Tetrachloride	0.63		0.19	0.19	ug/m3
Q3799-02	9B-IA-1	Air	Chloroform	0.68	J	0.49	2.44	ug/m3
Q3799-02	9B-IA-1	Air	Benzene	0.48	J	0.26	1.60	ug/m3
Q3799-02	9B-IA-1	Air	Hexane	3.10		0.56	1.76	ug/m3
Total Voc :				32.6				
Total Concentration:				32.6				
Client ID:	10-IA-1							
Q3799-03	10-IA-1	Air	Dichlorodifluoromethane	2.03	J	0.54	2.47	ug/m3
Q3799-03	10-IA-1	Air	Chloromethane	0.95	J	0.21	1.03	ug/m3
Q3799-03	10-IA-1	Air	Trichlorofluoromethane	1.29	J	0.96	2.81	ug/m3
Q3799-03	10-IA-1	Air	Heptane	0.70	J	0.70	2.05	ug/m3
Q3799-03	10-IA-1	Air	Acetone	689	E	0.43	1.19	ug/m3
Q3799-03	10-IA-1	Air	Methylene Chloride	55.2	E	0.80	1.74	ug/m3
Q3799-03	10-IA-1	Air	2-Butanone	1.62		0.18	1.47	ug/m3
Q3799-03	10-IA-1	Air	Carbon Tetrachloride	0.57		0.19	0.19	ug/m3
Q3799-03	10-IA-1	Air	Benzene	0.51	J	0.26	1.60	ug/m3
Q3799-03	10-IA-1	Air	Toluene	1.70	J	0.60	1.88	ug/m3

Hit Summary Sheet
SW-846

SDG No.: Q3799

Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3799-03	10-IA-1	Air	Tetrachloroethene	0.27		0.14	0.20	ug/m3
Q3799-03	10-IA-1	Air	Methyl Methacrylate	267	E	0.57	2.05	ug/m3
Total Voc :				1020				
Total Concentration:				1020				
Client ID:	10-IA-1DL							
Q3799-03DL	10-IA-1DL	Air	Acetone	3560	ED	4.28	11.9	ug/m3
Q3799-03DL	10-IA-1DL	Air	Methylene Chloride	50.4	D	7.99	17.4	ug/m3
Q3799-03DL	10-IA-1DL	Air	Methyl Methacrylate	235	D	5.73	20.5	ug/m3
Total Voc :				3850				
Total Concentration:				3850				
Client ID:	10-IA-1DL2							
Q3799-03DL2	10-IA-1DL2	Air	Acetone	7130	D	171	475	ug/m3
Total Voc :				7130				
Total Concentration:				7130				
Client ID:	11-IA-1							
Q3799-04	11-IA-1	Air	Dichlorodifluoromethane	1.98	J	0.54	2.47	ug/m3
Q3799-04	11-IA-1	Air	Chloromethane	1.03		0.21	1.03	ug/m3
Q3799-04	11-IA-1	Air	Trichlorofluoromethane	1.29	J	0.96	2.81	ug/m3
Q3799-04	11-IA-1	Air	Acetone	19.0		0.43	1.19	ug/m3
Q3799-04	11-IA-1	Air	Methylene Chloride	1.39	J	0.80	1.74	ug/m3
Q3799-04	11-IA-1	Air	2-Butanone	1.03	J	0.18	1.47	ug/m3
Q3799-04	11-IA-1	Air	Carbon Tetrachloride	0.57		0.19	0.19	ug/m3
Q3799-04	11-IA-1	Air	Benzene	0.48	J	0.26	1.60	ug/m3
Q3799-04	11-IA-1	Air	Toluene	0.98	J	0.60	1.88	ug/m3
Q3799-04	11-IA-1	Air	Tetrachloroethene	0.61		0.14	0.20	ug/m3
Q3799-04	11-IA-1	Air	Hexane	4.23		0.56	1.76	ug/m3
Total Voc :				32.6				
Total Concentration:				32.6				
Client ID:	13-IA-1							
Q3799-05	13-IA-1	Air	Dichlorodifluoromethane	1.98	J	0.54	2.47	ug/m3
Q3799-05	13-IA-1	Air	Chloromethane	0.95	J	0.21	1.03	ug/m3
Q3799-05	13-IA-1	Air	Trichlorofluoromethane	1.40	J	0.96	2.81	ug/m3
Q3799-05	13-IA-1	Air	Acetone	8.31		0.43	1.19	ug/m3
Q3799-05	13-IA-1	Air	Methylene Chloride	2.64		0.80	1.74	ug/m3
Q3799-05	13-IA-1	Air	2-Butanone	0.77	J	0.18	1.47	ug/m3
Q3799-05	13-IA-1	Air	Carbon Tetrachloride	0.44		0.19	0.19	ug/m3
Q3799-05	13-IA-1	Air	Benzene	0.67	J	0.26	1.60	ug/m3
Q3799-05	13-IA-1	Air	Toluene	0.72	J	0.60	1.88	ug/m3

Hit Summary Sheet
SW-846

SDG No.: Q3799

Client: RTP Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3799-05	13-IA-1	Air	Tetrachloroethene	0.34		0.14	0.20	ug/m3
Q3799-05	13-IA-1	Air	Hexane	8.11		0.56	1.76	ug/m3
Total Voc :				26.3				
Total Concentration:				26.3				
Client ID:	16-IA-1							
Q3799-06	16-IA-1	Air	Dichlorodifluoromethane	2.08	J	0.54	2.47	ug/m3
Q3799-06	16-IA-1	Air	Chloromethane	1.53		0.21	1.03	ug/m3
Q3799-06	16-IA-1	Air	Trichlorofluoromethane	1.18	J	0.96	2.81	ug/m3
Q3799-06	16-IA-1	Air	Acetone	70.8	E	0.43	1.19	ug/m3
Q3799-06	16-IA-1	Air	Methylene Chloride	25.4		0.80	1.74	ug/m3
Q3799-06	16-IA-1	Air	2-Butanone	1.30	J	0.18	1.47	ug/m3
Q3799-06	16-IA-1	Air	Carbon Tetrachloride	0.50		0.19	0.19	ug/m3
Q3799-06	16-IA-1	Air	Benzene	0.64	J	0.26	1.60	ug/m3
Q3799-06	16-IA-1	Air	4-Methyl-2-Pentanone	0.45	J	0.41	2.05	ug/m3
Q3799-06	16-IA-1	Air	Toluene	0.98	J	0.60	1.88	ug/m3
Q3799-06	16-IA-1	Air	Ethyl Benzene	0.96	J	0.83	2.17	ug/m3
Q3799-06	16-IA-1	Air	m/p-Xylene	3.91	J	1.78	4.34	ug/m3
Q3799-06	16-IA-1	Air	o-Xylene	1.04	J	0.91	2.17	ug/m3
Q3799-06	16-IA-1	Air	Hexane	9.52		0.56	1.76	ug/m3
Total Voc :				120				
Total Concentration:				120				
Client ID:	16-IA-1DL							
Q3799-06DL	16-IA-1DL	Air	Acetone	83.1	D	4.28	11.9	ug/m3
Q3799-06DL	16-IA-1DL	Air	Methylene Chloride	31.6	D	7.99	17.4	ug/m3
Q3799-06DL	16-IA-1DL	Air	Hexane	8.46	JD	5.64	17.6	ug/m3
Total Voc :				123				
Total Concentration:				123				

Project : CHADWICK SQUARE

Field Id Number : AA-1

Laboratory Id Number : Q3799-01

Sampling Date : 12/03/25

Analysis Date : 12/08/25

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.39	J	1.93		
Chloromethane	74-87-3	50.49	2.2		4.54		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.08	U	0.31		
Chloroethane	75-00-3	64.52	0.15	U	0.4		
Tetrahydrofuran	109-99-9	72.11	0.08	U	0.24		
Trichlorofluoromethane	75-69-4	137.4	0.21	J	1.18		
Dichlorotetrafluoroethane	76-14-2	170.9	0.15	U	1.05		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.14	U	1.07		
tert-Butyl alcohol	75-65-0	74.12	0.16	U	0.49		
Heptane	142-82-5	100.2	0.17	U	0.7		
1,1-Dichloroethene	75-35-4	96.94	0.15	U	0.59		
Acetone	67-64-1	58.08	4.8		11.4		
Carbon Disulfide	75-15-0	76.14	0.08	U	0.25		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.23	U	0.83		
Methylene Chloride	75-09-2	84.94	2.4		8.34		
trans-1,2-Dichloroethene	156-60-5	96.94	0.12	U	0.48		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.5		1.47		
Carbon Tetrachloride	56-23-5	153.8	0.09		0.57		
cis-1,2-Dichloroethene	156-59-2	96.94	0.1	U	0.4		
Chloroform	67-66-3	119.4	0.15	J	0.73		
1,1,1-Trichloroethane	71-55-6	133.4	0.02	U	0.11		
2,2,4-Trimethylpentane	540-84-1	114.2	0.14	U	0.65		
Benzene	71-43-2	78.11	0.18	J	0.58		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.13	U	0.6		
Bromodichloromethane	75-27-4	163.8	0.06	U	0.4		
4-Methyl-2-Pentanone	108-10-1	100.2	0.1	U	0.41		
Toluene	108-88-3	92.14	0.16	U	0.6		
t-1,3-Dichloropropene	10061-02-6	111	0.15	U	0.68		
cis-1,3-Dichloropropene	10061-01-5	111	0.11	U	0.5		
1,1,2-Trichloroethane	79-00-5	133.4	0.08	U	0.44		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : AA-1

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-01

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.09	U	0.77		
1,2-Dibromoethane	106-93-4	187.9	0.09	U	0.69		
Tetrachloroethene	127-18-4	165.8	0.02	U	0.14		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.19	U	0.83		
m/p-Xylene	179601-23-1	106.2	0.41	U	1.78		
o-Xylene	95-47-6	106.2	0.21	U	0.91		
Styrene	100-42-5	104.1	0.16	U	0.68		
Bromoform	75-25-2	252.8	0.05	U	0.52		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.02	U	0.14		
2-Chlorotoluene	95-49-8	126.6	0.17	U	0.88		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.18	U	0.88		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.18	U	0.88		
1,3-Dichlorobenzene	541-73-1	147	0.05	U	0.3		
1,4-Dichlorobenzene	106-46-7	147	0.13	U	0.78		
1,2-Dichlorobenzene	95-50-1	147	0.13	U	0.78		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.21	U	1.56		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.13	U	1.39		
Naphthalene	91-20-3	128.17	0.01	U	0.05		
1,3-Butadiene	106-99-0	54.09	0.05	U	0.11		
4-Ethyltoluene	622-96-8	120.2	0.21	U	1.03		
Hexane	110-54-3	86.17	1.2		4.23		
Allyl Chloride	107-05-1	76.53	0.11	U	0.34		
1,4-Dioxane	123-91-1	88.12	0.22	U	0.79		
Methyl Methacrylate	80-62-6	100.117	0.14	U	0.57		

Project : CHADWICK SQUARE

Field Id Number : 9B-IA-1

Laboratory Id Number : Q3799-02

Sampling Date : 12/03/25

Analysis Date : 12/08/25

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.4	J	1.98		
Chloromethane	74-87-3	50.49	0.55		1.14		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.08	U	0.31		
Chloroethane	75-00-3	64.52	0.15	U	0.4		
Tetrahydrofuran	109-99-9	72.11	0.08	U	0.24		
Trichlorofluoromethane	75-69-4	137.4	0.2	J	1.12		
Dichlorotetrafluoroethane	76-14-2	170.9	0.15	U	1.05		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.14	U	1.07		
tert-Butyl alcohol	75-65-0	74.12	0.16	U	0.49		
Heptane	142-82-5	100.2	0.17	U	0.7		
1,1-Dichloroethene	75-35-4	96.94	0.15	U	0.59		
Acetone	67-64-1	58.08	8.8		20.9		
Carbon Disulfide	75-15-0	76.14	0.08	U	0.25		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.23	U	0.83		
Methylene Chloride	75-09-2	84.94	0.48	J	1.67		
trans-1,2-Dichloroethene	156-60-5	96.94	0.12	U	0.48		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.3	J	0.88		
Carbon Tetrachloride	56-23-5	153.8	0.1		0.63		
cis-1,2-Dichloroethene	156-59-2	96.94	0.1	U	0.4		
Chloroform	67-66-3	119.4	0.14	J	0.68		
1,1,1-Trichloroethane	71-55-6	133.4	0.02	U	0.11		
2,2,4-Trimethylpentane	540-84-1	114.2	0.14	U	0.65		
Benzene	71-43-2	78.11	0.15	J	0.48		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.13	U	0.6		
Bromodichloromethane	75-27-4	163.8	0.06	U	0.4		
4-Methyl-2-Pentanone	108-10-1	100.2	0.1	U	0.41		
Toluene	108-88-3	92.14	0.16	U	0.6		
t-1,3-Dichloropropene	10061-02-6	111	0.15	U	0.68		
cis-1,3-Dichloropropene	10061-01-5	111	0.11	U	0.5		
1,1,2-Trichloroethane	79-00-5	133.4	0.08	U	0.44		

Project : CHADWICK SQUARE

Field Id Number : 9B-IA-1

Laboratory Id Number : Q3799-02

Sampling Date : 12/03/25

Analysis Date : 12/08/25

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.09	U	0.77		
1,2-Dibromoethane	106-93-4	187.9	0.09	U	0.69		
Tetrachloroethene	127-18-4	165.8	0.02	U	0.14		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.19	U	0.83		
m/p-Xylene	179601-23-1	106.2	0.41	U	1.78		
o-Xylene	95-47-6	106.2	0.21	U	0.91		
Styrene	100-42-5	104.1	0.16	U	0.68		
Bromoform	75-25-2	252.8	0.05	U	0.52		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.02	U	0.14		
2-Chlorotoluene	95-49-8	126.6	0.17	U	0.88		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.18	U	0.88		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.18	U	0.88		
1,3-Dichlorobenzene	541-73-1	147	0.05	U	0.3		
1,4-Dichlorobenzene	106-46-7	147	0.13	U	0.78		
1,2-Dichlorobenzene	95-50-1	147	0.13	U	0.78		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.21	U	1.56		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.13	U	1.39		
Naphthalene	91-20-3	128.17	0.01	U	0.05		
1,3-Butadiene	106-99-0	54.09	0.05	U	0.11		
4-Ethyltoluene	622-96-8	120.2	0.21	U	1.03		
Hexane	110-54-3	86.17	0.88		3.1		
Allyl Chloride	107-05-1	76.53	0.11	U	0.34		
1,4-Dioxane	123-91-1	88.12	0.22	U	0.79		
Methyl Methacrylate	80-62-6	100.117	0.14	U	0.57		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 10-IA-1

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-03

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.41	J	2.03		
Chloromethane	74-87-3	50.49	0.46	J	0.95		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.08	U	0.31		
Chloroethane	75-00-3	64.52	0.15	U	0.4		
Tetrahydrofuran	109-99-9	72.11	0.08	U	0.24		
Trichlorofluoromethane	75-69-4	137.4	0.23	J	1.29		
Dichlorotetrafluoroethane	76-14-2	170.9	0.15	U	1.05		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.14	U	1.07		
tert-Butyl alcohol	75-65-0	74.12	0.16	U	0.49		
Heptane	142-82-5	100.2	0.17	J	0.7		
1,1-Dichloroethene	75-35-4	96.94	0.15	U	0.59		
Acetone	67-64-1	58.08	290	E	688		
Carbon Disulfide	75-15-0	76.14	0.08	U	0.25		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.23	U	0.83		
Methylene Chloride	75-09-2	84.94	15.9	E	55.2		
trans-1,2-Dichloroethene	156-60-5	96.94	0.12	U	0.48		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.55		1.62		
Carbon Tetrachloride	56-23-5	153.8	0.09		0.57		
cis-1,2-Dichloroethene	156-59-2	96.94	0.1	U	0.4		
Chloroform	67-66-3	119.4	0.1	U	0.49		
1,1,1-Trichloroethane	71-55-6	133.4	0.02	U	0.11		
2,2,4-Trimethylpentane	540-84-1	114.2	0.14	U	0.65		
Benzene	71-43-2	78.11	0.16	J	0.51		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.13	U	0.6		
Bromodichloromethane	75-27-4	163.8	0.06	U	0.4		
4-Methyl-2-Pentanone	108-10-1	100.2	0.1	U	0.41		
Toluene	108-88-3	92.14	0.45	J	1.7		
t-1,3-Dichloropropene	10061-02-6	111	0.15	U	0.68		
cis-1,3-Dichloropropene	10061-01-5	111	0.11	U	0.5		
1,1,2-Trichloroethane	79-00-5	133.4	0.08	U	0.44		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 10-IA-1

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-03

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.09	U	0.77		
1,2-Dibromoethane	106-93-4	187.9	0.09	U	0.69		
Tetrachloroethene	127-18-4	165.8	0.04		0.27		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.19	U	0.83		
m/p-Xylene	179601-23-1	106.2	0.41	U	1.78		
o-Xylene	95-47-6	106.2	0.21	U	0.91		
Styrene	100-42-5	104.1	0.16	U	0.68		
Bromoform	75-25-2	252.8	0.05	U	0.52		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.02	U	0.14		
2-Chlorotoluene	95-49-8	126.6	0.17	U	0.88		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.18	U	0.88		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.18	U	0.88		
1,3-Dichlorobenzene	541-73-1	147	0.05	U	0.3		
1,4-Dichlorobenzene	106-46-7	147	0.13	U	0.78		
1,2-Dichlorobenzene	95-50-1	147	0.13	U	0.78		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.21	U	1.56		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.13	U	1.39		
Naphthalene	91-20-3	128.17	0.01	U	0.05		
1,3-Butadiene	106-99-0	54.09	0.05	U	0.11		
4-Ethyltoluene	622-96-8	120.2	0.21	U	1.03		
Hexane	110-54-3	86.17	0.16	U	0.56		
Allyl Chloride	107-05-1	76.53	0.11	U	0.34		
1,4-Dioxane	123-91-1	88.12	0.22	U	0.79		
Methyl Methacrylate	80-62-6	100.117	65.2	E	266		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 10-IA-1DL

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-03DL

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	1.1	UD	5.44		
Chloromethane	74-87-3	50.49	0.98	UD	2.02		
Vinyl Chloride	75-01-4	62.5	0.25	UD	0.64		
Bromomethane	74-83-9	94.94	0.77	UD	2.99		
Chloroethane	75-00-3	64.52	1.5	UD	3.96		
Tetrahydrofuran	109-99-9	72.11	0.8	UD	2.36		
Trichlorofluoromethane	75-69-4	137.4	1.7	UD	9.55		
Dichlorotetrafluoroethane	76-14-2	170.9	1.5	UD	10.5		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	1.4	UD	10.7		
tert-Butyl alcohol	75-65-0	74.12	1.6	UD	4.85		
Heptane	142-82-5	100.2	1.7	UD	6.97		
1,1-Dichloroethene	75-35-4	96.94	1.5	UD	5.95		
Acetone	67-64-1	58.08	1500	ED	3563		
Carbon Disulfide	75-15-0	76.14	0.8	UD	2.49		
Methyl tert-Butyl Ether	1634-04-4	88.15	2.3	UD	8.29		
Methylene Chloride	75-09-2	84.94	14.5	D	50.4		
trans-1,2-Dichloroethene	156-60-5	96.94	1.2	UD	4.76		
1,1-Dichloroethane	75-34-3	98.96	1.3	UD	5.26		
Cyclohexane	110-82-7	84.16	2.2	UD	7.57		
2-Butanone	78-93-3	72.11	0.56	UD	1.65		
Carbon Tetrachloride	56-23-5	153.8	0.25	UD	1.57		
cis-1,2-Dichloroethene	156-59-2	96.94	0.99	UD	3.93		
Chloroform	67-66-3	119.4	0.96	UD	4.69		
1,1,1-Trichloroethane	71-55-6	133.4	0.16	UD	0.87		
2,2,4-Trimethylpentane	540-84-1	114.2	1.4	UD	6.54		
Benzene	71-43-2	78.11	0.79	UD	2.52		
1,2-Dichloroethane	107-06-2	98.96	0.93	UD	3.76		
Trichloroethene	79-01-6	131.4	0.24	UD	1.29		
1,2-Dichloropropane	78-87-5	113	1.3	UD	6.01		
Bromodichloromethane	75-27-4	163.8	0.63	UD	4.22		
4-Methyl-2-Pentanone	108-10-1	100.2	0.97	UD	3.98		
Toluene	108-88-3	92.14	1.6	UD	6.03		
t-1,3-Dichloropropene	10061-02-6	111	1.5	UD	6.81		
cis-1,3-Dichloropropene	10061-01-5	111	1.1	UD	4.99		
1,1,2-Trichloroethane	79-00-5	133.4	0.81	UD	4.42		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 10-IA-1DL

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-03DL

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.85	UD	7.24		
1,2-Dibromoethane	106-93-4	187.9	0.85	UD	6.53		
Tetrachloroethene	127-18-4	165.8	0.15	UD	1.02		
Chlorobenzene	108-90-7	112.6	0.77	UD	3.55		
Ethyl Benzene	100-41-4	106.2	1.9	UD	8.25		
m/p-Xylene	179601-23-1	106.2	4.1	UD	17.8		
o-Xylene	95-47-6	106.2	2.1	UD	9.12		
Styrene	100-42-5	104.1	1.6	UD	6.81		
Bromoform	75-25-2	252.8	0.48	UD	4.96		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.18	UD	1.24		
2-Chlorotoluene	95-49-8	126.6	1.7	UD	8.8		
1,3,5-Trimethylbenzene	108-67-8	120.2	1.8	UD	8.85		
1,2,4-Trimethylbenzene	95-63-6	120.2	1.8	UD	8.85		
1,3-Dichlorobenzene	541-73-1	147	0.54	UD	3.25		
1,4-Dichlorobenzene	106-46-7	147	1.3	UD	7.82		
1,2-Dichlorobenzene	95-50-1	147	1.3	UD	7.82		
1,2,4-Trichlorobenzene	120-82-1	181.5	2.1	UD	15.6		
Hexachloro-1,3-Butadiene	87-68-3	260.8	1.3	UD	13.9		
Naphthalene	91-20-3	128.17	0.13	UD	0.68		
1,3-Butadiene	106-99-0	54.09	0.51	UD	1.13		
4-Ethyltoluene	622-96-8	120.2	2.1	UD	10.3		
Hexane	110-54-3	86.17	1.6	UD	5.64		
Allyl Chloride	107-05-1	76.53	1.1	UD	3.44		
1,4-Dioxane	123-91-1	88.12	2.2	UD	7.93		
Methyl Methacrylate	80-62-6	100.117	57.4	D	235		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 10-IA-1DL2

Analysis Date : 12/09/25

Laboratory Id Number : Q3799-03DL2

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	44	UD	217		
Chloromethane	74-87-3	50.49	39.2	UD	81.0		
Vinyl Chloride	75-01-4	62.5	10	UD	25.6		
Bromomethane	74-83-9	94.94	30.8	UD	119		
Chloroethane	75-00-3	64.52	60	UD	158		
Tetrahydrofuran	109-99-9	72.11	32	UD	94.4		
Trichlorofluoromethane	75-69-4	137.4	68	UD	382		
Dichlorotetrafluoroethane	76-14-2	170.9	60	UD	419		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	56	UD	429		
tert-Butyl alcohol	75-65-0	74.12	64	UD	194		
Heptane	142-82-5	100.2	68	UD	278		
1,1-Dichloroethene	75-35-4	96.94	60	UD	237		
Acetone	67-64-1	58.08	3000	D	7126		
Carbon Disulfide	75-15-0	76.14	32	UD	99.6		
Methyl tert-Butyl Ether	1634-04-4	88.15	92	UD	331		
Methylene Chloride	75-09-2	84.94	92	UD	319		
trans-1,2-Dichloroethene	156-60-5	96.94	48	UD	190		
1,1-Dichloroethane	75-34-3	98.96	52	UD	210		
Cyclohexane	110-82-7	84.16	88	UD	302		
2-Butanone	78-93-3	72.11	22.4	UD	66.1		
Carbon Tetrachloride	56-23-5	153.8	10	UD	62.9		
cis-1,2-Dichloroethene	156-59-2	96.94	39.6	UD	157		
Chloroform	67-66-3	119.4	38.4	UD	187		
1,1,1-Trichloroethane	71-55-6	133.4	6.4	UD	34.9		
2,2,4-Trimethylpentane	540-84-1	114.2	56	UD	261		
Benzene	71-43-2	78.11	31.6	UD	100		
1,2-Dichloroethane	107-06-2	98.96	37.2	UD	150		
Trichloroethene	79-01-6	131.4	9.6	UD	51.6		
1,2-Dichloropropane	78-87-5	113	52	UD	240		
Bromodichloromethane	75-27-4	163.8	25.2	UD	168		
4-Methyl-2-Pentanone	108-10-1	100.2	38.8	UD	159		
Toluene	108-88-3	92.14	64	UD	241		
t-1,3-Dichloropropene	10061-02-6	111	60	UD	272		
cis-1,3-Dichloropropene	10061-01-5	111	44	UD	199		
1,1,2-Trichloroethane	79-00-5	133.4	32.4	UD	176		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 10-IA-1DL2

Analysis Date : 12/09/25

Laboratory Id Number : Q3799-03DL2

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	34	UD	289		
1,2-Dibromoethane	106-93-4	187.9	34	UD	261		
Tetrachloroethene	127-18-4	165.8	6	UD	40.7		
Chlorobenzene	108-90-7	112.6	30.8	UD	141		
Ethyl Benzene	100-41-4	106.2	76	UD	330		
m/p-Xylene	179601-23-1	106.2	160	UD	694		
o-Xylene	95-47-6	106.2	84	UD	364		
Styrene	100-42-5	104.1	64	UD	272		
Bromoform	75-25-2	252.8	19.2	UD	198		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	7.2	UD	49.4		
2-Chlorotoluene	95-49-8	126.6	68	UD	352		
1,3,5-Trimethylbenzene	108-67-8	120.2	72	UD	353		
1,2,4-Trimethylbenzene	95-63-6	120.2	72	UD	353		
1,3-Dichlorobenzene	541-73-1	147	21.6	UD	129		
1,4-Dichlorobenzene	106-46-7	147	52	UD	312		
1,2-Dichlorobenzene	95-50-1	147	52	UD	312		
1,2,4-Trichlorobenzene	120-82-1	181.5	84	UD	623		
Hexachloro-1,3-Butadiene	87-68-3	260.8	52	UD	554		
Naphthalene	91-20-3	128.17	5.2	UD	27.3		
1,3-Butadiene	106-99-0	54.09	20.4	UD	45.1		
4-Ethyltoluene	622-96-8	120.2	84	UD	412		
Hexane	110-54-3	86.17	64	UD	225		
Allyl Chloride	107-05-1	76.53	44	UD	137		
1,4-Dioxane	123-91-1	88.12	88	UD	317		
Methyl Methacrylate	80-62-6	100.117	56	UD	229		

Project : CHADWICK SQUARE

Field Id Number : 11-IA-1

Laboratory Id Number : Q3799-04

Sampling Date : 12/03/25

Analysis Date : 12/08/25

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.4	J	1.98		
Chloromethane	74-87-3	50.49	0.5		1.03		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.08	U	0.31		
Chloroethane	75-00-3	64.52	0.15	U	0.4		
Tetrahydrofuran	109-99-9	72.11	0.08	U	0.24		
Trichlorofluoromethane	75-69-4	137.4	0.23	J	1.29		
Dichlorotetrafluoroethane	76-14-2	170.9	0.15	U	1.05		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.14	U	1.07		
tert-Butyl alcohol	75-65-0	74.12	0.16	U	0.49		
Heptane	142-82-5	100.2	0.17	U	0.7		
1,1-Dichloroethene	75-35-4	96.94	0.15	U	0.59		
Acetone	67-64-1	58.08	8		19		
Carbon Disulfide	75-15-0	76.14	0.08	U	0.25		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.23	U	0.83		
Methylene Chloride	75-09-2	84.94	0.4	J	1.39		
trans-1,2-Dichloroethene	156-60-5	96.94	0.12	U	0.48		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.35	J	1.03		
Carbon Tetrachloride	56-23-5	153.8	0.09		0.57		
cis-1,2-Dichloroethene	156-59-2	96.94	0.1	U	0.4		
Chloroform	67-66-3	119.4	0.1	U	0.49		
1,1,1-Trichloroethane	71-55-6	133.4	0.02	U	0.11		
2,2,4-Trimethylpentane	540-84-1	114.2	0.14	U	0.65		
Benzene	71-43-2	78.11	0.15	J	0.48		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.13	U	0.6		
Bromodichloromethane	75-27-4	163.8	0.06	U	0.4		
4-Methyl-2-Pentanone	108-10-1	100.2	0.1	U	0.41		
Toluene	108-88-3	92.14	0.26	J	0.98		
t-1,3-Dichloropropene	10061-02-6	111	0.15	U	0.68		
cis-1,3-Dichloropropene	10061-01-5	111	0.11	U	0.5		
1,1,2-Trichloroethane	79-00-5	133.4	0.08	U	0.44		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 11-IA-1

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-04

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.09	U	0.77		
1,2-Dibromoethane	106-93-4	187.9	0.09	U	0.69		
Tetrachloroethene	127-18-4	165.8	0.09		0.61		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.19	U	0.83		
m/p-Xylene	179601-23-1	106.2	0.41	U	1.78		
o-Xylene	95-47-6	106.2	0.21	U	0.91		
Styrene	100-42-5	104.1	0.16	U	0.68		
Bromoform	75-25-2	252.8	0.05	U	0.52		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.02	U	0.14		
2-Chlorotoluene	95-49-8	126.6	0.17	U	0.88		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.18	U	0.88		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.18	U	0.88		
1,3-Dichlorobenzene	541-73-1	147	0.05	U	0.3		
1,4-Dichlorobenzene	106-46-7	147	0.13	U	0.78		
1,2-Dichlorobenzene	95-50-1	147	0.13	U	0.78		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.21	U	1.56		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.13	U	1.39		
Naphthalene	91-20-3	128.17	0.01	U	0.05		
1,3-Butadiene	106-99-0	54.09	0.05	U	0.11		
4-Ethyltoluene	622-96-8	120.2	0.21	U	1.03		
Hexane	110-54-3	86.17	1.2		4.23		
Allyl Chloride	107-05-1	76.53	0.11	U	0.34		
1,4-Dioxane	123-91-1	88.12	0.22	U	0.79		
Methyl Methacrylate	80-62-6	100.117	0.14	U	0.57		

Project : CHADWICK SQUARE

Field Id Number : 13-IA-1

Laboratory Id Number : Q3799-05

Sampling Date : 12/03/25

Analysis Date : 12/08/25

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.4	J	1.98		
Chloromethane	74-87-3	50.49	0.46	J	0.95		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.08	U	0.31		
Chloroethane	75-00-3	64.52	0.15	U	0.4		
Tetrahydrofuran	109-99-9	72.11	0.08	U	0.24		
Trichlorofluoromethane	75-69-4	137.4	0.25	J	1.4		
Dichlorotetrafluoroethane	76-14-2	170.9	0.15	U	1.05		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.14	U	1.07		
tert-Butyl alcohol	75-65-0	74.12	0.16	U	0.49		
Heptane	142-82-5	100.2	0.17	U	0.7		
1,1-Dichloroethene	75-35-4	96.94	0.15	U	0.59		
Acetone	67-64-1	58.08	3.5		8.31		
Carbon Disulfide	75-15-0	76.14	0.08	U	0.25		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.23	U	0.83		
Methylene Chloride	75-09-2	84.94	0.76		2.64		
trans-1,2-Dichloroethene	156-60-5	96.94	0.12	U	0.48		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.26	J	0.77		
Carbon Tetrachloride	56-23-5	153.8	0.07		0.44		
cis-1,2-Dichloroethene	156-59-2	96.94	0.1	U	0.4		
Chloroform	67-66-3	119.4	0.1	U	0.49		
1,1,1-Trichloroethane	71-55-6	133.4	0.02	U	0.11		
2,2,4-Trimethylpentane	540-84-1	114.2	0.14	U	0.65		
Benzene	71-43-2	78.11	0.21	J	0.67		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.13	U	0.6		
Bromodichloromethane	75-27-4	163.8	0.06	U	0.4		
4-Methyl-2-Pentanone	108-10-1	100.2	0.1	U	0.41		
Toluene	108-88-3	92.14	0.19	J	0.72		
t-1,3-Dichloropropene	10061-02-6	111	0.15	U	0.68		
cis-1,3-Dichloropropene	10061-01-5	111	0.11	U	0.5		
1,1,2-Trichloroethane	79-00-5	133.4	0.08	U	0.44		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 13-IA-1

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-05

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.09	U	0.77		
1,2-Dibromoethane	106-93-4	187.9	0.09	U	0.69		
Tetrachloroethene	127-18-4	165.8	0.05		0.34		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.19	U	0.83		
m/p-Xylene	179601-23-1	106.2	0.41	U	1.78		
o-Xylene	95-47-6	106.2	0.21	U	0.91		
Styrene	100-42-5	104.1	0.16	U	0.68		
Bromoform	75-25-2	252.8	0.05	U	0.52		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.02	U	0.14		
2-Chlorotoluene	95-49-8	126.6	0.17	U	0.88		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.18	U	0.88		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.18	U	0.88		
1,3-Dichlorobenzene	541-73-1	147	0.05	U	0.3		
1,4-Dichlorobenzene	106-46-7	147	0.13	U	0.78		
1,2-Dichlorobenzene	95-50-1	147	0.13	U	0.78		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.21	U	1.56		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.13	U	1.39		
1,3-Butadiene	106-99-0	54.09	0.05	U	0.11		
Naphthalene	91-20-3	128.17	0.01	U	0.05		
4-Ethyltoluene	622-96-8	120.2	0.21	U	1.03		
Hexane	110-54-3	86.17	2.3		8.11		
Allyl Chloride	107-05-1	76.53	0.11	U	0.34		
1,4-Dioxane	123-91-1	88.12	0.22	U	0.79		
Methyl Methacrylate	80-62-6	100.117	0.14	U	0.57		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 16-IA-1

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-06

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.42	J	2.08		
Chloromethane	74-87-3	50.49	0.74		1.53		
Vinyl Chloride	75-01-4	62.5	0.03	U	0.08		
Bromomethane	74-83-9	94.94	0.08	U	0.31		
Chloroethane	75-00-3	64.52	0.15	U	0.4		
Tetrahydrofuran	109-99-9	72.11	0.08	U	0.24		
Trichlorofluoromethane	75-69-4	137.4	0.21	J	1.18		
Dichlorotetrafluoroethane	76-14-2	170.9	0.15	U	1.05		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.14	U	1.07		
tert-Butyl alcohol	75-65-0	74.12	0.16	U	0.49		
Heptane	142-82-5	100.2	0.17	U	0.7		
1,1-Dichloroethene	75-35-4	96.94	0.15	U	0.59		
Acetone	67-64-1	58.08	29.8	E	70.8		
Carbon Disulfide	75-15-0	76.14	0.08	U	0.25		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.23	U	0.83		
Methylene Chloride	75-09-2	84.94	7.3		25.4		
trans-1,2-Dichloroethene	156-60-5	96.94	0.12	U	0.48		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.44	J	1.3		
Carbon Tetrachloride	56-23-5	153.8	0.08		0.5		
cis-1,2-Dichloroethene	156-59-2	96.94	0.1	U	0.4		
Chloroform	67-66-3	119.4	0.1	U	0.49		
1,1,1-Trichloroethane	71-55-6	133.4	0.02	U	0.11		
2,2,4-Trimethylpentane	540-84-1	114.2	0.14	U	0.65		
Benzene	71-43-2	78.11	0.2	J	0.64		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.13	U	0.6		
Bromodichloromethane	75-27-4	163.8	0.06	U	0.4		
4-Methyl-2-Pentanone	108-10-1	100.2	0.11	J	0.45		
Toluene	108-88-3	92.14	0.26	J	0.98		
t-1,3-Dichloropropene	10061-02-6	111	0.15	U	0.68		
cis-1,3-Dichloropropene	10061-01-5	111	0.11	U	0.5		
1,1,2-Trichloroethane	79-00-5	133.4	0.08	U	0.44		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 16-IA-1

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-06

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.09	U	0.77		
1,2-Dibromoethane	106-93-4	187.9	0.09	U	0.69		
Tetrachloroethene	127-18-4	165.8	0.02	U	0.14		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.22	J	0.96		
m/p-Xylene	179601-23-1	106.2	0.9	J	3.91		
o-Xylene	95-47-6	106.2	0.24	J	1.04		
Styrene	100-42-5	104.1	0.16	U	0.68		
Bromoform	75-25-2	252.8	0.05	U	0.52		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.02	U	0.14		
2-Chlorotoluene	95-49-8	126.6	0.17	U	0.88		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.18	U	0.88		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.18	U	0.88		
1,3-Dichlorobenzene	541-73-1	147	0.05	U	0.3		
1,4-Dichlorobenzene	106-46-7	147	0.13	U	0.78		
1,2-Dichlorobenzene	95-50-1	147	0.13	U	0.78		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.21	U	1.56		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.13	U	1.39		
Naphthalene	91-20-3	128.17	0.01	U	0.05		
1,3-Butadiene	106-99-0	54.09	0.05	U	0.11		
4-Ethyltoluene	622-96-8	120.2	0.21	U	1.03		
Hexane	110-54-3	86.17	2.7		9.52		
Allyl Chloride	107-05-1	76.53	0.11	U	0.34		
1,4-Dioxane	123-91-1	88.12	0.22	U	0.79		
Methyl Methacrylate	80-62-6	100.117	0.14	U	0.57		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 16-IA-1DL

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-06DL

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	1.1	UD	5.44		
Chloromethane	74-87-3	50.49	0.98	UD	2.02		
Vinyl Chloride	75-01-4	62.5	0.25	UD	0.64		
Bromomethane	74-83-9	94.94	0.77	UD	2.99		
Chloroethane	75-00-3	64.52	1.5	UD	3.96		
Tetrahydrofuran	109-99-9	72.11	0.8	UD	2.36		
Trichlorofluoromethane	75-69-4	137.4	1.7	UD	9.55		
Dichlorotetrafluoroethane	76-14-2	170.9	1.5	UD	10.5		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	1.4	UD	10.7		
tert-Butyl alcohol	75-65-0	74.12	1.6	UD	4.85		
Heptane	142-82-5	100.2	1.7	UD	6.97		
1,1-Dichloroethene	75-35-4	96.94	1.5	UD	5.95		
Acetone	67-64-1	58.08	35	D	83.1		
Carbon Disulfide	75-15-0	76.14	0.8	UD	2.49		
Methyl tert-Butyl Ether	1634-04-4	88.15	2.3	UD	8.29		
Methylene Chloride	75-09-2	84.94	9.1	D	31.6		
trans-1,2-Dichloroethene	156-60-5	96.94	1.2	UD	4.76		
1,1-Dichloroethane	75-34-3	98.96	1.3	UD	5.26		
Cyclohexane	110-82-7	84.16	2.2	UD	7.57		
2-Butanone	78-93-3	72.11	0.56	UD	1.65		
Carbon Tetrachloride	56-23-5	153.8	0.25	UD	1.57		
cis-1,2-Dichloroethene	156-59-2	96.94	0.99	UD	3.93		
Chloroform	67-66-3	119.4	0.96	UD	4.69		
1,1,1-Trichloroethane	71-55-6	133.4	0.16	UD	0.87		
2,2,4-Trimethylpentane	540-84-1	114.2	1.4	UD	6.54		
Benzene	71-43-2	78.11	0.79	UD	2.52		
1,2-Dichloroethane	107-06-2	98.96	0.93	UD	3.76		
Trichloroethene	79-01-6	131.4	0.24	UD	1.29		
1,2-Dichloropropane	78-87-5	113	1.3	UD	6.01		
Bromodichloromethane	75-27-4	163.8	0.63	UD	4.22		
4-Methyl-2-Pentanone	108-10-1	100.2	0.97	UD	3.98		
Toluene	108-88-3	92.14	1.6	UD	6.03		
t-1,3-Dichloropropene	10061-02-6	111	1.5	UD	6.81		
cis-1,3-Dichloropropene	10061-01-5	111	1.1	UD	4.99		
1,1,2-Trichloroethane	79-00-5	133.4	0.81	UD	4.42		

Project : CHADWICK SQUARE

Sampling Date : 12/03/25

Field Id Number : 16-IA-1DL

Analysis Date : 12/08/25

Laboratory Id Number : Q3799-06DL

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.85	UD	7.24		
1,2-Dibromoethane	106-93-4	187.9	0.85	UD	6.53		
Tetrachloroethene	127-18-4	165.8	0.15	UD	1.02		
Chlorobenzene	108-90-7	112.6	0.77	UD	3.55		
Ethyl Benzene	100-41-4	106.2	1.9	UD	8.25		
m/p-Xylene	179601-23-1	106.2	4.1	UD	17.8		
o-Xylene	95-47-6	106.2	2.1	UD	9.12		
Styrene	100-42-5	104.1	1.6	UD	6.81		
Bromoform	75-25-2	252.8	0.48	UD	4.96		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.18	UD	1.24		
2-Chlorotoluene	95-49-8	126.6	1.7	UD	8.8		
1,3,5-Trimethylbenzene	108-67-8	120.2	1.8	UD	8.85		
1,2,4-Trimethylbenzene	95-63-6	120.2	1.8	UD	8.85		
1,3-Dichlorobenzene	541-73-1	147	0.54	UD	3.25		
1,4-Dichlorobenzene	106-46-7	147	1.3	UD	7.82		
1,2-Dichlorobenzene	95-50-1	147	1.3	UD	7.82		
1,2,4-Trichlorobenzene	120-82-1	181.5	2.1	UD	15.6		
Hexachloro-1,3-Butadiene	87-68-3	260.8	1.3	UD	13.9		
Naphthalene	91-20-3	128.17	0.13	UD	0.68		
1,3-Butadiene	106-99-0	54.09	0.51	UD	1.13		
4-Ethyltoluene	622-96-8	120.2	2.1	UD	10.3		
Hexane	110-54-3	86.17	2.4	JD	8.46		
Allyl Chloride	107-05-1	76.53	1.1	UD	3.44		
1,4-Dioxane	123-91-1	88.12	2.2	UD	7.93		
Methyl Methacrylate	80-62-6	100.117	1.4	UD	5.73		

Method	Matrix	Parameter	Instrument ID	Column ID	MDL (ppbv)	RL (ppbv)	MB Data	MB Datafile	StdevVal
TO-15	Air	1,1,1,2-Tetrachloroethane	MSVOA_L	RTX-1	0.089	0.50	0		0.028115408
TO-15	Air	1,1,1-Trichloroethane	MSVOA_L	RTX-1	0.016	0.030	0		0.005089672
TO-15	Air	1,1,2,2-Tetrachloroethane	MSVOA_L	RTX-1	0.018	0.030	0		0.005472877
TO-15	Air	1,1,2-Trichloroethane	MSVOA_L	RTX-1	0.081	0.50	0		0.02572751
TO-15	Air	1,1,2-Trichlorotrifluoroethane	MSVOA_L	RTX-1	0.14	0.50	0		0.042983939
TO-15	Air	1,1-Dichloroethane	MSVOA_L	RTX-1	0.13	0.50	0		0.040178175
TO-15	Air	1,1-Dichloroethene	MSVOA_L	RTX-1	0.15	0.50	0		0.047056197
TO-15	Air	1,2,4-Trichlorobenzene	MSVOA_L	RTX-1	0.21	0.50	0		0.064291005
TO-15	Air	1,2,4-Trimethylbenzene	MSVOA_L	RTX-1	0.18	0.50	0		0.054467115
TO-15	Air	1,2-Dibromoethane	MSVOA_L	RTX-1	0.085	0.10	0		0.026726124
TO-15	Air	1,2-Dichlorobenzene	MSVOA_L	RTX-1	0.13	0.50	0		0.040999419
TO-15	Air	1,2-Dichloroethane	MSVOA_L	RTX-1	0.093	0.50	0		0.029277002
TO-15	Air	1,2-Dichloropropane	MSVOA_L	RTX-1	0.13	0.50	0		0.039880775
TO-15	Air	1,3,5-Trimethylbenzene	MSVOA_L	RTX-1	0.18	0.50	0		0.054598099
TO-15	Air	1,3-Butadiene	MSVOA_L	RTX-1	0.051	0.50	0		0.016035675
TO-15	Air	1,3-Dichlorobenzene	MSVOA_L	RTX-1	0.054	0.50	0		0.017043362
TO-15	Air	1,4-Dichlorobenzene	MSVOA_L	RTX-1	0.13	0.50	0		0.039460649
TO-15	Air	1,4-Dioxane	MSVOA_L	RTX-1	0.22	0.50	0		0.067330033
TO-15	Air	2,2,4-Trimethylpentane	MSVOA_L	RTX-1	0.14	0.50	0		0.044131837
TO-15	Air	2-Butanone	MSVOA_L	RTX-1	0.056	0.50	0		0.017728105
TO-15	Air	2-Chlorotoluene	MSVOA_L	RTX-1	0.17	0.50	0		0.052599113
TO-15	Air	2-Hexanone	MSVOA_L	RTX-1	0.11	0.50	0		0.033523268
TO-15	Air	4-Ethyltoluene	MSVOA_L	RTX-1	0.21	0.50	0		0.066332496
TO-15	Air	4-Methyl-2-Pentanone	MSVOA_L	RTX-1	0.097	0.50	0		0.030783422
TO-15	Air	Acetone	MSVOA_L	RTX-1	0.18	0.50	0		0.055592052
TO-15	Air	Allyl Chloride	MSVOA_L	RTX-1	0.11	0.50	0		0.033380918
TO-15	Air	Benzene	MSVOA_L	RTX-1	0.079	0.50	0		0.025071327
TO-15	Air	Benzyl Chloride	MSVOA_L	RTX-1	0.089	0.50	0		0.028284271
TO-15	Air	Bromodichloromethane	MSVOA_L	RTX-1	0.063	0.50	0		0.019880596
TO-15	Air	Bromoethene	MSVOA_L	RTX-1	0.2	0.50	0		0.062678317
TO-15	Air	Bromoform	MSVOA_L	RTX-1	0.048	0.50	0		0.015118579
TO-15	Air	Bromomethane	MSVOA_L	RTX-1	0.077	0.50	0		0.024494897
TO-15	Air	Carbon Disulfide	MSVOA_L	RTX-1	0.08	0.50	0		0.02544836
TO-15	Air	Carbon Tetrachloride	MSVOA_L	RTX-1	0.025	0.030	0		0.007690439
TO-15	Air	Chlorobenzene	MSVOA_L	RTX-1	0.077	0.50	0		0.024397502
TO-15	Air	Chlorodifluoromethane	MSVOA_L	RTX-1	0.12	0.50	0		0.037859389
TO-15	Air	Chloroethane	MSVOA_L	RTX-1	0.15	0.50	0		0.045250625
TO-15	Air	Chloroform	MSVOA_L	RTX-1	0.096	0.50	0		0.030394235
TO-15	Air	Chloromethane	MSVOA_L	RTX-1	0.098	0.50	0		0.03101459
TO-15	Air	cis-1,2-Dichloroethene	MSVOA_L	RTX-1	0.099	0.50	0		0.031471832
TO-15	Air	cis-1,3-Dichloropropene	MSVOA_L	RTX-1	0.11	0.50	0		0.032071349
TO-15	Air	Cyclohexane	MSVOA_L	RTX-1	0.22	0.50	0		0.06972736
TO-15	Air	Dibromochloromethane	MSVOA_L	RTX-1	0.085	0.50	0		0.026726124
TO-15	Air	Dichlorodifluoromethane	MSVOA_L	RTX-1	0.11	0.50	0		0.033380918
TO-15	Air	Dichlorotetrafluoroethane	MSVOA_L	RTX-1	0.15	0.50	0		0.045460606
TO-15	Air	Ethanol	MSVOA_L	RTX-1	0.32	0.50	0		0.098848128
TO-15	Air	Ethyl Acetate	MSVOA_L	RTX-1	0.12	0.50	0		0.036384193
TO-15	Air	Ethyl Benzene	MSVOA_L	RTX-1	0.19	0.50	0		0.058554004
TO-15	Air	Heptane	MSVOA_L	RTX-1	0.17	0.50	0		0.051130086
TO-15	Air	Hexachloro-1,3-Butadiene	MSVOA_L	RTX-1	0.13	0.50	0		0.038297084
TO-15	Air	Hexane	MSVOA_L	RTX-1	0.16	0.50	0		0.04894117
TO-15	Air	Isopropyl Alcohol	MSVOA_L	RTX-1	0.23	0.50	0		0.070676998
TO-15	Air	Isopropylbenzene	MSVOA_L	RTX-1	0.16	0.50	0		0.048205908
TO-15	Air	m/p-Xylene	MSVOA_L	RTX-1	0.41	1.00	0		0.130091543
TO-15	Air	Methyl Methacrylate	MSVOA_L	RTX-1	0.14	0.50	0		0.044131837
TO-15	Air	Methyl tert-Butyl Ether	MSVOA_L	RTX-1	0.23	0.50	0		0.07296444
TO-15	Air	Methylene Chloride	MSVOA_L	RTX-1	0.23	0.50	0		0.07204496
TO-15	Air	Naphthalene	MSVOA_L	RTX-1	0.013	0.10	0		0.004076647
TO-15	Air	Naphthalene,2-methyl-	MSVOA_L	RTX-1	0.22	0.50	0		0.069213266
TO-15	Air	n-Butylbenzene	MSVOA_L	RTX-1	0.18	0.50	0		0.055805786
TO-15	Air	n-propylbenzene	MSVOA_L	RTX-1	0.2	0.50	0		0.06164441
TO-15	Air	o-Xylene	MSVOA_L	RTX-1	0.21	0.50	0		0.06491753
TO-15	Air	p-Isopropyltoluene	MSVOA_L	RTX-1	0.19	0.50	0		0.059561893
TO-15	Air	Propene	MSVOA_L	RTX-1	0.077	0.50	0		0.024397502
TO-15	Air	sec-Butylbenzene	MSVOA_L	RTX-1	0.2	0.50	0		0.061913919
TO-15	Air	Styrene	MSVOA_L	RTX-1	0.16	0.50	0		0.05
TO-15	Air	t-1,3-Dichloropropene	MSVOA_L	RTX-1	0.15	0.50	0		0.046496288
TO-15	Air	tert-Butyl alcohol	MSVOA_L	RTX-1	0.16	0.50	0		0.048892496
TO-15	Air	tert-Butylbenzene	MSVOA_L	RTX-1	0.22	0.50	0		0.067928534
TO-15	Air	Tetrachloroethene	MSVOA_L	RTX-1	0.015	0.030	0		0.004725816
TO-15	Air	Tetrahydrofuran	MSVOA_L	RTX-1	0.08	0.50	0		0.025166115
TO-15	Air	Toluene	MSVOA_L	RTX-1	0.16	0.50	0		0.050142654
TO-15	Air	trans-1,2-Dichloroethene	MSVOA_L	RTX-1	0.12	0.50	0		0.036968455
TO-15	Air	Trichloroethene	MSVOA_L	RTX-1	0.024	0.030	0		0.007587584
TO-15	Air	Trichlorofluoromethane	MSVOA_L	RTX-1	0.17	0.50	0		0.053452248
TO-15	Air	Vinyl Acetate	MSVOA_L	RTX-1	0.26	0.50	0		0.079671946
TO-15	Air	Vinyl Chloride	MSVOA_L	RTX-1	0.025	0.030	0		0.007955232

Students	Spike_Amt_A	Datafile1	Concentration1	Adate1	Datafile2	Concentration2	Adate2
Value	dded						
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.03	VL041883.D	0.037	01102025T00:17:00	VL041884.D	0.04	01102025T00:47:00
3.143	0.03	VL041883.D	0.045	01102025T00:17:00	VL041884.D	0.04	01102025T00:47:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.54	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.5	01092025T21:16:00	VL041878.D	0.53	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.52	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.52	01092025T21:16:00	VL041878.D	0.54	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.49	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.52	01092025T21:16:00	VL041878.D	0.54	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.41	01092025T21:16:00	VL041878.D	0.38	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.6	01092025T21:16:00	VL041878.D	0.62	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.31	01092025T21:16:00	VL041878.D	0.29	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.53	01092025T21:46:00
3.143	0.4	VL041877.D	0.36	01092025T21:16:00	VL041878.D	0.35	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.4	VL041877.D	0.41	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.03	VL041883.D	0.036	01102025T00:17:00	VL041884.D	0.038	01102025T00:47:00
3.143	0.4	VL041877.D	0.5	01092025T21:16:00	VL041878.D	0.48	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.57	01092025T21:16:00	VL041878.D	0.59	01092025T21:46:00
3.143	0.4	VL041877.D	0.51	01092025T21:16:00	VL041878.D	0.5	01092025T21:46:00
3.143	0.4	VL041877.D	0.54	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.41	01092025T21:16:00	VL041878.D	0.4	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.39	01092025T21:16:00	VL041878.D	0.4	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.4	VL042501.D	0.885	05062025T17:36:00	VL042502.D	0.76	05062025T18:07:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.43	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.43	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.5	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.59	01092025T21:16:00	VL041878.D	0.6	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.91	01092025T21:16:00	VL041878.D	0.91	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.41	01092025T21:46:00
3.143	0.4	VL041877.D	0.4	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.68	01092025T21:16:00	VL041878.D	0.61	01092025T21:46:00
3.143	0.04	VL042504.D	0.035	05062025T19:10:00	VL042505.D	0.04	05062025T19:42:00
3.143	0.4	VL041877.D	0.34	01092025T21:16:00	VL041878.D	0.34	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.42	01092025T21:16:00	VL041878.D	0.4	01092025T21:46:00
3.143	0.4	VL041877.D	0.41	01092025T21:16:00	VL041878.D	0.36	01092025T21:46:00
3.143	0.4	VL041877.D	0.5	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.03	VL041883.D	0.037	01102025T00:17:00	VL041884.D	0.04	01102025T00:47:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.03	VL041883.D	0.047	01102025T00:17:00	VL041884.D	0.047	01102025T00:47:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.4	VL041877.D	0.43	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.03	VL042498.D	0.018	05062025T16:01:00	VL042500.D	0.032	05062025T17:04:00

Datafile3	Concentration3	Adate3	Datafile4	Concentration4	Adate4	Datafile5	Concentration5
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.4
VL041885.D	0.038	01102025T01:18:00	VL041913.D	0.033	01152025T02:53:00	VL041914.D	0.037
VL041885.D	0.044	01102025T01:18:00	VL041913.D	0.033	01152025T02:53:00	VL041914.D	0.032
VL041879.D	0.48	01092025T22:17:00	VL041909.D	0.44	01152025T00:52:00	VL041910.D	0.47
VL041879.D	0.55	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.47
VL041879.D	0.57	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.48
VL041879.D	0.58	01092025T22:17:00	VL041909.D	0.49	01152025T00:52:00	VL041910.D	0.49
VL041879.D	0.48	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.47	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.39
VL041879.D	0.47	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.43
VL041879.D	0.53	01092025T22:17:00	VL041909.D	0.44	01152025T00:52:00	VL041910.D	0.43
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.46
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.43	01152025T00:52:00	VL041910.D	0.42
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.51	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.51
VL041879.D	0.48	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.44
VL041879.D	0.51	01092025T22:17:00	VL041909.D	0.43	01152025T00:52:00	VL041910.D	0.41
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.45	01152025T00:52:00	VL041910.D	0.47
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.39
VL041879.D	0.49	01092025T22:17:00	VL041909.D	0.44	01152025T00:52:00	VL041910.D	0.44
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.39	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.33
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.39	01152025T00:52:00	VL041910.D	0.37
VL041879.D	0.68	01092025T22:17:00	VL041909.D	0.53	01152025T00:52:00	VL041910.D	0.53
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.41
VL041879.D	0.28	01092025T22:17:00	VL041909.D	0.26	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.4	01152025T00:52:00	VL041910.D	0.4
VL041879.D	0.55	01092025T22:17:00	VL041909.D	0.44	01152025T00:52:00	VL041910.D	0.52
VL041879.D	0.37	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.53	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.41	01152025T00:52:00	VL041910.D	0.4
VL041885.D	0.048	01102025T01:18:00	VL041913.D	0.026	01152025T02:53:00	VL041914.D	0.033
VL041879.D	0.52	01092025T22:17:00	VL041909.D	0.45	01152025T00:52:00	VL041910.D	0.46
VL041879.D	0.51	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.42
VL041879.D	0.52	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.53	01092025T22:17:00	VL041909.D	0.45	01152025T00:52:00	VL041910.D	0.46
VL041879.D	0.53	01092025T22:17:00	VL041909.D	0.52	01152025T00:52:00	VL041910.D	0.45
VL041879.D	0.49	01092025T22:17:00	VL041909.D	0.43	01152025T00:52:00	VL041910.D	0.42
VL041879.D	0.38	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.35	01152025T00:52:00	VL041910.D	0.35
VL041879.D	0.41	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.56	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.56	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.46
VL042503.D	0.611	05062025T18:39:00	VL042518.D	0.831	05072025T15:02:00	VL042519.D	0.842
VL041879.D	0.47	01092025T22:17:00	VL041909.D	0.41	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.35
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.37
VL041879.D	0.51	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.43
VL041879.D	0.49	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.63	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.92	01092025T22:17:00	VL041909.D	0.72	01152025T00:52:00	VL041910.D	0.69
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.36	01152025T00:52:00	VL041910.D	0.35
VL041879.D	0.42	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.52
VL041879.D	0.63	01092025T22:17:00	VL041909.D	0.54	01152025T00:52:00	VL041910.D	0.57
VL042506.D	0.04	05062025T20:13:00	VL042521.D	0.031	05072025T16:49:00	VL042522.D	0.032
VL041879.D	0.33	01092025T22:17:00	VL041909.D	0.28	01152025T00:52:00	VL041910.D	0.25
VL041879.D	0.42	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.36	01152025T00:52:00	VL041910.D	0.33
VL041879.D	0.47	01092025T22:17:00	VL041909.D	0.35	01152025T00:52:00	VL041910.D	0.35
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.52	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.48
VL041879.D	0.45	01092025T22:17:00	VL041909.D	0.38	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.38	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.31
VL041879.D	0.38	01092025T22:17:00	VL041909.D	0.36	01152025T00:52:00	VL041910.D	0.33
VL041879.D	0.52	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.41
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.36	01152025T00:52:00	VL041910.D	0.37
VL041885.D	0.034	01102025T01:18:00	VL041913.D	0.037	01152025T02:53:00	VL041914.D	0.034
VL041879.D	0.41	01092025T22:17:00	VL041909.D	0.41	01152025T00:52:00	VL041910.D	0.42
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.54	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.48
VL041885.D	0.035	01102025T01:18:00	VL041913.D	0.029	01152025T02:53:00	VL041914.D	0.033
VL041879.D	0.54	01092025T22:17:00	VL041909.D	0.45	01152025T00:52:00	VL041910.D	0.47
VL041879.D	0.56	01092025T22:17:00	VL041909.D	0.35	01152025T00:52:00	VL041910.D	0.44
VL042515.D	0.021	05072025T13:17:00	VL042516.D	0.021	05072025T13:50:00	VL042517.D	0.014

Adate5	Datafile6	Concentration6	Adate6	Datafile7	Concentration7	Adate7	Department
01152025T01:23:00	VL041932.D	0.41	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T03:23:00	VL041936.D	0.028	01212025T00:31:00	VL041940.D	0.027	01212025T09:44:00	VOC
01152025T03:23:00	VL041936.D	0.034	01212025T00:31:00	VL041940.D	0.034	01212025T09:44:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.49	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.44	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.45	01202025T22:30:00	VL041933.D	0.47	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.47	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.34	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.35	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.45	01202025T22:30:00	VL041933.D	0.4	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.45	01202025T22:30:00	VL041933.D	0.49	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.49	01202025T22:30:00	VL041933.D	0.47	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.47	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.49	01202025T22:30:00	VL041933.D	0.52	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.45	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.3	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.41	01202025T22:30:00	VL041933.D	0.37	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.47	01202025T22:30:00	VL041933.D	0.46	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.36	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.35	01202025T22:30:00	VL041933.D	0.31	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.31	01202025T22:30:00	VL041933.D	0.3	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.38	01202025T22:30:00	VL041933.D	0.37	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.58	01202025T22:30:00	VL041933.D	0.54	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.43	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.47	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.29	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.42	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.39	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.34	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.47	01202025T22:30:00	VL041933.D	0.52	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.38	01202025T22:30:00	VL041933.D	0.38	01202025T23:00:00	VOC
01152025T03:23:00	VL041936.D	0.029	01212025T00:31:00	VL041940.D	0.027	01212025T09:44:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.48	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.51	01202025T22:30:00	VL041933.D	0.49	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.52	01202025T22:30:00	VL041933.D	0.49	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.47	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.48	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.35	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.37	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.38	01202025T22:30:00	VL041933.D	0.36	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.48	01202025T22:30:00	VL041933.D	0.51	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.45	01202025T22:30:00	VL041933.D	0.45	01202025T23:00:00	VOC
05072025T15:37:00	VL042520.D	0.766	05072025T16:13:00	VL042569.D	0.664	05212025T16:53:00	VOC
01152025T01:23:00	VL041932.D	0.39	01202025T22:30:00	VL041933.D	0.4	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.32	01202025T22:30:00	VL041933.D	0.31	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.37	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.44	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.4	01202025T22:30:00	VL041933.D	0.36	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.46	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.36	01202025T22:30:00	VL041933.D	0.35	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.64	01202025T22:30:00	VL041933.D	0.65	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.35	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.35	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.48	01202025T23:00:00	VOC
05072025T17:24:00	VL042523.D	0.03	05072025T17:58:00	VL042567.D	0.034	05212025T15:49:00	VOC
01152025T01:23:00	VL041932.D	0.2	01202025T22:30:00	VL041933.D	0.17	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.31	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.31	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.32	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.32	01202025T22:30:00	VL041933.D	0.32	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.48	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.34	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.3	01202025T22:30:00	VL041933.D	0.3	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.27	01202025T22:30:00	VL041933.D	0.31	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.44	01202025T22:30:00	VL041933.D	0.38	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.3	01202025T23:00:00	VOC
01152025T03:23:00	VL041936.D	0.026	01212025T00:31:00	VL041940.D	0.03	01212025T09:44:00	VOC
01152025T01:23:00	VL041932.D	0.39	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.34	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.44	01202025T23:00:00	VOC
01152025T03:23:00	VL041936.D	0.033	01212025T00:31:00	VL041940.D	0.03	01212025T09:44:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.37	01202025T23:00:00	VOC
05072025T14:29:00	VL042575.D	0.034	05222025T11:27:00	VL042576.D	0.032	05222025T11:59:00	VOC

AVG Concentration	Recovery	Effective Date
0.43	108	5/22/2025
0.03	114	5/22/2025
0.04	125	5/22/2025
0.47	119	5/22/2025
0.50	125	5/22/2025
0.50	125	5/22/2025
0.50	125	5/22/2025
0.40	100	5/22/2025
0.40	100	5/22/2025
0.45	112	5/22/2025
0.48	120	5/22/2025
0.49	123	5/22/2025
0.48	119	5/22/2025
0.38	95	5/22/2025
0.51	128	5/22/2025
0.46	116	5/22/2025
0.45	112	5/22/2025
0.44	110	5/22/2025
0.42	105	5/22/2025
0.46	115	5/22/2025
0.40	100	5/22/2025
0.36	91	5/22/2025
0.37	93	5/22/2025
0.40	100	5/22/2025
0.58	146	5/22/2025
0.47	118	5/22/2025
0.45	111	5/22/2025
0.30	75	5/22/2025
0.42	106	5/22/2025
0.48	119	5/22/2025
0.35	86	5/22/2025
0.50	125	5/22/2025
0.41	102	5/22/2025
0.03	113	5/22/2025
0.48	121	5/22/2025
0.47	118	5/22/2025
0.52	130	5/22/2025
0.48	121	5/22/2025
0.50	126	5/22/2025
0.44	111	5/22/2025
0.36	91	5/22/2025
0.40	101	5/22/2025
0.38	95	5/22/2025
0.50	125	5/22/2025
0.49	123	5/22/2025
0.77	191	5/22/2025
0.42	106	5/22/2025
0.38	94	5/22/2025
0.40	100	5/22/2025
0.46	115	5/22/2025
0.43	106	5/22/2025
0.53	134	5/22/2025
0.40	101	5/22/2025
0.78	194	5/22/2025
0.38	95	5/22/2025
0.40	101	5/22/2025
0.57	143	5/22/2025
0.03	86	5/22/2025
0.27	68	5/22/2025
0.37	93	5/22/2025
0.38	95	5/22/2025
0.39	97	5/22/2025
0.38	95	5/22/2025
0.50	124	5/22/2025
0.40	100	5/22/2025
0.35	88	5/22/2025
0.35	86	5/22/2025
0.45	113	5/22/2025
0.39	98	5/22/2025
0.03	113	5/22/2025
0.42	105	5/22/2025
0.39	97	5/22/2025
0.48	120	5/22/2025
0.04	121	5/22/2025
0.48	119	5/22/2025
0.42	105	5/22/2025
0.02	82	5/22/2025



QC SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Surrogate Summary

SDG No.: Q3799

Client: RTP Environmental

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3799-01	AA-1	1-Bromo-4-Fluorobenzene	10	10.2	102		70 (65)	130 (135)
Q3799-01DUP	AA-1DUP	1-Bromo-4-Fluorobenzene	10	9.82	98		70 (65)	130 (135)
Q3799-02	9B-IA-1	1-Bromo-4-Fluorobenzene	10	10.3	103		70 (65)	130 (135)
Q3799-03	10-IA-1	1-Bromo-4-Fluorobenzene	10	10.4	104		70 (65)	130 (135)
Q3799-03DL	10-IA-1DL	1-Bromo-4-Fluorobenzene	10	9.74	97		70 (65)	130 (135)
Q3799-03DL2	10-IA-1DL2	1-Bromo-4-Fluorobenzene	10	9.53	95		70 (65)	130 (135)
Q3799-04	11-IA-1	1-Bromo-4-Fluorobenzene	10	10.5	105		70 (65)	130 (135)
Q3799-05	13-IA-1	1-Bromo-4-Fluorobenzene	10	10.2	102		70 (65)	130 (135)
Q3799-06	16-IA-1	1-Bromo-4-Fluorobenzene	10	10.4	104		70 (65)	130 (135)
Q3799-06DL	16-IA-1DL	1-Bromo-4-Fluorobenzene	10	9.90	99		70 (65)	130 (135)
VL1208ABL01	VL1208ABL01	1-Bromo-4-Fluorobenzene	10	9.05	91		70 (65)	130 (135)
VL1208ABS01	VL1208ABS01	1-Bromo-4-Fluorobenzene	10	10.1	101		70 (65)	130 (135)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SDG No.:	Q3799	SW-846	Analytical Method:	SWTO-15
Client:	RTP Environmental		Datafile :	VL043307.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL1208ABS01	Dichlorodifluoromethane	10	10.8	ppbv	108			40 (70)	160 (130)	
	Chloromethane	10	10.4	ppbv	104			40 (70)	160 (130)	
	Vinyl Chloride	10	9.20	ppbv	92			70 (70)	130 (130)	
	Bromomethane	10	9.30	ppbv	93			40 (70)	160 (130)	
	Chloroethane	10	9.90	ppbv	99			40 (70)	160 (130)	
	Tetrahydrofuran	10	9.70	ppbv	97			70 (70)	130 (130)	
	Trichlorofluoromethane	10	11.2	ppbv	112			40 (70)	160 (130)	
	1,1,2-Trichlorotrifluoroethane	10	11.2	ppbv	112			70 (70)	130 (130)	
	Dichlorotetrafluoroethane	10	10.6	ppbv	106			70 (70)	130 (130)	
	tert-Butyl Alcohol	10	8.90	ppbv	89			70 (70)	130 (130)	
	Heptane	10	9.40	ppbv	94			70 (70)	130 (130)	
	1,1-Dichloroethene	10	10.2	ppbv	102			70 (70)	130 (130)	
	Acetone	10	8.60	ppbv	86			40 (70)	160 (130)	
	Carbon disulfide	10	11.1	ppbv	111			40 (70)	160 (130)	
	Methyl tert-butyl Ether	10	9.40	ppbv	94			70 (70)	130 (130)	
	Methylene Chloride	10	11.6	ppbv	116			70 (70)	130 (130)	
	trans-1,2-Dichloroethene	10	10.3	ppbv	103			70 (70)	130 (130)	
	1,1-Dichloroethane	10	10.5	ppbv	105			70 (70)	130 (130)	
	Cyclohexane	10	9.70	ppbv	97			70 (70)	130 (130)	
	2-Butanone	10	10.3	ppbv	103			40 (70)	160 (130)	
	Carbon Tetrachloride	10	12.1	ppbv	121			70 (70)	130 (130)	
	cis-1,2-Dichloroethene	10	9.90	ppbv	99			70 (70)	130 (130)	
	Chloroform	10	11.7	ppbv	117			70 (70)	130 (130)	
	1,1,1-Trichloroethane	10	10.9	ppbv	109			70 (70)	130 (130)	
	2,2,4-Trimethylpentane	10	10.8	ppbv	108			70 (70)	130 (130)	
	Benzene	10	11.2	ppbv	112			70 (70)	130 (130)	
	1,2-Dichloroethane	10	11.9	ppbv	119			70 (70)	130 (130)	
	Trichloroethene	10	11.5	ppbv	115			70 (70)	130 (130)	
	1,2-Dichloropropane	10	10.9	ppbv	109			70 (70)	130 (130)	
	Bromodichloromethane	10	12.0	ppbv	120			70 (70)	130 (130)	
	4-Methyl-2-Pentanone	10	10.0	ppbv	100			40 (70)	160 (130)	
	Toluene	10	10.3	ppbv	103			70 (70)	130 (130)	
	t-1,3-Dichloropropene	10	7.80	ppbv	78			70 (70)	130 (130)	
	cis-1,3-Dichloropropene	10	8.30	ppbv	83			70 (70)	130 (130)	
	1,1,2-Trichloroethane	10	11.6	ppbv	116			70 (70)	130 (130)	
	Dibromochloromethane	10	11.8	ppbv	118			70 (70)	130 (130)	
	1,2-Dibromoethane	10	10.5	ppbv	105			70 (70)	130 (130)	
	Tetrachloroethene	10	11.1	ppbv	111			70 (70)	130 (130)	
	Chlorobenzene	10	11.9	ppbv	119			70 (70)	130 (130)	
	Ethyl Benzene	10	11.2	ppbv	112			70 (70)	130 (130)	
	m/p-Xylene	20	23.5	ppbv	117			70 (70)	130 (130)	
	o-Xylene	10	11.6	ppbv	116			70 (70)	130 (130)	
	Styrene	10	9.90	ppbv	99			70 (70)	130 (130)	
	Bromoform	10	12.2	ppbv	122			70 (70)	130 (130)	
	1,1,2,2-Tetrachloroethane	10	11.6	ppbv	116			70 (70)	130 (130)	
	2-Chlorotoluene	10	11.0	ppbv	110			70 (70)	130 (130)	
	1,3,5-Trimethylbenzene	10	11.4	ppbv	114			70 (70)	130 (130)	
	1,2,4-Trimethylbenzene	10	11.5	ppbv	115			70 (70)	130 (130)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3799</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>RTP Environmental</u>	Datafile :	<u>VL043307.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL1208ABS01	1,3-Dichlorobenzene	10	11.8	ppbv	118			70 (70)	130 (130)	
	1,4-Dichlorobenzene	10	11.9	ppbv	119			70 (70)	130 (130)	
	1,2-Dichlorobenzene	10	11.8	ppbv	118			70 (70)	130 (130)	
	1,2,4-Trichlorobenzene	10	11.9	ppbv	119			70 (70)	130 (130)	
	Hexachloro-1,3-butadiene	10	10.9	ppbv	109			70 (70)	130 (130)	
	Naphthalene	10	10.7	ppbv	107			40 (70)	160 (130)	
	1,3-Butadiene	10	9.60	ppbv	96			70 (70)	130 (130)	
	4-Ethyltoluene	10	11.5	ppbv	115			70 (70)	130 (130)	
	Hexane	10	9.90	ppbv	99			70 (70)	130 (130)	
	Allyl Chloride	10	8.80	ppbv	88			70 (70)	130 (130)	
	1,4-Dioxane	10	8.70	ppbv	87			40 (70)	160 (130)	
	Methyl methacrylate	10	9.70	ppbv	97			70 (70)	130 (130)	

() = LABORATORY INHOUSE LIMIT

Duplicate Sample Summary

Lab Sample Id :	Q3799-01DUP	Q3799-01
Client Id :	AA-1DUP	AA-1
DF :	1	1
Datafile :	VL043311.D	VL043310.D
Anal Date & Time :	12/08/2025 13:16	12/08/2025 12:40

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.51	0.5	2
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	4.8	4.8	0
Allyl Chloride	0	0	0
Benzene	0.16	0.18	11.8
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.09	0.09	0

Duplicate Sample Summary

Lab Sample Id :	Q3799-01DUP	Q3799-01
Client Id :	AA-1DUP	AA-1
DF :	1	1
Datafile :	VL043311.D	VL043310.D
Anal Date & Time :	12/08/2025 13:16	12/08/2025 12:40

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0.15	0.15	0
Chloromethane	2.1	2.2	4.7
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.41	0.39	5
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	1.4	1.2	15.4
m/p-Xylene	0	0	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	3.3	2.4	31.6 *
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0	0	0
Tetrahydrofuran	0	0	0
Toluene	0	0	0
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.2	0.21	4.9
Vinyl Chloride	0	0	0

VOLATILE METHOD BLANK SUMMARY

Client ID

VL1208ABL01

Lab Name: Alliance

Contract: RTPE01

Lab Code: ACE

SDG NO.: Q3799

Lab File ID: VL043306.D

Lab Sample ID: VL1208ABL01

Date Analyzed: 12/08/2025

Time Analyzed: 09:36

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_L

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VL1208ABS01	VL1208ABS01	VL043307.D	12/08/2025
AA-1	Q3799-01	VL043310.D	12/08/2025
AA-1DUP	Q3799-01DUP	VL043311.D	12/08/2025
9B-IA-1	Q3799-02	VL043312.D	12/08/2025
10-IA-1	Q3799-03	VL043313.D	12/08/2025
11-IA-1	Q3799-04	VL043314.D	12/08/2025
13-IA-1	Q3799-05	VL043315.D	12/08/2025
16-IA-1	Q3799-06	VL043316.D	12/08/2025
10-IA-1DL	Q3799-03DL	VL043322.D	12/08/2025
16-IA-1DL	Q3799-06DL	VL043324.D	12/08/2025
10-IA-1DL2	Q3799-03DL2	VL043331.D	12/09/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: VL043202.D
Instrument ID: MSVOA_L
GC Column: RTX-1 ID: 0.32 (mm)

Contract: RTPE01
SDG NO.: Q3799
BFB Injection Date: 11/18/2025
BFB Injection Time: 08:46
Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.1
75	30.0 - 66.0% of mass 95	55.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 120.0% of mass 95	63.6
175	4.0 - 9.0% of mass 174	5 (7.8) 1
176	93.0 - 101.0% of mass 174	60.7 (95.5) 1
177	5.0 - 9.0% of mass 176	4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC0.5	VSTDICC0.5	VL043203.D	11/18/2025	09:29
VSTDICCC010	VSTDICCC010	VL043204.D	11/18/2025	10:02
VSTDICC002	VSTDICC002	VL043205.D	11/18/2025	10:37
VSTDICC001	VSTDICC001	VL043206.D	11/18/2025	11:10
VSTDICC0.1	VSTDICC0.1	VL043208.D	11/18/2025	12:16
VSTDICC0.03	VSTDICC0.03	VL043209.D	11/18/2025	12:49
VSTDICC015	VSTDICC015	VL043210.D	11/18/2025	13:24

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: VL043304.D
Instrument ID: MSVOA_L
GC Column: RTX-1 ID: 0.32 (mm)

Contract: RTPE01
SDG NO.: Q3799
BFB Injection Date: 12/08/2025
BFB Injection Time: 07:56
Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.8
75	30.0 - 66.0% of mass 95	55.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.6 (0.9) 1
174	50.0 - 120.0% of mass 95	67.2
175	4.0 - 9.0% of mass 174	5.2 (7.8) 1
176	93.0 - 101.0% of mass 174	63.1 (93.8) 1
177	5.0 - 9.0% of mass 176	4.3 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL043305.D	12/08/2025	08:44
VL1208ABL01	VL1208ABL01	VL043306.D	12/08/2025	09:36
VL1208ABS01	VL1208ABS01	VL043307.D	12/08/2025	10:24
AA-1	Q3799-01	VL043310.D	12/08/2025	12:40
AA-1DUP	Q3799-01DUP	VL043311.D	12/08/2025	13:16
9B-IA-1	Q3799-02	VL043312.D	12/08/2025	13:52
10-IA-1	Q3799-03	VL043313.D	12/08/2025	14:29
11-IA-1	Q3799-04	VL043314.D	12/08/2025	15:37
13-IA-1	Q3799-05	VL043315.D	12/08/2025	16:25
16-IA-1	Q3799-06	VL043316.D	12/08/2025	17:00
10-IA-1DL	Q3799-03DL	VL043322.D	12/08/2025	20:50
16-IA-1DL	Q3799-06DL	VL043324.D	12/08/2025	21:56
10-IA-1DL2	Q3799-03DL2	VL043331.D	12/09/2025	07:30

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: RTPE01
Lab Code: ACE SDG NO.: Q3799
Lab File ID: VL043305.D Date Analyzed: 12/08/2025
Instrument ID: MSVOA_L Time Analyzed: 08:44
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	129333	2.78	319801	3.95	246102	8.87
UPPER LIMIT	181066	3.11	447721	4.28	344543	9.20
LOWER LIMIT	77599.8	2.45	191881	3.62	147661	8.54
EPA SAMPLE NO.						
AA-1	151890	2.78	389995	3.95	311445	8.87
AA-1DUP	155357	2.78	397814	3.95	318793	8.87
9B-IA-1	137082	2.78	357699	3.95	281919	8.88
10-IA-1	142316	2.78	361997	3.96	299956	8.87
10-IA-1DL	129360	2.78	324163	3.94	261825	8.87
10-IA-1DL2	122404	2.78	307022	3.95	245820	8.87
11-IA-1	138499	2.78	354324	3.95	279503	8.87
13-IA-1	133616	2.78	344933	3.94	261904	8.87
16-IA-1	146093	2.78	393021	3.95	306652	8.87
16-IA-1DL	127959	2.78	316604	3.95	248177	8.87
VL1208ABL01	125653	2.78	312886	3.95	239807	8.87
VL1208ABS01	128373	2.78	319254	3.95	250711	8.87

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: AA-1
Lab Sample ID: Q3799-01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.39	1.93	J	1	0.54	2.47	12/08/25 12:40	VL120825
74-87-3	Chloromethane	2.20	4.54		1	0.21	1.03	12/08/25 12:40	VL120825
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 12:40	VL120825
74-83-9	Bromomethane	0.080	0.31	U	1	0.31	1.94	12/08/25 12:40	VL120825
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	12/08/25 12:40	VL120825
109-99-9	Tetrahydrofuran	0.080	0.24	U	1	0.24	1.47	12/08/25 12:40	VL120825
75-69-4	Trichlorofluoromethane	0.21	1.18	J	1	0.96	2.81	12/08/25 12:40	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1	1.07	3.83	12/08/25 12:40	VL120825
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1	1.05	3.49	12/08/25 12:40	VL120825
75-65-0	tert-Butyl alcohol	0.16	0.49	U	1	0.49	1.52	12/08/25 12:40	VL120825
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	12/08/25 12:40	VL120825
75-35-4	1,1-Dichloroethene	0.15	0.59	U	1	0.59	1.98	12/08/25 12:40	VL120825
67-64-1	Acetone	4.80	11.4		1	0.43	1.19	12/08/25 12:40	VL120825
75-15-0	Carbon Disulfide	0.080	0.25	U	1	0.25	1.56	12/08/25 12:40	VL120825
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	1	0.83	1.80	12/08/25 12:40	VL120825
75-09-2	Methylene Chloride	2.40	8.34		1	0.80	1.74	12/08/25 12:40	VL120825
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	1	0.48	1.98	12/08/25 12:40	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 12:40	VL120825
110-82-7	Cyclohexane	0.22	0.76	U	1	0.76	1.72	12/08/25 12:40	VL120825
78-93-3	2-Butanone	0.50	1.47		1	0.18	1.47	12/08/25 12:40	VL120825
56-23-5	Carbon Tetrachloride	0.090	0.57		1	0.19	0.19	12/08/25 12:40	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 12:40	VL120825
67-66-3	Chloroform	0.15	0.73	J	1	0.49	2.44	12/08/25 12:40	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 12:40	VL120825
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	1	0.65	2.34	12/08/25 12:40	VL120825
71-43-2	Benzene	0.18	0.58	J	1	0.26	1.60	12/08/25 12:40	VL120825
107-06-2	1,2-Dichloroethane	0.090	0.36	U	1	0.36	2.02	12/08/25 12:40	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 12:40	VL120825
78-87-5	1,2-Dichloropropane	0.13	0.60	U	1	0.60	2.31	12/08/25 12:40	VL120825
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	12/08/25 12:40	VL120825
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	1	0.41	2.05	12/08/25 12:40	VL120825
108-88-3	Toluene	0.16	0.60	U	1	0.60	1.88	12/08/25 12:40	VL120825
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	1	0.68	2.27	12/08/25 12:40	VL120825
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	1	0.50	2.27	12/08/25 12:40	VL120825
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	1	0.44	2.73	12/08/25 12:40	VL120825
124-48-1	Dibromochloromethane	0.090	0.77	U	1	0.77	4.26	12/08/25 12:40	VL120825
106-93-4	1,2-Dibromoethane	0.090	0.69	U	1	0.69	0.77	12/08/25 12:40	VL120825
127-18-4	Tetrachloroethene	0.020	0.14	U	1	0.14	0.20	12/08/25 12:40	VL120825
108-90-7	Chlorobenzene	0.080	0.37	U	1	0.37	2.30	12/08/25 12:40	VL120825
100-41-4	Ethyl Benzene	0.19	0.83	U	1	0.83	2.17	12/08/25 12:40	VL120825
179601-23-1	m/p-Xylene	0.41	1.78	U	1	1.78	4.34	12/08/25 12:40	VL120825
95-47-6	o-Xylene	0.21	0.91	U	1	0.91	2.17	12/08/25 12:40	VL120825
100-42-5	Styrene	0.16	0.68	U	1	0.68	2.13	12/08/25 12:40	VL120825
75-25-2	Bromoform	0.050	0.52	U	1	0.52	5.17	12/08/25 12:40	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	1	0.14	0.21	12/08/25 12:40	VL120825
95-49-8	2-Chlorotoluene	0.17	0.88	U	1	0.88	2.59	12/08/25 12:40	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: AA-1
Lab Sample ID: Q3799-01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 12:40	VL120825
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 12:40	VL120825
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	1	0.30	3.01	12/08/25 12:40	VL120825
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 12:40	VL120825
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 12:40	VL120825
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1	1.56	3.71	12/08/25 12:40	VL120825
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1	1.39	5.33	12/08/25 12:40	VL120825
106-99-0	1,3-Butadiene	0.050	0.11	U	1	0.11	1.11	12/08/25 12:40	VL120825
91-20-3	Naphthalene	0.010	0.050	U	1	0.050	0.52	12/08/25 12:40	VL120825
622-96-8	4-Ethyltoluene	0.21	1.03	U	1	1.03	2.46	12/08/25 12:40	VL120825
110-54-3	Hexane	1.20	4.23		1	0.56	1.76	12/08/25 12:40	VL120825
107-05-1	Allyl Chloride	0.11	0.34	U	1	0.34	1.57	12/08/25 12:40	VL120825
123-91-1	1,4-Dioxane	0.22	0.79	U	1	0.79	1.80	12/08/25 12:40	VL120825
80-62-6	Methyl Methacrylate	0.14	0.57	U	1	0.57	2.05	12/08/25 12:40	VL120825

SURROGATES

460-00-4 1-Bromo-4-Fluorobenzene 10.2 70 (65) - 130 (135) 102% SPK: 10

INTERNAL STANDARDS

Area Count
74-97-5 Bromochloromethane 152000
540-36-3 1,4-Difluorobenzene 390000
3114-55-4 Chlorobenzene-d5 311000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043310.D
Acq On : 08 Dec 2025 12:40
Operator : SY/MD
Sample : Q3799-01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AA-1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:10:41 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	151890	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	389995	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	311445	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.364	95	233482	10.227	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 102.300%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.496	85	8408m	0.392	ppbv	
3) Chlorodifluoromethane	1.473	51	4994m	0.237	ppbv	
4) Chloromethane	1.531	50	17448	2.178	ppbv	95
9) Propene	1.483	41	5806	0.649	ppbv #	1
11) Trichlorofluoromethane	1.874	101	4433	0.214	ppbv	95
13) Ethanol	1.719	45	11664	3.708	ppbv #	38
15) Acetone	1.832	43	100928	4.783	ppbv	98
19) Isopropyl Alcohol	1.884	45	9295	0.779	ppbv #	1
20) Methylene Chloride	2.059	84	18675	2.407	ppbv	89
26) Hexane	2.819	57	24551	1.230	ppbv #	44
29) Chloroform	2.836	83	4378m	0.154	ppbv	
34) 2-Butanone	2.551	43	13762	0.500	ppbv	99
35) Carbon Tetrachloride	3.755	117	2686m	0.091	ppbv	
36) Benzene	3.651	78	6717m	0.180	ppbv	
53) Toluene	6.910	91	5747m	0.130	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043310.D
Acq On : 08 Dec 2025 12:40
Operator : SY/MD
Sample : Q3799-01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
AA-1

Quant Time: Dec 09 00:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By : Semsettin Yesilyurt 12/11/2025
Supervised By : Mahesh Dadoda 12/11/2025





#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 832
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1




Tgt Ion: 49 Resp: 151890
Ion Ratio Lower Upper
49 100
128 48.5 23.0 69.0
130 62.9 29.9 89.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025




#2
Dichlorodifluoromethane
Concen: 0.392 ppbv m
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Tgt Ion: 85 Resp: 8408
Ion Ratio Lower Upper
85 100
87 43.2 25.4 38.2#





#3
Chlorodifluoromethane
Concen: 0.237 ppbv m
RT: 1.473 min Scan# 428
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1




Tgt Ion: 51 Resp: 4994
Ion Ratio Lower Upper
51 100
67 20.5 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025




#4
Chloromethane
Concen: 2.178 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40



Tgt Ion: 50 Resp: 17448
Ion Ratio Lower Upper
50 100
52 33.1 24.1 36.1





#9

Propene

Concen: 0.649 ppbv

RT: 1.483 min Scan# 41

Delta R.T. 0.003 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40

Instrument :

MSVOA_L

ClientSampleId :

AA-1

Tgt Ion: 41 Resp: 5800

Ion Ratio Lower Upper

41 100





42 64.4 3.7 5.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2025

Supervised By :Mahesh Dadoda 12/11/2025



#11

Trichlorofluoromethane

Concen: 0.214 ppbv

RT: 1.874 min Scan# 552

Delta R.T. -0.000 min

Lab File: VL043310.D




Acq: 08 Dec 2025 12:40


Tgt Ion:101 Resp: 4433

Ion Ratio Lower Upper

101 100

103 70.0 52.6 79.0









#13
Ethanol
Concen: 3.708 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Instrument :
MSVOA_L
ClientSampleId :
AA-1

Tgt Ion: 45 Resp: 11664
Ion Ratio Lower Upper
45 100
46 0.0 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025



#15
Acetone
Concen: 4.783 ppbv
RT: 1.832 min Scan# 539
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion: 43 Resp: 100928
Ion Ratio Lower Upper
43 100
58 26.8 20.7 31.1









#19
Isopropyl Alcohol
Concen: 0.779 ppbv
RT: 1.884 min Scan# 510
Delta R.T. 0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Instrument :
MSVOA_L
ClientSampleId :
AA-1

Tgt Ion: 45 Resp: 929
Ion Ratio Lower Upper
45 100
58 291.0 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025



#20
Methylene Chloride
Concen: 2.407 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40


Tgt Ion: 84 Resp: 18675
Ion Ratio Lower Upper
84 100
49 146.1 132.9 199.3
51 39.7 34.8 52.2
86 58.4 48.2 72.2





#26
Hexane
Concen: 1.230 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1




Tgt Ion: 57 Resp: 2455
Ion Ratio Lower Upper
57 100
41 101.6 74.6 111.8
43 83.3 179.0 268.4
56 62.7 42.3 63.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025




#29
Chloroform
Concen: 0.154 ppbv m
RT: 2.836 min Scan# 849
Delta R.T. -0.007 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Tgt Ion: 83 Resp: 4378
Ion Ratio Lower Upper
83 100
85 57.3 50.5 75.7





#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.946 min Scan# 1193
Delta R.T. -0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1



Tgt Ion:114 Resp: 38999
Ion Ratio Lower Upper
114 100
63 25.9 21.0 31.6
88 18.0 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025



#34
2-Butanone
Concen: 0.500 ppbv
RT: 2.551 min Scan# 761
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40


Tgt Ion: 43 Resp: 13762
Ion Ratio Lower Upper
43 100
72 21.9 18.1 27.1





#35
Carbon Tetrachloride
Concen: 0.091 ppbv m
RT: 3.755 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1




Tgt Ion: 117 Resp: 2680
Ion Ratio Lower Upper
117 100
119 93.5 74.6 112.0
121 28.5 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025




#36
Benzene
Concen: 0.180 ppbv m
RT: 3.651 min Scan# 1101
Delta R.T. 0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Tgt Ion: 78 Resp: 6717
Ion Ratio Lower Upper
78 100
77 21.0 17.5 26.3
50 16.0 15.5 23.3





#53
Toluene
Concen: 0.130 ppbv m
RT: 6.910 min Scan# 2111
Delta R.T. -0.010 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1




Tgt Ion: 91 Resp: 574
Ion Ratio Lower Upper
91 100
92 55.9 47.4 71.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025




#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.869 min Scan# 2713
Delta R.T. -0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Tgt Ion:117 Resp: 311445
Ion Ratio Lower Upper
117 100
82 66.3 56.1 84.1
119 32.4 25.0 37.6





#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.227 ppbv
 RT: 10.364 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: VL043310.D
 Acq: 08 Dec 2025 12:40



Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1



Tgt Ion: 95 Resp: 23348
 Ion Ratio Lower Upper
 95 100
 174 66.4 50.2 75.2
 175 5.1 4.1 6.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2025
 Supervised By :Mahesh Dadoda 12/11/2025



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 9B-IA-1
Lab Sample ID: Q3799-02
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.40	1.98	J	1	0.54	2.47	12/08/25 13:52	VL120825
74-87-3	Chloromethane	0.55	1.14		1	0.21	1.03	12/08/25 13:52	VL120825
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 13:52	VL120825
74-83-9	Bromomethane	0.080	0.31	U	1	0.31	1.94	12/08/25 13:52	VL120825
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	12/08/25 13:52	VL120825
109-99-9	Tetrahydrofuran	0.080	0.24	U	1	0.24	1.47	12/08/25 13:52	VL120825
75-69-4	Trichlorofluoromethane	0.20	1.12	J	1	0.96	2.81	12/08/25 13:52	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1	1.07	3.83	12/08/25 13:52	VL120825
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1	1.05	3.49	12/08/25 13:52	VL120825
75-65-0	tert-Butyl alcohol	0.16	0.49	U	1	0.49	1.52	12/08/25 13:52	VL120825
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	12/08/25 13:52	VL120825
75-35-4	1,1-Dichloroethene	0.15	0.59	U	1	0.59	1.98	12/08/25 13:52	VL120825
67-64-1	Acetone	8.80	20.9		1	0.43	1.19	12/08/25 13:52	VL120825
75-15-0	Carbon Disulfide	0.080	0.25	U	1	0.25	1.56	12/08/25 13:52	VL120825
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	1	0.83	1.80	12/08/25 13:52	VL120825
75-09-2	Methylene Chloride	0.48	1.67	J	1	0.80	1.74	12/08/25 13:52	VL120825
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	1	0.48	1.98	12/08/25 13:52	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 13:52	VL120825
110-82-7	Cyclohexane	0.22	0.76	U	1	0.76	1.72	12/08/25 13:52	VL120825
78-93-3	2-Butanone	0.30	0.88	J	1	0.18	1.47	12/08/25 13:52	VL120825
56-23-5	Carbon Tetrachloride	0.10	0.63		1	0.19	0.19	12/08/25 13:52	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 13:52	VL120825
67-66-3	Chloroform	0.14	0.68	J	1	0.49	2.44	12/08/25 13:52	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 13:52	VL120825
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	1	0.65	2.34	12/08/25 13:52	VL120825
71-43-2	Benzene	0.15	0.48	J	1	0.26	1.60	12/08/25 13:52	VL120825
107-06-2	1,2-Dichloroethane	0.090	0.36	U	1	0.36	2.02	12/08/25 13:52	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 13:52	VL120825
78-87-5	1,2-Dichloropropane	0.13	0.60	U	1	0.60	2.31	12/08/25 13:52	VL120825
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	12/08/25 13:52	VL120825
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	1	0.41	2.05	12/08/25 13:52	VL120825
108-88-3	Toluene	0.16	0.60	U	1	0.60	1.88	12/08/25 13:52	VL120825
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	1	0.68	2.27	12/08/25 13:52	VL120825
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	1	0.50	2.27	12/08/25 13:52	VL120825
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	1	0.44	2.73	12/08/25 13:52	VL120825
124-48-1	Dibromochloromethane	0.090	0.77	U	1	0.77	4.26	12/08/25 13:52	VL120825
106-93-4	1,2-Dibromoethane	0.090	0.69	U	1	0.69	0.77	12/08/25 13:52	VL120825
127-18-4	Tetrachloroethene	0.020	0.14	U	1	0.14	0.20	12/08/25 13:52	VL120825
108-90-7	Chlorobenzene	0.080	0.37	U	1	0.37	2.30	12/08/25 13:52	VL120825
100-41-4	Ethyl Benzene	0.19	0.83	U	1	0.83	2.17	12/08/25 13:52	VL120825
179601-23-1	m/p-Xylene	0.41	1.78	U	1	1.78	4.34	12/08/25 13:52	VL120825
95-47-6	o-Xylene	0.21	0.91	U	1	0.91	2.17	12/08/25 13:52	VL120825
100-42-5	Styrene	0.16	0.68	U	1	0.68	2.13	12/08/25 13:52	VL120825
75-25-2	Bromoform	0.050	0.52	U	1	0.52	5.17	12/08/25 13:52	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	1	0.14	0.21	12/08/25 13:52	VL120825
95-49-8	2-Chlorotoluene	0.17	0.88	U	1	0.88	2.59	12/08/25 13:52	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 9B-IA-1
Lab Sample ID: Q3799-02
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 13:52	VL120825
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 13:52	VL120825
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	1	0.30	3.01	12/08/25 13:52	VL120825
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 13:52	VL120825
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 13:52	VL120825
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1	1.56	3.71	12/08/25 13:52	VL120825
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1	1.39	5.33	12/08/25 13:52	VL120825
106-99-0	1,3-Butadiene	0.050	0.11	U	1	0.11	1.11	12/08/25 13:52	VL120825
91-20-3	Naphthalene	0.010	0.050	U	1	0.050	0.52	12/08/25 13:52	VL120825
622-96-8	4-Ethyltoluene	0.21	1.03	U	1	1.03	2.46	12/08/25 13:52	VL120825
110-54-3	Hexane	0.88	3.10		1	0.56	1.76	12/08/25 13:52	VL120825
107-05-1	Allyl Chloride	0.11	0.34	U	1	0.34	1.57	12/08/25 13:52	VL120825
123-91-1	1,4-Dioxane	0.22	0.79	U	1	0.79	1.80	12/08/25 13:52	VL120825
80-62-6	Methyl Methacrylate	0.14	0.57	U	1	0.57	2.05	12/08/25 13:52	VL120825

SURROGATES

460-00-4 1-Bromo-4-Fluorobenzene 10.3 70 (65) - 130 (135) 103% SPK: 10

INTERNAL STANDARDS

Area Count
74-97-5 Bromochloromethane 137000
540-36-3 1,4-Difluorobenzene 358000
3114-55-4 Chlorobenzene-d5 282000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043312.D
Acq On : 08 Dec 2025 13:52
Operator : SY/MD
Sample : Q3799-02
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:11:27 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	137082	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	357699	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	281919	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	213574	10.335	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 103.400%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.499	85	7699	0.398	ppbv	99
3) Chlorodifluoromethane	1.476	51	4363m	0.229	ppbv	
4) Chloromethane	1.531	50	3969	0.549	ppbv	92
9) Propene	1.486	41	4514	0.559	ppbv #	1
11) Trichlorofluoromethane	1.877	101	3724	0.199	ppbv	89
13) Ethanol	1.722	45	50530	17.797	ppbv	99
15) Acetone	1.835	43	166746	8.755	ppbv	99
19) Isopropyl Alcohol	1.884	45	109957	10.217	ppbv	99
20) Methylene Chloride	2.059	84	6881	0.479	ppbv	83
26) Hexane	2.826	57	15912	0.884	ppbv #	52
29) Chloroform	2.842	83	3468m	0.135	ppbv	
34) 2-Butanone	2.560	43	7477	0.296	ppbv	94
35) Carbon Tetrachloride	3.755	117	2649m	0.098	ppbv	
36) Benzene	3.648	78	5022	0.147	ppbv #	77
53) Toluene	6.923	91	4569m	0.113	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043312.D
Acq On : 08 Dec 2025 13:52
Operator : SY/MD
Sample : Q3799-02
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1

Quant Time: Dec 09 00:11:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/09/2025
Supervised By : Mahesh Dadoda 12/11/2025









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1

Tgt Ion: 49 Resp: 13708
Ion Ratio Lower Upper
49 100
128 49.8 23.0 69.0
130 63.2 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#2
Dichlorodifluoromethane
Concen: 0.398 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 85 Resp: 7699
Ion Ratio Lower Upper
85 100
87 31.3 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.229 ppbv m
RT: 1.476 min Scan# 428
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1

Tgt Ion: 51 Resp: 436
Ion Ratio Lower Upper
51 100
67 19.3 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#4
Chloromethane
Concen: 0.549 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 50 Resp: 3969
Ion Ratio Lower Upper
50 100
52 34.3 24.1 36.1





#9

Propene

Concen: 0.559 ppbv

RT: 1.486 min Scan# 41

Delta R.T. 0.006 min

Lab File: VL043312.D

Acq: 08 Dec 2025 13:52

Instrument :

MSVOA_L

ClientSampleId :





9B-IA-1

Tgt Ion: 41 Resp: 4514
Ion Ratio Lower Upper
41 100
42 61.9 3.7 5.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#11

Trichlorofluoromethane

Concen: 0.199 ppbv




RT: 1.877 min Scan# 553


Delta R.T. 0.003 min

Lab File: VL043312.D

Acq: 08 Dec 2025 13:52

Tgt Ion:101 Resp: 3724
Ion Ratio Lower Upper
101 100
103 74.7 52.6 79.0









#13
Ethanol
Concen: 17.797 ppbv
RT: 1.722 min Scan# 505
Delta R.T. 0.006 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1

Tgt Ion: 45 Resp: 50530
Ion Ratio Lower Upper
45 100
46 38.1 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#15
Acetone
Concen: 8.755 ppbv
RT: 1.835 min Scan# 540
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 43 Resp: 166746
Ion Ratio Lower Upper
43 100
58 26.3 20.7 31.1









#19
Isopropyl Alcohol
Concen: 10.217 ppbv
RT: 1.884 min Scan# 51
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1

Tgt Ion: 45 Resp: 10995
Ion Ratio Lower Upper
45 100
58 39.9 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#20
Methylene Chloride
Concen: 0.479 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 84 Resp: 6881
Ion Ratio Lower Upper
84 100
49 136.6 132.9 199.3
51 40.7 34.8 52.2
86 50.3 48.2 72.2









#26
Hexane
Concen: 0.884 ppbv
RT: 2.826 min Scan# 844
Delta R.T. 0.006 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1

Tgt Ion: 57 Resp: 15912
Ion Ratio Lower Upper
57 100
41 100.0 74.6 111.8
43 104.4 179.0 268.4
56 60.9 42.3 63.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#29
Chloroform
Concen: 0.135 ppbv m
RT: 2.842 min Scan# 851
Delta R.T. -0.000 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52


Tgt Ion: 83 Resp: 3468
Ion Ratio Lower Upper
83 100
85 59.5 50.5 75.7





#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1



Tgt Ion:114 Resp: 35769
Ion Ratio Lower Upper
114 100
63 24.7 21.0 31.6
88 18.3 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#34
2-Butanone
Concen: 0.296 ppbv
RT: 2.560 min Scan# 764
Delta R.T. 0.009 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52


Tgt Ion: 43 Resp: 7477
Ion Ratio Lower Upper
43 100
72 25.3 18.1 27.1





#35
Carbon Tetrachloride
Concen: 0.098 ppbv m
RT: 3.755 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1



Tgt Ion:117 Resp: 2649
Ion Ratio Lower Upper
117 100
119 96.0 74.6 112.0
121 31.2 23.2 34.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#36
Benzene
Concen: 0.147 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52


Tgt Ion: 78 Resp: 5022
Ion Ratio Lower Upper
78 100
77 32.5 17.5 26.3#
50 8.7 15.5 23.3#





#53
Toluene
Concen: 0.113 ppbv m
RT: 6.923 min Scan# 2111
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1




Tgt Ion: 91 Resp: 4569
Ion Ratio Lower Upper
91 100
92 58.1 47.4 71.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.875 min Scan# 2715
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Tgt Ion:117 Resp: 281919
Ion Ratio Lower Upper
117 100
82 64.8 56.1 84.1
119 32.1 25.0 37.6





#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.335 ppbv
 RT: 10.370 min Scan# 31
 Delta R.T. 0.003 min
 Lab File: VL043312.D
 Acq: 08 Dec 2025 13:52



Instrument :
 MSVOA_L
 ClientSampleId :
 9B-IA-1



Tgt Ion: 95 Resp: 213574
 Ion Ratio Lower Upper
 95 100
 174 68.6 50.2 75.2
 175 4.9 4.1 6.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 10-IA-1
Lab Sample ID: Q3799-03
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.41	2.03	J	1	0.54	2.47	12/08/25 14:29	VL120825
74-87-3	Chloromethane	0.46	0.95	J	1	0.21	1.03	12/08/25 14:29	VL120825
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 14:29	VL120825
74-83-9	Bromomethane	0.080	0.31	U	1	0.31	1.94	12/08/25 14:29	VL120825
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	12/08/25 14:29	VL120825
109-99-9	Tetrahydrofuran	0.080	0.24	U	1	0.24	1.47	12/08/25 14:29	VL120825
75-69-4	Trichlorofluoromethane	0.23	1.29	J	1	0.96	2.81	12/08/25 14:29	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1	1.07	3.83	12/08/25 14:29	VL120825
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1	1.05	3.49	12/08/25 14:29	VL120825
75-65-0	tert-Butyl alcohol	0.16	0.49	U	1	0.49	1.52	12/08/25 14:29	VL120825
142-82-5	Heptane	0.17	0.70	J	1	0.70	2.05	12/08/25 14:29	VL120825
75-35-4	1,1-Dichloroethene	0.15	0.59	U	1	0.59	1.98	12/08/25 14:29	VL120825
67-64-1	Acetone	290	689	E	1	0.43	1.19	12/08/25 14:29	VL120825
75-15-0	Carbon Disulfide	0.080	0.25	U	1	0.25	1.56	12/08/25 14:29	VL120825
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	1	0.83	1.80	12/08/25 14:29	VL120825
75-09-2	Methylene Chloride	15.9	55.2	E	1	0.80	1.74	12/08/25 14:29	VL120825
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	1	0.48	1.98	12/08/25 14:29	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 14:29	VL120825
110-82-7	Cyclohexane	0.22	0.76	U	1	0.76	1.72	12/08/25 14:29	VL120825
78-93-3	2-Butanone	0.55	1.62		1	0.18	1.47	12/08/25 14:29	VL120825
56-23-5	Carbon Tetrachloride	0.090	0.57		1	0.19	0.19	12/08/25 14:29	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 14:29	VL120825
67-66-3	Chloroform	0.10	0.49	U	1	0.49	2.44	12/08/25 14:29	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 14:29	VL120825
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	1	0.65	2.34	12/08/25 14:29	VL120825
71-43-2	Benzene	0.16	0.51	J	1	0.26	1.60	12/08/25 14:29	VL120825
107-06-2	1,2-Dichloroethane	0.090	0.36	U	1	0.36	2.02	12/08/25 14:29	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 14:29	VL120825
78-87-5	1,2-Dichloropropane	0.13	0.60	U	1	0.60	2.31	12/08/25 14:29	VL120825
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	12/08/25 14:29	VL120825
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	1	0.41	2.05	12/08/25 14:29	VL120825
108-88-3	Toluene	0.45	1.70	J	1	0.60	1.88	12/08/25 14:29	VL120825
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	1	0.68	2.27	12/08/25 14:29	VL120825
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	1	0.50	2.27	12/08/25 14:29	VL120825
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	1	0.44	2.73	12/08/25 14:29	VL120825
124-48-1	Dibromochloromethane	0.090	0.77	U	1	0.77	4.26	12/08/25 14:29	VL120825
106-93-4	1,2-Dibromoethane	0.090	0.69	U	1	0.69	0.77	12/08/25 14:29	VL120825
127-18-4	Tetrachloroethene	0.040	0.27		1	0.14	0.20	12/08/25 14:29	VL120825
108-90-7	Chlorobenzene	0.080	0.37	U	1	0.37	2.30	12/08/25 14:29	VL120825
100-41-4	Ethyl Benzene	0.19	0.83	U	1	0.83	2.17	12/08/25 14:29	VL120825
179601-23-1	m/p-Xylene	0.41	1.78	U	1	1.78	4.34	12/08/25 14:29	VL120825
95-47-6	o-Xylene	0.21	0.91	U	1	0.91	2.17	12/08/25 14:29	VL120825
100-42-5	Styrene	0.16	0.68	U	1	0.68	2.13	12/08/25 14:29	VL120825
75-25-2	Bromoform	0.050	0.52	U	1	0.52	5.17	12/08/25 14:29	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	1	0.14	0.21	12/08/25 14:29	VL120825
95-49-8	2-Chlorotoluene	0.17	0.88	U	1	0.88	2.59	12/08/25 14:29	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 10-IA-1
Lab Sample ID: Q3799-03
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 14:29	VL120825
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 14:29	VL120825
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	1	0.30	3.01	12/08/25 14:29	VL120825
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 14:29	VL120825
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 14:29	VL120825
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1	1.56	3.71	12/08/25 14:29	VL120825
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1	1.39	5.33	12/08/25 14:29	VL120825
106-99-0	1,3-Butadiene	0.050	0.11	U	1	0.11	1.11	12/08/25 14:29	VL120825
91-20-3	Naphthalene	0.010	0.050	U	1	0.050	0.52	12/08/25 14:29	VL120825
622-96-8	4-Ethyltoluene	0.21	1.03	U	1	1.03	2.46	12/08/25 14:29	VL120825
110-54-3	Hexane	0.16	0.56	U	1	0.56	1.76	12/08/25 14:29	VL120825
107-05-1	Allyl Chloride	0.11	0.34	U	1	0.34	1.57	12/08/25 14:29	VL120825
123-91-1	1,4-Dioxane	0.22	0.79	U	1	0.79	1.80	12/08/25 14:29	VL120825
80-62-6	Methyl Methacrylate	65.2	267	E	1	0.57	2.05	12/08/25 14:29	VL120825

SURROGATES

460-00-4	1-Bromo-4-Fluorobenzene	10.4		70 (65) - 130 (135)	104% SPK: 10
----------	-------------------------	------	--	---------------------	--------------

INTERNAL STANDARDS

	Area Count	
74-97-5	Bromochloromethane	142000
540-36-3	1,4-Difluorobenzene	362000
3114-55-4	Chlorobenzene-d5	300000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043313.D
 Acq On : 08 Dec 2025 14:29
 Operator : SY/MD
 Sample : Q3799-03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 10-IA-1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:11:53 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.784	49	142316	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	361997	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	299956	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	227860	10.363	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.600%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.495	85	8224	0.410	ppbv	96
3) Chlorodifluoromethane	1.476	51	4267m	0.216	ppbv	
4) Chloromethane	1.531	50	3421	0.456	ppbv	95
9) Propene	1.482	41	14885	1.776	ppbv #	86
10) Heptane	4.962	43	4267m	0.171	ppbv	
11) Trichlorofluoromethane	1.874	101	4374	0.225	ppbv	97
13) Ethanol	1.722	45	357639	121.331	ppbv	98
15) Acetone	1.842	43	5690107	287.779	ppbv #	17
19) Isopropyl Alcohol	1.897	45	4546385	406.914	ppbv #	36
20) Methylene Chloride	2.062	84	89672	15.850	ppbv	92
25) Ethyl Acetate	2.832	43	841516m	19.733	ppbv	
34) 2-Butanone	2.557	43	14010	0.548	ppbv	99
35) Carbon Tetrachloride	3.761	117	2513	0.091	ppbv #	84
36) Benzene	3.654	78	5665	0.163	ppbv	98
43) Methyl Methacrylate	4.871	69	1199833	65.204	ppbv	96
52) Tetrachloroethene	8.218	164	477m	0.037	ppbv	
53) Toluene	6.920	91	18366	0.447	ppbv	100
59) m/p-Xylene	9.526	91	5545m	0.120	ppbv	
61) Styrene	9.862	104	2236m	0.110	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043313.D
Acq On : 08 Dec 2025 14:29
Operator : SY/MD
Sample : Q3799-03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
10-IA-1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:11:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1

Tgt Ion: 49 Resp: 142310
Ion Ratio Lower Upper
49 100
128 49.2 23.0 69.0
130 63.6 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#2
Dichlorodifluoromethane
Concen: 0.410 ppbv
RT: 1.495 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29


Tgt Ion: 85 Resp: 8224
Ion Ratio Lower Upper
85 100
87 34.0 25.4 38.2





#3
Chlorodifluoromethane
Concen: 0.216 ppbv m
RT: 1.476 min Scan# 428
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1




Tgt Ion: 51 Resp: 426
Ion Ratio Lower Upper
51 100
67 23.6 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#4
Chloromethane
Concen: 0.456 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29



Tgt Ion: 50 Resp: 3421
Ion Ratio Lower Upper
50 100
52 27.5 24.1 36.1









#9
Propene
Concen: 1.776 ppbv
RT: 1.482 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1

Tgt Ion: 41 Resp: 1488
Ion Ratio Lower Upper
41 100
42 0.0 3.7 5.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#10
Heptane
Concen: 0.171 ppbv m
RT: 4.962 min Scan# 1506
Delta R.T. -0.004 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29


Tgt Ion: 43 Resp: 4267
Ion Ratio Lower Upper
43 100
57 19.5 39.7 59.5





#11
Trichlorofluoromethane
Concen: 0.225 ppbv
RT: 1.874 min Scan# 501
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1




Tgt Ion:101 Resp: 4374
Ion Ratio Lower Upper
101 100
103 63.7 52.6 79.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#13
Ethanol
Concen: 121.331 ppbv
RT: 1.722 min Scan# 505
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Tgt Ion: 45 Resp: 357639
Ion Ratio Lower Upper
45 100
46 38.4 15.0 59.8





#15
Acetone
Concen: 287.779 ppbv
RT: 1.842 min Scan# 54
Delta R.T. 0.009 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1




Tgt Ion: 43 Resp: 569010
Ion Ratio Lower Upper
43 100
58 68.0 20.7 31.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#19
Isopropyl Alcohol
Concen: 406.914 ppbv
RT: 1.897 min Scan# 559
Delta R.T. 0.016 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Tgt Ion: 45 Resp: 4546385
Ion Ratio Lower Upper
45 100
58 0.0 31.3 46.9





#20
Methylene Chloride
Concen: 15.850 ppbv
RT: 2.062 min Scan# 610
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1



Tgt Ion:	84	Resp:	89672
Ion	Ratio	Lower	Upper
84	100		
49	154.0	132.9	199.3
51	47.3	34.8	52.2
86	66.4	48.2	72.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#25
Ethyl Acetate
Concen: 19.733 ppbv m
RT: 2.832 min Scan# 848
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29


Tgt Ion:	43	Resp:	841516
Ion	Ratio	Lower	Upper
43	100		
45	0.0	8.1	12.1#
61	0.0	7.0	10.4#
70	0.0	7.0	10.6#





#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.955 min Scan# 1193
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1




Tgt Ion:114 Resp: 36199
Ion Ratio Lower Upper
114 100
63 25.6 21.0 31.6
88 18.1 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#34
2-Butanone
Concen: 0.548 ppbv
RT: 2.557 min Scan# 763
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Tgt Ion: 43 Resp: 14010
Ion Ratio Lower Upper
43 100
72 22.1 18.1 27.1





#35
Carbon Tetrachloride
Concen: 0.091 ppbv
RT: 3.761 min Scan# 1135
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1




Tgt Ion: 117 Resp: 251
Ion Ratio Lower Upper
117 100
119 74.3 74.6 112.0
121 31.5 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#36
Benzene
Concen: 0.163 ppbv
RT: 3.654 min Scan# 1102
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Tgt Ion: 78 Resp: 5665
Ion Ratio Lower Upper
78 100
77 22.2 17.5 26.3
50 18.3 15.5 23.3





#43
Methyl Methacrylate
Concen: 65.204 ppbv
RT: 4.871 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1



Tgt Ion: 69 Resp: 119983
Ion Ratio Lower Upper
69 100
41 151.4 126.7 190.1
100 36.1 28.6 43.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#52
Tetrachloroethene
Concen: 0.037 ppbv m
RT: 8.218 min Scan# 2512
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion:164 Resp: 477
Ion Ratio Lower Upper
164 100
166 86.3 99.4 149.2#
129 68.5 81.8 122.8#
131 48.0 77.4 116.2#









#53
Toluene
Concen: 0.447 ppbv
RT: 6.920 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1

Tgt Ion: 91 Resp: 1836
Ion Ratio Lower Upper
91 100
92 59.4 47.4 71.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 2714
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion:117 Resp: 299956
Ion Ratio Lower Upper
117 100
82 63.9 56.1 84.1
119 32.3 25.0 37.6





#59

m/p-Xylene

Concen: 0.120 ppbv m

RT: 9.526 min Scan# 2916

Delta R.T. -0.017 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29

Instrument :

MSVOA_L





ClientSampleId :

10-IA-1

Tgt Ion: 91 Resp: 5549
 Ion Ratio Lower Upper
 91 100
 106 0.0 36.1 54.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#61

Styrene

Concen: 0.110 ppbv m




RT: 9.862 min Scan# 3020


Delta R.T. -0.000 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29


Tgt Ion:104 Resp: 2236
 Ion Ratio Lower Upper
 104 100
 78 0.0 42.1 63.1#
 103 0.0 37.2 55.8#





#68
1-Bromo-4-Fluorobenzene
Concen: 10.363 ppbv
RT: 10.367 min Scan# 31
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29



Instrument :
MSVOA_L
ClientSampleId :
10-IA-1



Tgt Ion: 95 Resp: 227860
Ion Ratio Lower Upper
95 100
174 69.3 50.2 75.2
175 4.9 4.1 6.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 10-IA-1DL
Lab Sample ID: Q3799-03DL
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	1.10	5.44	UD	10	5.44	24.7	12/08/25 20:50	VL120825
74-87-3	Chloromethane	0.98	2.02	UD	10	2.02	10.3	12/08/25 20:50	VL120825
75-01-4	Vinyl Chloride	0.25	0.64	UD	10	0.64	0.77	12/08/25 20:50	VL120825
74-83-9	Bromomethane	0.77	2.99	UD	10	2.99	19.4	12/08/25 20:50	VL120825
75-00-3	Chloroethane	1.50	3.96	UD	10	3.96	13.2	12/08/25 20:50	VL120825
109-99-9	Tetrahydrofuran	0.80	2.36	UD	10	2.36	14.8	12/08/25 20:50	VL120825
75-69-4	Trichlorofluoromethane	1.70	9.55	UD	10	9.55	28.1	12/08/25 20:50	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	1.40	10.7	UD	10	10.7	38.3	12/08/25 20:50	VL120825
76-14-2	Dichlorotetrafluoroethane	1.50	10.5	UD	10	10.5	35.0	12/08/25 20:50	VL120825
75-65-0	tert-Butyl alcohol	1.60	4.85	UD	10	4.85	15.2	12/08/25 20:50	VL120825
142-82-5	Heptane	1.70	6.97	UD	10	6.97	20.5	12/08/25 20:50	VL120825
75-35-4	1,1-Dichloroethene	1.50	5.95	UD	10	5.95	19.8	12/08/25 20:50	VL120825
67-64-1	Acetone	1500	3560	ED	10	4.28	11.9	12/08/25 20:50	VL120825
75-15-0	Carbon Disulfide	0.80	2.49	UD	10	2.49	15.6	12/08/25 20:50	VL120825
1634-04-4	Methyl tert-Butyl Ether	2.30	8.29	UD	10	8.29	18.0	12/08/25 20:50	VL120825
75-09-2	Methylene Chloride	14.5	50.4	D	10	7.99	17.4	12/08/25 20:50	VL120825
156-60-5	trans-1,2-Dichloroethene	1.20	4.76	UD	10	4.76	19.8	12/08/25 20:50	VL120825
75-34-3	1,1-Dichloroethane	1.30	5.26	UD	10	5.26	20.2	12/08/25 20:50	VL120825
110-82-7	Cyclohexane	2.20	7.57	UD	10	7.57	17.2	12/08/25 20:50	VL120825
78-93-3	2-Butanone	0.56	1.65	UD	10	1.65	14.8	12/08/25 20:50	VL120825
56-23-5	Carbon Tetrachloride	0.25	1.57	UD	10	1.57	1.89	12/08/25 20:50	VL120825
156-59-2	cis-1,2-Dichloroethene	0.99	3.93	UD	10	3.93	19.8	12/08/25 20:50	VL120825
67-66-3	Chloroform	0.96	4.69	UD	10	4.69	24.4	12/08/25 20:50	VL120825
71-55-6	1,1,1-Trichloroethane	0.16	0.87	UD	10	0.87	1.64	12/08/25 20:50	VL120825
540-84-1	2,2,4-Trimethylpentane	1.40	6.54	UD	10	6.54	23.4	12/08/25 20:50	VL120825
71-43-2	Benzene	0.79	2.52	UD	10	2.52	16.0	12/08/25 20:50	VL120825
107-06-2	1,2-Dichloroethane	0.93	3.76	UD	10	3.76	20.2	12/08/25 20:50	VL120825
79-01-6	Trichloroethene	0.24	1.29	UD	10	1.29	1.61	12/08/25 20:50	VL120825
78-87-5	1,2-Dichloropropane	1.30	6.01	UD	10	6.01	23.1	12/08/25 20:50	VL120825
75-27-4	Bromodichloromethane	0.63	4.22	UD	10	4.22	33.5	12/08/25 20:50	VL120825
108-10-1	4-Methyl-2-Pentanone	0.97	3.98	UD	10	3.98	20.5	12/08/25 20:50	VL120825
108-88-3	Toluene	1.60	6.03	UD	10	6.03	18.8	12/08/25 20:50	VL120825
10061-02-6	t-1,3-Dichloropropene	1.50	6.81	UD	10	6.81	22.7	12/08/25 20:50	VL120825
10061-01-5	cis-1,3-Dichloropropene	1.10	4.99	UD	10	4.99	22.7	12/08/25 20:50	VL120825
79-00-5	1,1,2-Trichloroethane	0.81	4.42	UD	10	4.42	27.3	12/08/25 20:50	VL120825
124-48-1	Dibromochloromethane	0.85	7.24	UD	10	7.24	42.6	12/08/25 20:50	VL120825
106-93-4	1,2-Dibromoethane	0.85	6.53	UD	10	6.53	7.69	12/08/25 20:50	VL120825
127-18-4	Tetrachloroethene	0.15	1.02	UD	10	1.02	2.03	12/08/25 20:50	VL120825
108-90-7	Chlorobenzene	0.77	3.55	UD	10	3.55	23.0	12/08/25 20:50	VL120825
100-41-4	Ethyl Benzene	1.90	8.25	UD	10	8.25	21.7	12/08/25 20:50	VL120825
179601-23-1	m/p-Xylene	4.10	17.8	UD	10	17.8	43.4	12/08/25 20:50	VL120825
95-47-6	o-Xylene	2.10	9.12	UD	10	9.12	21.7	12/08/25 20:50	VL120825
100-42-5	Styrene	1.60	6.81	UD	10	6.81	21.3	12/08/25 20:50	VL120825
75-25-2	Bromoform	0.48	4.96	UD	10	4.96	51.7	12/08/25 20:50	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.18	1.24	UD	10	1.24	2.06	12/08/25 20:50	VL120825
95-49-8	2-Chlorotoluene	1.70	8.80	UD	10	8.80	25.9	12/08/25 20:50	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 10-IA-1DL
Lab Sample ID: Q3799-03DL
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	1.80	8.85	UD	10	8.85	24.6	12/08/25 20:50	VL120825
95-63-6	1,2,4-Trimethylbenzene	1.80	8.85	UD	10	8.85	24.6	12/08/25 20:50	VL120825
541-73-1	1,3-Dichlorobenzene	0.54	3.25	UD	10	3.25	30.1	12/08/25 20:50	VL120825
106-46-7	1,4-Dichlorobenzene	1.30	7.82	UD	10	7.82	30.1	12/08/25 20:50	VL120825
95-50-1	1,2-Dichlorobenzene	1.30	7.82	UD	10	7.82	30.1	12/08/25 20:50	VL120825
120-82-1	1,2,4-Trichlorobenzene	2.10	15.6	UD	10	15.6	37.1	12/08/25 20:50	VL120825
87-68-3	Hexachloro-1,3-Butadiene	1.30	13.9	UD	10	13.9	53.3	12/08/25 20:50	VL120825
106-99-0	1,3-Butadiene	0.51	1.13	UD	10	1.13	11.1	12/08/25 20:50	VL120825
91-20-3	Naphthalene	0.13	0.68	UD	10	0.68	5.24	12/08/25 20:50	VL120825
622-96-8	4-Ethyltoluene	2.10	10.3	UD	10	10.3	24.6	12/08/25 20:50	VL120825
110-54-3	Hexane	1.60	5.64	UD	10	5.64	17.6	12/08/25 20:50	VL120825
107-05-1	Allyl Chloride	1.10	3.44	UD	10	3.44	15.7	12/08/25 20:50	VL120825
123-91-1	1,4-Dioxane	2.20	7.93	UD	10	7.93	18.0	12/08/25 20:50	VL120825
80-62-6	Methyl Methacrylate	57.4	235	D	10	5.73	20.5	12/08/25 20:50	VL120825

SURROGATES

460-00-4	1-Bromo-4-Fluorobenzene	9.70		70 (65) - 130 (135)	97% SPK: 10
----------	-------------------------	------	--	---------------------	-------------

INTERNAL STANDARDS

	Area Count	
74-97-5	Bromochloromethane	129000
540-36-3	1,4-Difluorobenzene	324000
3114-55-4	Chlorobenzene-d5	262000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043322.D
Acq On : 08 Dec 2025 20:50
Operator : SY/MD
Sample : Q3799-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL

Quant Time: Dec 09 00:15:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Bromochloromethane	2.778	49	129360	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.943	114	324163	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.866	117	261825	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.361	95	186988	9.743	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.400%
Target Compounds						
					Qvalue	
13) Ethanol	1.716	45	34680	12.944	ppbv	100
15) Acetone	1.829	43	2615938	145.553	ppbv #	66
19) Isopropyl Alcohol	1.878	45	1280827	126.119	ppbv #	20
20) Methylene Chloride	2.056	84	11227	1.449	ppbv	86
25) Ethyl Acetate	2.829	43	67451	1.740	ppbv #	92
43) Methyl Methacrylate	4.859	69	94519	5.736	ppbv	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043322.D
Acq On : 08 Dec 2025 20:50
Operator : SY/MD
Sample : Q3799-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL

Quant Time: Dec 09 00:15:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.778 min Scan# 81
Delta R.T. -0.003 min
Lab File: VL043322.D
Acq: 08 Dec 2025 20:50




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL


Tgt Ion: 49 Resp: 129360
Ion Ratio Lower Upper
49 100
128 49.5 23.0 69.0
130 63.4 29.9 89.7



#13
Ethanol
Concen: 12.944 ppbv
RT: 1.716 min Scan# 503
Delta R.T. 0.000 min
Lab File: VL043322.D
Acq: 08 Dec 2025 20:50

Tgt Ion: 45 Resp: 34680
Ion Ratio Lower Upper
45 100
46 37.5 15.0 59.8









#15
Acetone
Concen: 145.553 ppbv
RT: 1.829 min Scan# 51
Delta R.T. -0.003 min
Lab File: VL043322.D
Acq: 08 Dec 2025 20:50




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL


Tgt Ion: 43 Resp: 2615938
Ion Ratio Lower Upper
43 100
58 43.0 20.7 31.1#



#19
Isopropyl Alcohol
Concen: 126.119 ppbv
RT: 1.878 min Scan# 553
Delta R.T. -0.003 min
Lab File: VL043322.D
Acq: 08 Dec 2025 20:50

Tgt Ion: 45 Resp: 1280827
Ion Ratio Lower Upper
45 100
58 87.9 31.3 46.9#









#20
Methylene Chloride
Concen: 1.449 ppbv
RT: 2.056 min Scan# 608
Delta R.T. -0.000 min
Lab File: VL043322.D
Acq: 08 Dec 2025 20:50

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL

Tgt Ion: 84 Resp: 11227




Ion	Ratio	Lower	Upper
84	100		
49	139.7	132.9	199.3
51	44.7	34.8	52.2
86	65.6	48.2	72.2




#25
Ethyl Acetate
Concen: 1.740 ppbv
RT: 2.829 min Scan# 847
Delta R.T. 0.000 min
Lab File: VL043322.D
Acq: 08 Dec 2025 20:50

Tgt Ion: 43 Resp: 67451

Ion	Ratio	Lower	Upper
43	100		
45	14.3	8.1	12.1#
61	11.1	7.0	10.4#
70	10.4	7.0	10.6









#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 3.943 min Scan# 1191
 Delta R.T. -0.006 min
 Lab File: VL043322.D
 Acq: 08 Dec 2025 20:50




Instrument :
 MSVOA_1
 ClientSampleId :
 10-IA-1DL


Tgt Ion:114 Resp: 324163
 Ion Ratio Lower Upper
 114 100
 63 25.5 21.0 31.6
 88 17.9 14.9 22.3



#43
 Methyl Methacrylate
 Concen: 5.736 ppbv
 RT: 4.859 min Scan# 1474
 Delta R.T. -0.006 min
 Lab File: VL043322.D
 Acq: 08 Dec 2025 20:50

Tgt Ion: 69 Resp: 94519
 Ion Ratio Lower Upper
 69 100
 41 152.9 126.7 190.1
 100 34.9 28.6 43.0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.866 min Scan# 21
Delta R.T. -0.006 min
Lab File: VL043322.D
Acq: 08 Dec 2025 20:50




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL

Tgt Ion:117 Resp: 261825
Ion Ratio Lower Upper
117 100
82 66.1 56.1 84.1
119 31.8 25.0 37.6



#68
1-Bromo-4-Fluorobenzene
Concen: 9.743 ppbv
RT: 10.361 min Scan# 3174
Delta R.T. -0.006 min
Lab File: VL043322.D
Acq: 08 Dec 2025 20:50

Tgt Ion: 95 Resp: 186988
Ion Ratio Lower Upper
95 100
174 69.2 50.2 75.2
175 4.9 4.1 6.1



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 10-IA-1DL2
Lab Sample ID: Q3799-03DL2
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	44.0	218	UD	400	218	989	12/09/25 07:30	VL120825
74-87-3	Chloromethane	39.2	81.0	UD	400	81.0	413	12/09/25 07:30	VL120825
75-01-4	Vinyl Chloride	10.0	25.6	UD	400	25.6	30.7	12/09/25 07:30	VL120825
74-83-9	Bromomethane	30.8	120	UD	400	120	777	12/09/25 07:30	VL120825
75-00-3	Chloroethane	60.0	158	UD	400	158	528	12/09/25 07:30	VL120825
109-99-9	Tetrahydrofuran	32.0	94.4	UD	400	94.4	590	12/09/25 07:30	VL120825
75-69-4	Trichlorofluoromethane	68.0	382	UD	400	382	1120	12/09/25 07:30	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	56.0	429	UD	400	429	1530	12/09/25 07:30	VL120825
76-14-2	Dichlorotetrafluoroethane	60.0	419	UD	400	419	1400	12/09/25 07:30	VL120825
75-65-0	tert-Butyl alcohol	64.0	194	UD	400	194	606	12/09/25 07:30	VL120825
142-82-5	Heptane	68.0	279	UD	400	279	820	12/09/25 07:30	VL120825
75-35-4	1,1-Dichloroethene	60.0	238	UD	400	238	793	12/09/25 07:30	VL120825
67-64-1	Acetone	3000	7130	D	400	171	475	12/09/25 07:30	VL120825
75-15-0	Carbon Disulfide	32.0	99.7	UD	400	99.7	623	12/09/25 07:30	VL120825
1634-04-4	Methyl tert-Butyl Ether	92.0	332	UD	400	332	721	12/09/25 07:30	VL120825
75-09-2	Methylene Chloride	92.0	320	UD	400	320	695	12/09/25 07:30	VL120825
156-60-5	trans-1,2-Dichloroethene	48.0	190	UD	400	190	793	12/09/25 07:30	VL120825
75-34-3	1,1-Dichloroethane	52.0	210	UD	400	210	809	12/09/25 07:30	VL120825
110-82-7	Cyclohexane	88.0	303	UD	400	303	688	12/09/25 07:30	VL120825
78-93-3	2-Butanone	22.4	66.1	UD	400	66.1	590	12/09/25 07:30	VL120825
56-23-5	Carbon Tetrachloride	10.0	62.9	UD	400	62.9	75.5	12/09/25 07:30	VL120825
156-59-2	cis-1,2-Dichloroethene	39.6	157	UD	400	157	793	12/09/25 07:30	VL120825
67-66-3	Chloroform	38.4	188	UD	400	188	977	12/09/25 07:30	VL120825
71-55-6	1,1,1-Trichloroethane	6.40	34.9	UD	400	34.9	65.5	12/09/25 07:30	VL120825
540-84-1	2,2,4-Trimethylpentane	56.0	262	UD	400	262	934	12/09/25 07:30	VL120825
71-43-2	Benzene	31.6	101	UD	400	101	639	12/09/25 07:30	VL120825
107-06-2	1,2-Dichloroethane	37.2	151	UD	400	151	809	12/09/25 07:30	VL120825
79-01-6	Trichloroethene	9.60	51.6	UD	400	51.6	64.5	12/09/25 07:30	VL120825
78-87-5	1,2-Dichloropropane	52.0	240	UD	400	240	924	12/09/25 07:30	VL120825
75-27-4	Bromodichloromethane	25.2	169	UD	400	169	1340	12/09/25 07:30	VL120825
108-10-1	4-Methyl-2-Pentanone	38.8	159	UD	400	159	820	12/09/25 07:30	VL120825
108-88-3	Toluene	64.0	241	UD	400	241	754	12/09/25 07:30	VL120825
10061-02-6	t-1,3-Dichloropropene	60.0	272	UD	400	272	908	12/09/25 07:30	VL120825
10061-01-5	cis-1,3-Dichloropropene	44.0	200	UD	400	200	908	12/09/25 07:30	VL120825
79-00-5	1,1,2-Trichloroethane	32.4	177	UD	400	177	1090	12/09/25 07:30	VL120825
124-48-1	Dibromochloromethane	34.0	290	UD	400	290	1700	12/09/25 07:30	VL120825
106-93-4	1,2-Dibromoethane	34.0	261	UD	400	261	307	12/09/25 07:30	VL120825
127-18-4	Tetrachloroethene	6.00	40.7	UD	400	40.7	81.4	12/09/25 07:30	VL120825
108-90-7	Chlorobenzene	30.8	142	UD	400	142	921	12/09/25 07:30	VL120825
100-41-4	Ethyl Benzene	76.0	330	UD	400	330	869	12/09/25 07:30	VL120825
179601-23-1	m/p-Xylene	160	695	UD	400	695	1740	12/09/25 07:30	VL120825
95-47-6	o-Xylene	84.0	365	UD	400	365	869	12/09/25 07:30	VL120825
100-42-5	Styrene	64.0	272	UD	400	272	852	12/09/25 07:30	VL120825
75-25-2	Bromoform	19.2	199	UD	400	199	2070	12/09/25 07:30	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	7.20	49.4	UD	400	49.4	82.4	12/09/25 07:30	VL120825
95-49-8	2-Chlorotoluene	68.0	352	UD	400	352	1040	12/09/25 07:30	VL120825

Report of Analysis

Client:	RTP Environmental	Date Collected:	12/03/25
Project:	CHADWICK SQUARE	Date Received:	12/04/25
Client Sample ID:	10-IA-1DL2	SDG No.:	Q3799
Lab Sample ID:	Q3799-03DL2	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 mL		

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	72.0	354	UD	400	354	983	12/09/25 07:30	VL120825
95-63-6	1,2,4-Trimethylbenzene	72.0	354	UD	400	354	983	12/09/25 07:30	VL120825
541-73-1	1,3-Dichlorobenzene	21.6	130	UD	400	130	1200	12/09/25 07:30	VL120825
106-46-7	1,4-Dichlorobenzene	52.0	313	UD	400	313	1200	12/09/25 07:30	VL120825
95-50-1	1,2-Dichlorobenzene	52.0	313	UD	400	313	1200	12/09/25 07:30	VL120825
120-82-1	1,2,4-Trichlorobenzene	84.0	624	UD	400	624	1480	12/09/25 07:30	VL120825
87-68-3	Hexachloro-1,3-Butadiene	52.0	555	UD	400	555	2130	12/09/25 07:30	VL120825
106-99-0	1,3-Butadiene	20.4	45.1	UD	400	45.1	442	12/09/25 07:30	VL120825
91-20-3	Naphthalene	5.20	27.3	UD	400	27.3	210	12/09/25 07:30	VL120825
622-96-8	4-Ethyltoluene	84.0	413	UD	400	413	983	12/09/25 07:30	VL120825
110-54-3	Hexane	64.0	226	UD	400	226	705	12/09/25 07:30	VL120825
107-05-1	Allyl Chloride	44.0	138	UD	400	138	626	12/09/25 07:30	VL120825
123-91-1	1,4-Dioxane	88.0	317	UD	400	317	721	12/09/25 07:30	VL120825
80-62-6	Methyl Methacrylate	56.0	229	UD	400	229	819	12/09/25 07:30	VL120825

SURROGATES

460-00-4	1-Bromo-4-Fluorobenzene	9.50		70 (65) - 130 (135)	95% SPK: 10
----------	-------------------------	------	--	---------------------	-------------

INTERNAL STANDARDS

	Area Count	
74-97-5	Bromochloromethane	122000
540-36-3	1,4-Difluorobenzene	307000
3114-55-4	Chlorobenzene-d5	246000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043331.D
Acq On : 09 Dec 2025 07:30
Operator : SY/MD
Sample : Q3799-03DL2 400X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL2

Quant Time: Dec 10 02:12:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.780	49	122404	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	307022	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	245820	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	171652	9.526	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.300%
Target Compounds						
13) Ethanol	1.719	45	876m	0.346	ppbv	Qvalue
15) Acetone	1.832	43	127495	7.497	ppbv	100
19) Isopropyl Alcohol	1.884	45	32181	3.349	ppbv #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043331.D
Acq On : 09 Dec 2025 07:30
Operator : SY/MD
Sample : Q3799-03DL2 400X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL2

Quant Time: Dec 10 02:12:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By : Semsettin Yesilyurt 12/11/2025
Supervised By : Mahesh Dadoda 12/11/2025





#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.780 min Scan# 81
Delta R.T. -0.001 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL2




Tgt Ion: 49 Resp: 122404
Ion Ratio Lower Upper
49 100
128 49.5 23.0 69.0
130 64.0 29.9 89.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025




#13
Ethanol
Concen: 0.346 ppbv m
RT: 1.719 min Scan# 504
Delta R.T. 0.003 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30



Tgt Ion: 45 Resp: 876
Ion Ratio Lower Upper
45 100
46 0.0 15.0 59.8#









#15
Acetone
Concen: 7.497 ppbv
RT: 1.832 min Scan# 511
Delta R.T. -0.000 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL2

Tgt Ion: 43 Resp: 127495
Ion Ratio Lower Upper
43 100
58 25.7 20.7 31.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025



#19
Isopropyl Alcohol
Concen: 3.349 ppbv
RT: 1.884 min Scan# 555
Delta R.T. 0.003 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30

Tgt Ion: 45 Resp: 32181
Ion Ratio Lower Upper
45 100
58 101.7 31.3 46.9#









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.949 min Scan# 1193
Delta R.T. -0.000 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL2

Tgt Ion:114 Resp: 30702
Ion Ratio Lower Upper
114 100
63 25.7 21.0 31.6
88 17.7 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025



#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.869 min Scan# 2713
Delta R.T. -0.004 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30


Tgt Ion:117 Resp: 245820
Ion Ratio Lower Upper
117 100
82 63.9 56.1 84.1
119 32.4 25.0 37.6





#68
1-Bromo-4-Fluorobenzene
Concen: 9.526 ppbv
RT: 10.367 min Scan# 31
Delta R.T. -0.000 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30



Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL2



Tgt Ion: 95 Resp: 171652
Ion Ratio Lower Upper
95 100
174 69.1 50.2 75.2
175 5.0 4.1 6.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2025
Supervised By :Mahesh Dadoda 12/11/2025



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 11-IA-1
Lab Sample ID: Q3799-04
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.40	1.98	J	1	0.54	2.47	12/08/25 15:37	VL120825
74-87-3	Chloromethane	0.50	1.03		1	0.21	1.03	12/08/25 15:37	VL120825
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 15:37	VL120825
74-83-9	Bromomethane	0.080	0.31	U	1	0.31	1.94	12/08/25 15:37	VL120825
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	12/08/25 15:37	VL120825
109-99-9	Tetrahydrofuran	0.080	0.24	U	1	0.24	1.47	12/08/25 15:37	VL120825
75-69-4	Trichlorofluoromethane	0.23	1.29	J	1	0.96	2.81	12/08/25 15:37	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1	1.07	3.83	12/08/25 15:37	VL120825
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1	1.05	3.49	12/08/25 15:37	VL120825
75-65-0	tert-Butyl alcohol	0.16	0.49	U	1	0.49	1.52	12/08/25 15:37	VL120825
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	12/08/25 15:37	VL120825
75-35-4	1,1-Dichloroethene	0.15	0.59	U	1	0.59	1.98	12/08/25 15:37	VL120825
67-64-1	Acetone	8.00	19.0		1	0.43	1.19	12/08/25 15:37	VL120825
75-15-0	Carbon Disulfide	0.080	0.25	U	1	0.25	1.56	12/08/25 15:37	VL120825
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	1	0.83	1.80	12/08/25 15:37	VL120825
75-09-2	Methylene Chloride	0.40	1.39	J	1	0.80	1.74	12/08/25 15:37	VL120825
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	1	0.48	1.98	12/08/25 15:37	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 15:37	VL120825
110-82-7	Cyclohexane	0.22	0.76	U	1	0.76	1.72	12/08/25 15:37	VL120825
78-93-3	2-Butanone	0.35	1.03	J	1	0.18	1.47	12/08/25 15:37	VL120825
56-23-5	Carbon Tetrachloride	0.090	0.57		1	0.19	0.19	12/08/25 15:37	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 15:37	VL120825
67-66-3	Chloroform	0.10	0.49	U	1	0.49	2.44	12/08/25 15:37	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 15:37	VL120825
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	1	0.65	2.34	12/08/25 15:37	VL120825
71-43-2	Benzene	0.15	0.48	J	1	0.26	1.60	12/08/25 15:37	VL120825
107-06-2	1,2-Dichloroethane	0.090	0.36	U	1	0.36	2.02	12/08/25 15:37	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 15:37	VL120825
78-87-5	1,2-Dichloropropane	0.13	0.60	U	1	0.60	2.31	12/08/25 15:37	VL120825
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	12/08/25 15:37	VL120825
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	1	0.41	2.05	12/08/25 15:37	VL120825
108-88-3	Toluene	0.26	0.98	J	1	0.60	1.88	12/08/25 15:37	VL120825
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	1	0.68	2.27	12/08/25 15:37	VL120825
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	1	0.50	2.27	12/08/25 15:37	VL120825
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	1	0.44	2.73	12/08/25 15:37	VL120825
124-48-1	Dibromochloromethane	0.090	0.77	U	1	0.77	4.26	12/08/25 15:37	VL120825
106-93-4	1,2-Dibromoethane	0.090	0.69	U	1	0.69	0.77	12/08/25 15:37	VL120825
127-18-4	Tetrachloroethene	0.090	0.61		1	0.14	0.20	12/08/25 15:37	VL120825
108-90-7	Chlorobenzene	0.080	0.37	U	1	0.37	2.30	12/08/25 15:37	VL120825
100-41-4	Ethyl Benzene	0.19	0.83	U	1	0.83	2.17	12/08/25 15:37	VL120825
179601-23-1	m/p-Xylene	0.41	1.78	U	1	1.78	4.34	12/08/25 15:37	VL120825
95-47-6	o-Xylene	0.21	0.91	U	1	0.91	2.17	12/08/25 15:37	VL120825
100-42-5	Styrene	0.16	0.68	U	1	0.68	2.13	12/08/25 15:37	VL120825
75-25-2	Bromoform	0.050	0.52	U	1	0.52	5.17	12/08/25 15:37	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	1	0.14	0.21	12/08/25 15:37	VL120825
95-49-8	2-Chlorotoluene	0.17	0.88	U	1	0.88	2.59	12/08/25 15:37	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 11-IA-1
Lab Sample ID: Q3799-04
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 15:37	VL120825
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 15:37	VL120825
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	1	0.30	3.01	12/08/25 15:37	VL120825
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 15:37	VL120825
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 15:37	VL120825
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1	1.56	3.71	12/08/25 15:37	VL120825
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1	1.39	5.33	12/08/25 15:37	VL120825
106-99-0	1,3-Butadiene	0.050	0.11	U	1	0.11	1.11	12/08/25 15:37	VL120825
91-20-3	Naphthalene	0.010	0.050	U	1	0.050	0.52	12/08/25 15:37	VL120825
622-96-8	4-Ethyltoluene	0.21	1.03	U	1	1.03	2.46	12/08/25 15:37	VL120825
110-54-3	Hexane	1.20	4.23		1	0.56	1.76	12/08/25 15:37	VL120825
107-05-1	Allyl Chloride	0.11	0.34	U	1	0.34	1.57	12/08/25 15:37	VL120825
123-91-1	1,4-Dioxane	0.22	0.79	U	1	0.79	1.80	12/08/25 15:37	VL120825
80-62-6	Methyl Methacrylate	0.14	0.57	U	1	0.57	2.05	12/08/25 15:37	VL120825

SURROGATES

460-00-4 1-Bromo-4-Fluorobenzene 10.5 70 (65) - 130 (135) 105% SPK: 10

INTERNAL STANDARDS

Area Count
74-97-5 Bromochloromethane 138000
540-36-3 1,4-Difluorobenzene 354000
3114-55-4 Chlorobenzene-d5 280000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043314.D
Acq On : 08 Dec 2025 15:37
Operator : SY/MD
Sample : Q3799-04
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
11-IA-1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:12:22 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.781	49	138499	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	354324	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	279503	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.364	95	214484	10.469	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.700%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.496	85	7849	0.402	ppbv	96
3) Chlorodifluoromethane	1.473	51	6392m	0.333	ppbv	
4) Chloromethane	1.531	50	3636	0.498	ppbv	94
9) Propene	1.483	41	6390	0.783	ppbv #	86
11) Trichlorofluoromethane	1.878	101	4305	0.227	ppbv	98
13) Ethanol	1.722	45	1527543	532.508	ppbv	91
15) Acetone	1.832	43	153023	7.952	ppbv	100
17) tert-Butyl alcohol	2.040	59	3031	0.120	ppbv	100
19) Isopropyl Alcohol	1.884	45	507615	46.685	ppbv #	37
20) Methylene Chloride	2.059	84	6519	0.396	ppbv	88
26) Hexane	2.823	57	22119	1.216	ppbv #	58
34) 2-Butanone	2.554	43	8673	0.347	ppbv #	90
35) Carbon Tetrachloride	3.755	117	2524	0.094	ppbv #	86
36) Benzene	3.648	78	5240	0.154	ppbv	94
52) Tetrachloroethene	8.215	164	1184m	0.094	ppbv	
53) Toluene	6.917	91	10274m	0.256	ppbv	
59) m/p-Xylene	9.519	91	5340m	0.124	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043314.D
Acq On : 08 Dec 2025 15:37
Operator : SY/MD
Sample : Q3799-04
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
11-IA-1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:12:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 832
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Instrument :
MSVOA_L
ClientSampleId :
11-IA-1

Tgt Ion: 49 Resp: 138499
Ion Ratio Lower Upper
49 100
128 49.7 23.0 69.0
130 63.4 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#2
Dichlorodifluoromethane
Concen: 0.402 ppbv
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37


Tgt Ion: 85 Resp: 7849
Ion Ratio Lower Upper
85 100
87 29.4 25.4 38.2





#3
Chlorodifluoromethane
Concen: 0.333 ppbv m
RT: 1.473 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_L
ClientSampleId :
11-IA-1




Tgt Ion: 51 Resp: 6392
Ion Ratio Lower Upper
51 100
67 15.6 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#4
Chloromethane
Concen: 0.498 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37



Tgt Ion: 50 Resp: 3636
Ion Ratio Lower Upper
50 100
52 33.1 24.1 36.1





#9

Propene

Concen: 0.783 ppbv

RT: 1.483 min Scan# 41

Delta R.T. 0.003 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Instrument :

MSVOA_L





ClientSampleId :

11-IA-1

Tgt Ion: 41 Resp: 6390
 Ion Ratio Lower Upper
 41 100
 42 0.0 3.7 5.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#11

Trichlorofluoromethane

Concen: 0.227 ppbv




RT: 1.878 min Scan# 553


Delta R.T. 0.003 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Tgt Ion:101 Resp: 4305
 Ion Ratio Lower Upper
 101 100
 103 67.4 52.6 79.0









#13
Ethanol
Concen: 532.508 ppbv
RT: 1.722 min Scan# 503
Delta R.T. 0.006 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Instrument :
MSVOA_L
ClientSampleId :
11-IA-1

Tgt Ion: 45 Resp: 152754
Ion Ratio Lower Upper
45 100
46 42.8 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#15
Acetone
Concen: 7.952 ppbv
RT: 1.832 min Scan# 539
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 43 Resp: 153023
Ion Ratio Lower Upper
43 100
58 25.9 20.7 31.1









#17
tert-Butyl alcohol
Concen: 0.120 ppbv
RT: 2.040 min Scan# 602
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Instrument :
MSVOA_L
ClientSampleId :
11-IA-1

Tgt Ion: 59 Resp: 303
Ion Ratio Lower Upper
59 100
57 11.1 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#19
Isopropyl Alcohol
Concen: 46.685 ppbv
RT: 1.884 min Scan# 555
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37


Tgt Ion: 45 Resp: 507615
Ion Ratio Lower Upper
45 100
58 0.9 31.3 46.9#





#20
Methylene Chloride
Concen: 0.396 ppbv
RT: 2.059 min Scan# 608
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Instrument :
MSVOA_L
ClientSampleId :
11-IA-1






Tgt Ion: 84 Resp: 6519

Ion	Ratio	Lower	Upper
84	100		
49	148.0	132.9	199.3
51	51.3	34.8	52.2
86	66.4	48.2	72.2

Manual Integrations
APPROVED




Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#26
Hexane
Concen: 1.216 ppbv
RT: 2.823 min Scan# 845
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 57 Resp: 22119


Ion	Ratio	Lower	Upper
57	100		
41	104.5	74.6	111.8
43	125.3	179.0	268.4#
56	66.2	42.3	63.5#





#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 3.946 min Scan# 1193
 Delta R.T. -0.003 min
 Lab File: VL043314.D
 Acq: 08 Dec 2025 15:37




Instrument :
 MSVOA_L
 ClientSampleId :
 11-IA-1



Tgt Ion: 114 Resp: 354324
 Ion Ratio Lower Upper
 114 100
 63 24.9 21.0 31.6
 88 18.1 14.9 22.3




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#34
 2-Butanone
 Concen: 0.347 ppbv
 RT: 2.554 min Scan# 762
 Delta R.T. 0.003 min
 Lab File: VL043314.D
 Acq: 08 Dec 2025 15:37

Tgt Ion: 43 Resp: 8673
 Ion Ratio Lower Upper
 43 100
 72 27.2 18.1 27.1#





#35

Carbon Tetrachloride

Concen: 0.094 ppbv

RT: 3.755 min Scan# 1133

Delta R.T. -0.000 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Instrument :

MSVOA_L

ClientSampleId :

11-IA-1





Tgt Ion: 117 Resp: 2524

Ion	Ratio	Lower	Upper
117	100		
119	92.0	74.6	112.0
121	0.0	23.2	34.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#36

Benzene

Concen: 0.154 ppbv

RT: 3.648 min Scan# 1100




Delta R.T. -0.000 min


Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Tgt Ion: 78 Resp: 5240


Ion	Ratio	Lower	Upper
78	100		
77	25.4	17.5	26.3
50	21.4	15.5	23.3





#52
Tetrachloroethene
Concen: 0.094 ppbv m
RT: 8.215 min Scan# 2110
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_L
ClientSampleId :
11-IA-1



Tgt Ion:164 Resp: 1184
Ion Ratio Lower Upper
164 100
166 105.0 99.4 149.2
129 167.6 81.8 122.8
131 124.6 77.4 116.2#




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#53
Toluene
Concen: 0.256 ppbv m
RT: 6.917 min Scan# 2110
Delta R.T. -0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 91 Resp: 10274
Ion Ratio Lower Upper
91 100
92 62.3 47.4 71.0





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.869 min Scan# 21

Delta R.T. -0.003 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Instrument :

MSVOA_L

ClientSampleId :





11-IA-1

Tgt Ion:117 Resp: 27950
 Ion Ratio Lower Upper
 117 100
 82 65.9 56.1 84.1
 119 32.6 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#59

m/p-Xylene

Concen: 0.124 ppbv m




RT: 9.519 min Scan# 2914


Delta R.T. -0.023 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37


Tgt Ion: 91 Resp: 5340
 Ion Ratio Lower Upper
 91 100
 106 0.0 36.1 54.1#





#68
1-Bromo-4-Fluorobenzene
Concen: 10.469 ppbv
RT: 10.364 min Scan# 3175
Delta R.T. -0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37



Instrument :
MSVOA_L
ClientSampleId :
11-IA-1



Tgt Ion: 95 Resp: 214484
Ion Ratio Lower Upper
95 100
174 68.8 50.2 75.2
175 4.8 4.1 6.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 13-IA-1
Lab Sample ID: Q3799-05
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.40	1.98	J	1	0.54	2.47	12/08/25 16:25	VL120825
74-87-3	Chloromethane	0.46	0.95	J	1	0.21	1.03	12/08/25 16:25	VL120825
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 16:25	VL120825
74-83-9	Bromomethane	0.080	0.31	U	1	0.31	1.94	12/08/25 16:25	VL120825
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	12/08/25 16:25	VL120825
109-99-9	Tetrahydrofuran	0.080	0.24	U	1	0.24	1.47	12/08/25 16:25	VL120825
75-69-4	Trichlorofluoromethane	0.25	1.40	J	1	0.96	2.81	12/08/25 16:25	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1	1.07	3.83	12/08/25 16:25	VL120825
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1	1.05	3.49	12/08/25 16:25	VL120825
75-65-0	tert-Butyl alcohol	0.16	0.49	U	1	0.49	1.52	12/08/25 16:25	VL120825
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	12/08/25 16:25	VL120825
75-35-4	1,1-Dichloroethene	0.15	0.59	U	1	0.59	1.98	12/08/25 16:25	VL120825
67-64-1	Acetone	3.50	8.31		1	0.43	1.19	12/08/25 16:25	VL120825
75-15-0	Carbon Disulfide	0.080	0.25	U	1	0.25	1.56	12/08/25 16:25	VL120825
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	1	0.83	1.80	12/08/25 16:25	VL120825
75-09-2	Methylene Chloride	0.76	2.64		1	0.80	1.74	12/08/25 16:25	VL120825
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	1	0.48	1.98	12/08/25 16:25	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 16:25	VL120825
110-82-7	Cyclohexane	0.22	0.76	U	1	0.76	1.72	12/08/25 16:25	VL120825
78-93-3	2-Butanone	0.26	0.77	J	1	0.18	1.47	12/08/25 16:25	VL120825
56-23-5	Carbon Tetrachloride	0.070	0.44		1	0.19	0.19	12/08/25 16:25	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 16:25	VL120825
67-66-3	Chloroform	0.10	0.49	U	1	0.49	2.44	12/08/25 16:25	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 16:25	VL120825
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	1	0.65	2.34	12/08/25 16:25	VL120825
71-43-2	Benzene	0.21	0.67	J	1	0.26	1.60	12/08/25 16:25	VL120825
107-06-2	1,2-Dichloroethane	0.090	0.36	U	1	0.36	2.02	12/08/25 16:25	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 16:25	VL120825
78-87-5	1,2-Dichloropropane	0.13	0.60	U	1	0.60	2.31	12/08/25 16:25	VL120825
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	12/08/25 16:25	VL120825
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	1	0.41	2.05	12/08/25 16:25	VL120825
108-88-3	Toluene	0.19	0.72	J	1	0.60	1.88	12/08/25 16:25	VL120825
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	1	0.68	2.27	12/08/25 16:25	VL120825
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	1	0.50	2.27	12/08/25 16:25	VL120825
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	1	0.44	2.73	12/08/25 16:25	VL120825
124-48-1	Dibromochloromethane	0.090	0.77	U	1	0.77	4.26	12/08/25 16:25	VL120825
106-93-4	1,2-Dibromoethane	0.090	0.69	U	1	0.69	0.77	12/08/25 16:25	VL120825
127-18-4	Tetrachloroethene	0.050	0.34		1	0.14	0.20	12/08/25 16:25	VL120825
108-90-7	Chlorobenzene	0.080	0.37	U	1	0.37	2.30	12/08/25 16:25	VL120825
100-41-4	Ethyl Benzene	0.19	0.83	U	1	0.83	2.17	12/08/25 16:25	VL120825
179601-23-1	m/p-Xylene	0.41	1.78	U	1	1.78	4.34	12/08/25 16:25	VL120825
95-47-6	o-Xylene	0.21	0.91	U	1	0.91	2.17	12/08/25 16:25	VL120825
100-42-5	Styrene	0.16	0.68	U	1	0.68	2.13	12/08/25 16:25	VL120825
75-25-2	Bromoform	0.050	0.52	U	1	0.52	5.17	12/08/25 16:25	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	1	0.14	0.21	12/08/25 16:25	VL120825
95-49-8	2-Chlorotoluene	0.17	0.88	U	1	0.88	2.59	12/08/25 16:25	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 13-IA-1
Lab Sample ID: Q3799-05
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 16:25	VL120825
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 16:25	VL120825
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	1	0.30	3.01	12/08/25 16:25	VL120825
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 16:25	VL120825
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 16:25	VL120825
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1	1.56	3.71	12/08/25 16:25	VL120825
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1	1.39	5.33	12/08/25 16:25	VL120825
106-99-0	1,3-Butadiene	0.050	0.11	U	1	0.11	1.11	12/08/25 16:25	VL120825
91-20-3	Naphthalene	0.010	0.050	U	1	0.050	0.52	12/08/25 16:25	VL120825
622-96-8	4-Ethyltoluene	0.21	1.03	U	1	1.03	2.46	12/08/25 16:25	VL120825
110-54-3	Hexane	2.30	8.11		1	0.56	1.76	12/08/25 16:25	VL120825
107-05-1	Allyl Chloride	0.11	0.34	U	1	0.34	1.57	12/08/25 16:25	VL120825
123-91-1	1,4-Dioxane	0.22	0.79	U	1	0.79	1.80	12/08/25 16:25	VL120825
80-62-6	Methyl Methacrylate	0.14	0.57	U	1	0.57	2.05	12/08/25 16:25	VL120825

SURROGATES

460-00-4 1-Bromo-4-Fluorobenzene 10.2 70 (65) - 130 (135) 102% SPK: 10

INTERNAL STANDARDS

Area Count
74-97-5 Bromochloromethane 134000
540-36-3 1,4-Difluorobenzene 345000
3114-55-4 Chlorobenzene-d5 262000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043315.D
 Acq On : 08 Dec 2025 16:25
 Operator : SY/MD
 Sample : Q3799-05
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 13-IA-1

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:12:49 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.777	49	133616	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.943	114	344933	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.865	117	261904	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.361	95	195928	10.206	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 102.100%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.496	85	7550	0.401	ppbv	94
3) Chlorodifluoromethane	1.473	51	11927m	0.644	ppbv	
4) Chloromethane	1.528	50	3225	0.458	ppbv	93
9) Propene	1.486	41	6937	0.881	ppbv #	1
11) Trichlorofluoromethane	1.874	101	4607	0.252	ppbv	94
13) Ethanol	1.716	45	117047	42.294	ppbv	99
15) Acetone	1.829	43	65404	3.523	ppbv	100
19) Isopropyl Alcohol	1.881	45	152976	14.583	ppbv #	38
20) Methylene Chloride	2.056	84	8141	0.763	ppbv #	96
26) Hexane	2.819	57	39916	2.274	ppbv #	44
34) 2-Butanone	2.551	43	6283	0.258	ppbv	99
35) Carbon Tetrachloride	3.745	117	1779m	0.068	ppbv	
36) Benzene	3.645	78	6800	0.206	ppbv	97
52) Tetrachloroethene	8.202	164	663m	0.054	ppbv	
53) Toluene	6.914	91	7587m	0.194	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043315.D
Acq On : 08 Dec 2025 16:25
Operator : SY/MD
Sample : Q3799-05
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
13-IA-1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:12:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.777 min Scan# 831
Delta R.T. -0.004 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Instrument :
MSVOA_L
ClientSampleId :
13-IA-1

Tgt Ion: 49 Resp: 133610
Ion Ratio Lower Upper
49 100
128 49.4 23.0 69.0
130 62.4 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#2
Dichlorodifluoromethane
Concen: 0.401 ppbv
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25


Tgt Ion: 85 Resp: 7550
Ion Ratio Lower Upper
85 100
87 35.4 25.4 38.2





#3
Chlorodifluoromethane
Concen: 0.644 ppbv m
RT: 1.473 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1




Tgt Ion: 51 Resp: 1192
Ion Ratio Lower Upper
51 100
67 20.8 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#4
Chloromethane
Concen: 0.458 ppbv
RT: 1.528 min Scan# 445
Delta R.T. -0.003 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25



Tgt Ion: 50 Resp: 3225
Ion Ratio Lower Upper
50 100
52 26.3 24.1 36.1





#9

Propene

Concen: 0.881 ppbv

RT: 1.486 min Scan# 41

Delta R.T. 0.006 min

Lab File: VL043315.D

Acq: 08 Dec 2025 16:25

Instrument :

MSVOA_L

ClientSampleId :

13-IA-1

Tgt Ion: 41 Resp: 693

Ion Ratio Lower Upper

41 100





42 45.2 3.7 5.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025

Supervised By :Mahesh Dadoda 12/11/2025



#11

Trichlorofluoromethane

Concen: 0.252 ppbv

RT: 1.874 min Scan# 552

Delta R.T. -0.000 min

Lab File: VL043315.D




Acq: 08 Dec 2025 16:25


Tgt Ion:101 Resp: 4607

Ion Ratio Lower Upper

101 100

103 60.7 52.6 79.0









#13
Ethanol
Concen: 42.294 ppbv
RT: 1.716 min Scan# 503
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Instrument :
MSVOA_L
ClientSampleId :
13-IA-1

Tgt Ion: 45 Resp: 11704
Ion Ratio Lower Upper
45 100
46 38.1 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#15
Acetone
Concen: 3.523 ppbv
RT: 1.829 min Scan# 538
Delta R.T. -0.003 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25


Tgt Ion: 43 Resp: 65404
Ion Ratio Lower Upper
43 100
58 25.8 20.7 31.1





#19
Isopropyl Alcohol
Concen: 14.583 ppbv
RT: 1.881 min Scan# 511
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1




Tgt Ion: 45 Resp: 152970
Ion Ratio Lower Upper
45 100
58 1.2 31.3 46.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#20
Methylene Chloride
Concen: 0.763 ppbv
RT: 2.056 min Scan# 608
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Tgt Ion: 84 Resp: 8141
Ion Ratio Lower Upper
84 100
49 166.8 132.9 199.3
51 52.5 34.8 52.2#
86 65.1 48.2 72.2





#26
Hexane
Concen: 2.274 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1




Tgt Ion:	57	Resp:	39910
Ion Ratio	100	Lower	Upper
57	100		
41	100.1	74.6	111.8
43	81.2	179.0	268.4
56	58.6	42.3	63.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.943 min Scan# 1191
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25



Tgt Ion:	114	Resp:	344933
Ion Ratio	100	Lower	Upper
114	100		
63	25.8	21.0	31.6
88	17.8	14.9	22.3









#34
2-Butanone
Concen: 0.258 ppbv
RT: 2.551 min Scan# 761
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Instrument :
MSVOA_L
ClientSampleId :
13-IA-1

Tgt Ion: 43 Resp: 628
Ion Ratio Lower Upper
43 100
72 23.0 18.1 27.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#35
Carbon Tetrachloride
Concen: 0.068 ppbv m
RT: 3.745 min Scan# 1130
Delta R.T. -0.010 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion:117 Resp: 1779
Ion Ratio Lower Upper
117 100
119 121.8 74.6 112.0#
121 15.3 23.2 34.8#





#36

Benzene

Concen: 0.206 ppbv

RT: 3.645 min Scan# 1100

Delta R.T. -0.003 min

Lab File: VL043315.D

Acq: 08 Dec 2025 16:25

Instrument :

MSVOA_L

ClientSampleId :





13-IA-1

Tgt Ion: 78	Resp: 6800
Ion Ratio	Lower Upper
78 100	
77 22.1	17.5 26.3
50 21.8	15.5 23.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#52

Tetrachloroethene

Concen: 0.054 ppbv m




RT: 8.202 min Scan# 2507


Delta R.T. -0.010 min

Lab File: VL043315.D

Acq: 08 Dec 2025 16:25

Tgt	Ion:164	Resp:	663
Ion	Ratio	Lower	Upper
164	100		
166	91.8	99.4	149.2#
129	41.4	81.8	122.8#
131	34.1	77.4	116.2#









#53
Toluene
Concen: 0.194 ppbv m
RT: 6.914 min Scan# 2109
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Instrument :
MSVOA_L
ClientSampleId :
13-IA-1

Tgt Ion: 91 Resp: 758
Ion Ratio Lower Upper
91 100
92 61.7 47.4 71.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.865 min Scan# 2712
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25


Tgt Ion:117 Resp: 261904
Ion Ratio Lower Upper
117 100
82 64.6 56.1 84.1
119 32.6 25.0 37.6





#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.206 ppbv
 RT: 10.361 min Scan# 31
 Delta R.T. -0.007 min
 Lab File: VL043315.D
 Acq: 08 Dec 2025 16:25



Instrument :
 MSVOA_L
 ClientSampleId :
 13-IA-1



Tgt Ion: 95 Resp: 195928
 Ion Ratio Lower Upper
 95 100
 174 67.0 50.2 75.2
 175 5.0 4.1 6.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 16-IA-1
Lab Sample ID: Q3799-06
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.42	2.08	J	1	0.54	2.47	12/08/25 17:00	VL120825
74-87-3	Chloromethane	0.74	1.53		1	0.21	1.03	12/08/25 17:00	VL120825
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 17:00	VL120825
74-83-9	Bromomethane	0.080	0.31	U	1	0.31	1.94	12/08/25 17:00	VL120825
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	12/08/25 17:00	VL120825
109-99-9	Tetrahydrofuran	0.080	0.24	U	1	0.24	1.47	12/08/25 17:00	VL120825
75-69-4	Trichlorofluoromethane	0.21	1.18	J	1	0.96	2.81	12/08/25 17:00	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1	1.07	3.83	12/08/25 17:00	VL120825
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1	1.05	3.49	12/08/25 17:00	VL120825
75-65-0	tert-Butyl alcohol	0.16	0.49	U	1	0.49	1.52	12/08/25 17:00	VL120825
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	12/08/25 17:00	VL120825
75-35-4	1,1-Dichloroethene	0.15	0.59	U	1	0.59	1.98	12/08/25 17:00	VL120825
67-64-1	Acetone	29.8	70.8	E	1	0.43	1.19	12/08/25 17:00	VL120825
75-15-0	Carbon Disulfide	0.080	0.25	U	1	0.25	1.56	12/08/25 17:00	VL120825
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	1	0.83	1.80	12/08/25 17:00	VL120825
75-09-2	Methylene Chloride	7.30	25.4		1	0.80	1.74	12/08/25 17:00	VL120825
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	1	0.48	1.98	12/08/25 17:00	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 17:00	VL120825
110-82-7	Cyclohexane	0.22	0.76	U	1	0.76	1.72	12/08/25 17:00	VL120825
78-93-3	2-Butanone	0.44	1.30	J	1	0.18	1.47	12/08/25 17:00	VL120825
56-23-5	Carbon Tetrachloride	0.080	0.50		1	0.19	0.19	12/08/25 17:00	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 17:00	VL120825
67-66-3	Chloroform	0.10	0.49	U	1	0.49	2.44	12/08/25 17:00	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 17:00	VL120825
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	1	0.65	2.34	12/08/25 17:00	VL120825
71-43-2	Benzene	0.20	0.64	J	1	0.26	1.60	12/08/25 17:00	VL120825
107-06-2	1,2-Dichloroethane	0.090	0.36	U	1	0.36	2.02	12/08/25 17:00	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 17:00	VL120825
78-87-5	1,2-Dichloropropane	0.13	0.60	U	1	0.60	2.31	12/08/25 17:00	VL120825
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	12/08/25 17:00	VL120825
108-10-1	4-Methyl-2-Pentanone	0.11	0.45	J	1	0.41	2.05	12/08/25 17:00	VL120825
108-88-3	Toluene	0.26	0.98	J	1	0.60	1.88	12/08/25 17:00	VL120825
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	1	0.68	2.27	12/08/25 17:00	VL120825
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	1	0.50	2.27	12/08/25 17:00	VL120825
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	1	0.44	2.73	12/08/25 17:00	VL120825
124-48-1	Dibromochloromethane	0.090	0.77	U	1	0.77	4.26	12/08/25 17:00	VL120825
106-93-4	1,2-Dibromoethane	0.090	0.69	U	1	0.69	0.77	12/08/25 17:00	VL120825
127-18-4	Tetrachloroethene	0.020	0.14	U	1	0.14	0.20	12/08/25 17:00	VL120825
108-90-7	Chlorobenzene	0.080	0.37	U	1	0.37	2.30	12/08/25 17:00	VL120825
100-41-4	Ethyl Benzene	0.22	0.96	J	1	0.83	2.17	12/08/25 17:00	VL120825
179601-23-1	m/p-Xylene	0.90	3.91	J	1	1.78	4.34	12/08/25 17:00	VL120825
95-47-6	o-Xylene	0.24	1.04	J	1	0.91	2.17	12/08/25 17:00	VL120825
100-42-5	Styrene	0.16	0.68	U	1	0.68	2.13	12/08/25 17:00	VL120825
75-25-2	Bromoform	0.050	0.52	U	1	0.52	5.17	12/08/25 17:00	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	1	0.14	0.21	12/08/25 17:00	VL120825
95-49-8	2-Chlorotoluene	0.17	0.88	U	1	0.88	2.59	12/08/25 17:00	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 16-IA-1
Lab Sample ID: Q3799-06
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 17:00	VL120825
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 17:00	VL120825
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	1	0.30	3.01	12/08/25 17:00	VL120825
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 17:00	VL120825
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 17:00	VL120825
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1	1.56	3.71	12/08/25 17:00	VL120825
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1	1.39	5.33	12/08/25 17:00	VL120825
106-99-0	1,3-Butadiene	0.050	0.11	U	1	0.11	1.11	12/08/25 17:00	VL120825
91-20-3	Naphthalene	0.010	0.050	U	1	0.050	0.52	12/08/25 17:00	VL120825
622-96-8	4-Ethyltoluene	0.21	1.03	U	1	1.03	2.46	12/08/25 17:00	VL120825
110-54-3	Hexane	2.70	9.52		1	0.56	1.76	12/08/25 17:00	VL120825
107-05-1	Allyl Chloride	0.11	0.34	U	1	0.34	1.57	12/08/25 17:00	VL120825
123-91-1	1,4-Dioxane	0.22	0.79	U	1	0.79	1.80	12/08/25 17:00	VL120825
80-62-6	Methyl Methacrylate	0.14	0.57	U	1	0.57	2.05	12/08/25 17:00	VL120825

SURROGATES

460-00-4	1-Bromo-4-Fluorobenzene	10.4		70 (65) - 130 (135)	104% SPK: 10
----------	-------------------------	------	--	---------------------	--------------

INTERNAL STANDARDS

	Area Count	
74-97-5	Bromochloromethane	146000
540-36-3	1,4-Difluorobenzene	393000
3114-55-4	Chlorobenzene-d5	307000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043316.D
 Acq On : 08 Dec 2025 17:00
 Operator : SY/MD
 Sample : Q3799-06
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 16-IA-1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025

Quant Time: Dec 09 00:13:13 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2


QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Bromochloromethane	2.784	49	146093	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	393021	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	306652	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	232936	10.363	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 103.600%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.499	85	8618	0.418	ppbv	92
3) Chlorodifluoromethane	1.473	51	2556m	0.126	ppbv	
4) Chloromethane	1.531	50	5664	0.735	ppbv	98
9) Propene	1.483	41	14559m	1.692	ppbv	
10) Heptane	4.965	43	3437m	0.134	ppbv	
11) Trichlorofluoromethane	1.874	101	4165	0.209	ppbv	86
13) Ethanol	1.722	45	462202	152.750	ppbv	99
15) Acetone	1.832	43	604849	29.800	ppbv	97
17) tert-Butyl alcohol	2.046	59	3190m	0.120	ppbv	
19) Isopropyl Alcohol	1.910	45	7888776	687.812	ppbv #	38
20) Methylene Chloride	2.062	84	44935	7.301	ppbv	89
26) Hexane	2.826	57	52020	2.710	ppbv #	72
34) 2-Butanone	2.554	43	12307	0.444	ppbv	99
35) Carbon Tetrachloride	3.755	117	2261m	0.076	ppbv	
36) Benzene	3.648	78	7554	0.201	ppbv	95
50) 4-Methyl-2-Pentanone	5.745	43	4131m	0.106	ppbv	
53) Toluene	6.920	91	11377	0.255	ppbv	100
58) Ethyl Benzene	9.344	91	13162	0.223	ppbv	93
59) m/p-Xylene	9.523	91	42461	0.897	ppbv	97
60) o-Xylene	9.953	91	11276	0.239	ppbv	90
61) Styrene	9.862	104	2498m	0.121	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1




Tgt Ion: 49 Resp: 14609
Ion Ratio Lower Upper
49 100
128 50.3 23.0 69.0
130 64.6 29.9 89.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#2
Dichlorodifluoromethane
Concen: 0.418 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Tgt Ion: 85 Resp: 8618
Ion Ratio Lower Upper
85 100
87 36.4 25.4 38.2





#3
Chlorodifluoromethane
Concen: 0.126 ppbv m
RT: 1.473 min Scan# 428
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1




Tgt Ion: 51 Resp: 2550
Ion Ratio Lower Upper
51 100
67 39.2 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#4
Chloromethane
Concen: 0.735 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Tgt Ion: 50 Resp: 5664
Ion Ratio Lower Upper
50 100
52 28.7 24.1 36.1





#9
 Propene
 Concen: 1.692 ppbv m
 RT: 1.483 min Scan# 41
 Delta R.T. 0.003 min
 Lab File: VL043316.D
 Acq: 08 Dec 2025 17:00




Instrument :
 MSVOA_L
 ClientSampleId :
 16-IA-1




Tgt Ion: 41 Resp: 1455
 Ion Ratio Lower Upper
 41 100
 42 294.1 3.7 5.5

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025




#10
 Heptane
 Concen: 0.134 ppbv m
 RT: 4.965 min Scan# 1507
 Delta R.T. -0.000 min
 Lab File: VL043316.D
 Acq: 08 Dec 2025 17:00



Tgt Ion: 43 Resp: 3437
 Ion Ratio Lower Upper
 43 100
 57 0.0 39.7 59.5









#11
Trichlorofluoromethane
Concen: 0.209 ppbv
RT: 1.874 min Scan# 505
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Instrument :
MSVOA_L
ClientSampleId :
16-IA-1

Tgt Ion:101 Resp: 416
Ion Ratio Lower Upper
101 100
103 54.8 52.6 79.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#13
Ethanol
Concen: 152.750 ppbv
RT: 1.722 min Scan# 505
Delta R.T. 0.006 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00


Tgt Ion: 45 Resp: 462202
Ion Ratio Lower Upper
45 100
46 37.9 15.0 59.8





#15
Acetone
Concen: 29.800 ppbv
RT: 1.832 min Scan# 51
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1




Tgt Ion: 43 Resp: 604849
Ion Ratio Lower Upper
43 100
58 27.2 20.7 31.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#17
tert-Butyl alcohol
Concen: 0.120 ppbv m
RT: 2.046 min Scan# 605
Delta R.T. 0.010 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Tgt Ion: 59 Resp: 3190
Ion Ratio Lower Upper
59 100
57 2730.3 9.0 13.4#





#19
Isopropyl Alcohol
Concen: 687.812 ppbv
RT: 1.910 min Scan# 50
Delta R.T. 0.029 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1



Tgt Ion: 45 Resp: 7888770
Ion Ratio Lower Upper
45 100
58 1.0 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#20
Methylene Chloride
Concen: 7.301 ppbv
RT: 2.062 min Scan# 610
Delta R.T. 0.006 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00


Tgt Ion: 84 Resp: 44935
Ion Ratio Lower Upper
84 100
49 144.4 132.9 199.3
51 47.5 34.8 52.2
86 58.7 48.2 72.2





#26
Hexane
Concen: 2.710 ppbv
RT: 2.826 min Scan# 844
Delta R.T. 0.006 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1




Tgt Ion: 57 Resp: 52020
Ion Ratio Lower Upper
57 100
41 97.1 74.6 111.8
43 154.4 179.0 268.4
56 58.0 42.3 63.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.949 min Scan# 1193
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00



Tgt Ion:114 Resp: 393021
Ion Ratio Lower Upper
114 100
63 24.6 21.0 31.6
88 17.7 14.9 22.3









#34
2-Butanone
Concen: 0.444 ppbv
RT: 2.554 min Scan# 761
Delta R.T. 0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Instrument :
MSVOA_L
ClientSampleId :
16-IA-1

Tgt Ion: 43 Resp: 1230
Ion Ratio Lower Upper
43 100
72 22.9 18.1 27.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#35
Carbon Tetrachloride
Concen: 0.076 ppbv m
RT: 3.755 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00


Tgt Ion:117 Resp: 2261
Ion Ratio Lower Upper
117 100
119 87.7 74.6 112.0
121 34.1 23.2 34.8





#36
Benzene
Concen: 0.201 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1



Tgt Ion: 78 Resp: 7554
Ion Ratio Lower Upper
78 100
77 19.3 17.5 26.3
50 21.4 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#50
4-Methyl-2-Pentanone
Concen: 0.106 ppbv m
RT: 5.745 min Scan# 1748
Delta R.T. 0.013 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 43 Resp: 4131
Ion Ratio Lower Upper
43 100
58 0.0 28.4 42.6#





#53

Toluene

Concen: 0.255 ppbv

RT: 6.920 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VL043316.D

Acq: 08 Dec 2025 17:00

Instrument :

MSVOA_L

ClientSampleId :

16-IA-1

Tgt Ion: 91 Resp: 1137

Ion Ratio Lower Upper

91 100





92 59.0 47.4 71.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025

Supervised By :Mahesh Dadoda 12/11/2025



#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.872 min Scan# 2714

Delta R.T. -0.000 min

Lab File: VL043316.D

Acq: 08 Dec 2025 17:00




Tgt Ion:117 Resp: 306652


Ion Ratio Lower Upper

117 100

82 64.3 56.1 84.1

119 32.1 25.0 37.6









#58
Ethyl Benzene
Concen: 0.223 ppbv
RT: 9.344 min Scan# 2115
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Instrument :
MSVOA_L
ClientSampleId :
16-IA-1

Tgt Ion: 91 Resp: 1316
Ion Ratio Lower Upper
91 100
106 32.4 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#59
m/p-Xylene
Concen: 0.897 ppbv
RT: 9.523 min Scan# 2915
Delta R.T. -0.020 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 91 Resp: 42461
Ion Ratio Lower Upper
91 100
106 47.0 36.1 54.1









#60
o-Xylene
Concen: 0.239 ppbv
RT: 9.953 min Scan# 3049
Delta R.T. -0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Instrument :
MSVOA_L
ClientSampleId :
16-IA-1

Tgt Ion: 91 Resp: 11270
Ion Ratio Lower Upper
91 100
106 48.2 20.9 62.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#61
Styrene
Concen: 0.121 ppbv m
RT: 9.862 min Scan# 3020
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00


Tgt Ion:104 Resp: 2498
Ion Ratio Lower Upper
104 100
78 51.9 42.1 63.1
103 44.7 37.2 55.8





#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.363 ppbv
 RT: 10.367 min Scan# 3176
 Delta R.T. -0.000 min
 Lab File: VL043316.D
 Acq: 08 Dec 2025 17:00



Instrument :
 MSVOA_L
 ClientSampleId :
 16-IA-1



Tgt Ion:	95	Resp:	232930
Ion Ratio	Lower	Upper	
95	100		
174	68.8	50.2	75.2
175	5.0	4.1	6.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: 16-IA-1DL
Lab Sample ID: Q3799-06DL
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected: 12/03/25
Date Received: 12/04/25
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	1.10	5.44	UD	10	5.44	24.7	12/08/25 21:56	VL120825
74-87-3	Chloromethane	0.98	2.02	UD	10	2.02	10.3	12/08/25 21:56	VL120825
75-01-4	Vinyl Chloride	0.25	0.64	UD	10	0.64	0.77	12/08/25 21:56	VL120825
74-83-9	Bromomethane	0.77	2.99	UD	10	2.99	19.4	12/08/25 21:56	VL120825
75-00-3	Chloroethane	1.50	3.96	UD	10	3.96	13.2	12/08/25 21:56	VL120825
109-99-9	Tetrahydrofuran	0.80	2.36	UD	10	2.36	14.8	12/08/25 21:56	VL120825
75-69-4	Trichlorofluoromethane	1.70	9.55	UD	10	9.55	28.1	12/08/25 21:56	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	1.40	10.7	UD	10	10.7	38.3	12/08/25 21:56	VL120825
76-14-2	Dichlorotetrafluoroethane	1.50	10.5	UD	10	10.5	35.0	12/08/25 21:56	VL120825
75-65-0	tert-Butyl alcohol	1.60	4.85	UD	10	4.85	15.2	12/08/25 21:56	VL120825
142-82-5	Heptane	1.70	6.97	UD	10	6.97	20.5	12/08/25 21:56	VL120825
75-35-4	1,1-Dichloroethene	1.50	5.95	UD	10	5.95	19.8	12/08/25 21:56	VL120825
67-64-1	Acetone	35.0	83.1	D	10	4.28	11.9	12/08/25 21:56	VL120825
75-15-0	Carbon Disulfide	0.80	2.49	UD	10	2.49	15.6	12/08/25 21:56	VL120825
1634-04-4	Methyl tert-Butyl Ether	2.30	8.29	UD	10	8.29	18.0	12/08/25 21:56	VL120825
75-09-2	Methylene Chloride	9.10	31.6	D	10	7.99	17.4	12/08/25 21:56	VL120825
156-60-5	trans-1,2-Dichloroethene	1.20	4.76	UD	10	4.76	19.8	12/08/25 21:56	VL120825
75-34-3	1,1-Dichloroethane	1.30	5.26	UD	10	5.26	20.2	12/08/25 21:56	VL120825
110-82-7	Cyclohexane	2.20	7.57	UD	10	7.57	17.2	12/08/25 21:56	VL120825
78-93-3	2-Butanone	0.56	1.65	UD	10	1.65	14.8	12/08/25 21:56	VL120825
56-23-5	Carbon Tetrachloride	0.25	1.57	UD	10	1.57	1.89	12/08/25 21:56	VL120825
156-59-2	cis-1,2-Dichloroethene	0.99	3.93	UD	10	3.93	19.8	12/08/25 21:56	VL120825
67-66-3	Chloroform	0.96	4.69	UD	10	4.69	24.4	12/08/25 21:56	VL120825
71-55-6	1,1,1-Trichloroethane	0.16	0.87	UD	10	0.87	1.64	12/08/25 21:56	VL120825
540-84-1	2,2,4-Trimethylpentane	1.40	6.54	UD	10	6.54	23.4	12/08/25 21:56	VL120825
71-43-2	Benzene	0.79	2.52	UD	10	2.52	16.0	12/08/25 21:56	VL120825
107-06-2	1,2-Dichloroethane	0.93	3.76	UD	10	3.76	20.2	12/08/25 21:56	VL120825
79-01-6	Trichloroethene	0.24	1.29	UD	10	1.29	1.61	12/08/25 21:56	VL120825
78-87-5	1,2-Dichloropropane	1.30	6.01	UD	10	6.01	23.1	12/08/25 21:56	VL120825
75-27-4	Bromodichloromethane	0.63	4.22	UD	10	4.22	33.5	12/08/25 21:56	VL120825
108-10-1	4-Methyl-2-Pentanone	0.97	3.98	UD	10	3.98	20.5	12/08/25 21:56	VL120825
108-88-3	Toluene	1.60	6.03	UD	10	6.03	18.8	12/08/25 21:56	VL120825
10061-02-6	t-1,3-Dichloropropene	1.50	6.81	UD	10	6.81	22.7	12/08/25 21:56	VL120825
10061-01-5	cis-1,3-Dichloropropene	1.10	4.99	UD	10	4.99	22.7	12/08/25 21:56	VL120825
79-00-5	1,1,2-Trichloroethane	0.81	4.42	UD	10	4.42	27.3	12/08/25 21:56	VL120825
124-48-1	Dibromochloromethane	0.85	7.24	UD	10	7.24	42.6	12/08/25 21:56	VL120825
106-93-4	1,2-Dibromoethane	0.85	6.53	UD	10	6.53	7.69	12/08/25 21:56	VL120825
127-18-4	Tetrachloroethene	0.15	1.02	UD	10	1.02	2.03	12/08/25 21:56	VL120825
108-90-7	Chlorobenzene	0.77	3.55	UD	10	3.55	23.0	12/08/25 21:56	VL120825
100-41-4	Ethyl Benzene	1.90	8.25	UD	10	8.25	21.7	12/08/25 21:56	VL120825
179601-23-1	m/p-Xylene	4.10	17.8	UD	10	17.8	43.4	12/08/25 21:56	VL120825
95-47-6	o-Xylene	2.10	9.12	UD	10	9.12	21.7	12/08/25 21:56	VL120825
100-42-5	Styrene	1.60	6.81	UD	10	6.81	21.3	12/08/25 21:56	VL120825
75-25-2	Bromoform	0.48	4.96	UD	10	4.96	51.7	12/08/25 21:56	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.18	1.24	UD	10	1.24	2.06	12/08/25 21:56	VL120825
95-49-8	2-Chlorotoluene	1.70	8.80	UD	10	8.80	25.9	12/08/25 21:56	VL120825

Report of Analysis

Client:	RTP Environmental	Date Collected:	12/03/25
Project:	CHADWICK SQUARE	Date Received:	12/04/25
Client Sample ID:	16-IA-1DL	SDG No.:	Q3799
Lab Sample ID:	Q3799-06DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 mL		

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	1.80	8.85	UD	10	8.85	24.6	12/08/25 21:56	VL120825
95-63-6	1,2,4-Trimethylbenzene	1.80	8.85	UD	10	8.85	24.6	12/08/25 21:56	VL120825
541-73-1	1,3-Dichlorobenzene	0.54	3.25	UD	10	3.25	30.1	12/08/25 21:56	VL120825
106-46-7	1,4-Dichlorobenzene	1.30	7.82	UD	10	7.82	30.1	12/08/25 21:56	VL120825
95-50-1	1,2-Dichlorobenzene	1.30	7.82	UD	10	7.82	30.1	12/08/25 21:56	VL120825
120-82-1	1,2,4-Trichlorobenzene	2.10	15.6	UD	10	15.6	37.1	12/08/25 21:56	VL120825
87-68-3	Hexachloro-1,3-Butadiene	1.30	13.9	UD	10	13.9	53.3	12/08/25 21:56	VL120825
106-99-0	1,3-Butadiene	0.51	1.13	UD	10	1.13	11.1	12/08/25 21:56	VL120825
91-20-3	Naphthalene	0.13	0.68	UD	10	0.68	5.24	12/08/25 21:56	VL120825
622-96-8	4-Ethyltoluene	2.10	10.3	UD	10	10.3	24.6	12/08/25 21:56	VL120825
110-54-3	Hexane	2.40	8.46	JD	10	5.64	17.6	12/08/25 21:56	VL120825
107-05-1	Allyl Chloride	1.10	3.44	UD	10	3.44	15.7	12/08/25 21:56	VL120825
123-91-1	1,4-Dioxane	2.20	7.93	UD	10	7.93	18.0	12/08/25 21:56	VL120825
80-62-6	Methyl Methacrylate	1.40	5.73	UD	10	5.73	20.5	12/08/25 21:56	VL120825

SURROGATES

460-00-4	1-Bromo-4-Fluorobenzene	9.90		70 (65) - 130 (135)	99% SPK: 10
----------	-------------------------	------	--	---------------------	-------------

INTERNAL STANDARDS

	Area Count	
74-97-5	Bromochloromethane	128000
540-36-3	1,4-Difluorobenzene	317000
3114-55-4	Chlorobenzene-d5	248000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043324.D
Acq On : 08 Dec 2025 21:56
Operator : SY/MD
Sample : Q3799-06DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
16-IA-1DL

Quant Time: Dec 09 01:25:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Bromochloromethane	2.780	49	127959	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	316604	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	248177	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.364	95	180171	9.904	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.000%
Target Compounds						
					Qvalue	
13) Ethanol	1.719	45	40793	15.392	ppbv	98
15) Acetone	1.832	43	62169	3.497	ppbv	98
19) Isopropyl Alcohol	1.884	45	2523636	251.215	ppbv #	37
20) Methylene Chloride	2.059	84	8511	0.911	ppbv	88
26) Hexane	2.819	57	3985	0.237	ppbv #	68


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043324.D
Acq On : 08 Dec 2025 21:56
Operator : SY/MD
Sample : Q3799-06DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
16-IA-1DL

Quant Time: Dec 09 01:25:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.780 min Scan# 81
Delta R.T. -0.001 min
Lab File: VL043324.D
Acq: 08 Dec 2025 21:56




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1DL


Tgt Ion: 49 Resp: 127959
Ion Ratio Lower Upper
49 100
128 49.0 23.0 69.0
130 63.6 29.9 89.7



#13
Ethanol
Concen: 15.392 ppbv
RT: 1.719 min Scan# 504
Delta R.T. 0.003 min
Lab File: VL043324.D
Acq: 08 Dec 2025 21:56

Tgt Ion: 45 Resp: 40793
Ion Ratio Lower Upper
45 100
46 36.1 15.0 59.8









#15
Acetone
Concen: 3.497 ppbv
RT: 1.832 min Scan# 51
Delta R.T. -0.000 min
Lab File: VL043324.D
Acq: 08 Dec 2025 21:56




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1DL


Tgt Ion: 43 Resp: 62169
Ion Ratio Lower Upper
43 100
58 25.0 20.7 31.1



#19
Isopropyl Alcohol
Concen: 251.215 ppbv
RT: 1.884 min Scan# 555
Delta R.T. 0.003 min
Lab File: VL043324.D
Acq: 08 Dec 2025 21:56

Tgt Ion: 45 Resp: 2523636
Ion Ratio Lower Upper
45 100
58 0.6 31.3 46.9#









#20
Methylene Chloride
Concen: 0.911 ppbv
RT: 2.059 min Scan# 608
Delta R.T. 0.003 min
Lab File: VL043324.D
Acq: 08 Dec 2025 21:56




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1DL


Tgt Ion:	84	Resp:	8511
Ion	Ratio	Lower	Upper
84	100		
49	144.1	132.9	199.3
51	41.6	34.8	52.2
86	66.5	48.2	72.2



#26
Hexane
Concen: 0.237 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043324.D
Acq: 08 Dec 2025 21:56

Tgt Ion:	57	Resp:	3985
Ion	Ratio	Lower	Upper
57	100		
41	114.6	74.6	111.8#
43	165.2	179.0	268.4#
56	75.2	42.3	63.5#









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.946 min Scan# 1192
Delta R.T. -0.004 min
Lab File: VL043324.D
Acq: 08 Dec 2025 21:56




Instrument :
MSVOA_1
ClientSampleId :
16-IA-1DL


Tgt Ion:114 Resp: 316604
Ion Ratio Lower Upper
114 100
63 24.9 21.0 31.6
88 18.2 14.9 22.3



#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.869 min Scan# 2713
Delta R.T. -0.004 min
Lab File: VL043324.D
Acq: 08 Dec 2025 21:56


Tgt Ion:117 Resp: 248177
Ion Ratio Lower Upper
117 100
82 65.4 56.1 84.1
119 32.3 25.0 37.6







#68
 1-Bromo-4-Fluorobenzene
 Concen: 9.904 ppbv
 RT: 10.364 min Scan# 31
 Delta R.T. -0.004 min
 Lab File: VL043324.D
 Acq: 08 Dec 2025 21:56

Instrument :
 MSVOA_L
 ClientSampleId :
 16-IA-1DL



Tgt Ion:	95	Resp:	180171
Ion	Ratio	Lower	Upper
95	100		
174	70.2	50.2	75.2
175	5.0	4.1	6.1





CALIBRATION SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL111825AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Wed Nov 19 01:11:25 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL043209.D 0.1 =VL043208.D 0.5 =VL043203.D 1 =VL043206.D 2 =VL043205.D 10 =VL043204.D 15 =VL043210.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.398	1.505	1.476	1.340	1.334	1.410	5.53
3) Chlorodifluoro...			1.545	1.404	1.437	1.317	1.231	1.387	8.60
4) Chloromethane			0.521	0.559	0.568	0.498	0.489	0.527	6.72
5) T Vinyl Chloride	0.724	0.655	0.533	0.580	0.575	0.486	0.483	0.577	15.33
6) T Bromomethane			0.300	0.302	0.286	0.253	0.258	0.280	8.30
7) Chloroethane			0.246	0.248	0.236	0.194	0.199	0.225	11.59
8) T Dichlorotetra...			1.156	1.241	1.262	1.091	1.116	1.173	6.43
9) T Propene			0.677	0.598	0.621	0.501	0.548	0.589	11.44
10) T Heptane			2.106	1.804	1.728	1.604	1.522	1.753	12.87
11) T Trichlorofluor...			1.238	1.515	1.487	1.240	1.352	1.367	9.62
12) T 1,1,2-Trichlor...			1.061	1.142	1.115	0.987	1.029	1.067	5.92
13) Ethanol			0.215	0.231	0.252	0.166	0.172	0.207	17.99
14) T Bromoethene			0.448	0.423	0.436	0.362	0.376	0.409	9.29
15) T Acetone			1.721	1.591	1.519	1.032	1.085	1.389	22.41
16) T 1,3-Butadiene			0.745	0.629	0.626	0.551	0.546	0.619	13.01
17) tert-Butyl alc...			1.784	2.014	1.880	1.705	1.735	1.824	6.87
18) T 1,1-Dichloroet...			0.555	0.530	0.515	0.441	0.443	0.497	10.52
19) T Isopropyl Alcohol			0.848	0.861	0.823	0.684	0.709	0.785	10.51
20) T Methylene Chlo...			0.906	0.710	0.585	0.398	0.403	0.600	35.87
21) T Allyl Chloride			1.139	0.895	0.936	0.849	0.855	0.935	12.75
22) T trans-1,2-Dich...			0.605	0.614	0.599	0.499	0.521	0.568	9.39
23) T Vinyl Acetate			1.885	1.632	1.682	1.622	1.565	1.677	7.36
24) T 1,1-Dichloroet...			1.050	1.173	1.194	1.018	1.044	1.096	7.42
25) T Ethyl Acetate			3.371	3.123	3.074	2.759	2.655	2.996	9.68
26) T Hexane			1.374	1.424	1.375	1.235	1.161	1.314	8.44
27) T Carbon Disulfide			1.303	1.424	1.432	1.257	1.291	1.341	6.03
28) T Methyl tert-Bu...			0.642	0.704	0.674	0.591	0.614	0.645	7.02
29) T Chloroform			1.955	1.898	1.928	1.785	1.770	1.867	4.52
30) T Cyclohexane			1.135	1.194	1.130	1.024	1.024	1.101	6.80
31) T cis-1,2-Dichlo...			1.399	1.351	1.322	1.240	1.217	1.306	5.83
32) T 1,1,1-Trichlor...	2.116	2.046	1.970	2.018	2.033	1.884	1.865	1.990	4.52


33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.744	0.751	0.736	0.644	0.655	0.706	7.38
35) T Carbon Tetrach...	0.832	0.765	0.779	0.771	0.737	0.708	0.721	0.759	5.51
36) T Benzene			0.996	1.001	0.986	0.905	0.900	0.958	5.28
37) T 1,2-Dichloroet...			0.572	0.580	0.565	0.531	0.553	0.560	3.39
38) T Trichloroethene	0.293	0.475	0.390	0.388	0.371	0.353	0.353	0.375	14.65
39) T 1,2-Dichloropr...			0.372	0.364	0.350	0.335	0.332	0.351	4.97

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\											
Method File : VL111825AIR.M											
40)	T	1,4-Dioxane			0.180	0.166	0.155	0.142	0.144	0.158	10.04
41)	T	Tetrahydrofuran			0.418	0.401	0.402	0.367	0.360	0.390	6.32
42)	T	Bromodichlorom...			0.782	0.795	0.783	0.745	0.758	0.773	2.63
43)		Methyl Methacr...			0.576	0.532	0.511	0.462	0.462	0.508	9.55
44)	T	2,2,4-Trimethy...			1.746	1.765	1.621	1.555	1.537	1.645	6.47
45)	T	t-1,3-Dichloro...			0.495	0.472	0.489	0.455	0.458	0.474	3.79
46)	T	cis-1,3-Dichlo...			0.599	0.557	0.562	0.534	0.529	0.556	5.00
47)	T	1,1,2-Trichlor...			0.424	0.372	0.378	0.344	0.342	0.372	8.91
48)	T	Dibromochlorom...			0.629	0.666	0.663	0.618	0.621	0.640	3.63
49)	T	Bromoform			0.583	0.569	0.581	0.555	0.545	0.566	2.86
50)	T	4-Methyl-2-Pen...			1.025	1.061	1.027	0.946	0.914	0.994	6.19
51)	T	2-Hexanone			0.782	0.805	0.816	0.734	0.724	0.772	5.37
52)	T	Tetrachloroethene	0.363	0.424	0.344	0.362	0.349	0.329	0.323	0.356	9.37
53)	T	Toluene			1.198	1.182	1.146	1.075	1.069	1.134	5.26
54)	T	1,2-Dibromoethane		0.528	0.587	0.569	0.549	0.525	0.526	0.547	4.79
-----ISTD-----											
55)	I	Chlorobenzene-d5									
56)		1,1,1,2-Tetrac...			0.791	0.769	0.761	0.677	0.677	0.735	7.34
57)	T	Chlorobenzene			1.035	1.088	1.089	0.969	0.953	1.027	6.24
58)	T	Ethyl Benzene			1.927	2.038	2.008	1.821	1.830	1.925	5.16
59)	T	m/p-Xylene			1.618	1.611	1.607	1.434	1.446	1.543	6.12
60)	T	o-Xylene			1.584	1.625	1.659	1.419	1.421	1.542	7.42
61)	T	Styrene			0.669	0.690	0.719	0.648	0.647	0.675	4.54
62)		Isopropylbenzene			2.807	2.946	2.945	2.556	2.536	2.758	7.31
63)	T	1,1,2,2-Tetrac...	0.999	1.034	1.016	1.012	1.028	0.891	0.884	0.981	6.58
64)		n-propylbenzene			0.722	0.758	0.757	0.657	0.649	0.709	7.47
65)		tert-Butylbenzene			2.583	2.687	2.646	2.224	2.176	2.463	9.90
66)	T	Benzyl Chloride			0.275	0.251	0.257	0.241	0.225	0.250	7.36
67)		sec-Butylbenzene			3.588	3.728	3.697	3.108	3.070	3.438	9.40
68)	S	1-Bromo-4-Fluo...	0.712	0.725	0.729	0.733	0.741	0.735	0.757	0.733	1.90
69)		p-Isopropyltol...			2.994	3.070	3.197	2.660	2.641	2.912	8.58
70)		n-Butylbenzene			2.676	3.041	3.125	2.639	2.647	2.825	8.39
71)		2-Chlorotoluene			2.198	2.257	2.210	1.960	1.975	2.120	6.65
72)	T	4-Ethyltoluene			1.934	1.948	2.009	1.749	1.747	1.877	6.46
73)	T	1,3,5-Trimethy...			1.623	1.682	1.662	1.456	1.454	1.575	7.09
74)	T	1,2,4-Trimethy...			1.796	1.844	1.871	1.548	1.537	1.719	9.51
75)	T	1,3-Dichlorobe...			1.043	1.053	1.061	0.937	0.922	1.003	6.78
76)	T	1,4-Dichlorobe...			1.056	1.033	1.060	0.940	0.930	1.003	6.34
77)	T	1,2-Dichlorobe...			0.974	1.064	1.050	0.872	0.863	0.965	9.83
78)	T	Hexachloro-1,3...			0.854	0.899	0.920	0.686	0.666	0.805	14.94
79)	T	Naphthalene		1.230	1.483	1.599	1.860	1.496	1.493	1.527	13.35
80)	T	Naphthalene,2-...			0.397	0.323	0.448	0.397	0.439	0.401	12.30
81)	T	1,2,4-Trichlor...			0.549	0.692	0.813	0.678	0.682	0.683	13.72

(#) = Out of Range

Methylene Chloride

Response Ratio



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043203.D
 Acq On : 18 Nov 2025 09:29
 Operator : SY/MD
 Sample : VSTDIC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5

Quant Time: Nov 19 00:48:47 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	110490	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.956	114	296493	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	247761	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.371 95 180529 9.941 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 99.400%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	7723	0.496	ppbv	99
3) Chlorodifluoromethane	1.473	51	8538	0.557	ppbv #	82
4) Chloromethane	1.531	50	2881	0.494	ppbv	94
5) Vinyl Chloride	1.580	62	2947	0.462	ppbv #	87
6) Bromomethane	1.667	94	1656	0.536	ppbv	97
7) Chloroethane	1.703	64	1357	0.547	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	6386	0.493	ppbv	100
9) Propene	1.483	41	3739	0.526	ppbv #	1
10) Heptane	4.972	43	11637m	0.597	ppbv	
11) Trichlorofluoromethane	1.878	101	6841	0.453	ppbv	92
12) 1,1,2-Trichlorotrifluo...	2.140	101	5863	0.497	ppbv	99
13) Ethanol	1.719	45	1190	0.520	ppbv	64
14) Bromoethene	1.781	108	2476	0.548	ppbv #	75
15) Acetone	1.836	43	9506	0.619	ppbv #	88
16) 1,3-Butadiene	1.606	39	4115	0.601	ppbv	95
17) tert-Butyl alcohol	2.043	59	9854	0.489	ppbv #	68
18) 1,1-Dichloroethene	2.033	96	3068	0.559	ppbv	96
19) Isopropyl Alcohol	1.887	45	4687	0.540	ppbv #	51
20) Methylene Chloride	2.059	84	5006	0.349	ppbv	92
21) Allyl Chloride	2.095	41	6291	0.609	ppbv	95
22) trans-1,2-Dichloroethene	2.337	96	3343	0.533	ppbv	80
23) Vinyl Acetate	2.464	43	10414	0.562	ppbv #	92
24) 1,1-Dichloroethane	2.409	63	5803	0.479	ppbv #	96
25) Ethyl Acetate	2.839	43	18625m	0.562	ppbv	
26) Hexane	2.823	57	7589	0.523	ppbv #	91
27) Carbon Disulfide	2.150	76	7198	0.486	ppbv #	1
28) Methyl tert-Butyl Ether	2.441	73	3545	0.498	ppbv #	15
29) Chloroform	2.849	83	10800	0.523	ppbv	95
30) Cyclohexane	3.849	84	6271	0.515	ppbv	85
31) cis-1,2-Dichloroethene	2.719	61	7728	0.536	ppbv	87
32) 1,1,1-Trichloroethane	3.363	97	10886	0.490	ppbv	98
34) 2-Butanone	2.557	43	11026	0.527	ppbv	95
35) Carbon Tetrachloride	3.761	117	11551	0.513	ppbv	94
36) Benzene	3.655	78	14759	0.520	ppbv #	91
37) 1,2-Dichloroethane	3.218	62	8477	0.510	ppbv #	93
38) Trichloroethene	4.542	130	5775	0.519	ppbv #	87
39) 1,2-Dichloropropane	4.312	63	5510	0.530	ppbv #	80
40) 1,4-Dioxane	4.597	88	2675m	0.573	ppbv	
41) Tetrahydrofuran	3.059	42	6193	0.536	ppbv	95
42) Bromodichloromethane	4.490	83	11597	0.506	ppbv #	84
43) Methyl Methacrylate	4.881	69	8534m	0.566	ppbv	
44) 2,2,4-Trimethylpentane	4.642	57	25890	0.531	ppbv	95
45) t-1,3-Dichloropropene	6.412	75	7343	0.523	ppbv	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043203.D
 Acq On : 18 Nov 2025 09:29
 Operator : SY/MD
 Sample : VSTDIC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:48:47 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M


Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	8886m	0.539	ppbv	
47) 1,1,2-Trichloroethane	6.581	97	6280m	0.567	ppbv	
48) Dibromochloromethane	7.380	129	9327	0.492	ppbv	97
49) Bromoform	9.481	173	8636	0.514	ppbv	100
50) 4-Methyl-2-Pentanone	5.755	43	15190	0.515	ppbv	92
51) 2-Hexanone	7.442	43	11594	0.506	ppbv #	93
52) Tetrachloroethene	8.218	164	5097	0.483	ppbv #	82
53) Toluene	6.924	91	17755m	0.525	ppbv	
54) 1,2-Dibromoethane	7.649	107	8708	0.537	ppbv	90
56) 1,1,1,2-Tetrachloroethane	8.917	131	9793m	0.538	ppbv	
57) Chlorobenzene	8.917	112	12818	0.504	ppbv	98
58) Ethyl Benzene	9.351	91	23866	0.500	ppbv	96
59) m/p-Xylene	9.526	91	40076m	1.044	ppbv	
60) o-Xylene	9.956	91	19621	0.514	ppbv	100
61) Styrene	9.869	104	8287	0.496	ppbv	96
62) Isopropylbenzene	10.536	105	34773	0.509	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.960	83	12591	0.519	ppbv	100
64) n-propylbenzene	10.986	120	8945	0.509	ppbv	100
65) tert-Butylbenzene	11.526	119	31998	0.524	ppbv	99
66) Benzyl Chloride	11.610	91	3403	0.550	ppbv	98
67) sec-Butylbenzene	11.766	105	44448	0.522	ppbv	97
69) p-Isopropyltoluene	11.921	119	37087	0.514	ppbv	100
70) n-Butylbenzene	12.277	91	33153	0.474	ppbv	94
71) 2-Chlorotoluene	10.895	91	27224	0.518	ppbv	99
72) 4-Ethyltoluene	11.122	105	23963	0.515	ppbv	97
73) 1,3,5-Trimethylbenzene	11.199	105	20104	0.515	ppbv	99
74) 1,2,4-Trimethylbenzene	11.533	105	22254	0.522	ppbv	91
75) 1,3-Dichlorobenzene	11.604	146	12926	0.520	ppbv	98
76) 1,4-Dichlorobenzene	11.672	146	13076	0.526	ppbv	98
77) 1,2-Dichlorobenzene	11.947	146	12069	0.505	ppbv	97
78) Hexachloro-1,3-Butadiene	13.879	225	10578	0.530	ppbv	98
79) Naphthalene	13.513	128	18377	0.486	ppbv	97
80) Naphthalene,2-methyl-	14.462	142	4917	0.495	ppbv	91
81) 1,2,4-Trichlorobenzene	13.439	180	6798	0.402	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 832
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5




Tgt Ion: 49 Resp: 110490
Ion Ratio Lower Upper
49 100
128 46.6 23.0 69.0
130 58.4 29.9 89.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#2
Dichlorodifluoromethane
Concen: 0.496 ppbv
RT: 1.496 min Scan# 435
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29



Tgt Ion: 85 Resp: 7723
Ion Ratio Lower Upper
85 100
87 32.5 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.557 ppbv
RT: 1.473 min Scan# 428
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 51 Resp: 8538
Ion Ratio Lower Upper
51 100
67 26.4 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#4
Chloromethane
Concen: 0.494 ppbv
RT: 1.531 min Scan# 446
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 50 Resp: 2881
Ion Ratio Lower Upper
50 100
52 33.5 24.1 36.1









#5
Vinyl Chloride
Concen: 0.462 ppbv
RT: 1.580 min Scan# 461
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 62 Resp: 2947
Ion Ratio Lower Upper
62 100
64 26.4 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#6
Bromomethane
Concen: 0.536 ppbv
RT: 1.667 min Scan# 488
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 94 Resp: 1656
Ion Ratio Lower Upper
94 100
96 93.9 72.9 109.3









#7
Chloroethane
Concen: 0.547 ppbv
RT: 1.703 min Scan# 498
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 64 Resp: 135
Ion Ratio Lower Upper
64 100
66 33.7 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#8
Dichlorotetrafluoroethane
Concen: 0.493 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 85 Resp: 6386
Ion Ratio Lower Upper
85 100
135 65.8 52.6 79.0









#9
 Propene
 Concen: 0.526 ppbv
 RT: 1.483 min Scan# 431
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Tgt Ion: 41 Resp: 3739
 Ion Ratio Lower Upper
 41 100
 42 121.9 3.7 5.5




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#10
 Heptane
 Concen: 0.597 ppbv m
 RT: 4.972 min Scan# 1509
 Delta R.T. 0.007 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29


Tgt Ion: 43 Resp: 11637
 Ion Ratio Lower Upper
 43 100
 57 0.0 39.7 59.5





#11
Trichlorofluoromethane
Concen: 0.453 ppbv
RT: 1.878 min Scan# 553
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5




Tgt Ion:101 Resp: 684
Ion Ratio Lower Upper
101 100
103 71.8 52.6 79.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#12
1,1,2-Trichlorotrifluoroethane
Concen: 0.497 ppbv
RT: 2.140 min Scan# 634
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29



Tgt Ion:101 Resp: 5863
Ion Ratio Lower Upper
101 100
151 72.8 59.2 88.8









#13
Ethanol
Concen: 0.520 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 45 Resp: 1190
Ion Ratio Lower Upper
45 100
46 58.7 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#14
Bromoethene
Concen: 0.548 ppbv
RT: 1.781 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:108 Resp: 2476
Ion Ratio Lower Upper
108 100
106 126.4 88.9 133.3
79 58.5 16.5 24.7#









#15
Acetone
Concen: 0.619 ppbv
RT: 1.836 min Scan# 540
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 43 Resp: 9500
Ion Ratio Lower Upper
43 100
58 32.0 20.7 31.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#16
1,3-Butadiene
Concen: 0.601 ppbv
RT: 1.606 min Scan# 469
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 39 Resp: 4115
Ion Ratio Lower Upper
39 100
54 70.0 59.4 89.2









#17
tert-Butyl alcohol
Concen: 0.489 ppbv
RT: 2.043 min Scan# 601
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 59 Resp: 9854
Ion Ratio Lower Upper
59 100
57 23.3 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#18
1,1-Dichloroethene
Concen: 0.559 ppbv
RT: 2.033 min Scan# 601
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 96 Resp: 3068
Ion Ratio Lower Upper
96 100
61 220.0 170.6 255.8
98 66.3 52.2 78.4









#19
Isopropyl Alcohol
Concen: 0.540 ppbv
RT: 1.887 min Scan# 556
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Tgt Ion: 45 Resp: 468
Ion Ratio Lower Upper
45 100
58 9.0 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#20
Methylene Chloride
Concen: 0.349 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29


Tgt Ion: 84 Resp: 5006
Ion Ratio Lower Upper
84 100
49 173.9 132.9 199.3
51 48.8 34.8 52.2
86 70.6 48.2 72.2





#21
 Allyl Chloride
 Concen: 0.609 ppbv
 RT: 2.095 min Scan# 619
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5






Tgt Ion: 41 Resp: 6291


Ion	Ratio	Lower	Upper
41	100		
76	22.2	22.1	33.1
39	73.5	61.0	91.4

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025






#22
 trans-1,2-Dichloroethene
 Concen: 0.533 ppbv
 RT: 2.337 min Scan# 695
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29



Tgt Ion: 96 Resp: 3343


Ion	Ratio	Lower	Upper
96	100		
61	141.8	139.0	208.6
98	51.7	49.4	74.2





#23
 Vinyl Acetate
 Concen: 0.562 ppbv
 RT: 2.464 min Scan# 734
 Delta R.T. 0.006 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5






Tgt Ion: 43 Resp: 10414

Ion	Ratio	Lower	Upper
43	100		
86	10.8	6.2	9.2
42	11.4	8.0	12.0
44	10.7	4.0	6.0

Manual Integrations
 APPROVED




Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#24
 1,1-Dichloroethane
 Concen: 0.479 ppbv
 RT: 2.409 min Scan# 717
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Tgt Ion: 63 Resp: 5803

Ion	Ratio	Lower	Upper
63	100		
98	6.6	2.8	8.4
100	4.2	1.3	3.9





#25

Ethyl Acetate

Concen: 0.562 ppbv m

RT: 2.839 min Scan# 847

Delta R.T. 0.010 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L





ClientSampleId :

VSTDIC0.5

Tgt Ion: 43	Resp: 18625
Ion Ratio	Lower Upper
43 100	
45 11.2	8.1 12.1
61 9.0	7.0 10.4
70 12.2	7.0 10.6

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#26

Hexane

Concen: 0.523 ppbv




RT: 2.823 min Scan# 845


Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29


Tgt Ion: 57	Resp: 7589
Ion Ratio	Lower Upper
57 100	
41 114.6	74.6 111.8
43 229.7	179.0 268.4
56 58.1	42.3 63.5





#27
Carbon Disulfide
Concen: 0.486 ppbv
RT: 2.150 min Scan# 636
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5




Tgt Ion: 76 Resp: 7198
Ion Ratio Lower Upper
76 100
44 115.9 16.1 24.1
78 26.4 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#28
Methyl tert-Butyl Ether
Concen: 0.498 ppbv
RT: 2.441 min Scan# 727
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29



Tgt Ion: 73 Resp: 3545
Ion Ratio Lower Upper
73 100
57 67.2 19.8 29.8





#29

Chloroform

Concen: 0.523 ppbv

RT: 2.849 min Scan# 851

Delta R.T. 0.006 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.5

Tgt Ion: 83 Resp: 10800

Ion Ratio Lower Upper

83 100





85 66.9 50.5 75.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#30

Cyclohexane

Concen: 0.515 ppbv

RT: 3.849 min Scan# 1162

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29




Tgt Ion: 84 Resp: 6271


Ion Ratio Lower Upper

84 100

56 148.3 104.7 157.1

41 99.3 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 0.536 ppbv
 RT: 2.719 min Scan# 81
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5

Tgt Ion: 61 Resp: 7728
 Ion Ratio Lower Upper
 61 100
 96 49.8 49.3 73.9
 98 46.5 32.1 48.1




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#32
 1,1,1-Trichloroethane
 Concen: 0.490 ppbv
 RT: 3.363 min Scan# 1012
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Tgt Ion: 97 Resp: 10886
 Ion Ratio Lower Upper
 97 100
 99 64.4 50.3 75.5
 61 49.5 38.9 58.3





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.956 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.5





Tgt Ion:114 Resp: 296493

Ion	Ratio	Lower	Upper
114	100		
63	25.7	21.0	31.6
88	18.8	14.9	22.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#34

2-Butanone

Concen: 0.527 ppbv

RT: 2.557 min Scan# 763




Delta R.T. 0.006 min


Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Tgt Ion: 43 Resp: 11026

Ion	Ratio	Lower	Upper
43	100		
72	24.9	18.1	27.1





#35

Carbon Tetrachloride

Concen: 0.513 ppbv

RT: 3.761 min Scan# 1135

Delta R.T. 0.006 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC0.5

Tgt Ion:	117	Resp:	1155
Ion Ratio	Lower	Upper	
117	100		
119	99.4	74.6	112.0
121	25.7	23.2	34.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#36

Benzene

Concen: 0.520 ppbv




RT: 3.655 min Scan# 1102


Delta R.T. 0.006 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Tgt Ion:	78	Resp:	14759
Ion Ratio	Lower	Upper	
78	100		
77	24.6	17.5	26.3
50	24.7	15.5	23.3









#37
1,2-Dichloroethane
Concen: 0.510 ppbv
RT: 3.218 min Scan# 965
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 62 Resp: 847
Ion Ratio Lower Upper
62 100
98 9.2 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#38
Trichloroethene
Concen: 0.519 ppbv
RT: 4.542 min Scan# 1376
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:130 Resp: 5775
Ion Ratio Lower Upper
130 100
95 96.0 85.7 128.5
132 72.9 73.4 110.0#
97 59.9 52.7 79.1





#39

1,2-Dichloropropane

Concen: 0.530 ppbv

RT: 4.312 min Scan# 1305

Delta R.T. 0.010 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.5





Tgt Ion: 63	Resp: 5510
Ion Ratio	Lower Upper
63	100
41	98.4 62.6 93.8
62	76.7 51.0 76.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#40

1,4-Dioxane

Concen: 0.573 ppbv m




RT: 4.597 min Scan# 1393


Delta R.T. 0.029 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Tgt Ion: 88	Resp: 2675
Ion Ratio	Lower Upper
88	100
58	0.0 63.0 94.6
43	19.9 32.5 48.7
57	0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 0.536 ppbv
RT: 3.059 min Scan# 914
Delta R.T. 0.013 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion:	42	Resp:	6193
Ion	Ratio	Lower	Upper
42	100		
72	40.2	30.6	46.0
71	42.1	30.5	45.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#42
Bromodichloromethane
Concen: 0.506 ppbv
RT: 4.490 min Scan# 1360
Delta R.T. 0.013 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:	83	Resp:	11597
Ion	Ratio	Lower	Upper
83	100		
85	51.4	51.7	77.5#
127	4.8	6.3	9.5#









#43
Methyl Methacrylate
Concen: 0.566 ppbv m
RT: 4.881 min Scan# 1476
Delta R.T. 0.016 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 69 Resp: 8534
Ion Ratio Lower Upper
69 100
41 0.0 126.7 190.1
100 0.0 28.6 43.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#44
2,2,4-Trimethylpentane
Concen: 0.531 ppbv
RT: 4.642 min Scan# 1407
Delta R.T. 0.016 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 57 Resp: 25890
Ion Ratio Lower Upper
57 100
41 36.2 28.0 42.0
56 36.9 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 0.523 ppbv
RT: 6.412 min Scan# 1954
Delta R.T. 0.013 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 75 Resp: 7343
Ion Ratio Lower Upper
75 100
77 26.9 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#46
cis-1,3-Dichloropropene
Concen: 0.539 ppbv m
RT: 5.593 min Scan# 1701
Delta R.T. 0.010 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 75 Resp: 8886
Ion Ratio Lower Upper
75 100
39 69.0 59.1 88.7
77 33.2 24.2 36.2





#47
1,1,2-Trichloroethane
Concen: 0.567 ppbv m
RT: 6.581 min Scan# 2006
Delta R.T. 0.016 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29





Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 97 Resp: 6280

Ion	Ratio	Lower	Upper
97	100		
83	78.3	64.6	97.0
85	38.7	39.4	59.2
99	45.8	47.0	70.6

Manual Integrations
APPROVED




Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#48
Dibromochloromethane
Concen: 0.492 ppbv
RT: 7.380 min Scan# 2253
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:129 Resp: 9327

Ion	Ratio	Lower	Upper
129	100		
127	80.8	39.3	117.9









#49
Bromoform
Concen: 0.514 ppbv
RT: 9.481 min Scan# 2900
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Tgt Ion:173 Resp: 8636
Ion Ratio Lower Upper
173 100
175 48.4 38.6 58.0
171 50.9 40.6 61.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#50
4-Methyl-2-Pentanone
Concen: 0.515 ppbv
RT: 5.755 min Scan# 1751
Delta R.T. 0.023 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 43 Resp: 15190
Ion Ratio Lower Upper
43 100
58 40.3 28.4 42.6









#51
2-Hexanone
Concen: 0.506 ppbv
RT: 7.442 min Scan# 211594
Delta R.T. 0.023 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Tgt Ion: 43	Resp: 11594
Ion Ratio	Lower Upper
43 100	
58 52.6	38.4 57.6
98 0.7	0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#52
Tetrachloroethene
Concen: 0.483 ppbv
RT: 8.218 min Scan# 2512
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt	Ion:164	Resp:	5097
Ion	Ratio	Lower	Upper
164	100		
166	147.2	99.4	149.2
129	127.2	81.8	122.8#
131	103.9	77.4	116.2









#53
Toluene
Concen: 0.525 ppbv m
RT: 6.924 min Scan# 2111
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 91 Resp: 1775
Ion Ratio Lower Upper
91 100
92 0.0 47.4 71.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#54
1,2-Dibromoethane
Concen: 0.537 ppbv
RT: 7.649 min Scan# 2336
Delta R.T. 0.010 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:107 Resp: 8708
Ion Ratio Lower Upper
107 100
109 80.8 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.875 min Scan# 2715
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Tgt Ion:117 Resp: 24776
Ion Ratio Lower Upper
117 100
82 65.5 56.1 84.1
119 31.6 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#56
1,1,1,2-Tetrachloroethane
Concen: 0.538 ppbv m
RT: 8.917 min Scan# 2728
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:131 Resp: 9793
Ion Ratio Lower Upper
131 100
133 98.4 77.4 116.2
119 50.6 49.1 73.7









#57
Chlorobenzene
Concen: 0.504 ppbv
RT: 8.917 min Scan# 2726
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion:112 Resp: 12818
Ion Ratio Lower Upper
112 100
77 69.0 55.5 83.3
114 29.2 29.1 35.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#58
Ethyl Benzene
Concen: 0.500 ppbv
RT: 9.351 min Scan# 2862
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 91 Resp: 23866
Ion Ratio Lower Upper
91 100
106 30.7 22.8 34.2









#59
m/p-Xylene
Concen: 1.044 ppbv m
RT: 9.526 min Scan# 2916
Delta R.T. -0.016 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 91 Resp: 40076
Ion Ratio Lower Upper
91 100
106 46.0 36.1 54.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#60
o-Xylene
Concen: 0.514 ppbv
RT: 9.956 min Scan# 3049
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 91 Resp: 19621
Ion Ratio Lower Upper
91 100
106 42.0 20.9 62.8









#61
Styrene
Concen: 0.496 ppbv
RT: 9.869 min Scan# 3020
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion:104 Resp: 828
Ion Ratio Lower Upper
104 100
78 56.3 42.1 63.1
103 47.8 37.2 55.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#62
Isopropylbenzene
Concen: 0.509 ppbv
RT: 10.536 min Scan# 3228
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:105 Resp: 34773
Ion Ratio Lower Upper
105 100
120 25.6 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 0.519 ppbv
RT: 9.960 min Scan# 3048
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5

Tgt Ion: 83 Resp: 1259
Ion Ratio Lower Upper
83 100
131 9.8 4.8 14.3
85 64.0 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#64
n-propylbenzene
Concen: 0.509 ppbv
RT: 10.986 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:120 Resp: 8945
Ion Ratio Lower Upper
120 100
91 472.3 377.1 565.7









#65
tert-Butylbenzene
Concen: 0.524 ppbv
RT: 11.526 min Scan# 3533
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion:	119	Resp:	31998
Ion Ratio	Lower	Upper	
119	100		
91	81.6	65.3	97.9
134	23.4	17.0	25.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#66
Benzyl Chloride
Concen: 0.550 ppbv
RT: 11.610 min Scan# 3560
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:	91	Resp:	3403
Ion Ratio	Lower	Upper	
91	100		
126	16.8	14.4	21.6
128	6.1	4.6	6.8





#67

sec-Butylbenzene

Concen: 0.522 ppbv

RT: 11.766 min Scan# 3607

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC0.5

Tgt Ion:105 Resp: 44448

Ion Ratio Lower Upper

105 100





134 20.0 15.0 22.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#68

1-Bromo-4-Fluorobenzene

Concen: 9.941 ppbv

RT: 10.371 min Scan# 3177

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29




Tgt Ion: 95 Resp: 180529


Ion Ratio Lower Upper

95 100

174 63.2 50.2 75.2

175 4.9 4.1 6.1









#69
p-Isopropyltoluene
Concen: 0.514 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 119 Resp: 3708
Ion Ratio Lower Upper
119 100
134 25.4 20.2 30.2
91 29.9 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#70
n-Butylbenzene
Concen: 0.474 ppbv
RT: 12.277 min Scan# 3766
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 91 Resp: 33153
Ion Ratio Lower Upper
91 100
92 59.3 43.5 65.3
134 26.3 19.0 28.6









#71
2-Chlorotoluene
Concen: 0.518 ppbv
RT: 10.895 min Scan# 3339
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion: 91 Resp: 27224
Ion Ratio Lower Upper
91 100
126 28.6 11.7 46.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#72
4-Ethyltoluene
Concen: 0.515 ppbv
RT: 11.122 min Scan# 3409
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29


Tgt Ion:105 Resp: 23963
Ion Ratio Lower Upper
105 100
120 31.3 23.8 35.8
91 14.6 10.3 15.5
77 15.6 11.1 16.7





#73
1,3,5-Trimethylbenzene
Concen: 0.515 ppbv
RT: 11.199 min Scan# 3432
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5




Tgt Ion:105 Resp: 20104
Ion Ratio Lower Upper
105 100
120 48.8 38.6 57.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#74
1,2,4-Trimethylbenzene
Concen: 0.522 ppbv
RT: 11.533 min Scan# 3536
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29



Tgt Ion:105 Resp: 22254
Ion Ratio Lower Upper
105 100
120 46.6 41.7 62.5
91 77.8 70.0 105.0
77 25.6 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 0.520 ppbv
RT: 11.604 min Scan# 3557
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion:146 Resp: 12926
Ion Ratio Lower Upper
146 100
111 45.1 23.4 70.3
148 64.7 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#76
1,4-Dichlorobenzene
Concen: 0.526 ppbv
RT: 11.672 min Scan# 3579
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:146 Resp: 13076
Ion Ratio Lower Upper
146 100
111 42.1 22.8 68.3
148 63.8 31.8 95.4









#77
1,2-Dichlorobenzene
Concen: 0.505 ppbv
RT: 11.947 min Scan# 3662
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Tgt Ion:146 Resp: 12069
Ion Ratio Lower Upper
146 100
111 49.5 24.3 73.0
148 66.3 31.6 94.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#78
Hexachloro-1,3-Butadiene
Concen: 0.530 ppbv
RT: 13.879 min Scan# 4261
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:225 Resp: 10578
Ion Ratio Lower Upper
225 100
223 62.7 50.2 75.4
227 61.0 51.1 76.7









#79
Naphthalene
Concen: 0.486 ppbv
RT: 13.513 min Scan# 4146
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5

Tgt Ion:	128	Resp:	1837
Ion Ratio	Lower	Upper	
128	100		
127	11.2	9.9	14.9
129	11.5	8.4	12.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#80
Naphthalene,2-methyl-
Concen: 0.495 ppbv
RT: 14.462 min Scan# 4441
Delta R.T. 0.010 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29


Tgt Ion:	142	Resp:	4917
Ion Ratio	Lower	Upper	
142	100		
141	90.2	67.1	100.7
115	47.3	31.9	47.9





#81
1,2,4-Trichlorobenzene
Concen: 0.402 ppbv
RT: 13.439 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29



Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5



Tgt Ion:180 Resp: 6798
Ion Ratio Lower Upper
180 100
182 95.7 77.4 116.2
184 34.3 24.2 36.4

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043204.D
 Acq On : 18 Nov 2025 10:02
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Quant Time: Nov 19 00:49:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	105095	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	283830	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	234772	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.367 95 172524 10.025 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	140789	9.498	ppbv	100
3) Chlorodifluoromethane	1.473	51	138443	9.497	ppbv	100
4) Chloromethane	1.531	50	52372	9.450	ppbv	100
5) Vinyl Chloride	1.577	62	51045	8.422	ppbv	100
6) Bromomethane	1.664	94	26542	9.034	ppbv	100
7) Chloroethane	1.700	64	20375	8.632	ppbv	100
8) Dichlorotetrafluoroethane	1.551	85	114670	9.300	ppbv	100
9) Propene	1.480	41	52671	7.789	ppbv #	1
10) Heptane	4.965	43	168611	9.090	ppbv	100
11) Trichlorofluoromethane	1.875	101	130362	9.077	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.137	101	103679	9.247	ppbv	100
13) Ethanol	1.716	45	17464	8.023	ppbv	100
14) Bromoethene	1.777	108	38037	8.849	ppbv	100
15) Acetone	1.832	43	108422	7.426	ppbv	100
16) 1,3-Butadiene	1.606	39	57871	8.890	ppbv	100
17) tert-Butyl alcohol	2.036	59	179201	9.350	ppbv	100
18) 1,1-Dichloroethene	2.030	96	46313	8.871	ppbv	100
19) Isopropyl Alcohol	1.881	45	71862	8.710	ppbv	100
20) Methylene Chloride	2.056	84	41868	9.708	ppbv	100
21) Allyl Chloride	2.091	41	89266	9.087	ppbv	100
22) trans-1,2-Dichloroethene	2.334	96	52470	8.793	ppbv	100
23) Vinyl Acetate	2.457	43	170459	9.670	ppbv	100
24) 1,1-Dichloroethane	2.405	63	106970	9.287	ppbv	100
25) Ethyl Acetate	2.829	43	289921	9.190	ppbv	100
26) Hexane	2.820	57	129809	9.402	ppbv	100
27) Carbon Disulfide	2.146	76	132086	9.371	ppbv	100
28) Methyl tert-Butyl Ether	2.434	73	62115	9.165	ppbv	100
29) Chloroform	2.842	83	187620	9.561	ppbv	100
30) Cyclohexane	3.846	84	107621	9.297	ppbv	100
31) cis-1,2-Dichloroethene	2.716	61	130357	9.499	ppbv	100
32) 1,1,1-Trichloroethane	3.360	97	198036	9.381	ppbv	100
34) 2-Butanone	2.551	43	182749	9.121	ppbv	100
35) Carbon Tetrachloride	3.755	117	200831	9.322	ppbv	100
36) Benzene	3.648	78	256954	9.453	ppbv	100
37) 1,2-Dichloroethane	3.211	62	150769	9.483	ppbv	100
38) Trichloroethene	4.535	130	100136	9.398	ppbv	100
39) 1,2-Dichloropropane	4.302	63	95097	9.559	ppbv	100
40) 1,4-Dioxane	4.567	88	40441	9.043	ppbv #	100
41) Tetrahydrofuran	3.046	42	104294	9.426	ppbv	100
42) Bromodichloromethane	4.477	83	211504	9.643	ppbv	100
43) Methyl Methacrylate	4.865	69	131003	9.080	ppbv	100
44) 2,2,4-Trimethylpentane	4.626	57	441234	9.451	ppbv	100
45) t-1,3-Dichloropropene	6.399	75	129122	9.601	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043204.D
 Acq On : 18 Nov 2025 10:02
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Quant Time: Nov 19 00:49:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025


Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.584	75	151655	9.603	ppbv	100
47) 1,1,2-Trichloroethane	6.564	97	97512	9.193	ppbv	100
48) Dibromochloromethane	7.377	129	175529	9.670	ppbv	100
49) Bromoform	9.474	173	157563	9.800	ppbv	100
50) 4-Methyl-2-Pentanone	5.733	43	268369	9.509	ppbv	100
51) 2-Hexanone	7.419	43	208268	9.503	ppbv #	100
52) Tetrachloroethene	8.212	164	93518	9.249	ppbv	100
53) Toluene	6.920	91	305169	9.422	ppbv	100
54) 1,2-Dibromoethane	7.639	107	148973	9.588	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.914	131	158856	9.207	ppbv	100
57) Chlorobenzene	8.911	112	227490	9.438	ppbv	100
58) Ethyl Benzene	9.345	91	427530	9.461	ppbv	100
59) m/p-Xylene	9.542	91	673371m	18.519	ppbv	
60) o-Xylene	9.956	91	333044	9.202	ppbv	100
61) Styrene	9.863	104	152172	9.607	ppbv	100
62) Isopropylbenzene	10.533	105	600155	9.269	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.953	83	209272	9.109	ppbv	100
64) n-propylbenzene	10.982	120	154239	9.271	ppbv	100
65) tert-Butylbenzene	11.523	119	522106	9.029	ppbv	100
66) Benzyl Chloride	11.607	91	56620	9.649	ppbv	100
67) sec-Butylbenzene	11.762	105	729573	9.039	ppbv	100
69) p-Isopropyltoluene	11.918	119	624586	9.135	ppbv	100
70) n-Butylbenzene	12.274	91	619610	9.341	ppbv	100
71) 2-Chlorotoluene	10.895	91	460118	9.245	ppbv	100
72) 4-Ethyltoluene	11.118	105	410587	9.316	ppbv	100
73) 1,3,5-Trimethylbenzene	11.196	105	341900	9.244	ppbv	100
74) 1,2,4-Trimethylbenzene	11.529	105	363476	9.005	ppbv	100
75) 1,3-Dichlorobenzene	11.601	146	219894	9.335	ppbv	100
76) 1,4-Dichlorobenzene	11.665	146	220576	9.363	ppbv	100
77) 1,2-Dichlorobenzene	11.940	146	204826	9.044	ppbv	100
78) Hexachloro-1,3-Butadiene	13.876	225	161034	8.520	ppbv	100
79) Naphthalene	13.507	128	351148	9.795	ppbv	100
80) Naphthalene,2-methyl-	14.452	142	93113	9.895	ppbv	100
81) 1,2,4-Trichlorobenzene	13.436	180	159091	9.923	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 832
Delta R.T. -0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 49 Resp: 105095
Ion Ratio Lower Upper
49 100
128 46.0 23.0 69.0
130 59.8 29.9 89.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#2
Dichlorodifluoromethane
Concen: 9.498 ppbv
RT: 1.496 min Scan# 435
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02



Tgt Ion: 85 Resp: 140789
Ion Ratio Lower Upper
85 100
87 31.8 25.4 38.2









#3
Chlorodifluoromethane
Concen: 9.497 ppbv
RT: 1.473 min Scan# 428
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 51 Resp: 138443
Ion Ratio Lower Upper
51 100
67 18.6 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#4
Chloromethane
Concen: 9.450 ppbv
RT: 1.531 min Scan# 446
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion: 50 Resp: 52372
Ion Ratio Lower Upper
50 100
52 30.1 24.1 36.1





#5
 Vinyl Chloride
 Concen: 8.422 ppbv
 RT: 1.577 min Scan# 460
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICCC010




Tgt Ion: 62 Resp: 51045
 Ion Ratio Lower Upper
 62 100
 64 34.1 27.3 40.9

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#6
 Bromomethane
 Concen: 9.034 ppbv
 RT: 1.664 min Scan# 487
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Tgt Ion: 94 Resp: 26542
 Ion Ratio Lower Upper
 94 100
 96 91.1 72.9 109.3





#7
Chloroethane
Concen: 8.632 ppbv
RT: 1.700 min Scan# 498
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 64 Resp: 20375
Ion Ratio Lower Upper
64 100
66 33.5 26.8 40.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#8
Dichlorotetrafluoroethane
Concen: 9.300 ppbv
RT: 1.551 min Scan# 452
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02



Tgt Ion: 85 Resp: 114670
Ion Ratio Lower Upper
85 100
135 65.8 52.6 79.0









#9
Propene
Concen: 7.789 ppbv
RT: 1.480 min Scan# 430
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 41 Resp: 5267
Ion Ratio Lower Upper
41 100
42 105.2 3.7 5.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#10
Heptane
Concen: 9.090 ppbv
RT: 4.965 min Scan# 1507
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion: 43 Resp: 168611
Ion Ratio Lower Upper
43 100
57 49.6 39.7 59.5





#11
 Trichlorofluoromethane
 Concen: 9.077 ppbv
 RT: 1.875 min Scan# 552
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010




Tgt Ion:101 Resp: 130362
 Ion Ratio Lower Upper
 101 100
 103 65.8 52.6 79.0

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 9.247 ppbv
 RT: 2.137 min Scan# 633
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02



Tgt Ion:101 Resp: 103679
 Ion Ratio Lower Upper
 101 100
 151 74.0 59.2 88.8









#13
Ethanol
Concen: 8.023 ppbv
RT: 1.716 min Scan# 503
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 45 Resp: 17464
Ion Ratio Lower Upper
45 100
46 37.4 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#14
Bromoethene
Concen: 8.849 ppbv
RT: 1.777 min Scan# 522
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion:108 Resp: 38037
Ion Ratio Lower Upper
108 100
106 111.1 88.9 133.3
79 20.6 16.5 24.7





#15
Acetone
Concen: 7.426 ppbv
RT: 1.832 min Scan# 539
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 43 Resp: 108422
Ion Ratio Lower Upper
43 100
58 25.9 20.7 31.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#16
1,3-Butadiene
Concen: 8.890 ppbv
RT: 1.606 min Scan# 469
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02



Tgt Ion: 39 Resp: 57871
Ion Ratio Lower Upper
39 100
54 74.3 59.4 89.2









#17
tert-Butyl alcohol
Concen: 9.350 ppbv
RT: 2.036 min Scan# 602
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 59 Resp: 17920
Ion Ratio Lower Upper
59 100
57 11.2 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#18
1,1-Dichloroethene
Concen: 8.871 ppbv
RT: 2.030 min Scan# 600
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 96 Resp: 46313
Ion Ratio Lower Upper
96 100
61 213.2 170.6 255.8
98 65.3 52.2 78.4









#19
Isopropyl Alcohol
Concen: 8.710 ppbv
RT: 1.881 min Scan# 554
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 45 Resp: 71862
Ion Ratio Lower Upper
45 100
58 39.1 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#20
Methylene Chloride
Concen: 9.708 ppbv
RT: 2.056 min Scan# 608
Delta R.T. -0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion: 84 Resp: 41868
Ion Ratio Lower Upper
84 100
49 166.1 132.9 199.3
51 43.5 34.8 52.2
86 60.2 48.2 72.2





#21
 Allyl Chloride
 Concen: 9.087 ppbv
 RT: 2.091 min Scan# 619
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010




Tgt Ion: 41 Resp: 89266
 Ion Ratio Lower Upper
 41 100
 76 27.6 22.1 33.1
 39 76.2 61.0 91.4

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#22
 trans-1,2-Dichloroethene
 Concen: 8.793 ppbv
 RT: 2.334 min Scan# 694
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02



Tgt Ion: 96 Resp: 52470
 Ion Ratio Lower Upper
 96 100
 61 173.8 139.0 208.6
 98 61.8 49.4 74.2









#23
Vinyl Acetate
Concen: 9.670 ppbv
RT: 2.457 min Scan# 732
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010

Tgt Ion: 43	Resp: 170459
Ion Ratio	Lower Upper
43 100	
86 7.7	6.2 9.2
42 10.0	8.0 12.0
44 5.0	4.0 6.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#24
1,1-Dichloroethane
Concen: 9.287 ppbv
RT: 2.405 min Scan# 716
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 63	Resp: 106970
Ion Ratio	Lower Upper
63 100	
98 5.6	2.8 8.4
100 2.6	1.3 3.9









#25
Ethyl Acetate
Concen: 9.190 ppbv
RT: 2.829 min Scan# 847
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 43	Resp: 28992
Ion Ratio	Lower Upper
43 100	
45 10.1	8.1 12.1
61 8.7	7.0 10.4
70 8.8	7.0 10.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#26
Hexane
Concen: 9.402 ppbv
RT: 2.820 min Scan# 844
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion: 57	Resp: 129809
Ion Ratio	Lower Upper
57 100	
41 93.2	74.6 111.8
43 223.7	179.0 268.4
56 52.9	42.3 63.5





#27
Carbon Disulfide
Concen: 9.371 ppbv
RT: 2.146 min Scan# 636
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 76 Resp: 132086
Ion Ratio Lower Upper
76 100
44 20.1 16.1 24.1
78 12.0 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#28
Methyl tert-Butyl Ether
Concen: 9.165 ppbv
RT: 2.434 min Scan# 725
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02



Tgt Ion: 73 Resp: 62115
Ion Ratio Lower Upper
73 100
57 24.8 19.8 29.8









#29
Chloroform
Concen: 9.561 ppbv
RT: 2.842 min Scan# 851
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
Client Sample Id :
VSTDICCC010

Tgt Ion: 83 Resp: 187620
Ion Ratio Lower Upper
83 100
85 63.1 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#30
Cyclohexane
Concen: 9.297 ppbv
RT: 3.846 min Scan# 1161
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 84 Resp: 107621
Ion Ratio Lower Upper
84 100
56 130.9 104.7 157.1
41 86.3 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 9.499 ppbv
 RT: 2.716 min Scan# 812
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Tgt Ion: 61 Resp: 130352
 Ion Ratio Lower Upper
 61 100
 96 61.6 49.3 73.9
 98 40.1 32.1 48.1




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#32
 1,1,1-Trichloroethane
 Concen: 9.381 ppbv
 RT: 3.360 min Scan# 1011
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02


Tgt Ion: 97 Resp: 198036
 Ion Ratio Lower Upper
 97 100
 99 62.9 50.3 75.5
 61 48.6 38.9 58.3





#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.949 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion:114 Resp: 283830
Ion Ratio Lower Upper
114 100
63 26.3 21.0 31.6
88 18.6 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#34
2-Butanone
Concen: 9.121 ppbv
RT: 2.551 min Scan# 761
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 43 Resp: 182749
Ion Ratio Lower Upper
43 100
72 22.6 18.1 27.1





#35

Carbon Tetrachloride

Concen: 9.322 ppbv

RT: 3.755 min Scan# 1133

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion:117 Resp: 20083

Ion Ratio Lower Upper

117 100

119 93.3 74.6 112.0





121 29.0 23.2 34.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#36

Benzene

Concen: 9.453 ppbv

RT: 3.648 min Scan# 1100

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02




Tgt Ion: 78 Resp: 256954


Ion Ratio Lower Upper

78 100

77 21.9 17.5 26.3

50 19.4 15.5 23.3









#37
1,2-Dichloroethane
Concen: 9.483 ppbv
RT: 3.211 min Scan# 965
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 62 Resp: 150769
Ion Ratio Lower Upper
62 100
98 6.9 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#38
Trichloroethene
Concen: 9.398 ppbv
RT: 4.535 min Scan# 1374
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion:130 Resp: 100136
Ion Ratio Lower Upper
130 100
95 107.1 85.7 128.5
132 91.7 73.4 110.0
97 65.9 52.7 79.1





#39
1,2-Dichloropropane
Concen: 9.559 ppbv
RT: 4.302 min Scan# 1302
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 63 Resp: 95092
Ion Ratio Lower Upper
63 100
41 78.2 62.6 93.8
62 63.8 51.0 76.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#40
1,4-Dioxane
Concen: 9.043 ppbv
RT: 4.567 min Scan# 1384
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02



Tgt Ion: 88 Resp: 40441
Ion Ratio Lower Upper
88 100
58 78.8 63.0 94.6
43 40.6 32.5 48.7
57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 9.426 ppbv
RT: 3.046 min Scan# 914
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion:	42	Resp:	104294
Ion Ratio	Lower	Upper	
42	100		
72	38.3	30.6	46.0
71	38.1	30.5	45.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#42
Bromodichloromethane
Concen: 9.643 ppbv
RT: 4.477 min Scan# 1356
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:	83	Resp:	211504
Ion Ratio	Lower	Upper	
83	100		
85	64.6	51.7	77.5
127	7.9	6.3	9.5









#43
Methyl Methacrylate
Concen: 9.080 ppbv
RT: 4.865 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 69 Resp: 131003
Ion Ratio Lower Upper
69 100
41 158.4 126.7 190.1
100 35.8 28.6 43.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#44
2,2,4-Trimethylpentane
Concen: 9.451 ppbv
RT: 4.626 min Scan# 1402
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 57 Resp: 441234
Ion Ratio Lower Upper
57 100
41 35.0 28.0 42.0
56 32.0 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 9.601 ppbv
RT: 6.399 min Scan# 1950
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010

Tgt Ion: 75 Resp: 12912
Ion Ratio Lower Upper
75 100
77 30.4 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#46
cis-1,3-Dichloropropene
Concen: 9.603 ppbv
RT: 5.584 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 75 Resp: 151655
Ion Ratio Lower Upper
75 100
39 73.9 59.1 88.7
77 30.2 24.2 36.2









#47
1,1,2-Trichloroethane
Concen: 9.193 ppbv
RT: 6.564 min Scan# 2001
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion	Ratio	Lower	Upper
97	100		
83	80.8	64.6	97.0
85	49.3	39.4	59.2
99	58.8	47.0	70.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#48
Dibromochloromethane
Concen: 9.670 ppbv
RT: 7.377 min Scan# 2252
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion	Ratio	Lower	Upper
129	100		
127	78.6	39.3	117.9





#49
Bromoform
Concen: 9.800 ppbv
RT: 9.474 min Scan# 2900
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 173 Resp: 157563
Ion Ratio Lower Upper
173 100
175 48.3 38.6 58.0
171 50.8 40.6 61.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#50
4-Methyl-2-Pentanone
Concen: 9.509 ppbv
RT: 5.733 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Tgt Ion: 43 Resp: 268369
Ion Ratio Lower Upper
43 100
58 35.5 28.4 42.6





#51
2-Hexanone
Concen: 9.503 ppbv
RT: 7.419 min Scan# 2265
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion: 43 Resp: 208268
Ion Ratio Lower Upper
43 100
58 48.0 38.4 57.6
98 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#52
Tetrachloroethene
Concen: 9.249 ppbv
RT: 8.212 min Scan# 2510
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:164 Resp: 93518
Ion Ratio Lower Upper
164 100
166 124.3 99.4 149.2
129 102.3 81.8 122.8
131 96.8 77.4 116.2









#53
Toluene
Concen: 9.422 ppbv
RT: 6.920 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010

Tgt Ion: 91 Resp: 305169
Ion Ratio Lower Upper
91 100
92 59.2 47.4 71.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#54
1,2-Dibromoethane
Concen: 9.588 ppbv
RT: 7.639 min Scan# 2333
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:107 Resp: 148973
Ion Ratio Lower Upper
107 100
109 90.1 72.1 108.1





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.872 min Scan# 2714

Delta R.T. 0.000 min

Lab File: VL043204.D


Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010






Tgt Ion:117 Resp: 234772

Ion	Ratio	Lower	Upper
117	100		
82	70.1	56.1	84.1
119	31.3	25.0	37.6

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#56

1,1,1,2-Tetrachloroethane

Concen: 9.207 ppbv

RT: 8.914 min Scan# 2727




Delta R.T. 0.000 min


Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Tgt Ion:131 Resp: 158856


Ion	Ratio	Lower	Upper
131	100		
133	96.8	77.4	116.2
119	61.4	49.1	73.7





#57
Chlorobenzene
Concen: 9.438 ppbv
RT: 8.911 min Scan# 2726
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010




Tgt Ion:112 Resp: 227490
Ion Ratio Lower Upper
112 100
77 69.4 55.5 83.3
114 32.3 29.1 35.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#58
Ethyl Benzene
Concen: 9.461 ppbv
RT: 9.345 min Scan# 2860
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02



Tgt Ion: 91 Resp: 427530
Ion Ratio Lower Upper
91 100
106 28.5 22.8 34.2









#59
m/p-Xylene
Concen: 18.519 ppbv m
RT: 9.542 min Scan# 2921
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 91 Resp: 67337
Ion Ratio Lower Upper
91 100
106 45.1 36.1 54.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#60
o-Xylene
Concen: 9.202 ppbv
RT: 9.956 min Scan# 3049
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 91 Resp: 333044
Ion Ratio Lower Upper
91 100
106 41.9 20.9 62.8









#61
Styrene
Concen: 9.607 ppbv
RT: 9.863 min Scan# 3020
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion:104 Resp: 152172
Ion Ratio Lower Upper
104 100
78 52.6 42.1 63.1
103 46.5 37.2 55.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#62
Isopropylbenzene
Concen: 9.269 ppbv
RT: 10.533 min Scan# 3227
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:105 Resp: 600155
Ion Ratio Lower Upper
105 100
120 24.8 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 9.109 ppbv
RT: 9.953 min Scan# 3048
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 83 Resp: 209272
Ion Ratio Lower Upper
83 100
131 9.5 4.8 14.3
85 63.9 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#64
n-propylbenzene
Concen: 9.271 ppbv
RT: 10.982 min Scan# 3366
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:120 Resp: 154239
Ion Ratio Lower Upper
120 100
91 471.4 377.1 565.7









#65
tert-Butylbenzene
Concen: 9.029 ppbv
RT: 11.523 min Scan# 3533
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion:119 Resp: 522106
Ion Ratio Lower Upper
119 100
91 81.6 65.3 97.9
134 21.2 17.0 25.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#66
Benzyl Chloride
Concen: 9.649 ppbv
RT: 11.607 min Scan# 3559
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion: 91 Resp: 56620
Ion Ratio Lower Upper
91 100
126 18.0 14.4 21.6
128 5.7 4.6 6.8





#67
 sec-Butylbenzene
 Concen: 9.039 ppbv
 RT: 11.762 min Scan# 3607
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICCC010




Tgt Ion:105 Resp: 729573
 Ion Ratio Lower Upper
 105 100
 134 18.8 15.0 22.6

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.025 ppbv
 RT: 10.367 min Scan# 3176
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Tgt Ion: 95 Resp: 172524
 Ion Ratio Lower Upper
 95 100
 174 62.7 50.2 75.2
 175 5.1 4.1 6.1





#69
p-Isopropyltoluene
Concen: 9.135 ppbv
RT: 11.918 min Scan# 3655
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 119 Resp: 624586
Ion Ratio Lower Upper
119 100
134 25.2 20.2 30.2
91 29.8 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#70
n-Butylbenzene
Concen: 9.341 ppbv
RT: 12.274 min Scan# 3765
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02



Tgt Ion: 91 Resp: 619610
Ion Ratio Lower Upper
91 100
92 54.4 43.5 65.3
134 23.8 19.0 28.6









#71
2-Chlorotoluene
Concen: 9.245 ppbv
RT: 10.895 min Scan# 3339
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 91 Resp: 460118
Ion Ratio Lower Upper
91 100
126 29.2 11.7 46.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#72
4-Ethyltoluene
Concen: 9.316 ppbv
RT: 11.118 min Scan# 3408
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:105 Resp: 410587
Ion Ratio Lower Upper
105 100
120 29.8 23.8 35.8
91 12.9 10.3 15.5
77 13.9 11.1 16.7









#73
1,3,5-Trimethylbenzene
Concen: 9.244 ppbv
RT: 11.196 min Scan# 3432
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion:105 Resp: 341900
Ion Ratio Lower Upper
105 100
120 48.2 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#74
1,2,4-Trimethylbenzene
Concen: 9.005 ppbv
RT: 11.529 min Scan# 3535
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:105 Resp: 363476
Ion Ratio Lower Upper
105 100
120 52.1 41.7 62.5
91 87.5 70.0 105.0
77 29.8 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 9.335 ppbv
RT: 11.601 min Scan# 3557
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010

Tgt Ion:146 Resp: 219894
Ion Ratio Lower Upper
146 100
111 46.9 23.4 70.3
148 63.3 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#76
1,4-Dichlorobenzene
Concen: 9.363 ppbv
RT: 11.665 min Scan# 3577
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:146 Resp: 220576
Ion Ratio Lower Upper
146 100
111 45.5 22.8 68.3
148 63.6 31.8 95.4





#77

1,2-Dichlorobenzene

Concen: 9.044 ppbv

RT: 11.940 min Scan# 3662

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_L





ClientSampleId :

VSTDICCC010

Tgt Ion:	146	Resp:	204826
Ion Ratio	Lower	Upper	
146	100		
111	48.7	24.3	73.0
148	63.1	31.6	94.7

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#78

Hexachloro-1,3-Butadiene

Concen: 8.520 ppbv




RT: 13.876 min Scan# 4260


Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Tgt Ion:	225	Resp:	161034
Ion Ratio	Lower	Upper	
225	100		
223	62.8	50.2	75.4
227	63.9	51.1	76.7









#79
Naphthalene
Concen: 9.795 ppbv
RT: 13.507 min Scan# 4146
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion:128 Resp: 351148
Ion Ratio Lower Upper
128 100
127 12.4 9.9 14.9
129 10.5 8.4 12.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#80
Naphthalene,2-methyl-
Concen: 9.895 ppbv
RT: 14.452 min Scan# 4438
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02


Tgt Ion:142 Resp: 93113
Ion Ratio Lower Upper
142 100
141 83.9 67.1 100.7
115 39.9 31.9 47.9





#81
1,2,4-Trichlorobenzene
Concen: 9.923 ppbv
RT: 13.436 min Scan# 41
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02



Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion:180 Resp: 159091
Ion Ratio Lower Upper
180 100
182 96.8 77.4 116.2
184 30.3 24.2 36.4

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043205.D
 Acq On : 18 Nov 2025 10:37
 Operator : SY/MD
 Sample : VSTDIC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002

Quant Time: Nov 19 00:50:19 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.787	49	104644	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	284629	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	230646	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	171009	10.115	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.499	85	30881	2.092	ppbv	98
3) Chlorodifluoromethane	1.476	51	30069	2.072	ppbv	98
4) Chloromethane	1.534	50	11890	2.155	ppbv	96
5) Vinyl Chloride	1.580	62	12040	1.995	ppbv	96
6) Bromomethane	1.670	94	5981	2.044	ppbv	99
7) Chloroethane	1.706	64	4943	2.103	ppbv	95
8) Dichlorotetrafluoroethane	1.554	85	26409	2.151	ppbv	98
9) Propene	1.486	41	12993	1.930	ppbv #	1
10) Heptane	4.972	43	36171m	1.958	ppbv	
11) Trichlorofluoromethane	1.878	101	31124	2.177	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	23336	2.090	ppbv	100
13) Ethanol	1.722	45	5266	2.430	ppbv	86
14) Bromoethene	1.784	108	9129	2.133	ppbv	95
15) Acetone	1.835	43	31785	2.186	ppbv	91
16) 1,3-Butadiene	1.609	39	13101	2.021	ppbv	89
17) tert-Butyl alcohol	2.039	59	39348	2.062	ppbv	95
18) 1,1-Dichloroethene	2.036	96	10780	2.074	ppbv	94
19) Isopropyl Alcohol	1.887	45	17229	2.097	ppbv #	71
20) Methylene Chloride	2.062	84	12238	2.248	ppbv	96
21) Allyl Chloride	2.098	41	19580	2.002	ppbv	95
22) trans-1,2-Dichloroethene	2.340	96	12537	2.110	ppbv	94
23) Vinyl Acetate	2.460	43	35205	2.006	ppbv #	97
24) 1,1-Dichloroethane	2.408	63	24987	2.179	ppbv	99
25) Ethyl Acetate	2.832	43	64343	2.048	ppbv	99
26) Hexane	2.826	57	28776	2.093	ppbv	97
27) Carbon Disulfide	2.149	76	29964	2.135	ppbv #	91
28) Methyl tert-Butyl Ether	2.438	73	14101	2.090	ppbv #	85
29) Chloroform	2.849	83	40355	2.065	ppbv	98
30) Cyclohexane	3.855	84	23659	2.053	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	27663	2.024	ppbv	98
32) 1,1,1-Trichloroethane	3.366	97	42547	2.024	ppbv	99
34) 2-Butanone	2.557	43	41884	2.085	ppbv	99
35) Carbon Tetrachloride	3.765	117	41949	1.942	ppbv	96
36) Benzene	3.654	78	56141m	2.060	ppbv	
37) 1,2-Dichloroethane	3.218	62	32188	2.019	ppbv	100
38) Trichloroethene	4.548	130	21105m	1.975	ppbv	
39) 1,2-Dichloropropane	4.312	63	19913	1.996	ppbv	93
40) 1,4-Dioxane	4.580	88	8829m	1.969	ppbv	
41) Tetrahydrofuran	3.056	42	22912	2.065	ppbv	99
42) Bromodichloromethane	4.490	83	44584	2.027	ppbv	98
43) Methyl Methacrylate	4.875	69	29071	2.009	ppbv	99
44) 2,2,4-Trimethylpentane	4.635	57	92261	1.971	ppbv	96
45) t-1,3-Dichloropropene	6.406	75	27833m	2.064	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043205.D
 Acq On : 18 Nov 2025 10:37
 Operator : SY/MD
 Sample : VSTDIC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Quant Time: Nov 19 00:50:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	31986	2.020	ppbv	91
47) 1,1,2-Trichloroethane	6.577	97	21542m	2.025	ppbv	
48) Dibromochloromethane	7.386	129	37758m	2.074	ppbv	
49) Bromoform	9.480	173	33057	2.050	ppbv	98
50) 4-Methyl-2-Pentanone	5.745	43	58461	2.066	ppbv	98
51) 2-Hexanone	7.432	43	46444	2.113	ppbv #	99
52) Tetrachloroethene	8.218	164	19856	1.958	ppbv	96
53) Toluene	6.927	91	65238	2.008	ppbv	97
54) 1,2-Dibromoethane	7.648	107	31266	2.007	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.920	131	35099	2.071	ppbv	99
57) Chlorobenzene	8.917	112	50228	2.121	ppbv	97
58) Ethyl Benzene	9.351	91	92649	2.087	ppbv	98
59) m/p-Xylene	9.529	91	148300m	4.151	ppbv	
60) o-Xylene	9.959	91	76540	2.153	ppbv	100
61) Styrene	9.866	104	33188	2.133	ppbv	98
62) Isopropylbenzene	10.536	105	135853	2.136	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.963	83	47439	2.102	ppbv	99
64) n-propylbenzene	10.985	120	34915	2.136	ppbv	99
65) tert-Butylbenzene	11.529	119	122064	2.149	ppbv	99
66) Benzyl Chloride	11.610	91	11870	2.059	ppbv	99
67) sec-Butylbenzene	11.765	105	170529	2.151	ppbv	100
69) p-Isopropyltoluene	11.921	119	147482	2.195	ppbv	98
70) n-Butylbenzene	12.280	91	144148	2.212	ppbv	99
71) 2-Chlorotoluene	10.898	91	101951	2.085	ppbv	100
72) 4-Ethyltoluene	11.121	105	92654	2.140	ppbv	99
73) 1,3,5-Trimethylbenzene	11.202	105	76655	2.110	ppbv	96
74) 1,2,4-Trimethylbenzene	11.532	105	86286	2.176	ppbv	98
75) 1,3-Dichlorobenzene	11.604	146	48961	2.116	ppbv	98
76) 1,4-Dichlorobenzene	11.668	146	48881	2.112	ppbv	98
77) 1,2-Dichlorobenzene	11.944	146	48431	2.177	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	42456	2.287	ppbv	99
79) Naphthalene	13.510	128	85788	2.436	ppbv	97
80) Naphthalene,2-methyl-	14.458	142	20685	2.237	ppbv	98
81) 1,2,4-Trichlorobenzene	13.439	180	37521	2.382	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.787 min Scan# 832
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion: 49 Resp: 104644
Ion Ratio Lower Upper
49 100
128 46.4 23.0 69.0
130 59.1 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#2
Dichlorodifluoromethane
Concen: 2.092 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 85 Resp: 30881
Ion Ratio Lower Upper
85 100
87 33.1 25.4 38.2









#3
Chlorodifluoromethane
Concen: 2.072 ppbv
RT: 1.476 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion: 51 Resp: 30069
Ion Ratio Lower Upper
51 100
67 19.4 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#4
Chloromethane
Concen: 2.155 ppbv
RT: 1.534 min Scan# 447
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 50 Resp: 11890
Ion Ratio Lower Upper
50 100
52 32.0 24.1 36.1









#5
 Vinyl Chloride
 Concen: 1.995 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37

Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICC002

Tgt Ion: 62 Resp: 12040
 Ion Ratio Lower Upper
 62 100
 64 32.1 27.3 40.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#6
 Bromomethane
 Concen: 2.044 ppbv
 RT: 1.670 min Scan# 489
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37

Tgt Ion: 94 Resp: 5981
 Ion Ratio Lower Upper
 94 100
 96 92.4 72.9 109.3









#7
Chloroethane
Concen: 2.103 ppbv
RT: 1.706 min Scan# 500
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Tgt Ion: 64 Resp: 4943
Ion Ratio Lower Upper
64 100
66 30.8 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#8
Dichlorotetrafluoroethane
Concen: 2.151 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 85 Resp: 26409
Ion Ratio Lower Upper
85 100
135 64.3 52.6 79.0









#9
Propene
Concen: 1.930 ppbv
RT: 1.486 min Scan# 431
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Tgt Ion: 41 Resp: 12993
Ion Ratio Lower Upper
41 100
42 75.1 3.7 5.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#10
Heptane
Concen: 1.958 ppbv m
RT: 4.972 min Scan# 1509
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 43 Resp: 36171
Ion Ratio Lower Upper
43 100
57 0.0 39.7 59.5









#11
Trichlorofluoromethane
Concen: 2.177 ppbv
RT: 1.878 min Scan# 511
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion:101 Resp: 31124
Ion Ratio Lower Upper
101 100
103 63.6 52.6 79.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#12
1,1,2-Trichlorotrifluoroethane
Concen: 2.090 ppbv
RT: 2.140 min Scan# 634
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:101 Resp: 23336
Ion Ratio Lower Upper
101 100
151 73.8 59.2 88.8









#13
Ethanol
Concen: 2.430 ppbv
RT: 1.722 min Scan# 503
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Tgt Ion: 45 Resp: 5266
Ion Ratio Lower Upper
45 100
46 45.7 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#14
Bromoethene
Concen: 2.133 ppbv
RT: 1.784 min Scan# 524
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37


Tgt Ion:108 Resp: 9129
Ion Ratio Lower Upper
108 100
106 106.6 88.9 133.3
79 24.3 16.5 24.7





#15
Acetone
Concen: 2.186 ppbv
RT: 1.835 min Scan# 54
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002




Tgt Ion: 43 Resp: 31785
Ion Ratio Lower Upper
43 100
58 30.6 20.7 31.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#16
1,3-Butadiene
Concen: 2.021 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Tgt Ion: 39 Resp: 13101
Ion Ratio Lower Upper
39 100
54 83.3 59.4 89.2





#17
tert-Butyl alcohol
Concen: 2.062 ppbv
RT: 2.039 min Scan# 603
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002




Tgt Ion: 59 Resp: 39348
Ion Ratio Lower Upper
59 100
57 13.3 9.0 13.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#18
1,1-Dichloroethene
Concen: 2.074 ppbv
RT: 2.036 min Scan# 602
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37



Tgt Ion: 96 Resp: 10780
Ion Ratio Lower Upper
96 100
61 207.0 170.6 255.8
98 74.8 52.2 78.4









#19
Isopropyl Alcohol
Concen: 2.097 ppbv
RT: 1.887 min Scan# 554
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Tgt Ion: 45 Resp: 17229
Ion Ratio Lower Upper
45 100
58 56.6 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#20
Methylene Chloride
Concen: 2.248 ppbv
RT: 2.062 min Scan# 610
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37


Tgt Ion: 84 Resp: 12238
Ion Ratio Lower Upper
84 100
49 166.8 132.9 199.3
51 46.5 34.8 52.2
86 68.8 48.2 72.2





#21
 Allyl Chloride
 Concen: 2.002 ppbv
 RT: 2.098 min Scan# 619
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002




Tgt Ion: 41 Resp: 19580
 Ion Ratio Lower Upper
 41 100
 76 28.6 22.1 33.1
 39 81.8 61.0 91.4

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#22
 trans-1,2-Dichloroethene
 Concen: 2.110 ppbv
 RT: 2.340 min Scan# 696
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Tgt Ion: 96 Resp: 12537
 Ion Ratio Lower Upper
 96 100
 61 183.6 139.0 208.6
 98 64.3 49.4 74.2





#23
 Vinyl Acetate
 Concen: 2.006 ppbv
 RT: 2.460 min Scan# 731
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002



Tgt Ion: 43	Resp: 35205
Ion Ratio	Lower Upper
43 100	
86 6.8	6.2 9.2
42 9.5	8.0 12.0
44 6.5	4.0 6.0




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#24
 1,1-Dichloroethane
 Concen: 2.179 ppbv
 RT: 2.408 min Scan# 717
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37

Tgt Ion: 63	Resp: 24987
Ion Ratio	Lower Upper
63 100	
98 5.3	2.8 8.4
100 3.3	1.3 3.9









#25
Ethyl Acetate
Concen: 2.048 ppbv
RT: 2.832 min Scan# 847
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion: 43	Resp: 6434
Ion Ratio	Lower Upper
43	100
45	9.6 8.1 12.1
61	8.5 7.0 10.4
70	8.5 7.0 10.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#26
Hexane
Concen: 2.093 ppbv
RT: 2.826 min Scan# 846
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 57	Resp: 28776
Ion Ratio	Lower Upper
57	100
41	96.3 74.6 111.8
43	221.0 179.0 268.4
56	56.5 42.3 63.5









#27
Carbon Disulfide
Concen: 2.135 ppbv
RT: 2.149 min Scan# 637
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Tgt Ion: 76 Resp: 29964
Ion Ratio Lower Upper
76 100
44 26.0 16.1 24.1
78 13.5 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#28
Methyl tert-Butyl Ether
Concen: 2.090 ppbv
RT: 2.438 min Scan# 726
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37


Tgt Ion: 73 Resp: 14101
Ion Ratio Lower Upper
73 100
57 32.5 19.8 29.8





#29
Chloroform
Concen: 2.065 ppbv
RT: 2.849 min Scan# 851
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002




Tgt Ion: 83 Resp: 40355
Ion Ratio Lower Upper
83 100
85 65.0 50.5 75.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#30
Cyclohexane
Concen: 2.053 ppbv
RT: 3.855 min Scan# 1164
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Tgt Ion: 84 Resp: 23659
Ion Ratio Lower Upper
84 100
56 129.3 104.7 157.1
41 89.8 69.0 103.6





#31
 cis-1,2-Dichloroethene
 Concen: 2.024 ppbv
 RT: 2.722 min Scan# 81
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002




Tgt Ion: 61 Resp: 27663
 Ion Ratio Lower Upper
 61 100
 96 59.9 49.3 73.9
 98 39.6 32.1 48.1

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#32
 1,1,1-Trichloroethane
 Concen: 2.024 ppbv
 RT: 3.366 min Scan# 1013
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Tgt Ion: 97 Resp: 42547
 Ion Ratio Lower Upper
 97 100
 99 62.8 50.3 75.5
 61 50.4 38.9 58.3





#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.955 min Scan# 1193
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002



Tgt Ion:114 Resp: 284629
Ion Ratio Lower Upper
114 100
63 25.7 21.0 31.6
88 18.2 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#34
2-Butanone
Concen: 2.085 ppbv
RT: 2.557 min Scan# 763
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37


Tgt Ion: 43 Resp: 41884
Ion Ratio Lower Upper
43 100
72 22.9 18.1 27.1





#35
Carbon Tetrachloride
Concen: 1.942 ppbv
RT: 3.765 min Scan# 1133
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002




Tgt Ion: 117 Resp: 41949
Ion Ratio Lower Upper
117 100
119 97.8 74.6 112.0
121 29.6 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#36
Benzene
Concen: 2.060 ppbv m
RT: 3.654 min Scan# 1102
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37



Tgt Ion: 78 Resp: 56141
Ion Ratio Lower Upper
78 100
77 23.3 17.5 26.3
50 24.1 15.5 23.3#









#37
1,2-Dichloroethane
Concen: 2.019 ppbv
RT: 3.218 min Scan# 965
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion: 62 Resp: 32188
Ion Ratio Lower Upper
62 100
98 7.0 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#38
Trichloroethene
Concen: 1.975 ppbv m
RT: 4.548 min Scan# 1378
Delta R.T. 0.013 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:130 Resp: 21105
Ion Ratio Lower Upper
130 100
95 104.0 85.7 128.5
132 106.9 73.4 110.0
97 83.4 52.7 79.1#





#39

1,2-Dichloropropane

Concen: 1.996 ppbv

RT: 4.312 min Scan# 1302

Delta R.T. 0.010 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC002

Tgt Ion:	63	Resp:	1991
Ion	Ratio	Lower	Upper
63	100		
41	83.9	62.6	93.8
62	69.5	51.0	76.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#40

1,4-Dioxane

Concen: 1.969 ppbv m




RT: 4.580 min Scan# 1388


Delta R.T. 0.013 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37


Tgt Ion:	88	Resp:	8829
Ion	Ratio	Lower	Upper
88	100		
58	0.0	63.0	94.6#
43	0.0	32.5	48.7#
57	0.0	0.0	0.0





#41
Tetrahydrofuran
Concen: 2.065 ppbv
RT: 3.056 min Scan# 914
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002




Tgt Ion: 42 Resp: 22912
Ion Ratio Lower Upper
42 100
72 38.8 30.6 46.0
71 37.9 30.5 45.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#42
Bromodichloromethane
Concen: 2.027 ppbv
RT: 4.490 min Scan# 1360
Delta R.T. 0.013 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37



Tgt Ion: 83 Resp: 44584
Ion Ratio Lower Upper
83 100
85 66.6 51.7 77.5
127 7.9 6.3 9.5









#43
Methyl Methacrylate
Concen: 2.009 ppbv
RT: 4.875 min Scan# 1476
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion: 69 Resp: 2907
Ion Ratio Lower Upper
69 100
41 157.2 126.7 190.1
100 36.0 28.6 43.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#44
2,2,4-Trimethylpentane
Concen: 1.971 ppbv
RT: 4.635 min Scan# 1405
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 57 Resp: 92261
Ion Ratio Lower Upper
57 100
41 37.8 28.0 42.0
56 33.9 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 2.064 ppbv m
RT: 6.406 min Scan# 1952
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion: 75 Resp: 27833
Ion Ratio Lower Upper
75 100
77 27.8 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#46
cis-1,3-Dichloropropene
Concen: 2.020 ppbv
RT: 5.593 min Scan# 1701
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 75 Resp: 31986
Ion Ratio Lower Upper
75 100
39 67.2 59.1 88.7
77 24.2 24.2 36.2





#47

1,1,2-Trichloroethane

Concen: 2.025 ppbv m

RT: 6.577 min Scan# 2001

Delta R.T. 0.013 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC002

Tgt Ion	Ratio	Lower	Upper
97	100		
83	83.8	64.6	97.0
85	54.2	39.4	59.2
99	55.6	47.0	70.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#48

Dibromochloromethane

Concen: 2.074 ppbv m




RT: 7.386 min Scan# 2255


Delta R.T. 0.010 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion	Ratio	Lower	Upper
129	100		
127	0.0	39.3	117.9









#49
Bromoform
Concen: 2.050 ppbv
RT: 9.480 min Scan# 2900
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion: 173 Resp: 3305
Ion Ratio Lower Upper
173 100
175 46.6 38.6 58.0
171 52.4 40.6 61.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#50
4-Methyl-2-Pentanone
Concen: 2.066 ppbv
RT: 5.745 min Scan# 1748
Delta R.T. 0.013 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37


Tgt Ion: 43 Resp: 58461
Ion Ratio Lower Upper
43 100
58 36.7 28.4 42.6





#51
2-Hexanone
Concen: 2.113 ppbv
RT: 7.432 min Scan# 21
Delta R.T. 0.013 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002



Tgt Ion: 43 Resp: 4644
Ion Ratio Lower Upper
43 100
58 48.4 38.4 57.6
98 0.2 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#52
Tetrachloroethene
Concen: 1.958 ppbv
RT: 8.218 min Scan# 2512
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:164 Resp: 19856
Ion Ratio Lower Upper
164 100
166 117.2 99.4 149.2
129 96.5 81.8 122.8
131 97.0 77.4 116.2





#53

Toluene

Concen: 2.008 ppbv

RT: 6.927 min Scan# 2111

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC002

Tgt Ion: 91 Resp: 65238

Ion Ratio Lower Upper

91 100





92 61.5 47.4 71.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#54

1,2-Dibromoethane

Concen: 2.007 ppbv

RT: 7.648 min Scan# 2336

Delta R.T. 0.010 min

Lab File: VL043205.D




Acq: 18 Nov 2025 10:37


Tgt Ion:107 Resp: 31266

Ion Ratio Lower Upper

107 100


109 88.3 72.1 108.1





#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.878 min Scan# 2714
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002




Tgt Ion:117 Resp: 230646
Ion Ratio Lower Upper
117 100
82 67.2 56.1 84.1
119 32.0 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#56
1,1,1,2-Tetrachloroethane
Concen: 2.071 ppbv
RT: 8.920 min Scan# 2729
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37



Tgt Ion:131 Resp: 35099
Ion Ratio Lower Upper
131 100
133 97.4 77.4 116.2
119 60.1 49.1 73.7





#57

Chlorobenzene

Concen: 2.121 ppbv

RT: 8.917 min Scan# 2728

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_1





ClientSampleId :

VSTDICC002

Tgt Ion:	112	Resp:	50228
Ion Ratio	Lower	Upper	
112	100		
77	71.6	55.5	83.3
114	33.6	29.1	35.5

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#58

Ethyl Benzene

Concen: 2.087 ppbv




RT: 9.351 min Scan# 2862


Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion:	91	Resp:	92649
Ion Ratio	Lower	Upper	
91	100		
106	29.8	22.8	34.2





#59

m/p-Xylene

Concen: 4.151 ppbv m

RT: 9.529 min Scan# 29

Delta R.T. -0.013 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC002

Tgt Ion: 91 Resp: 148300

Ion Ratio Lower Upper

91 100





106 0.0 36.1 54.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#60

o-Xylene

Concen: 2.153 ppbv

RT: 9.959 min Scan# 3050

Delta R.T. 0.003 min

Lab File: VL043205.D




Acq: 18 Nov 2025 10:37


Tgt Ion: 91 Resp: 76540

Ion Ratio Lower Upper

91 100

106 41.9 20.9 62.8





#61

Styrene

Concen: 2.133 ppbv

RT: 9.866 min Scan# 3020

Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC002

Tgt Ion:104	Resp:	33188
Ion Ratio	Lower	Upper
104	100	
78	52.1	42.1 63.1
103	44.9	37.2 55.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#62

Isopropylbenzene

Concen: 2.136 ppbv




RT: 10.536 min Scan# 3228


Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion:105	Resp:	135853
Ion Ratio	Lower	Upper
105	100	
120	25.6	19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 2.102 ppbv
RT: 9.963 min Scan# 3048
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_1
ClientSampleId :
VSTDICC002

Tgt Ion: 83 Resp: 47439
Ion Ratio Lower Upper
83 100
131 10.0 4.8 14.3
85 63.4 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#64
n-propylbenzene
Concen: 2.136 ppbv
RT: 10.985 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:120 Resp: 34915
Ion Ratio Lower Upper
120 100
91 467.6 377.1 565.7









#65
tert-Butylbenzene
Concen: 2.149 ppbv
RT: 11.529 min Scan# 3533
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion:119 Resp: 122064
Ion Ratio Lower Upper
119 100
91 82.0 65.3 97.9
134 22.1 17.0 25.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#66
Benzyl Chloride
Concen: 2.059 ppbv
RT: 11.610 min Scan# 3560
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37


Tgt Ion: 91 Resp: 11870
Ion Ratio Lower Upper
91 100
126 18.3 14.4 21.6
128 5.9 4.6 6.8





#67
 sec-Butylbenzene
 Concen: 2.151 ppbv
 RT: 11.765 min Scan# 3607
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC002



Tgt Ion:105 Resp: 170529
 Ion Ratio Lower Upper
 105 100
 134 18.9 15.0 22.6




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.115 ppbv
 RT: 10.374 min Scan# 3178
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37


Tgt Ion: 95 Resp: 171009
 Ion Ratio Lower Upper
 95 100
 174 62.1 50.2 75.2
 175 4.9 4.1 6.1





#69
p-Isopropyltoluene
Concen: 2.195 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002




Tgt Ion: 119 Resp: 147482
Ion Ratio Lower Upper
119 100
134 24.2 20.2 30.2
91 29.1 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#70
n-Butylbenzene
Concen: 2.212 ppbv
RT: 12.280 min Scan# 3767
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37



Tgt Ion: 91 Resp: 144148
Ion Ratio Lower Upper
91 100
92 53.3 43.5 65.3
134 23.8 19.0 28.6









#71
2-Chlorotoluene
Concen: 2.085 ppbv
RT: 10.898 min Scan# 3339
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Tgt Ion: 91 Resp: 10195
Ion Ratio Lower Upper
91 100
126 29.0 11.7 46.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#72
4-Ethyltoluene
Concen: 2.140 ppbv
RT: 11.121 min Scan# 3409
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37


Tgt Ion:105 Resp: 92654
Ion Ratio Lower Upper
105 100
120 29.1 23.8 35.8
91 13.7 10.3 15.5
77 13.8 11.1 16.7





#73
1,3,5-Trimethylbenzene
Concen: 2.110 ppbv
RT: 11.202 min Scan# 3432
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002



Tgt Ion:105 Resp: 76658
Ion Ratio Lower Upper
105 100
120 45.5 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#74
1,2,4-Trimethylbenzene
Concen: 2.176 ppbv
RT: 11.532 min Scan# 3536
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:105 Resp: 86286
Ion Ratio Lower Upper
105 100
120 50.0 41.7 62.5
91 85.2 70.0 105.0
77 29.5 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 2.116 ppbv
RT: 11.604 min Scan# 3557
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion:146 Resp: 48961
Ion Ratio Lower Upper
146 100
111 45.7 23.4 70.3
148 64.6 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#76
1,4-Dichlorobenzene
Concen: 2.112 ppbv
RT: 11.668 min Scan# 3578
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:146 Resp: 48881
Ion Ratio Lower Upper
146 100
111 46.3 22.8 68.3
148 65.3 31.8 95.4





#77

1,2-Dichlorobenzene

Concen: 2.177 ppbv

RT: 11.944 min Scan# 3662

Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC002

Tgt Ion:	146	Resp:	4843
Ion Ratio	Lower	Upper	
146	100		
111	48.5	24.3	73.0
148	61.3	31.6	94.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#78

Hexachloro-1,3-Butadiene

Concen: 2.287 ppbv




RT: 13.879 min Scan# 4261


Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion:	225	Resp:	42456
Ion Ratio	Lower	Upper	
225	100		
223	62.1	50.2	75.4
227	63.4	51.1	76.7









#79
Naphthalene
Concen: 2.436 ppbv
RT: 13.510 min Scan# 4146
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002

Tgt Ion:128 Resp: 85788
Ion Ratio Lower Upper
128 100
127 13.3 9.9 14.9
129 11.7 8.4 12.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#80
Naphthalene,2-methyl-
Concen: 2.237 ppbv
RT: 14.458 min Scan# 4440
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37


Tgt Ion:142 Resp: 20685
Ion Ratio Lower Upper
142 100
141 82.9 67.1 100.7
115 36.9 31.9 47.9





#81
 1,2,4-Trichlorobenzene
 Concen: 2.382 ppbv
 RT: 13.439 min Scan# 41
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37



Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002



Tgt Ion:180 Resp: 3752
 Ion Ratio Lower Upper
 180 100
 182 95.8 77.4 116.2
 184 30.4 24.2 36.4

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043206.D
 Acq On : 18 Nov 2025 11:10
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Quant Time: Nov 19 00:51:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	102795	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	278732	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	227998	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	167061	9.996	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 100.000%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.499	85	15474	1.067	ppbv	100
3) Chlorodifluoromethane	1.476	51	14437	1.013	ppbv #	92
4) Chloromethane	1.531	50	5751	1.061	ppbv	92
5) Vinyl Chloride	1.580	62	5967	1.007	ppbv	95
6) Bromomethane	1.667	94	3104	1.080	ppbv	94
7) Chloroethane	1.706	64	2550	1.104	ppbv	97
8) Dichlorotetrafluoroethane	1.554	85	12759	1.058	ppbv	98
9) Propene	1.483	41	6147	0.929	ppbv #	1
10) Heptane	4.972	43	18544	1.022	ppbv	100
11) Trichlorofluoromethane	1.877	101	15571	1.108	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.140	101	11743	1.071	ppbv	99
13) Ethanol	1.719	45	2373	1.115	ppbv	67
14) Bromoethene	1.780	108	4345	1.033	ppbv #	91
15) Acetone	1.835	43	16353	1.145	ppbv	90
16) 1,3-Butadiene	1.609	39	6469	1.016	ppbv	96
17) tert-Butyl alcohol	2.039	59	20703	1.104	ppbv #	92
18) 1,1-Dichloroethene	2.033	96	5447	1.067	ppbv	97
19) Isopropyl Alcohol	1.884	45	8849	1.097	ppbv #	43
20) Methylene Chloride	2.059	84	7297	1.030	ppbv	85
21) Allyl Chloride	2.094	41	9198	0.957	ppbv #	84
22) trans-1,2-Dichloroethene	2.340	96	6313	1.082	ppbv	94
23) Vinyl Acetate	2.460	43	16779	0.973	ppbv	98
24) 1,1-Dichloroethane	2.405	63	12062	1.071	ppbv	94
25) Ethyl Acetate	2.832	43	32108	1.041	ppbv	98
26) Hexane	2.819	57	14639	1.084	ppbv	99
27) Carbon Disulfide	2.149	76	14637	1.062	ppbv #	30
28) Methyl tert-Butyl Ether	2.441	73	7236	1.092	ppbv #	83
29) Chloroform	2.845	83	19509	1.016	ppbv	91
30) Cyclohexane	3.852	84	12269	1.084	ppbv	96
31) cis-1,2-Dichloroethene	2.716	61	13889	1.035	ppbv	97
32) 1,1,1-Trichloroethane	3.363	97	20745	1.005	ppbv	99
34) 2-Butanone	2.554	43	20940	1.064	ppbv	96
35) Carbon Tetrachloride	3.758	117	21497	1.016	ppbv	96
36) Benzene	3.648	78	27911	1.046	ppbv	99
37) 1,2-Dichloroethane	3.214	62	16159	1.035	ppbv	99
38) Trichloroethene	4.538	130	10809m	1.033	ppbv	
39) 1,2-Dichloropropane	4.311	63	10147	1.039	ppbv	91
40) 1,4-Dioxane	4.577	88	4616m	1.051	ppbv	
41) Tetrahydrofuran	3.052	42	11176	1.029	ppbv	96
42) Bromodichloromethane	4.480	83	22150	1.028	ppbv #	98
43) Methyl Methacrylate	4.871	69	14821	1.046	ppbv	94
44) 2,2,4-Trimethylpentane	4.632	57	49203m	1.073	ppbv	
45) t-1,3-Dichloropropene	6.402	75	13149m	0.996	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043206.D
 Acq On : 18 Nov 2025 11:10
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:51:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M


Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.590	75	15527	1.001	ppbv #	95
47) 1,1,2-Trichloroethane	6.577	97	10369	0.995	ppbv #	85
48) Dibromochloromethane	7.377	129	18560	1.041	ppbv	93
49) Bromoform	9.477	173	15858	1.004	ppbv	98
50) 4-Methyl-2-Pentanone	5.745	43	29560	1.067	ppbv	99
51) 2-Hexanone	7.428	43	22438	1.043	ppbv #	98
52) Tetrachloroethene	8.215	164	10080	1.015	ppbv	90
53) Toluene	6.923	91	32951	1.036	ppbv	100
54) 1,2-Dibromoethane	7.645	107	15865	1.040	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.917	131	17535	1.046	ppbv	99
57) Chlorobenzene	8.914	112	24806	1.060	ppbv	98
58) Ethyl Benzene	9.344	91	46455	1.059	ppbv	96
59) m/p-Xylene	9.545	91	73461m	2.080	ppbv	
60) o-Xylene	9.956	91	37054	1.054	ppbv	97
61) Styrene	9.866	104	15730	1.023	ppbv	95
62) Isopropylbenzene	10.532	105	67163	1.068	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.956	83	23083	1.035	ppbv	97
64) n-propylbenzene	10.985	120	17289	1.070	ppbv	100
65) tert-Butylbenzene	11.526	119	61258	1.091	ppbv	99
66) Benzyl Chloride	11.610	91	5729	1.005	ppbv	96
67) sec-Butylbenzene	11.762	105	84989	1.084	ppbv	99
69) p-Isopropyltoluene	11.921	119	69993	1.054	ppbv	99
70) n-Butylbenzene	12.277	91	69323	1.076	ppbv	99
71) 2-Chlorotoluene	10.895	91	51452	1.065	ppbv	100
72) 4-Ethyltoluene	11.118	105	44403	1.037	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	38350	1.068	ppbv	91
74) 1,2,4-Trimethylbenzene	11.532	105	42051	1.073	ppbv #	81
75) 1,3-Dichlorobenzene	11.600	146	24011	1.050	ppbv	98
76) 1,4-Dichlorobenzene	11.665	146	23543	1.029	ppbv	97
77) 1,2-Dichlorobenzene	11.940	146	24248	1.103	ppbv	98
78) Hexachloro-1,3-Butadiene	13.879	225	20489	1.116	ppbv	99
79) Naphthalene	13.507	128	36462	1.047	ppbv	95
80) Naphthalene,2-methyl-	14.455	142	7375	0.807	ppbv	95
81) 1,2,4-Trichlorobenzene	13.436	180	15785	1.014	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 832
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001




Tgt Ion: 49 Resp: 102795
Ion Ratio Lower Upper
49 100
128 46.9 23.0 69.0
130 60.1 29.9 89.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#2
Dichlorodifluoromethane
Concen: 1.067 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10



Tgt Ion: 85 Resp: 15474
Ion Ratio Lower Upper
85 100
87 31.9 25.4 38.2









#3
Chlorodifluoromethane
Concen: 1.013 ppbv
RT: 1.476 min Scan# 428
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 51 Resp: 1443
Ion Ratio Lower Upper
51 100
67 22.3 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#4
Chloromethane
Concen: 1.061 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 50 Resp: 5751
Ion Ratio Lower Upper
50 100
52 34.2 24.1 36.1









#5
 Vinyl Chloride
 Concen: 1.007 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Tgt Ion: 62 Resp: 596
 Ion Ratio Lower Upper
 62 100
 64 31.3 27.3 40.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#6
 Bromomethane
 Concen: 1.080 ppbv
 RT: 1.667 min Scan# 488
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10


Tgt Ion: 94 Resp: 3104
 Ion Ratio Lower Upper
 94 100
 96 85.7 72.9 109.3





#7
Chloroethane
Concen: 1.104 ppbv
RT: 1.706 min Scan# 500
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001




Tgt Ion: 64 Resp: 2550
Ion Ratio Lower Upper
64 100
66 31.5 26.8 40.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#8
Dichlorotetrafluoroethane
Concen: 1.058 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Tgt Ion: 85 Resp: 12759
Ion Ratio Lower Upper
85 100
135 67.7 52.6 79.0





#9
Propene
Concen: 0.929 ppbv
RT: 1.483 min Scan# 431
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001




Tgt Ion: 41 Resp: 6147
Ion Ratio Lower Upper
41 100
42 80.0 3.7 5.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#10
Heptane
Concen: 1.022 ppbv
RT: 4.972 min Scan# 1509
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Tgt Ion: 43 Resp: 18544
Ion Ratio Lower Upper
43 100
57 49.3 39.7 59.5





#11
Trichlorofluoromethane
Concen: 1.108 ppbv
RT: 1.877 min Scan# 552
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001




Tgt Ion:101 Resp: 1557
Ion Ratio Lower Upper
101 100
103 65.2 52.6 79.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#12
1,1,2-Trichlorotrifluoroethane
Concen: 1.071 ppbv
RT: 2.140 min Scan# 634
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10



Tgt Ion:101 Resp: 11743
Ion Ratio Lower Upper
101 100
151 74.6 59.2 88.8









#13
Ethanol
Concen: 1.115 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 45 Resp: 2373
Ion Ratio Lower Upper
45 100
46 57.0 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#14
Bromoethene
Concen: 1.033 ppbv
RT: 1.780 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10


Tgt Ion:108 Resp: 4345
Ion Ratio Lower Upper
108 100
106 114.0 88.9 133.3
79 39.2 16.5 24.7#





#15
Acetone
Concen: 1.145 ppbv
RT: 1.835 min Scan# 54
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001




Tgt Ion: 43 Resp: 1635
Ion Ratio Lower Upper
43 100
58 30.9 20.7 31.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#16
1,3-Butadiene
Concen: 1.016 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10



Tgt Ion: 39 Resp: 6469
Ion Ratio Lower Upper
39 100
54 77.8 59.4 89.2









#17
tert-Butyl alcohol
Concen: 1.104 ppbv
RT: 2.039 min Scan# 601
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 59 Resp: 20703
Ion Ratio Lower Upper
59 100
57 14.1 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#18
1,1-Dichloroethene
Concen: 1.067 ppbv
RT: 2.033 min Scan# 601
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 96 Resp: 5447
Ion Ratio Lower Upper
96 100
61 219.3 170.6 255.8
98 63.7 52.2 78.4









#19
Isopropyl Alcohol
Concen: 1.097 ppbv
RT: 1.884 min Scan# 554
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Tgt Ion: 45 Resp: 8849
Ion Ratio Lower Upper
45 100
58 4.0 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#20
Methylene Chloride
Concen: 1.030 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10


Tgt Ion: 84 Resp: 7297
Ion Ratio Lower Upper
84 100
49 142.0 132.9 199.3
51 47.3 34.8 52.2
86 71.4 48.2 72.2





#21
 Allyl Chloride
 Concen: 0.957 ppbv
 RT: 2.094 min Scan# 619
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001




Tgt Ion:	41	Resp:	9198
Ion Ratio	Lower	Upper	
41	100		
76	33.5	22.1	33.1
39	91.7	61.0	91.4

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#22
 trans-1,2-Dichloroethene
 Concen: 1.082 ppbv
 RT: 2.340 min Scan# 696
 Delta R.T. 0.006 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Tgt Ion:	96	Resp:	6313
Ion Ratio	Lower	Upper	
96	100		
61	165.0	139.0	208.6
98	65.1	49.4	74.2





#23
 Vinyl Acetate
 Concen: 0.973 ppbv
 RT: 2.460 min Scan# 733
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001




Tgt Ion: 43	Resp: 16779
Ion Ratio	Lower Upper
43 100	
86 8.8	6.2 9.2
42 10.4	8.0 12.0
44 5.9	4.0 6.0

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#24
 1,1-Dichloroethane
 Concen: 1.071 ppbv
 RT: 2.405 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10



Tgt Ion: 63	Resp: 12062
Ion Ratio	Lower Upper
63 100	
98 2.9	2.8 8.4
100 2.5	1.3 3.9









#25
Ethyl Acetate
Concen: 1.041 ppbv
RT: 2.832 min Scan# 847
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 43	Resp: 32108
Ion Ratio	Lower Upper
43 100	
45 11.1	8.1 12.1
61 8.7	7.0 10.4
70 9.9	7.0 10.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#26
Hexane
Concen: 1.084 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 57	Resp: 14639
Ion Ratio	Lower Upper
57 100	
41 96.8	74.6 111.8
43 222.8	179.0 268.4
56 52.3	42.3 63.5









#27
Carbon Disulfide
Concen: 1.062 ppbv
RT: 2.149 min Scan# 636
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001

Tgt Ion: 76 Resp: 1463
Ion Ratio Lower Upper
76 100
44 69.6 16.1 24.1
78 14.8 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#28
Methyl tert-Butyl Ether
Concen: 1.092 ppbv
RT: 2.441 min Scan# 727
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 73 Resp: 7236
Ion Ratio Lower Upper
73 100
57 33.1 19.8 29.8









#29
Chloroform
Concen: 1.016 ppbv
RT: 2.845 min Scan# 851
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 83 Resp: 19509
Ion Ratio Lower Upper
83 100
85 56.3 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#30
Cyclohexane
Concen: 1.084 ppbv
RT: 3.852 min Scan# 1163
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 84 Resp: 12269
Ion Ratio Lower Upper
84 100
56 127.5 104.7 157.1
41 90.4 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 1.035 ppbv
 RT: 2.716 min Scan# 812
 Delta R.T. -0.000 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Tgt Ion: 61	Resp: 13889
Ion Ratio	Lower Upper
61 100	
96 60.4	49.3 73.9
98 43.0	32.1 48.1




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#32
 1,1,1-Trichloroethane
 Concen: 1.005 ppbv
 RT: 3.363 min Scan# 1012
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Tgt Ion: 97	Resp: 20745
Ion Ratio	Lower Upper
97 100	
99 63.6	50.3 75.5
61 48.3	38.9 58.3





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.952 min Scan# 1194

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC001

Tgt Ion:	114	Resp:	278732
Ion Ratio	Lower	Upper	
114	100		
63	26.0	21.0	31.6
88	18.6	14.9	22.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#34

2-Butanone

Concen: 1.064 ppbv




RT: 2.554 min Scan# 762


Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Tgt Ion:	43	Resp:	20940
Ion Ratio	Lower	Upper	
43	100		
72	24.7	18.1	27.1





#35

Carbon Tetrachloride

Concen: 1.016 ppbv

RT: 3.758 min Scan# 1134

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC001

Tgt Ion: 117 Resp: 2149

Ion	Ratio	Lower	Upper
117	100		
119	89.6	74.6	112.0
121	30.9	23.2	34.8

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#36

Benzene

Concen: 1.046 ppbv

RT: 3.648 min Scan# 1100




Delta R.T. -0.000 min


Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Tgt Ion: 78 Resp: 27911

Ion	Ratio	Lower	Upper
78	100		
77	21.5	17.5	26.3
50	18.6	15.5	23.3









#37
1,2-Dichloroethane
Concen: 1.035 ppbv
RT: 3.214 min Scan# 965
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Tgt Ion: 62 Resp: 16159
Ion Ratio Lower Upper
62 100
98 6.7 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#38
Trichloroethene
Concen: 1.033 ppbv m
RT: 4.538 min Scan# 1375
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10


Tgt Ion:130 Resp: 10809
Ion Ratio Lower Upper
130 100
95 110.4 85.7 128.5
132 73.3 73.4 110.0#
97 68.9 52.7 79.1





#39
1,2-Dichloropropane
Concen: 1.039 ppbv
RT: 4.311 min Scan# 1305
Delta R.T. 0.009 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001




Tgt Ion: 63 Resp: 1014
Ion Ratio Lower Upper
63 100
41 70.5 62.6 93.8
62 56.2 51.0 76.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#40
1,4-Dioxane
Concen: 1.051 ppbv m
RT: 4.577 min Scan# 1387
Delta R.T. 0.009 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10



Tgt Ion: 88 Resp: 4616
Ion Ratio Lower Upper
88 100
58 69.6 63.0 94.6
43 0.0 32.5 48.7#
57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 1.029 ppbv
RT: 3.052 min Scan# 914
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 42 Resp: 11176
Ion Ratio Lower Upper
42 100
72 40.2 30.6 46.0
71 40.4 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#42
Bromodichloromethane
Concen: 1.028 ppbv
RT: 4.480 min Scan# 1357
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 83 Resp: 22150
Ion Ratio Lower Upper
83 100
85 63.8 51.7 77.5
127 5.6 6.3 9.5#









#43
Methyl Methacrylate
Concen: 1.046 ppbv
RT: 4.871 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Tgt Ion: 69 Resp: 1482
Ion Ratio Lower Upper
69 100
41 148.1 126.7 190.1
100 35.7 28.6 43.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#44
2,2,4-Trimethylpentane
Concen: 1.073 ppbv m
RT: 4.632 min Scan# 1404
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 57 Resp: 49203
Ion Ratio Lower Upper
57 100
41 37.7 28.0 42.0
56 32.0 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 0.996 ppbv m
RT: 6.402 min Scan# 1951
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 75 Resp: 13149
Ion Ratio Lower Upper
75 100
77 28.5 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#46
cis-1,3-Dichloropropene
Concen: 1.001 ppbv
RT: 5.590 min Scan# 1700
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10


Tgt Ion: 75 Resp: 15527
Ion Ratio Lower Upper
75 100
39 72.5 59.1 88.7
77 37.9 24.2 36.2#





#47
1,1,2-Trichloroethane
Concen: 0.995 ppbv
RT: 6.577 min Scan# 2005
Delta R.T. 0.013 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001




Tgt Ion: 97 Resp: 10369
Ion Ratio Lower Upper
97 100
83 97.9 64.6 97.0
85 56.4 39.4 59.2
99 67.4 47.0 70.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#48
Dibromochloromethane
Concen: 1.041 ppbv
RT: 7.377 min Scan# 2252
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10



Tgt Ion:129 Resp: 18560
Ion Ratio Lower Upper
129 100
127 72.8 39.3 117.9





#49

Bromoform

Concen: 1.004 ppbv

RT: 9.477 min Scan# 2900

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC001

Tgt Ion:	173	Resp:	1585
Ion Ratio	Lower	Upper	
173	100		
175	47.9	38.6	58.0
171	53.5	40.6	61.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#50

4-Methyl-2-Pentanone

Concen: 1.067 ppbv




RT: 5.745 min Scan# 1748


Delta R.T. 0.013 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Tgt Ion:	43	Resp:	29560
Ion Ratio	Lower	Upper	
43	100		
58	36.1	28.4	42.6









#51
2-Hexanone
Concen: 1.043 ppbv
RT: 7.428 min Scan# 22438
Delta R.T. 0.009 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 43 Resp: 22438
Ion Ratio Lower Upper
43 100
58 49.4 38.4 57.6
98 0.2 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#52
Tetrachloroethene
Concen: 1.015 ppbv
RT: 8.215 min Scan# 2511
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10


Tgt Ion:164 Resp: 10080
Ion Ratio Lower Upper
164 100
166 116.5 99.4 149.2
129 96.3 81.8 122.8
131 113.5 77.4 116.2





#53
Toluene
Concen: 1.036 ppbv
RT: 6.923 min Scan# 2111
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001




Tgt Ion: 91 Resp: 3295
Ion Ratio Lower Upper
91 100
92 59.0 47.4 71.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#54
1,2-Dibromoethane
Concen: 1.040 ppbv
RT: 7.645 min Scan# 2335
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Tgt Ion:107 Resp: 15865
Ion Ratio Lower Upper
107 100
109 93.3 72.1 108.1





#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 2714
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001




Tgt Ion:117 Resp: 227998
Ion Ratio Lower Upper
117 100
82 66.2 56.1 84.1
119 32.1 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#56
1,1,1,2-Tetrachloroethane
Concen: 1.046 ppbv
RT: 8.917 min Scan# 2728
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Tgt Ion:131 Resp: 17535
Ion Ratio Lower Upper
131 100
133 97.3 77.4 116.2
119 60.6 49.1 73.7





#57
Chlorobenzene
Concen: 1.060 ppbv
RT: 8.914 min Scan# 2726
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001



Tgt Ion: 112 Resp: 24806
Ion Ratio Lower Upper
112 100
77 68.6 55.5 83.3
114 29.8 29.1 35.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#58
Ethyl Benzene
Concen: 1.059 ppbv
RT: 9.344 min Scan# 2860
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 91 Resp: 46455
Ion Ratio Lower Upper
91 100
106 26.6 22.8 34.2









#59
m/p-Xylene
Concen: 2.080 ppbv m
RT: 9.545 min Scan# 2921
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion: 91 Resp: 7346
Ion Ratio Lower Upper
91 100
106 0.0 36.1 54.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#60
o-Xylene
Concen: 1.054 ppbv
RT: 9.956 min Scan# 3049
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 91 Resp: 37054
Ion Ratio Lower Upper
91 100
106 44.1 20.9 62.8





#61

Styrene

Concen: 1.023 ppbv

RT: 9.866 min Scan# 3020

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC001

Tgt Ion:104	Resp:	15730
Ion Ratio	Lower	Upper
104	100	
78	56.9	42.1 63.1
103	48.6	37.2 55.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#62

Isopropylbenzene

Concen: 1.068 ppbv




RT: 10.532 min Scan# 3227


Delta R.T. -0.000 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Tgt Ion:105	Resp:	67163
Ion Ratio	Lower	Upper
105	100	
120	25.2	19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 1.035 ppbv
RT: 9.956 min Scan# 3048
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001

Tgt Ion: 83 Resp: 23083
Ion Ratio Lower Upper
83 100
131 10.0 4.8 14.3
85 66.2 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#64
n-propylbenzene
Concen: 1.070 ppbv
RT: 10.985 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:120 Resp: 17289
Ion Ratio Lower Upper
120 100
91 471.4 377.1 565.7









#65
tert-Butylbenzene
Concen: 1.091 ppbv
RT: 11.526 min Scan# 3533
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Tgt Ion:119 Resp: 61258
Ion Ratio Lower Upper
119 100
91 81.7 65.3 97.9
134 20.0 17.0 25.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#66
Benzyl Chloride
Concen: 1.005 ppbv
RT: 11.610 min Scan# 3560
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10


Tgt Ion: 91 Resp: 5729
Ion Ratio Lower Upper
91 100
126 16.3 14.4 21.6
128 6.5 4.6 6.8





#67
 sec-Butylbenzene
 Concen: 1.084 ppbv
 RT: 11.762 min Scan# 3607
 Delta R.T. -0.000 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC001




Tgt Ion: 105 Resp: 84989
 Ion Ratio Lower Upper
 105 100
 134 19.3 15.0 22.6

Manual Integrations
 APPROVED



Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025




#68
 1-Bromo-4-Fluorobenzene
 Concen: 9.996 ppbv
 RT: 10.370 min Scan# 3177
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10



Tgt Ion: 95 Resp: 167061
 Ion Ratio Lower Upper
 95 100
 174 62.1 50.2 75.2
 175 4.6 4.1 6.1






#69
p-Isopropyltoluene
Concen: 1.054 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :




VSTDIC001




Tgt Ion: 119 Resp: 69993
Ion Ratio Lower Upper
119 100
134 25.0 20.2 30.2
91 28.9 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#70
n-Butylbenzene
Concen: 1.076 ppbv
RT: 12.277 min Scan# 3766
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10



Tgt Ion: 91 Resp: 69323
Ion Ratio Lower Upper
91 100
92 55.0 43.5 65.3
134 23.8 19.0 28.6









#71
2-Chlorotoluene
Concen: 1.065 ppbv
RT: 10.895 min Scan# 3339
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument : MSVOA_L
ClientSampleId : VSTDIC001

Tgt Ion: 91 Resp: 51452
Ion Ratio Lower Upper
91 100
126 29.3 11.7 46.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#72
4-Ethyltoluene
Concen: 1.037 ppbv
RT: 11.118 min Scan# 3408
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:105 Resp: 44403
Ion Ratio Lower Upper
105 100
120 29.8 23.8 35.8
91 13.2 10.3 15.5
77 14.6 11.1 16.7









#73
1,3,5-Trimethylbenzene
Concen: 1.068 ppbv
RT: 11.199 min Scan# 3432
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Tgt Ion:105 Resp: 38350
Ion Ratio Lower Upper
105 100
120 42.3 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#74
1,2,4-Trimethylbenzene
Concen: 1.073 ppbv
RT: 11.532 min Scan# 3536
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:105 Resp: 42051
Ion Ratio Lower Upper
105 100
120 46.8 41.7 62.5
91 61.4 70.0 105.0#
77 21.7 23.8 35.8#









#75
1,3-Dichlorobenzene
Concen: 1.050 ppbv
RT: 11.600 min Scan# 3557
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Tgt Ion:146 Resp: 2401
Ion Ratio Lower Upper
146 100
111 45.9 23.4 70.3
148 64.7 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#76
1,4-Dichlorobenzene
Concen: 1.029 ppbv
RT: 11.665 min Scan# 3577
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:146 Resp: 23543
Ion Ratio Lower Upper
146 100
111 44.5 22.8 68.3
148 66.4 31.8 95.4





#77

1,2-Dichlorobenzene

Concen: 1.103 ppbv

RT: 11.940 min Scan# 3662

Delta R.T. -0.000 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC001

Tgt Ion:	146	Resp:	24248
Ion Ratio	Lower	Upper	
146	100		
111	46.3	24.3	73.0
148	62.3	31.6	94.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#78

Hexachloro-1,3-Butadiene

Concen: 1.116 ppbv




RT: 13.879 min Scan# 4261


Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Tgt Ion:	225	Resp:	20489
Ion Ratio	Lower	Upper	
225	100		
223	63.3	50.2	75.4
227	62.5	51.1	76.7









#79
Naphthalene
Concen: 1.047 ppbv
RT: 13.507 min Scan# 4146
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Tgt Ion:128 Resp: 36462
Ion Ratio Lower Upper
128 100
127 14.2 9.9 14.9
129 12.2 8.4 12.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#80
Naphthalene,2-methyl-
Concen: 0.807 ppbv
RT: 14.455 min Scan# 4439
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10


Tgt Ion:142 Resp: 7375
Ion Ratio Lower Upper
142 100
141 77.8 67.1 100.7
115 40.8 31.9 47.9





#81
1,2,4-Trichlorobenzene
Concen: 1.014 ppbv
RT: 13.436 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10



Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001



Tgt Ion:180 Resp: 15785
Ion Ratio Lower Upper
180 100
182 98.2 77.4 116.2
184 33.4 24.2 36.4

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043208.D
Acq On : 18 Nov 2025 12:16
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Quant Time: Nov 19 00:51:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.784	49	101366	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	273996	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	222034	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	160968	9.890	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.900%
Target Compounds						
5) Vinyl Chloride	1.580	62	664	0.114	ppbv	Qvalue 97
32) 1,1,1-Trichloroethane	3.360	97	2074m	0.102	ppbv	
35) Carbon Tetrachloride	3.761	117	2097m	0.101	ppbv	
38) Trichloroethene	4.541	130	1302m	0.127	ppbv	
52) Tetrachloroethene	8.218	164	1161m	0.119	ppbv	
54) 1,2-Dibromoethane	7.648	107	1446m	0.096	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.953	83	2296m	0.106	ppbv	
79) Naphthalene	13.510	128	2732m	0.081	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043208.D
Acq On : 18 Nov 2025 12:16
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Quant Time: Nov 19 00:51:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By : Semsettin Yesilyurt 11/19/2025
Supervised By : Mahesh Dadoda 11/19/2025





#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 832
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1



Tgt Ion: 49 Resp: 101366
Ion Ratio Lower Upper
49 100
128 46.8 23.0 69.0
130 58.3 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#5
Vinyl Chloride
Concen: 0.114 ppbv
RT: 1.580 min Scan# 461
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16


Tgt Ion: 62 Resp: 664
Ion Ratio Lower Upper
62 100
64 32.2 27.3 40.9





#32
1,1,1-Trichloroethane
Concen: 0.102 ppbv m
RT: 3.360 min Scan# 1011
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.1




Tgt Ion: 97 Resp: 2074
Ion Ratio Lower Upper
97 100
99 62.7 50.3 75.5
61 53.5 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Tgt Ion:114 Resp: 273996
Ion Ratio Lower Upper
114 100
63 25.3 21.0 31.6
88 18.3 14.9 22.3





#35
Carbon Tetrachloride
Concen: 0.101 ppbv m
RT: 3.761 min Scan# 1133
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1




Tgt Ion:117 Resp: 2097
Ion Ratio Lower Upper
117 100
119 88.9 74.6 112.0
121 16.7 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#38
Trichloroethene
Concen: 0.127 ppbv m
RT: 4.541 min Scan# 1376
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Tgt Ion:130 Resp: 1302
Ion Ratio Lower Upper
130 100
95 75.4 85.7 128.5#
132 32.6 73.4 110.0#
97 27.9 52.7 79.1#





#52
Tetrachloroethene
Concen: 0.119 ppbv m
RT: 8.218 min Scan# 2510
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1




Tgt Ion:164 Resp: 1163
Ion Ratio Lower Upper
164 100
166 209.0 99.4 149.2
129 166.4 81.8 122.8
131 194.1 77.4 116.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#54
1,2-Dibromoethane
Concen: 0.096 ppbv m
RT: 7.648 min Scan# 2336
Delta R.T. 0.009 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16



Tgt Ion:107 Resp: 1446
Ion Ratio Lower Upper
107 100
109 85.5 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.875 min Scan# 2117
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Tgt Ion:117 Resp: 222034
Ion Ratio Lower Upper
117 100
82 66.4 56.1 84.1
119 32.2 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#63
1,1,2,2-Tetrachloroethane
Concen: 0.106 ppbv m
RT: 9.953 min Scan# 3048
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16


Tgt Ion: 83 Resp: 2296
Ion Ratio Lower Upper
83 100
131 10.5 4.8 14.3
85 59.6 31.9 95.9





#68
1-Bromo-4-Fluorobenzene
Concen: 9.890 ppbv
RT: 10.370 min Scan# 3176
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1




Tgt Ion: 95 Resp: 160968
Ion Ratio Lower Upper
95 100
174 62.5 50.2 75.2
175 4.8 4.1 6.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#79
Naphthalene
Concen: 0.081 ppbv m
RT: 13.510 min Scan# 4147
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16



Tgt Ion:128 Resp: 2732
Ion Ratio Lower Upper
128 100
127 9.7 9.9 14.9#
129 18.1 8.4 12.6#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043209.D
 Acq On : 18 Nov 2025 12:49
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:52:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	98952	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	266012	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	213547	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.371	95	152015	9.712	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.100%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	215m	0.038	ppbv	
32) 1,1,1-Trichloroethane	3.357	97	628m	0.032	ppbv	
35) Carbon Tetrachloride	3.758	117	664m	0.033	ppbv	
38) Trichloroethene	4.542	130	234m	0.023	ppbv	
52) Tetrachloroethene	8.222	164	290m	0.031	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.963	83	640m	0.031	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043209.D
Acq On : 18 Nov 2025 12:49
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Nov 19 00:52:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
Supervised By : Mahesh Dadoda 11/19/2025









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Tgt Ion: 49 Resp: 9895
Ion Ratio Lower Upper
49 100
128 45.9 23.0 69.0
130 58.8 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#5
Vinyl Chloride
Concen: 0.038 ppbv m
RT: 1.580 min Scan# 461
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 62 Resp: 215
Ion Ratio Lower Upper
62 100
64 32.1 27.3 40.9









#32
1,1,1-Trichloroethane
Concen: 0.032 ppbv m
RT: 3.357 min Scan# 1011
Delta R.T. -0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Tgt Ion: 97 Resp: 628
Ion Ratio Lower Upper
97 100
99 56.8 50.3 75.5
61 0.0 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49


Tgt Ion:114 Resp: 266012
Ion Ratio Lower Upper
114 100
63 25.7 21.0 31.6
88 18.2 14.9 22.3





#35
Carbon Tetrachloride
Concen: 0.033 ppbv m
RT: 3.758 min Scan# 1133
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03



Tgt Ion:117 Resp: 664
Ion Ratio Lower Upper
117 100
119 46.9 74.6 112.0#
121 33.5 23.2 34.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#38
Trichloroethene
Concen: 0.023 ppbv m
RT: 4.542 min Scan# 1376
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49


Tgt Ion:130 Resp: 234
Ion Ratio Lower Upper
130 100
95 87.4 85.7 128.5
132 225.8 73.4 110.0#
97 110.1 52.7 79.1#





#52
Tetrachloroethene
Concen: 0.031 ppbv m
RT: 8.222 min Scan# 21
Delta R.T. 0.010 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03






Tgt Ion:164 Resp: 290

Ion	Ratio	Lower	Upper
164	100		
166	102.3	99.4	149.2
129	104.6	81.8	122.8
131	87.5	77.4	116.2

Manual Integrations
APPROVED




Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.875 min Scan# 2715
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion:117 Resp: 213547

Ion	Ratio	Lower	Upper
117	100		
82	66.2	56.1	84.1
119	32.3	25.0	37.6









#63
1,1,2,2-Tetrachloroethane
Concen: 0.031 ppbv m
RT: 9.963 min Scan# 3048
Delta R.T. 0.010 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03

Tgt Ion:	83	Resp:	640
Ion	Ratio	Lower	Upper
83	100		
131	6.1	4.8	14.3
85	52.8	31.9	95.9




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#68
1-Bromo-4-Fluorobenzene
Concen: 9.712 ppbv
RT: 10.371 min Scan# 3177
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion:	95	Resp:	152015
Ion	Ratio	Lower	Upper
95	100		
174	62.0	50.2	75.2
175	4.6	4.1	6.1



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043210.D
 Acq On : 18 Nov 2025 13:24
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Nov 19 00:53:26 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	98968	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	258461	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	211247	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	159814	10.321	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.200%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.499	85	197987	14.184	ppbv	99
3) Chlorodifluoromethane	1.476	51	182813	13.317	ppbv	98
4) Chloromethane	1.531	50	72636	13.918	ppbv	97
5) Vinyl Chloride	1.580	62	71671	12.557	ppbv	96
6) Bromomethane	1.667	94	38265	13.830	ppbv	98
7) Chloroethane	1.703	64	29584	13.309	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	165709	14.271	ppbv	97
9) Propene	1.483	41	81393m	12.782	ppbv	
10) Heptane	4.972	43	225935	12.934	ppbv	100
11) Trichlorofluoromethane	1.874	101	200703	14.840	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.137	101	152772	14.469	ppbv	98
13) Ethanol	1.719	45	25468	12.425	ppbv	99
14) Bromoethene	1.781	108	55813	13.789	ppbv	96
15) Acetone	1.832	43	161036	11.712	ppbv	99
16) 1,3-Butadiene	1.609	39	81077	13.226	ppbv	99
17) tert-Butyl alcohol	2.036	59	257617	14.274	ppbv	99
18) 1,1-Dichloroethene	2.033	96	65731	13.370	ppbv	96
19) Isopropyl Alcohol	1.881	45	105271	13.549	ppbv	99
20) Methylene Chloride	2.059	84	59798	15.164	ppbv	93
21) Allyl Chloride	2.094	41	126959	13.724	ppbv	96
22) trans-1,2-Dichloroethene	2.337	96	77386	13.772	ppbv	94
23) Vinyl Acetate	2.457	43	232315	13.995	ppbv	99
24) 1,1-Dichloroethane	2.405	63	155017	14.292	ppbv	98
25) Ethyl Acetate	2.829	43	394074	13.265	ppbv	99
26) Hexane	2.823	57	172303	13.252	ppbv	97
27) Carbon Disulfide	2.150	76	191612	14.435	ppbv	97
28) Methyl tert-Butyl Ether	2.434	73	91132	14.279	ppbv	97
29) Chloroform	2.845	83	262803	14.221	ppbv	100
30) Cyclohexane	3.852	84	152015	13.946	ppbv	94
31) cis-1,2-Dichloroethene	2.716	61	180647	13.978	ppbv	98
32) 1,1,1-Trichloroethane	3.363	97	276915	13.930	ppbv	98
34) 2-Butanone	2.551	43	253855	13.914	ppbv	100
35) Carbon Tetrachloride	3.755	117	279560	14.250	ppbv	96
36) Benzene	3.651	78	348892	14.096	ppbv	99
37) 1,2-Dichloroethane	3.214	62	214248	14.798	ppbv	98
38) Trichloroethene	4.541	130	137027	14.123	ppbv	93
39) 1,2-Dichloropropane	4.302	63	128700	14.206	ppbv	98
40) 1,4-Dioxane	4.571	88	55905	13.728	ppbv #	95
41) Tetrahydrofuran	3.043	42	139722	13.868	ppbv	99
42) Bromodichloromethane	4.480	83	294008	14.721	ppbv	97
43) Methyl Methacrylate	4.872	69	179110	13.633	ppbv	100
44) 2,2,4-Trimethylpentane	4.632	57	595778	14.014	ppbv	99
45) t-1,3-Dichloropropene	6.406	75	177678	14.508	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043210.D
 Acq On : 18 Nov 2025 13:24
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Nov 19 00:53:26 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025


Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.587	75	205226	14.270	ppbv	100
47) 1,1,2-Trichloroethane	6.571	97	132701	13.738	ppbv	92
48) Dibromochloromethane	7.377	129	240736	14.564	ppbv	99
49) Bromoform	9.477	173	211325	14.434	ppbv	99
50) 4-Methyl-2-Pentanone	5.732	43	354421	13.790	ppbv	99
51) 2-Hexanone	7.422	43	280742	14.067	ppbv #	100
52) Tetrachloroethene	8.218	164	125132	13.591	ppbv	93
53) Toluene	6.924	91	414411	14.050	ppbv	99
54) 1,2-Dibromoethane	7.642	107	203913	14.413	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.917	131	214673	13.827	ppbv	99
57) Chlorobenzene	8.914	112	301986	13.923	ppbv	98
58) Ethyl Benzene	9.348	91	579890	14.262	ppbv	99
59) m/p-Xylene	9.545	91	916113m	28.000	ppbv	
60) o-Xylene	9.956	91	450260	13.826	ppbv	99
61) Styrene	9.862	104	204972	14.382	ppbv	99
62) Isopropylbenzene	10.532	105	803487	13.791	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.956	83	280254	13.557	ppbv	99
64) n-propylbenzene	10.986	120	205637	13.737	ppbv	100
65) tert-Butylbenzene	11.526	119	689352	13.249	ppbv	98
66) Benzyl Chloride	11.610	91	71386	13.520	ppbv	100
67) sec-Butylbenzene	11.762	105	972780	13.394	ppbv	100
69) p-Isopropyltoluene	11.921	119	836888	13.602	ppbv	99
70) n-Butylbenzene	12.277	91	838598	14.050	ppbv	99
71) 2-Chlorotoluene	10.895	91	625884	13.976	ppbv	100
72) 4-Ethyltoluene	11.121	105	553654	13.961	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	460783	13.846	ppbv	97
74) 1,2,4-Trimethylbenzene	11.533	105	486995	13.409	ppbv	97
75) 1,3-Dichlorobenzene	11.604	146	292143	13.784	ppbv	99
76) 1,4-Dichlorobenzene	11.668	146	294723	13.903	ppbv	99
77) 1,2-Dichlorobenzene	11.944	146	273467	13.420	ppbv	98
78) Hexachloro-1,3-Butadiene	13.876	225	211153	12.416	ppbv	99
79) Naphthalene	13.510	128	473229	14.670	ppbv	99
80) Naphthalene,2-methyl-	14.452	142	139019	16.418	ppbv	99
81) 1,2,4-Trichlorobenzene	13.436	180	216260	14.990	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 832
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 49 Resp: 98968
Ion Ratio Lower Upper
49 100
128 46.7 23.0 69.0
130 58.4 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#2
Dichlorodifluoromethane
Concen: 14.184 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 85 Resp: 197987
Ion Ratio Lower Upper
85 100
87 32.4 25.4 38.2









#3
Chlorodifluoromethane
Concen: 13.317 ppbv
RT: 1.476 min Scan# 428
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015

Tgt Ion: 51 Resp: 182813
Ion Ratio Lower Upper
51 100
67 19.7 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#4
Chloromethane
Concen: 13.918 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 50 Resp: 72636
Ion Ratio Lower Upper
50 100
52 31.7 24.1 36.1









#5
 Vinyl Chloride
 Concen: 12.557 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Tgt Ion: 62 Resp: 7167
 Ion Ratio Lower Upper
 62 100
 64 32.0 27.3 40.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#6
 Bromomethane
 Concen: 13.830 ppbv
 RT: 1.667 min Scan# 488
 Delta R.T. 0.003 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24


Tgt Ion: 94 Resp: 38265
 Ion Ratio Lower Upper
 94 100
 96 89.0 72.9 109.3





#7
Chloroethane
Concen: 13.309 ppbv
RT: 1.703 min Scan# 49
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015




Tgt Ion: 64 Resp: 29584
Ion Ratio Lower Upper
64 100
66 32.6 26.8 40.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#8
Dichlorotetrafluoroethane
Concen: 14.271 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24



Tgt Ion: 85 Resp: 165709
Ion Ratio Lower Upper
85 100
135 63.7 52.6 79.0









#9
 Propene
 Concen: 12.782 ppbv m
 RT: 1.483 min Scan# 431
 Delta R.T. 0.003 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Tgt Ion: 41 Resp: 81393
 Ion Ratio Lower Upper
 41 100
 42 4.8 3.7 5.5




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#10
 Heptane
 Concen: 12.934 ppbv
 RT: 4.972 min Scan# 1509
 Delta R.T. 0.006 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24


Tgt Ion: 43 Resp: 225935
 Ion Ratio Lower Upper
 43 100
 57 49.4 39.7 59.5





#11
Trichlorofluoromethane
Concen: 14.840 ppbv
RT: 1.874 min Scan# 552
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015




Tgt Ion:101 Resp: 200703
Ion Ratio Lower Upper
101 100
103 62.3 52.6 79.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#12
1,1,2-Trichlorotrifluoroethane
Concen: 14.469 ppbv
RT: 2.137 min Scan# 633
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24



Tgt Ion:101 Resp: 152772
Ion Ratio Lower Upper
101 100
151 72.3 59.2 88.8









#13
Ethanol
Concen: 12.425 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 45 Resp: 25468
Ion Ratio Lower Upper
45 100
46 37.9 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#14
Bromoethene
Concen: 13.789 ppbv
RT: 1.781 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:108 Resp: 55813
Ion Ratio Lower Upper
108 100
106 105.8 88.9 133.3
79 20.6 16.5 24.7









#15
Acetone
Concen: 11.712 ppbv
RT: 1.832 min Scan# 539
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015

Tgt Ion: 43 Resp: 161036
Ion Ratio Lower Upper
43 100
58 25.1 20.7 31.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#16
1,3-Butadiene
Concen: 13.226 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 39 Resp: 81077
Ion Ratio Lower Upper
39 100
54 73.5 59.4 89.2









#17
tert-Butyl alcohol
Concen: 14.274 ppbv
RT: 2.036 min Scan# 601
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015

Tgt Ion: 59 Resp: 25761
Ion Ratio Lower Upper
59 100
57 10.9 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#18
1,1-Dichloroethene
Concen: 13.370 ppbv
RT: 2.033 min Scan# 601
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24


Tgt Ion: 96 Resp: 65731
Ion Ratio Lower Upper
96 100
61 204.3 170.6 255.8
98 64.7 52.2 78.4





#19
Isopropyl Alcohol
Concen: 13.549 ppbv
RT: 1.881 min Scan# 554
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015




Tgt Ion: 45 Resp: 10527
Ion Ratio Lower Upper
45 100
58 38.5 31.3 46.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#20
Methylene Chloride
Concen: 15.164 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Tgt Ion: 84 Resp: 59798
Ion Ratio Lower Upper
84 100
49 153.8 132.9 199.3
51 47.0 34.8 52.2
86 58.2 48.2 72.2





#21
 Allyl Chloride
 Concen: 13.724 ppbv
 RT: 2.094 min Scan# 619
 Delta R.T. 0.003 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015




Tgt Ion: 41 Resp: 126959
 Ion Ratio Lower Upper
 41 100
 76 27.5 22.1 33.1
 39 81.3 61.0 91.4

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#22
 trans-1,2-Dichloroethene
 Concen: 13.772 ppbv
 RT: 2.337 min Scan# 695
 Delta R.T. 0.003 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24




Tgt Ion: 96 Resp: 77386
 Ion Ratio Lower Upper
 96 100
 61 184.6 139.0 208.6
 98 63.5 49.4 74.2





#23
 Vinyl Acetate
 Concen: 13.995 ppbv
 RT: 2.457 min Scan# 732
 Delta R.T. -0.000 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC015



Tgt Ion	Ratio	Lower	Upper
43	100		
86	8.2	6.2	9.2
42	10.1	8.0	12.0
44	5.1	4.0	6.0




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#24
 1,1-Dichloroethane
 Concen: 14.292 ppbv
 RT: 2.405 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24


Tgt Ion	Ratio	Lower	Upper
63	100		
98	4.9	2.8	8.4
100	3.0	1.3	3.9





#25
Ethyl Acetate
Concen: 13.265 ppbv
RT: 2.829 min Scan# 845
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015




Tgt Ion:	43	Resp:	394074
Ion Ratio	100	Lower	Upper
43	100		
45	9.5	8.1	12.1
61	8.4	7.0	10.4
70	8.2	7.0	10.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#26
Hexane
Concen: 13.252 ppbv
RT: 2.823 min Scan# 845
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24



Tgt Ion:	57	Resp:	172303
Ion Ratio	100	Lower	Upper
57	100		
41	95.4	74.6	111.8
43	229.0	179.0	268.4
56	53.5	42.3	63.5









#27
Carbon Disulfide
Concen: 14.435 ppbv
RT: 2.150 min Scan# 636
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 76 Resp: 191612
Ion Ratio Lower Upper
76 100
44 22.2 16.1 24.1
78 11.5 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#28
Methyl tert-Butyl Ether
Concen: 14.279 ppbv
RT: 2.434 min Scan# 725
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 73 Resp: 91132
Ion Ratio Lower Upper
73 100
57 23.5 19.8 29.8









#29
Chloroform
Concen: 14.221 ppbv
RT: 2.845 min Scan# 851
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 83 Resp: 262803
Ion Ratio Lower Upper
83 100
85 63.2 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#30
Cyclohexane
Concen: 13.946 ppbv
RT: 3.852 min Scan# 1163
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24


Tgt Ion: 84 Resp: 152015
Ion Ratio Lower Upper
84 100
56 123.0 104.7 157.1
41 82.7 69.0 103.6





#31
cis-1,2-Dichloroethene
Concen: 13.978 ppbv
RT: 2.716 min Scan# 81
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015




Tgt Ion: 61 Resp: 18064
Ion Ratio Lower Upper
61 100
96 60.9 49.3 73.9
98 38.3 32.1 48.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#32
1,1,1-Trichloroethane
Concen: 13.930 ppbv
RT: 3.363 min Scan# 1012
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Tgt Ion: 97 Resp: 276915
Ion Ratio Lower Upper
97 100
99 64.4 50.3 75.5
61 49.3 38.9 58.3





#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015



Tgt Ion:114 Resp: 258463
Ion Ratio Lower Upper
114 100
63 26.2 21.0 31.6
88 18.8 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#34
2-Butanone
Concen: 13.914 ppbv
RT: 2.551 min Scan# 761
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 43 Resp: 253855
Ion Ratio Lower Upper
43 100
72 22.4 18.1 27.1





#35

Carbon Tetrachloride

Concen: 14.250 ppbv

RT: 3.755 min Scan# 1133

Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC015

Tgt Ion:117 Resp: 279560

Ion Ratio Lower Upper

117 100

119 97.4 74.6 112.0





121 30.4 23.2 34.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#36

Benzene

Concen: 14.096 ppbv

RT: 3.651 min Scan# 1101

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion: 78 Resp: 348892


Ion Ratio Lower Upper

78 100

77 22.1 17.5 26.3

50 19.0 15.5 23.3









#37
1,2-Dichloroethane
Concen: 14.798 ppbv
RT: 3.214 min Scan# 965
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 62 Resp: 214248
Ion Ratio Lower Upper
62 100
98 6.1 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#38
Trichloroethene
Concen: 14.123 ppbv
RT: 4.541 min Scan# 1376
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:130 Resp: 137027
Ion Ratio Lower Upper
130 100
95 109.6 85.7 128.5
132 97.4 73.4 110.0
97 77.0 52.7 79.1





#39
1,2-Dichloropropane
Concen: 14.206 ppbv
RT: 4.302 min Scan# 1302
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L





ClientSampleId :

VSTDICC015

Tgt Ion: 63 Resp: 128700
Ion Ratio Lower Upper
63 100
41 75.9 62.6 93.8
62 65.6 51.0 76.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#40
1,4-Dioxane
Concen: 13.728 ppbv
RT: 4.571 min Scan# 1385
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 88 Resp: 55905
Ion Ratio Lower Upper
88 100
58 76.7 63.0 94.6
43 34.0 32.5 48.7
57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 13.868 ppbv
RT: 3.043 min Scan# 914
Delta R.T. -0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 42 Resp: 139722
Ion Ratio Lower Upper
42 100
72 39.1 30.6 46.0
71 37.3 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#42
Bromodichloromethane
Concen: 14.721 ppbv
RT: 4.480 min Scan# 1357
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 83 Resp: 294008
Ion Ratio Lower Upper
83 100
85 61.7 51.7 77.5
127 7.6 6.3 9.5









#43
Methyl Methacrylate
Concen: 13.633 ppbv
RT: 4.872 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 69 Resp: 179110
Ion Ratio Lower Upper
69 100
41 157.9 126.7 190.1
100 35.7 28.6 43.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#44
2,2,4-Trimethylpentane
Concen: 14.014 ppbv
RT: 4.632 min Scan# 1404
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 57 Resp: 595778
Ion Ratio Lower Upper
57 100
41 35.9 28.0 42.0
56 31.6 25.6 38.4





#45

t-1,3-Dichloropropene

Concen: 14.508 ppbv

RT: 6.406 min Scan# 1950

Delta R.T. 0.006 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC015

Tgt Ion: 75 Resp: 177678

Ion Ratio Lower Upper

75 100





77 30.2 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#46

cis-1,3-Dichloropropene

Concen: 14.270 ppbv

RT: 5.587 min Scan# 1699

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion: 75 Resp: 205226


Ion Ratio Lower Upper

75 100

39 73.9 59.1 88.7

77 29.8 24.2 36.2









#47
1,1,2-Trichloroethane
Concen: 13.738 ppbv
RT: 6.571 min Scan# 2003
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 97 Resp: 132701
Ion Ratio Lower Upper
97 100
83 90.4 64.6 97.0
85 53.2 39.4 59.2
99 62.2 47.0 70.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#48
Dibromochloromethane
Concen: 14.564 ppbv
RT: 7.377 min Scan# 2252
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:129 Resp: 240736
Ion Ratio Lower Upper
129 100
127 77.5 39.3 117.9









#49
Bromoform
Concen: 14.434 ppbv
RT: 9.477 min Scan# 2900
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015

Tgt Ion: 173 Resp: 211325
Ion Ratio Lower Upper
173 100
175 48.6 38.6 58.0
171 51.5 40.6 61.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#50
4-Methyl-2-Pentanone
Concen: 13.790 ppbv
RT: 5.732 min Scan# 1744
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24


Tgt Ion: 43 Resp: 354421
Ion Ratio Lower Upper
43 100
58 35.9 28.4 42.6





#51
2-Hexanone
Concen: 14.067 ppbv
RT: 7.422 min Scan# 21
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015



Tgt Ion: 43 Resp: 280742
Ion Ratio Lower Upper
43 100
58 47.7 38.4 57.6
98 0.1 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#52
Tetrachloroethene
Concen: 13.591 ppbv
RT: 8.218 min Scan# 2512
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:164 Resp: 125132
Ion Ratio Lower Upper
164 100
166 128.8 99.4 149.2
129 110.8 81.8 122.8
131 104.7 77.4 116.2





#53

Toluene

Concen: 14.050 ppbv

RT: 6.924 min Scan# 2111

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC015

Tgt Ion: 91 Resp: 41441

Ion Ratio Lower Upper

91 100





92 58.8 47.4 71.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#54

1,2-Dibromoethane

Concen: 14.413 ppbv

RT: 7.642 min Scan# 2334

Delta R.T. 0.003 min

Lab File: VL043210.D




Acq: 18 Nov 2025 13:24


Tgt Ion:107 Resp: 203913

Ion Ratio Lower Upper

107 100

109 90.6 72.1 108.1





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.872 min Scan# 21124

Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC015

Tgt Ion:117 Resp: 21124

Ion	Ratio	Lower	Upper
117	100		
82	68.4	56.1	84.1
119	32.6	25.0	37.6

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#56

1,1,1,2-Tetrachloroethane

Concen: 13.827 ppbv

RT: 8.917 min Scan# 2728




Delta R.T. 0.003 min


Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Tgt Ion:131 Resp: 214673

Ion	Ratio	Lower	Upper
131	100		
133	94.9	77.4	116.2
119	61.5	49.1	73.7









#57
Chlorobenzene
Concen: 13.923 ppbv
RT: 8.914 min Scan# 2726
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015

Tgt Ion:112 Resp: 301986
Ion Ratio Lower Upper
112 100
77 71.1 55.5 83.3
114 31.5 29.1 35.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#58
Ethyl Benzene
Concen: 14.262 ppbv
RT: 9.348 min Scan# 2861
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 91 Resp: 579890
Ion Ratio Lower Upper
91 100
106 29.2 22.8 34.2









#59
m/p-Xylene
Concen: 28.000 ppbv m
RT: 9.545 min Scan# 2921
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 91 Resp: 916113
Ion Ratio Lower Upper
91 100
106 45.1 36.1 54.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#60
o-Xylene
Concen: 13.826 ppbv
RT: 9.956 min Scan# 3049
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 91 Resp: 450260
Ion Ratio Lower Upper
91 100
106 42.4 20.9 62.8









#61
Styrene
Concen: 14.382 ppbv
RT: 9.862 min Scan# 3020
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion:104 Resp: 204972
Ion Ratio Lower Upper
104 100
78 53.7 42.1 63.1
103 46.9 37.2 55.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#62
Isopropylbenzene
Concen: 13.791 ppbv
RT: 10.532 min Scan# 3227
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:105 Resp: 803487
Ion Ratio Lower Upper
105 100
120 25.3 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 13.557 ppbv
RT: 9.956 min Scan# 3048
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 83 Resp: 280254
Ion Ratio Lower Upper
83 100
131 9.7 4.8 14.3
85 64.9 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#64
n-propylbenzene
Concen: 13.737 ppbv
RT: 10.986 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:120 Resp: 205637
Ion Ratio Lower Upper
120 100
91 471.6 377.1 565.7









#65
tert-Butylbenzene
Concen: 13.249 ppbv
RT: 11.526 min Scan# 3533
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion: 119 Resp: 689352
Ion Ratio Lower Upper
119 100
91 83.7 65.3 97.9
134 20.9 17.0 25.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#66
Benzyl Chloride
Concen: 13.520 ppbv
RT: 11.610 min Scan# 3560
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 91 Resp: 71386
Ion Ratio Lower Upper
91 100
126 17.9 14.4 21.6
128 5.9 4.6 6.8





#67

sec-Butylbenzene

Concen: 13.394 ppbv

RT: 11.762 min Scan# 3607

Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC015

Tgt Ion:105 Resp: 972780

Ion Ratio Lower Upper

105 100





134 18.9 15.0 22.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#68

1-Bromo-4-Fluorobenzene

Concen: 10.321 ppbv

RT: 10.367 min Scan# 3176

Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion: 95 Resp: 159814


Ion Ratio Lower Upper

95 100

174 60.3 50.2 75.2


175 4.7 4.1 6.1





#69
p-Isopropyltoluene
Concen: 13.602 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015




Tgt Ion:119 Resp: 836888
Ion Ratio Lower Upper
119 100
134 24.8 20.2 30.2
91 30.3 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#70
n-Butylbenzene
Concen: 14.050 ppbv
RT: 12.277 min Scan# 3766
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24



Tgt Ion: 91 Resp: 838598
Ion Ratio Lower Upper
91 100
92 54.9 43.5 65.3
134 23.4 19.0 28.6









#71
2-Chlorotoluene
Concen: 13.976 ppbv
RT: 10.895 min Scan# 3339
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015

Tgt Ion: 91 Resp: 625884
Ion Ratio Lower Upper
91 100
126 28.9 11.7 46.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#72
4-Ethyltoluene
Concen: 13.961 ppbv
RT: 11.121 min Scan# 3409
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:105 Resp: 553654
Ion Ratio Lower Upper
105 100
120 29.3 23.8 35.8
91 13.2 10.3 15.5
77 14.1 11.1 16.7









#73
1,3,5-Trimethylbenzene
Concen: 13.846 ppbv
RT: 11.199 min Scan# 3432
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Tgt Ion:105 Resp: 460783
Ion Ratio Lower Upper
105 100
120 45.9 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#74
1,2,4-Trimethylbenzene
Concen: 13.409 ppbv
RT: 11.533 min Scan# 3536
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:105 Resp: 486995
Ion Ratio Lower Upper
105 100
120 50.4 41.7 62.5
91 83.4 70.0 105.0
77 28.5 23.8 35.8





#75

1,3-Dichlorobenzene

Concen: 13.784 ppbv

RT: 11.604 min Scan# 3557

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_1





ClientSampleId :

VSTDIC015

Tgt Ion:146	Resp:	292143
Ion Ratio	Lower	Upper
146	100	
111	48.4	23.4 70.3
148	63.1	31.6 95.0

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#76

1,4-Dichlorobenzene

Concen: 13.903 ppbv




RT: 11.668 min Scan# 3578


Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Tgt Ion:146	Resp:	294723
Ion Ratio	Lower	Upper
146	100	
111	47.2	22.8 68.3
148	63.2	31.8 95.4





#77

1,2-Dichlorobenzene

Concen: 13.420 ppbv

RT: 11.944 min Scan# 3663

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC015

Tgt Ion:146 Resp: 273462

Ion Ratio Lower Upper

146 100

111 50.2 24.3 73.0





148 63.9 31.6 94.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#78

Hexachloro-1,3-Butadiene

Concen: 12.416 ppbv

RT: 13.876 min Scan# 4260

Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion:225 Resp: 211153


Ion Ratio Lower Upper

225 100

223 64.9 50.2 75.4

227 63.9 51.1 76.7





#79

Naphthalene

Concen: 14.670 ppbv

RT: 13.510 min Scan# 4146

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC015

Tgt Ion:	128	Resp:	473229
Ion Ratio	Lower	Upper	
128	100		
127	12.4	9.9	14.9
129	11.1	8.4	12.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#80

Naphthalene, 2-methyl-

Concen: 16.418 ppbv




RT: 14.452 min Scan# 4438


Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Tgt Ion:	142	Resp:	139019
Ion Ratio	Lower	Upper	
142	100		
141	82.6	67.1	100.7
115	40.2	31.9	47.9





#81

1,2,4-Trichlorobenzene

Concen: 14.990 ppbv

RT: 13.436 min Scan# 41

Delta R.T. -0.000 min

Lab File: VL043210.D


Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC015



Tgt Ion:180 Resp: 216260

Ion Ratio Lower Upper

180 100

182 95.3 77.4 116.2



184 30.1 24.2 36.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 01:34:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	97854	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.956	114	261914	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.879	117	216212	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.374 95 159242 10.048 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	133374	9.664	ppbv	98
3) Chlorodifluoromethane	1.476	51	124858	9.199	ppbv	98
4) Chloromethane	1.535	50	46906	9.090	ppbv	95
5) Vinyl Chloride	1.580	62	47932	8.493	ppbv	95
6) Bromomethane	1.667	94	22230	8.126	ppbv	99
7) Chloroethane	1.703	64	19854	9.033	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	110587	9.632	ppbv	97
9) Propene	1.483	41	54006m	9.370	ppbv	
10) Heptane	4.975	43	149202	8.698	ppbv	99
11) Trichlorofluoromethane	1.878	101	134746	10.077	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	100984	9.673	ppbv	99
13) Ethanol	1.719	45	15371	7.584	ppbv	92
14) Bromoethene	1.781	108	36979	9.240	ppbv	98
15) Acetone	1.832	43	104830	7.711	ppbv	99
16) 1,3-Butadiene	1.609	39	54795	9.041	ppbv	96
17) tert-Butyl alcohol	2.036	59	164987	9.245	ppbv	99
18) 1,1-Dichloroethene	2.033	96	43833	9.017	ppbv	99
19) Isopropyl Alcohol	1.884	45	64772	8.431	ppbv	97
20) Methylene Chloride	2.062	84	41254	10.323	ppbv	95
21) Allyl Chloride	2.095	41	75516	8.256	ppbv	97
22) trans-1,2-Dichloroethene	2.341	96	50228	9.041	ppbv	92
23) Vinyl Acetate	2.460	43	140156	8.539	ppbv	98
24) 1,1-Dichloroethane	2.409	63	103187	9.622	ppbv	97
25) Ethyl Acetate	2.829	43	258609	8.820	ppbv	99
26) Hexane	2.826	57	114312	8.892	ppbv	98
27) Carbon Disulfide	2.150	76	125048	9.528	ppbv	97
28) Methyl tert-Butyl Ether	2.438	73	64183	10.171	ppbv	98
29) Chloroform	2.846	83	173977	9.521	ppbv	99
30) Cyclohexane	3.852	84	98598	9.148	ppbv	97
31) cis-1,2-Dichloroethene	2.719	61	119843	9.379	ppbv	98
32) 1,1,1-Trichloroethane	3.363	97	182262	9.358	ppbv	98
34) 2-Butanone	2.551	43	160977	8.707	ppbv	98
35) Carbon Tetrachloride	3.761	117	182902	9.200	ppbv	98
36) Benzene	3.655	78	233094	9.293	ppbv	99
37) 1,2-Dichloroethane	3.214	62	143363	9.772	ppbv	98
38) Trichloroethene	4.542	130	90993	9.272	ppbv	95
39) 1,2-Dichloropropane	4.305	63	83416	9.086	ppbv	95
40) 1,4-Dioxane	4.571	88	32109	7.780	ppbv #	95
41) Tetrahydrofuran	3.046	42	90769	8.891	ppbv	98
42) Bromodichloromethane	4.486	83	195442	9.657	ppbv	96
43) Methyl Methacrylate	4.872	69	116325	8.737	ppbv	98
44) 2,2,4-Trimethylpentane	4.635	57	393438	9.133	ppbv	99
45) t-1,3-Dichloropropene	6.409	75	97587	7.863	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 01:34:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M


Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	115476	7.924	ppbv	97
47) 1,1,2-Trichloroethane	6.577	97	88525	9.086	ppbv	95
48) Dibromochloromethane	7.380	129	159365	9.514	ppbv	99
49) Bromoform	9.477	173	134790	9.085	ppbv	99
50) 4-Methyl-2-Pentanone	5.742	43	224791	8.631	ppbv	99
51) 2-Hexanone	7.429	43	166506	8.233	ppbv #	100
52) Tetrachloroethene	8.222	164	82890	8.884	ppbv	98
53) Toluene	6.924	91	273172	9.197	ppbv	99
54) 1,2-Dibromoethane	7.645	107	124804	8.705	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.921	131	142512	8.969	ppbv	98
57) Chlorobenzene	8.917	112	199373	8.981	ppbv	96
58) Ethyl Benzene	9.351	91	381474	9.167	ppbv	97
59) m/p-Xylene	9.549	91	604165m	18.108	ppbv	
60) o-Xylene	9.960	91	295836	8.876	ppbv	99
61) Styrene	9.866	104	127117	8.714	ppbv	97
62) Isopropylbenzene	10.536	105	525717	8.816	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.960	83	178591	8.421	ppbv	100
64) n-propylbenzene	10.986	120	131634	8.591	ppbv	96
65) tert-Butylbenzene	11.529	119	450819	8.465	ppbv	99
66) Benzyl Chloride	11.614	91	27878	5.159	ppbv	99
67) sec-Butylbenzene	11.766	105	623520	8.388	ppbv	100
69) p-Isopropyltoluene	11.924	119	520880	8.272	ppbv	99
70) n-Butylbenzene	12.280	91	513027	8.398	ppbv	99
71) 2-Chlorotoluene	10.898	91	396434	8.649	ppbv	99
72) 4-Ethyltoluene	11.122	105	356325	8.779	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	294977	8.660	ppbv	95
74) 1,2,4-Trimethylbenzene	11.533	105	314606	8.463	ppbv	94
75) 1,3-Dichlorobenzene	11.604	146	180491	8.320	ppbv	99
76) 1,4-Dichlorobenzene	11.669	146	179491	8.273	ppbv	99
77) 1,2-Dichlorobenzene	11.944	146	166137	7.966	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	125707	7.222	ppbv	100
79) Naphthalene	13.510	128	273029	8.270	ppbv	99
80) Naphthalene,2-methyl-	14.455	142	74802	8.631	ppbv	97
81) 1,2,4-Trichlorobenzene	13.436	180	125470	8.497	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 832
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion: 49 Resp: 97854
Ion Ratio Lower Upper
49 100
128 46.7 23.0 69.0
130 59.0 29.9 89.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#2
Dichlorodifluoromethane
Concen: 9.664 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion: 85 Resp: 133374
Ion Ratio Lower Upper
85 100
87 33.0 25.4 38.2





#3
Chlorodifluoromethane
Concen: 9.199 ppbv
RT: 1.476 min Scan# 428
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion: 51 Resp: 124858
Ion Ratio Lower Upper
51 100
67 19.6 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#4
Chloromethane
Concen: 9.090 ppbv
RT: 1.535 min Scan# 447
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion: 50 Resp: 46906
Ion Ratio Lower Upper
50 100
52 32.8 24.1 36.1





#5
 Vinyl Chloride
 Concen: 8.493 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825




Tgt Ion: 62 Resp: 47932
 Ion Ratio Lower Upper
 62 100
 64 31.4 27.3 40.9

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#6
 Bromomethane
 Concen: 8.126 ppbv
 RT: 1.667 min Scan# 488
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56



Tgt Ion: 94 Resp: 22230
 Ion Ratio Lower Upper
 94 100
 96 92.2 72.9 109.3









#7
Chloroethane
Concen: 9.033 ppbv
RT: 1.703 min Scan# 498
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument : MSVOA_L
ClientSampleId : ICVVL111825

Tgt Ion: 64 Resp: 19854
Ion Ratio Lower Upper
64 100
66 32.4 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#8
Dichlorotetrafluoroethane
Concen: 9.632 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 85 Resp: 110587
Ion Ratio Lower Upper
85 100
135 63.5 52.6 79.0









#9
Propene
Concen: 9.370 ppbv m
RT: 1.483 min Scan# 431
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion: 41 Resp: 54006
Ion Ratio Lower Upper
41 100
42 4.4 3.7 5.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#10
Heptane
Concen: 8.698 ppbv
RT: 4.975 min Scan# 1510
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion: 43 Resp: 149202
Ion Ratio Lower Upper
43 100
57 49.1 39.7 59.5





#11
Trichlorofluoromethane
Concen: 10.077 ppbv
RT: 1.878 min Scan# 552
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion:101 Resp: 134746
Ion Ratio Lower Upper
101 100
103 63.8 52.6 79.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#12
1,1,2-Trichlorotrifluoroethane
Concen: 9.673 ppbv
RT: 2.140 min Scan# 634
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56



Tgt Ion:101 Resp: 100984
Ion Ratio Lower Upper
101 100
151 73.1 59.2 88.8









#13
Ethanol
Concen: 7.584 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion: 45 Resp: 1537
Ion Ratio Lower Upper
45 100
46 42.2 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#14
Bromoethene
Concen: 9.240 ppbv
RT: 1.781 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion:108 Resp: 36979
Ion Ratio Lower Upper
108 100
106 109.3 88.9 133.3
79 21.9 16.5 24.7





#15
Acetone
Concen: 7.711 ppbv
RT: 1.832 min Scan# 539
Delta R.T. 0.000 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion: 43 Resp: 104830
Ion Ratio Lower Upper
43 100
58 25.2 20.7 31.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#16
1,3-Butadiene
Concen: 9.041 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion: 39 Resp: 54795
Ion Ratio Lower Upper
39 100
54 71.3 59.4 89.2





#17
tert-Butyl alcohol
Concen: 9.245 ppbv
RT: 2.036 min Scan# 601
Delta R.T. 0.000 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion: 59 Resp: 16498
Ion Ratio Lower Upper
59 100
57 10.8 9.0 13.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#18
1,1-Dichloroethene
Concen: 9.017 ppbv
RT: 2.033 min Scan# 601
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56



Tgt Ion: 96 Resp: 43833
Ion Ratio Lower Upper
96 100
61 214.0 170.6 255.8
98 62.9 52.2 78.4









#19
Isopropyl Alcohol
Concen: 8.431 ppbv
RT: 1.884 min Scan# 511
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion: 45 Resp: 64772
Ion Ratio Lower Upper
45 100
58 40.8 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#20
Methylene Chloride
Concen: 10.323 ppbv
RT: 2.062 min Scan# 610
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion: 84 Resp: 41254
Ion Ratio Lower Upper
84 100
49 155.1 132.9 199.3
51 43.9 34.8 52.2
86 60.8 48.2 72.2





#21
 Allyl Chloride
 Concen: 8.256 ppbv
 RT: 2.095 min Scan# 61
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825




Tgt Ion: 41 Resp: 75516
 Ion Ratio Lower Upper
 41 100
 76 27.9 22.1 33.1
 39 79.3 61.0 91.4

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#22
 trans-1,2-Dichloroethene
 Concen: 9.041 ppbv
 RT: 2.341 min Scan# 696
 Delta R.T. 0.006 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Tgt Ion: 96 Resp: 50228
 Ion Ratio Lower Upper
 96 100
 61 185.7 139.0 208.6
 98 66.6 49.4 74.2





#23
 Vinyl Acetate
 Concen: 8.539 ppbv
 RT: 2.460 min Scan# 732
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825




Tgt Ion: 43 Resp: 140156
 Ion Ratio Lower Upper
 43 100
 86 8.4 6.2 9.2
 42 9.1 8.0 12.0
 44 4.4 4.0 6.0

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#24
 1,1-Dichloroethane
 Concen: 9.622 ppbv
 RT: 2.409 min Scan# 717
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Tgt Ion: 63 Resp: 103187
 Ion Ratio Lower Upper
 63 100
 98 4.1 2.8 8.4
 100 2.6 1.3 3.9





#25
Ethyl Acetate
Concen: 8.820 ppbv
RT: 2.829 min Scan# 847
Delta R.T. 0.000 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion: 43 Resp: 258609
Ion Ratio Lower Upper
43 100
45 9.6 8.1 12.1
61 8.5 7.0 10.4
70 8.2 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#26
Hexane
Concen: 8.892 ppbv
RT: 2.826 min Scan# 846
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion: 57 Resp: 114312
Ion Ratio Lower Upper
57 100
41 96.0 74.6 111.8
43 226.4 179.0 268.4
56 53.6 42.3 63.5





#27
Carbon Disulfide
Concen: 9.528 ppbv
RT: 2.150 min Scan# 637
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument : MSVOA_1
ClientSampleId : ICVVL111825




Tgt Ion: 76 Resp: 125048
Ion Ratio Lower Upper
76 100
44 21.9 16.1 24.1
78 11.6 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#28
Methyl tert-Butyl Ether
Concen: 10.171 ppbv
RT: 2.438 min Scan# 726
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56



Tgt Ion: 73 Resp: 64183
Ion Ratio Lower Upper
73 100
57 23.6 19.8 29.8









#29
Chloroform
Concen: 9.521 ppbv
RT: 2.846 min Scan# 851
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion: 83 Resp: 173972
Ion Ratio Lower Upper
83 100
85 64.2 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#30
Cyclohexane
Concen: 9.148 ppbv
RT: 3.852 min Scan# 1163
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion: 84 Resp: 98598
Ion Ratio Lower Upper
84 100
56 127.4 104.7 157.1
41 84.5 69.0 103.6





#31
 cis-1,2-Dichloroethene
 Concen: 9.379 ppbv
 RT: 2.719 min Scan# 81
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825




Tgt Ion: 61 Resp: 119843
 Ion Ratio Lower Upper
 61 100
 96 62.9 49.3 73.9
 98 42.0 32.1 48.1

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025




#32
 1,1,1-Trichloroethane
 Concen: 9.358 ppbv
 RT: 3.363 min Scan# 1012
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56



Tgt Ion: 97 Resp: 182262
 Ion Ratio Lower Upper
 97 100
 99 64.2 50.3 75.5
 61 49.9 38.9 58.3





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.956 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :





ICVVL111825

Tgt Ion: 114 Resp: 261914
 Ion Ratio Lower Upper
 114 100
 63 26.0 21.0 31.6
 88 18.3 14.9 22.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#34

2-Butanone

Concen: 8.707 ppbv




RT: 2.551 min Scan# 761


Delta R.T. 0.000 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Tgt Ion: 43 Resp: 160977
 Ion Ratio Lower Upper
 43 100
 72 21.8 18.1 27.1









#35
Carbon Tetrachloride
Concen: 9.200 ppbv
RT: 3.761 min Scan# 1135
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument : MSVOA_L
ClientSampleId : ICVVL111825

Tgt Ion: 117 Resp: 182902
Ion Ratio Lower Upper
117 100
119 94.7 74.6 112.0
121 30.4 23.2 34.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#36
Benzene
Concen: 9.293 ppbv
RT: 3.655 min Scan# 1102
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 78 Resp: 233094
Ion Ratio Lower Upper
78 100
77 21.8 17.5 26.3
50 20.4 15.5 23.3









#37
1,2-Dichloroethane
Concen: 9.772 ppbv
RT: 3.214 min Scan# 965
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion: 62 Resp: 143363
Ion Ratio Lower Upper
62 100
98 6.3 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#38
Trichloroethene
Concen: 9.272 ppbv
RT: 4.542 min Scan# 1376
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion:130 Resp: 90993
Ion Ratio Lower Upper
130 100
95 112.7 85.7 128.5
132 94.1 73.4 110.0
97 70.6 52.7 79.1





#39

1,2-Dichloropropane

Concen: 9.086 ppbv

RT: 4.305 min Scan# 1302

Delta R.T. 0.003 min

Lab File: VL043211.D


Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :

ICVVL111825



Tgt Ion: 63 Resp: 83416

Ion Ratio Lower Upper

63 100

41 85.0 62.6 93.8




62 64.6 51.0 76.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025

Supervised By :Mahesh Dadoda 11/19/2025



#40

1,4-Dioxane

Concen: 7.780 ppbv

RT: 4.571 min Scan# 1385

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Tgt Ion: 88 Resp: 32109




Ion Ratio Lower Upper


88 100

58 74.8 63.0 94.6

43 37.4 32.5 48.7


57 0.0 0.0 0.0





#41
Tetrahydrofuran
Concen: 8.891 ppbv
RT: 3.046 min Scan# 914
Delta R.T. 0.000 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion: 42 Resp: 90769
Ion Ratio Lower Upper
42 100
72 38.0 30.6 46.0
71 36.1 30.5 45.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#42
Bromodichloromethane
Concen: 9.657 ppbv
RT: 4.486 min Scan# 1359
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion: 83 Resp: 195442
Ion Ratio Lower Upper
83 100
85 61.0 51.7 77.5
127 7.6 6.3 9.5





#43
Methyl Methacrylate
Concen: 8.737 ppbv
RT: 4.872 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825




Tgt Ion: 69 Resp: 116325
Ion Ratio Lower Upper
69 100
41 155.2 126.7 190.1
100 36.5 28.6 43.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#44
2,2,4-Trimethylpentane
Concen: 9.133 ppbv
RT: 4.635 min Scan# 1405
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56



Tgt Ion: 57 Resp: 393438
Ion Ratio Lower Upper
57 100
41 35.8 28.0 42.0
56 31.5 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 7.863 ppbv
RT: 6.409 min Scan# 1950
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion: 75 Resp: 9758
Ion Ratio Lower Upper
75 100
77 30.3 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#46
cis-1,3-Dichloropropene
Concen: 7.924 ppbv
RT: 5.593 min Scan# 1701
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion: 75 Resp: 115476
Ion Ratio Lower Upper
75 100
39 70.4 59.1 88.7
77 30.4 24.2 36.2





#47
1,1,2-Trichloroethane
Concen: 9.086 ppbv
RT: 6.577 min Scan# 2005
Delta R.T. 0.013 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion	Ratio	Lower	Upper
97	100		
83	85.4	64.6	97.0
85	52.6	39.4	59.2
99	62.6	47.0	70.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#48
Dibromochloromethane
Concen: 9.514 ppbv
RT: 7.380 min Scan# 2253
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion	Ratio	Lower	Upper
129	100		
127	77.3	39.3	117.9





#49
Bromoform
Concen: 9.085 ppbv
RT: 9.477 min Scan# 2900
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
Client Sample Id :
ICVVL111825




Tgt Ion: 173 Resp: 134790
Ion Ratio Lower Upper
173 100
175 47.2 38.6 58.0
171 50.2 40.6 61.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#50
4-Methyl-2-Pentanone
Concen: 8.631 ppbv
RT: 5.742 min Scan# 1747
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56



Tgt Ion: 43 Resp: 224791
Ion Ratio Lower Upper
43 100
58 36.0 28.4 42.6









#51
2-Hexanone
Concen: 8.233 ppbv
RT: 7.429 min Scan# 2113
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument : MSVOA_L
ClientSampleId : ICVVL111825

Tgt Ion: 43 Resp: 166500
Ion Ratio Lower Upper
43 100
58 47.9 38.4 57.6
98 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#52
Tetrachloroethene
Concen: 8.884 ppbv
RT: 8.222 min Scan# 2513
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion:164 Resp: 82890
Ion Ratio Lower Upper
164 100
166 121.3 99.4 149.2
129 100.1 81.8 122.8
131 94.6 77.4 116.2









#53
Toluene
Concen: 9.197 ppbv
RT: 6.924 min Scan# 2112
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
Client Sample Id :
ICVVL111825

Tgt Ion: 91 Resp: 273172
Ion Ratio Lower Upper
91 100
92 59.8 47.4 71.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#54
1,2-Dibromoethane
Concen: 8.705 ppbv
RT: 7.645 min Scan# 2335
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion:107 Resp: 124804
Ion Ratio Lower Upper
107 100
109 91.1 72.1 108.1





#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.879 min Scan# 2714
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion:117 Resp: 216212
Ion Ratio Lower Upper
117 100
82 70.6 56.1 84.1
119 32.3 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#56
1,1,1,2-Tetrachloroethane
Concen: 8.969 ppbv
RT: 8.921 min Scan# 2729
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion:131 Resp: 142512
Ion Ratio Lower Upper
131 100
133 94.3 77.4 116.2
119 61.9 49.1 73.7





#57
Chlorobenzene
Concen: 8.981 ppbv
RT: 8.917 min Scan# 211
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion: 112 Resp: 199373
Ion Ratio Lower Upper
112 100
77 72.7 55.5 83.3
114 35.1 29.1 35.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#58
Ethyl Benzene
Concen: 9.167 ppbv
RT: 9.351 min Scan# 2862
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56



Tgt Ion: 91 Resp: 381474
Ion Ratio Lower Upper
91 100
106 30.2 22.8 34.2









#59
m/p-Xylene
Concen: 18.108 ppbv m
RT: 9.549 min Scan# 2921
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion: 91 Resp: 604165
Ion Ratio Lower Upper
91 100
106 44.8 36.1 54.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#60
o-Xylene
Concen: 8.876 ppbv
RT: 9.960 min Scan# 3050
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 91 Resp: 295836
Ion Ratio Lower Upper
91 100
106 42.8 20.9 62.8





#61

Styrene

Concen: 8.714 ppbv

RT: 9.866 min Scan# 3020

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :





ICVVL111825

Tgt Ion:104	Resp:	127117
Ion Ratio	Lower	Upper
104	100	
78	54.7	42.1 63.1
103	48.0	37.2 55.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#62

Isopropylbenzene

Concen: 8.816 ppbv




RT: 10.536 min Scan# 3228


Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Tgt Ion:105	Resp:	525717
Ion Ratio	Lower	Upper
105	100	
120	25.1	19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 8.421 ppbv
RT: 9.960 min Scan# 3048
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion: 83 Resp: 17859
Ion Ratio Lower Upper
83 100
131 9.6 4.8 14.3
85 63.9 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#64
n-propylbenzene
Concen: 8.591 ppbv
RT: 10.986 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion:120 Resp: 131634
Ion Ratio Lower Upper
120 100
91 482.2 377.1 565.7









#65
tert-Butylbenzene
Concen: 8.465 ppbv
RT: 11.529 min Scan# 3533
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion:119 Resp: 450819
Ion Ratio Lower Upper
119 100
91 83.0 65.3 97.9
134 20.8 17.0 25.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#66
Benzyl Chloride
Concen: 5.159 ppbv
RT: 11.614 min Scan# 3561
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion: 91 Resp: 27878
Ion Ratio Lower Upper
91 100
126 17.3 14.4 21.6
128 5.6 4.6 6.8





#67
 sec-Butylbenzene
 Concen: 8.388 ppbv
 RT: 11.766 min Scan# 3608
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_1
 ClientSampleId :
 ICVVL111825



Tgt Ion: 105 Resp: 623520
 Ion Ratio Lower Upper
 105 100
 134 19.0 15.0 22.6




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.048 ppbv
 RT: 10.374 min Scan# 3178
 Delta R.T. 0.006 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56


Tgt Ion: 95 Resp: 159242
 Ion Ratio Lower Upper
 95 100
 174 60.4 50.2 75.2
 175 4.7 4.1 6.1





#69
p-Isopropyltoluene
Concen: 8.272 ppbv
RT: 11.924 min Scan# 3657
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion: 119 Resp: 520880
Ion Ratio Lower Upper
119 100
134 25.2 20.2 30.2
91 30.6 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#70
n-Butylbenzene
Concen: 8.398 ppbv
RT: 12.280 min Scan# 3767
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion: 91 Resp: 513027
Ion Ratio Lower Upper
91 100
92 54.0 43.5 65.3
134 23.1 19.0 28.6





#71
2-Chlorotoluene
Concen: 8.649 ppbv
RT: 10.898 min Scan# 3339
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825



Tgt Ion: 91 Resp: 396434
Ion Ratio Lower Upper
91 100
126 28.9 11.7 46.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#72
4-Ethyltoluene
Concen: 8.779 ppbv
RT: 11.122 min Scan# 3409
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion:105 Resp: 356325
Ion Ratio Lower Upper
105 100
120 29.3 23.8 35.8
91 13.5 10.3 15.5
77 13.9 11.1 16.7





#73
1,3,5-Trimethylbenzene
Concen: 8.660 ppbv
RT: 11.199 min Scan# 3432
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion:105 Resp: 294972
Ion Ratio Lower Upper
105 100
120 45.0 38.6 57.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#74
1,2,4-Trimethylbenzene
Concen: 8.463 ppbv
RT: 11.533 min Scan# 3536
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56



Tgt Ion:105 Resp: 314606
Ion Ratio Lower Upper
105 100
120 51.8 41.7 62.5
91 96.9 70.0 105.0
77 32.1 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 8.320 ppbv
RT: 11.604 min Scan# 3557
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion:146 Resp: 180491
Ion Ratio Lower Upper
146 100
111 47.7 23.4 70.3
148 62.7 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#76
1,4-Dichlorobenzene
Concen: 8.273 ppbv
RT: 11.669 min Scan# 3578
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion:146 Resp: 179491
Ion Ratio Lower Upper
146 100
111 46.3 22.8 68.3
148 63.4 31.8 95.4









#77
1,2-Dichlorobenzene
Concen: 7.966 ppbv
RT: 11.944 min Scan# 3662
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Tgt Ion:146 Resp: 166137
Ion Ratio Lower Upper
146 100
111 49.2 24.3 73.0
148 63.9 31.6 94.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025



#78
Hexachloro-1,3-Butadiene
Concen: 7.222 ppbv
RT: 13.879 min Scan# 4261
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion:225 Resp: 125707
Ion Ratio Lower Upper
225 100
223 62.7 50.2 75.4
227 63.7 51.1 76.7





#79
Naphthalene
Concen: 8.270 ppbv
RT: 13.510 min Scan# 4146
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825




Tgt Ion:128 Resp: 273029
Ion Ratio Lower Upper
128 100
127 12.7 9.9 14.9
129 11.0 8.4 12.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025




#80
Naphthalene,2-methyl-
Concen: 8.631 ppbv
RT: 14.455 min Scan# 4439
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Tgt Ion:142 Resp: 74802
Ion Ratio Lower Upper
142 100
141 86.8 67.1 100.7
115 41.7 31.9 47.9





#81
 1,2,4-Trichlorobenzene
 Concen: 8.497 ppbv
 RT: 13.436 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56



Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825



Tgt Ion:	180	Resp:	125476
Ion Ratio	Lower	Upper	
180	100		
182	96.1	77.4	116.2
184	30.9	24.2	36.4

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043211.D
Acq On : 18 Nov 2025 13:56
Operator : SY/MD
Sample : VSTDICV010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Quant Time: Nov 19 01:34:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	1.410	1.363	3.3	95	0.00
3	Chlorodifluoromethane	1.387	1.276	8.0	90	0.00
4	Chloromethane	0.527	0.479	9.1	90	0.00
5 T	Vinyl Chloride	0.577	0.490	15.1	94	0.00
6 T	Bromomethane	0.280	0.227	18.9	84	0.00
7	Chloroethane	0.225	0.203	9.8	97	0.00
8 T	Dichlorotetrafluoroethane	1.173	1.130	3.7	96	0.00
9 T	Propene	0.589	0.552	6.3	103	0.00
10 T	Heptane	1.753	1.525	13.0	88	0.00
11 T	Trichlorofluoromethane	1.367	1.377	-0.7	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.067	1.032	3.3	97	0.00
13	Ethanol	0.207	0.157	24.2	88	0.00
14 T	Bromoethene	0.409	0.378	7.6	97	0.00
15 T	Acetone	1.389	1.071	22.9	97	0.00
16 T	1,3-Butadiene	0.619	0.560	9.5	95	0.00
17	tert-Butyl alcohol	1.824	1.686	7.6	92	0.00
18 T	1,1-Dichloroethene	0.497	0.448	9.9	95	0.00
19 T	Isopropyl Alcohol	0.785	0.662	15.7	90	0.00
20 T	Methylene Chloride	0.600	0.422	29.7	99	0.00
21 T	Allyl Chloride	0.935	0.772	17.4	85	0.00
22 T	trans-1,2-Dichloroethene	0.568	0.513	9.7	96	0.00
23 T	Vinyl Acetate	1.677	1.432	14.6	82	0.00
24 T	1,1-Dichloroethane	1.096	1.054	3.8	96	0.00
25 T	Ethyl Acetate	2.996	2.643	11.8	89	0.00
26 T	Hexane	1.314	1.168	11.1	88	0.00
27 T	Carbon Disulfide	1.341	1.278	4.7	95	0.00
28 T	Methyl tert-Butyl Ether	0.645	0.656	-1.7	103	0.00
29 T	Chloroform	1.867	1.778	4.8	93	0.00
30 T	Cyclohexane	1.101	1.008	8.4	92	0.00
31 T	cis-1,2-Dichloroethene	1.306	1.225	6.2	92	0.00
32 T	1,1,1-Trichloroethane	1.990	1.863	6.4	92	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
34 T	2-Butanone	0.706	0.615	12.9	88	0.00
35 T	Carbon Tetrachloride	0.759	0.698	8.0	91	0.00
36 T	Benzene	0.958	0.890	7.1	91	0.00
37 T	1,2-Dichloroethane	0.560	0.547	2.3	95	0.00
38 T	Trichloroethene	0.375	0.347	7.5	91	0.00
39 T	1,2-Dichloropropane	0.351	0.318	9.4	88	0.00
40 T	1,4-Dioxane	0.158	0.123	22.2	79	0.00
41 T	Tetrahydrofuran	0.390	0.347	11.0	87	0.00
42 T	Bromodichloromethane	0.773	0.746	3.5	92	0.00
43	Methyl Methacrylate	0.508	0.444	12.6	89	0.00
44 T	2,2,4-Trimethylpentane	1.645	1.502	8.7	89	0.00
45 T	t-1,3-Dichloropropene	0.474	0.373	21.3	76	0.00
46 T	cis-1,3-Dichloropropene	0.556	0.441	20.7	76	0.00
47 T	1,1,2-Trichloroethane	0.372	0.338	9.1	91	0.01
48 T	Dibromochloromethane	0.640	0.608	5.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.566	0.515	9.0	86	0.00
50 T	4-Methyl-2-Pentanone	0.994	0.858	13.7	84	0.00
51 T	2-Hexanone	0.772	0.636	17.6	80	0.00
52 T	Tetrachloroethene	0.356	0.316	11.2	89	0.00
53 T	Toluene	1.134	1.043	8.0	90	0.00
54 T	1,2-Dibromoethane	0.547	0.477	12.8	84	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
56	1,1,1,2-Tetrachloroethane	0.735	0.659	10.3	90	0.00
57 T	Chlorobenzene	1.027	0.922	10.2	88	0.00
58 T	Ethyl Benzene	1.925	1.764	8.4	89	0.00
59 T	m/p-Xylene	1.543	1.397	9.5	90	0.00
60 T	o-Xylene	1.542	1.368	11.3	89	0.00
61 T	Styrene	0.675	0.588	12.9	84	0.00
62	Isopropylbenzene	2.758	2.431	11.9	88	0.00
63 T	1,1,2,2-Tetrachloroethane	0.981	0.826	15.8	85	0.00
64	n-propylbenzene	0.709	0.609	14.1	85	0.00
65	tert-Butylbenzene	2.463	2.085	15.3	86	0.00
66 T	Benzyl Chloride	0.250	0.129	48.4#	49	0.00
67	sec-Butylbenzene	3.438	2.884	16.1	85	0.00
68 S	1-Bromo-4-Fluorobenzene	0.733	0.737	-0.5	92	0.00
69	p-Isopropyltoluene	2.912	2.409	17.3	83	0.00
70	n-Butylbenzene	2.825	2.373	16.0	83	0.00
71	2-Chlorotoluene	2.120	1.834	13.5	86	0.00
72 T	4-Ethyltoluene	1.877	1.648	12.2	87	0.00
73 T	1,3,5-Trimethylbenzene	1.575	1.364	13.4	86	0.00
74 T	1,2,4-Trimethylbenzene	1.719	1.455	15.4	87	0.00
75 T	1,3-Dichlorobenzene	1.003	0.835	16.7	82	0.00
76 T	1,4-Dichlorobenzene	1.003	0.830	17.2	81	0.00
77 T	1,2-Dichlorobenzene	0.965	0.768	20.4	81	0.00
78 T	Hexachloro-1,3-Butadiene	0.805	0.581	27.8	78	0.00
79 T	Naphthalene	1.527	1.263	17.3	78	0.00
80 T	Naphthalene,2-methyl-	0.401	0.346	13.7	80	0.00
81 T	1,2,4-Trichlorobenzene	0.683	0.580	15.1	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043211.D
Acq On : 18 Nov 2025 13:56
Operator : SY/MD
Sample : VSTDICV010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Quant Time: Nov 19 01:34:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	10.000	9.664	3.4	95	0.00
3	Chlorodifluoromethane	10.000	9.199	8.0	90	0.00
4	Chloromethane	10.000	9.090	9.1	90	0.00
5 T	Vinyl Chloride	10.000	8.493	15.1	94	0.00
6 T	Bromomethane	10.000	8.126	18.7	84	0.00
7	Chloroethane	10.000	9.033	9.7	97	0.00
8 T	Dichlorotetrafluoroethane	10.000	9.632	3.7	96	0.00
9 T	Propene	10.000	9.370	6.3	103	0.00
10 T	Heptane	10.000	8.698	13.0	88	0.00
11 T	Trichlorofluoromethane	10.000	10.077	-0.8	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	9.673	3.3	97	0.00
13	Ethanol	10.000	7.584	24.2	88	0.00
14 T	Bromoethene	10.000	9.240	7.6	97	0.00
15 T	Acetone	10.000	7.711	22.9	97	0.00
16 T	1,3-Butadiene	10.000	9.041	9.6	95	0.00
17	tert-Butyl alcohol	10.000	9.245	7.6	92	0.00
18 T	1,1-Dichloroethene	10.000	9.017	9.8	95	0.00
19 T	Isopropyl Alcohol	10.000	8.431	15.7	90	0.00
20 T	Methylene Chloride	10.000	10.323	-3.2	99	0.00
21 T	Allyl Chloride	10.000	8.256	17.4	85	0.00
22 T	trans-1,2-Dichloroethene	10.000	9.041	9.6	96	0.00
23 T	Vinyl Acetate	10.000	8.539	14.6	82	0.00
24 T	1,1-Dichloroethane	10.000	9.622	3.8	96	0.00
25 T	Ethyl Acetate	10.000	8.820	11.8	89	0.00
26 T	Hexane	10.000	8.892	11.1	88	0.00
27 T	Carbon Disulfide	10.000	9.528	4.7	95	0.00
28 T	Methyl tert-Butyl Ether	10.000	10.171	-1.7	103	0.00
29 T	Chloroform	10.000	9.521	4.8	93	0.00
30 T	Cyclohexane	10.000	9.148	8.5	92	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.379	6.2	92	0.00
32 T	1,1,1-Trichloroethane	10.000	9.358	6.4	92	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	92	0.00
34 T	2-Butanone	10.000	8.707	12.9	88	0.00
35 T	Carbon Tetrachloride	10.000	9.200	8.0	91	0.00
36 T	Benzene	10.000	9.293	7.1	91	0.00
37 T	1,2-Dichloroethane	10.000	9.772	2.3	95	0.00
38 T	Trichloroethene	10.000	9.272	7.3	91	0.00
39 T	1,2-Dichloropropane	10.000	9.086	9.1	88	0.00
40 T	1,4-Dioxane	10.000	7.780	22.2	79	0.00
41 T	Tetrahydrofuran	10.000	8.891	11.1	87	0.00
42 T	Bromodichloromethane	10.000	9.657	3.4	92	0.00
43	Methyl Methacrylate	10.000	8.737	12.6	89	0.00
44 T	2,2,4-Trimethylpentane	10.000	9.133	8.7	89	0.00
45 T	t-1,3-Dichloropropene	10.000	7.863	21.4	76	0.00
46 T	cis-1,3-Dichloropropene	10.000	7.924	20.8	76	0.00
47 T	1,1,2-Trichloroethane	10.000	9.086	9.1	91	0.01
48 T	Dibromochloromethane	10.000	9.514	4.9	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	Bromoform	10.000	9.085	9.1	86	0.00
50 T	4-Methyl-2-Pentanone	10.000	8.631	13.7	84	0.00
51 T	2-Hexanone	10.000	8.233	17.7	80	0.00
52 T	Tetrachloroethene	10.000	8.884	11.2	89	0.00
53 T	Toluene	10.000	9.197	8.0	90	0.00
54 T	1,2-Dibromoethane	10.000	8.705	13.0	84	0.00
55 I	Chlorobenzene-d5	10.000	10.000	0.0	92	0.00
56	1,1,1,2-Tetrachloroethane	10.000	8.969	10.3	90	0.00
57 T	Chlorobenzene	10.000	8.981	10.2	88	0.00
58 T	Ethyl Benzene	10.000	9.167	8.3	89	0.00
59 T	m/p-Xylene	20.000	18.108	9.5	90	0.00
60 T	o-Xylene	10.000	8.876	11.2	89	0.00
61 T	Styrene	10.000	8.714	12.9	84	0.00
62	Isopropylbenzene	10.000	8.816	11.8	88	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	8.421	15.8	85	0.00
64	n-propylbenzene	10.000	8.591	14.1	85	0.00
65	tert-Butylbenzene	10.000	8.465	15.4	86	0.00
66 T	Benzyl Chloride	10.000	5.159	48.4#	49	0.00
67	sec-Butylbenzene	10.000	8.388	16.1	85	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.048	-0.5	92	0.00
69	p-Isopropyltoluene	10.000	8.272	17.3	83	0.00
70	n-Butylbenzene	10.000	8.398	16.0	83	0.00
71	2-Chlorotoluene	10.000	8.649	13.5	86	0.00
72 T	4-Ethyltoluene	10.000	8.779	12.2	87	0.00
73 T	1,3,5-Trimethylbenzene	10.000	8.660	13.4	86	0.00
74 T	1,2,4-Trimethylbenzene	10.000	8.463	15.4	87	0.00
75 T	1,3-Dichlorobenzene	10.000	8.320	16.8	82	0.00
76 T	1,4-Dichlorobenzene	10.000	8.273	17.3	81	0.00
77 T	1,2-Dichlorobenzene	10.000	7.966	20.3	81	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	7.222	27.8	78	0.00
79 T	Naphthalene	10.000	8.270	17.3	78	0.00
80 T	Naphthalene,2-methyl-	10.000	8.631	13.7	80	0.00
81 T	1,2,4-Trichlorobenzene	10.000	8.497	15.0	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>RTPE01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3799</u>
Instrument ID:	<u>MSVOA_L</u>	Calibration Date/Time:	<u>12/08/2025 08:44</u>
Lab File ID:	<u>VL043305.D</u>	Init. Calib. Date(s):	<u>11/18/2025 11/18/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:29 13:24</u>
GC Column:	<u>RTX-1</u>	ID:	<u>0.32</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.410	1.501		6.45	30
Chloromethane	0.527	0.540		2.47	30
Vinyl Chloride	0.577	0.526		-8.84	30
Bromomethane	0.280	0.255		-8.93	30
Chloroethane	0.225	0.217		-3.56	30
Tetrahydrofuran	0.390	0.375		-3.85	30
Trichlorofluoromethane	1.367	1.556		13.83	30
1,1,2-Trichlorotrifluoroethane	1.067	1.193		11.81	30
Dichlorotetrafluoroethane	1.173	1.224		4.35	30
tert-Butyl alcohol	1.824	1.625		-10.91	30
Heptane	1.753	1.655		-5.59	30
1,1-Dichloroethene	0.497	0.521		4.83	30
Acetone	1.389	1.192		-14.18	30
Carbon Disulfide	1.341	1.459		8.8	30
Methyl tert-Butyl Ether	0.645	0.654		1.39	30
Methylene Chloride	0.600	0.470		-21.67	30
trans-1,2-Dichloroethene	0.568	0.584		2.82	30
1,1-Dichloroethane	1.096	1.133		3.38	30
Cyclohexane	1.101	1.072		-2.63	30
2-Butanone	0.706	0.734		3.97	30
Carbon Tetrachloride	0.759	0.922		21.48	30
cis-1,2-Dichloroethene	1.306	1.287		-1.46	30
Chloroform	1.867	2.168		16.12	30
1,1,1-Trichloroethane	1.990	2.167		8.89	30
2,2,4-Trimethylpentane	1.645	1.794		9.06	30
Benzene	0.958	1.061		10.75	30
1,2-Dichloroethane	0.560	0.657		17.32	30
Trichloroethene	0.375	0.424		13.07	30
1,2-Dichloropropane	0.351	0.379		7.98	30
Bromodichloromethane	0.773	0.938		21.34	30
4-Methyl-2-Pentanone	0.994	0.969		-2.52	30
Toluene	1.134	1.153		1.67	30
t-1,3-Dichloropropene	0.474	0.366		-22.78	30
cis-1,3-Dichloropropene	0.556	0.457		-17.81	30
1,1,2-Trichloroethane	0.372	0.423		13.71	30
Dibromochloromethane	0.640	0.752		17.5	30
1,2-Dibromoethane	0.547	0.579		5.85	30
Tetrachloroethene	0.356	0.385		8.15	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>RTPE01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3799</u>
Instrument ID:	<u>MSVOA_L</u>	Calibration Date/Time:	<u>12/08/2025 08:44</u>
Lab File ID:	<u>VL043305.D</u>	Init. Calib. Date(s):	<u>11/18/2025 11/18/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:29 13:24</u>
GC Column:	<u>RTX-1</u>	ID:	<u>0.32</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chlorobenzene	1.027	1.226		19.38	30
Ethyl Benzene	1.925	2.145		11.43	30
m/p-Xylene	1.543	1.823		18.15	30
o-Xylene	1.542	1.819		17.96	30
Styrene	0.675	0.679		0.59	30
Bromoform	0.566	0.681		20.32	30
1,1,2,2-Tetrachloroethane	0.981	1.179		20.18	30
2-Chlorotoluene	2.120	2.393		12.88	30
1,3,5-Trimethylbenzene	1.575	1.812		15.05	30
1,2,4-Trimethylbenzene	1.719	2.029		18.03	30
1,3-Dichlorobenzene	1.003	1.213		20.94	30
1,4-Dichlorobenzene	1.003	1.212		20.84	30
1,2-Dichlorobenzene	0.965	1.164		20.62	30
1,2,4-Trichlorobenzene	0.683	0.821		20.2	30
Hexachloro-1,3-Butadiene	0.805	0.897		11.43	30
1,3-Butadiene	0.619	0.595		-3.88	30
Naphthalene	1.527	1.639		7.34	30
4-Ethyltoluene	1.877	2.192		16.78	30
1-Bromo-4-Fluorobenzene	0.733	0.753		2.73	30
Hexane	1.314	1.300		-1.07	30
Allyl Chloride	0.935	0.818		-12.51	30
1,4-Dioxane	157.566	139.218		-11.65	30
Methyl Methacrylate	0.508	0.476		-6.3	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043305.D
 Acq On : 08 Dec 2025 08:44
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Dec 09 00:07:49 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 12/09/2025
 Supervised By : Mahesh Dadoda 12/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	129333	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	319801	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	246102	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	185245	10.269	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.700%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.496	85	194161	10.644	ppbv	98
3) Chlorodifluoromethane	1.473	51	181461	10.115	ppbv	95
4) Chloromethane	1.531	50	69866	10.244	ppbv	96
5) Vinyl Chloride	1.580	62	68051	9.124	ppbv	98
6) Bromomethane	1.667	94	33036	9.137	ppbv	97
7) Chloroethane	1.703	64	28027	9.648	ppbv	93
8) Dichlorotetrafluoroethane	1.554	85	158307	10.433	ppbv	97
9) Propene	1.479	41	71565	9.394	ppbv #	1
10) Heptane	4.969	43	214107	9.444	ppbv	99
11) Trichlorofluoromethane	1.874	101	201211	11.385	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.136	101	154240	11.178	ppbv	97
13) Ethanol	1.719	45	22019	8.220	ppbv	97
14) Bromoethene	1.780	108	55354	10.465	ppbv	96
15) Acetone	1.832	43	154196	8.581	ppbv	99
16) 1,3-Butadiene	1.606	39	76902	9.600	ppbv	97
17) tert-Butyl alcohol	2.036	59	210170	8.911	ppbv	98
18) 1,1-Dichloroethene	2.030	96	67377	10.487	ppbv	94
19) Isopropyl Alcohol	1.884	45	84553	8.327	ppbv #	86
20) Methylene Chloride	2.059	84	60771	11.604	ppbv	96
21) Allyl Chloride	2.091	41	105856	8.756	ppbv	98
22) trans-1,2-Dichloroethene	2.337	96	75565	10.291	ppbv	94
23) Vinyl Acetate	2.457	43	204427	9.424	ppbv #	96
24) 1,1-Dichloroethane	2.405	63	146480	10.334	ppbv	98
25) Ethyl Acetate	2.829	43	373216	9.630	ppbv	99
26) Hexane	2.819	57	168150	9.897	ppbv	98
27) Carbon Disulfide	2.146	76	188702	10.878	ppbv	99
28) Methyl tert-Butyl Ether	2.434	73	84588	10.142	ppbv	96
29) Chloroform	2.842	83	280385	11.610	ppbv	97
30) Cyclohexane	3.845	84	138616	9.731	ppbv	94
31) cis-1,2-Dichloroethene	2.716	61	166393	9.853	ppbv	96
32) 1,1,1-Trichloroethane	3.360	97	280239	10.886	ppbv	99
34) 2-Butanone	2.551	43	234600	10.392	ppbv	100
35) Carbon Tetrachloride	3.752	117	294900	12.149	ppbv	98
36) Benzene	3.648	78	339149	11.074	ppbv	99
37) 1,2-Dichloroethane	3.211	62	209977	11.721	ppbv	98
38) Trichloroethene	4.535	130	135509	11.309	ppbv	98
39) 1,2-Dichloropropane	4.302	63	121360	10.827	ppbv	87
40) 1,4-Dioxane	4.564	88	44522	8.836	ppbv #	98
41) Tetrahydrofuran	3.046	42	120028	9.628	ppbv	98
42) Bromodichloromethane	4.477	83	299904	12.136	ppbv	99
43) Methyl Methacrylate	4.865	69	152174	9.361	ppbv	98
44) 2,2,4-Trimethylpentane	4.629	57	573790	10.909	ppbv	99
45) t-1,3-Dichloropropene	6.402	75	117128	7.729	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043305.D
 Acq On : 08 Dec 2025 08:44
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Dec 09 00:07:49 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025


Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 12/09/2025
 Supervised By : Mahesh Dadoda 12/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.580	75	146255m	8.219	ppbv	
47) 1,1,2-Trichloroethane	6.564	97	135385	11.381	ppbv	95
48) Dibromochloromethane	7.370	129	240386	11.753	ppbv	99
49) Bromoform	9.471	173	217719	12.018	ppbv	98
50) 4-Methyl-2-Pentanone	5.729	43	310023	9.749	ppbv	97
51) 2-Hexanone	7.419	43	219906	8.905	ppbv #	99
52) Tetrachloroethene	8.212	164	123188	10.813	ppbv	98
53) Toluene	6.920	91	368638	10.165	ppbv	100
54) 1,2-Dibromoethane	7.636	107	185272	10.583	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.914	131	231860	12.819	ppbv	97
57) Chlorobenzene	8.911	112	301763	11.943	ppbv	95
58) Ethyl Benzene	9.344	91	527984	11.146	ppbv	95
59) m/p-Xylene	9.522	91	897418m	23.631	ppbv	
60) o-Xylene	9.953	91	447628	11.799	ppbv	99
61) Styrene	9.859	104	167047	10.061	ppbv	99
62) Isopropylbenzene	10.529	105	797107	11.744	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.950	83	290145	12.020	ppbv	99
64) n-propylbenzene	10.979	120	205431	11.779	ppbv	99
65) tert-Butylbenzene	11.523	119	694372	11.455	ppbv	100
66) Benzyl Chloride	11.607	91	30971	5.035	ppbv	98
67) sec-Butylbenzene	11.759	105	982704	11.615	ppbv	99
69) p-Isopropyltoluene	11.918	119	812059	11.330	ppbv	99
70) n-Butylbenzene	12.274	91	780968	11.231	ppbv	99
71) 2-Chlorotoluene	10.892	91	588943	11.289	ppbv	99
72) 4-Ethyltoluene	11.118	105	539530	11.678	ppbv	99
73) 1,3,5-Trimethylbenzene	11.196	105	445981	11.503	ppbv	97
74) 1,2,4-Trimethylbenzene	11.529	105	499268	11.800	ppbv	91
75) 1,3-Dichlorobenzene	11.597	146	298569	12.092	ppbv	98
76) 1,4-Dichlorobenzene	11.662	146	298168	12.074	ppbv	99
77) 1,2-Dichlorobenzene	11.937	146	286499	12.068	ppbv	99
78) Hexachloro-1,3-Butadiene	13.876	225	220765	11.143	ppbv	99
79) Naphthalene	13.507	128	403341	10.733	ppbv	99
80) Naphthalene,2-methyl-	14.452	142	91763	9.302	ppbv	97
81) 1,2,4-Trichlorobenzene	13.432	180	202078	12.024	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 832
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 49 Resp: 12933
Ion Ratio Lower Upper
49 100
128 47.5 23.0 69.0
130 61.0 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#2
Dichlorodifluoromethane
Concen: 10.644 ppbv
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 85 Resp: 194161
Ion Ratio Lower Upper
85 100
87 30.9 25.4 38.2









#3
Chlorodifluoromethane
Concen: 10.115 ppbv
RT: 1.473 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 51 Resp: 18146
Ion Ratio Lower Upper
51 100
67 20.9 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#4
Chloromethane
Concen: 10.244 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44


Tgt Ion: 50 Resp: 69866
Ion Ratio Lower Upper
50 100
52 32.0 24.1 36.1





#5
 Vinyl Chloride
 Concen: 9.124 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010




Tgt Ion: 62 Resp: 6805
 Ion Ratio Lower Upper
 62 100
 64 32.9 27.3 40.9

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025




#6
 Bromomethane
 Concen: 9.137 ppbv
 RT: 1.667 min Scan# 488
 Delta R.T. 0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44



Tgt Ion: 94 Resp: 33036
 Ion Ratio Lower Upper
 94 100
 96 94.1 72.9 109.3









#7
Chloroethane
Concen: 9.648 ppbv
RT: 1.703 min Scan# 499
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 64 Resp: 2802
Ion Ratio Lower Upper
64 100
66 29.5 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#8
Dichlorotetrafluoroethane
Concen: 10.433 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44


Tgt Ion: 85 Resp: 158307
Ion Ratio Lower Upper
85 100
135 67.8 52.6 79.0





#9
 Propene
 Concen: 9.394 ppbv
 RT: 1.479 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010




Tgt Ion: 41 Resp: 7156
 Ion Ratio Lower Upper
 41 100
 42 68.4 3.7 5.5

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025




#10
 Heptane
 Concen: 9.444 ppbv
 RT: 4.969 min Scan# 1508
 Delta R.T. 0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44



Tgt Ion: 43 Resp: 214107
 Ion Ratio Lower Upper
 43 100
 57 50.0 39.7 59.5









#11
Trichlorofluoromethane
Concen: 11.385 ppbv
RT: 1.874 min Scan# 511
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion:101 Resp: 20121
Ion Ratio Lower Upper
101 100
103 63.0 52.6 79.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#12
1,1,2-Trichlorotrifluoroethane
Concen: 11.178 ppbv
RT: 2.136 min Scan# 633
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:101 Resp: 154240
Ion Ratio Lower Upper
101 100
151 76.6 59.2 88.8









#13
Ethanol
Concen: 8.220 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 45 Resp: 22019
Ion Ratio Lower Upper
45 100
46 39.2 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#14
Bromoethene
Concen: 10.465 ppbv
RT: 1.780 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:108 Resp: 55354
Ion Ratio Lower Upper
108 100
106 106.7 88.9 133.3
79 19.4 16.5 24.7









#15
Acetone
Concen: 8.581 ppbv
RT: 1.832 min Scan# 539
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 43 Resp: 154190
Ion Ratio Lower Upper
43 100
58 26.2 20.7 31.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#16
1,3-Butadiene
Concen: 9.600 ppbv
RT: 1.606 min Scan# 469
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 39 Resp: 76902
Ion Ratio Lower Upper
39 100
54 71.7 59.4 89.2









#17
tert-Butyl alcohol
Concen: 8.911 ppbv
RT: 2.036 min Scan# 602
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 59 Resp: 210170
Ion Ratio Lower Upper
59 100
57 10.5 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#18
1,1-Dichloroethene
Concen: 10.487 ppbv
RT: 2.030 min Scan# 600
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 96 Resp: 67377
Ion Ratio Lower Upper
96 100
61 202.7 170.6 255.8
98 62.0 52.2 78.4









#19
Isopropyl Alcohol
Concen: 8.327 ppbv
RT: 1.884 min Scan# 51
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 45 Resp: 8455
Ion Ratio Lower Upper
45 100
58 47.9 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#20
Methylene Chloride
Concen: 11.604 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 84 Resp: 60771
Ion Ratio Lower Upper
84 100
49 160.8 132.9 199.3
51 47.0 34.8 52.2
86 64.2 48.2 72.2









#21
 Allyl Chloride
 Concen: 8.756 ppbv
 RT: 2.091 min Scan# 619
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Tgt Ion: 41 Resp: 105850
 Ion Ratio Lower Upper
 41 100
 76 27.6 22.1 33.1
 39 78.1 61.0 91.4




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#22
 trans-1,2-Dichloroethene
 Concen: 10.291 ppbv
 RT: 2.337 min Scan# 695
 Delta R.T. 0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion: 96 Resp: 75565
 Ion Ratio Lower Upper
 96 100
 61 163.0 139.0 208.6
 98 61.8 49.4 74.2





#23
 Vinyl Acetate
 Concen: 9.424 ppbv
 RT: 2.457 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44





Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Tgt Ion: 43 Resp: 20442

Ion	Ratio	Lower	Upper
43	100		
86	9.5	6.2	9.2
42	8.6	8.0	12.0
44	6.4	4.0	6.0

Manual Integrations
 APPROVED




Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025




#24
 1,1-Dichloroethane
 Concen: 10.334 ppbv
 RT: 2.405 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion: 63 Resp: 146480


Ion	Ratio	Lower	Upper
63	100		
98	5.0	2.8	8.4
100	3.2	1.3	3.9





#25
Ethyl Acetate
Concen: 9.630 ppbv
RT: 2.829 min Scan# 847
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion: 43 Resp: 373210
Ion Ratio Lower Upper
43 100
45 9.5 8.1 12.1
61 8.8 7.0 10.4
70 8.2 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#26
Hexane
Concen: 9.897 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Tgt Ion: 57 Resp: 168150
Ion Ratio Lower Upper
57 100
41 95.7 74.6 111.8
43 222.8 179.0 268.4
56 56.4 42.3 63.5





#27
Carbon Disulfide
Concen: 10.878 ppbv
RT: 2.146 min Scan# 611
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion: 76 Resp: 188702
Ion Ratio Lower Upper
76 100
44 20.4 16.1 24.1
78 11.8 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#28
Methyl tert-Butyl Ether
Concen: 10.142 ppbv
RT: 2.434 min Scan# 725
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44



Tgt Ion: 73 Resp: 84588
Ion Ratio Lower Upper
73 100
57 22.7 19.8 29.8









#29
Chloroform
Concen: 11.610 ppbv
RT: 2.842 min Scan# 851
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 83 Resp: 280385
Ion Ratio Lower Upper
83 100
85 60.9 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#30
Cyclohexane
Concen: 9.731 ppbv
RT: 3.845 min Scan# 1161
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44


Tgt Ion: 84 Resp: 138616
Ion Ratio Lower Upper
84 100
56 121.7 104.7 157.1
41 84.0 69.0 103.6





#31
 cis-1,2-Dichloroethene
 Concen: 9.853 ppbv
 RT: 2.716 min Scan# 812
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010




Tgt Ion: 61 Resp: 16639
 Ion Ratio Lower Upper
 61 100
 96 64.4 49.3 73.9
 98 42.2 32.1 48.1

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025




#32
 1,1,1-Trichloroethane
 Concen: 10.886 ppbv
 RT: 3.360 min Scan# 1011
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44



Tgt Ion: 97 Resp: 280239
 Ion Ratio Lower Upper
 97 100
 99 64.0 50.3 75.5
 61 48.7 38.9 58.3









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.949 min Scan# 1193
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion:114 Resp: 31980
Ion Ratio Lower Upper
114 100
63 25.6 21.0 31.6
88 18.0 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#34
2-Butanone
Concen: 10.392 ppbv
RT: 2.551 min Scan# 761
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 43 Resp: 234600
Ion Ratio Lower Upper
43 100
72 22.7 18.1 27.1









#35
Carbon Tetrachloride
Concen: 12.149 ppbv
RT: 3.752 min Scan# 1133
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion:117 Resp: 294900
Ion Ratio Lower Upper
117 100
119 95.5 74.6 112.0
121 30.1 23.2 34.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#36
Benzene
Concen: 11.074 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 78 Resp: 339149
Ion Ratio Lower Upper
78 100
77 22.7 17.5 26.3
50 19.0 15.5 23.3









#37
1,2-Dichloroethane
Concen: 11.721 ppbv
RT: 3.211 min Scan# 90
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 62 Resp: 20997
Ion Ratio Lower Upper
62 100
98 6.3 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#38
Trichloroethene
Concen: 11.309 ppbv
RT: 4.535 min Scan# 1374
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:130 Resp: 135509
Ion Ratio Lower Upper
130 100
95 106.7 85.7 128.5
132 90.9 73.4 110.0
97 70.3 52.7 79.1





#39

1,2-Dichloropropane

Concen: 10.827 ppbv

RT: 4.302 min Scan# 11

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Instrument :

MSVOA_L

ClientSampleId :

VSTDCCC010

Tgt Ion: 63 Resp: 121360

Ion Ratio Lower Upper

63 100

41 91.0 62.6 93.8





62 72.5 51.0 76.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025

Supervised By :Mahesh Dadoda 12/11/2025



#40

1,4-Dioxane

Concen: 8.836 ppbv

RT: 4.564 min Scan# 1383

Delta R.T. -0.003 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Tgt Ion: 88 Resp: 44522




Ion Ratio Lower Upper


88 100

58 80.4 63.0 94.6

43 38.4 32.5 48.7

57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 9.628 ppbv
RT: 3.046 min Scan# 914
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 42 Resp: 120028
Ion Ratio Lower Upper
42 100
72 38.4 30.6 46.0
71 36.1 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#42
Bromodichloromethane
Concen: 12.136 ppbv
RT: 4.477 min Scan# 1356
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44


Tgt Ion: 83 Resp: 299904
Ion Ratio Lower Upper
83 100
85 63.9 51.7 77.5
127 8.0 6.3 9.5





#43
Methyl Methacrylate
Concen: 9.361 ppbv
RT: 4.865 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion: 69 Resp: 152174
Ion Ratio Lower Upper
69 100
41 155.9 126.7 190.1
100 35.6 28.6 43.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#44
2,2,4-Trimethylpentane
Concen: 10.909 ppbv
RT: 4.629 min Scan# 1403
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Tgt Ion: 57 Resp: 573790
Ion Ratio Lower Upper
57 100
41 36.1 28.0 42.0
56 32.4 25.6 38.4





#45
 t-1,3-Dichloropropene
 Concen: 7.729 ppbv
 RT: 6.402 min Scan# 1951
 Delta R.T. 0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010




Tgt Ion: 75 Resp: 117128
 Ion Ratio Lower Upper
 75 100
 77 30.1 24.3 36.5

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025




#46
 cis-1,3-Dichloropropene
 Concen: 8.219 ppbv m
 RT: 5.580 min Scan# 1697
 Delta R.T. -0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Tgt Ion: 75 Resp: 146255
 Ion Ratio Lower Upper
 75 100
 39 67.3 59.1 88.7
 77 29.8 24.2 36.2





#47
1,1,2-Trichloroethane
Concen: 11.381 ppbv
RT: 6.564 min Scan# 2001
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion: 97 Resp: 13538
Ion Ratio Lower Upper
97 100
83 84.4 64.6 97.0
85 52.5 39.4 59.2
99 62.6 47.0 70.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#48
Dibromochloromethane
Concen: 11.753 ppbv
RT: 7.370 min Scan# 2250
Delta R.T. -0.007 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Tgt Ion:129 Resp: 240386
Ion Ratio Lower Upper
129 100
127 77.6 39.3 117.9





#49
Bromoform
Concen: 12.018 ppbv
RT: 9.471 min Scan# 217719
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion: 173 Resp: 217719
Ion Ratio Lower Upper
173 100
175 49.1 38.6 58.0
171 52.5 40.6 61.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#50
4-Methyl-2-Pentanone
Concen: 9.749 ppbv
RT: 5.729 min Scan# 1743
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Tgt Ion: 43 Resp: 310023
Ion Ratio Lower Upper
43 100
58 34.0 28.4 42.6





#51
2-Hexanone
Concen: 8.905 ppbv
RT: 7.419 min Scan# 219900
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010



Tgt Ion: 43 Resp: 219900
Ion Ratio Lower Upper
43 100
58 47.6 38.4 57.6
98 1.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#52
Tetrachloroethene
Concen: 10.813 ppbv
RT: 8.212 min Scan# 2510
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44


Tgt Ion:164 Resp: 123188
Ion Ratio Lower Upper
164 100
166 124.8 99.4 149.2
129 96.7 81.8 122.8
131 97.9 77.4 116.2





#53
Toluene
Concen: 10.165 ppbv
RT: 6.920 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion: 91 Resp: 368638
Ion Ratio Lower Upper
91 100
92 59.4 47.4 71.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#54
1,2-Dibromoethane
Concen: 10.583 ppbv
RT: 7.636 min Scan# 2332
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44



Tgt Ion:107 Resp: 185272
Ion Ratio Lower Upper
107 100
109 88.4 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 2714
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion:117 Resp: 24610
Ion Ratio Lower Upper
117 100
82 71.1 56.1 84.1
119 30.4 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#56
1,1,1,2-Tetrachloroethane
Concen: 12.819 ppbv
RT: 8.914 min Scan# 2727
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:131 Resp: 231860
Ion Ratio Lower Upper
131 100
133 94.4 77.4 116.2
119 58.4 49.1 73.7









#57
Chlorobenzene
Concen: 11.943 ppbv
RT: 8.911 min Scan# 21
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion:112 Resp: 30176
Ion Ratio Lower Upper
112 100
77 65.2 55.5 83.3
114 30.6 29.1 35.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#58
Ethyl Benzene
Concen: 11.146 ppbv
RT: 9.344 min Scan# 2860
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 91 Resp: 527984
Ion Ratio Lower Upper
91 100
106 31.2 22.8 34.2









#59
m/p-Xylene
Concen: 23.631 ppbv m
RT: 9.522 min Scan# 2915
Delta R.T. -0.020 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 91 Resp: 897418
Ion Ratio Lower Upper
91 100
106 45.6 36.1 54.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#60
o-Xylene
Concen: 11.799 ppbv
RT: 9.953 min Scan# 3048
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44


Tgt Ion: 91 Resp: 447628
Ion Ratio Lower Upper
91 100
106 42.5 20.9 62.8





#61
Styrene
Concen: 10.061 ppbv
RT: 9.859 min Scan# 3019
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion:104 Resp: 16704
Ion Ratio Lower Upper
104 100
78 51.8 42.1 63.1
103 45.3 37.2 55.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#62
Isopropylbenzene
Concen: 11.744 ppbv
RT: 10.529 min Scan# 3226
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44



Tgt Ion:105 Resp: 797107
Ion Ratio Lower Upper
105 100
120 25.3 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 12.020 ppbv
RT: 9.950 min Scan# 3048
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument : MSVOA_L
ClientSampleId : VSTDCCC010

Tgt Ion: 83 Resp: 29014
Ion Ratio Lower Upper
83 100
131 9.9 4.8 14.3
85 63.4 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#64
n-propylbenzene
Concen: 11.779 ppbv
RT: 10.979 min Scan# 3365
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44


Tgt Ion:120 Resp: 205431
Ion Ratio Lower Upper
120 100
91 467.7 377.1 565.7





#65
tert-Butylbenzene
Concen: 11.455 ppbv
RT: 11.523 min Scan# 3533
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion: 119 Resp: 694372
Ion Ratio Lower Upper
119 100
91 81.7 65.3 97.9
134 22.0 17.0 25.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#66
Benzyl Chloride
Concen: 5.035 ppbv
RT: 11.607 min Scan# 3559
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Tgt Ion: 91 Resp: 30971
Ion Ratio Lower Upper
91 100
126 17.2 14.4 21.6
128 6.0 4.6 6.8





#67
 sec-Butylbenzene
 Concen: 11.615 ppbv
 RT: 11.759 min Scan# 3606
 Delta R.T. -0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010



Tgt Ion:105 Resp: 982704
 Ion Ratio Lower Upper
 105 100
 134 19.4 15.0 22.6




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.269 ppbv
 RT: 10.367 min Scan# 3176
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion: 95 Resp: 185245
 Ion Ratio Lower Upper
 95 100
 174 68.3 50.2 75.2
 175 5.0 4.1 6.1









#69
p-Isopropyltoluene
Concen: 11.330 ppbv
RT: 11.918 min Scan# 3655
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 119 Resp: 812059
Ion Ratio Lower Upper
119 100
134 25.9 20.2 30.2
91 29.0 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#70
n-Butylbenzene
Concen: 11.231 ppbv
RT: 12.274 min Scan# 3765
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44


Tgt Ion: 91 Resp: 780968
Ion Ratio Lower Upper
91 100
92 55.2 43.5 65.3
134 24.3 19.0 28.6





#71
2-Chlorotoluene
Concen: 11.289 ppbv
RT: 10.892 min Scan# 31
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion: 91 Resp: 58894
Ion Ratio Lower Upper
91 100
126 30.0 11.7 46.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#72
4-Ethyltoluene
Concen: 11.678 ppbv
RT: 11.118 min Scan# 3408
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Tgt Ion:105 Resp: 539530
Ion Ratio Lower Upper
105 100
120 28.8 23.8 35.8
91 12.8 10.3 15.5
77 13.6 11.1 16.7





#73
1,3,5-Trimethylbenzene
Concen: 11.503 ppbv
RT: 11.196 min Scan# 3432
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010




Tgt Ion:105 Resp: 44598
Ion Ratio Lower Upper
105 100
120 46.5 38.6 57.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#74
1,2,4-Trimethylbenzene
Concen: 11.800 ppbv
RT: 11.529 min Scan# 3535
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44



Tgt Ion:105 Resp: 499268
Ion Ratio Lower Upper
105 100
120 50.4 41.7 62.5
91 74.6 70.0 105.0
77 26.3 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 12.092 ppbv
RT: 11.597 min Scan# 3556
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion:146 Resp: 298569
Ion Ratio Lower Upper
146 100
111 44.8 23.4 70.3
148 64.8 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#76
1,4-Dichlorobenzene
Concen: 12.074 ppbv
RT: 11.662 min Scan# 3576
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:146 Resp: 298168
Ion Ratio Lower Upper
146 100
111 44.2 22.8 68.3
148 64.0 31.8 95.4









#77
1,2-Dichlorobenzene
Concen: 12.068 ppbv
RT: 11.937 min Scan# 3661
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion:146 Resp: 286499
Ion Ratio Lower Upper
146 100
111 47.3 24.3 73.0
148 63.7 31.6 94.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#78
Hexachloro-1,3-Butadiene
Concen: 11.143 ppbv
RT: 13.876 min Scan# 4260
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:225 Resp: 220765
Ion Ratio Lower Upper
225 100
223 63.4 50.2 75.4
227 64.1 51.1 76.7





#79

Naphthalene

Concen: 10.733 ppbv

RT: 13.507 min Scan# 4146

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Instrument :

MSVOA_L

ClientSampleId :

VSTDCCC010





Tgt Ion:128 Resp: 40334

Ion	Ratio	Lower	Upper
128	100		
127	12.1	9.9	14.9
129	11.2	8.4	12.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#80

Naphthalene,2-methyl-

Concen: 9.302 ppbv

RT: 14.452 min Scan# 4438




Delta R.T. -0.000 min


Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Tgt Ion:142 Resp: 91763


Ion	Ratio	Lower	Upper
142	100		
141	81.6	67.1	100.7
115	37.9	31.9	47.9





#81
 1,2,4-Trichlorobenzene
 Concen: 12.024 ppbv
 RT: 13.432 min Scan# 41
 Delta R.T. -0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44



Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010



Tgt Ion:180 Resp: 202078
 Ion Ratio Lower Upper
 180 100
 182 97.5 77.4 116.2
 184 30.7 24.2 36.4

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043305.D
Acq On : 08 Dec 2025 08:44
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Dec 09 00:07:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	123	0.00
2 T	Dichlorodifluoromethane	1.410	1.501	-6.5	138	0.00
3	Chlorodifluoromethane	1.387	1.403	-1.2	131	0.00
4	Chloromethane	0.527	0.540	-2.5	133	0.00
5 T	Vinyl Chloride	0.577	0.526	8.8	133	0.00
6 T	Bromomethane	0.280	0.255	8.9	124	0.00
7	Chloroethane	0.225	0.217	3.6	138	0.00
8 T	Dichlorotetrafluoroethane	1.173	1.224	-4.3	138	0.00
9 T	Propene	0.589	0.553	6.1	136	0.00
10 T	Heptane	1.753	1.655	5.6	127	0.00
11 T	Trichlorofluoromethane	1.367	1.556	-13.8	154#	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.067	1.193	-11.8	149	0.00
13	Ethanol	0.207	0.170	17.9	126	0.00
14 T	Bromoethene	0.409	0.428	-4.6	146	0.00
15 T	Acetone	1.389	1.192	14.2	142	0.00
16 T	1,3-Butadiene	0.619	0.595	3.9	133	0.00
17	tert-Butyl alcohol	1.824	1.625	10.9	117	0.00
18 T	1,1-Dichloroethene	0.497	0.521	-4.8	145	0.00
19 T	Isopropyl Alcohol	0.785	0.654	16.7	118	0.00
20 T	Methylene Chloride	0.600	0.470	21.7	145	0.00
21 T	Allyl Chloride	0.935	0.818	12.5	119	0.00
22 T	trans-1,2-Dichloroethene	0.568	0.584	-2.8	144	0.00
23 T	Vinyl Acetate	1.677	1.581	5.7	120	0.00
24 T	1,1-Dichloroethane	1.096	1.133	-3.4	137	0.00
25 T	Ethyl Acetate	2.996	2.886	3.7	129	0.00
26 T	Hexane	1.314	1.300	1.1	130	0.00
27 T	Carbon Disulfide	1.341	1.459	-8.8	143	0.00
28 T	Methyl tert-Butyl Ether	0.645	0.654	-1.4	136	0.00
29 T	Chloroform	1.867	2.168	-16.1	149	0.00
30 T	Cyclohexane	1.101	1.072	2.6	129	0.00
31 T	cis-1,2-Dichloroethene	1.306	1.287	1.5	128	0.00
32 T	1,1,1-Trichloroethane	1.990	2.167	-8.9	142	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	113	0.00
34 T	2-Butanone	0.706	0.734	-4.0	128	0.00
35 T	Carbon Tetrachloride	0.759	0.922	-21.5	147	0.00
36 T	Benzene	0.958	1.061	-10.8	132	0.00
37 T	1,2-Dichloroethane	0.560	0.657	-17.3	139	0.00
38 T	Trichloroethene	0.375	0.424	-13.1	135	0.00
39 T	1,2-Dichloropropane	0.351	0.379	-8.0	128	0.00
40 T	1,4-Dioxane	0.158	0.139	12.0	110	0.00
41 T	Tetrahydrofuran	0.390	0.375	3.8	115	0.00
42 T	Bromodichloromethane	0.773	0.938	-21.3	142	0.00
43	Methyl Methacrylate	0.508	0.476	6.3	116	0.00
44 T	2,2,4-Trimethylpentane	1.645	1.794	-9.1	130	0.00
45 T	t-1,3-Dichloropropene	0.474	0.366	22.8	91	0.00
46 T	cis-1,3-Dichloropropene	0.556	0.457	17.8	96	0.00
47 T	1,1,2-Trichloroethane	0.372	0.423	-13.7	139	0.00
48 T	Dibromochloromethane	0.640	0.752	-17.5	137	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043305.D
Acq On : 08 Dec 2025 08:44
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Dec 09 00:07:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.566	0.681	-20.3	138	0.00
50 T	4-Methyl-2-Pentanone	0.994	0.969	2.5	116	0.00
51 T	2-Hexanone	0.772	0.688	10.9	106	0.00
52 T	Tetrachloroethene	0.356	0.385	-8.1	132	0.00
53 T	Toluene	1.134	1.153	-1.7	121	0.00
54 T	1,2-Dibromoethane	0.547	0.579	-5.9	124	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
56	1,1,1,2-Tetrachloroethane	0.735	0.942	-28.2	146	0.00
57 T	Chlorobenzene	1.027	1.226	-19.4	133	0.00
58 T	Ethyl Benzene	1.925	2.145	-11.4	123	0.00
59 T	m/p-Xylene	1.543	1.823	-18.1	133	-0.02
60 T	o-Xylene	1.542	1.819	-18.0	134	0.00
61 T	Styrene	0.675	0.679	-0.6	110	0.00
62	Isopropylbenzene	2.758	3.239	-17.4	133	0.00
63 T	1,1,2,2-Tetrachloroethane	0.981	1.179	-20.2	139	0.00
64	n-propylbenzene	0.709	0.835	-17.8	133	0.00
65	tert-Butylbenzene	2.463	2.821	-14.5	133	0.00
66 T	Benzyl Chloride	0.250	0.126	49.6#	55	0.00
67	sec-Butylbenzene	3.438	3.993	-16.1	135	0.00
68 S	1-Bromo-4-Fluorobenzene	0.733	0.753	-2.7	107	0.00
69	p-Isopropyltoluene	2.912	3.300	-13.3	130	0.00
70	n-Butylbenzene	2.825	3.173	-12.3	126	0.00
71	2-Chlorotoluene	2.120	2.393	-12.9	128	0.00
72 T	4-Ethyltoluene	1.877	2.192	-16.8	131	0.00
73 T	1,3,5-Trimethylbenzene	1.575	1.812	-15.0	130	0.00
74 T	1,2,4-Trimethylbenzene	1.719	2.029	-18.0	137	0.00
75 T	1,3-Dichlorobenzene	1.003	1.213	-20.9	136	0.00
76 T	1,4-Dichlorobenzene	1.003	1.212	-20.8	135	0.00
77 T	1,2-Dichlorobenzene	0.965	1.164	-20.6	140	0.00
78 T	Hexachloro-1,3-Butadiene	0.805	0.897	-11.4	137	0.00
79 T	Naphthalene	1.527	1.639	-7.3	115	0.00
80 T	Naphthalene,2-methyl-	0.401	0.373	7.0	99	0.00
81 T	1,2,4-Trichlorobenzene	0.683	0.821	-20.2	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043305.D
Acq On : 08 Dec 2025 08:44
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Dec 09 00:07:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	123	0.00
2 T	Dichlorodifluoromethane	10.000	10.644	-6.4	138	0.00
3	Chlorodifluoromethane	10.000	10.115	-1.2	131	0.00
4	Chloromethane	10.000	10.244	-2.4	133	0.00
5 T	Vinyl Chloride	10.000	9.124	8.8	133	0.00
6 T	Bromomethane	10.000	9.137	8.6	124	0.00
7	Chloroethane	10.000	9.648	3.5	138	0.00
8 T	Dichlorotetrafluoroethane	10.000	10.433	-4.3	138	0.00
9 T	Propene	10.000	9.394	6.1	136	0.00
10 T	Heptane	10.000	9.444	5.6	127	0.00
11 T	Trichlorofluoromethane	10.000	11.385	-13.8	154	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	11.178	-11.8	149	0.00
13	Ethanol	10.000	8.220	17.8	126	0.00
14 T	Bromoethene	10.000	10.465	-4.6	146	0.00
15 T	Acetone	10.000	8.581	14.2	142	0.00
16 T	1,3-Butadiene	10.000	9.600	4.0	133	0.00
17	tert-Butyl alcohol	10.000	8.911	10.9	117	0.00
18 T	1,1-Dichloroethene	10.000	10.487	-4.9	145	0.00
19 T	Isopropyl Alcohol	10.000	8.327	16.7	118	0.00
20 T	Methylene Chloride	10.000	11.604	-16.0	145	0.00
21 T	Allyl Chloride	10.000	8.756	12.4	119	0.00
22 T	trans-1,2-Dichloroethene	10.000	10.291	-2.9	144	0.00
23 T	Vinyl Acetate	10.000	9.424	5.8	120	0.00
24 T	1,1-Dichloroethane	10.000	10.334	-3.3	137	0.00
25 T	Ethyl Acetate	10.000	9.630	3.7	129	0.00
26 T	Hexane	10.000	9.897	1.0	130	0.00
27 T	Carbon Disulfide	10.000	10.878	-8.8	143	0.00
28 T	Methyl tert-Butyl Ether	10.000	10.142	-1.4	136	0.00
29 T	Chloroform	10.000	11.610	-16.1	149	0.00
30 T	Cyclohexane	10.000	9.731	2.7	129	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.853	1.5	128	0.00
32 T	1,1,1-Trichloroethane	10.000	10.886	-8.9	142	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	113	0.00
34 T	2-Butanone	10.000	10.392	-3.9	128	0.00
35 T	Carbon Tetrachloride	10.000	12.149	-21.5	147	0.00
36 T	Benzene	10.000	11.074	-10.7	132	0.00
37 T	1,2-Dichloroethane	10.000	11.721	-17.2	139	0.00
38 T	Trichloroethene	10.000	11.309	-13.1	135	0.00
39 T	1,2-Dichloropropane	10.000	10.827	-8.3	128	0.00
40 T	1,4-Dioxane	10.000	8.836	11.6	110	0.00
41 T	Tetrahydrofuran	10.000	9.628	3.7	115	0.00
42 T	Bromodichloromethane	10.000	12.136	-21.4	142	0.00
43	Methyl Methacrylate	10.000	9.361	6.4	116	0.00
44 T	2,2,4-Trimethylpentane	10.000	10.909	-9.1	130	0.00
45 T	t-1,3-Dichloropropene	10.000	7.729	22.7	91	0.00
46 T	cis-1,3-Dichloropropene	10.000	8.219	17.8	96	0.00
47 T	1,1,2-Trichloroethane	10.000	11.381	-13.8	139	0.00
48 T	Dibromochloromethane	10.000	11.753	-17.5	137	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043305.D
Acq On : 08 Dec 2025 08:44
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Dec 09 00:07:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	Bromoform	10.000	12.018	-20.2	138	0.00
50 T	4-Methyl-2-Pentanone	10.000	9.749	2.5	116	0.00
51 T	2-Hexanone	10.000	8.905	11.0	106	0.00
52 T	Tetrachloroethene	10.000	10.813	-8.1	132	0.00
53 T	Toluene	10.000	10.165	-1.6	121	0.00
54 T	1,2-Dibromoethane	10.000	10.583	-5.8	124	0.00
55 I	Chlorobenzene-d5	10.000	10.000	0.0	105	0.00
56	1,1,1,2-Tetrachloroethane	10.000	12.819	-28.2	146	0.00
57 T	Chlorobenzene	10.000	11.943	-19.4	133	0.00
58 T	Ethyl Benzene	10.000	11.146	-11.5	123	0.00
59 T	m/p-Xylene	20.000	23.631	-18.2	133	-0.02
60 T	o-Xylene	10.000	11.799	-18.0	134	0.00
61 T	Styrene	10.000	10.061	-0.6	110	0.00
62	Isopropylbenzene	10.000	11.744	-17.4	133	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	12.020	-20.2	139	0.00
64	n-propylbenzene	10.000	11.779	-17.8	133	0.00
65	tert-Butylbenzene	10.000	11.455	-14.6	133	0.00
66 T	Benzyl Chloride	10.000	5.035	49.6#	55	0.00
67	sec-Butylbenzene	10.000	11.615	-16.2	135	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.269	-2.7	107	0.00
69	p-Isopropyltoluene	10.000	11.330	-13.3	130	0.00
70	n-Butylbenzene	10.000	11.231	-12.3	126	0.00
71	2-Chlorotoluene	10.000	11.289	-12.9	128	0.00
72 T	4-Ethyltoluene	10.000	11.678	-16.8	131	0.00
73 T	1,3,5-Trimethylbenzene	10.000	11.503	-15.0	130	0.00
74 T	1,2,4-Trimethylbenzene	10.000	11.800	-18.0	137	0.00
75 T	1,3-Dichlorobenzene	10.000	12.092	-20.9	136	0.00
76 T	1,4-Dichlorobenzene	10.000	12.074	-20.7	135	0.00
77 T	1,2-Dichlorobenzene	10.000	12.068	-20.7	140	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	11.143	-11.4	137	0.00
79 T	Naphthalene	10.000	10.733	-7.3	115	0.00
80 T	Naphthalene,2-methyl-	10.000	9.302	7.0	99	0.00
81 T	1,2,4-Trichlorobenzene	10.000	12.024	-20.2	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



QC SAMPLE DATA



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043202.D
Acq On : 18 Nov 2025 08:46
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Wed Nov 19 01:11:25 2025



AutoFind: Scans 3175, 3176, 3177; Background Corrected with Scan 3156



Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	22.1	30285	PASS
75	95	30	66	55.6	76296	PASS
95	95	100	100	100.0	137266	PASS
96	95	5	9	6.7	9227	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	63.6	87269	PASS
175	174	4	9	7.8	6847	PASS
176	174	93	101	95.5	83301	PASS
177	176	5	9	6.6	5509	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043304.D
Acq On : 08 Dec 2025 07:56
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Wed Nov 19 01:11:25 2025



AutoFind: Scans 3176, 3177, 3178; Background Corrected with Scan 3162

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	22.8	30677	PASS
75	95	30	66	55.8	74976	PASS
95	95	100	100	100.0	134320	PASS
96	95	5	9	6.7	8968	PASS
173	174	0.00	2	0.9	774	PASS
174	95	50	120	67.2	90299	PASS
175	174	4	9	7.8	7010	PASS
176	174	93	101	93.8	84736	PASS
177	176	5	9	6.8	5800	PASS

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: VL1208ABL01
Lab Sample ID: VL1208ABL01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected:
Date Received:
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.11	0.54	U	1	0.54	2.47	12/08/25 09:36	VL120825
74-87-3	Chloromethane	0.10	0.21	U	1	0.21	1.03	12/08/25 09:36	VL120825
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 09:36	VL120825
74-83-9	Bromomethane	0.080	0.31	U	1	0.31	1.94	12/08/25 09:36	VL120825
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	12/08/25 09:36	VL120825
109-99-9	Tetrahydrofuran	0.080	0.24	U	1	0.24	1.47	12/08/25 09:36	VL120825
75-69-4	Trichlorofluoromethane	0.17	0.96	U	1	0.96	2.81	12/08/25 09:36	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1	1.07	3.83	12/08/25 09:36	VL120825
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1	1.05	3.49	12/08/25 09:36	VL120825
75-65-0	tert-Butyl alcohol	0.16	0.49	U	1	0.49	1.52	12/08/25 09:36	VL120825
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	12/08/25 09:36	VL120825
75-35-4	1,1-Dichloroethene	0.15	0.59	U	1	0.59	1.98	12/08/25 09:36	VL120825
67-64-1	Acetone	0.18	0.43	U	1	0.43	1.19	12/08/25 09:36	VL120825
75-15-0	Carbon Disulfide	0.080	0.25	U	1	0.25	1.56	12/08/25 09:36	VL120825
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	1	0.83	1.80	12/08/25 09:36	VL120825
75-09-2	Methylene Chloride	0.23	0.80	U	1	0.80	1.74	12/08/25 09:36	VL120825
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	1	0.48	1.98	12/08/25 09:36	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 09:36	VL120825
110-82-7	Cyclohexane	0.22	0.76	U	1	0.76	1.72	12/08/25 09:36	VL120825
78-93-3	2-Butanone	0.060	0.18	U	1	0.18	1.47	12/08/25 09:36	VL120825
56-23-5	Carbon Tetrachloride	0.030	0.19	U	1	0.19	0.19	12/08/25 09:36	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 09:36	VL120825
67-66-3	Chloroform	0.10	0.49	U	1	0.49	2.44	12/08/25 09:36	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 09:36	VL120825
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	1	0.65	2.34	12/08/25 09:36	VL120825
71-43-2	Benzene	0.080	0.26	U	1	0.26	1.60	12/08/25 09:36	VL120825
107-06-2	1,2-Dichloroethane	0.090	0.36	U	1	0.36	2.02	12/08/25 09:36	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 09:36	VL120825
78-87-5	1,2-Dichloropropane	0.13	0.60	U	1	0.60	2.31	12/08/25 09:36	VL120825
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	12/08/25 09:36	VL120825
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	1	0.41	2.05	12/08/25 09:36	VL120825
108-88-3	Toluene	0.16	0.60	U	1	0.60	1.88	12/08/25 09:36	VL120825
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	1	0.68	2.27	12/08/25 09:36	VL120825
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	1	0.50	2.27	12/08/25 09:36	VL120825
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	1	0.44	2.73	12/08/25 09:36	VL120825
124-48-1	Dibromochloromethane	0.090	0.77	U	1	0.77	4.26	12/08/25 09:36	VL120825
106-93-4	1,2-Dibromoethane	0.090	0.69	U	1	0.69	0.77	12/08/25 09:36	VL120825
127-18-4	Tetrachloroethene	0.020	0.14	U	1	0.14	0.20	12/08/25 09:36	VL120825
108-90-7	Chlorobenzene	0.080	0.37	U	1	0.37	2.30	12/08/25 09:36	VL120825
100-41-4	Ethyl Benzene	0.19	0.83	U	1	0.83	2.17	12/08/25 09:36	VL120825
179601-23-1	m/p-Xylene	0.41	1.78	U	1	1.78	4.34	12/08/25 09:36	VL120825
95-47-6	o-Xylene	0.21	0.91	U	1	0.91	2.17	12/08/25 09:36	VL120825
100-42-5	Styrene	0.16	0.68	U	1	0.68	2.13	12/08/25 09:36	VL120825
75-25-2	Bromoform	0.050	0.52	U	1	0.52	5.17	12/08/25 09:36	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	1	0.14	0.21	12/08/25 09:36	VL120825
95-49-8	2-Chlorotoluene	0.17	0.88	U	1	0.88	2.59	12/08/25 09:36	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: VL1208ABL01
Lab Sample ID: VL1208ABL01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected:
Date Received:
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 09:36	VL120825
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 09:36	VL120825
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	1	0.30	3.01	12/08/25 09:36	VL120825
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 09:36	VL120825
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 09:36	VL120825
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1	1.56	3.71	12/08/25 09:36	VL120825
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1	1.39	5.33	12/08/25 09:36	VL120825
106-99-0	1,3-Butadiene	0.050	0.11	U	1	0.11	1.11	12/08/25 09:36	VL120825
91-20-3	Naphthalene	0.010	0.050	U	1	0.050	0.52	12/08/25 09:36	VL120825
622-96-8	4-Ethyltoluene	0.21	1.03	U	1	1.03	2.46	12/08/25 09:36	VL120825
110-54-3	Hexane	0.16	0.56	U	1	0.56	1.76	12/08/25 09:36	VL120825
107-05-1	Allyl Chloride	0.11	0.34	U	1	0.34	1.57	12/08/25 09:36	VL120825
123-91-1	1,4-Dioxane	0.22	0.79	U	1	0.79	1.80	12/08/25 09:36	VL120825
80-62-6	Methyl Methacrylate	0.14	0.57	U	1	0.57	2.05	12/08/25 09:36	VL120825

SURROGATES

460-00-4	1-Bromo-4-Fluorobenzene	9.10		70 (65) - 130 (135)	91% SPK: 10
----------	-------------------------	------	--	---------------------	-------------

INTERNAL STANDARDS

	Area Count	
74-97-5	Bromochloromethane	126000
540-36-3	1,4-Difluorobenzene	313000
3114-55-4	Chlorobenzene-d5	240000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043306.D
Acq On : 08 Dec 2025 09:36
Operator : SY/MD
Sample : VL1208ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABL01

Quant Time: Dec 09 00:08:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.781	49	125653	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	312886	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	239807	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.368	95	159096	9.051	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	90.500%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043306.D
Acq On : 08 Dec 2025 09:36
Operator : SY/MD
Sample : VL1208ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABL01

Quant Time: Dec 09 00:08:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 81
Delta R.T. -0.000 min
Lab File: VL043306.D
Acq: 08 Dec 2025 09:36




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABL01


Tgt Ion: 49 Resp: 125653
Ion Ratio Lower Upper
49 100
128 49.3 23.0 69.0
130 63.3 29.9 89.7



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.949 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VL043306.D
Acq: 08 Dec 2025 09:36

Tgt Ion: 114 Resp: 312886
Ion Ratio Lower Upper
114 100
63 25.2 21.0 31.6
88 17.7 14.9 22.3









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.869 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL043306.D
Acq: 08 Dec 2025 09:36




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABL01

Tgt Ion:117 Resp: 239807
Ion Ratio Lower Upper
117 100
82 64.8 56.1 84.1
119 31.5 25.0 37.6



#68
1-Bromo-4-Fluorobenzene
Concen: 9.051 ppbv
RT: 10.368 min Scan# 3176
Delta R.T. 0.000 min
Lab File: VL043306.D
Acq: 08 Dec 2025 09:36

Tgt Ion: 95 Resp: 159096
Ion Ratio Lower Upper
95 100
174 68.7 50.2 75.2
175 5.0 4.1 6.1



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: VL1208ABS01
Lab Sample ID: VL1208ABS01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected:
Date Received:
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	10.8	53.4		1	0.54	2.47	12/08/25 10:24	VL120825
74-87-3	Chloromethane	10.4	21.5		1	0.21	1.03	12/08/25 10:24	VL120825
75-01-4	Vinyl Chloride	9.20	23.5		1	0.080	0.080	12/08/25 10:24	VL120825
74-83-9	Bromomethane	9.30	36.1		1	0.31	1.94	12/08/25 10:24	VL120825
75-00-3	Chloroethane	9.90	26.1		1	0.40	1.32	12/08/25 10:24	VL120825
109-99-9	Tetrahydrofuran	9.70	28.6		1	0.24	1.47	12/08/25 10:24	VL120825
75-69-4	Trichlorofluoromethane	11.2	62.9		1	0.96	2.81	12/08/25 10:24	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	11.2	85.8		1	1.07	3.83	12/08/25 10:24	VL120825
76-14-2	Dichlorotetrafluoroethane	10.6	74.1		1	1.05	3.49	12/08/25 10:24	VL120825
75-65-0	tert-Butyl alcohol	8.90	27.0		1	0.49	1.52	12/08/25 10:24	VL120825
142-82-5	Heptane	9.40	38.5		1	0.70	2.05	12/08/25 10:24	VL120825
75-35-4	1,1-Dichloroethene	10.2	40.4		1	0.59	1.98	12/08/25 10:24	VL120825
67-64-1	Acetone	8.60	20.4		1	0.43	1.19	12/08/25 10:24	VL120825
75-15-0	Carbon Disulfide	11.1	34.6		1	0.25	1.56	12/08/25 10:24	VL120825
1634-04-4	Methyl tert-Butyl Ether	9.40	33.9		1	0.83	1.80	12/08/25 10:24	VL120825
75-09-2	Methylene Chloride	11.6	40.3		1	0.80	1.74	12/08/25 10:24	VL120825
156-60-5	trans-1,2-Dichloroethene	10.3	40.8		1	0.48	1.98	12/08/25 10:24	VL120825
75-34-3	1,1-Dichloroethane	10.5	42.5		1	0.53	2.02	12/08/25 10:24	VL120825
110-82-7	Cyclohexane	9.70	33.4		1	0.76	1.72	12/08/25 10:24	VL120825
78-93-3	2-Butanone	10.3	30.4		1	0.18	1.47	12/08/25 10:24	VL120825
56-23-5	Carbon Tetrachloride	12.1	76.1		1	0.19	0.19	12/08/25 10:24	VL120825
156-59-2	cis-1,2-Dichloroethene	9.90	39.3		1	0.40	1.98	12/08/25 10:24	VL120825
67-66-3	Chloroform	11.7	57.1		1	0.49	2.44	12/08/25 10:24	VL120825
71-55-6	1,1,1-Trichloroethane	10.9	59.5		1	0.11	0.16	12/08/25 10:24	VL120825
540-84-1	2,2,4-Trimethylpentane	10.8	50.4		1	0.65	2.34	12/08/25 10:24	VL120825
71-43-2	Benzene	11.2	35.8		1	0.26	1.60	12/08/25 10:24	VL120825
107-06-2	1,2-Dichloroethane	11.9	48.2		1	0.36	2.02	12/08/25 10:24	VL120825
79-01-6	Trichloroethene	11.5	61.8		1	0.11	0.16	12/08/25 10:24	VL120825
78-87-5	1,2-Dichloropropane	10.9	50.4		1	0.60	2.31	12/08/25 10:24	VL120825
75-27-4	Bromodichloromethane	12.0	80.4		1	0.40	3.35	12/08/25 10:24	VL120825
108-10-1	4-Methyl-2-Pentanone	10.0	41.0		1	0.41	2.05	12/08/25 10:24	VL120825
108-88-3	Toluene	10.3	38.8		1	0.60	1.88	12/08/25 10:24	VL120825
10061-02-6	t-1,3-Dichloropropene	7.80	35.4		1	0.68	2.27	12/08/25 10:24	VL120825
10061-01-5	cis-1,3-Dichloropropene	8.30	37.7		1	0.50	2.27	12/08/25 10:24	VL120825
79-00-5	1,1,2-Trichloroethane	11.6	63.3		1	0.44	2.73	12/08/25 10:24	VL120825
124-48-1	Dibromochloromethane	11.8	101		1	0.77	4.26	12/08/25 10:24	VL120825
106-93-4	1,2-Dibromoethane	10.5	80.7		1	0.69	0.77	12/08/25 10:24	VL120825
127-18-4	Tetrachloroethene	11.1	75.3		1	0.14	0.20	12/08/25 10:24	VL120825
108-90-7	Chlorobenzene	11.9	54.8		1	0.37	2.30	12/08/25 10:24	VL120825
100-41-4	Ethyl Benzene	11.2	48.6		1	0.83	2.17	12/08/25 10:24	VL120825
179601-23-1	m/p-Xylene	23.5	102		1	1.78	4.34	12/08/25 10:24	VL120825
95-47-6	o-Xylene	11.6	50.4		1	0.91	2.17	12/08/25 10:24	VL120825
100-42-5	Styrene	9.90	42.1		1	0.68	2.13	12/08/25 10:24	VL120825
75-25-2	Bromoform	12.2	126		1	0.52	5.17	12/08/25 10:24	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	11.6	79.7		1	0.14	0.21	12/08/25 10:24	VL120825
95-49-8	2-Chlorotoluene	11.0	57.0		1	0.88	2.59	12/08/25 10:24	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: VL1208ABS01
Lab Sample ID: VL1208ABS01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected:
Date Received:
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	11.4	56.0	1	0.88		2.46	12/08/25 10:24	VL120825
95-63-6	1,2,4-Trimethylbenzene	11.5	56.5	1	0.88		2.46	12/08/25 10:24	VL120825
541-73-1	1,3-Dichlorobenzene	11.8	70.9	1	0.30		3.01	12/08/25 10:24	VL120825
106-46-7	1,4-Dichlorobenzene	11.9	71.5	1	0.78		3.01	12/08/25 10:24	VL120825
95-50-1	1,2-Dichlorobenzene	11.8	70.9	1	0.78		3.01	12/08/25 10:24	VL120825
120-82-1	1,2,4-Trichlorobenzene	11.9	88.3	1	1.56		3.71	12/08/25 10:24	VL120825
87-68-3	Hexachloro-1,3-Butadiene	10.9	116	1	1.39		5.33	12/08/25 10:24	VL120825
106-99-0	1,3-Butadiene	9.60	21.2	1	0.11		1.11	12/08/25 10:24	VL120825
91-20-3	Naphthalene	10.7	56.1	1	0.050		0.52	12/08/25 10:24	VL120825
622-96-8	4-Ethyltoluene	11.5	56.5	1	1.03		2.46	12/08/25 10:24	VL120825
110-54-3	Hexane	9.90	34.9	1	0.56		1.76	12/08/25 10:24	VL120825
107-05-1	Allyl Chloride	8.80	27.5	1	0.34		1.57	12/08/25 10:24	VL120825
123-91-1	1,4-Dioxane	8.70	31.4	1	0.79		1.80	12/08/25 10:24	VL120825
80-62-6	Methyl Methacrylate	9.70	39.7	1	0.57		2.05	12/08/25 10:24	VL120825

SURROGATES

460-00-4 1-Bromo-4-Fluorobenzene 10.1 70 (65) - 130 (135) 101% SPK: 10

INTERNAL STANDARDS

Area Count
74-97-5 Bromochloromethane 128000
540-36-3 1,4-Difluorobenzene 319000
3114-55-4 Chlorobenzene-d5 251000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043307.D
 Acq On : 08 Dec 2025 10:24
 Operator : SY/MD
 Sample : VL1208ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01

Quant Time: Dec 09 00:09:13 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 12/09/2025
 Supervised By : Mahesh Dadoda 12/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.777	49	128373	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	319254	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	250711	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.364	95	185879	10.115	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.496	85	195547	10.800	ppbv	100
3) Chlorodifluoromethane	1.473	51	181736	10.206	ppbv	97
4) Chloromethane	1.531	50	70563	10.423	ppbv	100
5) Vinyl Chloride	1.577	62	67921	9.174	ppbv	93
6) Bromomethane	1.664	94	33521	9.340	ppbv	98
7) Chloroethane	1.700	64	28601	9.919	ppbv	91
8) Dichlorotetrafluoroethane	1.551	85	159668	10.601	ppbv	97
9) Propene	1.479	41	71417	9.445	ppbv #	1
10) Heptane	4.962	43	210516	9.355	ppbv	98
11) Trichlorofluoromethane	1.874	101	196801	11.219	ppbv	99
12) 1,1,2-Trichlorotrifluoro...	2.133	101	153996	11.244	ppbv	97
13) Ethanol	1.716	45	23202	8.726	ppbv	97
14) Bromoethene	1.777	108	55304	10.533	ppbv	94
15) Acetone	1.829	43	152789	8.567	ppbv	99
16) 1,3-Butadiene	1.606	39	76687	9.645	ppbv	97
17) tert-Butyl alcohol	2.033	59	209336	8.942	ppbv	100
18) 1,1-Dichloroethene	2.030	96	65282	10.237	ppbv	93
19) Isopropyl Alcohol	1.881	45	85723	8.506	ppbv #	86
20) Methylene Chloride	2.056	84	60065	11.551	ppbv	95
21) Allyl Chloride	2.091	41	105272	8.773	ppbv	100
22) trans-1,2-Dichloroethene	2.334	96	75422	10.348	ppbv	95
23) Vinyl Acetate	2.454	43	197206	9.159	ppbv #	98
24) 1,1-Dichloroethane	2.402	63	147058	10.453	ppbv	98
25) Ethyl Acetate	2.826	43	372295	9.678	ppbv	99
26) Hexane	2.816	57	166920	9.898	ppbv	98
27) Carbon Disulfide	2.146	76	190369	11.057	ppbv	99
28) Methyl tert-Butyl Ether	2.431	73	77723	9.389	ppbv	98
29) Chloroform	2.839	83	280544	11.703	ppbv	98
30) Cyclohexane	3.842	84	136887	9.681	ppbv	95
31) cis-1,2-Dichloroethene	2.713	61	166197	9.915	ppbv	94
32) 1,1,1-Trichloroethane	3.357	97	278163	10.886	ppbv	99
34) 2-Butanone	2.548	43	232285	10.307	ppbv	99
35) Carbon Tetrachloride	3.752	117	293430	12.109	ppbv	100
36) Benzene	3.645	78	342251	11.194	ppbv	99
37) 1,2-Dichloroethane	3.208	62	212457	11.880	ppbv	98
38) Trichloroethene	4.532	130	137474	11.493	ppbv	98
39) 1,2-Dichloropropane	4.295	63	122348	10.934	ppbv	92
40) 1,4-Dioxane	4.564	88	43982	8.743	ppbv #	97
41) Tetrahydrofuran	3.043	42	121301	9.747	ppbv	98
42) Bromodichloromethane	4.473	83	296233	12.008	ppbv	97
43) Methyl Methacrylate	4.862	69	156667	9.654	ppbv	95
44) 2,2,4-Trimethylpentane	4.625	57	569168	10.839	ppbv	98
45) t-1,3-Dichloropropene	6.393	75	118391	7.826	ppbv	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043307.D
 Acq On : 08 Dec 2025 10:24
 Operator : SY/MD
 Sample : VL1208ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01

Quant Time: Dec 09 00:09:13 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025


Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 12/09/2025
 Supervised By : Mahesh Dadoda 12/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.580	75	146637	8.255	ppbv	94
47) 1,1,2-Trichloroethane	6.561	97	137247	11.557	ppbv	98
48) Dibromochloromethane	7.370	129	240784	11.793	ppbv	99
49) Bromoform	9.471	173	220417	12.188	ppbv	99
50) 4-Methyl-2-Pentanone	5.723	43	315888	9.951	ppbv	98
51) 2-Hexanone	7.415	43	223708	9.075	ppbv #	98
52) Tetrachloroethene	8.208	164	125681	11.051	ppbv	96
53) Toluene	6.914	91	374608	10.347	ppbv	99
54) 1,2-Dibromoethane	7.636	107	183811	10.518	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.911	131	229368	12.449	ppbv	99
57) Chlorobenzene	8.908	112	305509	11.869	ppbv	99
58) Ethyl Benzene	9.341	91	538550	11.161	ppbv	97
59) m/p-Xylene	9.539	91	909288m	23.503	ppbv	
60) o-Xylene	9.950	91	446789	11.560	ppbv	98
61) Styrene	9.859	104	167877	9.925	ppbv	99
62) Isopropylbenzene	10.529	105	805674	11.652	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.950	83	285542	11.611	ppbv	99
64) n-propylbenzene	10.979	120	208180	11.718	ppbv	96
65) tert-Butylbenzene	11.523	119	696094	11.273	ppbv	100
66) Benzyl Chloride	11.604	91	31638	5.049	ppbv	98
67) sec-Butylbenzene	11.759	105	973259	11.291	ppbv	99
69) p-Isopropyltoluene	11.914	119	808286	11.070	ppbv	99
70) n-Butylbenzene	12.270	91	780553	11.019	ppbv	100
71) 2-Chlorotoluene	10.888	91	584474	10.997	ppbv	99
72) 4-Ethyltoluene	11.115	105	540638	11.487	ppbv	99
73) 1,3,5-Trimethylbenzene	11.193	105	451214	11.424	ppbv	96
74) 1,2,4-Trimethylbenzene	11.526	105	496984	11.530	ppbv	100
75) 1,3-Dichlorobenzene	11.597	146	297680	11.834	ppbv	98
76) 1,4-Dichlorobenzene	11.662	146	298352	11.859	ppbv	98
77) 1,2-Dichlorobenzene	11.937	146	285835	11.819	ppbv	98
78) Hexachloro-1,3-Butadiene	13.873	225	220056	10.903	ppbv	99
79) Naphthalene	13.504	128	409312	10.692	ppbv	97
80) Naphthalene,2-methyl-	14.449	142	92975	9.252	ppbv	99
81) 1,2,4-Trichlorobenzene	13.432	180	204005	11.915	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.777 min Scan# 81
Delta R.T. -0.004 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 49 Resp: 12837
Ion Ratio Lower Upper
49 100
128 48.0 23.0 69.0
130 62.5 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#2
Dichlorodifluoromethane
Concen: 10.800 ppbv
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 85 Resp: 195547
Ion Ratio Lower Upper
85 100
87 32.0 25.4 38.2









#3
Chlorodifluoromethane
Concen: 10.206 ppbv
RT: 1.473 min Scan# 428
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 51 Resp: 181730
Ion Ratio Lower Upper
51 100
67 19.9 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#4
Chloromethane
Concen: 10.423 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 50 Resp: 70563
Ion Ratio Lower Upper
50 100
52 30.2 24.1 36.1









#5
 Vinyl Chloride
 Concen: 9.174 ppbv
 RT: 1.577 min Scan# 460
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01

Tgt Ion: 62 Resp: 6792
 Ion Ratio Lower Upper
 62 100
 64 30.2 27.3 40.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#6
 Bromomethane
 Concen: 9.340 ppbv
 RT: 1.664 min Scan# 487
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion: 94 Resp: 33521
 Ion Ratio Lower Upper
 94 100
 96 93.0 72.9 109.3









#7
Chloroethane
Concen: 9.919 ppbv
RT: 1.700 min Scan# 498
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument : MSVOA_L
ClientSampleId : VL1208ABS01

Tgt Ion: 64 Resp: 2860
Ion Ratio Lower Upper
64 100
66 28.6 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#8
Dichlorotetrafluoroethane
Concen: 10.601 ppbv
RT: 1.551 min Scan# 452
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 85 Resp: 159668
Ion Ratio Lower Upper
85 100
135 68.3 52.6 79.0









#9
Propene
Concen: 9.445 ppbv
RT: 1.479 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 41 Resp: 7141
Ion Ratio Lower Upper
41 100
42 67.1 3.7 5.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#10
Heptane
Concen: 9.355 ppbv
RT: 4.962 min Scan# 1506
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 43 Resp: 210516
Ion Ratio Lower Upper
43 100
57 50.9 39.7 59.5









#11
 Trichlorofluoromethane
 Concen: 11.219 ppbv
 RT: 1.874 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01

Tgt Ion:101 Resp: 19680
 Ion Ratio Lower Upper
 101 100
 103 65.2 52.6 79.0




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 11.244 ppbv
 RT: 2.133 min Scan# 632
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion:101 Resp: 153996
 Ion Ratio Lower Upper
 101 100
 151 76.4 59.2 88.8









#13
Ethanol
Concen: 8.726 ppbv
RT: 1.716 min Scan# 503
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 45 Resp: 2320
Ion Ratio Lower Upper
45 100
46 35.5 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#14
Bromoethene
Concen: 10.533 ppbv
RT: 1.777 min Scan# 522
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24


Tgt Ion:108 Resp: 55304
Ion Ratio Lower Upper
108 100
106 104.9 88.9 133.3
79 17.7 16.5 24.7





#15
Acetone
Concen: 8.567 ppbv
RT: 1.829 min Scan# 51
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion: 43 Resp: 152789
Ion Ratio Lower Upper
43 100
58 26.6 20.7 31.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#16
1,3-Butadiene
Concen: 9.645 ppbv
RT: 1.606 min Scan# 469
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24



Tgt Ion: 39 Resp: 76687
Ion Ratio Lower Upper
39 100
54 71.8 59.4 89.2









#17
tert-Butyl alcohol
Concen: 8.942 ppbv
RT: 2.033 min Scan# 602
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 59 Resp: 209330
Ion Ratio Lower Upper
59 100
57 11.2 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#18
1,1-Dichloroethene
Concen: 10.237 ppbv
RT: 2.030 min Scan# 600
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 96 Resp: 65282
Ion Ratio Lower Upper
96 100
61 200.9 170.6 255.8
98 62.7 52.2 78.4









#19
Isopropyl Alcohol
Concen: 8.506 ppbv
RT: 1.881 min Scan# 554
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 45 Resp: 8572
Ion Ratio Lower Upper
45 100
58 47.5 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#20
Methylene Chloride
Concen: 11.551 ppbv
RT: 2.056 min Scan# 608
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 84 Resp: 60065
Ion Ratio Lower Upper
84 100
49 158.6 132.9 199.3
51 48.9 34.8 52.2
86 60.8 48.2 72.2









#21
 Allyl Chloride
 Concen: 8.773 ppbv
 RT: 2.091 min Scan# 619
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01

Tgt Ion: 41 Resp: 105272
 Ion Ratio Lower Upper
 41 100
 76 27.6 22.1 33.1
 39 76.5 61.0 91.4




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#22
 trans-1,2-Dichloroethene
 Concen: 10.348 ppbv
 RT: 2.334 min Scan# 694
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion: 96 Resp: 75422
 Ion Ratio Lower Upper
 96 100
 61 165.0 139.0 208.6
 98 63.3 49.4 74.2









#23
 Vinyl Acetate
 Concen: 9.159 ppbv
 RT: 2.454 min Scan# 715
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01

Tgt Ion: 43 Resp: 197200
 Ion Ratio Lower Upper
 43 100
 86 8.3 6.2 9.2
 42 10.1 8.0 12.0
 44 6.8 4.0 6.0#




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#24
 1,1-Dichloroethane
 Concen: 10.453 ppbv
 RT: 2.402 min Scan# 715
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24


Tgt Ion: 63 Resp: 147058
 Ion Ratio Lower Upper
 63 100
 98 4.8 2.8 8.4
 100 3.1 1.3 3.9





#25
Ethyl Acetate
Concen: 9.678 ppbv
RT: 2.826 min Scan# 84
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion: 43 Resp: 37229
Ion Ratio Lower Upper
43 100
45 9.8 8.1 12.1
61 8.6 7.0 10.4
70 8.0 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#26
Hexane
Concen: 9.898 ppbv
RT: 2.816 min Scan# 843
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Tgt Ion: 57 Resp: 166920
Ion Ratio Lower Upper
57 100
41 96.6 74.6 111.8
43 224.2 179.0 268.4
56 56.9 42.3 63.5





#27
Carbon Disulfide
Concen: 11.057 ppbv
RT: 2.146 min Scan# 61
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion: 76 Resp: 190369
Ion Ratio Lower Upper
76 100
44 19.1 16.1 24.1
78 11.8 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#28
Methyl tert-Butyl Ether
Concen: 9.389 ppbv
RT: 2.431 min Scan# 724
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24



Tgt Ion: 73 Resp: 77723
Ion Ratio Lower Upper
73 100
57 23.8 19.8 29.8





#29

Chloroform

Concen: 11.703 ppbv

RT: 2.839 min Scan# 81

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :

VL1208ABS01

Tgt Ion: 83 Resp: 280544

Ion Ratio Lower Upper

83 100





85 64.6 50.5 75.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025

Supervised By :Mahesh Dadoda 12/11/2025



#30

Cyclohexane

Concen: 9.681 ppbv

RT: 3.842 min Scan# 1160

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24




Tgt Ion: 84 Resp: 136887


Ion Ratio Lower Upper

84 100

56 123.4 104.7 157.1


41 84.6 69.0 103.6





#31
 cis-1,2-Dichloroethene
 Concen: 9.915 ppbv
 RT: 2.713 min Scan# 81
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24




Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01




Tgt Ion: 61 Resp: 16619
 Ion Ratio Lower Upper
 61 100
 96 67.8 49.3 73.9
 98 42.0 32.1 48.1

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025




#32
 1,1,1-Trichloroethane
 Concen: 10.886 ppbv
 RT: 3.357 min Scan# 1010
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24



Tgt Ion: 97 Resp: 278163
 Ion Ratio Lower Upper
 97 100
 99 64.3 50.3 75.5
 61 49.0 38.9 58.3





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.946 min Scan# 1193

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :

VL1208ABS01





Tgt Ion:	114	Resp:	319254
Ion Ratio	Lower	Upper	
114	100		
63	25.8	21.0	31.6
88	17.9	14.9	22.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025

Supervised By :Mahesh Dadoda 12/11/2025



#34

2-Butanone

Concen: 10.307 ppbv




RT: 2.548 min Scan# 760


Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24


Tgt Ion:	43	Resp:	232285
Ion Ratio	Lower	Upper	
43	100		
72	22.3	18.1	27.1





#35
Carbon Tetrachloride
Concen: 12.109 ppbv
RT: 3.752 min Scan# 1133
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion:117 Resp: 293430
Ion Ratio Lower Upper
117 100
119 92.9 74.6 112.0
121 29.0 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#36
Benzene
Concen: 11.194 ppbv
RT: 3.645 min Scan# 1099
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24



Tgt Ion: 78 Resp: 342251
Ion Ratio Lower Upper
78 100
77 21.3 17.5 26.3
50 19.1 15.5 23.3









#37
1,2-Dichloroethane
Concen: 11.880 ppbv
RT: 3.208 min Scan# 964
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 62 Resp: 21245
Ion Ratio Lower Upper
62 100
98 6.2 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#38
Trichloroethene
Concen: 11.493 ppbv
RT: 4.532 min Scan# 1373
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:130 Resp: 137474
Ion Ratio Lower Upper
130 100
95 104.4 85.7 128.5
132 88.8 73.4 110.0
97 65.2 52.7 79.1





#39

1,2-Dichloropropane

Concen: 10.934 ppbv

RT: 4.295 min Scan# 11

Delta R.T. -0.007 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :

VL1208ABS01

Tgt Ion: 63 Resp: 12234

Ion Ratio Lower Upper

63 100

41 86.4 62.6 93.8





62 68.2 51.0 76.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025

Supervised By :Mahesh Dadoda 12/11/2025



#40

1,4-Dioxane

Concen: 8.743 ppbv

RT: 4.564 min Scan# 1383

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Tgt Ion: 88 Resp: 43982




Ion Ratio Lower Upper


88 100

58 76.2 63.0 94.6

43 37.8 32.5 48.7

57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 9.747 ppbv
RT: 3.043 min Scan# 914
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 42 Resp: 12130
Ion Ratio Lower Upper
42 100
72 38.9 30.6 46.0
71 36.7 30.5 45.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#42
Bromodichloromethane
Concen: 12.008 ppbv
RT: 4.473 min Scan# 1355
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 83 Resp: 296233
Ion Ratio Lower Upper
83 100
85 62.5 51.7 77.5
127 8.5 6.3 9.5









#43
Methyl Methacrylate
Concen: 9.654 ppbv
RT: 4.862 min Scan# 1476
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 69 Resp: 15666
Ion Ratio Lower Upper
69 100
41 150.8 126.7 190.1
100 35.4 28.6 43.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#44
2,2,4-Trimethylpentane
Concen: 10.839 ppbv
RT: 4.625 min Scan# 1402
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 57 Resp: 569168
Ion Ratio Lower Upper
57 100
41 36.5 28.0 42.0
56 32.5 25.6 38.4









#45
 t-1,3-Dichloropropene
 Concen: 7.826 ppbv
 RT: 6.393 min Scan# 1950
 Delta R.T. -0.007 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01

Tgt Ion: 75 Resp: 11839
 Ion Ratio Lower Upper
 75 100
 77 29.3 24.3 36.5




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#46
 cis-1,3-Dichloropropene
 Concen: 8.255 ppbv
 RT: 5.580 min Scan# 1697
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion: 75 Resp: 146637
 Ion Ratio Lower Upper
 75 100
 39 69.0 59.1 88.7
 77 27.6 24.2 36.2









#47
1,1,2-Trichloroethane
Concen: 11.557 ppbv
RT: 6.561 min Scan# 2000
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 97 Resp: 13724
Ion Ratio Lower Upper
97 100
83 81.7 64.6 97.0
85 51.0 39.4 59.2
99 61.1 47.0 70.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#48
Dibromochloromethane
Concen: 11.793 ppbv
RT: 7.370 min Scan# 2250
Delta R.T. -0.007 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:129 Resp: 240784
Ion Ratio Lower Upper
129 100
127 79.2 39.3 117.9









#49
Bromoform
Concen: 12.188 ppbv
RT: 9.471 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion:173 Resp: 22041
Ion Ratio Lower Upper
173 100
175 48.8 38.6 58.0
171 50.4 40.6 61.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#50
4-Methyl-2-Pentanone
Concen: 9.951 ppbv
RT: 5.723 min Scan# 1741
Delta R.T. -0.010 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 43 Resp: 315888
Ion Ratio Lower Upper
43 100
58 34.5 28.4 42.6









#51
2-Hexanone
Concen: 9.075 ppbv
RT: 7.415 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 43 Resp: 223708
Ion Ratio Lower Upper
43 100
58 46.6 38.4 57.6
98 0.9 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#52
Tetrachloroethene
Concen: 11.051 ppbv
RT: 8.208 min Scan# 2509
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24


Tgt Ion:164 Resp: 125681
Ion Ratio Lower Upper
164 100
166 125.0 99.4 149.2
129 95.8 81.8 122.8
131 91.5 77.4 116.2





#53
Toluene
Concen: 10.347 ppbv
RT: 6.914 min Scan# 2109
Delta R.T. -0.007 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion: 91 Resp: 374608
Ion Ratio Lower Upper
91 100
92 58.7 47.4 71.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#54
1,2-Dibromoethane
Concen: 10.518 ppbv
RT: 7.636 min Scan# 2332
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24



Tgt Ion:107 Resp: 183811
Ion Ratio Lower Upper
107 100
109 94.4 72.1 108.1





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.869 min Scan# 21

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :





VL1208ABS01

Tgt Ion:117 Resp: 25071

Ion	Ratio	Lower	Upper
117	100		
82	69.4	56.1	84.1
119	32.6	25.0	37.6

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#56

1,1,1,2-Tetrachloroethane

Concen: 12.449 ppbv

RT: 8.911 min Scan# 2726




Delta R.T. -0.003 min


Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Tgt Ion:131 Resp: 229368

Ion	Ratio	Lower	Upper
131	100		
133	96.8	77.4	116.2
119	60.5	49.1	73.7









#57
Chlorobenzene
Concen: 11.869 ppbv
RT: 8.908 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion:112 Resp: 305509
Ion Ratio Lower Upper
112 100
77 69.6 55.5 83.3
114 30.5 29.1 35.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#58
Ethyl Benzene
Concen: 11.161 ppbv
RT: 9.341 min Scan# 2859
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 91 Resp: 538550
Ion Ratio Lower Upper
91 100
106 30.1 22.8 34.2





#59

m/p-Xylene

Concen: 23.503 ppbv m

RT: 9.539 min Scan# 2921

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :

VL1208ABS01

Tgt Ion: 91 Resp: 90928

Ion Ratio Lower Upper

91 100





106 45.1 36.1 54.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025

Supervised By :Mahesh Dadoda 12/11/2025



#60

o-Xylene

Concen: 11.560 ppbv

RT: 9.950 min Scan# 3047

Delta R.T. -0.007 min

Lab File: VL043307.D




Acq: 08 Dec 2025 10:24


Tgt Ion: 91 Resp: 446789

Ion Ratio Lower Upper

91 100

106 43.1 20.9 62.8





#61

Styrene

Concen: 9.925 ppbv

RT: 9.859 min Scan# 3020

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L





ClientSampleId :

VL1208ABS01

Tgt Ion:104 Resp: 16787
 Ion Ratio Lower Upper
 104 100
 78 51.5 42.1 63.1
 103 47.4 37.2 55.8

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#62

Isopropylbenzene

Concen: 11.652 ppbv




RT: 10.529 min Scan# 3226


Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Tgt Ion:105 Resp: 805674
 Ion Ratio Lower Upper
 105 100
 120 25.6 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 11.611 ppbv
RT: 9.950 min Scan# 3048
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion: 83 Resp: 28554
Ion Ratio Lower Upper
83 100
131 9.6 4.8 14.3
85 64.5 31.9 95.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#64
n-propylbenzene
Concen: 11.718 ppbv
RT: 10.979 min Scan# 3365
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24


Tgt Ion:120 Resp: 208180
Ion Ratio Lower Upper
120 100
91 461.2 377.1 565.7





#65
tert-Butylbenzene
Concen: 11.273 ppbv
RT: 11.523 min Scan# 3533
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion: 119 Resp: 696094
Ion Ratio Lower Upper
119 100
91 81.5 65.3 97.9
134 21.7 17.0 25.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#66
Benzyl Chloride
Concen: 5.049 ppbv
RT: 11.604 min Scan# 3558
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Tgt Ion: 91 Resp: 31638
Ion Ratio Lower Upper
91 100
126 16.9 14.4 21.6
128 6.4 4.6 6.8





#67
 sec-Butylbenzene
 Concen: 11.291 ppbv
 RT: 11.759 min Scan# 3066
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24




Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01




Tgt Ion:105 Resp: 973259
 Ion Ratio Lower Upper
 105 100
 134 19.4 15.0 22.6

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025




#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.115 ppbv
 RT: 10.364 min Scan# 3175
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24




Tgt Ion: 95 Resp: 185879
 Ion Ratio Lower Upper
 95 100
 174 69.2 50.2 75.2
 175 5.1 4.1 6.1





#69
p-Isopropyltoluene
Concen: 11.070 ppbv
RT: 11.914 min Scan# 3654
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion: 119 Resp: 808280
Ion Ratio Lower Upper
119 100
134 25.9 20.2 30.2
91 29.1 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#70
n-Butylbenzene
Concen: 11.019 ppbv
RT: 12.270 min Scan# 3764
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Tgt Ion: 91 Resp: 780553
Ion Ratio Lower Upper
91 100
92 54.6 43.5 65.3
134 24.1 19.0 28.6





#71
2-Chlorotoluene
Concen: 10.997 ppbv
RT: 10.888 min Scan# 31
Delta R.T. -0.007 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion: 91 Resp: 584474
Ion Ratio Lower Upper
91 100
126 30.0 11.7 46.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#72
4-Ethyltoluene
Concen: 11.487 ppbv
RT: 11.115 min Scan# 3407
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24



Tgt Ion:105 Resp: 540638
Ion Ratio Lower Upper
105 100
120 28.8 23.8 35.8
91 12.9 10.3 15.5
77 13.7 11.1 16.7









#73
1,3,5-Trimethylbenzene
Concen: 11.424 ppbv
RT: 11.193 min Scan# 3432
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion:105 Resp: 451214
Ion Ratio Lower Upper
105 100
120 45.6 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#74
1,2,4-Trimethylbenzene
Concen: 11.530 ppbv
RT: 11.526 min Scan# 3534
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:105 Resp: 496984
Ion Ratio Lower Upper
105 100
120 52.1 41.7 62.5
91 88.0 70.0 105.0
77 29.7 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 11.834 ppbv
RT: 11.597 min Scan# 3556
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Tgt Ion:146 Resp: 297680
Ion Ratio Lower Upper
146 100
111 46.0 23.4 70.3
148 64.9 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#76
1,4-Dichlorobenzene
Concen: 11.859 ppbv
RT: 11.662 min Scan# 3576
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:146 Resp: 298352
Ion Ratio Lower Upper
146 100
111 43.2 22.8 68.3
148 64.2 31.8 95.4





#77

1,2-Dichlorobenzene

Concen: 11.819 ppbv

RT: 11.937 min Scan# 30

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :





VL1208ABS01

Tgt Ion:	146	Resp:	28583
Ion Ratio	Lower	Upper	
146	100		
111	46.1	24.3	73.0
148	63.9	31.6	94.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#78

Hexachloro-1,3-Butadiene

Concen: 10.903 ppbv




RT: 13.873 min Scan# 4259


Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24


Tgt Ion:	225	Resp:	220056
Ion Ratio	Lower	Upper	
225	100		
223	63.6	50.2	75.4
227	64.7	51.1	76.7





#79
Naphthalene
Concen: 10.692 ppbv
RT: 13.504 min Scan# 4146
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01




Tgt Ion:128 Resp: 409312
Ion Ratio Lower Upper
128 100
127 13.5 9.9 14.9
129 11.4 8.4 12.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#80
Naphthalene,2-methyl-
Concen: 9.252 ppbv
RT: 14.449 min Scan# 4437
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24



Tgt Ion:142 Resp: 92975
Ion Ratio Lower Upper
142 100
141 84.6 67.1 100.7
115 39.8 31.9 47.9





#81

1,2,4-Trichlorobenzene

Concen: 11.915 ppbv

RT: 13.432 min Scan# 41

Delta R.T. -0.003 min

Lab File: VL043307.D


Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :

VL1208ABS01



Tgt Ion:180 Resp: 204000

Ion Ratio Lower Upper

180 100

182 94.3 77.4 116.2



184 30.3 24.2 36.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025

Supervised By :Mahesh Dadoda 12/11/2025



Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: AA-1DUP
Lab Sample ID: Q3799-01DUP
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected:
Date Received:
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.41	2.03	J	1	0.54	2.47	12/08/25 13:16	VL120825
74-87-3	Chloromethane	2.10	4.34		1	0.21	1.03	12/08/25 13:16	VL120825
75-01-4	Vinyl Chloride	0.030	0.080	U	1	0.080	0.080	12/08/25 13:16	VL120825
74-83-9	Bromomethane	0.080	0.31	U	1	0.31	1.94	12/08/25 13:16	VL120825
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	12/08/25 13:16	VL120825
109-99-9	Tetrahydrofuran	0.080	0.24	U	1	0.24	1.47	12/08/25 13:16	VL120825
75-69-4	Trichlorofluoromethane	0.20	1.12	J	1	0.96	2.81	12/08/25 13:16	VL120825
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1	1.07	3.83	12/08/25 13:16	VL120825
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1	1.05	3.49	12/08/25 13:16	VL120825
75-65-0	tert-Butyl alcohol	0.16	0.49	U	1	0.49	1.52	12/08/25 13:16	VL120825
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	12/08/25 13:16	VL120825
75-35-4	1,1-Dichloroethene	0.15	0.59	U	1	0.59	1.98	12/08/25 13:16	VL120825
67-64-1	Acetone	4.80	11.4		1	0.43	1.19	12/08/25 13:16	VL120825
75-15-0	Carbon Disulfide	0.080	0.25	U	1	0.25	1.56	12/08/25 13:16	VL120825
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	1	0.83	1.80	12/08/25 13:16	VL120825
75-09-2	Methylene Chloride	3.30	11.5		1	0.80	1.74	12/08/25 13:16	VL120825
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	1	0.48	1.98	12/08/25 13:16	VL120825
75-34-3	1,1-Dichloroethane	0.13	0.53	U	1	0.53	2.02	12/08/25 13:16	VL120825
110-82-7	Cyclohexane	0.22	0.76	U	1	0.76	1.72	12/08/25 13:16	VL120825
78-93-3	2-Butanone	0.51	1.50		1	0.18	1.47	12/08/25 13:16	VL120825
56-23-5	Carbon Tetrachloride	0.090	0.57		1	0.19	0.19	12/08/25 13:16	VL120825
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	1	0.40	1.98	12/08/25 13:16	VL120825
67-66-3	Chloroform	0.15	0.73	J	1	0.49	2.44	12/08/25 13:16	VL120825
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	1	0.11	0.16	12/08/25 13:16	VL120825
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	1	0.65	2.34	12/08/25 13:16	VL120825
71-43-2	Benzene	0.16	0.51	J	1	0.26	1.60	12/08/25 13:16	VL120825
107-06-2	1,2-Dichloroethane	0.090	0.36	U	1	0.36	2.02	12/08/25 13:16	VL120825
79-01-6	Trichloroethene	0.020	0.11	U	1	0.11	0.16	12/08/25 13:16	VL120825
78-87-5	1,2-Dichloropropane	0.13	0.60	U	1	0.60	2.31	12/08/25 13:16	VL120825
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	12/08/25 13:16	VL120825
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	1	0.41	2.05	12/08/25 13:16	VL120825
108-88-3	Toluene	0.16	0.60	U	1	0.60	1.88	12/08/25 13:16	VL120825
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	1	0.68	2.27	12/08/25 13:16	VL120825
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	1	0.50	2.27	12/08/25 13:16	VL120825
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	1	0.44	2.73	12/08/25 13:16	VL120825
124-48-1	Dibromochloromethane	0.090	0.77	U	1	0.77	4.26	12/08/25 13:16	VL120825
106-93-4	1,2-Dibromoethane	0.090	0.69	U	1	0.69	0.77	12/08/25 13:16	VL120825
127-18-4	Tetrachloroethene	0.020	0.14	U	1	0.14	0.20	12/08/25 13:16	VL120825
108-90-7	Chlorobenzene	0.080	0.37	U	1	0.37	2.30	12/08/25 13:16	VL120825
100-41-4	Ethyl Benzene	0.19	0.83	U	1	0.83	2.17	12/08/25 13:16	VL120825
179601-23-1	m/p-Xylene	0.41	1.78	U	1	1.78	4.34	12/08/25 13:16	VL120825
95-47-6	o-Xylene	0.21	0.91	U	1	0.91	2.17	12/08/25 13:16	VL120825
100-42-5	Styrene	0.16	0.68	U	1	0.68	2.13	12/08/25 13:16	VL120825
75-25-2	Bromoform	0.050	0.52	U	1	0.52	5.17	12/08/25 13:16	VL120825
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	1	0.14	0.21	12/08/25 13:16	VL120825
95-49-8	2-Chlorotoluene	0.17	0.88	U	1	0.88	2.59	12/08/25 13:16	VL120825

Report of Analysis

Client: RTP Environmental
Project: CHADWICK SQUARE
Client Sample ID: AA-1DUP
Lab Sample ID: Q3799-01DUP
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected:
Date Received:
SDG No.: Q3799
Matrix: Air
Test: TO-15

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 13:16	VL120825
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	1	0.88	2.46	12/08/25 13:16	VL120825
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	1	0.30	3.01	12/08/25 13:16	VL120825
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 13:16	VL120825
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	1	0.78	3.01	12/08/25 13:16	VL120825
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1	1.56	3.71	12/08/25 13:16	VL120825
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1	1.39	5.33	12/08/25 13:16	VL120825
106-99-0	1,3-Butadiene	0.050	0.11	U	1	0.11	1.11	12/08/25 13:16	VL120825
91-20-3	Naphthalene	0.010	0.050	U	1	0.050	0.52	12/08/25 13:16	VL120825
622-96-8	4-Ethyltoluene	0.21	1.03	U	1	1.03	2.46	12/08/25 13:16	VL120825
110-54-3	Hexane	1.40	4.93		1	0.56	1.76	12/08/25 13:16	VL120825
107-05-1	Allyl Chloride	0.11	0.34	U	1	0.34	1.57	12/08/25 13:16	VL120825
123-91-1	1,4-Dioxane	0.22	0.79	U	1	0.79	1.80	12/08/25 13:16	VL120825
80-62-6	Methyl Methacrylate	0.14	0.57	U	1	0.57	2.05	12/08/25 13:16	VL120825

SURROGATES

460-00-4 1-Bromo-4-Fluorobenzene 9.80 70 (65) - 130 (135) 98% SPK: 10

INTERNAL STANDARDS

Area Count
74-97-5 Bromochloromethane 155000
540-36-3 1,4-Difluorobenzene 398000
3114-55-4 Chlorobenzene-d5 319000

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043311.D
Acq On : 08 Dec 2025 13:16
Operator : SY/MD
Sample : Q3799-01DUP
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP

Quant Time: Dec 09 00:11:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.784	49	155357	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	397814	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	318793	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	229577	9.825	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.496	85	8883	0.405	ppbv	96
3) Chlorodifluoromethane	1.473	51	5370m	0.249	ppbv	
4) Chloromethane	1.531	50	17124	2.090	ppbv #	87
9) Propene	1.486	41	6430	0.703	ppbv #	1
11) Trichlorofluoromethane	1.875	101	4183	0.197	ppbv	90
13) Ethanol	1.719	45	12812	3.982	ppbv	99
15) Acetone	1.832	43	103724	4.806	ppbv	95
19) Isopropyl Alcohol	1.884	45	9039	0.741	ppbv #	1
20) Methylene Chloride	2.059	84	24089	3.258	ppbv	95
26) Hexane	2.823	57	29316	1.436	ppbv #	41
29) Chloroform	2.846	83	4392	0.151	ppbv	92
34) 2-Butanone	2.557	43	14218	0.506	ppbv	98
35) Carbon Tetrachloride	3.758	117	2821m	0.093	ppbv	
36) Benzene	3.648	78	6108	0.160	ppbv	94
53) Toluene	6.924	91	5234m	0.116	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043311.D
Acq On : 08 Dec 2025 13:16
Operator : SY/MD
Sample : Q3799-01DUP
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP

Quant Time: Dec 09 00:11:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/09/2025
Supervised By : Mahesh Dadoda 12/11/2025









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 811
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP

Tgt Ion: 49 Resp: 15535
Ion Ratio Lower Upper
49 100
128 49.7 23.0 69.0
130 61.6 29.9 89.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#2
Dichlorodifluoromethane
Concen: 0.405 ppbv
RT: 1.496 min Scan# 435
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16


Tgt Ion: 85 Resp: 8883
Ion Ratio Lower Upper
85 100
87 29.8 25.4 38.2





#3
Chlorodifluoromethane
Concen: 0.249 ppbv m
RT: 1.473 min Scan# 428
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP




Tgt Ion: 51 Resp: 5370
Ion Ratio Lower Upper
51 100
67 25.8 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#4
Chloromethane
Concen: 2.090 ppbv
RT: 1.531 min Scan# 446
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16



Tgt Ion: 50 Resp: 17124
Ion Ratio Lower Upper
50 100
52 37.1 24.1 36.1









#9
 Propene
 Concen: 0.703 ppbv
 RT: 1.486 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DUP

Tgt Ion: 41 Resp: 6430
 Ion Ratio Lower Upper
 41 100
 42 50.0 3.7 5.5




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



#11
 Trichlorofluoromethane
 Concen: 0.197 ppbv
 RT: 1.875 min Scan# 552
 Delta R.T. 0.000 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16

Tgt Ion:101 Resp: 4183
 Ion Ratio Lower Upper
 101 100
 103 57.7 52.6 79.0









#13
Ethanol
Concen: 3.982 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP

Tgt Ion: 45 Resp: 1281
Ion Ratio Lower Upper
45 100
46 38.2 15.0 59.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#15
Acetone
Concen: 4.806 ppbv
RT: 1.832 min Scan# 539
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16


Tgt Ion: 43 Resp: 103724
Ion Ratio Lower Upper
43 100
58 28.4 20.7 31.1





#19
Isopropyl Alcohol
Concen: 0.741 ppbv
RT: 1.884 min Scan# 511
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP



Tgt Ion: 45 Resp: 9039
Ion Ratio Lower Upper
45 100
58 326.3 31.3 46.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#20
Methylene Chloride
Concen: 3.258 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16


Tgt Ion: 84 Resp: 24089
Ion Ratio Lower Upper
84 100
49 157.1 132.9 199.3
51 45.8 34.8 52.2
86 61.9 48.2 72.2





#26
Hexane
Concen: 1.436 ppbv
RT: 2.823 min Scan# 844
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP



Tgt Ion: 57 Resp: 29310
Ion Ratio Lower Upper
57 100
41 101.3 74.6 111.8
43 75.4 179.0 268.4
56 61.6 42.3 63.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#29
Chloroform
Concen: 0.151 ppbv
RT: 2.846 min Scan# 852
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16


Tgt Ion: 83 Resp: 4392
Ion Ratio Lower Upper
83 100
85 57.2 50.5 75.7





#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.949 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP




Tgt Ion: 114 Resp: 397814
Ion Ratio Lower Upper
114 100
63 25.5 21.0 31.6
88 17.8 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#34
2-Butanone
Concen: 0.506 ppbv
RT: 2.557 min Scan# 763
Delta R.T. 0.006 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Tgt Ion: 43 Resp: 14218
Ion Ratio Lower Upper
43 100
72 21.7 18.1 27.1





#35
Carbon Tetrachloride
Concen: 0.093 ppbv m
RT: 3.758 min Scan# 1133
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP




Tgt Ion: 117 Resp: 282
Ion Ratio Lower Upper
117 100
119 72.9 74.6 112.0
121 18.1 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025




#36
Benzene
Concen: 0.160 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16



Tgt Ion: 78 Resp: 6108
Ion Ratio Lower Upper
78 100
77 19.3 17.5 26.3
50 16.5 15.5 23.3









#53
Toluene
Concen: 0.116 ppbv m
RT: 6.924 min Scan# 2111
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP

Tgt Ion: 91 Resp: 5234
Ion Ratio Lower Upper
91 100
92 69.6 47.4 71.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 12/09/2025
Supervised By :Mahesh Dadoda 12/11/2025



#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 2714
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16


Tgt Ion:117 Resp: 318793
Ion Ratio Lower Upper
117 100
82 65.7 56.1 84.1
119 32.2 25.0 37.6





#68
 1-Bromo-4-Fluorobenzene
 Concen: 9.825 ppbv
 RT: 10.367 min Scan# 3176
 Delta R.T. 0.000 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DUP





Tgt Ion: 95 Resp: 22957

Ion	Ratio	Lower	Upper
95	100		
174	68.8	50.2	75.2
175	5.2	4.1	6.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/09/2025
 Supervised By :Mahesh Dadoda 12/11/2025



Manual Integration Report

Sequence:	VL11825	Instrument	MSVOA_I
-----------	---------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC0.5	VL043203.D	1,1,1,2-Tetrachloroethane	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICC0.5	VL043203.D	1,1,2-Trichloroethane	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICC0.5	VL043203.D	1,4-Dioxane	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICC0.5	VL043203.D	cis-1,3-Dichloropropene	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICC0.5	VL043203.D	Ethyl Acetate	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICC0.5	VL043203.D	Heptane	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICC0.5	VL043203.D	m/p-Xylene	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICC0.5	VL043203.D	Methyl Methacrylate	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICC0.5	VL043203.D	Toluene	sam	11/19/2025 8:29:16 AM	MMDadoda	11/19/2025 9:42:12 AM	Peak Integrated by Software
VSTDICCC010	VL043204.D	m/p-Xylene	sam	11/19/2025 8:29:20 AM	MMDadoda	11/19/2025 9:42:15 AM	Peak Integrated by Software
VSTDICC002	VL043205.D	1,1,2-Trichloroethane	sam	11/19/2025 8:30:50 AM	MMDadoda	11/19/2025 9:42:20 AM	Peak Integrated by Software
VSTDICC002	VL043205.D	1,4-Dioxane	sam	11/19/2025 8:30:50 AM	MMDadoda	11/19/2025 9:42:20 AM	Peak Integrated by Software
VSTDICC002	VL043205.D	Benzene	sam	11/19/2025 8:30:50 AM	MMDadoda	11/19/2025 9:42:20 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	VL11825	Instrument	MSVOA_I
-----------	---------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC002	VL043205.D	Dibromochloromethane	sam	11/19/2025 8:30:50 AM	MMDadoda	11/19/2025 9:42:20 AM	Peak Integrated by Software
VSTDICC002	VL043205.D	Heptane	sam	11/19/2025 8:30:50 AM	MMDadoda	11/19/2025 9:42:20 AM	Peak Integrated by Software
VSTDICC002	VL043205.D	m/p-Xylene	sam	11/19/2025 8:30:50 AM	MMDadoda	11/19/2025 9:42:20 AM	Peak Integrated by Software
VSTDICC002	VL043205.D	t-1,3-Dichloropropene	sam	11/19/2025 8:30:50 AM	MMDadoda	11/19/2025 9:42:20 AM	Peak Integrated by Software
VSTDICC002	VL043205.D	Trichloroethene	sam	11/19/2025 8:30:50 AM	MMDadoda	11/19/2025 9:42:20 AM	Peak Integrated by Software
VSTDICC001	VL043206.D	1,4-Dioxane	sam	11/19/2025 8:31:27 AM	MMDadoda	11/19/2025 9:42:23 AM	Peak Integrated by Software
VSTDICC001	VL043206.D	2,2,4-Trimethylpentane	sam	11/19/2025 8:31:27 AM	MMDadoda	11/19/2025 9:42:23 AM	Peak Integrated by Software
VSTDICC001	VL043206.D	m/p-Xylene	sam	11/19/2025 8:31:27 AM	MMDadoda	11/19/2025 9:42:23 AM	Peak Integrated by Software
VSTDICC001	VL043206.D	t-1,3-Dichloropropene	sam	11/19/2025 8:31:27 AM	MMDadoda	11/19/2025 9:42:23 AM	Peak Integrated by Software
VSTDICC001	VL043206.D	Trichloroethene	sam	11/19/2025 8:31:27 AM	MMDadoda	11/19/2025 9:42:23 AM	Peak Integrated by Software
VSTDICC0.1	VL043208.D	1,1,1-Trichloroethane	sam	11/19/2025 8:31:38 AM	MMDadoda	11/19/2025 9:42:32 AM	Peak Integrated by Software
VSTDICC0.1	VL043208.D	1,1,2,2-Tetrachloroethane	sam	11/19/2025 8:31:38 AM	MMDadoda	11/19/2025 9:42:32 AM	Peak Integrated by Software
VSTDICC0.1	VL043208.D	1,2-Dibromoethane	sam	11/19/2025 8:31:38 AM	MMDadoda	11/19/2025 9:42:32 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	VL11825	Instrument	MSVOA_I
-----------	---------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC0.1	VL043208.D	Carbon Tetrachloride	sam	11/19/2025 8:31:38 AM	MMDadoda	11/19/2025 9:42:32 AM	Peak Integrated by Software
VSTDICC0.1	VL043208.D	Naphthalene	sam	11/19/2025 8:31:38 AM	MMDadoda	11/19/2025 9:42:32 AM	Peak Integrated by Software
VSTDICC0.1	VL043208.D	Tetrachloroethene	sam	11/19/2025 8:31:38 AM	MMDadoda	11/19/2025 9:42:32 AM	Peak Integrated by Software
VSTDICC0.1	VL043208.D	Trichloroethene	sam	11/19/2025 8:31:38 AM	MMDadoda	11/19/2025 9:42:32 AM	Peak Integrated by Software
VSTDICC0.03	VL043209.D	1,1,1-Trichloroethane	sam	11/19/2025 8:31:44 AM	MMDadoda	11/19/2025 9:42:35 AM	Peak Integrated by Software
VSTDICC0.03	VL043209.D	1,1,2,2-Tetrachloroethane	sam	11/19/2025 8:31:44 AM	MMDadoda	11/19/2025 9:42:35 AM	Peak Integrated by Software
VSTDICC0.03	VL043209.D	Carbon Tetrachloride	sam	11/19/2025 8:31:44 AM	MMDadoda	11/19/2025 9:42:35 AM	Peak Integrated by Software
VSTDICC0.03	VL043209.D	Tetrachloroethene	sam	11/19/2025 8:31:44 AM	MMDadoda	11/19/2025 9:42:35 AM	Peak Integrated by Software
VSTDICC0.03	VL043209.D	Trichloroethene	sam	11/19/2025 8:31:44 AM	MMDadoda	11/19/2025 9:42:35 AM	Peak Integrated by Software
VSTDICC0.03	VL043209.D	Vinyl Chloride	sam	11/19/2025 8:31:44 AM	MMDadoda	11/19/2025 9:42:35 AM	Peak Integrated by Software
VSTDICC015	VL043210.D	m/p-Xylene	sam	11/19/2025 8:32:26 AM	MMDadoda	11/19/2025 9:42:39 AM	Peak Integrated by Software
VSTDICC015	VL043210.D	Propene	sam	11/19/2025 8:32:26 AM	MMDadoda	11/19/2025 9:42:39 AM	Peak Integrated by Software
VSTDICV010	VL043211.D	m/p-Xylene	sam	11/19/2025 8:31:49 AM	MMDadoda	11/19/2025 9:43:04 AM	Peak Integrated by Software

Manual Integration Report

Sequence:

VL111825

Instrument

MSVOA_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICV010	VL043211.D	Propene	sam	11/19/2025 8:31:49 AM	MMDadoda	11/19/2025 9:43:04 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	VL120825	Instrument	MSVOA_I
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC010	VL043305.D	cis-1,3-Dichloropropene	sam	12/9/2025 7:58:25 AM	MMDadoda	12/11/2025 1:43:47 PM	Peak Integrated by Software
VSTDCCC010	VL043305.D	m/p-Xylene	sam	12/9/2025 7:58:25 AM	MMDadoda	12/11/2025 1:43:47 PM	Peak Integrated by Software
VL1208ABS01	VL043307.D	m/p-Xylene	sam	12/9/2025 7:58:29 AM	MMDadoda	12/11/2025 1:43:50 PM	Peak Integrated by Software
Q3799-01	VL043310.D	Benzene	sam	12/11/2025 12:30:22 PM	MMDadoda	12/11/2025 1:43:54 PM	Peak Integrated by Software
Q3799-01	VL043310.D	Carbon Tetrachloride	sam	12/11/2025 12:30:22 PM	MMDadoda	12/11/2025 1:43:54 PM	Peak Integrated by Software
Q3799-01	VL043310.D	Chlorodifluoromethane	sam	12/11/2025 12:30:22 PM	MMDadoda	12/11/2025 1:43:54 PM	Peak Integrated by Software
Q3799-01	VL043310.D	Chloroform	sam	12/11/2025 12:30:22 PM	MMDadoda	12/11/2025 1:43:54 PM	Peak Integrated by Software
Q3799-01	VL043310.D	Dichlorodifluoromethane	sam	12/11/2025 12:30:22 PM	MMDadoda	12/11/2025 1:43:54 PM	Peak Integrated by Software
Q3799-01	VL043310.D	Toluene	sam	12/11/2025 12:30:22 PM	MMDadoda	12/11/2025 1:43:54 PM	Peak Integrated by Software
Q3799-01DUP	VL043311.D	Carbon Tetrachloride	sam	12/9/2025 8:03:02 AM	MMDadoda	12/11/2025 1:43:58 PM	Peak Integrated by Software
Q3799-01DUP	VL043311.D	Chlorodifluoromethane	sam	12/9/2025 8:03:02 AM	MMDadoda	12/11/2025 1:43:58 PM	Peak Integrated by Software
Q3799-01DUP	VL043311.D	Toluene	sam	12/9/2025 8:03:02 AM	MMDadoda	12/11/2025 1:43:58 PM	Peak Integrated by Software
Q3799-02	VL043312.D	Carbon Tetrachloride	sam	12/9/2025 8:02:18 AM	MMDadoda	12/11/2025 1:44:01 PM	Peak Integrated by Software

Manual Integration Report

Sequence:	VL120825	Instrument	MSVOA_I
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3799-02	VL043312.D	Chlorodifluoromethane	sam	12/9/2025 8:02:18 AM	MMDadoda	12/11/2025 1:44:01 PM	Peak Integrated by Software
Q3799-02	VL043312.D	Chloroform	sam	12/9/2025 8:02:18 AM	MMDadoda	12/11/2025 1:44:01 PM	Peak Integrated by Software
Q3799-02	VL043312.D	Toluene	sam	12/9/2025 8:02:18 AM	MMDadoda	12/11/2025 1:44:01 PM	Peak Integrated by Software
Q3799-03	VL043313.D	Chlorodifluoromethane	sam	12/9/2025 8:03:09 AM	MMDadoda	12/11/2025 1:44:05 PM	Peak Integrated by Software
Q3799-03	VL043313.D	Heptane	sam	12/9/2025 8:03:09 AM	MMDadoda	12/11/2025 1:44:05 PM	Peak Integrated by Software
Q3799-03	VL043313.D	m/p-Xylene	sam	12/9/2025 8:03:09 AM	MMDadoda	12/11/2025 1:44:05 PM	Peak Integrated by Software
Q3799-04	VL043314.D	Chlorodifluoromethane	sam	12/9/2025 8:03:14 AM	MMDadoda	12/11/2025 1:44:08 PM	Peak Integrated by Software
Q3799-04	VL043314.D	m/p-Xylene	sam	12/9/2025 8:03:14 AM	MMDadoda	12/11/2025 1:44:08 PM	Peak Integrated by Software
Q3799-04	VL043314.D	Tetrachloroethene	sam	12/9/2025 8:03:14 AM	MMDadoda	12/11/2025 1:44:08 PM	Peak Integrated by Software
Q3799-04	VL043314.D	Toluene	sam	12/9/2025 8:03:14 AM	MMDadoda	12/11/2025 1:44:08 PM	Peak Integrated by Software
Q3799-05	VL043315.D	Carbon Tetrachloride	sam	12/9/2025 8:04:01 AM	MMDadoda	12/11/2025 1:44:12 PM	Peak Integrated by Software
Q3799-05	VL043315.D	Chlorodifluoromethane	sam	12/9/2025 8:04:01 AM	MMDadoda	12/11/2025 1:44:12 PM	Peak Integrated by Software
Q3799-05	VL043315.D	Tetrachloroethene	sam	12/9/2025 8:04:01 AM	MMDadoda	12/11/2025 1:44:12 PM	Peak Integrated by Software

Manual Integration Report

Sequence:	VL120825	Instrument	MSVOA_I
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3799-05	VL043315.D	Toluene	sam	12/9/2025 8:04:01 AM	MMDadoda	12/11/2025 1:44:12 PM	Peak Integrated by Software
Q3799-06	VL043316.D	4-Methyl-2-Pentanone	sam	12/9/2025 8:03:20 AM	MMDadoda	12/11/2025 1:44:20 PM	Peak Integrated by Software
Q3799-06	VL043316.D	Carbon Tetrachloride	sam	12/9/2025 8:03:20 AM	MMDadoda	12/11/2025 1:44:20 PM	Peak Integrated by Software
Q3799-06	VL043316.D	Chlorodifluoromethane	sam	12/9/2025 8:03:20 AM	MMDadoda	12/11/2025 1:44:20 PM	Peak Integrated by Software
Q3799-06	VL043316.D	Heptane	sam	12/9/2025 8:03:20 AM	MMDadoda	12/11/2025 1:44:20 PM	Peak Integrated by Software
Q3799-06	VL043316.D	Propene	sam	12/9/2025 8:03:20 AM	MMDadoda	12/11/2025 1:44:20 PM	Peak Integrated by Software
Q3799-06	VL043316.D	Styrene	sam	12/9/2025 8:03:20 AM	MMDadoda	12/11/2025 1:44:20 PM	Peak Integrated by Software
Q3799-06	VL043316.D	tert-Butyl alcohol	sam	12/9/2025 8:03:20 AM	MMDadoda	12/11/2025 1:44:20 PM	Peak Integrated by Software
Q3799-03DL2	VL043331.D	Ethanol	sam	12/11/2025 12:30:13 PM	MMDadoda	12/11/2025 1:45:04 PM	Peak Integrated by Software

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL111825

Review By	Semsettin Yesilyurt	Review On	11/19/2025 8:33:42 AM		
Supervise By	Maresh Dadoda	Supervise On	11/19/2025 9:44:32 AM		
SubDirectory	VL111825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2691			
Initial Calibration Stds		AP2694,AP2695,AP2696			
CCC		AP2694			
Internal Standard/PEM		AP2691			
ICV/I.BLK		AP2697			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL043202.D	18 Nov 2025 08:46	SY/MD	Ok
2	VSTDIC0.5	VL043203.D	18 Nov 2025 09:29	SY/MD	Ok,M
3	VSTDICCC010	VL043204.D	18 Nov 2025 10:02	SY/MD	Ok,M
4	VSTDIC002	VL043205.D	18 Nov 2025 10:37	SY/MD	Ok,M
5	VSTDIC001	VL043206.D	18 Nov 2025 11:10	SY/MD	Ok,M
6	VSTDIC0.5	VL043207.D	18 Nov 2025 11:43	SY/MD	Not Ok
7	VSTDIC0.1	VL043208.D	18 Nov 2025 12:16	SY/MD	Ok,M
8	VSTDIC0.03	VL043209.D	18 Nov 2025 12:49	SY/MD	Ok,M
9	VSTDIC015	VL043210.D	18 Nov 2025 13:24	SY/MD	Ok,M
10	VSTDICV010	VL043211.D	18 Nov 2025 13:56	SY/MD	Ok,M
11	VL1118ABL01	VL043212.D	18 Nov 2025 14:41	SY/MD	Ok
12	VL1118ABS01	VL043213.D	18 Nov 2025 15:24	SY/MD	Ok,M
13	Q3651-03	VL043214.D	18 Nov 2025 15:57	SY/MD	Dilution
14	Q3651-03DUP	VL043215.D	18 Nov 2025 16:31	SY/MD	Ok,M
15	Q3651-04	VL043216.D	18 Nov 2025 17:06	SY/MD	Dilution
16	Q3651-06	VL043217.D	18 Nov 2025 17:40	SY/MD	Dilution
17	Q3650-03	VL043218.D	18 Nov 2025 18:15	SY/MD	Ok,M
18	Q3650-04	VL043219.D	18 Nov 2025 18:49	SY/MD	Ok,M
19	Q3650-05	VL043220.D	18 Nov 2025 19:23	SY/MD	Ok,M
20	Q3650-06	VL043221.D	18 Nov 2025 19:58	SY/MD	Ok,M
21	Q3651-01	VL043222.D	18 Nov 2025 20:29	SY/MD	Ok,M

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QCBatch ID # VL111825

Review By	Semsettin Yesilyurt	Review On	11/19/2025 8:33:42 AM		
Supervise By	Mahesh Dadoda	Supervise On	11/19/2025 9:44:32 AM		
SubDirectory	VL111825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2691			
Initial Calibration Stds		AP2694,AP2695,AP2696			
CCC		AP2694			
Internal Standard/PEM		AP2691			
ICV/I.BLK		AP2697			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q3651-02	VL043223.D	18 Nov 2025 21:01	SY/MD	Ok,M
23	Q3651-05	VL043224.D	18 Nov 2025 21:33	SY/MD	Ok,M
24	Q3650-01	VL043225.D	18 Nov 2025 22:05	SY/MD	Ok,M
25	Q3650-02	VL043226.D	18 Nov 2025 22:37	SY/MD	Ok,M
26	Q3651-03DL	VL043227.D	18 Nov 2025 23:09	SY/MD	Ok,M
27	Q3651-03DL	VL043228.D	18 Nov 2025 23:41	SY/MD	Not Ok
28	Q3651-04DL	VL043229.D	19 Nov 2025 00:13	SY/MD	Ok,M
29	Q3651-06DL	VL043230.D	19 Nov 2025 00:44	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL120825

Review By	Semsettin Yesilyurt	Review On	12/9/2025 8:08:46 AM		
Supervise By	John Carlone	Supervise On	12/9/2025 1:27:26 PM		
SubDirectory	VL120825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2692			
Initial Calibration Stds		AP2694,AP2695,AP2696			
CCC		AP2694			
Internal Standard/PEM		AP2692			
ICV/I.BLK		AP2697			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL043304.D	08 Dec 2025 07:56	SY/MD	Ok
2	VSTDCCC010	VL043305.D	08 Dec 2025 08:44	SY/MD	Ok,M
3	VL1208ABL01	VL043306.D	08 Dec 2025 09:36	SY/MD	Ok
4	VL1208ABS01	VL043307.D	08 Dec 2025 10:24	SY/MD	Ok,M
5	QC120225 CAN 10297	VL043308.D	08 Dec 2025 11:11	SY/MD	Ok
6	QC120325 CAN 10606	VL043309.D	08 Dec 2025 11:44	SY/MD	Ok
7	Q3799-01	VL043310.D	08 Dec 2025 12:40	SY/MD	Ok,M
8	Q3799-01DUP	VL043311.D	08 Dec 2025 13:16	SY/MD	Ok,M
9	Q3799-02	VL043312.D	08 Dec 2025 13:52	SY/MD	Ok,M
10	Q3799-03	VL043313.D	08 Dec 2025 14:29	SY/MD	Dilution
11	Q3799-04	VL043314.D	08 Dec 2025 15:37	SY/MD	Ok,M
12	Q3799-05	VL043315.D	08 Dec 2025 16:25	SY/MD	Ok,M
13	Q3799-06	VL043316.D	08 Dec 2025 17:00	SY/MD	Dilution
14	Q3797-02	VL043317.D	08 Dec 2025 17:36	SY/MD	Ok,M
15	Q3797-04	VL043318.D	08 Dec 2025 18:11	SY/MD	Ok,M
16	Q3797-05	VL043319.D	08 Dec 2025 18:46	SY/MD	Ok,M
17	Q3798-02	VL043320.D	08 Dec 2025 19:21	SY/MD	Ok,M
18	Q3798-03	VL043321.D	08 Dec 2025 19:56	SY/MD	Ok,M
19	Q3799-03DL	VL043322.D	08 Dec 2025 20:50	SY/MD	Dilution
20	Q3799-03DL2	VL043323.D	08 Dec 2025 21:23	SY/MD	Not Ok
21	Q3799-06DL	VL043324.D	08 Dec 2025 21:56	SY/MD	Ok

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL120825

Review By	Semsettin Yesilyurt	Review On	12/9/2025 8:08:46 AM		
Supervise By	John Carlone	Supervise On	12/9/2025 1:27:26 PM		
SubDirectory	VL120825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME	STD REF.#				
Tune/Reschk	AP2692				
Initial Calibration Stds	AP2694,AP2695,AP2696				
CCC	AP2694				
Internal Standard/PEM	AP2692				
ICV/I.BLK	AP2697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q3797-01	VL043325.D	08 Dec 2025 22:28	SY/MD	Ok,M
23	Q3797-01DL	VL043326.D	08 Dec 2025 23:01	SY/MD	Not Ok
24	Q3797-03	VL043327.D	08 Dec 2025 23:34	SY/MD	Ok,M
25	Q3797-03DL	VL043328.D	09 Dec 2025 00:06	SY/MD	Not Ok
26	Q3798-01	VL043329.D	09 Dec 2025 00:38	SY/MD	Ok,M
27	Q3798-01DL	VL043330.D	09 Dec 2025 01:11	SY/MD	Not Ok
28	Q3799-03DL2	VL043331.D	09 Dec 2025 07:30	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL111825

Review By	Semsettin Yesilyurt	Review On	11/19/2025 8:33:42 AM		
Supervise By	Mahesh Dadoda	Supervise On	11/19/2025 9:44:32 AM		
SubDirectory	VL111825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2691
Initial Calibration Stds	AP2694,AP2695,AP2696
CCC	AP2694
Internal Standard/PEM	AP2691
ICV/I.BLK	AP2697
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL043202.D	18 Nov 2025 08:46		SY/MD	Ok
2	VSTDIC0.5	VSTDIC0.5	VL043203.D	18 Nov 2025 09:29		SY/MD	Ok,M
3	VSTDICCC010	VSTDICCC010	VL043204.D	18 Nov 2025 10:02	LR-20	SY/MD	Ok,M
4	VSTDIC002	VSTDIC002	VL043205.D	18 Nov 2025 10:37		SY/MD	Ok,M
5	VSTDIC001	VSTDIC001	VL043206.D	18 Nov 2025 11:10		SY/MD	Ok,M
6	VSTDIC0.5	VSTDIC0.5	VL043207.D	18 Nov 2025 11:43	Not used	SY/MD	Not Ok
7	VSTDIC0.1	VSTDIC0.1	VL043208.D	18 Nov 2025 12:16		SY/MD	Ok,M
8	VSTDIC0.03	VSTDIC0.03	VL043209.D	18 Nov 2025 12:49		SY/MD	Ok,M
9	VSTDIC015	VSTDIC015	VL043210.D	18 Nov 2025 13:24		SY/MD	Ok,M
10	VSTDICV010	ICVVL111825	VL043211.D	18 Nov 2025 13:56	ICV failed for com.#66	SY/MD	Ok,M
11	VL1118ABL01	VL1118ABL01	VL043212.D	18 Nov 2025 14:41		SY/MD	Ok
12	VL1118ABS01	VL1118ABS01	VL043213.D	18 Nov 2025 15:24		SY/MD	Ok,M
13	Q3651-03	IA1	VL043214.D	18 Nov 2025 15:57	Need 20X	SY/MD	Dilution
14	Q3651-03DUP	IA1DUP	VL043215.D	18 Nov 2025 16:31		SY/MD	Ok,M
15	Q3651-04	IA2	VL043216.D	18 Nov 2025 17:06	Need 10X	SY/MD	Dilution
16	Q3651-06	IA3	VL043217.D	18 Nov 2025 17:40	Need 10X	SY/MD	Dilution
17	Q3650-03	IA1	VL043218.D	18 Nov 2025 18:15		SY/MD	Ok,M
18	Q3650-04	IA2	VL043219.D	18 Nov 2025 18:49		SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QCBatch ID # VL111825

Review By	Semsettin Yesilyurt	Review On	11/19/2025 8:33:42 AM		
Supervise By	Mahesh Dadoda	Supervise On	11/19/2025 9:44:32 AM		
SubDirectory	VL111825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2691			
Initial Calibration Stds		AP2694,AP2695,AP2696			
CCC		AP2694			
Internal Standard/PEM		AP2691			
ICV/I.BLK		AP2697			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3650-05	IA3	VL043220.D	18 Nov 2025 19:23		SY/MD	Ok,M
20	Q3650-06	OA1	VL043221.D	18 Nov 2025 19:58		SY/MD	Ok,M
21	Q3651-01	SV1	VL043222.D	18 Nov 2025 20:29		SY/MD	Ok,M
22	Q3651-02	SV2	VL043223.D	18 Nov 2025 21:01		SY/MD	Ok,M
23	Q3651-05	SV3	VL043224.D	18 Nov 2025 21:33		SY/MD	Ok,M
24	Q3650-01	SV1	VL043225.D	18 Nov 2025 22:05		SY/MD	Ok,M
25	Q3650-02	SV2	VL043226.D	18 Nov 2025 22:37		SY/MD	Ok,M
26	Q3651-03DL	IA1DL	VL043227.D	18 Nov 2025 23:09		SY/MD	Ok,M
27	Q3651-03DL	IA1DL	VL043228.D	18 Nov 2025 23:41	Not Required	SY/MD	Not Ok
28	Q3651-04DL	IA2DL	VL043229.D	19 Nov 2025 00:13		SY/MD	Ok,M
29	Q3651-06DL	IA3DL	VL043230.D	19 Nov 2025 00:44		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL120825

Review By	Semsettin Yesilyurt	Review On	12/9/2025 8:08:46 AM		
Supervise By	John Carlone	Supervise On	12/9/2025 1:27:26 PM		
SubDirectory	VL120825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2692
Initial Calibration Stds	AP2694,AP2695,AP2696
CCC	AP2694
Internal Standard/PEM	AP2692
ICV/I.BLK	AP2697
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL043304.D	08 Dec 2025 07:56		SY/MD	Ok
2	VSTDCCC010	VSTDCCC010	VL043305.D	08 Dec 2025 08:44		SY/MD	Ok,M
3	VL1208ABL01	VL1208ABL01	VL043306.D	08 Dec 2025 09:36		SY/MD	Ok
4	VL1208ABS01	VL1208ABS01	VL043307.D	08 Dec 2025 10:24		SY/MD	Ok,M
5	QC120225 CAN 10297	QC120225 CAN 10297	VL043308.D	08 Dec 2025 11:11		SY/MD	Ok
6	QC120325 CAN 10606	QC120325 CAN 10606	VL043309.D	08 Dec 2025 11:44		SY/MD	Ok
7	Q3799-01	AA-1	VL043310.D	08 Dec 2025 12:40		SY/MD	Ok,M
8	Q3799-01DUP	AA-1DUP	VL043311.D	08 Dec 2025 13:16		SY/MD	Ok,M
9	Q3799-02	9B-IA-1	VL043312.D	08 Dec 2025 13:52		SY/MD	Ok,M
10	Q3799-03	10-IA-1	VL043313.D	08 Dec 2025 14:29	Need 10X	SY/MD	Dilution
11	Q3799-04	11-IA-1	VL043314.D	08 Dec 2025 15:37		SY/MD	Ok,M
12	Q3799-05	13-IA-1	VL043315.D	08 Dec 2025 16:25		SY/MD	Ok,M
13	Q3799-06	16-IA-1	VL043316.D	08 Dec 2025 17:00	Need 10X	SY/MD	Dilution
14	Q3797-02	IA1	VL043317.D	08 Dec 2025 17:36		SY/MD	Ok,M
15	Q3797-04	IA2	VL043318.D	08 Dec 2025 18:11		SY/MD	Ok,M
16	Q3797-05	OA1	VL043319.D	08 Dec 2025 18:46		SY/MD	Ok,M
17	Q3798-02	IA1	VL043320.D	08 Dec 2025 19:21		SY/MD	Ok,M
18	Q3798-03	OA1	VL043321.D	08 Dec 2025 19:56		SY/MD	Ok,M

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL120825

Review By	Semsettin Yesilyurt	Review On	12/9/2025 8:08:46 AM		
Supervise By	John Carlone	Supervise On	12/9/2025 1:27:26 PM		
SubDirectory	VL120825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2692			
Initial Calibration Stds		AP2694,AP2695,AP2696			
CCC		AP2694			
Internal Standard/PEM		AP2692			
ICV/I.BLK		AP2697			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3799-03DL	10-IA-1DL	VL043322.D	08 Dec 2025 20:50	Need 40X;Not Match exactly please confirm	SY/MD	Dilution
20	Q3799-03DL2	10-IA-1DL2	VL043323.D	08 Dec 2025 21:23	Need further dilution, rerun 400X	SY/MD	Not Ok
21	Q3799-06DL	16-IA-1DL	VL043324.D	08 Dec 2025 21:56		SY/MD	Ok
22	Q3797-01	SV1	VL043325.D	08 Dec 2025 22:28		SY/MD	Ok,M
23	Q3797-01DL	SV1DL	VL043326.D	08 Dec 2025 23:01	Not Required	SY/MD	Not Ok
24	Q3797-03	SV2	VL043327.D	08 Dec 2025 23:34		SY/MD	Ok,M
25	Q3797-03DL	SV2DL	VL043328.D	09 Dec 2025 00:06	Not Required	SY/MD	Not Ok
26	Q3798-01	SV1	VL043329.D	09 Dec 2025 00:38		SY/MD	Ok,M
27	Q3798-01DL	SV1DL	VL043330.D	09 Dec 2025 01:11	Not Required	SY/MD	Not Ok
28	Q3799-03DL2	10-IA-1DL2	VL043331.D	09 Dec 2025 07:30	Not match exactly please check	SY/MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q3799

Test : TO-15

Prepbatch ID :

Sequence ID/Qc Batch ID: VL120825,

Standard ID :

AP2692,AP2694,AP2695,AP2696,AP2697,

Chemical ID :

A1117,A1141,A1142,A1143,

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Air STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
47	Internal Standard/Surrogate Mix-80 ppbv	AP2692	11/18/2025	12/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
11/25/2025								

FROM 2.40000psi of A1141 + 27.60000psi of A1117 = Final Quantity: 30.000 psi

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2396	TO-15 15 PPBV CAL MIX	AP2694	11/18/2025	12/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
11/25/2025								

FROM 1455.00000SCCM of A1117 + 45.00000SCCM of A1142 = Final Quantity: 30.000 psi

Air STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2397	TO-15 -2 PPBV CAL.MIX	AP2695	11/18/2025	12/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
11/25/2025								

FROM 26.00000psi of A1117 + 4.00000psi of AP2694 = Final Quantity: 30.000 psi

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2668	0.5 PPBV CAL.MIX	AP2696	11/18/2025	12/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
11/25/2025								

FROM 29.00000psi of A1117 + 1.00000psi of AP2694 = Final Quantity: 30.000 psi

Air STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2396	TO-15 15 PPBV CAL MIX	AP2697	11/18/2025	12/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
FROM 1455.00000SCCM of A1117 + 45.00000SCCM of A1143 = Final Quantity: 30.000 psi								

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

CHEMICAL RECEIPT LOG BOOK


Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
AIR LIQUIDE	365A-49 / AIR, Compressed	90402401186-01	04/01/2026	04/01/2022 / apatel	04/01/2022 / SAM	A1117

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CUSTOMGAS Solutions	TB500009-110 / TO-15 Internal Standard/Surrogate Standard	BC884004	10/01/2026	10/14/2025 / sam	10/13/2025 / sam	A1141

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CUSTOMGAS Solutions	TO15-80-6R-07092 / TO-15 Modified (80 comp) in Nitrogen (addition of 2-methylnaphthalene)	051925-001A	10/01/2026	10/13/2025 / sam	10/13/2025 / sam	A1142

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CUSTOMGAS Solutions	TO15-80-6R-07092 / TO-15 Modified (80 comp) in Nitrogen (addition of 2-methylnaphthalene)	051925-001B	10/01/2026	10/13/2025 / sam	10/13/2025 / sam	A1143

CUSTOMGAS SOLUTIONS



1750 East Club Boulevard

Durham, NC 27704

Phone: (919) 220-2570

Fax: (919) 220-4540

A 1141 ~~1141~~
A 1141

Certificate of Analysis

Customer:

Chem Tech
284 Sheffield Street
Mountainside, NJ 07092

Tel: (908) 789-8900

Cylinder Number: BC884004
Cylinder Size/CGA: 170/180SS
Fill Pressure: 1815 PSIA
Gas Volume: ~170 liters
Date of Mfg: 10/01/2025
Expiration Date: 10/01/2026

Ship To : Chem Tech
284 Sheffield Street
Mountainside, NJ 07092

Customer Number	Ship VIA	Job No.	Customer PO	Mixture Type
00107092NJ	Best Way	051925-001RRR	Email051925	Gravimetric

Component	Nominal Concentration	Actual Concentration*	Mixture Type
Bromochloromethane	1 ppm	1.019 ppm +/- 0.02 ppm	Gravimetric Master Gas
4-Bromofluorobenzene	1 ppm	1.013 ppm +/- 0.02 ppm	
Chlorobenzene-D5	1 ppm	0.993 ppm +/- 0.02 ppm	
1,4-Difluorobenzene	1 ppm	0.981 ppm +/- 0.02 ppm	
Nitrogen	balance	balance	

NOTES: Blend Tolerance: +/- 10 %

Analytical Tolerance: +/- 5 %

Traceability: NIST by weight set. NIST Traceability No MT001810.
Internal Standards by analysis

Reactive Mixtures: Analyzed twice with required agreement between analyses of 2%.
Required wait time between analyses of >7 days.

Caution: Do not use below 150 PSIG.

Analyst Name: Joseph A. Ernst

QA Signature:

Signature:

Date: 10/01/2025

*Every effort has been made to establish the actual concentration of the components using master gas blending technology however, Custom Gas Solutions shall have no liability in excess of the established charge for this material.

CUSTOMGAS SOLUTIONS



1750 East Club Boulevard
Durham, NC 27704
Phone: (919) 220-2570
Fax: (919) 220-4540

A1142
Rec 10/13/25

Certificate of Analysis

Customer:

ChemTech
284 Sheffield Street
Mountianside, NJ 07092

Cylinder Number: BC993219
Cylinder Size/CGA: 170/180
Fill Pressure: 1815 PSIA
Gas Volume: 110 liters
Date of Mfg: 10/01/2025
Expiration Date: 10/01/2026
Lot Number: 051925-001A

Ship To : Chemtech
284 Sheffield Street
Mountainside, NJ 07092

Customer Number	Ship VIA	Job No.	Customer PO	Mixture Type
00107092NJ	Best Way	051925-001	Email051925	Gravimetric

Component	Nominal Concentration	Actual Concentration*	Mixture Type
Acetone	500 ppb	508 ppb +/- 50 ppb	Gravimetric Master Gas
Acetonitrile	500 ppb	515 ppb +/- 50 ppb	
Acrolein	500 ppb	514 ppb +/- 50 ppb	
Acrylonitrile	500 ppb	519 ppb +/- 50 ppb	
Allyl chloride	500 ppb	499 ppb +/- 50 ppb	
Benzene	500 ppb	475 ppb +/- 50 ppb	
Benzyl Chloride	500 ppb	480 ppb +/- 50 ppb	
Bromodichloromethane	500 ppb	500 ppb +/- 50 ppb	
Bromoform	500 ppb	486 ppb +/- 50 ppb	
1,3-Butadiene	500 ppb	495 ppb +/- 50 ppb	
tert-Butyl alcohol	500 ppb	512 ppb +/- 50 ppb	
n-Butyl benzene	500 ppb	510 ppb +/- 50 ppb	
sec-Butyl benzene	500 ppb	510 ppb +/- 50 ppb	
tert-Butyl benzene	500 ppb	529 ppb +/- 50 ppb	
Carbon disulfide	500 ppb	476 ppb +/- 50 ppb	
Carbon tetrachloride	500 ppb	497 ppb +/- 50 ppb	
Chlorobenzene	500 ppb	483 ppb +/- 50 ppb	
Chlorodibromomethane	500 ppb	488 ppb +/- 50 ppb	
Chloroform	500 ppb	483 ppb +/- 50 ppb	
2-Chlorotoluene	500 ppb	512 ppb +/- 50 ppb	
Cyclohexane	500 ppb	474 ppb +/- 50 ppb	
1,2-Dibromoethane	500 ppb	482 ppb +/- 50 ppb	
1,2-Dichlorobenzene	500 ppb	500 ppb +/- 50 ppb	
1,3-Dichlorobenzene	500 ppb	480 ppb +/- 50 ppb	
1,4-Dichlorobenzene	500 ppb	481 ppb +/- 50 ppb	
Dichlorodifluoromethane (R12)	500 ppb	499 ppb +/- 50 ppb	

1,1-Dichloroethane	500 ppb	483 ppb +/- 50 ppb	
1,2-Dichloroethane	500 ppb	488 ppb +/- 50 ppb	
1,1-Dichloroethylene	500 ppb	484 ppb +/- 50 ppb	
cis 1,2-Dichloroethylene	500 ppb	480 ppb +/- 50 ppb	
trans 1,2-Dichloroethylene	500 ppb	479 ppb +/- 50 ppb	
1,2-Dichloropropane	500 ppb	481 ppb +/- 50 ppb	
cis 1,3-Dichloropropylene	500 ppb	507 ppb +/- 50 ppb	
trans 1,3-Dichloropropylene	500 ppb	457 ppb +/- 50 ppb	
1,2-Dichlorotetrafluoroethane	500 ppb	496 ppb +/- 50 ppb	
1,4-Dioxane	500 ppb	475 ppb +/- 50 ppb	
Ethyl acetate	500 ppb	477 ppb +/- 50 ppb	
Ethyl Alcohol	500 ppb	513 ppb +/- 50 ppb	
Ethyl benzene	500 ppb	488 ppb +/- 50 ppb	
Ethyl Chloride	500 ppb	497 ppb +/- 50 ppb	
4-Ethyltoluene	500 ppb	477 ppb +/- 50 ppb	
n-Heptane	500 ppb	478 ppb +/- 50 ppb	
Hexachloro-1,3-butadiene	500 ppb	480 ppb +/- 50 ppb	
2-Hexanone	500 ppb	482 ppb +/- 50 ppb	
n-Hexane	500 ppb	476 ppb +/- 50 ppb	
Isopropyl alcohol	500 ppb	502 ppb +/- 50 ppb	
Isopropyl benzene	500 ppb	507 ppb +/- 50 ppb	
p-Isopropyl toluene	500 ppb	515 ppb +/- 50 ppb	
Methyl Bromide	500 ppb	496 ppb +/- 50 ppb	
Methyl Chloride	500 ppb	499 ppb +/- 50 ppb	
Methyl ethyl ketone	500 ppb	487 ppb +/- 50 ppb	
Methyl isobutyl ketone	500 ppb	484 ppb +/- 50 ppb	
Methyl methacrylate	500 ppb	512 ppb +/- 50 ppb	
Methyl tertiary butyl ether	500 ppb	474 ppb +/- 50 ppb	
Methylene chloride	500 ppb	489 ppb +/- 50 ppb	
Naphthalene	500 ppb	506 ppb +/- 50 ppb	
n-Propylbenzene	500 ppb	510 ppb +/- 50 ppb	
Propylene	500 ppb	499 ppb +/- 50 ppb	
Styrene	500 ppb	477 ppb +/- 50 ppb	
1,1,1,2-Tetrachloroethane	500 ppb	515 ppb +/- 50 ppb	
1,1,2,2-Tetrachloroethane	500 ppb	475 ppb +/- 50 ppb	
Tetrachloroethylene	500 ppb	478 ppb +/- 50 ppb	
Tetrahydrofuran	500 ppb	504 ppb +/- 50 ppb	
Toluene	500 ppb	482 ppb +/- 50 ppb	
1,2,4-Trichlorobenzene	500 ppb	485 ppb +/- 50 ppb	
1,1,1-Trichloroethane	500 ppb	485 ppb +/- 50 ppb	
1,1,2-Trichloroethane	500 ppb	485 ppb +/- 50 ppb	
Trichloroethylene	500 ppb	479 ppb +/- 50 ppb	
Trichlorofluoromethane	500 ppb	499 ppb +/- 50 ppb	
1,1,2-Trichlorotrifluoroethane	500 ppb	486 ppb +/- 50 ppb	
1,2,4-Trimethylbenzene	500 ppb	485 ppb +/- 50 ppb	
1,3,5-Trimethylbenzene	500 ppb	489 ppb +/- 50 ppb	
2,2,4-Trimethylpentane	500 ppb	480 ppb +/- 50 ppb	
Vinyl acetate	500 ppb	490 ppb +/- 50 ppb	
Vinyl bromide	500 ppb	496 ppb +/- 50 ppb	
Vinyl chloride	500 ppb	507 ppb +/- 50 ppb	
m-Xylene	500 ppb	484 ppb +/- 50 ppb	
o-Xylene	500 ppb	471 ppb +/- 50 ppb	
p-Xylene	500 ppb	477 ppb +/- 50 ppb	
2-Methyl Naphthalene	500 ppb	498 ppb +/- 50 ppb	
R-22	500 ppb	590 ppb +/- 50 ppb	
Nitrogen	balance	balance	


NOTES: Blend Tolerance: +/- 10 %
Analytical Tolerance: +/- 10 %
Traceability: NIST by weight set. NIST Traceability No MT001810.
Internal Standards by analysis
Reactive Mixtures: Analyzed twice with required agreement between analyses of 2%.
Required wait time between analyses of >7 days.
Caution: Do not use below 150 PSIG.

Analyst Name: Joseph A. Ernst

QA Signature:

Signature:

Date: 10/01/2025



*Every effort has been made to establish the actual concentration of the components using master gas blending technology however, Custom Gas Solutions shall have no liability in excess of the established charge for this material.

CUSTOMGAS SOLUTIONS



1750 East Club Boulevard
Durham, NC 27704
Phone: (919) 220-2570
Fax: (919) 220-4540

A 1143
Rec 10/13/25

Certificate of Analysis

Customer:

ChemTech
284 Sheffield Street
Mountianside, NJ 07092

Cylinder Number: BC785565
Cylinder Size/CGA: 170/180
Fill Pressure: 1815 PSIA
Gas Volume: 110 liters
Date of Mfg: 10/01/2025
Expiration Date: 10/01/2026
Lot Number: 051925-001B

Ship To : Chemtech
284 Sheffield Street
Mountainside, NJ 07092

Customer Number	Ship VIA	Job No.	Customer PO	Mixture Type
00107092NJ	Best Way	051925-001	Email051925	Gravimetric

Component	Nominal Concentration	Actual Concentration*	Mixture Type
Acetone	500 ppb	525 ppb +/- 50 ppb	Gravimetric Master Gas
Acetonitrile	500 ppb	525 ppb +/- 50 ppb	
Acrolein	500 ppb	534 ppb +/- 50 ppb	
Acrylonitrile	500 ppb	530 ppb +/- 50 ppb	
Allyl chloride	500 ppb	517 ppb +/- 50 ppb	
Benzene	500 ppb	492 ppb +/- 50 ppb	
Benzyl Chloride	500 ppb	497 ppb +/- 50 ppb	
Bromodichloromethane	500 ppb	518 ppb +/- 50 ppb	
Bromoform	500 ppb	503 ppb +/- 50 ppb	
1,3-Butadiene	500 ppb	512 ppb +/- 50 ppb	
tert-Butyl alcohol	500 ppb	523 ppb +/- 50 ppb	
n-Butyl benzene	500 ppb	510 ppb +/- 50 ppb	
sec-Butyl benzene	500 ppb	520 ppb +/- 50 ppb	
tert-Butyl benzene	500 ppb	520 ppb +/- 50 ppb	
Carbon disulfide	500 ppb	493 ppb +/- 50 ppb	
Carbon tetrachloride	500 ppb	515 ppb +/- 50 ppb	
Chlorobenzene	500 ppb	500 ppb +/- 50 ppb	
Chlorodibromomethane	500 ppb	496 ppb +/- 50 ppb	
Chloroform	500 ppb	500 ppb +/- 50 ppb	
2-Chlorotoluene	500 ppb	523 ppb +/- 50 ppb	
Cyclohexane	500 ppb	491 ppb +/- 50 ppb	
1,2-Dibromoethane	500 ppb	499 ppb +/- 50 ppb	
1,2-Dichlorobenzene	500 ppb	518 ppb +/- 50 ppb	
1,3-Dichlorobenzene	500 ppb	497 ppb +/- 50 ppb	
1,4-Dichlorobenzene	500 ppb	498 ppb +/- 50 ppb	
Dichlorodifluoromethane (R12)	500 ppb	516 ppb +/- 50 ppb	

1,1-Dichloroethane	500 ppb	500 ppb +/- 50 ppb	
1,2-Dichloroethane	500 ppb	505 ppb +/- 50 ppb	
1,1-Dichloroethylene	500 ppb	501 ppb +/- 50 ppb	
cis 1,2-Dichloroethylene	500 ppb	497 ppb +/- 50 ppb	
trans 1,2-Dichloroethylene	500 ppb	496 ppb +/- 50 ppb	
1,2-Dichloropropane	500 ppb	498 ppb +/- 50 ppb	
cis 1,3-Dichloropropylene	500 ppb	525 ppb +/- 50 ppb	
trans 1,3-Dichloropropylene	500 ppb	473 ppb +/- 50 ppb	
1,2-Dichlorotetrafluoroethane	500 ppb	513 ppb +/- 50 ppb	
1,4-Dioxane	500 ppb	492 ppb +/- 50 ppb	
Ethyl acetate	500 ppb	494 ppb +/- 50 ppb	
Ethyl Alcohol	500 ppb	523 ppb +/- 50 ppb	
Ethyl benzene	500 ppb	505 ppb +/- 50 ppb	
Ethyl Chloride	500 ppb	515 ppb +/- 50 ppb	
4-Ethyltoluene	500 ppb	494 ppb +/- 50 ppb	
n-Heptane	500 ppb	495 ppb +/- 50 ppb	
Hexachloro-1,3-butadiene	500 ppb	497 ppb +/- 50 ppb	
2-Hexanone	500 ppb	499 ppb +/- 50 ppb	
n-Hexane	500 ppb	493 ppb +/- 50 ppb	
Isopropyl alcohol	500 ppb	520 ppb +/- 50 ppb	
Isopropyl benzene	500 ppb	518 ppb +/- 50 ppb	
p-Isopropyl toluene	500 ppb	524 ppb +/- 50 ppb	
Methyl Bromide	500 ppb	514 ppb +/- 50 ppb	
Methyl Chloride	500 ppb	517 ppb +/- 50 ppb	
Methyl ethyl ketone	500 ppb	504 ppb +/- 50 ppb	
Methyl isobutyl ketone	500 ppb	501 ppb +/- 50 ppb	
Methyl methacrylate	500 ppb	523 ppb +/- 50 ppb	
Methyl tertiary butyl ether	500 ppb	491 ppb +/- 50 ppb	
Methylene chloride	500 ppb	506 ppb +/- 50 ppb	
Naphthalene	500 ppb	517 ppb +/- 50 ppb	
n-Propylbenzene	500 ppb	519 ppb +/- 50 ppb	
Propylene	500 ppb	516 ppb +/- 50 ppb	
Styrene	500 ppb	494 ppb +/- 50 ppb	
1,1,1,2-Tetrachloroethane	500 ppb	525 ppb +/- 50 ppb	
1,1,2,2-Tetrachloroethane	500 ppb	492 ppb +/- 50 ppb	
Tetrachloroethylene	500 ppb	495 ppb +/- 50 ppb	
Tetrahydrofuran	500 ppb	510 ppb +/- 50 ppb	
Toluene	500 ppb	499 ppb +/- 50 ppb	
1,2,4-Trichlorobenzene	500 ppb	502 ppb +/- 50 ppb	
1,1,1-Trichloroethane	500 ppb	502 ppb +/- 50 ppb	
1,1,2-Trichloroethane	500 ppb	503 ppb +/- 50 ppb	
Trichloroethylene	500 ppb	496 ppb +/- 50 ppb	
Trichlorofluoromethane	500 ppb	517 ppb +/- 50 ppb	
1,1,2-Trichlorotrifluoroethane	500 ppb	503 ppb +/- 50 ppb	
1,2,4-Trimethylbenzene	500 ppb	502 ppb +/- 50 ppb	
1,3,5-Trimethylbenzene	500 ppb	506 ppb +/- 50 ppb	
2,2,4-Trimethylpentane	500 ppb	497 ppb +/- 50 ppb	
Vinyl acetate	500 ppb	507 ppb +/- 50 ppb	
Vinyl bromide	500 ppb	514 ppb +/- 50 ppb	
Vinyl chloride	500 ppb	515 ppb +/- 50 ppb	
m-Xylene	500 ppb	501 ppb +/- 50 ppb	
o-Xylene	500 ppb	488 ppb +/- 50 ppb	
p-Xylene	500 ppb	494 ppb +/- 50 ppb	
2-Methyl Naphthalene	500 ppb	498 ppb +/- 50 ppb	
R-22	500 ppb	590 ppb +/- 50 ppb	
Nitrogen	balance	balance	


NOTES: Blend Tolerance: +/- 10 %
Analytical Tolerance: +/- 10 %
Traceability: NIST by weight set. NIST Traceability No MT001810.
Internal Standards by analysis
Reactive Mixtures: Analyzed twice with required agreement between analyses of 2%.
Required wait time between analyses of >7 days.
Caution: Do not use below 150 PSIG.

Analyst Name: Joseph A. Ernst

QA Signature:

Signature:

Date: 10/01/2025



*Every effort has been made to establish the actual concentration of the components using master gas blending technology however, Custom Gas Solutions shall have no liability in excess of the established charge for this material.

Order ID : Q3799

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
12/08/2025	B2512003	Q3799-01	-3.9	10280	6 L	10520		C2511003	VL043290.D Seq :VL112425 TUNE : VL043277.D ICAL : VL043209.D CCAL : VL043278.D MB : VL043279.D
12/08/2025	B2512003	Q3799-02	-5.5	10608	6 L	10771		C2511003	VL043290.D Seq :VL112425 TUNE : VL043277.D ICAL : VL043209.D CCAL : VL043278.D MB : VL043279.D
12/08/2025	B2512003	Q3799-05	-14.1	10320	6 L	10626		C2511003	VL043290.D Seq :VL112425 TUNE : VL043277.D ICAL : VL043209.D CCAL : VL043278.D MB : VL043279.D
12/08/2025	B2512003	Q3799-04	-6.3	10315	6 L	10240		C2511003	VL043290.D Seq :VL112425 TUNE : VL043277.D ICAL : VL043209.D CCAL : VL043278.D MB : VL043279.D
12/08/2025	B2512003	Q3799-03	-7.5	10283	6 L	10207		C2511002	VL043259.D Seq :VL112025 TUNE : VL043256.D ICAL : VL043209.D CCAL : VL043257.D MB : VL043258.D
12/08/2025	B2512003	Q3799-06	-5.9	10266	6 L	10482		C2511002	VL043259.D Seq :VL112025 TUNE : VL043256.D ICAL : VL043209.D CCAL : VL043257.D MB : VL043258.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2511003

Type of Canister : 6 L

Cleaning Date : 11/24/2025

Analyst : sam

Oven Temperature : 125.0 c

Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10315	-30	11/25/2025	11/25/2025	-30	12/26/2025	Q3799-04	10315	VL043290.D
10590	-30	11/25/2025	11/25/2025	-30	12/26/2025		10315	VL043290.D
10320	-30	11/25/2025	11/25/2025	-30	12/26/2025	Q3799-05	10315	VL043290.D
10595	-30	11/25/2025	11/25/2025	-30	12/26/2025		10315	VL043290.D
10321	-30	11/25/2025	11/25/2025	-30	12/26/2025	Q3836-02	10315	VL043290.D
10608	-30	11/25/2025	11/25/2025	-30	12/26/2025	Q3799-02	10315	VL043290.D
10281	-30	11/25/2025	11/25/2025	-30	12/26/2025		10315	VL043290.D
10280	-30	11/25/2025	11/25/2025	-30	12/26/2025	Q3799-01	10315	VL043290.D



Canister Cleaning

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Canister Batch : C2511002

Type of Canister : 6 L

Cleaning Date : 11/17/2025

Analyst : sam

Oven Temperature : 125.0 c



Canister	Vaccum After Cleaning ("Hg)	Vaccum After Cleaning Date	Analysis Date	VAC 24Hr ("Hg)	Vaccum After 24hr Date	New Sample Number	QC Canister	File
10588	-30	11/18/2025	11/20/2025	-30	11/19/2025	Q3798-02	10588	VL043259.D
10296	-30	11/18/2025	11/20/2025	-30	11/19/2025	Q3797-03	10588	VL043259.D
10266	-30	11/18/2025	11/20/2025	-30	11/19/2025	Q3799-06	10588	VL043259.D
10060	-30	11/18/2025	11/20/2025	-30	11/19/2025	Q3797-02	10588	VL043259.D
10283	-30	11/18/2025	11/20/2025	-30	11/19/2025	Q3799-03	10588	VL043259.D
10300	-30	11/18/2025	11/20/2025	-30	11/19/2025	Q3797-05	10588	VL043259.D
10153	-30	11/18/2025	11/20/2025	-30	11/19/2025	Q3797-01	10588	VL043259.D
10492	-30	11/18/2025	11/20/2025	-30	11/19/2025	Q3798-01	10588	VL043259.D

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043202.D
Acq On : 18 Nov 2025 08:46
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Wed Nov 19 01:11:25 2025



AutoFind: Scans 3175, 3176, 3177; Background Corrected with Scan 3156

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	22.1	30285	PASS
75	95	30	66	55.6	76296	PASS
95	95	100	100	100.0	137266	PASS
96	95	5	9	6.7	9227	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	63.6	87269	PASS
175	174	4	9	7.8	6847	PASS
176	174	93	101	95.5	83301	PASS
177	176	5	9	6.6	5509	PASS

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL111825AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Wed Nov 19 01:11:25 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL043209.D 0.1 =VL043208.D 0.5 =VL043203.D 1 =VL043206.D 2 =VL043205.D 10 =VL043204.D 15 =VL043210.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.398	1.505	1.476	1.340	1.334	1.410	5.53
3) Chlorodifluoro...			1.545	1.404	1.437	1.317	1.231	1.387	8.60
4) Chloromethane			0.521	0.559	0.568	0.498	0.489	0.527	6.72
5) T Vinyl Chloride	0.724	0.655	0.533	0.580	0.575	0.486	0.483	0.577	15.33
6) T Bromomethane			0.300	0.302	0.286	0.253	0.258	0.280	8.30
7) Chloroethane			0.246	0.248	0.236	0.194	0.199	0.225	11.59
8) T Dichlorotetra...			1.156	1.241	1.262	1.091	1.116	1.173	6.43
9) T Propene			0.677	0.598	0.621	0.501	0.548	0.589	11.44
10) T Heptane			2.106	1.804	1.728	1.604	1.522	1.753	12.87
11) T Trichlorofluor...			1.238	1.515	1.487	1.240	1.352	1.367	9.62
12) T 1,1,2-Trichlor...			1.061	1.142	1.115	0.987	1.029	1.067	5.92
13) Ethanol			0.215	0.231	0.252	0.166	0.172	0.207	17.99
14) T Bromoethene			0.448	0.423	0.436	0.362	0.376	0.409	9.29
15) T Acetone			1.721	1.591	1.519	1.032	1.085	1.389	22.41
16) T 1,3-Butadiene			0.745	0.629	0.626	0.551	0.546	0.619	13.01
17) tert-Butyl alc...			1.784	2.014	1.880	1.705	1.735	1.824	6.87
18) T 1,1-Dichloroet...			0.555	0.530	0.515	0.441	0.443	0.497	10.52
19) T Isopropyl Alcohol			0.848	0.861	0.823	0.684	0.709	0.785	10.51
20) T Methylene Chlo...			0.906	0.710	0.585	0.398	0.403	0.600	35.87
21) T Allyl Chloride			1.139	0.895	0.936	0.849	0.855	0.935	12.75
22) T trans-1,2-Dich...			0.605	0.614	0.599	0.499	0.521	0.568	9.39
23) T Vinyl Acetate			1.885	1.632	1.682	1.622	1.565	1.677	7.36
24) T 1,1-Dichloroet...			1.050	1.173	1.194	1.018	1.044	1.096	7.42
25) T Ethyl Acetate			3.371	3.123	3.074	2.759	2.655	2.996	9.68
26) T Hexane			1.374	1.424	1.375	1.235	1.161	1.314	8.44
27) T Carbon Disulfide			1.303	1.424	1.432	1.257	1.291	1.341	6.03
28) T Methyl tert-Bu...			0.642	0.704	0.674	0.591	0.614	0.645	7.02
29) T Chloroform			1.955	1.898	1.928	1.785	1.770	1.867	4.52
30) T Cyclohexane			1.135	1.194	1.130	1.024	1.024	1.101	6.80
31) T cis-1,2-Dichlo...			1.399	1.351	1.322	1.240	1.217	1.306	5.83
32) T 1,1,1-Trichlor...	2.116	2.046	1.970	2.018	2.033	1.884	1.865	1.990	4.52

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.744	0.751	0.736	0.644	0.655	0.706	7.38
35) T Carbon Tetrach...	0.832	0.765	0.779	0.771	0.737	0.708	0.721	0.759	5.51
36) T Benzene			0.996	1.001	0.986	0.905	0.900	0.958	5.28
37) T 1,2-Dichloroet...			0.572	0.580	0.565	0.531	0.553	0.560	3.39
38) T Trichloroethene	0.293	0.475	0.390	0.388	0.371	0.353	0.353	0.375	14.65
39) T 1,2-Dichloropr...			0.372	0.364	0.350	0.335	0.332	0.351	4.97

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL111825AIR.M										
40)	T	1,4-Dioxane		0.180	0.166	0.155	0.142	0.144	0.158	10.04
41)	T	Tetrahydrofuran		0.418	0.401	0.402	0.367	0.360	0.390	6.32
42)	T	Bromodichlorom...		0.782	0.795	0.783	0.745	0.758	0.773	2.63
43)		Methyl Methacr...		0.576	0.532	0.511	0.462	0.462	0.508	9.55
44)	T	2,2,4-Trimethy...		1.746	1.765	1.621	1.555	1.537	1.645	6.47
45)	T	t-1,3-Dichloro...		0.495	0.472	0.489	0.455	0.458	0.474	3.79
46)	T	cis-1,3-Dichlo...		0.599	0.557	0.562	0.534	0.529	0.556	5.00
47)	T	1,1,2-Trichlor...		0.424	0.372	0.378	0.344	0.342	0.372	8.91
48)	T	Dibromochlorom...		0.629	0.666	0.663	0.618	0.621	0.640	3.63
49)	T	Bromoform		0.583	0.569	0.581	0.555	0.545	0.566	2.86
50)	T	4-Methyl-2-Pen...		1.025	1.061	1.027	0.946	0.914	0.994	6.19
51)	T	2-Hexanone		0.782	0.805	0.816	0.734	0.724	0.772	5.37
52)	T	Tetrachloroethene	0.363 0.424	0.344	0.362	0.349	0.329	0.323	0.356	9.37
53)	T	Toluene		1.198	1.182	1.146	1.075	1.069	1.134	5.26
54)	T	1,2-Dibromoethane	0.528	0.587	0.569	0.549	0.525	0.526	0.547	4.79
55)	I	Chlorobenzene-d5	-----ISTD-----							
56)		1,1,1,2-Tetrac...		0.791	0.769	0.761	0.677	0.677	0.735	7.34
57)	T	Chlorobenzene		1.035	1.088	1.089	0.969	0.953	1.027	6.24
58)	T	Ethyl Benzene		1.927	2.038	2.008	1.821	1.830	1.925	5.16
59)	T	m/p-Xylene		1.618	1.611	1.607	1.434	1.446	1.543	6.12
60)	T	o-Xylene		1.584	1.625	1.659	1.419	1.421	1.542	7.42
61)	T	Styrene		0.669	0.690	0.719	0.648	0.647	0.675	4.54
62)		Isopropylbenzene		2.807	2.946	2.945	2.556	2.536	2.758	7.31
63)	T	1,1,2,2-Tetrac...	0.999 1.034	1.016	1.012	1.028	0.891	0.884	0.981	6.58
64)		n-propylbenzene		0.722	0.758	0.757	0.657	0.649	0.709	7.47
65)		tert-Butylbenzene		2.583	2.687	2.646	2.224	2.176	2.463	9.90
66)	T	Benzyl Chloride		0.275	0.251	0.257	0.241	0.225	0.250	7.36
67)		sec-Butylbenzene		3.588	3.728	3.697	3.108	3.070	3.438	9.40
68)	S	1-Bromo-4-Fluo...	0.712 0.725	0.729	0.733	0.741	0.735	0.757	0.733	1.90
69)		p-Isopropyltol...		2.994	3.070	3.197	2.660	2.641	2.912	8.58
70)		n-Butylbenzene		2.676	3.041	3.125	2.639	2.647	2.825	8.39
71)		2-Chlorotoluene		2.198	2.257	2.210	1.960	1.975	2.120	6.65
72)	T	4-Ethyltoluene		1.934	1.948	2.009	1.749	1.747	1.877	6.46
73)	T	1,3,5-Trimethy...		1.623	1.682	1.662	1.456	1.454	1.575	7.09
74)	T	1,2,4-Trimethy...		1.796	1.844	1.871	1.548	1.537	1.719	9.51
75)	T	1,3-Dichlorobe...		1.043	1.053	1.061	0.937	0.922	1.003	6.78
76)	T	1,4-Dichlorobe...		1.056	1.033	1.060	0.940	0.930	1.003	6.34
77)	T	1,2-Dichlorobe...		0.974	1.064	1.050	0.872	0.863	0.965	9.83
78)	T	Hexachloro-1,3...		0.854	0.899	0.920	0.686	0.666	0.805	14.94
79)	T	Naphthalene	1.230	1.483	1.599	1.860	1.496	1.493	1.527	13.35
80)	T	Naphthalene,2-...		0.397	0.323	0.448	0.397	0.439	0.401	12.30
81)	T	1,2,4-Trichlor...		0.549	0.692	0.813	0.678	0.682	0.683	13.72

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043203.D
 Acq On : 18 Nov 2025 09:29
 Operator : SY/MD
 Sample : VSTDIC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5

Quant Time: Nov 19 00:48:47 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.784	49	110490	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.956	114	296493	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	247761	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.371	95	180529	9.941	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.400%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.496	85	7723	0.496	ppbv	99
3) Chlorodifluoromethane	1.473	51	8538	0.557	ppbv #	82
4) Chloromethane	1.531	50	2881	0.494	ppbv	94
5) Vinyl Chloride	1.580	62	2947	0.462	ppbv #	87
6) Bromomethane	1.667	94	1656	0.536	ppbv	97
7) Chloroethane	1.703	64	1357	0.547	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	6386	0.493	ppbv	100
9) Propene	1.483	41	3739	0.526	ppbv #	1
10) Heptane	4.972	43	11637m	0.597	ppbv	
11) Trichlorofluoromethane	1.878	101	6841	0.453	ppbv	92
12) 1,1,2-Trichlorotrifluo...	2.140	101	5863	0.497	ppbv	99
13) Ethanol	1.719	45	1190	0.520	ppbv	64
14) Bromoethene	1.781	108	2476	0.548	ppbv #	75
15) Acetone	1.836	43	9506	0.619	ppbv #	88
16) 1,3-Butadiene	1.606	39	4115	0.601	ppbv	95
17) tert-Butyl alcohol	2.043	59	9854	0.489	ppbv #	68
18) 1,1-Dichloroethene	2.033	96	3068	0.559	ppbv	96
19) Isopropyl Alcohol	1.887	45	4687	0.540	ppbv #	51
20) Methylene Chloride	2.059	84	5006	0.349	ppbv	92
21) Allyl Chloride	2.095	41	6291	0.609	ppbv	95
22) trans-1,2-Dichloroethene	2.337	96	3343	0.533	ppbv	80
23) Vinyl Acetate	2.464	43	10414	0.562	ppbv #	92
24) 1,1-Dichloroethane	2.409	63	5803	0.479	ppbv #	96
25) Ethyl Acetate	2.839	43	18625m	0.562	ppbv	
26) Hexane	2.823	57	7589	0.523	ppbv #	91
27) Carbon Disulfide	2.150	76	7198	0.486	ppbv #	1
28) Methyl tert-Butyl Ether	2.441	73	3545	0.498	ppbv #	15
29) Chloroform	2.849	83	10800	0.523	ppbv	95
30) Cyclohexane	3.849	84	6271	0.515	ppbv	85
31) cis-1,2-Dichloroethene	2.719	61	7728	0.536	ppbv	87
32) 1,1,1-Trichloroethane	3.363	97	10886	0.490	ppbv	98
34) 2-Butanone	2.557	43	11026	0.527	ppbv	95
35) Carbon Tetrachloride	3.761	117	11551	0.513	ppbv	94
36) Benzene	3.655	78	14759	0.520	ppbv #	91
37) 1,2-Dichloroethane	3.218	62	8477	0.510	ppbv #	93
38) Trichloroethene	4.542	130	5775	0.519	ppbv #	87
39) 1,2-Dichloropropane	4.312	63	5510	0.530	ppbv #	80
40) 1,4-Dioxane	4.597	88	2675m	0.573	ppbv	
41) Tetrahydrofuran	3.059	42	6193	0.536	ppbv	95
42) Bromodichloromethane	4.490	83	11597	0.506	ppbv #	84
43) Methyl Methacrylate	4.881	69	8534m	0.566	ppbv	
44) 2,2,4-Trimethylpentane	4.642	57	25890	0.531	ppbv	95
45) t-1,3-Dichloropropene	6.412	75	7343	0.523	ppbv	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043203.D
 Acq On : 18 Nov 2025 09:29
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Quant Time: Nov 19 00:48:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	8886m	0.539	ppbv	
47) 1,1,2-Trichloroethane	6.581	97	6280m	0.567	ppbv	
48) Dibromochloromethane	7.380	129	9327	0.492	ppbv	97
49) Bromoform	9.481	173	8636	0.514	ppbv	100
50) 4-Methyl-2-Pentanone	5.755	43	15190	0.515	ppbv	92
51) 2-Hexanone	7.442	43	11594	0.506	ppbv #	93
52) Tetrachloroethene	8.218	164	5097	0.483	ppbv #	82
53) Toluene	6.924	91	17755m	0.525	ppbv	
54) 1,2-Dibromoethane	7.649	107	8708	0.537	ppbv	90
56) 1,1,1,2-Tetrachloroethane	8.917	131	9793m	0.538	ppbv	
57) Chlorobenzene	8.917	112	12818	0.504	ppbv	98
58) Ethyl Benzene	9.351	91	23866	0.500	ppbv	96
59) m/p-Xylene	9.526	91	40076m	1.044	ppbv	
60) o-Xylene	9.956	91	19621	0.514	ppbv	100
61) Styrene	9.869	104	8287	0.496	ppbv	96
62) Isopropylbenzene	10.536	105	34773	0.509	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.960	83	12591	0.519	ppbv	100
64) n-propylbenzene	10.986	120	8945	0.509	ppbv	100
65) tert-Butylbenzene	11.526	119	31998	0.524	ppbv	99
66) Benzyl Chloride	11.610	91	3403	0.550	ppbv	98
67) sec-Butylbenzene	11.766	105	44448	0.522	ppbv	97
69) p-Isopropyltoluene	11.921	119	37087	0.514	ppbv	100
70) n-Butylbenzene	12.277	91	33153	0.474	ppbv	94
71) 2-Chlorotoluene	10.895	91	27224	0.518	ppbv	99
72) 4-Ethyltoluene	11.122	105	23963	0.515	ppbv	97
73) 1,3,5-Trimethylbenzene	11.199	105	20104	0.515	ppbv	99
74) 1,2,4-Trimethylbenzene	11.533	105	22254	0.522	ppbv	91
75) 1,3-Dichlorobenzene	11.604	146	12926	0.520	ppbv	98
76) 1,4-Dichlorobenzene	11.672	146	13076	0.526	ppbv	98
77) 1,2-Dichlorobenzene	11.947	146	12069	0.505	ppbv	97
78) Hexachloro-1,3-Butadiene	13.879	225	10578	0.530	ppbv	98
79) Naphthalene	13.513	128	18377	0.486	ppbv	97
80) Naphthalene,2-methyl-	14.462	142	4917	0.495	ppbv	91
81) 1,2,4-Trichlorobenzene	13.439	180	6798	0.402	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043203.D
Acq On : 18 Nov 2025 09:29
Operator : SY/MD
Sample : VSTDICC0.5
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:48:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043204.D
 Acq On : 18 Nov 2025 10:02
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Quant Time: Nov 19 00:49:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	105095	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	283830	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	234772	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	172524	10.025	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.300%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.496	85	140789	9.498	ppbv	100
3) Chlorodifluoromethane	1.473	51	138443	9.497	ppbv	100
4) Chloromethane	1.531	50	52372	9.450	ppbv	100
5) Vinyl Chloride	1.577	62	51045	8.422	ppbv	100
6) Bromomethane	1.664	94	26542	9.034	ppbv	100
7) Chloroethane	1.700	64	20375	8.632	ppbv	100
8) Dichlorotetrafluoroethane	1.551	85	114670	9.300	ppbv	100
9) Propene	1.480	41	52671	7.789	ppbv #	1
10) Heptane	4.965	43	168611	9.090	ppbv	100
11) Trichlorofluoromethane	1.875	101	130362	9.077	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.137	101	103679	9.247	ppbv	100
13) Ethanol	1.716	45	17464	8.023	ppbv	100
14) Bromoethene	1.777	108	38037	8.849	ppbv	100
15) Acetone	1.832	43	108422	7.426	ppbv	100
16) 1,3-Butadiene	1.606	39	57871	8.890	ppbv	100
17) tert-Butyl alcohol	2.036	59	179201	9.350	ppbv	100
18) 1,1-Dichloroethene	2.030	96	46313	8.871	ppbv	100
19) Isopropyl Alcohol	1.881	45	71862	8.710	ppbv	100
20) Methylene Chloride	2.056	84	41868	9.708	ppbv	100
21) Allyl Chloride	2.091	41	89266	9.087	ppbv	100
22) trans-1,2-Dichloroethene	2.334	96	52470	8.793	ppbv	100
23) Vinyl Acetate	2.457	43	170459	9.670	ppbv	100
24) 1,1-Dichloroethane	2.405	63	106970	9.287	ppbv	100
25) Ethyl Acetate	2.829	43	289921	9.190	ppbv	100
26) Hexane	2.820	57	129809	9.402	ppbv	100
27) Carbon Disulfide	2.146	76	132086	9.371	ppbv	100
28) Methyl tert-Butyl Ether	2.434	73	62115	9.165	ppbv	100
29) Chloroform	2.842	83	187620	9.561	ppbv	100
30) Cyclohexane	3.846	84	107621	9.297	ppbv	100
31) cis-1,2-Dichloroethene	2.716	61	130357	9.499	ppbv	100
32) 1,1,1-Trichloroethane	3.360	97	198036	9.381	ppbv	100
34) 2-Butanone	2.551	43	182749	9.121	ppbv	100
35) Carbon Tetrachloride	3.755	117	200831	9.322	ppbv	100
36) Benzene	3.648	78	256954	9.453	ppbv	100
37) 1,2-Dichloroethane	3.211	62	150769	9.483	ppbv	100
38) Trichloroethene	4.535	130	100136	9.398	ppbv	100
39) 1,2-Dichloropropane	4.302	63	95097	9.559	ppbv	100
40) 1,4-Dioxane	4.567	88	40441	9.043	ppbv #	100
41) Tetrahydrofuran	3.046	42	104294	9.426	ppbv	100
42) Bromodichloromethane	4.477	83	211504	9.643	ppbv	100
43) Methyl Methacrylate	4.865	69	131003	9.080	ppbv	100
44) 2,2,4-Trimethylpentane	4.626	57	441234	9.451	ppbv	100
45) t-1,3-Dichloropropene	6.399	75	129122	9.601	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043204.D
 Acq On : 18 Nov 2025 10:02
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Quant Time: Nov 19 00:49:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.584	75	151655	9.603	ppbv	100
47) 1,1,2-Trichloroethane	6.564	97	97512	9.193	ppbv	100
48) Dibromochloromethane	7.377	129	175529	9.670	ppbv	100
49) Bromoform	9.474	173	157563	9.800	ppbv	100
50) 4-Methyl-2-Pentanone	5.733	43	268369	9.509	ppbv	100
51) 2-Hexanone	7.419	43	208268	9.503	ppbv #	100
52) Tetrachloroethene	8.212	164	93518	9.249	ppbv	100
53) Toluene	6.920	91	305169	9.422	ppbv	100
54) 1,2-Dibromoethane	7.639	107	148973	9.588	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.914	131	158856	9.207	ppbv	100
57) Chlorobenzene	8.911	112	227490	9.438	ppbv	100
58) Ethyl Benzene	9.345	91	427530	9.461	ppbv	100
59) m/p-Xylene	9.542	91	673371m	18.519	ppbv	
60) o-Xylene	9.956	91	333044	9.202	ppbv	100
61) Styrene	9.863	104	152172	9.607	ppbv	100
62) Isopropylbenzene	10.533	105	600155	9.269	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.953	83	209272	9.109	ppbv	100
64) n-propylbenzene	10.982	120	154239	9.271	ppbv	100
65) tert-Butylbenzene	11.523	119	522106	9.029	ppbv	100
66) Benzyl Chloride	11.607	91	56620	9.649	ppbv	100
67) sec-Butylbenzene	11.762	105	729573	9.039	ppbv	100
69) p-Isopropyltoluene	11.918	119	624586	9.135	ppbv	100
70) n-Butylbenzene	12.274	91	619610	9.341	ppbv	100
71) 2-Chlorotoluene	10.895	91	460118	9.245	ppbv	100
72) 4-Ethyltoluene	11.118	105	410587	9.316	ppbv	100
73) 1,3,5-Trimethylbenzene	11.196	105	341900	9.244	ppbv	100
74) 1,2,4-Trimethylbenzene	11.529	105	363476	9.005	ppbv	100
75) 1,3-Dichlorobenzene	11.601	146	219894	9.335	ppbv	100
76) 1,4-Dichlorobenzene	11.665	146	220576	9.363	ppbv	100
77) 1,2-Dichlorobenzene	11.940	146	204826	9.044	ppbv	100
78) Hexachloro-1,3-Butadiene	13.876	225	161034	8.520	ppbv	100
79) Naphthalene	13.507	128	351148	9.795	ppbv	100
80) Naphthalene,2-methyl-	14.452	142	93113	9.895	ppbv	100
81) 1,2,4-Trichlorobenzene	13.436	180	159091	9.923	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:49:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043205.D
 Acq On : 18 Nov 2025 10:37
 Operator : SY/MD
 Sample : VSTDIC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002

Quant Time: Nov 19 00:50:19 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.787	49	104644	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	284629	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	230646	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	171009	10.115	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.499	85	30881	2.092	ppbv	98
3) Chlorodifluoromethane	1.476	51	30069	2.072	ppbv	98
4) Chloromethane	1.534	50	11890	2.155	ppbv	96
5) Vinyl Chloride	1.580	62	12040	1.995	ppbv	96
6) Bromomethane	1.670	94	5981	2.044	ppbv	99
7) Chloroethane	1.706	64	4943	2.103	ppbv	95
8) Dichlorotetrafluoroethane	1.554	85	26409	2.151	ppbv	98
9) Propene	1.486	41	12993	1.930	ppbv #	1
10) Heptane	4.972	43	36171m	1.958	ppbv	
11) Trichlorofluoromethane	1.878	101	31124	2.177	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	23336	2.090	ppbv	100
13) Ethanol	1.722	45	5266	2.430	ppbv	86
14) Bromoethene	1.784	108	9129	2.133	ppbv	95
15) Acetone	1.835	43	31785	2.186	ppbv	91
16) 1,3-Butadiene	1.609	39	13101	2.021	ppbv	89
17) tert-Butyl alcohol	2.039	59	39348	2.062	ppbv	95
18) 1,1-Dichloroethene	2.036	96	10780	2.074	ppbv	94
19) Isopropyl Alcohol	1.887	45	17229	2.097	ppbv #	71
20) Methylene Chloride	2.062	84	12238	2.248	ppbv	96
21) Allyl Chloride	2.098	41	19580	2.002	ppbv	95
22) trans-1,2-Dichloroethene	2.340	96	12537	2.110	ppbv	94
23) Vinyl Acetate	2.460	43	35205	2.006	ppbv #	97
24) 1,1-Dichloroethane	2.408	63	24987	2.179	ppbv	99
25) Ethyl Acetate	2.832	43	64343	2.048	ppbv	99
26) Hexane	2.826	57	28776	2.093	ppbv	97
27) Carbon Disulfide	2.149	76	29964	2.135	ppbv #	91
28) Methyl tert-Butyl Ether	2.438	73	14101	2.090	ppbv #	85
29) Chloroform	2.849	83	40355	2.065	ppbv	98
30) Cyclohexane	3.855	84	23659	2.053	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	27663	2.024	ppbv	98
32) 1,1,1-Trichloroethane	3.366	97	42547	2.024	ppbv	99
34) 2-Butanone	2.557	43	41884	2.085	ppbv	99
35) Carbon Tetrachloride	3.765	117	41949	1.942	ppbv	96
36) Benzene	3.654	78	56141m	2.060	ppbv	
37) 1,2-Dichloroethane	3.218	62	32188	2.019	ppbv	100
38) Trichloroethene	4.548	130	21105m	1.975	ppbv	
39) 1,2-Dichloropropane	4.312	63	19913	1.996	ppbv	93
40) 1,4-Dioxane	4.580	88	8829m	1.969	ppbv	
41) Tetrahydrofuran	3.056	42	22912	2.065	ppbv	99
42) Bromodichloromethane	4.490	83	44584	2.027	ppbv	98
43) Methyl Methacrylate	4.875	69	29071	2.009	ppbv	99
44) 2,2,4-Trimethylpentane	4.635	57	92261	1.971	ppbv	96
45) t-1,3-Dichloropropene	6.406	75	27833m	2.064	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043205.D
 Acq On : 18 Nov 2025 10:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Quant Time: Nov 19 00:50:19 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	31986	2.020	ppbv	91
47) 1,1,2-Trichloroethane	6.577	97	21542m	2.025	ppbv	
48) Dibromochloromethane	7.386	129	37758m	2.074	ppbv	
49) Bromoform	9.480	173	33057	2.050	ppbv	98
50) 4-Methyl-2-Pentanone	5.745	43	58461	2.066	ppbv	98
51) 2-Hexanone	7.432	43	46444	2.113	ppbv #	99
52) Tetrachloroethene	8.218	164	19856	1.958	ppbv	96
53) Toluene	6.927	91	65238	2.008	ppbv	97
54) 1,2-Dibromoethane	7.648	107	31266	2.007	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.920	131	35099	2.071	ppbv	99
57) Chlorobenzene	8.917	112	50228	2.121	ppbv	97
58) Ethyl Benzene	9.351	91	92649	2.087	ppbv	98
59) m/p-Xylene	9.529	91	148300m	4.151	ppbv	
60) o-Xylene	9.959	91	76540	2.153	ppbv	100
61) Styrene	9.866	104	33188	2.133	ppbv	98
62) Isopropylbenzene	10.536	105	135853	2.136	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.963	83	47439	2.102	ppbv	99
64) n-propylbenzene	10.985	120	34915	2.136	ppbv	99
65) tert-Butylbenzene	11.529	119	122064	2.149	ppbv	99
66) Benzyl Chloride	11.610	91	11870	2.059	ppbv	99
67) sec-Butylbenzene	11.765	105	170529	2.151	ppbv	100
69) p-Isopropyltoluene	11.921	119	147482	2.195	ppbv	98
70) n-Butylbenzene	12.280	91	144148	2.212	ppbv	99
71) 2-Chlorotoluene	10.898	91	101951	2.085	ppbv	100
72) 4-Ethyltoluene	11.121	105	92654	2.140	ppbv	99
73) 1,3,5-Trimethylbenzene	11.202	105	76655	2.110	ppbv	96
74) 1,2,4-Trimethylbenzene	11.532	105	86286	2.176	ppbv	98
75) 1,3-Dichlorobenzene	11.604	146	48961	2.116	ppbv	98
76) 1,4-Dichlorobenzene	11.668	146	48881	2.112	ppbv	98
77) 1,2-Dichlorobenzene	11.944	146	48431	2.177	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	42456	2.287	ppbv	99
79) Naphthalene	13.510	128	85788	2.436	ppbv	97
80) Naphthalene,2-methyl-	14.458	142	20685	2.237	ppbv	98
81) 1,2,4-Trichlorobenzene	13.439	180	37521	2.382	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043206.D
 Acq On : 18 Nov 2025 11:10
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Quant Time: Nov 19 00:51:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	102795	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	278732	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	227998	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 167061 9.996 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.000%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	15474	1.067	ppbv	100
3) Chlorodifluoromethane	1.476	51	14437	1.013	ppbv #	92
4) Chloromethane	1.531	50	5751	1.061	ppbv	92
5) Vinyl Chloride	1.580	62	5967	1.007	ppbv	95
6) Bromomethane	1.667	94	3104	1.080	ppbv	94
7) Chloroethane	1.706	64	2550	1.104	ppbv	97
8) Dichlorotetrafluoroethane	1.554	85	12759	1.058	ppbv	98
9) Propene	1.483	41	6147	0.929	ppbv #	1
10) Heptane	4.972	43	18544	1.022	ppbv	100
11) Trichlorofluoromethane	1.877	101	15571	1.108	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.140	101	11743	1.071	ppbv	99
13) Ethanol	1.719	45	2373	1.115	ppbv	67
14) Bromoethene	1.780	108	4345	1.033	ppbv #	91
15) Acetone	1.835	43	16353	1.145	ppbv	90
16) 1,3-Butadiene	1.609	39	6469	1.016	ppbv	96
17) tert-Butyl alcohol	2.039	59	20703	1.104	ppbv #	92
18) 1,1-Dichloroethene	2.033	96	5447	1.067	ppbv	97
19) Isopropyl Alcohol	1.884	45	8849	1.097	ppbv #	43
20) Methylene Chloride	2.059	84	7297	1.030	ppbv	85
21) Allyl Chloride	2.094	41	9198	0.957	ppbv #	84
22) trans-1,2-Dichloroethene	2.340	96	6313	1.082	ppbv	94
23) Vinyl Acetate	2.460	43	16779	0.973	ppbv	98
24) 1,1-Dichloroethane	2.405	63	12062	1.071	ppbv	94
25) Ethyl Acetate	2.832	43	32108	1.041	ppbv	98
26) Hexane	2.819	57	14639	1.084	ppbv	99
27) Carbon Disulfide	2.149	76	14637	1.062	ppbv #	30
28) Methyl tert-Butyl Ether	2.441	73	7236	1.092	ppbv #	83
29) Chloroform	2.845	83	19509	1.016	ppbv	91
30) Cyclohexane	3.852	84	12269	1.084	ppbv	96
31) cis-1,2-Dichloroethene	2.716	61	13889	1.035	ppbv	97
32) 1,1,1-Trichloroethane	3.363	97	20745	1.005	ppbv	99
34) 2-Butanone	2.554	43	20940	1.064	ppbv	96
35) Carbon Tetrachloride	3.758	117	21497	1.016	ppbv	96
36) Benzene	3.648	78	27911	1.046	ppbv	99
37) 1,2-Dichloroethane	3.214	62	16159	1.035	ppbv	99
38) Trichloroethene	4.538	130	10809m	1.033	ppbv	
39) 1,2-Dichloropropane	4.311	63	10147	1.039	ppbv	91
40) 1,4-Dioxane	4.577	88	4616m	1.051	ppbv	
41) Tetrahydrofuran	3.052	42	11176	1.029	ppbv	96
42) Bromodichloromethane	4.480	83	22150	1.028	ppbv #	98
43) Methyl Methacrylate	4.871	69	14821	1.046	ppbv	94
44) 2,2,4-Trimethylpentane	4.632	57	49203m	1.073	ppbv	
45) t-1,3-Dichloropropene	6.402	75	13149m	0.996	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043206.D
 Acq On : 18 Nov 2025 11:10
 Operator : SY/MD
 Sample : VSTDIC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC001

Quant Time: Nov 19 00:51:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.590	75	15527	1.001	ppbv #	95
47) 1,1,2-Trichloroethane	6.577	97	10369	0.995	ppbv #	85
48) Dibromochloromethane	7.377	129	18560	1.041	ppbv	93
49) Bromoform	9.477	173	15858	1.004	ppbv	98
50) 4-Methyl-2-Pentanone	5.745	43	29560	1.067	ppbv	99
51) 2-Hexanone	7.428	43	22438	1.043	ppbv #	98
52) Tetrachloroethene	8.215	164	10080	1.015	ppbv	90
53) Toluene	6.923	91	32951	1.036	ppbv	100
54) 1,2-Dibromoethane	7.645	107	15865	1.040	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.917	131	17535	1.046	ppbv	99
57) Chlorobenzene	8.914	112	24806	1.060	ppbv	98
58) Ethyl Benzene	9.344	91	46455	1.059	ppbv	96
59) m/p-Xylene	9.545	91	73461m	2.080	ppbv	
60) o-Xylene	9.956	91	37054	1.054	ppbv	97
61) Styrene	9.866	104	15730	1.023	ppbv	95
62) Isopropylbenzene	10.532	105	67163	1.068	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.956	83	23083	1.035	ppbv	97
64) n-propylbenzene	10.985	120	17289	1.070	ppbv	100
65) tert-Butylbenzene	11.526	119	61258	1.091	ppbv	99
66) Benzyl Chloride	11.610	91	5729	1.005	ppbv	96
67) sec-Butylbenzene	11.762	105	84989	1.084	ppbv	99
69) p-Isopropyltoluene	11.921	119	69993	1.054	ppbv	99
70) n-Butylbenzene	12.277	91	69323	1.076	ppbv	99
71) 2-Chlorotoluene	10.895	91	51452	1.065	ppbv	100
72) 4-Ethyltoluene	11.118	105	44403	1.037	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	38350	1.068	ppbv	91
74) 1,2,4-Trimethylbenzene	11.532	105	42051	1.073	ppbv #	81
75) 1,3-Dichlorobenzene	11.600	146	24011	1.050	ppbv	98
76) 1,4-Dichlorobenzene	11.665	146	23543	1.029	ppbv	97
77) 1,2-Dichlorobenzene	11.940	146	24248	1.103	ppbv	98
78) Hexachloro-1,3-Butadiene	13.879	225	20489	1.116	ppbv	99
79) Naphthalene	13.507	128	36462	1.047	ppbv	95
80) Naphthalene,2-methyl-	14.455	142	7375	0.807	ppbv	95
81) 1,2,4-Trichlorobenzene	13.436	180	15785	1.014	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043208.D
Acq On : 18 Nov 2025 12:16
Operator : SY/MD
Sample : VSTDIC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Quant Time: Nov 19 00:51:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.784	49	101366	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	273996	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	222034	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	160968	9.890	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.900%
Target Compounds						
5) Vinyl Chloride	1.580	62	664	0.114	ppbv	Qvalue 97
32) 1,1,1-Trichloroethane	3.360	97	2074m	0.102	ppbv	
35) Carbon Tetrachloride	3.761	117	2097m	0.101	ppbv	
38) Trichloroethene	4.541	130	1302m	0.127	ppbv	
52) Tetrachloroethene	8.218	164	1161m	0.119	ppbv	
54) 1,2-Dibromoethane	7.648	107	1446m	0.096	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.953	83	2296m	0.106	ppbv	
79) Naphthalene	13.510	128	2732m	0.081	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043208.D
Acq On : 18 Nov 2025 12:16
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:51:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043209.D
Acq On : 18 Nov 2025 12:49
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 00:52:39 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.784	49	98952	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	266012	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	213547	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.371	95	152015	9.712	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.100%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	215m	0.038	ppbv	
32) 1,1,1-Trichloroethane	3.357	97	628m	0.032	ppbv	
35) Carbon Tetrachloride	3.758	117	664m	0.033	ppbv	
38) Trichloroethene	4.542	130	234m	0.023	ppbv	
52) Tetrachloroethene	8.222	164	290m	0.031	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.963	83	640m	0.031	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043209.D
Acq On : 18 Nov 2025 12:49
Operator : SY/MD
Sample : VSTDIC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03

Quant Time: Nov 19 00:52:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
Supervised By : Mahesh Dadoda 11/19/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043210.D
 Acq On : 18 Nov 2025 13:24
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Nov 19 00:53:26 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	98968	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	258461	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	211247	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	159814	10.321	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.200%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.499	85	197987	14.184	ppbv	99
3) Chlorodifluoromethane	1.476	51	182813	13.317	ppbv	98
4) Chloromethane	1.531	50	72636	13.918	ppbv	97
5) Vinyl Chloride	1.580	62	71671	12.557	ppbv	96
6) Bromomethane	1.667	94	38265	13.830	ppbv	98
7) Chloroethane	1.703	64	29584	13.309	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	165709	14.271	ppbv	97
9) Propene	1.483	41	81393m	12.782	ppbv	
10) Heptane	4.972	43	225935	12.934	ppbv	100
11) Trichlorofluoromethane	1.874	101	200703	14.840	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.137	101	152772	14.469	ppbv	98
13) Ethanol	1.719	45	25468	12.425	ppbv	99
14) Bromoethene	1.781	108	55813	13.789	ppbv	96
15) Acetone	1.832	43	161036	11.712	ppbv	99
16) 1,3-Butadiene	1.609	39	81077	13.226	ppbv	99
17) tert-Butyl alcohol	2.036	59	257617	14.274	ppbv	99
18) 1,1-Dichloroethene	2.033	96	65731	13.370	ppbv	96
19) Isopropyl Alcohol	1.881	45	105271	13.549	ppbv	99
20) Methylene Chloride	2.059	84	59798	15.164	ppbv	93
21) Allyl Chloride	2.094	41	126959	13.724	ppbv	96
22) trans-1,2-Dichloroethene	2.337	96	77386	13.772	ppbv	94
23) Vinyl Acetate	2.457	43	232315	13.995	ppbv	99
24) 1,1-Dichloroethane	2.405	63	155017	14.292	ppbv	98
25) Ethyl Acetate	2.829	43	394074	13.265	ppbv	99
26) Hexane	2.823	57	172303	13.252	ppbv	97
27) Carbon Disulfide	2.150	76	191612	14.435	ppbv	97
28) Methyl tert-Butyl Ether	2.434	73	91132	14.279	ppbv	97
29) Chloroform	2.845	83	262803	14.221	ppbv	100
30) Cyclohexane	3.852	84	152015	13.946	ppbv	94
31) cis-1,2-Dichloroethene	2.716	61	180647	13.978	ppbv	98
32) 1,1,1-Trichloroethane	3.363	97	276915	13.930	ppbv	98
34) 2-Butanone	2.551	43	253855	13.914	ppbv	100
35) Carbon Tetrachloride	3.755	117	279560	14.250	ppbv	96
36) Benzene	3.651	78	348892	14.096	ppbv	99
37) 1,2-Dichloroethane	3.214	62	214248	14.798	ppbv	98
38) Trichloroethene	4.541	130	137027	14.123	ppbv	93
39) 1,2-Dichloropropane	4.302	63	128700	14.206	ppbv	98
40) 1,4-Dioxane	4.571	88	55905	13.728	ppbv #	95
41) Tetrahydrofuran	3.043	42	139722	13.868	ppbv	99
42) Bromodichloromethane	4.480	83	294008	14.721	ppbv	97
43) Methyl Methacrylate	4.872	69	179110	13.633	ppbv	100
44) 2,2,4-Trimethylpentane	4.632	57	595778	14.014	ppbv	99
45) t-1,3-Dichloropropene	6.406	75	177678	14.508	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043210.D
 Acq On : 18 Nov 2025 13:24
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Nov 19 00:53:26 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.587	75	205226	14.270	ppbv	100
47) 1,1,2-Trichloroethane	6.571	97	132701	13.738	ppbv	92
48) Dibromochloromethane	7.377	129	240736	14.564	ppbv	99
49) Bromoform	9.477	173	211325	14.434	ppbv	99
50) 4-Methyl-2-Pentanone	5.732	43	354421	13.790	ppbv	99
51) 2-Hexanone	7.422	43	280742	14.067	ppbv #	100
52) Tetrachloroethene	8.218	164	125132	13.591	ppbv	93
53) Toluene	6.924	91	414411	14.050	ppbv	99
54) 1,2-Dibromoethane	7.642	107	203913	14.413	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.917	131	214673	13.827	ppbv	99
57) Chlorobenzene	8.914	112	301986	13.923	ppbv	98
58) Ethyl Benzene	9.348	91	579890	14.262	ppbv	99
59) m/p-Xylene	9.545	91	916113m	28.000	ppbv	
60) o-Xylene	9.956	91	450260	13.826	ppbv	99
61) Styrene	9.862	104	204972	14.382	ppbv	99
62) Isopropylbenzene	10.532	105	803487	13.791	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.956	83	280254	13.557	ppbv	99
64) n-propylbenzene	10.986	120	205637	13.737	ppbv	100
65) tert-Butylbenzene	11.526	119	689352	13.249	ppbv	98
66) Benzyl Chloride	11.610	91	71386	13.520	ppbv	100
67) sec-Butylbenzene	11.762	105	972780	13.394	ppbv	100
69) p-Isopropyltoluene	11.921	119	836888	13.602	ppbv	99
70) n-Butylbenzene	12.277	91	838598	14.050	ppbv	99
71) 2-Chlorotoluene	10.895	91	625884	13.976	ppbv	100
72) 4-Ethyltoluene	11.121	105	553654	13.961	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	460783	13.846	ppbv	97
74) 1,2,4-Trimethylbenzene	11.533	105	486995	13.409	ppbv	97
75) 1,3-Dichlorobenzene	11.604	146	292143	13.784	ppbv	99
76) 1,4-Dichlorobenzene	11.668	146	294723	13.903	ppbv	99
77) 1,2-Dichlorobenzene	11.944	146	273467	13.420	ppbv	98
78) Hexachloro-1,3-Butadiene	13.876	225	211153	12.416	ppbv	99
79) Naphthalene	13.510	128	473229	14.670	ppbv	99
80) Naphthalene,2-methyl-	14.452	142	139019	16.418	ppbv	99
81) 1,2,4-Trichlorobenzene	13.436	180	216260	14.990	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 11/19/2025
 Supervised By :Mahesh Dadoda 11/19/2025

Quant Time: Nov 19 01:34:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	97854	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.956	114	261914	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.879	117	216212	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.374 95 159242 10.048 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	133374	9.664	ppbv	98
3) Chlorodifluoromethane	1.476	51	124858	9.199	ppbv	98
4) Chloromethane	1.535	50	46906	9.090	ppbv	95
5) Vinyl Chloride	1.580	62	47932	8.493	ppbv	95
6) Bromomethane	1.667	94	22230	8.126	ppbv	99
7) Chloroethane	1.703	64	19854	9.033	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	110587	9.632	ppbv	97
9) Propene	1.483	41	54006m	9.370	ppbv	
10) Heptane	4.975	43	149202	8.698	ppbv	99
11) Trichlorofluoromethane	1.878	101	134746	10.077	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	100984	9.673	ppbv	99
13) Ethanol	1.719	45	15371	7.584	ppbv	92
14) Bromoethene	1.781	108	36979	9.240	ppbv	98
15) Acetone	1.832	43	104830	7.711	ppbv	99
16) 1,3-Butadiene	1.609	39	54795	9.041	ppbv	96
17) tert-Butyl alcohol	2.036	59	164987	9.245	ppbv	99
18) 1,1-Dichloroethene	2.033	96	43833	9.017	ppbv	99
19) Isopropyl Alcohol	1.884	45	64772	8.431	ppbv	97
20) Methylene Chloride	2.062	84	41254	10.323	ppbv	95
21) Allyl Chloride	2.095	41	75516	8.256	ppbv	97
22) trans-1,2-Dichloroethene	2.341	96	50228	9.041	ppbv	92
23) Vinyl Acetate	2.460	43	140156	8.539	ppbv	98
24) 1,1-Dichloroethane	2.409	63	103187	9.622	ppbv	97
25) Ethyl Acetate	2.829	43	258609	8.820	ppbv	99
26) Hexane	2.826	57	114312	8.892	ppbv	98
27) Carbon Disulfide	2.150	76	125048	9.528	ppbv	97
28) Methyl tert-Butyl Ether	2.438	73	64183	10.171	ppbv	98
29) Chloroform	2.846	83	173977	9.521	ppbv	99
30) Cyclohexane	3.852	84	98598	9.148	ppbv	97
31) cis-1,2-Dichloroethene	2.719	61	119843	9.379	ppbv	98
32) 1,1,1-Trichloroethane	3.363	97	182262	9.358	ppbv	98
34) 2-Butanone	2.551	43	160977	8.707	ppbv	98
35) Carbon Tetrachloride	3.761	117	182902	9.200	ppbv	98
36) Benzene	3.655	78	233094	9.293	ppbv	99
37) 1,2-Dichloroethane	3.214	62	143363	9.772	ppbv	98
38) Trichloroethene	4.542	130	90993	9.272	ppbv	95
39) 1,2-Dichloropropane	4.305	63	83416	9.086	ppbv	95
40) 1,4-Dioxane	4.571	88	32109	7.780	ppbv #	95
41) Tetrahydrofuran	3.046	42	90769	8.891	ppbv	98
42) Bromodichloromethane	4.486	83	195442	9.657	ppbv	96
43) Methyl Methacrylate	4.872	69	116325	8.737	ppbv	98
44) 2,2,4-Trimethylpentane	4.635	57	393438	9.133	ppbv	99
45) t-1,3-Dichloropropene	6.409	75	97587	7.863	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/19/2025
 Supervised By : Mahesh Dadoda 11/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	115476	7.924	ppbv	97
47) 1,1,2-Trichloroethane	6.577	97	88525	9.086	ppbv	95
48) Dibromochloromethane	7.380	129	159365	9.514	ppbv	99
49) Bromoform	9.477	173	134790	9.085	ppbv	99
50) 4-Methyl-2-Pentanone	5.742	43	224791	8.631	ppbv	99
51) 2-Hexanone	7.429	43	166506	8.233	ppbv #	100
52) Tetrachloroethene	8.222	164	82890	8.884	ppbv	98
53) Toluene	6.924	91	273172	9.197	ppbv	99
54) 1,2-Dibromoethane	7.645	107	124804	8.705	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.921	131	142512	8.969	ppbv	98
57) Chlorobenzene	8.917	112	199373	8.981	ppbv	96
58) Ethyl Benzene	9.351	91	381474	9.167	ppbv	97
59) m/p-Xylene	9.549	91	604165m	18.108	ppbv	
60) o-Xylene	9.960	91	295836	8.876	ppbv	99
61) Styrene	9.866	104	127117	8.714	ppbv	97
62) Isopropylbenzene	10.536	105	525717	8.816	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.960	83	178591	8.421	ppbv	100
64) n-propylbenzene	10.986	120	131634	8.591	ppbv	96
65) tert-Butylbenzene	11.529	119	450819	8.465	ppbv	99
66) Benzyl Chloride	11.614	91	27878	5.159	ppbv	99
67) sec-Butylbenzene	11.766	105	623520	8.388	ppbv	100
69) p-Isopropyltoluene	11.924	119	520880	8.272	ppbv	99
70) n-Butylbenzene	12.280	91	513027	8.398	ppbv	99
71) 2-Chlorotoluene	10.898	91	396434	8.649	ppbv	99
72) 4-Ethyltoluene	11.122	105	356325	8.779	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	294977	8.660	ppbv	95
74) 1,2,4-Trimethylbenzene	11.533	105	314606	8.463	ppbv	94
75) 1,3-Dichlorobenzene	11.604	146	180491	8.320	ppbv	99
76) 1,4-Dichlorobenzene	11.669	146	179491	8.273	ppbv	99
77) 1,2-Dichlorobenzene	11.944	146	166137	7.966	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	125707	7.222	ppbv	100
79) Naphthalene	13.510	128	273029	8.270	ppbv	99
80) Naphthalene,2-methyl-	14.455	142	74802	8.631	ppbv	97
81) 1,2,4-Trichlorobenzene	13.436	180	125470	8.497	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	1.410	1.363	3.3	95	0.00
3	Chlorodifluoromethane	1.387	1.276	8.0	90	0.00
4	Chloromethane	0.527	0.479	9.1	90	0.00
5 T	Vinyl Chloride	0.577	0.490	15.1	94	0.00
6 T	Bromomethane	0.280	0.227	18.9	84	0.00
7	Chloroethane	0.225	0.203	9.8	97	0.00
8 T	Dichlorotetrafluoroethane	1.173	1.130	3.7	96	0.00
9 T	Propene	0.589	0.552	6.3	103	0.00
10 T	Heptane	1.753	1.525	13.0	88	0.00
11 T	Trichlorofluoromethane	1.367	1.377	-0.7	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.067	1.032	3.3	97	0.00
13	Ethanol	0.207	0.157	24.2	88	0.00
14 T	Bromoethene	0.409	0.378	7.6	97	0.00
15 T	Acetone	1.389	1.071	22.9	97	0.00
16 T	1,3-Butadiene	0.619	0.560	9.5	95	0.00
17	tert-Butyl alcohol	1.824	1.686	7.6	92	0.00
18 T	1,1-Dichloroethene	0.497	0.448	9.9	95	0.00
19 T	Isopropyl Alcohol	0.785	0.662	15.7	90	0.00
20 T	Methylene Chloride	0.600	0.422	29.7	99	0.00
21 T	Allyl Chloride	0.935	0.772	17.4	85	0.00
22 T	trans-1,2-Dichloroethene	0.568	0.513	9.7	96	0.00
23 T	Vinyl Acetate	1.677	1.432	14.6	82	0.00
24 T	1,1-Dichloroethane	1.096	1.054	3.8	96	0.00
25 T	Ethyl Acetate	2.996	2.643	11.8	89	0.00
26 T	Hexane	1.314	1.168	11.1	88	0.00
27 T	Carbon Disulfide	1.341	1.278	4.7	95	0.00
28 T	Methyl tert-Butyl Ether	0.645	0.656	-1.7	103	0.00
29 T	Chloroform	1.867	1.778	4.8	93	0.00
30 T	Cyclohexane	1.101	1.008	8.4	92	0.00
31 T	cis-1,2-Dichloroethene	1.306	1.225	6.2	92	0.00
32 T	1,1,1-Trichloroethane	1.990	1.863	6.4	92	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
34 T	2-Butanone	0.706	0.615	12.9	88	0.00
35 T	Carbon Tetrachloride	0.759	0.698	8.0	91	0.00
36 T	Benzene	0.958	0.890	7.1	91	0.00
37 T	1,2-Dichloroethane	0.560	0.547	2.3	95	0.00
38 T	Trichloroethene	0.375	0.347	7.5	91	0.00
39 T	1,2-Dichloropropane	0.351	0.318	9.4	88	0.00
40 T	1,4-Dioxane	0.158	0.123	22.2	79	0.00
41 T	Tetrahydrofuran	0.390	0.347	11.0	87	0.00
42 T	Bromodichloromethane	0.773	0.746	3.5	92	0.00
43	Methyl Methacrylate	0.508	0.444	12.6	89	0.00
44 T	2,2,4-Trimethylpentane	1.645	1.502	8.7	89	0.00
45 T	t-1,3-Dichloropropene	0.474	0.373	21.3	76	0.00
46 T	cis-1,3-Dichloropropene	0.556	0.441	20.7	76	0.00
47 T	1,1,2-Trichloroethane	0.372	0.338	9.1	91	0.01
48 T	Dibromochloromethane	0.640	0.608	5.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.566	0.515	9.0	86	0.00
50 T	4-Methyl-2-Pentanone	0.994	0.858	13.7	84	0.00
51 T	2-Hexanone	0.772	0.636	17.6	80	0.00
52 T	Tetrachloroethene	0.356	0.316	11.2	89	0.00
53 T	Toluene	1.134	1.043	8.0	90	0.00
54 T	1,2-Dibromoethane	0.547	0.477	12.8	84	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
56	1,1,1,2-Tetrachloroethane	0.735	0.659	10.3	90	0.00
57 T	Chlorobenzene	1.027	0.922	10.2	88	0.00
58 T	Ethyl Benzene	1.925	1.764	8.4	89	0.00
59 T	m/p-Xylene	1.543	1.397	9.5	90	0.00
60 T	o-Xylene	1.542	1.368	11.3	89	0.00
61 T	Styrene	0.675	0.588	12.9	84	0.00
62	Isopropylbenzene	2.758	2.431	11.9	88	0.00
63 T	1,1,2,2-Tetrachloroethane	0.981	0.826	15.8	85	0.00
64	n-propylbenzene	0.709	0.609	14.1	85	0.00
65	tert-Butylbenzene	2.463	2.085	15.3	86	0.00
66 T	Benzyl Chloride	0.250	0.129	48.4#	49	0.00
67	sec-Butylbenzene	3.438	2.884	16.1	85	0.00
68 S	1-Bromo-4-Fluorobenzene	0.733	0.737	-0.5	92	0.00
69	p-Isopropyltoluene	2.912	2.409	17.3	83	0.00
70	n-Butylbenzene	2.825	2.373	16.0	83	0.00
71	2-Chlorotoluene	2.120	1.834	13.5	86	0.00
72 T	4-Ethyltoluene	1.877	1.648	12.2	87	0.00
73 T	1,3,5-Trimethylbenzene	1.575	1.364	13.4	86	0.00
74 T	1,2,4-Trimethylbenzene	1.719	1.455	15.4	87	0.00
75 T	1,3-Dichlorobenzene	1.003	0.835	16.7	82	0.00
76 T	1,4-Dichlorobenzene	1.003	0.830	17.2	81	0.00
77 T	1,2-Dichlorobenzene	0.965	0.768	20.4	81	0.00
78 T	Hexachloro-1,3-Butadiene	0.805	0.581	27.8	78	0.00
79 T	Naphthalene	1.527	1.263	17.3	78	0.00
80 T	Naphthalene,2-methyl-	0.401	0.346	13.7	80	0.00
81 T	1,2,4-Trichlorobenzene	0.683	0.580	15.1	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043211.D
Acq On : 18 Nov 2025 13:56
Operator : SY/MD
Sample : VSTDICV010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Quant Time: Nov 19 01:34:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	10.000	9.664	3.4	95	0.00
3	Chlorodifluoromethane	10.000	9.199	8.0	90	0.00
4	Chloromethane	10.000	9.090	9.1	90	0.00
5 T	Vinyl Chloride	10.000	8.493	15.1	94	0.00
6 T	Bromomethane	10.000	8.126	18.7	84	0.00
7	Chloroethane	10.000	9.033	9.7	97	0.00
8 T	Dichlorotetrafluoroethane	10.000	9.632	3.7	96	0.00
9 T	Propene	10.000	9.370	6.3	103	0.00
10 T	Heptane	10.000	8.698	13.0	88	0.00
11 T	Trichlorofluoromethane	10.000	10.077	-0.8	103	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	9.673	3.3	97	0.00
13	Ethanol	10.000	7.584	24.2	88	0.00
14 T	Bromoethene	10.000	9.240	7.6	97	0.00
15 T	Acetone	10.000	7.711	22.9	97	0.00
16 T	1,3-Butadiene	10.000	9.041	9.6	95	0.00
17	tert-Butyl alcohol	10.000	9.245	7.6	92	0.00
18 T	1,1-Dichloroethene	10.000	9.017	9.8	95	0.00
19 T	Isopropyl Alcohol	10.000	8.431	15.7	90	0.00
20 T	Methylene Chloride	10.000	10.323	-3.2	99	0.00
21 T	Allyl Chloride	10.000	8.256	17.4	85	0.00
22 T	trans-1,2-Dichloroethene	10.000	9.041	9.6	96	0.00
23 T	Vinyl Acetate	10.000	8.539	14.6	82	0.00
24 T	1,1-Dichloroethane	10.000	9.622	3.8	96	0.00
25 T	Ethyl Acetate	10.000	8.820	11.8	89	0.00
26 T	Hexane	10.000	8.892	11.1	88	0.00
27 T	Carbon Disulfide	10.000	9.528	4.7	95	0.00
28 T	Methyl tert-Butyl Ether	10.000	10.171	-1.7	103	0.00
29 T	Chloroform	10.000	9.521	4.8	93	0.00
30 T	Cyclohexane	10.000	9.148	8.5	92	0.00
31 T	cis-1,2-Dichloroethene	10.000	9.379	6.2	92	0.00
32 T	1,1,1-Trichloroethane	10.000	9.358	6.4	92	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	92	0.00
34 T	2-Butanone	10.000	8.707	12.9	88	0.00
35 T	Carbon Tetrachloride	10.000	9.200	8.0	91	0.00
36 T	Benzene	10.000	9.293	7.1	91	0.00
37 T	1,2-Dichloroethane	10.000	9.772	2.3	95	0.00
38 T	Trichloroethene	10.000	9.272	7.3	91	0.00
39 T	1,2-Dichloropropane	10.000	9.086	9.1	88	0.00
40 T	1,4-Dioxane	10.000	7.780	22.2	79	0.00
41 T	Tetrahydrofuran	10.000	8.891	11.1	87	0.00
42 T	Bromodichloromethane	10.000	9.657	3.4	92	0.00
43	Methyl Methacrylate	10.000	8.737	12.6	89	0.00
44 T	2,2,4-Trimethylpentane	10.000	9.133	8.7	89	0.00
45 T	t-1,3-Dichloropropene	10.000	7.863	21.4	76	0.00
46 T	cis-1,3-Dichloropropene	10.000	7.924	20.8	76	0.00
47 T	1,1,2-Trichloroethane	10.000	9.086	9.1	91	0.01
48 T	Dibromochloromethane	10.000	9.514	4.9	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043211.D
Acq On : 18 Nov 2025 13:56
Operator : SY/MD
Sample : VSTDICV010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825

Quant Time: Nov 19 01:34:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	Bromoform	10.000	9.085	9.1	86	0.00
50 T	4-Methyl-2-Pentanone	10.000	8.631	13.7	84	0.00
51 T	2-Hexanone	10.000	8.233	17.7	80	0.00
52 T	Tetrachloroethene	10.000	8.884	11.2	89	0.00
53 T	Toluene	10.000	9.197	8.0	90	0.00
54 T	1,2-Dibromoethane	10.000	8.705	13.0	84	0.00
55 I	Chlorobenzene-d5	10.000	10.000	0.0	92	0.00
56	1,1,1,2-Tetrachloroethane	10.000	8.969	10.3	90	0.00
57 T	Chlorobenzene	10.000	8.981	10.2	88	0.00
58 T	Ethyl Benzene	10.000	9.167	8.3	89	0.00
59 T	m/p-Xylene	20.000	18.108	9.5	90	0.00
60 T	o-Xylene	10.000	8.876	11.2	89	0.00
61 T	Styrene	10.000	8.714	12.9	84	0.00
62	Isopropylbenzene	10.000	8.816	11.8	88	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	8.421	15.8	85	0.00
64	n-propylbenzene	10.000	8.591	14.1	85	0.00
65	tert-Butylbenzene	10.000	8.465	15.4	86	0.00
66 T	Benzyl Chloride	10.000	5.159	48.4#	49	0.00
67	sec-Butylbenzene	10.000	8.388	16.1	85	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.048	-0.5	92	0.00
69	p-Isopropyltoluene	10.000	8.272	17.3	83	0.00
70	n-Butylbenzene	10.000	8.398	16.0	83	0.00
71	2-Chlorotoluene	10.000	8.649	13.5	86	0.00
72 T	4-Ethyltoluene	10.000	8.779	12.2	87	0.00
73 T	1,3,5-Trimethylbenzene	10.000	8.660	13.4	86	0.00
74 T	1,2,4-Trimethylbenzene	10.000	8.463	15.4	87	0.00
75 T	1,3-Dichlorobenzene	10.000	8.320	16.8	82	0.00
76 T	1,4-Dichlorobenzene	10.000	8.273	17.3	81	0.00
77 T	1,2-Dichlorobenzene	10.000	7.966	20.3	81	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	7.222	27.8	78	0.00
79 T	Naphthalene	10.000	8.270	17.3	78	0.00
80 T	Naphthalene,2-methyl-	10.000	8.631	13.7	80	0.00
81 T	1,2,4-Trichlorobenzene	10.000	8.497	15.0	79	0.00

(#) = Out of Range


SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043256.D
Acq On : 20 Nov 2025 08:41
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Wed Nov 19 01:11:25 2025



AutoFind: Scans 3175, 3176, 3177; Background Corrected with Scan 3162

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	8	40	25.5	29432	PASS
75	95	30	66	62.5	72149	PASS
95	95	100	100	100.0	115432	PASS
96	95	5	9	6.5	7511	PASS
173	174	0.00	2	1.0	742	PASS
174	95	50	120	63.3	73064	PASS
175	174	4	9	7.3	5336	PASS
176	174	93	101	96.5	70507	PASS
177	176	5	9	6.1	4312	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
 Data File : VL043257.D
 Acq On : 20 Nov 2025 10:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Nov 21 00:33:53 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/21/2025
 Supervised By : Mahesh Dadoda 11/21/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.777	49	102759	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.943	114	275595	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.866	117	222776	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.364 95 166532 10.198 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.000%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	161177	11.121	ppbv	98
3) Chlorodifluoromethane	1.473	51	146232	10.259	ppbv	95
4) Chloromethane	1.531	50	54380	10.035	ppbv	95
5) Vinyl Chloride	1.577	62	55244	9.322	ppbv	95
6) Bromomethane	1.664	94	27639	9.621	ppbv	99
7) Chloroethane	1.700	64	22384	9.698	ppbv	100
8) Dichlorotetrafluoroethane	1.551	85	129834	10.769	ppbv	99
9) Propene	1.480	41	55697m	9.202	ppbv	
10) Heptane	4.959	43	156253	8.674	ppbv	95
11) Trichlorofluoromethane	1.874	101	169670	12.083	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.133	101	126654	11.553	ppbv	97
13) Ethanol	1.716	45	16970	7.973	ppbv	95
14) Bromoethene	1.777	108	45797	10.897	ppbv	96
15) Acetone	1.829	43	122027	8.547	ppbv	99
16) 1,3-Butadiene	1.606	39	62334	9.794	ppbv	97
17) tert-Butyl alcohol	2.033	59	175626	9.372	ppbv	100
18) 1,1-Dichloroethene	2.030	96	55419	10.857	ppbv	88
19) Isopropyl Alcohol	1.881	45	72573	8.996	ppbv	94
20) Methylene Chloride	2.056	84	48499	11.659	ppbv	95
21) Allyl Chloride	2.091	41	83761	8.720	ppbv	93
22) trans-1,2-Dichloroethene	2.334	96	62490	10.711	ppbv	90
23) Vinyl Acetate	2.454	43	157670	9.148	ppbv	99
24) 1,1-Dichloroethane	2.402	63	119663	10.625	ppbv	97
25) Ethyl Acetate	2.823	43	285972	9.287	ppbv	99
26) Hexane	2.819	57	130004	9.630	ppbv	98
27) Carbon Disulfide	2.146	76	150498	10.920	ppbv	99
28) Methyl tert-Butyl Ether	2.431	73	68600	10.352	ppbv	95
29) Chloroform	2.839	83	219297	11.429	ppbv	99
30) Cyclohexane	3.845	84	109617	9.685	ppbv	95
31) cis-1,2-Dichloroethene	2.713	61	137895	10.277	ppbv	93
32) 1,1,1-Trichloroethane	3.357	97	226600	11.079	ppbv	99
34) 2-Butanone	2.548	43	180979	9.303	ppbv	100
35) Carbon Tetrachloride	3.752	117	237548	11.356	ppbv	98
36) Benzene	3.645	78	265981	10.078	ppbv	99
37) 1,2-Dichloroethane	3.208	62	173345	11.229	ppbv	98
38) Trichloroethene	4.528	130	107877	10.447	ppbv	94
39) 1,2-Dichloropropane	4.299	63	92453	9.571	ppbv	95
40) 1,4-Dioxane	4.558	88	36555	8.418	ppbv #	91
41) Tetrahydrofuran	3.040	42	94776	8.822	ppbv	98
42) Bromodichloromethane	4.470	83	237498	11.152	ppbv	99
43) Methyl Methacrylate	4.862	69	125480	8.957	ppbv	96
44) 2,2,4-Trimethylpentane	4.622	57	437108	9.643	ppbv	99
45) t-1,3-Dichloropropene	6.389	75	100231	7.675	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
 Data File : VL043257.D
 Acq On : 20 Nov 2025 10:14
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Nov 21 00:33:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/21/2025
 Supervised By : Mahesh Dadoda 11/21/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.577	75	122504	7.989	ppbv	98
47) 1,1,2-Trichloroethane	6.561	97	104460m	10.190	ppbv	
48) Dibromochloromethane	7.370	129	193819	10.997	ppbv	99
49) Bromoform	9.471	173	171644	10.994	ppbv	100
50) 4-Methyl-2-Pentanone	5.723	43	238084	8.688	ppbv	99
51) 2-Hexanone	7.415	43	172317	8.097	ppbv #	99
52) Tetrachloroethene	8.208	164	99129	10.097	ppbv	94
53) Toluene	6.914	91	300158	9.604	ppbv	99
54) 1,2-Dibromoethane	7.636	107	146099	9.684	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.911	131	176606	10.787	ppbv	98
57) Chlorobenzene	8.908	112	236347	10.333	ppbv	98
58) Ethyl Benzene	9.341	91	430385	10.037	ppbv	99
59) m/p-Xylene	9.539	91	709378m	20.635	ppbv	
60) o-Xylene	9.950	91	353166	10.284	ppbv	100
61) Styrene	9.859	104	140574	9.353	ppbv	99
62) Isopropylbenzene	10.529	105	637682	10.379	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.950	83	214523	9.817	ppbv	100
64) n-propylbenzene	10.979	120	158685	10.052	ppbv	97
65) tert-Butylbenzene	11.520	119	544959	9.932	ppbv	98
66) Benzyl Chloride	11.604	91	23441	4.210	ppbv	99
67) sec-Butylbenzene	11.756	105	764527	9.982	ppbv	100
69) p-Isopropyltoluene	11.914	119	641411	9.886	ppbv	100
70) n-Butylbenzene	12.270	91	626151	9.948	ppbv	99
71) 2-Chlorotoluene	10.888	91	466353	9.875	ppbv	99
72) 4-Ethyltoluene	11.115	105	424355	10.147	ppbv	98
73) 1,3,5-Trimethylbenzene	11.193	105	355138	10.119	ppbv	96
74) 1,2,4-Trimethylbenzene	11.526	105	386671	10.095	ppbv	99
75) 1,3-Dichlorobenzene	11.597	146	227896	10.196	ppbv	99
76) 1,4-Dichlorobenzene	11.662	146	226164	10.117	ppbv	100
77) 1,2-Dichlorobenzene	11.937	146	215578	10.032	ppbv	98
78) Hexachloro-1,3-Butadiene	13.873	225	172235	9.604	ppbv	100
79) Naphthalene	13.504	128	324391	9.536	ppbv	98
80) Naphthalene,2-methyl-	14.449	142	78637	8.806	ppbv	98
81) 1,2,4-Trichlorobenzene	13.429	180	160725	10.564	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043257.D
Acq On : 20 Nov 2025 10:14
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2025
Supervised By :Mahesh Dadoda 11/21/2025


Quant Time: Nov 21 00:33:53 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043257.D
Acq On : 20 Nov 2025 10:14
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Nov 21 00:33:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	1.410	1.568	-11.2	114	0.00
3	Chlorodifluoromethane	1.387	1.423	-2.6	106	0.00
4	Chloromethane	0.527	0.529	-0.4	104	0.00
5 T	Vinyl Chloride	0.577	0.538	6.8	108	0.00
6 T	Bromomethane	0.280	0.269	3.9	104	0.00
7	Chloroethane	0.225	0.218	3.1	110	0.00
8 T	Dichlorotetrafluoroethane	1.173	1.263	-7.7	113	0.00
9 T	Propene	0.589	0.542	8.0	106	0.00
10 T	Heptane	1.753	1.521	13.2	93	0.00
11 T	Trichlorofluoromethane	1.367	1.651	-20.8	130	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.067	1.233	-15.6	122	0.00
13	Ethanol	0.207	0.165	20.3	97	0.00
14 T	Bromoethene	0.409	0.446	-9.0	120	0.00
15 T	Acetone	1.389	1.188	14.5	113	0.00
16 T	1,3-Butadiene	0.619	0.607	1.9	108	0.00
17	tert-Butyl alcohol	1.824	1.709	6.3	98	0.00
18 T	1,1-Dichloroethene	0.497	0.539	-8.5	120	0.00
19 T	Isopropyl Alcohol	0.785	0.706	10.1	101	0.00
20 T	Methylene Chloride	0.600	0.472	21.3	116	0.00
21 T	Allyl Chloride	0.935	0.815	12.8	94	0.00
22 T	trans-1,2-Dichloroethene	0.568	0.608	-7.0	119	0.00
23 T	Vinyl Acetate	1.677	1.534	8.5	92	0.00
24 T	1,1-Dichloroethane	1.096	1.165	-6.3	112	0.00
25 T	Ethyl Acetate	2.996	2.783	7.1	99	0.00
26 T	Hexane	1.314	1.265	3.7	100	0.00
27 T	Carbon Disulfide	1.341	1.465	-9.2	114	0.00
28 T	Methyl tert-Butyl Ether	0.645	0.668	-3.6	110	0.00
29 T	Chloroform	1.867	2.134	-14.3	117	0.00
30 T	Cyclohexane	1.101	1.067	3.1	102	0.00
31 T	cis-1,2-Dichloroethene	1.306	1.342	-2.8	106	0.00
32 T	1,1,1-Trichloroethane	1.990	2.205	-10.8	114	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
34 T	2-Butanone	0.706	0.657	6.9	99	0.00
35 T	Carbon Tetrachloride	0.759	0.862	-13.6	118	0.00
36 T	Benzene	0.958	0.965	-0.7	104	0.00
37 T	1,2-Dichloroethane	0.560	0.629	-12.3	115	0.00
38 T	Trichloroethene	0.375	0.391	-4.3	108	0.00
39 T	1,2-Dichloropropane	0.351	0.335	4.6	97	0.00
40 T	1,4-Dioxane	0.158	0.133	15.8	90	0.00
41 T	Tetrahydrofuran	0.390	0.344	11.8	91	0.00
42 T	Bromodichloromethane	0.773	0.862	-11.5	112	0.00
43	Methyl Methacrylate	0.508	0.455	10.4	96	0.00
44 T	2,2,4-Trimethylpentane	1.645	1.586	3.6	99	0.00
45 T	t-1,3-Dichloropropene	0.474	0.364	23.2	78	0.00
46 T	cis-1,3-Dichloropropene	0.556	0.445	20.0	81	0.00
47 T	1,1,2-Trichloroethane	0.372	0.379	-1.9	107	0.00
48 T	Dibromochloromethane	0.640	0.703	-9.8	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043257.D
Acq On : 20 Nov 2025 10:14
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Nov 21 00:33:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.566	0.623	-10.1	109	0.00
50 T	4-Methyl-2-Pentanone	0.994	0.864	13.1	89	0.00
51 T	2-Hexanone	0.772	0.625	19.0	83	0.00
52 T	Tetrachloroethene	0.356	0.360	-1.1	106	0.00
53 T	Toluene	1.134	1.089	4.0	98	0.00
54 T	1,2-Dibromoethane	0.547	0.530	3.1	98	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
56	1,1,1,2-Tetrachloroethane	0.735	0.793	-7.9	111	0.00
57 T	Chlorobenzene	1.027	1.061	-3.3	104	0.00
58 T	Ethyl Benzene	1.925	1.932	-0.4	101	0.00
59 T	m/p-Xylene	1.543	1.592	-3.2	105	0.00
60 T	o-Xylene	1.542	1.585	-2.8	106	0.00
61 T	Styrene	0.675	0.631	6.5	92	0.00
62	Isopropylbenzene	2.758	2.862	-3.8	106	0.00
63 T	1,1,2,2-Tetrachloroethane	0.981	0.963	1.8	103	0.00
64	n-propylbenzene	0.709	0.712	-0.4	103	0.00
65	tert-Butylbenzene	2.463	2.446	0.7	104	0.00
66 T	Benzyl Chloride	0.250	0.105	58.0#	41	0.00
67	sec-Butylbenzene	3.438	3.432	0.2	105	0.00
68 S	1-Bromo-4-Fluorobenzene	0.733	0.748	-2.0	97	0.00
69	p-Isopropyltoluene	2.912	2.879	1.1	103	0.00
70	n-Butylbenzene	2.825	2.811	0.5	101	0.00
71	2-Chlorotoluene	2.120	2.093	1.3	101	0.00
72 T	4-Ethyltoluene	1.877	1.905	-1.5	103	0.00
73 T	1,3,5-Trimethylbenzene	1.575	1.594	-1.2	104	0.00
74 T	1,2,4-Trimethylbenzene	1.719	1.736	-1.0	106	0.00
75 T	1,3-Dichlorobenzene	1.003	1.023	-2.0	104	0.00
76 T	1,4-Dichlorobenzene	1.003	1.015	-1.2	103	0.00
77 T	1,2-Dichlorobenzene	0.965	0.968	-0.3	105	0.00
78 T	Hexachloro-1,3-Butadiene	0.805	0.773	4.0	107	0.00
79 T	Naphthalene	1.527	1.456	4.6	92	0.00
80 T	Naphthalene,2-methyl-	0.401	0.353	12.0	84	0.00
81 T	1,2,4-Trichlorobenzene	0.683	0.721	-5.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043257.D
Acq On : 20 Nov 2025 10:14
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Nov 21 00:33:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	10.000	11.121	-11.2	114	0.00
3	Chlorodifluoromethane	10.000	10.259	-2.6	106	0.00
4	Chloromethane	10.000	10.035	-0.4	104	0.00
5 T	Vinyl Chloride	10.000	9.322	6.8	108	0.00
6 T	Bromomethane	10.000	9.621	3.8	104	0.00
7	Chloroethane	10.000	9.698	3.0	110	0.00
8 T	Dichlorotetrafluoroethane	10.000	10.769	-7.7	113	0.00
9 T	Propene	10.000	9.202	8.0	106	0.00
10 T	Heptane	10.000	8.674	13.3	93	0.00
11 T	Trichlorofluoromethane	10.000	12.083	-20.8	130	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	11.553	-15.5	122	0.00
13	Ethanol	10.000	7.973	20.3	97	0.00
14 T	Bromoethene	10.000	10.897	-9.0	120	0.00
15 T	Acetone	10.000	8.547	14.5	113	0.00
16 T	1,3-Butadiene	10.000	9.794	2.1	108	0.00
17	tert-Butyl alcohol	10.000	9.372	6.3	98	0.00
18 T	1,1-Dichloroethene	10.000	10.857	-8.6	120	0.00
19 T	Isopropyl Alcohol	10.000	8.996	10.0	101	0.00
20 T	Methylene Chloride	10.000	11.659	-16.6	116	0.00
21 T	Allyl Chloride	10.000	8.720	12.8	94	0.00
22 T	trans-1,2-Dichloroethene	10.000	10.711	-7.1	119	0.00
23 T	Vinyl Acetate	10.000	9.148	8.5	92	0.00
24 T	1,1-Dichloroethane	10.000	10.625	-6.3	112	0.00
25 T	Ethyl Acetate	10.000	9.287	7.1	99	0.00
26 T	Hexane	10.000	9.630	3.7	100	0.00
27 T	Carbon Disulfide	10.000	10.920	-9.2	114	0.00
28 T	Methyl tert-Butyl Ether	10.000	10.352	-3.5	110	0.00
29 T	Chloroform	10.000	11.429	-14.3	117	0.00
30 T	Cyclohexane	10.000	9.685	3.1	102	0.00
31 T	cis-1,2-Dichloroethene	10.000	10.277	-2.8	106	0.00
32 T	1,1,1-Trichloroethane	10.000	11.079	-10.8	114	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	97	0.00
34 T	2-Butanone	10.000	9.303	7.0	99	0.00
35 T	Carbon Tetrachloride	10.000	11.356	-13.6	118	0.00
36 T	Benzene	10.000	10.078	-0.8	104	0.00
37 T	1,2-Dichloroethane	10.000	11.229	-12.3	115	0.00
38 T	Trichloroethene	10.000	10.447	-4.5	108	0.00
39 T	1,2-Dichloropropane	10.000	9.571	4.3	97	0.00
40 T	1,4-Dioxane	10.000	8.418	15.8	90	0.00
41 T	Tetrahydrofuran	10.000	8.822	11.8	91	0.00
42 T	Bromodichloromethane	10.000	11.152	-11.5	112	0.00
43	Methyl Methacrylate	10.000	8.957	10.4	96	0.00
44 T	2,2,4-Trimethylpentane	10.000	9.643	3.6	99	0.00
45 T	t-1,3-Dichloropropene	10.000	7.675	23.3	78	0.00
46 T	cis-1,3-Dichloropropene	10.000	7.989	20.1	81	0.00
47 T	1,1,2-Trichloroethane	10.000	10.190	-1.9	107	0.00
48 T	Dibromochloromethane	10.000	10.997	-10.0	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043257.D
Acq On : 20 Nov 2025 10:14
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Nov 21 00:33:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	Bromoform	10.000	10.994	-9.9	109	0.00
50 T	4-Methyl-2-Pentanone	10.000	8.688	13.1	89	0.00
51 T	2-Hexanone	10.000	8.097	19.0	83	0.00
52 T	Tetrachloroethene	10.000	10.097	-1.0	106	0.00
53 T	Toluene	10.000	9.604	4.0	98	0.00
54 T	1,2-Dibromoethane	10.000	9.684	3.2	98	0.00
55 I	Chlorobenzene-d5	10.000	10.000	0.0	95	0.00
56	1,1,1,2-Tetrachloroethane	10.000	10.787	-7.9	111	0.00
57 T	Chlorobenzene	10.000	10.333	-3.3	104	0.00
58 T	Ethyl Benzene	10.000	10.037	-0.4	101	0.00
59 T	m/p-Xylene	20.000	20.635	-3.2	105	0.00
60 T	o-Xylene	10.000	10.284	-2.8	106	0.00
61 T	Styrene	10.000	9.353	6.5	92	0.00
62	Isopropylbenzene	10.000	10.379	-3.8	106	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	9.817	1.8	103	0.00
64	n-propylbenzene	10.000	10.052	-0.5	103	0.00
65	tert-Butylbenzene	10.000	9.932	0.7	104	0.00
66 T	Benzyl Chloride	10.000	4.210	57.9#	41	0.00
67	sec-Butylbenzene	10.000	9.982	0.2	105	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.198	-2.0	97	0.00
69	p-Isopropyltoluene	10.000	9.886	1.1	103	0.00
70	n-Butylbenzene	10.000	9.948	0.5	101	0.00
71	2-Chlorotoluene	10.000	9.875	1.3	101	0.00
72 T	4-Ethyltoluene	10.000	10.147	-1.5	103	0.00
73 T	1,3,5-Trimethylbenzene	10.000	10.119	-1.2	104	0.00
74 T	1,2,4-Trimethylbenzene	10.000	10.095	-1.0	106	0.00
75 T	1,3-Dichlorobenzene	10.000	10.196	-2.0	104	0.00
76 T	1,4-Dichlorobenzene	10.000	10.117	-1.2	103	0.00
77 T	1,2-Dichlorobenzene	10.000	10.032	-0.3	105	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	9.604	4.0	107	0.00
79 T	Naphthalene	10.000	9.536	4.6	92	0.00
80 T	Naphthalene,2-methyl-	10.000	8.806	11.9	84	0.00
81 T	1,2,4-Trichlorobenzene	10.000	10.564	-5.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043258.D
Acq On : 20 Nov 2025 11:01
Operator : SY/MD
Sample : VL1120ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1120ABL01

Quant Time: Nov 21 00:35:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.781	49	105521	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	275399	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	211341	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.364	95	154256	9.958	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043258.D
Acq On : 20 Nov 2025 11:01
Operator : SY/MD
Sample : VL1120ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1120ABL01

Quant Time: Nov 21 00:35:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043259.D
Acq On : 20 Nov 2025 11:34
Operator : SY/MD
Sample : QC111725 CAN 10588
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC111725 CAN 10588

Quant Time: Nov 21 00:35:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	97406	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	258018	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	201755	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.370	95	144519	9.772	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.700%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043259.D
Acq On : 20 Nov 2025 11:34
Operator : SY/MD
Sample : QC111725 CAN 10588
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC111725 CAN 10588

Quant Time: Nov 21 00:35:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
 Data File : VL043260.D
 Acq On : 20 Nov 2025 12:07
 Operator : SY/MD
 Sample : VL1120ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1120ABS01

Quant Time: Nov 21 00:36:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/21/2025
 Supervised By : Mahesh Dadoda 11/21/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	101372	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	270113	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	214473	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.371 95 164533 10.466 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 104.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	164798	11.526	ppbv	99
3) Chlorodifluoromethane	1.476	51	142815	10.157	ppbv	95
4) Chloromethane	1.534	50	55072	10.302	ppbv	96
5) Vinyl Chloride	1.580	62	55657	9.520	ppbv	93
6) Bromomethane	1.667	94	25862	9.126	ppbv	82
7) Chloroethane	1.703	64	22017	9.670	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	131076	11.021	ppbv	98
9) Propene	1.483	41	53732m	8.999	ppbv	
10) Heptane	4.972	43	158256	8.906	ppbv	99
11) Trichlorofluoromethane	1.878	101	171548	12.384	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.140	101	126812	11.726	ppbv	99
13) Ethanol	1.719	45	17475	8.323	ppbv	99
14) Bromoethene	1.780	108	44833	10.813	ppbv	97
15) Acetone	1.832	43	119678	8.497	ppbv	100
16) 1,3-Butadiene	1.609	39	61711	9.828	ppbv	95
17) tert-Butyl alcohol	2.036	59	171332	9.268	ppbv	100
18) 1,1-Dichloroethene	2.033	96	53337	10.592	ppbv	98
19) Isopropyl Alcohol	1.884	45	68773	8.642	ppbv	90
20) Methylene Chloride	2.062	84	48243	11.763	ppbv	93
21) Allyl Chloride	2.094	41	83462	8.808	ppbv	92
22) trans-1,2-Dichloroethene	2.340	96	61279	10.647	ppbv	93
23) Vinyl Acetate	2.460	43	142709	8.393	ppbv	98
24) 1,1-Dichloroethane	2.408	63	121792	10.962	ppbv	98
25) Ethyl Acetate	2.829	43	281330	9.262	ppbv	99
26) Hexane	2.823	57	126532	9.501	ppbv	99
27) Carbon Disulfide	2.149	76	148765	10.942	ppbv	100
28) Methyl tert-Butyl Ether	2.434	73	64675	9.894	ppbv	98
29) Chloroform	2.845	83	220567	11.652	ppbv	96
30) Cyclohexane	3.852	84	107088	9.591	ppbv	95
31) cis-1,2-Dichloroethene	2.719	61	131746	9.953	ppbv	96
32) 1,1,1-Trichloroethane	3.363	97	225329	11.168	ppbv	99
34) 2-Butanone	2.551	43	177069	9.287	ppbv	100
35) Carbon Tetrachloride	3.758	117	236089	11.515	ppbv	98
36) Benzene	3.651	78	257481	9.954	ppbv	98
37) 1,2-Dichloroethane	3.214	62	171416	11.329	ppbv	97
38) Trichloroethene	4.541	130	107097	10.582	ppbv	93
39) 1,2-Dichloropropane	4.308	63	92063	9.724	ppbv	87
40) 1,4-Dioxane	4.570	88	35055	8.236	ppbv #	93
41) Tetrahydrofuran	3.046	42	92013	8.739	ppbv	100
42) Bromodichloromethane	4.480	83	232959	11.161	ppbv	100
43) Methyl Methacrylate	4.865	69	122481	8.920	ppbv	96
44) 2,2,4-Trimethylpentane	4.632	57	428558	9.646	ppbv	98
45) t-1,3-Dichloropropene	6.402	75	96138	7.511	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
 Data File : VL043260.D
 Acq On : 20 Nov 2025 12:07
 Operator : SY/MD
 Sample : VL1120ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1120ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2025
 Supervised By :Mahesh Dadoda 11/21/2025

Quant Time: Nov 21 00:36:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.590	75	117380	7.810	ppbv	98
47) 1,1,2-Trichloroethane	6.571	97	104185	10.369	ppbv	97
48) Dibromochloromethane	7.377	129	188165	10.892	ppbv	99
49) Bromoform	9.474	173	168175	10.991	ppbv	99
50) 4-Methyl-2-Pentanone	5.736	43	237956	8.859	ppbv	99
51) 2-Hexanone	7.425	43	170769	8.187	ppbv #	99
52) Tetrachloroethene	8.218	164	99982	10.391	ppbv	98
53) Toluene	6.927	91	294482	9.614	ppbv	100
54) 1,2-Dibromoethane	7.642	107	143721	9.720	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.917	131	176862	11.221	ppbv	99
57) Chlorobenzene	8.914	112	234816	10.664	ppbv	99
58) Ethyl Benzene	9.348	91	427546	10.357	ppbv	99
59) m/p-Xylene	9.542	91	708974m	21.422	ppbv	
60) o-Xylene	9.956	91	353595	10.695	ppbv	99
61) Styrene	9.862	104	135477	9.363	ppbv	98
62) Isopropylbenzene	10.532	105	634112	10.720	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.953	83	218173	10.371	ppbv	99
64) n-propylbenzene	10.982	120	159286	10.480	ppbv	97
65) tert-Butylbenzene	11.526	119	552915	10.467	ppbv	98
66) Benzyl Chloride	11.610	91	22396	4.178	ppbv	98
67) sec-Butylbenzene	11.762	105	779369	10.570	ppbv	99
69) p-Isopropyltoluene	11.921	119	649941	10.405	ppbv	100
70) n-Butylbenzene	12.277	91	622858	10.278	ppbv	100
71) 2-Chlorotoluene	10.895	91	469961	10.337	ppbv	100
72) 4-Ethyltoluene	11.118	105	423927	10.529	ppbv	98
73) 1,3,5-Trimethylbenzene	11.199	105	358599	10.613	ppbv	95
74) 1,2,4-Trimethylbenzene	11.532	105	392516	10.645	ppbv	94
75) 1,3-Dichlorobenzene	11.604	146	231654	10.765	ppbv	99
76) 1,4-Dichlorobenzene	11.665	146	229039	10.642	ppbv	100
77) 1,2-Dichlorobenzene	11.944	146	220159	10.642	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	171979	9.961	ppbv	100
79) Naphthalene	13.507	128	319763	9.764	ppbv	99
80) Naphthalene,2-methyl-	14.455	142	75079	8.734	ppbv	97
81) 1,2,4-Trichlorobenzene	13.436	180	157618	10.761	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112025\
Data File : VL043260.D
Acq On : 20 Nov 2025 12:07
Operator : SY/MD
Sample : VL1120ABS01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1120ABS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2025
Supervised By :Mahesh Dadoda 11/21/2025


Quant Time: Nov 21 00:36:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration




Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
 Data File : VL043277.D
 Acq On : 24 Nov 2025 08:53
 Operator : SY/MD
 Sample : BFB
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 Last Update : Wed Nov 19 01:11:25 2025



AutoFind: Scans 3174, 3175, 3176; Background Corrected with Scan 3159

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.3	29189	PASS
75	95	30	66	57.3	71715	PASS
95	95	100	100	100.0	125165	PASS
96	95	5	9	6.3	7928	PASS
173	174	0.00	2	0.8	642	PASS
174	95	50	120	64.5	80763	PASS
175	174	4	9	7.3	5908	PASS
176	174	93	101	96.3	77795	PASS
177	176	5	9	7.0	5416	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
 Data File : VL043278.D
 Acq On : 24 Nov 2025 12:27
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Nov 24 23:38:34 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/25/2025
 Supervised By : Mahesh Dadoda 11/25/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	107061	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	275969	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	215083	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 163444 10.367 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 103.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	165790	10.979	ppbv	100
3) Chlorodifluoromethane	1.476	51	154655	10.414	ppbv	96
4) Chloromethane	1.531	50	57140	10.121	ppbv	92
5) Vinyl Chloride	1.580	62	57379	9.293	ppbv	94
6) Bromomethane	1.667	94	28470	9.512	ppbv	95
7) Chloroethane	1.703	64	23590	9.810	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	137737	10.965	ppbv	99
9) Propene	1.482	41	60577m	9.606	ppbv	
10) Heptane	4.968	43	170551	9.087	ppbv	99
11) Trichlorofluoromethane	1.877	101	175925	12.025	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.136	101	130475	11.423	ppbv	96
13) Ethanol	1.719	45	17888	8.067	ppbv	95
14) Bromoethene	1.780	108	47510	10.850	ppbv	96
15) Acetone	1.832	43	125343	8.427	ppbv	99
16) 1,3-Butadiene	1.609	39	64565	9.736	ppbv	96
17) tert-Butyl alcohol	2.036	59	188025	9.630	ppbv	100
18) 1,1-Dichloroethene	2.033	96	57383	10.790	ppbv	98
19) Isopropyl Alcohol	1.884	45	75837	9.023	ppbv	93
20) Methylene Chloride	2.059	84	52345	12.108	ppbv	91
21) Allyl Chloride	2.094	41	88510	8.844	ppbv	94
22) trans-1,2-Dichloroethene	2.337	96	64226	10.566	ppbv	99
23) Vinyl Acetate	2.460	43	162160	9.030	ppbv #	99
24) 1,1-Dichloroethane	2.405	63	126001	10.739	ppbv	98
25) Ethyl Acetate	2.829	43	302952	9.443	ppbv	99
26) Hexane	2.822	57	136961	9.738	ppbv	98
27) Carbon Disulfide	2.149	76	160712	11.192	ppbv	99
28) Methyl tert-Butyl Ether	2.434	73	69559	10.075	ppbv	94
29) Chloroform	2.845	83	232873	11.649	ppbv	100
30) Cyclohexane	3.852	84	115484	9.793	ppbv	95
31) cis-1,2-Dichloroethene	2.716	61	143927	10.295	ppbv	98
32) 1,1,1-Trichloroethane	3.360	97	239307	11.230	ppbv	99
34) 2-Butanone	2.551	43	193336	9.925	ppbv	99
35) Carbon Tetrachloride	3.758	117	251772	12.019	ppbv	98
36) Benzene	3.651	78	281056	10.634	ppbv	98
37) 1,2-Dichloroethane	3.211	62	179931	11.639	ppbv	98
38) Trichloroethene	4.538	130	112004	10.832	ppbv	95
39) 1,2-Dichloropropane	4.305	63	100819	10.423	ppbv	89
40) 1,4-Dioxane	4.570	88	37926	8.722	ppbv #	97
41) Tetrahydrofuran	3.046	42	100803	9.370	ppbv	97
42) Bromodichloromethane	4.480	83	247223	11.593	ppbv	98
43) Methyl Methacrylate	4.868	69	133803	9.538	ppbv	94
44) 2,2,4-Trimethylpentane	4.632	57	465164	10.248	ppbv	98
45) t-1,3-Dichloropropene	6.405	75	104473	7.989	ppbv	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
 Data File : VL043278.D
 Acq On : 24 Nov 2025 12:27
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Nov 24 23:38:34 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/25/2025
 Supervised By : Mahesh Dadoda 11/25/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.590	75	128727	8.383	ppbv	93
47) 1,1,2-Trichloroethane	6.567	97	112927	11.001	ppbv	97
48) Dibromochloromethane	7.376	129	204344	11.578	ppbv	100
49) Bromoform	9.474	173	180678	11.557	ppbv	99
50) 4-Methyl-2-Pentanone	5.729	43	252700	9.209	ppbv	100
51) 2-Hexanone	7.422	43	183438	8.608	ppbv #	99
52) Tetrachloroethene	8.218	164	102830	10.460	ppbv	91
53) Toluene	6.923	91	312291	9.979	ppbv	99
54) 1,2-Dibromoethane	7.642	107	152418	10.090	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.917	131	189737	12.003	ppbv	99
57) Chlorobenzene	8.914	112	249546	11.300	ppbv	97
58) Ethyl Benzene	9.348	91	448888	10.843	ppbv	97
59) m/p-Xylene	9.545	91	735995m	22.175	ppbv	
60) o-Xylene	9.956	91	373234	11.257	ppbv	100
61) Styrene	9.862	104	142379	9.812	ppbv	99
62) Isopropylbenzene	10.532	105	663175	11.180	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.956	83	229515	10.879	ppbv	100
64) n-propylbenzene	10.982	120	162848	10.684	ppbv	93
65) tert-Butylbenzene	11.526	119	575690	10.867	ppbv	99
66) Benzyl Chloride	11.610	91	27757	5.163	ppbv	98
67) sec-Butylbenzene	11.762	105	808129	10.929	ppbv	99
69) p-Isopropyltoluene	11.921	119	665442	10.623	ppbv	99
70) n-Butylbenzene	12.277	91	647136	10.649	ppbv	100
71) 2-Chlorotoluene	10.895	91	495463	10.867	ppbv	99
72) 4-Ethyltoluene	11.118	105	442267	10.953	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	372473	10.992	ppbv	96
74) 1,2,4-Trimethylbenzene	11.532	105	409684	11.079	ppbv	90
75) 1,3-Dichlorobenzene	11.600	146	242050	11.217	ppbv	99
76) 1,4-Dichlorobenzene	11.665	146	239021	11.074	ppbv	99
77) 1,2-Dichlorobenzene	11.940	146	231049	11.136	ppbv	99
78) Hexachloro-1,3-Butadiene	13.876	225	181252	10.468	ppbv	99
79) Naphthalene	13.507	128	339967	10.351	ppbv	99
80) Naphthalene,2-methyl-	14.452	142	85751	9.947	ppbv	99
81) 1,2,4-Trichlorobenzene	13.435	180	170083	11.579	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043278.D
Acq On : 24 Nov 2025 12:27
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Nov 24 23:38:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	1.410	1.549	-9.9	118	0.00
3	Chlorodifluoromethane	1.387	1.445	-4.2	112	0.00
4	Chloromethane	0.527	0.534	-1.3	109	0.00
5 T	Vinyl Chloride	0.577	0.536	7.1	112	0.00
6 T	Bromomethane	0.280	0.266	5.0	107	0.00
7	Chloroethane	0.225	0.220	2.2	116	0.00
8 T	Dichlorotetrafluoroethane	1.173	1.287	-9.7	120	0.00
9 T	Propene	0.589	0.566	3.9	115	0.00
10 T	Heptane	1.753	1.593	9.1	101	0.00
11 T	Trichlorofluoromethane	1.367	1.643	-20.2	135	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.067	1.219	-14.2	126	0.00
13	Ethanol	0.207	0.167	19.3	102	0.00
14 T	Bromoethene	0.409	0.444	-8.6	125	0.00
15 T	Acetone	1.389	1.171	15.7	116	0.00
16 T	1,3-Butadiene	0.619	0.603	2.6	112	0.00
17	tert-Butyl alcohol	1.824	1.756	3.7	105	0.00
18 T	1,1-Dichloroethene	0.497	0.536	-7.8	124	0.00
19 T	Isopropyl Alcohol	0.785	0.708	9.8	106	0.00
20 T	Methylene Chloride	0.600	0.489	18.5	125	0.00
21 T	Allyl Chloride	0.935	0.827	11.6	99	0.00
22 T	trans-1,2-Dichloroethene	0.568	0.600	-5.6	122	0.00
23 T	Vinyl Acetate	1.677	1.515	9.7	95	0.00
24 T	1,1-Dichloroethane	1.096	1.177	-7.4	118	0.00
25 T	Ethyl Acetate	2.996	2.830	5.5	104	0.00
26 T	Hexane	1.314	1.279	2.7	106	0.00
27 T	Carbon Disulfide	1.341	1.501	-11.9	122	0.00
28 T	Methyl tert-Butyl Ether	0.645	0.650	-0.8	112	0.00
29 T	Chloroform	1.867	2.175	-16.5	124	0.00
30 T	Cyclohexane	1.101	1.079	2.0	107	0.00
31 T	cis-1,2-Dichloroethene	1.306	1.344	-2.9	110	0.00
32 T	1,1,1-Trichloroethane	1.990	2.235	-12.3	121	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
34 T	2-Butanone	0.706	0.701	0.7	106	0.00
35 T	Carbon Tetrachloride	0.759	0.912	-20.2	125	0.00
36 T	Benzene	0.958	1.018	-6.3	109	0.00
37 T	1,2-Dichloroethane	0.560	0.652	-16.4	119	0.00
38 T	Trichloroethene	0.375	0.406	-8.3	112	0.00
39 T	1,2-Dichloropropane	0.351	0.365	-4.0	106	0.00
40 T	1,4-Dioxane	0.158	0.137	13.3	94	0.00
41 T	Tetrahydrofuran	0.390	0.365	6.4	97	0.00
42 T	Bromodichloromethane	0.773	0.896	-15.9	117	0.00
43	Methyl Methacrylate	0.508	0.485	4.5	102	0.00
44 T	2,2,4-Trimethylpentane	1.645	1.686	-2.5	105	0.00
45 T	t-1,3-Dichloropropene	0.474	0.379	20.0	81	0.00
46 T	cis-1,3-Dichloropropene	0.556	0.466	16.2	85	0.00
47 T	1,1,2-Trichloroethane	0.372	0.409	-9.9	116	0.00
48 T	Dibromochloromethane	0.640	0.740	-15.6	116	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043278.D
Acq On : 24 Nov 2025 12:27
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Nov 24 23:38:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.566	0.655	-15.7	115	0.00
50 T	4-Methyl-2-Pentanone	0.994	0.916	7.8	94	0.00
51 T	2-Hexanone	0.772	0.665	13.9	88	0.00
52 T	Tetrachloroethene	0.356	0.373	-4.8	110	0.00
53 T	Toluene	1.134	1.132	0.2	102	0.00
54 T	1,2-Dibromoethane	0.547	0.552	-0.9	102	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
56	1,1,1,2-Tetrachloroethane	0.735	0.882	-20.0	119	0.00
57 T	Chlorobenzene	1.027	1.160	-13.0	110	0.00
58 T	Ethyl Benzene	1.925	2.087	-8.4	105	0.00
59 T	m/p-Xylene	1.543	1.711	-10.9	109	0.00
60 T	o-Xylene	1.542	1.735	-12.5	112	0.00
61 T	Styrene	0.675	0.662	1.9	94	0.00
62	Isopropylbenzene	2.758	3.083	-11.8	111	0.00
63 T	1,1,2,2-Tetrachloroethane	0.981	1.067	-8.8	110	0.00
64	n-propylbenzene	0.709	0.757	-6.8	106	0.00
65	tert-Butylbenzene	2.463	2.677	-8.7	110	0.00
66 T	Benzyl Chloride	0.250	0.129	48.4#	49	0.00
67	sec-Butylbenzene	3.438	3.757	-9.3	111	0.00
68 S	1-Bromo-4-Fluorobenzene	0.733	0.760	-3.7	95	0.00
69	p-Isopropyltoluene	2.912	3.094	-6.2	107	0.00
70	n-Butylbenzene	2.825	3.009	-6.5	104	0.00
71	2-Chlorotoluene	2.120	2.304	-8.7	108	0.00
72 T	4-Ethyltoluene	1.877	2.056	-9.5	108	0.00
73 T	1,3,5-Trimethylbenzene	1.575	1.732	-10.0	109	0.00
74 T	1,2,4-Trimethylbenzene	1.719	1.905	-10.8	113	0.00
75 T	1,3-Dichlorobenzene	1.003	1.125	-12.2	110	0.00
76 T	1,4-Dichlorobenzene	1.003	1.111	-10.8	108	0.00
77 T	1,2-Dichlorobenzene	0.965	1.074	-11.3	113	0.00
78 T	Hexachloro-1,3-Butadiene	0.805	0.843	-4.7	113	0.00
79 T	Naphthalene	1.527	1.581	-3.5	97	0.00
80 T	Naphthalene,2-methyl-	0.401	0.399	0.5	92	0.00
81 T	1,2,4-Trichlorobenzene	0.683	0.791	-15.8	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043278.D
Acq On : 24 Nov 2025 12:27
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Nov 24 23:38:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	10.000	10.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	10.000	10.979	-9.8	118	0.00
3	Chlorodifluoromethane	10.000	10.414	-4.1	112	0.00
4	Chloromethane	10.000	10.121	-1.2	109	0.00
5 T	Vinyl Chloride	10.000	9.293	7.1	112	0.00
6 T	Bromomethane	10.000	9.512	4.9	107	0.00
7	Chloroethane	10.000	9.810	1.9	116	0.00
8 T	Dichlorotetrafluoroethane	10.000	10.965	-9.6	120	0.00
9 T	Propene	10.000	9.606	3.9	115	0.00
10 T	Heptane	10.000	9.087	9.1	101	0.00
11 T	Trichlorofluoromethane	10.000	12.025	-20.3	135	0.00
12 T	1,1,2-Trichlorotrifluoroeth	10.000	11.423	-14.2	126	0.00
13	Ethanol	10.000	8.067	19.3	102	0.00
14 T	Bromoethene	10.000	10.850	-8.5	125	0.00
15 T	Acetone	10.000	8.427	15.7	116	0.00
16 T	1,3-Butadiene	10.000	9.736	2.6	112	0.00
17	tert-Butyl alcohol	10.000	9.630	3.7	105	0.00
18 T	1,1-Dichloroethene	10.000	10.790	-7.9	124	0.00
19 T	Isopropyl Alcohol	10.000	9.023	9.8	106	0.00
20 T	Methylene Chloride	10.000	12.108	-21.1	125	0.00
21 T	Allyl Chloride	10.000	8.844	11.6	99	0.00
22 T	trans-1,2-Dichloroethene	10.000	10.566	-5.7	122	0.00
23 T	Vinyl Acetate	10.000	9.030	9.7	95	0.00
24 T	1,1-Dichloroethane	10.000	10.739	-7.4	118	0.00
25 T	Ethyl Acetate	10.000	9.443	5.6	104	0.00
26 T	Hexane	10.000	9.738	2.6	106	0.00
27 T	Carbon Disulfide	10.000	11.192	-11.9	122	0.00
28 T	Methyl tert-Butyl Ether	10.000	10.075	-0.7	112	0.00
29 T	Chloroform	10.000	11.649	-16.5	124	0.00
30 T	Cyclohexane	10.000	9.793	2.1	107	0.00
31 T	cis-1,2-Dichloroethene	10.000	10.295	-2.9	110	0.00
32 T	1,1,1-Trichloroethane	10.000	11.230	-12.3	121	0.00
33 I	1,4-Difluorobenzene	10.000	10.000	0.0	97	0.00
34 T	2-Butanone	10.000	9.925	0.7	106	0.00
35 T	Carbon Tetrachloride	10.000	12.019	-20.2	125	0.00
36 T	Benzene	10.000	10.634	-6.3	109	0.00
37 T	1,2-Dichloroethane	10.000	11.639	-16.4	119	0.00
38 T	Trichloroethene	10.000	10.832	-8.3	112	0.00
39 T	1,2-Dichloropropane	10.000	10.423	-4.2	106	0.00
40 T	1,4-Dioxane	10.000	8.722	12.8	94	0.00
41 T	Tetrahydrofuran	10.000	9.370	6.3	97	0.00
42 T	Bromodichloromethane	10.000	11.593	-15.9	117	0.00
43	Methyl Methacrylate	10.000	9.538	4.6	102	0.00
44 T	2,2,4-Trimethylpentane	10.000	10.248	-2.5	105	0.00
45 T	t-1,3-Dichloropropene	10.000	7.989	20.1	81	0.00
46 T	cis-1,3-Dichloropropene	10.000	8.383	16.2	85	0.00
47 T	1,1,2-Trichloroethane	10.000	11.001	-10.0	116	0.00
48 T	Dibromochloromethane	10.000	11.578	-15.8	116	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043278.D
Acq On : 24 Nov 2025 12:27
Operator : SY/MD
Sample : VSTDCCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
LabSampleId :
VSTDCCC010

Quant Time: Nov 24 23:38:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	Bromoform	10.000	11.557	-15.6	115	0.00
50 T	4-Methyl-2-Pentanone	10.000	9.209	7.9	94	0.00
51 T	2-Hexanone	10.000	8.608	13.9	88	0.00
52 T	Tetrachloroethene	10.000	10.460	-4.6	110	0.00
53 T	Toluene	10.000	9.979	0.2	102	0.00
54 T	1,2-Dibromoethane	10.000	10.090	-0.9	102	0.00
55 I	Chlorobenzene-d5	10.000	10.000	0.0	92	0.00
56	1,1,1,2-Tetrachloroethane	10.000	12.003	-20.0	119	0.00
57 T	Chlorobenzene	10.000	11.300	-13.0	110	0.00
58 T	Ethyl Benzene	10.000	10.843	-8.4	105	0.00
59 T	m/p-Xylene	20.000	22.175	-10.9	109	0.00
60 T	o-Xylene	10.000	11.257	-12.6	112	0.00
61 T	Styrene	10.000	9.812	1.9	94	0.00
62	Isopropylbenzene	10.000	11.180	-11.8	111	0.00
63 T	1,1,2,2-Tetrachloroethane	10.000	10.879	-8.8	110	0.00
64	n-propylbenzene	10.000	10.684	-6.8	106	0.00
65	tert-Butylbenzene	10.000	10.867	-8.7	110	0.00
66 T	Benzyl Chloride	10.000	5.163	48.4#	49	0.00
67	sec-Butylbenzene	10.000	10.929	-9.3	111	0.00
68 S	1-Bromo-4-Fluorobenzene	10.000	10.367	-3.7	95	0.00
69	p-Isopropyltoluene	10.000	10.623	-6.2	107	0.00
70	n-Butylbenzene	10.000	10.649	-6.5	104	0.00
71	2-Chlorotoluene	10.000	10.867	-8.7	108	0.00
72 T	4-Ethyltoluene	10.000	10.953	-9.5	108	0.00
73 T	1,3,5-Trimethylbenzene	10.000	10.992	-9.9	109	0.00
74 T	1,2,4-Trimethylbenzene	10.000	11.079	-10.8	113	0.00
75 T	1,3-Dichlorobenzene	10.000	11.217	-12.2	110	0.00
76 T	1,4-Dichlorobenzene	10.000	11.074	-10.7	108	0.00
77 T	1,2-Dichlorobenzene	10.000	11.136	-11.4	113	0.00
78 T	Hexachloro-1,3-Butadiene	10.000	10.468	-4.7	113	0.00
79 T	Naphthalene	10.000	10.351	-3.5	97	0.00
80 T	Naphthalene,2-methyl-	10.000	9.947	0.5	92	0.00
81 T	1,2,4-Trichlorobenzene	10.000	11.579	-15.8	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043279.D
Acq On : 24 Nov 2025 13:11
Operator : SY/MD
Sample : VL1124ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1124ABL01

Quant Time: Nov 24 23:39:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	106137	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	277570	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	209566	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	151293	9.849	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.500%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043279.D
Acq On : 24 Nov 2025 13:11
Operator : SY/MD
Sample : VL1124ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1124ABL01

Quant Time: Nov 24 23:39:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
 Data File : VL043280.D
 Acq On : 24 Nov 2025 13:52
 Operator : SY/MD
 Sample : VL1124ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1124ABS01

Quant Time: Nov 24 23:39:50 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/25/2025
 Supervised By : Mahesh Dadoda 11/25/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	105147	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	272854	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	221036	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.367 95 164851 10.175 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	162546	10.960	ppbv	99
3) Chlorodifluoromethane	1.476	51	149762	10.268	ppbv	95
4) Chloromethane	1.531	50	55864	10.075	ppbv	96
5) Vinyl Chloride	1.580	62	57122	9.420	ppbv	98
6) Bromomethane	1.667	94	28296	9.626	ppbv	98
7) Chloroethane	1.703	64	22649	9.590	ppbv	93
8) Dichlorotetrafluoroethane	1.554	85	135608	10.992	ppbv	98
9) Propene	1.483	41	58566m	9.456	ppbv	
10) Heptane	4.969	43	170817	9.267	ppbv	99
11) Trichlorofluoromethane	1.875	101	169809	11.818	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.137	101	132086	11.775	ppbv	99
13) Ethanol	1.719	45	17266	7.928	ppbv	97
14) Bromoethene	1.781	108	46318	10.771	ppbv	98
15) Acetone	1.832	43	126744	8.676	ppbv	98
16) 1,3-Butadiene	1.609	39	64230	9.862	ppbv	96
17) tert-Butyl alcohol	2.036	59	179940	9.384	ppbv	100
18) 1,1-Dichloroethene	2.030	96	55652	10.655	ppbv	92
19) Isopropyl Alcohol	1.884	45	71853	8.704	ppbv	92
20) Methylene Chloride	2.059	84	50105	11.780	ppbv	91
21) Allyl Chloride	2.095	41	88132	8.967	ppbv	96
22) trans-1,2-Dichloroethene	2.337	96	64065	10.731	ppbv	98
23) Vinyl Acetate	2.457	43	171121	9.703	ppbv #	99
24) 1,1-Dichloroethane	2.405	63	124688	10.820	ppbv	98
25) Ethyl Acetate	2.826	43	301985	9.585	ppbv	99
26) Hexane	2.820	57	135935	9.841	ppbv	99
27) Carbon Disulfide	2.146	76	158318	11.226	ppbv	100
28) Methyl tert-Butyl Ether	2.434	73	75424	11.124	ppbv	95
29) Chloroform	2.842	83	228576	11.642	ppbv	98
30) Cyclohexane	3.852	84	116565	10.065	ppbv	93
31) cis-1,2-Dichloroethene	2.716	61	143211	10.430	ppbv	98
32) 1,1,1-Trichloroethane	3.360	97	235081	11.233	ppbv	99
34) 2-Butanone	2.551	43	193105	10.026	ppbv	99
35) Carbon Tetrachloride	3.755	117	246117	11.883	ppbv	99
36) Benzene	3.648	78	279506	10.697	ppbv	98
37) 1,2-Dichloroethane	3.211	62	178873	11.703	ppbv	98
38) Trichloroethene	4.538	130	112238	10.979	ppbv	96
39) 1,2-Dichloropropane	4.302	63	101403	10.603	ppbv	93
40) 1,4-Dioxane	4.564	88	38115	8.866	ppbv #	94
41) Tetrahydrofuran	3.043	42	103010	9.685	ppbv	98
42) Bromodichloromethane	4.477	83	246932	11.712	ppbv	95
43) Methyl Methacrylate	4.868	69	132570	9.558	ppbv	96
44) 2,2,4-Trimethylpentane	4.632	57	459765	10.245	ppbv	99
45) t-1,3-Dichloropropene	6.403	75	103648	8.017	ppbv	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
 Data File : VL043280.D
 Acq On : 24 Nov 2025 13:52
 Operator : SY/MD
 Sample : VL1124ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1124ABS01

Quant Time: Nov 24 23:39:50 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/25/2025
 Supervised By : Mahesh Dadoda 11/25/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.584	75	125652	8.277	ppbv	94
47) 1,1,2-Trichloroethane	6.568	97	111403	10.976	ppbv	92
48) Dibromochloromethane	7.377	129	203117	11.640	ppbv	98
49) Bromoform	9.474	173	179808	11.633	ppbv	99
50) 4-Methyl-2-Pentanone	5.729	43	257888	9.505	ppbv	98
51) 2-Hexanone	7.422	43	184928	8.777	ppbv #	98
52) Tetrachloroethene	8.212	164	102674	10.563	ppbv	94
53) Toluene	6.920	91	315386	10.193	ppbv	99
54) 1,2-Dibromoethane	7.642	107	154824	10.366	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.914	131	190059	11.700	ppbv	98
57) Chlorobenzene	8.914	112	250652	11.045	ppbv	98
58) Ethyl Benzene	9.348	91	451322	10.609	ppbv	98
59) m/p-Xylene	9.523	91	743261m	21.791	ppbv	
60) o-Xylene	9.956	91	367807	10.794	ppbv	99
61) Styrene	9.863	104	142907	9.583	ppbv	99
62) Isopropylbenzene	10.533	105	669597	10.984	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.953	83	227409	10.489	ppbv	99
64) n-propylbenzene	10.982	120	167620	10.701	ppbv	99
65) tert-Butylbenzene	11.523	119	573049	10.526	ppbv	99
66) Benzyl Chloride	11.610	91	25332	4.585	ppbv	99
67) sec-Butylbenzene	11.759	105	802399	10.559	ppbv	99
69) p-Isopropyltoluene	11.918	119	666208	10.349	ppbv	100
70) n-Butylbenzene	12.274	91	653544	10.465	ppbv	99
71) 2-Chlorotoluene	10.895	91	492497	10.511	ppbv	100
72) 4-Ethyltoluene	11.118	105	448275	10.803	ppbv	98
73) 1,3,5-Trimethylbenzene	11.196	105	372510	10.697	ppbv	96
74) 1,2,4-Trimethylbenzene	11.529	105	413769	10.888	ppbv	96
75) 1,3-Dichlorobenzene	11.601	146	239632	10.806	ppbv	97
76) 1,4-Dichlorobenzene	11.665	146	239094	10.779	ppbv	98
77) 1,2-Dichlorobenzene	11.940	146	227478	10.669	ppbv	99
78) Hexachloro-1,3-Butadiene	13.876	225	183653	10.321	ppbv	99
79) Naphthalene	13.507	128	340028	10.074	ppbv	100
80) Naphthalene,2-methyl-	14.452	142	83577	9.433	ppbv	100
81) 1,2,4-Trichlorobenzene	13.433	180	167890	11.122	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043290.D
Acq On : 25 Nov 2025 08:21
Operator : SY/MD
Sample : QC112425 CAN 10315
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC112425 CAN 10315

Quant Time: Nov 26 00:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	109733	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	274585	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	212841	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.370	95	147790	9.473	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.700%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112425\
Data File : VL043290.D
Acq On : 25 Nov 2025 08:21
Operator : SY/MD
Sample : QC112425 CAN 10315
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC112425 CAN 10315

Quant Time: Nov 26 00:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration



Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL111825

Review By	Semsettin Yesilyurt	Review On	11/19/2025 8:33:42 AM		
Supervise By	Maresh Dadoda	Supervise On	11/19/2025 9:44:32 AM		
SubDirectory	VL111825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2691			
Initial Calibration Stds		AP2694,AP2695,AP2696			
CCC		AP2694			
Internal Standard/PEM		AP2691			
ICV/I.BLK		AP2697			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL043202.D	18 Nov 2025 08:46	SY/MD	Ok
2	VSTDIC0.5	VL043203.D	18 Nov 2025 09:29	SY/MD	Ok,M
3	VSTDICCC010	VL043204.D	18 Nov 2025 10:02	SY/MD	Ok,M
4	VSTDIC002	VL043205.D	18 Nov 2025 10:37	SY/MD	Ok,M
5	VSTDIC001	VL043206.D	18 Nov 2025 11:10	SY/MD	Ok,M
6	VSTDIC0.5	VL043207.D	18 Nov 2025 11:43	SY/MD	Not Ok
7	VSTDIC0.1	VL043208.D	18 Nov 2025 12:16	SY/MD	Ok,M
8	VSTDIC0.03	VL043209.D	18 Nov 2025 12:49	SY/MD	Ok,M
9	VSTDIC015	VL043210.D	18 Nov 2025 13:24	SY/MD	Ok,M
10	VSTDICV010	VL043211.D	18 Nov 2025 13:56	SY/MD	Ok,M
11	VL1118ABL01	VL043212.D	18 Nov 2025 14:41	SY/MD	Ok
12	VL1118ABS01	VL043213.D	18 Nov 2025 15:24	SY/MD	Ok,M
13	Q3651-03	VL043214.D	18 Nov 2025 15:57	SY/MD	Dilution
14	Q3651-03DUP	VL043215.D	18 Nov 2025 16:31	SY/MD	Ok,M
15	Q3651-04	VL043216.D	18 Nov 2025 17:06	SY/MD	Dilution
16	Q3651-06	VL043217.D	18 Nov 2025 17:40	SY/MD	Dilution
17	Q3650-03	VL043218.D	18 Nov 2025 18:15	SY/MD	Ok,M
18	Q3650-04	VL043219.D	18 Nov 2025 18:49	SY/MD	Ok,M
19	Q3650-05	VL043220.D	18 Nov 2025 19:23	SY/MD	Ok,M
20	Q3650-06	VL043221.D	18 Nov 2025 19:58	SY/MD	Ok,M
21	Q3651-01	VL043222.D	18 Nov 2025 20:29	SY/MD	Ok,M

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL111825

Review By	Semsettin Yesilyurt	Review On	11/19/2025 8:33:42 AM		
Supervise By	Maresh Dadoda	Supervise On	11/19/2025 9:44:32 AM		
SubDirectory	VL111825	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2691			
Initial Calibration Stds		AP2694,AP2695,AP2696			
CCC		AP2694			
Internal Standard/PEM		AP2691			
ICV/I.BLK		AP2697			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q3651-02	VL043223.D	18 Nov 2025 21:01	SY/MD	Ok,M
23	Q3651-05	VL043224.D	18 Nov 2025 21:33	SY/MD	Ok,M
24	Q3650-01	VL043225.D	18 Nov 2025 22:05	SY/MD	Ok,M
25	Q3650-02	VL043226.D	18 Nov 2025 22:37	SY/MD	Ok,M
26	Q3651-03DL	VL043227.D	18 Nov 2025 23:09	SY/MD	Ok,M
27	Q3651-03DL	VL043228.D	18 Nov 2025 23:41	SY/MD	Not Ok
28	Q3651-04DL	VL043229.D	19 Nov 2025 00:13	SY/MD	Ok,M
29	Q3651-06DL	VL043230.D	19 Nov 2025 00:44	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL112025

Review By	Amit Patel	Review On	11/21/2025 11:53:43 AM		
Supervise By	Mahesh Dadoda	Supervise On	11/21/2025 12:23:05 PM		
SubDirectory	VL112025	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk	AP2691				
Initial Calibration Stds	AP2694,AP2695,AP2696				
CCC	AP2694				
Internal Standard/PEM	AP2691				
ICV/I.BLK	AP2697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL043256.D	20 Nov 2025 08:41	SY/MD	Ok
2	VSTDCCC010	VL043257.D	20 Nov 2025 10:14	SY/MD	Ok,M
3	VL1120ABL01	VL043258.D	20 Nov 2025 11:01	SY/MD	Ok
4	QC111725 CAN 10588	VL043259.D	20 Nov 2025 11:34	SY/MD	Ok
5	VL1120ABS01	VL043260.D	20 Nov 2025 12:07	SY/MD	Ok,M
6	Q3655-01	VL043261.D	20 Nov 2025 12:49	SY/MD	Ok,M
7	Q3655-02	VL043262.D	20 Nov 2025 13:24	SY/MD	Dilution
8	Q3655-03	VL043263.D	20 Nov 2025 13:58	SY/MD	Dilution
9	Q3655-04	VL043264.D	20 Nov 2025 14:32	SY/MD	Ok,M
10	Q3655-05	VL043265.D	20 Nov 2025 15:06	SY/MD	Ok,M
11	Q3655-06	VL043266.D	20 Nov 2025 15:40	SY/MD	Ok,M
12	Q3655-01DUP	VL043267.D	20 Nov 2025 16:20	SY/MD	Ok,M
13	Q3655-02DL	VL043268.D	20 Nov 2025 16:54	SY/MD	Ok,M
14	Q3655-03DL	VL043269.D	20 Nov 2025 17:27	SY/MD	Ok,M
15	Q3698-03	VL043270.D	20 Nov 2025 18:02	SY/MD	Dilution
16	Q3698-04	VL043271.D	20 Nov 2025 18:38	SY/MD	Ok,M
17	Q3698-01	VL043272.D	20 Nov 2025 19:11	SY/MD	Ok,M
18	Q3698-02	VL043273.D	20 Nov 2025 19:44	SY/MD	Ok,M
19	Q3698-03DL	VL043274.D	20 Nov 2025 20:43	SY/MD	Ok,M
20	Q3626-09DL	VL043275.D	21 Nov 2025 07:53	SY/MD	Ok,M
21	Q3698-04DUP	VL043276.D	21 Nov 2025 08:28	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL112425

Review By	Semsettin Yesilyurt	Review On	11/25/2025 7:23:41 AM		
Supervise By	Mahesh Dadoda	Supervise On	11/25/2025 8:00:09 AM		
SubDirectory	VL112425	HP Acquire Method	TO15AIR	HP Processing Method	VL111825AIR.M
STD. NAME		STD REF.#			
Tune/Reschk	AP2691				
Initial Calibration Stds	AP2694,AP2695,AP2696				
CCC	AP2694				
Internal Standard/PEM	AP2691				
ICV/I.BLK	AP2697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL043277.D	24 Nov 2025 08:53	SY/MD	Ok
2	VSTDCCC010	VL043278.D	24 Nov 2025 12:27	SY/MD	Ok,M
3	VL1124ABL01	VL043279.D	24 Nov 2025 13:11	SY/MD	Ok
4	VL1124ABS01	VL043280.D	24 Nov 2025 13:52	SY/MD	Ok,M
5	Q3696-02	VL043281.D	24 Nov 2025 14:26	SY/MD	Ok,M
6	Q3696-02DUP	VL043282.D	24 Nov 2025 15:00	SY/MD	Ok,M
7	Q3696-04	VL043283.D	24 Nov 2025 15:35	SY/MD	Ok,M
8	Q3696-06	VL043284.D	24 Nov 2025 16:09	SY/MD	Ok,M
9	Q3696-07	VL043285.D	24 Nov 2025 16:43	SY/MD	Ok,M
10	Q3696-01	VL043286.D	24 Nov 2025 17:15	SY/MD	Ok,M
11	Q3696-03	VL043287.D	24 Nov 2025 17:47	SY/MD	Ok,M
12	Q3696-05	VL043288.D	24 Nov 2025 18:20	SY/MD	Ok,M
13	Q3696-01DL	VL043289.D	24 Nov 2025 20:07	SY/MD	Not Ok
14	QC112425 CAN 10315	VL043290.D	25 Nov 2025 08:21	SY/MD	Ok

M : Manual Integration

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043203.D
 Acq On : 18 Nov 2025 09:29
 Operator : SY/MD
 Sample : VSTDIC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5

Quant Time: Nov 19 00:48:47 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	110490	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.956	114	296493	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	247761	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.371	95	180529	9.941	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.400%

Target Compounds

					Qvalue	
2) Dichlorodifluoromethane	1.496	85	7723	0.496	ppbv	99
3) Chlorodifluoromethane	1.473	51	8538	0.557	ppbv #	82
4) Chloromethane	1.531	50	2881	0.494	ppbv	94
5) Vinyl Chloride	1.580	62	2947	0.462	ppbv #	87
6) Bromomethane	1.667	94	1656	0.536	ppbv	97
7) Chloroethane	1.703	64	1357	0.547	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	6386	0.493	ppbv	100
9) Propene	1.483	41	3739	0.526	ppbv #	1
10) Heptane	4.972	43	5844	0.300	ppbv #	28
11) Trichlorofluoromethane	1.878	101	6841	0.453	ppbv	92
12) 1,1,2-Trichlorotrifluo...	2.140	101	5863	0.497	ppbv	99
13) Ethanol	1.719	45	1190	0.520	ppbv	64
14) Bromoethene	1.781	108	2476	0.548	ppbv #	75
15) Acetone	1.836	43	9506	0.619	ppbv #	88
16) 1,3-Butadiene	1.606	39	4115	0.601	ppbv	95
17) tert-Butyl alcohol	2.043	59	9854	0.489	ppbv #	68
18) 1,1-Dichloroethene	2.033	96	3068	0.559	ppbv	96
19) Isopropyl Alcohol	1.887	45	4687	0.540	ppbv #	51
20) Methylene Chloride	2.059	84	5006	0.349	ppbv	92
21) Allyl Chloride	2.095	41	6291	0.609	ppbv	95
22) trans-1,2-Dichloroethene	2.337	96	3343	0.533	ppbv	80
23) Vinyl Acetate	2.464	43	10414	0.562	ppbv #	92
24) 1,1-Dichloroethane	2.409	63	5803	0.479	ppbv #	96
25) Ethyl Acetate	2.839	43	13478	0.406	ppbv #	84
26) Hexane	2.823	57	7589	0.523	ppbv #	91
27) Carbon Disulfide	2.150	76	7198	0.486	ppbv #	1
28) Methyl tert-Butyl Ether	2.441	73	3545	0.498	ppbv #	15
29) Chloroform	2.849	83	10800	0.523	ppbv	95
30) Cyclohexane	3.849	84	6271	0.515	ppbv	85
31) cis-1,2-Dichloroethene	2.719	61	7728	0.536	ppbv	87
32) 1,1,1-Trichloroethane	3.363	97	10886	0.490	ppbv	98
34) 2-Butanone	2.557	43	11026	0.527	ppbv	95
35) Carbon Tetrachloride	3.761	117	11551	0.513	ppbv	94
36) Benzene	3.655	78	14759	0.520	ppbv #	91
37) 1,2-Dichloroethane	3.218	62	8477	0.510	ppbv #	93
38) Trichloroethene	4.542	130	5775	0.519	ppbv #	87
39) 1,2-Dichloropropane	4.312	63	5510	0.530	ppbv #	80
40) 1,4-Dioxane	4.597	88	1334	0.286	ppbv #	40
41) Tetrahydrofuran	3.059	42	6193	0.536	ppbv	95
42) Bromodichloromethane	4.490	83	11597	0.506	ppbv #	84
43) Methyl Methacrylate	4.865	69	2151	0.143	ppbv #	1
44) 2,2,4-Trimethylpentane	4.642	57	25890	0.531	ppbv	95
45) t-1,3-Dichloropropene	6.412	75	7343	0.523	ppbv	94


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043203.D
 Acq On : 18 Nov 2025 09:29
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Quant Time: Nov 19 00:48:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	7004	0.425	ppbv #	87
47) 1,1,2-Trichloroethane	6.581	97	3211	0.290	ppbv #	64
48) Dibromochloromethane	7.380	129	9327	0.492	ppbv	97
49) Bromoform	9.481	173	8636	0.514	ppbv	100
50) 4-Methyl-2-Pentanone	5.755	43	15190	0.515	ppbv	92
51) 2-Hexanone	7.442	43	11594	0.506	ppbv #	93
52) Tetrachloroethene	8.218	164	5097	0.483	ppbv #	82
53) Toluene	6.933	91	6863	0.203	ppbv #	21
54) 1,2-Dibromoethane	7.649	107	8708	0.537	ppbv	90
56) 1,1,1,2-Tetrachloroethane	8.917	131	5419	0.298	ppbv #	34
57) Chlorobenzene	8.917	112	12818	0.504	ppbv	98
58) Ethyl Benzene	9.351	91	23866	0.500	ppbv	96
59) m/p-Xylene	9.549	91	19470	0.507	ppbv #	24
60) o-Xylene	9.956	91	19621	0.514	ppbv	100
61) Styrene	9.869	104	8287	0.496	ppbv	96
62) Isopropylbenzene	10.536	105	34773	0.509	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.960	83	12591	0.519	ppbv	100
64) n-propylbenzene	10.986	120	8945	0.509	ppbv	100
65) tert-Butylbenzene	11.526	119	31998	0.524	ppbv	99
66) Benzyl Chloride	11.610	91	3403	0.550	ppbv	98
67) sec-Butylbenzene	11.766	105	44448	0.522	ppbv	97
69) p-Isopropyltoluene	11.921	119	37087	0.514	ppbv	100
70) n-Butylbenzene	12.277	91	33153	0.474	ppbv	94
71) 2-Chlorotoluene	10.895	91	27224	0.518	ppbv	99
72) 4-Ethyltoluene	11.122	105	23963	0.515	ppbv	97
73) 1,3,5-Trimethylbenzene	11.199	105	20104	0.515	ppbv	99
74) 1,2,4-Trimethylbenzene	11.533	105	22254	0.522	ppbv	91
75) 1,3-Dichlorobenzene	11.604	146	12926	0.520	ppbv	98
76) 1,4-Dichlorobenzene	11.672	146	13076	0.526	ppbv	98
77) 1,2-Dichlorobenzene	11.947	146	12069	0.505	ppbv	97
78) Hexachloro-1,3-Butadiene	13.879	225	10578	0.530	ppbv	98
79) Naphthalene	13.513	128	18377	0.486	ppbv	97
80) Naphthalene,2-methyl-	14.462	142	4917	0.495	ppbv	91
81) 1,2,4-Trichlorobenzene	13.439	180	6798	0.402	ppbv	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion: 49 Resp: 110490
Ion Ratio Lower Upper
49 100
128 46.6 23.0 69.0
130 58.4 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.496 ppbv
RT: 1.496 min Scan# 435
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 85 Resp: 7723
Ion Ratio Lower Upper
85 100
87 32.5 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.557 ppbv
RT: 1.473 min Scan# 428
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion: 51 Resp: 8538
Ion Ratio Lower Upper
51 100
67 26.4 14.9 22.3#



#4
Chloromethane
Concen: 0.494 ppbv
RT: 1.531 min Scan# 446
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 50 Resp: 2881
Ion Ratio Lower Upper
50 100
52 33.5 24.1 36.1









#5
Vinyl Chloride
Concen: 0.462 ppbv
RT: 1.580 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion: 62 Resp: 2947
Ion Ratio Lower Upper
62 100
64 26.4 27.3 40.9#



#6
Bromomethane
Concen: 0.536 ppbv
RT: 1.667 min Scan# 488
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 94 Resp: 1656
Ion Ratio Lower Upper
94 100
96 93.9 72.9 109.3









#7
Chloroethane
Concen: 0.547 ppbv
RT: 1.703 min Scan# 499
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion: 64 Resp: 1357
Ion Ratio Lower Upper
64 100
66 33.7 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 0.493 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 85 Resp: 6386
Ion Ratio Lower Upper
85 100
135 65.8 52.6 79.0





#9

Propene

Concen: 0.526 ppbv

RT: 1.483 min Scan# 41

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_1

ClientSampleId :





VSTDICC0.5

Tgt Ion: 41 Resp: 3739

Ion Ratio Lower Upper

41 100

42 121.9 3.7 5.5#



#10

Heptane

Concen: 0.300 ppbv

RT: 4.972 min Scan# 1509

Delta R.T. 0.007 min

Lab File: VL043203.D




Acq: 18 Nov 2025 09:29


Tgt Ion: 43 Resp: 5844

Ion Ratio Lower Upper

43 100

57 0.0 39.7 59.5#









#11
 Trichlorofluoromethane
 Concen: 0.453 ppbv
 RT: 1.878 min Scan# 51
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5


Tgt Ion:101 Resp: 6841
 Ion Ratio Lower Upper
 101 100
 103 71.8 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.497 ppbv
 RT: 2.140 min Scan# 634
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Tgt Ion:101 Resp: 5863
 Ion Ratio Lower Upper
 101 100
 151 72.8 59.2 88.8









#13
Ethanol
Concen: 0.520 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion: 45 Resp: 1190
Ion Ratio Lower Upper
45 100
46 58.7 15.0 59.8



#14
Bromoethene
Concen: 0.548 ppbv
RT: 1.781 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 108 Resp: 2476
Ion Ratio Lower Upper
108 100
106 126.4 88.9 133.3
79 58.5 16.5 24.7#









#15
Acetone
Concen: 0.619 ppbv
RT: 1.836 min Scan# 540
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.5


Tgt Ion: 43 Resp: 9506
Ion Ratio Lower Upper
43 100
58 32.0 20.7 31.1#



#16
1,3-Butadiene
Concen: 0.601 ppbv
RT: 1.606 min Scan# 469
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 39 Resp: 4115
Ion Ratio Lower Upper
39 100
54 70.0 59.4 89.2





#17

tert-Butyl alcohol

Concen: 0.489 ppbv

RT: 2.043 min Scan# 602

Delta R.T. 0.006 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC0.5

Tgt Ion: 59 Resp: 9854

Ion Ratio Lower Upper

59 100

57 23.3 9.0 13.4#



#18

1,1-Dichloroethene

Concen: 0.559 ppbv

RT: 2.033 min Scan# 601

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29




Tgt Ion: 96 Resp: 3068


Ion Ratio Lower Upper

96 100

61 220.0 170.6 255.8

98 66.3 52.2 78.4









#19
Isopropyl Alcohol
Concen: 0.540 ppbv
RT: 1.887 min Scan# 51
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5


Tgt Ion: 45 Resp: 4687
Ion Ratio Lower Upper
45 100
58 9.0 31.3 46.9#



#20
Methylene Chloride
Concen: 0.349 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 84 Resp: 5006
Ion Ratio Lower Upper
84 100
49 173.9 132.9 199.3
51 48.8 34.8 52.2
86 70.6 48.2 72.2









#21
 Allyl Chloride
 Concen: 0.609 ppbv
 RT: 2.095 min Scan# 619
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.5


Tgt Ion:	41	Resp:	6291
Ion	Ratio	Lower	Upper
41	100		
76	22.2	22.1	33.1
39	73.5	61.0	91.4



#22
 trans-1,2-Dichloroethene
 Concen: 0.533 ppbv
 RT: 2.337 min Scan# 695
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Tgt Ion:	96	Resp:	3343
Ion	Ratio	Lower	Upper
96	100		
61	141.8	139.0	208.6
98	51.7	49.4	74.2









#23
 Vinyl Acetate
 Concen: 0.562 ppbv
 RT: 2.464 min Scan# 71
 Delta R.T. 0.006 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.5


Tgt Ion:	43	Resp:	10414
Ion	Ratio	Lower	Upper
43	100		
86	10.8	6.2	9.2#
42	11.4	8.0	12.0
44	10.7	4.0	6.0#



#24
 1,1-Dichloroethane
 Concen: 0.479 ppbv
 RT: 2.409 min Scan# 717
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Tgt Ion:	63	Resp:	5803
Ion	Ratio	Lower	Upper
63	100		
98	6.6	2.8	8.4
100	4.2	1.3	3.9#





#25

Ethyl Acetate

Concen: 0.406 ppbv

RT: 2.839 min Scan# 847

Delta R.T. 0.010 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC0.5

Tgt Ion: 43 Resp: 13478





Ion Ratio Lower Upper

43 100

45 15.5 8.1 12.1#

61 12.4 7.0 10.4#

70 16.8 7.0 10.6#



#26

Hexane

Concen: 0.523 ppbv

RT: 2.823 min Scan# 845

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Tgt Ion: 57 Resp: 7589




Ion Ratio Lower Upper


57 100

41 114.6 74.6 111.8#

43 229.7 179.0 268.4

56 58.1 42.3 63.5





#27

Carbon Disulfide

Concen: 0.486 ppbv

RT: 2.150 min Scan# 61

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_1

ClientSampleId :

VSTDICC0.5





Tgt Ion: 76 Resp: 7198

Ion Ratio Lower Upper

76 100

44 115.9 16.1 24.1#

78 26.4 9.6 14.4#



#28

Methyl tert-Butyl Ether

Concen: 0.498 ppbv

RT: 2.441 min Scan# 727

Delta R.T. 0.006 min

Lab File: VL043203.D




Acq: 18 Nov 2025 09:29


Tgt Ion: 73 Resp: 3545

Ion Ratio Lower Upper

73 100

57 67.2 19.8 29.8#





#29

Chloroform

Concen: 0.523 ppbv

RT: 2.849 min Scan# 851

Delta R.T. 0.006 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC0.5

Tgt Ion: 83 Resp: 10800

Ion Ratio Lower Upper

83 100

85 66.9 50.5 75.7



#30

Cyclohexane

Concen: 0.515 ppbv

RT: 3.849 min Scan# 1162

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29




Tgt Ion: 84 Resp: 6271


Ion Ratio Lower Upper

84 100

56 148.3 104.7 157.1

41 99.3 69.0 103.6









#31
cis-1,2-Dichloroethene
Concen: 0.536 ppbv
RT: 2.719 min Scan# 812
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.5


Tgt Ion:	61	Resp:	7728
Ion	Ratio	Lower	Upper
61	100		
96	49.8	49.3	73.9
98	46.5	32.1	48.1



#32
1,1,1-Trichloroethane
Concen: 0.490 ppbv
RT: 3.363 min Scan# 1012
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:	97	Resp:	10886
Ion	Ratio	Lower	Upper
97	100		
99	64.4	50.3	75.5
61	49.5	38.9	58.3









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.956 min Scan# 1195
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion:114 Resp: 296493
Ion Ratio Lower Upper
114 100
63 25.7 21.0 31.6
88 18.8 14.9 22.3



#34
2-Butanone
Concen: 0.527 ppbv
RT: 2.557 min Scan# 763
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 43 Resp: 11026
Ion Ratio Lower Upper
43 100
72 24.9 18.1 27.1









#35
Carbon Tetrachloride
Concen: 0.513 ppbv
RT: 3.761 min Scan# 1135
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion:117 Resp: 11551
Ion Ratio Lower Upper
117 100
119 99.4 74.6 112.0
121 25.7 23.2 34.8



#36
Benzene
Concen: 0.520 ppbv
RT: 3.655 min Scan# 1102
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 78 Resp: 14759
Ion Ratio Lower Upper
78 100
77 24.6 17.5 26.3
50 24.7 15.5 23.3#





#37

1,2-Dichloroethane

Concen: 0.510 ppbv

RT: 3.218 min Scan# 9

Delta R.T. 0.006 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC0.5

Tgt Ion: 62 Resp: 8477

Ion Ratio Lower Upper

62 100

98 9.2 5.5 8.3#



#38

Trichloroethene

Concen: 0.519 ppbv

RT: 4.542 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Tgt Ion: 130 Resp: 5775




Ion Ratio Lower Upper


130 100

95 96.0 85.7 128.5

132 72.9 73.4 110.0#

97 59.9 52.7 79.1





#39

1,2-Dichloropropane

Concen: 0.530 ppbv

RT: 4.312 min Scan# 11

Delta R.T. 0.010 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.5





Tgt Ion: 63 Resp: 5510

Ion Ratio Lower Upper

63 100

41 98.4 62.6 93.8#

62 76.7 51.0 76.6#



#40

1,4-Dioxane

Concen: 0.286 ppbv

RT: 4.597 min Scan# 1393

Delta R.T. 0.029 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Tgt Ion: 88 Resp: 1334




Ion Ratio Lower Upper


88 100

58 0.0 63.0 94.6#

43 39.8 32.5 48.7

57 0.0 0.0 0.0





#41

Tetrahydrofuran

Concen: 0.536 ppbv

RT: 3.059 min Scan# 914

Delta R.T. 0.013 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.5





Tgt Ion: 42 Resp: 6193

Ion Ratio Lower Upper

42 100

72 40.2 30.6 46.0

71 42.1 30.5 45.7



#42

Bromodichloromethane

Concen: 0.506 ppbv

RT: 4.490 min Scan# 1360

Delta R.T. 0.013 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29




Tgt Ion: 83 Resp: 11597


Ion Ratio Lower Upper

83 100

85 51.4 51.7 77.5#

127 4.8 6.3 9.5#









#43
Methyl Methacrylate
Concen: 0.143 ppbv
RT: 4.865 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.5


Tgt Ion: 69 Resp: 2151
Ion Ratio Lower Upper
69 100
41 0.0 126.7 190.1#
100 0.0 28.6 43.0#



#44
2,2,4-Trimethylpentane
Concen: 0.531 ppbv
RT: 4.642 min Scan# 1407
Delta R.T. 0.016 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 57 Resp: 25890
Ion Ratio Lower Upper
57 100
41 36.2 28.0 42.0
56 36.9 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 0.523 ppbv
RT: 6.412 min Scan# 1954
Delta R.T. 0.013 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion: 75 Resp: 7343
Ion Ratio Lower Upper
75 100
77 26.9 24.3 36.5



#46
cis-1,3-Dichloropropene
Concen: 0.425 ppbv
RT: 5.593 min Scan# 1701
Delta R.T. 0.010 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 75 Resp: 7004
Ion Ratio Lower Upper
75 100
39 58.9 59.1 88.7#
77 30.6 24.2 36.2





#47

1,1,2-Trichloroethane

Concen: 0.290 ppbv

RT: 6.581 min Scan# 2001

Delta R.T. 0.016 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC0.5

Tgt Ion: 97 Resp: 3211





Ion Ratio Lower Upper

97 100

83 70.2 64.6 97.0

85 11.0 39.4 59.2#

99 18.5 47.0 70.6#



#48

Dibromochloromethane

Concen: 0.492 ppbv

RT: 7.380 min Scan# 2253

Delta R.T. 0.003 min

Lab File: VL043203.D




Acq: 18 Nov 2025 09:29


Tgt Ion: 129 Resp: 9327

Ion Ratio Lower Upper

129 100

127 80.8 39.3 117.9









#49
Bromoform
Concen: 0.514 ppbv
RT: 9.481 min Scan# 2900
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion:173	Resp:	8636
Ion Ratio	Lower	Upper
173	100	
175	48.4	38.6 58.0
171	50.9	40.6 61.0



#50
4-Methyl-2-Pentanone
Concen: 0.515 ppbv
RT: 5.755 min Scan# 1751
Delta R.T. 0.023 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 43	Resp:	15190
Ion Ratio	Lower	Upper
43	100	
58	40.3	28.4 42.6









#51
2-Hexanone
Concen: 0.506 ppbv
RT: 7.442 min Scan# 2115
Delta R.T. 0.023 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5


Tgt Ion: 43 Resp: 11594
Ion Ratio Lower Upper
43 100
58 52.6 38.4 57.6
98 0.7 0.0 0.0#



#52
Tetrachloroethene
Concen: 0.483 ppbv
RT: 8.218 min Scan# 2512
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 164 Resp: 5097
Ion Ratio Lower Upper
164 100
166 147.2 99.4 149.2
129 127.2 81.8 122.8#
131 103.9 77.4 116.2









#53
Toluene
Concen: 0.203 ppbv
RT: 6.933 min Scan# 2115
Delta R.T. 0.013 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion: 91 Resp: 6863
Ion Ratio Lower Upper
91 100
92 0.0 47.4 71.0#



#54
1,2-Dibromoethane
Concen: 0.537 ppbv
RT: 7.649 min Scan# 2336
Delta R.T. 0.010 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 107 Resp: 8708
Ion Ratio Lower Upper
107 100
109 80.8 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.875 min Scan# 21
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion:117 Resp: 247761
Ion Ratio Lower Upper
117 100
82 65.5 56.1 84.1
119 31.6 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 0.298 ppbv
RT: 8.917 min Scan# 2728
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:131 Resp: 5419
Ion Ratio Lower Upper
131 100
133 177.8 77.4 116.2#
119 91.4 49.1 73.7#









#57
Chlorobenzene
Concen: 0.504 ppbv
RT: 8.917 min Scan# 21
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion:112 Resp: 12818
Ion Ratio Lower Upper
112 100
77 69.0 55.5 83.3
114 29.2 29.1 35.5



#58
Ethyl Benzene
Concen: 0.500 ppbv
RT: 9.351 min Scan# 2862
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 91 Resp: 23866
Ion Ratio Lower Upper
91 100
106 30.7 22.8 34.2





#59

m/p-Xylene

Concen: 0.507 ppbv

RT: 9.549 min Scan# 2921

Delta R.T. 0.007 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_1

ClientSampleId :





VSTDICC0.5

Tgt Ion: 91 Resp: 19470

Ion Ratio Lower Upper

91 100

106 94.6 36.1 54.1#



#60

o-Xylene

Concen: 0.514 ppbv

RT: 9.956 min Scan# 3049

Delta R.T. 0.000 min

Lab File: VL043203.D




Acq: 18 Nov 2025 09:29


Tgt Ion: 91 Resp: 19621

Ion Ratio Lower Upper

91 100

106 42.0 20.9 62.8





#61

Styrene

Concen: 0.496 ppbv

RT: 9.869 min Scan# 3020

Delta R.T. 0.007 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC0.5





Tgt Ion:104 Resp: 8287

Ion Ratio Lower Upper

104 100

78 56.3 42.1 63.1

103 47.8 37.2 55.8



#62

Isopropylbenzene

Concen: 0.509 ppbv

RT: 10.536 min Scan# 3228

Delta R.T. 0.003 min

Lab File: VL043203.D




Acq: 18 Nov 2025 09:29


Tgt Ion:105 Resp: 34773

Ion Ratio Lower Upper

105 100

120 25.6 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 0.519 ppbv
RT: 9.960 min Scan# 3048
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion:	83	Resp:	12591
Ion	Ratio	Lower	Upper
83	100		
131	9.8	4.8	14.3
85	64.0	31.9	95.9



#64
n-propylbenzene
Concen: 0.509 ppbv
RT: 10.986 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:	120	Resp:	8945
Ion	Ratio	Lower	Upper
120	100		
91	472.3	377.1	565.7





#65

tert-Butylbenzene

Concen: 0.524 ppbv

RT: 11.526 min Scan# 31998

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC0.5





Tgt Ion:119 Resp: 31998

Ion Ratio Lower Upper

119 100

91 81.6 65.3 97.9

134 23.4 17.0 25.4



#66

Benzyl Chloride

Concen: 0.550 ppbv

RT: 11.610 min Scan# 3560

Delta R.T. 0.003 min

Lab File: VL043203.D

Acq: 18 Nov 2025 09:29




Tgt Ion: 91 Resp: 3403


Ion Ratio Lower Upper

91 100

126 16.8 14.4 21.6

128 6.1 4.6 6.8









#67
 sec-Butylbenzene
 Concen: 0.522 ppbv
 RT: 11.766 min Scan# 3177
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.5


Tgt Ion:105 Resp: 44448
 Ion Ratio Lower Upper
 105 100
 134 20.0 15.0 22.6



#68
 1-Bromo-4-Fluorobenzene
 Concen: 9.941 ppbv
 RT: 10.371 min Scan# 3177
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Tgt Ion: 95 Resp: 180529
 Ion Ratio Lower Upper
 95 100
 174 63.2 50.2 75.2
 175 4.9 4.1 6.1









#69
p-Isopropyltoluene
Concen: 0.514 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.5


Tgt Ion:	119	Resp:	37087
Ion	Ratio	Lower	Upper
119	100		
134	25.4	20.2	30.2
91	29.9	23.8	35.8



#70
n-Butylbenzene
Concen: 0.474 ppbv
RT: 12.277 min Scan# 3766
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:	91	Resp:	33153
Ion	Ratio	Lower	Upper
91	100		
92	59.3	43.5	65.3
134	26.3	19.0	28.6









#71
2-Chlorotoluene
Concen: 0.518 ppbv
RT: 10.895 min Scan# 31
Delta R.T. 0.000 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion: 91 Resp: 27224
Ion Ratio Lower Upper
91 100
126 28.6 11.7 46.7



#72
4-Ethyltoluene
Concen: 0.515 ppbv
RT: 11.122 min Scan# 3409
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion: 105 Resp: 23963
Ion Ratio Lower Upper
105 100
120 31.3 23.8 35.8
91 14.6 10.3 15.5
77 15.6 11.1 16.7









#73
 1,3,5-Trimethylbenzene
 Concen: 0.515 ppbv
 RT: 11.199 min Scan# 3432
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.5


Tgt Ion:105 Resp: 20104
 Ion Ratio Lower Upper
 105 100
 120 48.8 38.6 57.8



#74
 1,2,4-Trimethylbenzene
 Concen: 0.522 ppbv
 RT: 11.533 min Scan# 3536
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Tgt Ion:105 Resp: 22254
 Ion Ratio Lower Upper
 105 100
 120 46.6 41.7 62.5
 91 77.8 70.0 105.0
 77 25.6 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 0.520 ppbv
RT: 11.604 min Scan# 3557
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion:146 Resp: 12926
Ion Ratio Lower Upper
146 100
111 45.1 23.4 70.3
148 64.7 31.6 95.0



#76
1,4-Dichlorobenzene
Concen: 0.526 ppbv
RT: 11.672 min Scan# 3579
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:146 Resp: 13076
Ion Ratio Lower Upper
146 100
111 42.1 22.8 68.3
148 63.8 31.8 95.4









#77
1,2-Dichlorobenzene
Concen: 0.505 ppbv
RT: 11.947 min Scan# 30
Delta R.T. 0.006 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.5


Tgt Ion:	146	Resp:	12069
Ion Ratio	Lower	Upper	
146	100		
111	49.5	24.3	73.0
148	66.3	31.6	94.7



#78
Hexachloro-1,3-Butadiene
Concen: 0.530 ppbv
RT: 13.879 min Scan# 4261
Delta R.T. 0.003 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29

Tgt Ion:	225	Resp:	10578
Ion Ratio	Lower	Upper	
225	100		
223	62.7	50.2	75.4
227	61.0	51.1	76.7









#79
Naphthalene
Concen: 0.486 ppbv
RT: 13.513 min Scan# 4148
Delta R.T. 0.007 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.5


Tgt Ion:	128	Resp:	18377
Ion Ratio	Lower	Upper	
128	100		
127	11.2	9.9	14.9
129	11.5	8.4	12.6



#80
Naphthalene,2-methyl-
Concen: 0.495 ppbv
RT: 14.462 min Scan# 4441
Delta R.T. 0.010 min
Lab File: VL043203.D
Acq: 18 Nov 2025 09:29


Tgt Ion:	142	Resp:	4917
Ion Ratio	Lower	Upper	
142	100		
141	90.2	67.1	100.7
115	47.3	31.9	47.9







#81
 1,2,4-Trichlorobenzene
 Concen: 0.402 ppbv
 RT: 13.439 min Scan# 41
 Delta R.T. 0.003 min
 Lab File: VL043203.D
 Acq: 18 Nov 2025 09:29

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.5



Tgt Ion:	180	Resp:	6798
Ion Ratio	Lower	Upper	
180	100		
182	95.7	77.4	116.2
184	34.3	24.2	36.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043204.D
 Acq On : 18 Nov 2025 10:02
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Quant Time: Nov 19 00:49:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	105095	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	283830	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	234772	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.367 95 172524 10.025 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.300%

Target Compounds

					Qvalue
2) Dichlorodifluoromethane	1.496	85	140789	9.498 ppbv	100
3) Chlorodifluoromethane	1.473	51	138443	9.497 ppbv	100
4) Chloromethane	1.531	50	52372	9.450 ppbv	100
5) Vinyl Chloride	1.577	62	51045	8.422 ppbv	100
6) Bromomethane	1.664	94	26542	9.034 ppbv	100
7) Chloroethane	1.700	64	20375	8.632 ppbv	100
8) Dichlorotetrafluoroethane	1.551	85	114670	9.300 ppbv	100
9) Propene	1.480	41	52671	7.789 ppbv #	1
10) Heptane	4.965	43	168611	9.090 ppbv	100
11) Trichlorofluoromethane	1.875	101	130362	9.077 ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.137	101	103679	9.247 ppbv	100
13) Ethanol	1.716	45	17464	8.023 ppbv	100
14) Bromoethene	1.777	108	38037	8.849 ppbv	100
15) Acetone	1.832	43	108422	7.426 ppbv	100
16) 1,3-Butadiene	1.606	39	57871	8.890 ppbv	100
17) tert-Butyl alcohol	2.036	59	179201	9.350 ppbv	100
18) 1,1-Dichloroethene	2.030	96	46313	8.871 ppbv	100
19) Isopropyl Alcohol	1.881	45	71862	8.710 ppbv	100
20) Methylene Chloride	2.056	84	41868	9.708 ppbv	100
21) Allyl Chloride	2.091	41	89266	9.087 ppbv	100
22) trans-1,2-Dichloroethene	2.334	96	52470	8.793 ppbv	100
23) Vinyl Acetate	2.457	43	170459	9.670 ppbv	100
24) 1,1-Dichloroethane	2.405	63	106970	9.287 ppbv	100
25) Ethyl Acetate	2.829	43	289921	9.190 ppbv	100
26) Hexane	2.820	57	129809	9.402 ppbv	100
27) Carbon Disulfide	2.146	76	132086	9.371 ppbv	100
28) Methyl tert-Butyl Ether	2.434	73	62115	9.165 ppbv	100
29) Chloroform	2.842	83	187620	9.561 ppbv	100
30) Cyclohexane	3.846	84	107621	9.297 ppbv	100
31) cis-1,2-Dichloroethene	2.716	61	130357	9.499 ppbv	100
32) 1,1,1-Trichloroethane	3.360	97	198036	9.381 ppbv	100
34) 2-Butanone	2.551	43	182749	9.121 ppbv	100
35) Carbon Tetrachloride	3.755	117	200831	9.322 ppbv	100
36) Benzene	3.648	78	256954	9.453 ppbv	100
37) 1,2-Dichloroethane	3.211	62	150769	9.483 ppbv	100
38) Trichloroethene	4.535	130	100136	9.398 ppbv	100
39) 1,2-Dichloropropane	4.302	63	95097	9.559 ppbv	100
40) 1,4-Dioxane	4.567	88	40441	9.043 ppbv #	100
41) Tetrahydrofuran	3.046	42	104294	9.426 ppbv	100
42) Bromodichloromethane	4.477	83	211504	9.643 ppbv	100
43) Methyl Methacrylate	4.865	69	131003	9.080 ppbv	100
44) 2,2,4-Trimethylpentane	4.626	57	441234	9.451 ppbv	100
45) t-1,3-Dichloropropene	6.399	75	129122	9.601 ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043204.D
 Acq On : 18 Nov 2025 10:02
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010


Quant Time: Nov 19 00:49:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.584	75	151655	9.603	ppbv	100
47) 1,1,2-Trichloroethane	6.564	97	97512	9.193	ppbv	100
48) Dibromochloromethane	7.377	129	175529	9.670	ppbv	100
49) Bromoform	9.474	173	157563	9.800	ppbv	100
50) 4-Methyl-2-Pentanone	5.733	43	268369	9.509	ppbv	100
51) 2-Hexanone	7.419	43	208268	9.503	ppbv #	100
52) Tetrachloroethene	8.212	164	93518	9.249	ppbv	100
53) Toluene	6.920	91	305169	9.422	ppbv	100
54) 1,2-Dibromoethane	7.639	107	148973	9.588	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.914	131	158856	9.207	ppbv	100
57) Chlorobenzene	8.911	112	227490	9.438	ppbv	100
58) Ethyl Benzene	9.345	91	427530	9.461	ppbv	100
59) m/p-Xylene	9.542	91	316603	8.707	ppbv #	22
60) o-Xylene	9.956	91	333044	9.202	ppbv	100
61) Styrene	9.863	104	152172	9.607	ppbv	100
62) Isopropylbenzene	10.533	105	600155	9.269	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.953	83	209272	9.109	ppbv	100
64) n-propylbenzene	10.982	120	154239	9.271	ppbv	100
65) tert-Butylbenzene	11.523	119	522106	9.029	ppbv	100
66) Benzyl Chloride	11.607	91	56620	9.649	ppbv	100
67) sec-Butylbenzene	11.762	105	729573	9.039	ppbv	100
69) p-Isopropyltoluene	11.918	119	624586	9.135	ppbv	100
70) n-Butylbenzene	12.274	91	619610	9.341	ppbv	100
71) 2-Chlorotoluene	10.895	91	460118	9.245	ppbv	100
72) 4-Ethyltoluene	11.118	105	410587	9.316	ppbv	100
73) 1,3,5-Trimethylbenzene	11.196	105	341900	9.244	ppbv	100
74) 1,2,4-Trimethylbenzene	11.529	105	363476	9.005	ppbv	100
75) 1,3-Dichlorobenzene	11.601	146	219894	9.335	ppbv	100
76) 1,4-Dichlorobenzene	11.665	146	220576	9.363	ppbv	100
77) 1,2-Dichlorobenzene	11.940	146	204826	9.044	ppbv	100
78) Hexachloro-1,3-Butadiene	13.876	225	161034	8.520	ppbv	100
79) Naphthalene	13.507	128	351148	9.795	ppbv	100
80) Naphthalene,2-methyl-	14.452	142	93113	9.895	ppbv	100
81) 1,2,4-Trichlorobenzene	13.436	180	159091	9.923	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 81
Delta R.T. -0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion: 49 Resp: 105095
Ion Ratio Lower Upper
49 100
128 46.0 23.0 69.0
130 59.8 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 9.498 ppbv
RT: 1.496 min Scan# 435
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 85 Resp: 140789
Ion Ratio Lower Upper
85 100
87 31.8 25.4 38.2









#3
Chlorodifluoromethane
Concen: 9.497 ppbv
RT: 1.473 min Scan# 41
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 51 Resp: 138443
Ion Ratio Lower Upper
51 100
67 18.6 14.9 22.3



#4
Chloromethane
Concen: 9.450 ppbv
RT: 1.531 min Scan# 446
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 50 Resp: 52372
Ion Ratio Lower Upper
50 100
52 30.1 24.1 36.1









#5
Vinyl Chloride
Concen: 8.422 ppbv
RT: 1.577 min Scan# 460
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 62 Resp: 51045
Ion Ratio Lower Upper
62 100
64 34.1 27.3 40.9



#6
Bromomethane
Concen: 9.034 ppbv
RT: 1.664 min Scan# 487
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 94 Resp: 26542
Ion Ratio Lower Upper
94 100
96 91.1 72.9 109.3









#7
Chloroethane
Concen: 8.632 ppbv
RT: 1.700 min Scan# 498
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 64 Resp: 20375
Ion Ratio Lower Upper
64 100
66 33.5 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 9.300 ppbv
RT: 1.551 min Scan# 452
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 85 Resp: 114670
Ion Ratio Lower Upper
85 100
135 65.8 52.6 79.0





#9

Propene

Concen: 7.789 ppbv

RT: 1.480 min Scan# 41

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_1

ClientSampleId :





VSTDICCC010

Tgt Ion: 41 Resp: 52671

Ion Ratio Lower Upper

41 100

42 105.2 3.7 5.5#



#10

Heptane

Concen: 9.090 ppbv

RT: 4.965 min Scan# 1507

Delta R.T. 0.000 min

Lab File: VL043204.D




Acq: 18 Nov 2025 10:02


Tgt Ion: 43 Resp: 168611

Ion Ratio Lower Upper

43 100

57 49.6 39.7 59.5









#11
 Trichlorofluoromethane
 Concen: 9.077 ppbv
 RT: 1.875 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICCC010


Tgt Ion:101 Resp: 130362
 Ion Ratio Lower Upper
 101 100
 103 65.8 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 9.247 ppbv
 RT: 2.137 min Scan# 633
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02

Tgt Ion:101 Resp: 103679
 Ion Ratio Lower Upper
 101 100
 151 74.0 59.2 88.8









#13
Ethanol
Concen: 8.023 ppbv
RT: 1.716 min Scan# 503
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion: 45 Resp: 17464
Ion Ratio Lower Upper
45 100
46 37.4 15.0 59.8



#14
Bromoethene
Concen: 8.849 ppbv
RT: 1.777 min Scan# 522
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 108 Resp: 38037
Ion Ratio Lower Upper
108 100
106 111.1 88.9 133.3
79 20.6 16.5 24.7









#15
Acetone
Concen: 7.426 ppbv
RT: 1.832 min Scan# 51
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion: 43 Resp: 108422
Ion Ratio Lower Upper
43 100
58 25.9 20.7 31.1



#16
1,3-Butadiene
Concen: 8.890 ppbv
RT: 1.606 min Scan# 469
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 39 Resp: 57871
Ion Ratio Lower Upper
39 100
54 74.3 59.4 89.2





#17

tert-Butyl alcohol

Concen: 9.350 ppbv

RT: 2.036 min Scan# 602

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_1

ClientSampleId :





VSTDICCC010

Tgt Ion: 59 Resp: 179201

Ion Ratio Lower Upper

59 100

57 11.2 9.0 13.4



#18

1,1-Dichloroethene

Concen: 8.871 ppbv

RT: 2.030 min Scan# 600

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02




Tgt Ion: 96 Resp: 46313


Ion Ratio Lower Upper

96 100

61 213.2 170.6 255.8

98 65.3 52.2 78.4









#19
Isopropyl Alcohol
Concen: 8.710 ppbv
RT: 1.881 min Scan# 511
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion: 45 Resp: 71862
Ion Ratio Lower Upper
45 100
58 39.1 31.3 46.9



#20
Methylene Chloride
Concen: 9.708 ppbv
RT: 2.056 min Scan# 608
Delta R.T. -0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 84 Resp: 41868
Ion Ratio Lower Upper
84 100
49 166.1 132.9 199.3
51 43.5 34.8 52.2
86 60.2 48.2 72.2









#21
 Allyl Chloride
 Concen: 9.087 ppbv
 RT: 2.091 min Scan# 619
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010


Tgt Ion:	41	Resp:	89266
Ion	Ratio	Lower	Upper
41	100		
76	27.6	22.1	33.1
39	76.2	61.0	91.4



#22
 trans-1,2-Dichloroethene
 Concen: 8.793 ppbv
 RT: 2.334 min Scan# 694
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02

Tgt Ion:	96	Resp:	52470
Ion	Ratio	Lower	Upper
96	100		
61	173.8	139.0	208.6
98	61.8	49.4	74.2









#23
 Vinyl Acetate
 Concen: 9.670 ppbv
 RT: 2.457 min Scan# 716
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICCC010


Tgt Ion:	43	Resp:	170459
Ion	Ratio	Lower	Upper
43	100		
86	7.7	6.2	9.2
42	10.0	8.0	12.0
44	5.0	4.0	6.0



#24
 1,1-Dichloroethane
 Concen: 9.287 ppbv
 RT: 2.405 min Scan# 716
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02

Tgt Ion:	63	Resp:	106970
Ion	Ratio	Lower	Upper
63	100		
98	5.6	2.8	8.4
100	2.6	1.3	3.9









#25
Ethyl Acetate
Concen: 9.190 ppbv
RT: 2.829 min Scan# 847
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion:	43	Resp:	289921
Ion	Ratio	Lower	Upper
43	100		
45	10.1	8.1	12.1
61	8.7	7.0	10.4
70	8.8	7.0	10.6



#26
Hexane
Concen: 9.402 ppbv
RT: 2.820 min Scan# 844
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:	57	Resp:	129809
Ion	Ratio	Lower	Upper
57	100		
41	93.2	74.6	111.8
43	223.7	179.0	268.4
56	52.9	42.3	63.5









#27
Carbon Disulfide
Concen: 9.371 ppbv
RT: 2.146 min Scan# 611
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 76 Resp: 132086
Ion Ratio Lower Upper
76 100
44 20.1 16.1 24.1
78 12.0 9.6 14.4



#28
Methyl tert-Butyl Ether
Concen: 9.165 ppbv
RT: 2.434 min Scan# 725
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 73 Resp: 62115
Ion Ratio Lower Upper
73 100
57 24.8 19.8 29.8





#29

Chloroform

Concen: 9.561 ppbv

RT: 2.842 min Scan# 851

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_L

ClientSampleId :





VSTDICCC010

Tgt Ion: 83 Resp: 187620

Ion Ratio Lower Upper

83 100

85 63.1 50.5 75.7



#30

Cyclohexane

Concen: 9.297 ppbv

RT: 3.846 min Scan# 1161

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02




Tgt Ion: 84 Resp: 107621


Ion Ratio Lower Upper

84 100

56 130.9 104.7 157.1

41 86.3 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 9.499 ppbv
 RT: 2.716 min Scan# 812
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICCC010


Tgt Ion: 61 Resp: 130357
 Ion Ratio Lower Upper
 61 100
 96 61.6 49.3 73.9
 98 40.1 32.1 48.1



#32
 1,1,1-Trichloroethane
 Concen: 9.381 ppbv
 RT: 3.360 min Scan# 1011
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02

Tgt Ion: 97 Resp: 198036
 Ion Ratio Lower Upper
 97 100
 99 62.9 50.3 75.5
 61 48.6 38.9 58.3









#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 3.949 min Scan# 1193
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICCC010


Tgt Ion:114 Resp: 283830
 Ion Ratio Lower Upper
 114 100
 63 26.3 21.0 31.6
 88 18.6 14.9 22.3



#34
 2-Butanone
 Concen: 9.121 ppbv
 RT: 2.551 min Scan# 761
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02

Tgt Ion: 43 Resp: 182749
 Ion Ratio Lower Upper
 43 100
 72 22.6 18.1 27.1









#35
Carbon Tetrachloride
Concen: 9.322 ppbv
RT: 3.755 min Scan# 1133
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion:117 Resp: 200831
Ion Ratio Lower Upper
117 100
119 93.3 74.6 112.0
121 29.0 23.2 34.8



#36
Benzene
Concen: 9.453 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 78 Resp: 256954
Ion Ratio Lower Upper
78 100
77 21.9 17.5 26.3
50 19.4 15.5 23.3









#37
1,2-Dichloroethane
Concen: 9.483 ppbv
RT: 3.211 min Scan# 965
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 62 Resp: 150769
Ion Ratio Lower Upper
62 100
98 6.9 5.5 8.3



#38
Trichloroethene
Concen: 9.398 ppbv
RT: 4.535 min Scan# 1374
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:130 Resp: 100136
Ion Ratio Lower Upper
130 100
95 107.1 85.7 128.5
132 91.7 73.4 110.0
97 65.9 52.7 79.1





#39

1,2-Dichloropropane

Concen: 9.559 ppbv

RT: 4.302 min Scan# 1302

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010





Tgt Ion: 63 Resp: 95097

Ion Ratio Lower Upper

63 100

41 78.2 62.6 93.8

62 63.8 51.0 76.6



#40

1,4-Dioxane

Concen: 9.043 ppbv

RT: 4.567 min Scan# 1384

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Tgt Ion: 88 Resp: 40441




Ion Ratio Lower Upper


88 100

58 78.8 63.0 94.6

43 40.6 32.5 48.7

57 0.0 0.0 0.0





#41

Tetrahydrofuran

Concen: 9.426 ppbv

RT: 3.046 min Scan# 914

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010





Tgt Ion: 42 Resp: 104294

Ion Ratio Lower Upper

42 100

72 38.3 30.6 46.0

71 38.1 30.5 45.7



#42

Bromodichloromethane

Concen: 9.643 ppbv

RT: 4.477 min Scan# 1356

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02




Tgt Ion: 83 Resp: 211504


Ion Ratio Lower Upper

83 100

85 64.6 51.7 77.5

127 7.9 6.3 9.5









#43
Methyl Methacrylate
Concen: 9.080 ppbv
RT: 4.865 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 69 Resp: 131003
Ion Ratio Lower Upper
69 100
41 158.4 126.7 190.1
100 35.8 28.6 43.0



#44
2,2,4-Trimethylpentane
Concen: 9.451 ppbv
RT: 4.626 min Scan# 1402
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 57 Resp: 441234
Ion Ratio Lower Upper
57 100
41 35.0 28.0 42.0
56 32.0 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 9.601 ppbv
RT: 6.399 min Scan# 1950
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 75 Resp: 129122
Ion Ratio Lower Upper
75 100
77 30.4 24.3 36.5



#46
cis-1,3-Dichloropropene
Concen: 9.603 ppbv
RT: 5.584 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 75 Resp: 151655
Ion Ratio Lower Upper
75 100
39 73.9 59.1 88.7
77 30.2 24.2 36.2





#47

1,1,2-Trichloroethane

Concen: 9.193 ppbv

RT: 6.564 min Scan# 2001

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion: 97 Resp: 97512





Ion Ratio Lower Upper

97 100

83 80.8 64.6 97.0

85 49.3 39.4 59.2

99 58.8 47.0 70.6



#48

Dibromochloromethane

Concen: 9.670 ppbv

RT: 7.377 min Scan# 2252

Delta R.T. 0.000 min

Lab File: VL043204.D




Acq: 18 Nov 2025 10:02


Tgt Ion: 129 Resp: 175529

Ion Ratio Lower Upper

129 100

127 78.6 39.3 117.9





#49

Bromoform

Concen: 9.800 ppbv

RT: 9.474 min Scan# 2900

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_1

ClientSampleId :

VSTDICCC010





Tgt Ion:173 Resp: 157563

Ion Ratio Lower Upper

173 100

175 48.3 38.6 58.0

171 50.8 40.6 61.0



#50

4-Methyl-2-Pentanone

Concen: 9.509 ppbv

RT: 5.733 min Scan# 1744

Delta R.T. 0.000 min

Lab File: VL043204.D




Acq: 18 Nov 2025 10:02


Tgt Ion: 43 Resp: 268369

Ion Ratio Lower Upper

43 100

58 35.5 28.4 42.6









#51
2-Hexanone
Concen: 9.503 ppbv
RT: 7.419 min Scan# 21
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion: 43 Resp: 208268
Ion Ratio Lower Upper
43 100
58 48.0 38.4 57.6
98 0.0 0.0 0.0#



#52
Tetrachloroethene
Concen: 9.249 ppbv
RT: 8.212 min Scan# 2510
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 164 Resp: 93518
Ion Ratio Lower Upper
164 100
166 124.3 99.4 149.2
129 102.3 81.8 122.8
131 96.8 77.4 116.2









#53
Toluene
Concen: 9.422 ppbv
RT: 6.920 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 91 Resp: 305169
Ion Ratio Lower Upper
91 100
92 59.2 47.4 71.0



#54
1,2-Dibromoethane
Concen: 9.588 ppbv
RT: 7.639 min Scan# 2333
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 107 Resp: 148973
Ion Ratio Lower Upper
107 100
109 90.1 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 21
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion:117 Resp: 234772
Ion Ratio Lower Upper
117 100
82 70.1 56.1 84.1
119 31.3 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 9.207 ppbv
RT: 8.914 min Scan# 2727
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:131 Resp: 158856
Ion Ratio Lower Upper
131 100
133 96.8 77.4 116.2
119 61.4 49.1 73.7









#57
Chlorobenzene
Concen: 9.438 ppbv
RT: 8.911 min Scan# 21
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion:112 Resp: 227490
Ion Ratio Lower Upper
112 100
77 69.4 55.5 83.3
114 32.3 29.1 35.5



#58
Ethyl Benzene
Concen: 9.461 ppbv
RT: 9.345 min Scan# 2860
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 91 Resp: 427530
Ion Ratio Lower Upper
91 100
106 28.5 22.8 34.2





#59

m/p-Xylene

Concen: 8.707 ppbv

RT: 9.542 min Scan# 2921

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_1

ClientSampleId :





VSTDICCC010

Tgt Ion: 91 Resp: 316603

Ion Ratio Lower Upper

91 100

106 95.9 36.1 54.1#



#60

o-Xylene

Concen: 9.202 ppbv

RT: 9.956 min Scan# 3049

Delta R.T. 0.000 min

Lab File: VL043204.D




Acq: 18 Nov 2025 10:02


Tgt Ion: 91 Resp: 333044

Ion Ratio Lower Upper

91 100

106 41.9 20.9 62.8









#61
Styrene
Concen: 9.607 ppbv
RT: 9.863 min Scan# 3020
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion:104 Resp: 152172
Ion Ratio Lower Upper
104 100
78 52.6 42.1 63.1
103 46.5 37.2 55.8



#62
Isopropylbenzene
Concen: 9.269 ppbv
RT: 10.533 min Scan# 3227
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:105 Resp: 600155
Ion Ratio Lower Upper
105 100
120 24.8 19.8 29.8









#63
1,1,2-Tetrachloroethane
Concen: 9.109 ppbv
RT: 9.953 min Scan# 3048
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010


Tgt Ion: 83 Resp: 209272
Ion Ratio Lower Upper
83 100
131 9.5 4.8 14.3
85 63.9 31.9 95.9



#64
n-propylbenzene
Concen: 9.271 ppbv
RT: 10.982 min Scan# 3366
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 120 Resp: 154239
Ion Ratio Lower Upper
120 100
91 471.4 377.1 565.7





#65

tert-Butylbenzene

Concen: 9.029 ppbv

RT: 11.523 min Scan# 3533

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_1

ClientSampleId :

VSTDICCC010





Tgt Ion:119 Resp: 522106

Ion Ratio Lower Upper

119 100

91 81.6 65.3 97.9

134 21.2 17.0 25.4



#66

Benzyl Chloride

Concen: 9.649 ppbv

RT: 11.607 min Scan# 3559

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02




Tgt Ion: 91 Resp: 56620


Ion Ratio Lower Upper

91 100

126 18.0 14.4 21.6

128 5.7 4.6 6.8









#67
 sec-Butylbenzene
 Concen: 9.039 ppbv
 RT: 11.762 min Scan# 3176
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICCC010


Tgt Ion:105 Resp: 729573
 Ion Ratio Lower Upper
 105 100
 134 18.8 15.0 22.6



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.025 ppbv
 RT: 10.367 min Scan# 3176
 Delta R.T. 0.000 min
 Lab File: VL043204.D
 Acq: 18 Nov 2025 10:02

Tgt Ion: 95 Resp: 172524
 Ion Ratio Lower Upper
 95 100
 174 62.7 50.2 75.2
 175 5.1 4.1 6.1









#69
p-Isopropyltoluene
Concen: 9.135 ppbv
RT: 11.918 min Scan# 3655
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion:119 Resp: 624586
Ion Ratio Lower Upper
119 100
134 25.2 20.2 30.2
91 29.8 23.8 35.8



#70
n-Butylbenzene
Concen: 9.341 ppbv
RT: 12.274 min Scan# 3765
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 91 Resp: 619610
Ion Ratio Lower Upper
91 100
92 54.4 43.5 65.3
134 23.8 19.0 28.6









#71
2-Chlorotoluene
Concen: 9.245 ppbv
RT: 10.895 min Scan# 311
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion: 91 Resp: 460118
Ion Ratio Lower Upper
91 100
126 29.2 11.7 46.7



#72
4-Ethyltoluene
Concen: 9.316 ppbv
RT: 11.118 min Scan# 3408
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion: 105 Resp: 410587
Ion Ratio Lower Upper
105 100
120 29.8 23.8 35.8
91 12.9 10.3 15.5
77 13.9 11.1 16.7





#73

1,3,5-Trimethylbenzene

Concen: 9.244 ppbv

RT: 11.196 min Scan# 3432

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_1

ClientSampleId :





VSTDICCC010

Tgt Ion:105 Resp: 341900

Ion Ratio Lower Upper

105 100

120 48.2 38.6 57.8



#74

1,2,4-Trimethylbenzene

Concen: 9.005 ppbv

RT: 11.529 min Scan# 3535

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Tgt Ion:105 Resp: 363476




Ion Ratio Lower Upper


105 100

120 52.1 41.7 62.5

91 87.5 70.0 105.0

77 29.8 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 9.335 ppbv
RT: 11.601 min Scan# 3557
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion:146 Resp: 219894
Ion Ratio Lower Upper
146 100
111 46.9 23.4 70.3
148 63.3 31.6 95.0



#76
1,4-Dichlorobenzene
Concen: 9.363 ppbv
RT: 11.665 min Scan# 3577
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:146 Resp: 220576
Ion Ratio Lower Upper
146 100
111 45.5 22.8 68.3
148 63.6 31.8 95.4









#77
1,2-Dichlorobenzene
Concen: 9.044 ppbv
RT: 11.940 min Scan# 3662
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02




Instrument :
MSVOA_1
ClientSampleId :
VSTDICCC010


Tgt Ion:146 Resp: 204826
Ion Ratio Lower Upper
146 100
111 48.7 24.3 73.0
148 63.1 31.6 94.7



#78
Hexachloro-1,3-Butadiene
Concen: 8.520 ppbv
RT: 13.876 min Scan# 4260
Delta R.T. 0.000 min
Lab File: VL043204.D
Acq: 18 Nov 2025 10:02

Tgt Ion:225 Resp: 161034
Ion Ratio Lower Upper
225 100
223 62.8 50.2 75.4
227 63.9 51.1 76.7





#79

Naphthalene

Concen: 9.795 ppbv

RT: 13.507 min Scan# 4146

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_1

ClientSampleId :

VSTDICCC010





Tgt Ion:128 Resp: 351148

Ion Ratio Lower Upper

128 100

127 12.4 9.9 14.9

129 10.5 8.4 12.6



#80

Naphthalene,2-methyl-

Concen: 9.895 ppbv

RT: 14.452 min Scan# 4438

Delta R.T. 0.000 min

Lab File: VL043204.D

Acq: 18 Nov 2025 10:02




Tgt Ion:142 Resp: 93113


Ion Ratio Lower Upper

142 100

141 83.9 67.1 100.7

115 39.9 31.9 47.9





#81

1,2,4-Trichlorobenzene

Concen: 9.923 ppbv

RT: 13.436 min Scan# 41

Delta R.T. 0.000 min

Lab File: VL043204.D


Acq: 18 Nov 2025 10:02

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010





Tgt Ion:180 Resp: 159091

Ion Ratio Lower Upper

180 100

182 96.8 77.4 116.2

184 30.3 24.2 36.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043205.D
 Acq On : 18 Nov 2025 10:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Quant Time: Nov 19 00:50:19 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	104644	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	284629	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	230646	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.374 95 171009 10.115 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	30881	2.092	ppbv	98
3) Chlorodifluoromethane	1.476	51	30069	2.072	ppbv	98
4) Chloromethane	1.534	50	11890	2.155	ppbv	96
5) Vinyl Chloride	1.580	62	12040	1.995	ppbv	96
6) Bromomethane	1.670	94	5981	2.044	ppbv	99
7) Chloroethane	1.706	64	4943	2.103	ppbv	95
8) Dichlorotetrafluoroethane	1.554	85	26409	2.151	ppbv	98
9) Propene	1.486	41	12993	1.930	ppbv #	1
10) Heptane	4.978	43	17701	0.958	ppbv #	28
11) Trichlorofluoromethane	1.878	101	31124	2.177	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	23336	2.090	ppbv	100
13) Ethanol	1.722	45	5266	2.430	ppbv	86
14) Bromoethene	1.784	108	9129	2.133	ppbv	95
15) Acetone	1.835	43	31785	2.186	ppbv	91
16) 1,3-Butadiene	1.609	39	13101	2.021	ppbv	89
17) tert-Butyl alcohol	2.039	59	39348	2.062	ppbv	95
18) 1,1-Dichloroethene	2.036	96	10780	2.074	ppbv	94
19) Isopropyl Alcohol	1.887	45	17229	2.097	ppbv #	71
20) Methylene Chloride	2.062	84	12238	2.248	ppbv	96
21) Allyl Chloride	2.098	41	19580	2.002	ppbv	95
22) trans-1,2-Dichloroethene	2.340	96	12537	2.110	ppbv	94
23) Vinyl Acetate	2.460	43	35205	2.006	ppbv #	97
24) 1,1-Dichloroethane	2.408	63	24987	2.179	ppbv	99
25) Ethyl Acetate	2.832	43	64343	2.048	ppbv	99
26) Hexane	2.826	57	28776	2.093	ppbv	97
27) Carbon Disulfide	2.149	76	29964	2.135	ppbv #	91
28) Methyl tert-Butyl Ether	2.438	73	14101	2.090	ppbv #	85
29) Chloroform	2.849	83	40355	2.065	ppbv	98
30) Cyclohexane	3.855	84	23659	2.053	ppbv	98
31) cis-1,2-Dichloroethene	2.722	61	27663	2.024	ppbv	98
32) 1,1,1-Trichloroethane	3.366	97	42547	2.024	ppbv	99
34) 2-Butanone	2.557	43	41884	2.085	ppbv	99
35) Carbon Tetrachloride	3.765	117	41949	1.942	ppbv	96
36) Benzene	3.654	78	36362	1.334	ppbv #	94
37) 1,2-Dichloroethane	3.218	62	32188	2.019	ppbv	100
38) Trichloroethene	4.541	130	10841	1.015	ppbv	89
39) 1,2-Dichloropropane	4.312	63	19913	1.996	ppbv	93
40) 1,4-Dioxane	4.580	88	3232	0.721	ppbv #	18
41) Tetrahydrofuran	3.056	42	22912	2.065	ppbv	99
42) Bromodichloromethane	4.490	83	44584	2.027	ppbv	98
43) Methyl Methacrylate	4.875	69	29071	2.009	ppbv	99
44) 2,2,4-Trimethylpentane	4.635	57	92261	1.971	ppbv	96
45) t-1,3-Dichloropropene	6.406	75	14338	1.063	ppbv	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043205.D
 Acq On : 18 Nov 2025 10:37
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002


Quant Time: Nov 19 00:50:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	31986	2.020	ppbv	91
47) 1,1,2-Trichloroethane	6.577	97	12775	1.201	ppbv	96
48) Dibromochloromethane	7.386	129	16152	0.887	ppbv #	10
49) Bromoform	9.480	173	33057	2.050	ppbv	98
50) 4-Methyl-2-Pentanone	5.745	43	58461	2.066	ppbv	98
51) 2-Hexanone	7.432	43	46444	2.113	ppbv #	99
52) Tetrachloroethene	8.218	164	19856	1.958	ppbv	96
53) Toluene	6.927	91	65238	2.008	ppbv	97
54) 1,2-Dibromoethane	7.648	107	31266	2.007	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.920	131	35099	2.071	ppbv	99
57) Chlorobenzene	8.917	112	50228	2.121	ppbv	97
58) Ethyl Benzene	9.351	91	92649	2.087	ppbv	98
59) m/p-Xylene	9.545	91	65660	1.838	ppbv #	31
60) o-Xylene	9.959	91	76540	2.153	ppbv	100
61) Styrene	9.866	104	33188	2.133	ppbv	98
62) Isopropylbenzene	10.536	105	135853	2.136	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.963	83	47439	2.102	ppbv	99
64) n-propylbenzene	10.985	120	34915	2.136	ppbv	99
65) tert-Butylbenzene	11.529	119	122064	2.149	ppbv	99
66) Benzyl Chloride	11.610	91	11870	2.059	ppbv	99
67) sec-Butylbenzene	11.765	105	170529	2.151	ppbv	100
69) p-Isopropyltoluene	11.921	119	147482	2.195	ppbv	98
70) n-Butylbenzene	12.280	91	144148	2.212	ppbv	99
71) 2-Chlorotoluene	10.898	91	101951	2.085	ppbv	100
72) 4-Ethyltoluene	11.121	105	92654	2.140	ppbv	99
73) 1,3,5-Trimethylbenzene	11.202	105	76655	2.110	ppbv	96
74) 1,2,4-Trimethylbenzene	11.532	105	86286	2.176	ppbv	98
75) 1,3-Dichlorobenzene	11.604	146	48961	2.116	ppbv	98
76) 1,4-Dichlorobenzene	11.668	146	48881	2.112	ppbv	98
77) 1,2-Dichlorobenzene	11.944	146	48431	2.177	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	42456	2.287	ppbv	99
79) Naphthalene	13.510	128	85788	2.436	ppbv	97
80) Naphthalene,2-methyl-	14.458	142	20685	2.237	ppbv	98
81) 1,2,4-Trichlorobenzene	13.439	180	37521	2.382	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002





- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.787 min Scan# 81
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002


Tgt Ion: 49 Resp: 104644
Ion Ratio Lower Upper
49 100
128 46.4 23.0 69.0
130 59.1 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 2.092 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 85 Resp: 30881
Ion Ratio Lower Upper
85 100
87 33.1 25.4 38.2









#3
Chlorodifluoromethane
Concen: 2.072 ppbv
RT: 1.476 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002


Tgt Ion: 51 Resp: 30069
Ion Ratio Lower Upper
51 100
67 19.4 14.9 22.3



#4
Chloromethane
Concen: 2.155 ppbv
RT: 1.534 min Scan# 447
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 50 Resp: 11890
Ion Ratio Lower Upper
50 100
52 32.0 24.1 36.1









#5
 Vinyl Chloride
 Concen: 1.995 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC002


Tgt Ion: 62 Resp: 12040
 Ion Ratio Lower Upper
 62 100
 64 32.1 27.3 40.9



#6
 Bromomethane
 Concen: 2.044 ppbv
 RT: 1.670 min Scan# 489
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37

Tgt Ion: 94 Resp: 5981
 Ion Ratio Lower Upper
 94 100
 96 92.4 72.9 109.3









#7
Chloroethane
Concen: 2.103 ppbv
RT: 1.706 min Scan# 50
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion: 64 Resp: 4943
Ion Ratio Lower Upper
64 100
66 30.8 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 2.151 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 85 Resp: 26409
Ion Ratio Lower Upper
85 100
135 64.3 52.6 79.0





#9

Propene

Concen: 1.930 ppbv

RT: 1.486 min Scan# 41

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 41 Resp: 12993

Ion Ratio Lower Upper

41 100

42 75.1 3.7 5.5#



#10

Heptane

Concen: 0.958 ppbv

RT: 4.978 min Scan# 1511

Delta R.T. 0.013 min

Lab File: VL043205.D




Acq: 18 Nov 2025 10:37


Tgt Ion: 43 Resp: 17701

Ion Ratio Lower Upper

43 100

57 0.0 39.7 59.5#









#11
Trichlorofluoromethane
Concen: 2.177 ppbv
RT: 1.878 min Scan# 511
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion:101 Resp: 31124
Ion Ratio Lower Upper
101 100
103 63.6 52.6 79.0



#12
1,1,2-Trichlorotrifluoroethane
Concen: 2.090 ppbv
RT: 2.140 min Scan# 634
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:101 Resp: 23336
Ion Ratio Lower Upper
101 100
151 73.8 59.2 88.8









#13
Ethanol
Concen: 2.430 ppbv
RT: 1.722 min Scan# 503
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 45 Resp: 5266
Ion Ratio Lower Upper
45 100
46 45.7 15.0 59.8



#14
Bromoethene
Concen: 2.133 ppbv
RT: 1.784 min Scan# 524
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 108 Resp: 9129
Ion Ratio Lower Upper
108 100
106 106.6 88.9 133.3
79 24.3 16.5 24.7









#15
Acetone
Concen: 2.186 ppbv
RT: 1.835 min Scan# 54
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion: 43 Resp: 31785
Ion Ratio Lower Upper
43 100
58 30.6 20.7 31.1



#16
1,3-Butadiene
Concen: 2.021 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 39 Resp: 13101
Ion Ratio Lower Upper
39 100
54 83.3 59.4 89.2









#17
 tert-Butyl alcohol
 Concen: 2.062 ppbv
 RT: 2.039 min Scan# 603
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC002


Tgt Ion: 59 Resp: 39348
 Ion Ratio Lower Upper
 59 100
 57 13.3 9.0 13.4



#18
 1,1-Dichloroethene
 Concen: 2.074 ppbv
 RT: 2.036 min Scan# 602
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37

Tgt Ion: 96 Resp: 10780
 Ion Ratio Lower Upper
 96 100
 61 207.0 170.6 255.8
 98 74.8 52.2 78.4









#19
Isopropyl Alcohol
Concen: 2.097 ppbv
RT: 1.887 min Scan# 51
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC002


Tgt Ion: 45 Resp: 17229
Ion Ratio Lower Upper
45 100
58 56.6 31.3 46.9#



#20
Methylene Chloride
Concen: 2.248 ppbv
RT: 2.062 min Scan# 610
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 84 Resp: 12238
Ion Ratio Lower Upper
84 100
49 166.8 132.9 199.3
51 46.5 34.8 52.2
86 68.8 48.2 72.2





#21

Allyl Chloride

Concen: 2.002 ppbv

RT: 2.098 min Scan# 619

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC002





Tgt Ion: 41 Resp: 19580

Ion Ratio Lower Upper

41 100

76 28.6 22.1 33.1

39 81.8 61.0 91.4



#22

trans-1,2-Dichloroethene

Concen: 2.110 ppbv

RT: 2.340 min Scan# 696

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37




Tgt Ion: 96 Resp: 12537


Ion Ratio Lower Upper

96 100

61 183.6 139.0 208.6

98 64.3 49.4 74.2









#23
 Vinyl Acetate
 Concen: 2.006 ppbv
 RT: 2.460 min Scan# 717
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC002


Tgt Ion:	43	Resp:	35205
Ion	Ratio	Lower	Upper
43	100		
86	6.8	6.2	9.2
42	9.5	8.0	12.0
44	6.5	4.0	6.0



#24
 1,1-Dichloroethane
 Concen: 2.179 ppbv
 RT: 2.408 min Scan# 717
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37

Tgt Ion:	63	Resp:	24987
Ion	Ratio	Lower	Upper
63	100		
98	5.3	2.8	8.4
100	3.3	1.3	3.9





#25

Ethyl Acetate

Concen: 2.048 ppbv

RT: 2.832 min Scan# 84

Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC002

Tgt Ion: 43 Resp: 64343





Ion Ratio Lower Upper

43 100

45 9.6 8.1 12.1

61 8.5 7.0 10.4

70 8.5 7.0 10.6



#26

Hexane

Concen: 2.093 ppbv

RT: 2.826 min Scan# 846

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion: 57 Resp: 28776




Ion Ratio Lower Upper


57 100

41 96.3 74.6 111.8

43 221.0 179.0 268.4

56 56.5 42.3 63.5





#27

Carbon Disulfide

Concen: 2.135 ppbv

RT: 2.149 min Scan# 61

Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37





Instrument :

MSVOA_L

ClientSampleId :

VSTDICC002

Tgt Ion: 76	Resp: 29964
Ion Ratio	Lower Upper
76 100	
44 26.0	16.1 24.1#
78 13.5	9.6 14.4



#28

Methyl tert-Butyl Ether

Concen: 2.090 ppbv




RT: 2.438 min Scan# 726


Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion: 73	Resp: 14101
Ion Ratio	Lower Upper
73 100	
57 32.5	19.8 29.8#





#29

Chloroform

Concen: 2.065 ppbv

RT: 2.849 min Scan# 851

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 83 Resp: 40355

Ion Ratio Lower Upper

83 100

85 65.0 50.5 75.7



#30

Cyclohexane

Concen: 2.053 ppbv

RT: 3.855 min Scan# 1164

Delta R.T. 0.010 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37




Tgt Ion: 84 Resp: 23659


Ion Ratio Lower Upper

84 100

56 129.3 104.7 157.1

41 89.8 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 2.024 ppbv
 RT: 2.722 min Scan# 81
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC002


Tgt Ion: 61 Resp: 27663
 Ion Ratio Lower Upper
 61 100
 96 59.9 49.3 73.9
 98 39.6 32.1 48.1



#32
 1,1,1-Trichloroethane
 Concen: 2.024 ppbv
 RT: 3.366 min Scan# 1013
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37

Tgt Ion: 97 Resp: 42547
 Ion Ratio Lower Upper
 97 100
 99 62.8 50.3 75.5
 61 50.4 38.9 58.3





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.955 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_1

ClientSampleId :

VSTDICC002





Tgt Ion:114 Resp: 284629

Ion Ratio Lower Upper

114 100

63 25.7 21.0 31.6

88 18.2 14.9 22.3



#34

2-Butanone

Concen: 2.085 ppbv

RT: 2.557 min Scan# 763

Delta R.T. 0.006 min

Lab File: VL043205.D




Acq: 18 Nov 2025 10:37


Tgt Ion: 43 Resp: 41884

Ion Ratio Lower Upper

43 100

72 22.9 18.1 27.1









#35
Carbon Tetrachloride
Concen: 1.942 ppbv
RT: 3.765 min Scan# 1136
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument : MSVOA_L
ClientSampleId : VSTDIC002


Tgt Ion:117 Resp: 41949
Ion Ratio Lower Upper
117 100
119 97.8 74.6 112.0
121 29.6 23.2 34.8



#36
Benzene
Concen: 1.334 ppbv
RT: 3.654 min Scan# 1102
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 78 Resp: 36362
Ion Ratio Lower Upper
78 100
77 23.0 17.5 26.3
50 24.1 15.5 23.3#





#37

1,2-Dichloroethane

Concen: 2.019 ppbv

RT: 3.218 min Scan# 9

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 62 Resp: 32188

Ion Ratio Lower Upper

62 100

98 7.0 5.5 8.3



#38

Trichloroethene

Concen: 1.015 ppbv

RT: 4.541 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion:130 Resp: 10841




Ion Ratio Lower Upper


130 100

95 120.4 85.7 128.5

132 103.5 73.4 110.0

97 71.6 52.7 79.1





#39

1,2-Dichloropropane

Concen: 1.996 ppbv

RT: 4.312 min Scan# 11

Delta R.T. 0.010 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002





Tgt Ion: 63 Resp: 19913

Ion Ratio Lower Upper

63 100

41 83.9 62.6 93.8

62 69.5 51.0 76.6



#40

1,4-Dioxane

Concen: 0.721 ppbv

RT: 4.580 min Scan# 1388

Delta R.T. 0.013 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion: 88 Resp: 3232




Ion Ratio Lower Upper


88 100

58 0.0 63.0 94.6#

43 0.0 32.5 48.7#

57 0.0 0.0 0.0





#41

Tetrahydrofuran

Concen: 2.065 ppbv

RT: 3.056 min Scan# 914

Delta R.T. 0.010 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC002





Tgt Ion: 42 Resp: 22912

Ion Ratio Lower Upper

42 100

72 38.8 30.6 46.0

71 37.9 30.5 45.7



#42

Bromodichloromethane

Concen: 2.027 ppbv

RT: 4.490 min Scan# 1360

Delta R.T. 0.013 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37




Tgt Ion: 83 Resp: 44584


Ion Ratio Lower Upper

83 100

85 66.6 51.7 77.5

127 7.9 6.3 9.5









#43
Methyl Methacrylate
Concen: 2.009 ppbv
RT: 4.875 min Scan# 1476
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion: 69 Resp: 29071
Ion Ratio Lower Upper
69 100
41 157.2 126.7 190.1
100 36.0 28.6 43.0



#44
2,2,4-Trimethylpentane
Concen: 1.971 ppbv
RT: 4.635 min Scan# 1405
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 57 Resp: 92261
Ion Ratio Lower Upper
57 100
41 37.8 28.0 42.0
56 33.9 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 1.063 ppbv
RT: 6.406 min Scan# 1952
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion: 75 Resp: 14338
Ion Ratio Lower Upper
75 100
77 27.8 24.3 36.5



#46
cis-1,3-Dichloropropene
Concen: 2.020 ppbv
RT: 5.593 min Scan# 1701
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 75 Resp: 31986
Ion Ratio Lower Upper
75 100
39 67.2 59.1 88.7
77 24.2 24.2 36.2





#47

1,1,2-Trichloroethane

Concen: 1.201 ppbv

RT: 6.577 min Scan# 2001

Delta R.T. 0.013 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002

Tgt Ion: 97 Resp: 12775





Ion Ratio Lower Upper

97 100

83 83.8 64.6 97.0

85 52.8 39.4 59.2

99 55.6 47.0 70.6



#48

Dibromochloromethane

Concen: 0.887 ppbv

RT: 7.386 min Scan# 2255

Delta R.T. 0.010 min

Lab File: VL043205.D




Acq: 18 Nov 2025 10:37


Tgt Ion: 129 Resp: 16152

Ion Ratio Lower Upper

129 100

127 0.0 39.3 117.9#









#49
Bromoform
Concen: 2.050 ppbv
RT: 9.480 min Scan# 2900
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion:173 Resp: 33057
Ion Ratio Lower Upper
173 100
175 46.6 38.6 58.0
171 52.4 40.6 61.0



#50
4-Methyl-2-Pentanone
Concen: 2.066 ppbv
RT: 5.745 min Scan# 1748
Delta R.T. 0.013 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 43 Resp: 58461
Ion Ratio Lower Upper
43 100
58 36.7 28.4 42.6









#51
2-Hexanone
Concen: 2.113 ppbv
RT: 7.432 min Scan# 21
Delta R.T. 0.013 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 43 Resp: 46444
Ion Ratio Lower Upper
43 100
58 48.4 38.4 57.6
98 0.2 0.0 0.0#



#52
Tetrachloroethene
Concen: 1.958 ppbv
RT: 8.218 min Scan# 2512
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 164 Resp: 19856
Ion Ratio Lower Upper
164 100
166 117.2 99.4 149.2
129 96.5 81.8 122.8
131 97.0 77.4 116.2









#53
Toluene
Concen: 2.008 ppbv
RT: 6.927 min Scan# 2111
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion: 91 Resp: 65238
Ion Ratio Lower Upper
91 100
92 61.5 47.4 71.0



#54
1,2-Dibromoethane
Concen: 2.007 ppbv
RT: 7.648 min Scan# 2336
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 107 Resp: 31266
Ion Ratio Lower Upper
107 100
109 88.3 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.878 min Scan# 2117
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC002


Tgt Ion:117 Resp: 230646
Ion Ratio Lower Upper
117 100
82 67.2 56.1 84.1
119 32.0 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 2.071 ppbv
RT: 8.920 min Scan# 2729
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:131 Resp: 35099
Ion Ratio Lower Upper
131 100
133 97.4 77.4 116.2
119 60.1 49.1 73.7









#57
Chlorobenzene
Concen: 2.121 ppbv
RT: 8.917 min Scan# 21
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion:	112	Resp:	50228
Ion Ratio	Lower	Upper	
112	100		
77	71.6	55.5	83.3
114	33.6	29.1	35.5



#58
Ethyl Benzene
Concen: 2.087 ppbv
RT: 9.351 min Scan# 2862
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:	91	Resp:	92649
Ion Ratio	Lower	Upper	
91	100		
106	29.8	22.8	34.2





#59

m/p-Xylene

Concen: 1.838 ppbv

RT: 9.545 min Scan# 29

Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 91 Resp: 65660

Ion Ratio Lower Upper

91 100

106 0.0 36.1 54.1#



#60

o-Xylene

Concen: 2.153 ppbv

RT: 9.959 min Scan# 3050

Delta R.T. 0.003 min

Lab File: VL043205.D




Acq: 18 Nov 2025 10:37


Tgt Ion: 91 Resp: 76540

Ion Ratio Lower Upper

91 100

106 41.9 20.9 62.8





#61

Styrene

Concen: 2.133 ppbv

RT: 9.866 min Scan# 3020

Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC002





Tgt Ion:104 Resp: 33188

Ion Ratio Lower Upper

104 100

78 52.1 42.1 63.1

103 44.9 37.2 55.8



#62

Isopropylbenzene

Concen: 2.136 ppbv

RT: 10.536 min Scan# 3228

Delta R.T. 0.003 min

Lab File: VL043205.D




Acq: 18 Nov 2025 10:37


Tgt Ion:105 Resp: 135853

Ion Ratio Lower Upper

105 100

120 25.6 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 2.102 ppbv
RT: 9.963 min Scan# 3048
Delta R.T. 0.010 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument : MSVOA_1
ClientSampleId : VSTDIC002


Tgt Ion: 83 Resp: 47439
Ion Ratio Lower Upper
83 100
131 10.0 4.8 14.3
85 63.4 31.9 95.9



#64
n-propylbenzene
Concen: 2.136 ppbv
RT: 10.985 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 120 Resp: 34915
Ion Ratio Lower Upper
120 100
91 467.6 377.1 565.7









#65
tert-Butylbenzene
Concen: 2.149 ppbv
RT: 11.529 min Scan# 3535
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion:119 Resp: 122064
Ion Ratio Lower Upper
119 100
91 82.0 65.3 97.9
134 22.1 17.0 25.4



#66
Benzyl Chloride
Concen: 2.059 ppbv
RT: 11.610 min Scan# 3560
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 91 Resp: 11870
Ion Ratio Lower Upper
91 100
126 18.3 14.4 21.6
128 5.9 4.6 6.8









#67
 sec-Butylbenzene
 Concen: 2.151 ppbv
 RT: 11.765 min Scan# 3178
 Delta R.T. 0.003 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC002


Tgt Ion:105 Resp: 170529
 Ion Ratio Lower Upper
 105 100
 134 18.9 15.0 22.6



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.115 ppbv
 RT: 10.374 min Scan# 3178
 Delta R.T. 0.006 min
 Lab File: VL043205.D
 Acq: 18 Nov 2025 10:37

Tgt Ion: 95 Resp: 171009
 Ion Ratio Lower Upper
 95 100
 174 62.1 50.2 75.2
 175 4.9 4.1 6.1









#69
p-Isopropyltoluene
Concen: 2.195 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion:119 Resp: 147482
Ion Ratio Lower Upper
119 100
134 24.2 20.2 30.2
91 29.1 23.8 35.8



#70
n-Butylbenzene
Concen: 2.212 ppbv
RT: 12.280 min Scan# 3767
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 91 Resp: 144148
Ion Ratio Lower Upper
91 100
92 53.3 43.5 65.3
134 23.8 19.0 28.6









#71
2-Chlorotoluene
Concen: 2.085 ppbv
RT: 10.898 min Scan# 31
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion: 91 Resp: 101951
Ion Ratio Lower Upper
91 100
126 29.0 11.7 46.7



#72
4-Ethyltoluene
Concen: 2.140 ppbv
RT: 11.121 min Scan# 3409
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion: 105 Resp: 92654
Ion Ratio Lower Upper
105 100
120 29.1 23.8 35.8
91 13.7 10.3 15.5
77 13.8 11.1 16.7





#73

1,3,5-Trimethylbenzene

Concen: 2.110 ppbv

RT: 11.202 min Scan# 3432

Delta R.T. 0.006 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion:105 Resp: 76655

Ion Ratio Lower Upper

105 100

120 45.5 38.6 57.8



#74

1,2,4-Trimethylbenzene

Concen: 2.176 ppbv

RT: 11.532 min Scan# 3536

Delta R.T. 0.003 min

Lab File: VL043205.D

Acq: 18 Nov 2025 10:37

Tgt Ion:105 Resp: 86286




Ion Ratio Lower Upper


105 100

120 50.0 41.7 62.5

91 85.2 70.0 105.0

77 29.5 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 2.116 ppbv
RT: 11.604 min Scan# 3557
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion:146 Resp: 48961
Ion Ratio Lower Upper
146 100
111 45.7 23.4 70.3
148 64.6 31.6 95.0



#76
1,4-Dichlorobenzene
Concen: 2.112 ppbv
RT: 11.668 min Scan# 3578
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:146 Resp: 48881
Ion Ratio Lower Upper
146 100
111 46.3 22.8 68.3
148 65.3 31.8 95.4









#77
1,2-Dichlorobenzene
Concen: 2.177 ppbv
RT: 11.944 min Scan# 3663
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion:	146	Resp:	48431
Ion	Ratio	Lower	Upper
146	100		
111	48.5	24.3	73.0
148	61.3	31.6	94.7



#78
Hexachloro-1,3-Butadiene
Concen: 2.287 ppbv
RT: 13.879 min Scan# 4261
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:	225	Resp:	42456
Ion	Ratio	Lower	Upper
225	100		
223	62.1	50.2	75.4
227	63.4	51.1	76.7









#79
Naphthalene
Concen: 2.436 ppbv
RT: 13.510 min Scan# 4146
Delta R.T. 0.003 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC002


Tgt Ion:	128	Resp:	85788
Ion Ratio	Lower	Upper	
128	100		
127	13.3	9.9	14.9
129	11.7	8.4	12.6



#80
Naphthalene, 2-methyl-
Concen: 2.237 ppbv
RT: 14.458 min Scan# 4440
Delta R.T. 0.006 min
Lab File: VL043205.D
Acq: 18 Nov 2025 10:37

Tgt Ion:	142	Resp:	20685
Ion Ratio	Lower	Upper	
142	100		
141	82.9	67.1	100.7
115	36.9	31.9	47.9





#81

1,2,4-Trichlorobenzene

Concen: 2.382 ppbv

RT: 13.439 min Scan# 41

Delta R.T. 0.003 min

Lab File: VL043205.D


Acq: 18 Nov 2025 10:37

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002





Tgt Ion:180 Resp: 37521

Ion Ratio Lower Upper

180 100

182 95.8 77.4 116.2

184 30.4 24.2 36.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043206.D
 Acq On : 18 Nov 2025 11:10
 Operator : SY/MD
 Sample : VSTDIC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC001

Quant Time: Nov 19 00:51:06 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	102795	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	278732	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	227998	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 167061 9.996 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.000%

Target Compounds

					Qvalue	
2) Dichlorodifluoromethane	1.499	85	15474	1.067	ppbv	100
3) Chlorodifluoromethane	1.476	51	14437	1.013	ppbv #	92
4) Chloromethane	1.531	50	5751	1.061	ppbv	92
5) Vinyl Chloride	1.580	62	5967	1.007	ppbv	95
6) Bromomethane	1.667	94	3104	1.080	ppbv	94
7) Chloroethane	1.706	64	2550	1.104	ppbv	97
8) Dichlorotetrafluoroethane	1.554	85	12759	1.058	ppbv	98
9) Propene	1.483	41	6147	0.929	ppbv #	1
10) Heptane	4.972	43	18544	1.022	ppbv	100
11) Trichlorofluoromethane	1.877	101	15571	1.108	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.140	101	11743	1.071	ppbv	99
13) Ethanol	1.719	45	2373	1.115	ppbv	67
14) Bromoethene	1.780	108	4345	1.033	ppbv #	91
15) Acetone	1.835	43	16353	1.145	ppbv	90
16) 1,3-Butadiene	1.609	39	6469	1.016	ppbv	96
17) tert-Butyl alcohol	2.039	59	20703	1.104	ppbv #	92
18) 1,1-Dichloroethene	2.033	96	5447	1.067	ppbv	97
19) Isopropyl Alcohol	1.884	45	8849	1.097	ppbv #	43
20) Methylene Chloride	2.059	84	7297	1.030	ppbv	85
21) Allyl Chloride	2.094	41	9198	0.957	ppbv #	84
22) trans-1,2-Dichloroethene	2.340	96	6313	1.082	ppbv	94
23) Vinyl Acetate	2.460	43	16779	0.973	ppbv	98
24) 1,1-Dichloroethane	2.405	63	12062	1.071	ppbv	94
25) Ethyl Acetate	2.832	43	32108	1.041	ppbv	98
26) Hexane	2.819	57	14639	1.084	ppbv	99
27) Carbon Disulfide	2.149	76	14637	1.062	ppbv #	30
28) Methyl tert-Butyl Ether	2.441	73	7236	1.092	ppbv #	83
29) Chloroform	2.845	83	19509	1.016	ppbv	91
30) Cyclohexane	3.852	84	12269	1.084	ppbv	96
31) cis-1,2-Dichloroethene	2.716	61	13889	1.035	ppbv	97
32) 1,1,1-Trichloroethane	3.363	97	20745	1.005	ppbv	99
34) 2-Butanone	2.554	43	20940	1.064	ppbv	96
35) Carbon Tetrachloride	3.758	117	21497	1.016	ppbv	96
36) Benzene	3.648	78	27911	1.046	ppbv	99
37) 1,2-Dichloroethane	3.214	62	16159	1.035	ppbv	99
38) Trichloroethene	4.544	130	4331	0.414	ppbv	94
39) 1,2-Dichloropropane	4.311	63	10147	1.039	ppbv	91
40) 1,4-Dioxane	4.577	88	1369	0.312	ppbv #	1
41) Tetrahydrofuran	3.052	42	11176	1.029	ppbv	96
42) Bromodichloromethane	4.480	83	22150	1.028	ppbv #	98
43) Methyl Methacrylate	4.871	69	14821	1.046	ppbv	94
44) 2,2,4-Trimethylpentane	4.632	57	30372	0.662	ppbv #	60
45) t-1,3-Dichloropropene	6.402	75	8750	0.663	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043206.D
 Acq On : 18 Nov 2025 11:10
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Quant Time: Nov 19 00:51:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.590	75	15527	1.001	ppbv #	95
47) 1,1,2-Trichloroethane	6.577	97	10369	0.995	ppbv #	85
48) Dibromochloromethane	7.377	129	18560	1.041	ppbv	93
49) Bromoform	9.477	173	15858	1.004	ppbv	98
50) 4-Methyl-2-Pentanone	5.745	43	29560	1.067	ppbv	99
51) 2-Hexanone	7.428	43	22438	1.043	ppbv #	98
52) Tetrachloroethene	8.215	164	10080	1.015	ppbv	90
53) Toluene	6.923	91	32951	1.036	ppbv	100
54) 1,2-Dibromoethane	7.645	107	15865	1.040	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.917	131	17535	1.046	ppbv	99
57) Chlorobenzene	8.914	112	24806	1.060	ppbv	98
58) Ethyl Benzene	9.344	91	46455	1.059	ppbv	96
59) m/p-Xylene	9.545	91	32159	0.911	ppbv #	31
60) o-Xylene	9.956	91	37054	1.054	ppbv	97
61) Styrene	9.866	104	15730	1.023	ppbv	95
62) Isopropylbenzene	10.532	105	67163	1.068	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.956	83	23083	1.035	ppbv	97
64) n-propylbenzene	10.985	120	17289	1.070	ppbv	100
65) tert-Butylbenzene	11.526	119	61258	1.091	ppbv	99
66) Benzyl Chloride	11.610	91	5729	1.005	ppbv	96
67) sec-Butylbenzene	11.762	105	84989	1.084	ppbv	99
69) p-Isopropyltoluene	11.921	119	69993	1.054	ppbv	99
70) n-Butylbenzene	12.277	91	69323	1.076	ppbv	99
71) 2-Chlorotoluene	10.895	91	51452	1.065	ppbv	100
72) 4-Ethyltoluene	11.118	105	44403	1.037	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	38350	1.068	ppbv	91
74) 1,2,4-Trimethylbenzene	11.532	105	42051	1.073	ppbv #	81
75) 1,3-Dichlorobenzene	11.600	146	24011	1.050	ppbv	98
76) 1,4-Dichlorobenzene	11.665	146	23543	1.029	ppbv	97
77) 1,2-Dichlorobenzene	11.940	146	24248	1.103	ppbv	98
78) Hexachloro-1,3-Butadiene	13.879	225	20489	1.116	ppbv	99
79) Naphthalene	13.507	128	36462	1.047	ppbv	95
80) Naphthalene,2-methyl-	14.455	142	7375	0.807	ppbv	95
81) 1,2,4-Trichlorobenzene	13.436	180	15785	1.014	ppbv	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
Data File : VL043206.D
Acq On : 18 Nov 2025 11:10
Operator : SY/MD
Sample : VSTDIC001
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Quant Time: Nov 19 00:51:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC001


Tgt Ion: 49 Resp: 102795
Ion Ratio Lower Upper
49 100
128 46.9 23.0 69.0
130 60.1 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 1.067 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 85 Resp: 15474
Ion Ratio Lower Upper
85 100
87 31.9 25.4 38.2









#3
Chlorodifluoromethane
Concen: 1.013 ppbv
RT: 1.476 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion: 51 Resp: 14437
Ion Ratio Lower Upper
51 100
67 22.3 14.9 22.3#



#4
Chloromethane
Concen: 1.061 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 50 Resp: 5751
Ion Ratio Lower Upper
50 100
52 34.2 24.1 36.1









#5
Vinyl Chloride
Concen: 1.007 ppbv
RT: 1.580 min Scan# 461
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC001


Tgt Ion: 62 Resp: 5967
Ion Ratio Lower Upper
62 100
64 31.3 27.3 40.9



#6
Bromomethane
Concen: 1.080 ppbv
RT: 1.667 min Scan# 488
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 94 Resp: 3104
Ion Ratio Lower Upper
94 100
96 85.7 72.9 109.3









#7
Chloroethane
Concen: 1.104 ppbv
RT: 1.706 min Scan# 50
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion: 64 Resp: 2550
Ion Ratio Lower Upper
64 100
66 31.5 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 1.058 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 85 Resp: 12759
Ion Ratio Lower Upper
85 100
135 67.7 52.6 79.0









#9
 Propene
 Concen: 0.929 ppbv
 RT: 1.483 min Scan# 41
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC001


Tgt Ion: 41 Resp: 6147
 Ion Ratio Lower Upper
 41 100
 42 80.0 3.7 5.5#



#10
 Heptane
 Concen: 1.022 ppbv
 RT: 4.972 min Scan# 1509
 Delta R.T. 0.006 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Tgt Ion: 43 Resp: 18544
 Ion Ratio Lower Upper
 43 100
 57 49.3 39.7 59.5









#11
 Trichlorofluoromethane
 Concen: 1.108 ppbv
 RT: 1.877 min Scan# 51
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICC001


Tgt Ion:101 Resp: 15571
 Ion Ratio Lower Upper
 101 100
 103 65.2 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 1.071 ppbv
 RT: 2.140 min Scan# 634
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Tgt Ion:101 Resp: 11743
 Ion Ratio Lower Upper
 101 100
 151 74.6 59.2 88.8









#13
Ethanol
Concen: 1.115 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC001


Tgt Ion: 45 Resp: 2373
Ion Ratio Lower Upper
45 100
46 57.0 15.0 59.8



#14
Bromoethene
Concen: 1.033 ppbv
RT: 1.780 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 108 Resp: 4345
Ion Ratio Lower Upper
108 100
106 114.0 88.9 133.3
79 39.2 16.5 24.7#









#15
Acetone
Concen: 1.145 ppbv
RT: 1.835 min Scan# 54
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC001


Tgt Ion: 43 Resp: 16353
Ion Ratio Lower Upper
43 100
58 30.9 20.7 31.1



#16
1,3-Butadiene
Concen: 1.016 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 39 Resp: 6469
Ion Ratio Lower Upper
39 100
54 77.8 59.4 89.2





#17

tert-Butyl alcohol

Concen: 1.104 ppbv

RT: 2.039 min Scan# 601

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC001

Tgt Ion: 59 Resp: 20703

Ion Ratio Lower Upper

59 100

57 14.1 9.0 13.4#



#18

1,1-Dichloroethene

Concen: 1.067 ppbv

RT: 2.033 min Scan# 601

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10




Tgt Ion: 96 Resp: 5447


Ion Ratio Lower Upper

96 100

61 219.3 170.6 255.8

98 63.7 52.2 78.4









#19
Isopropyl Alcohol
Concen: 1.097 ppbv
RT: 1.884 min Scan# 51
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001


Tgt Ion: 45 Resp: 8849
Ion Ratio Lower Upper
45 100
58 4.0 31.3 46.9#



#20
Methylene Chloride
Concen: 1.030 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 84 Resp: 7297
Ion Ratio Lower Upper
84 100
49 142.0 132.9 199.3
51 47.3 34.8 52.2
86 71.4 48.2 72.2









#21
Allyl Chloride
Concen: 0.957 ppbv
RT: 2.094 min Scan# 619
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion: 41 Resp: 9198
Ion Ratio Lower Upper
41 100
76 33.5 22.1 33.1#
39 91.7 61.0 91.4#



#22
trans-1,2-Dichloroethene
Concen: 1.082 ppbv
RT: 2.340 min Scan# 696
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 96 Resp: 6313
Ion Ratio Lower Upper
96 100
61 165.0 139.0 208.6
98 65.1 49.4 74.2









#23
 Vinyl Acetate
 Concen: 0.973 ppbv
 RT: 2.460 min Scan# 716
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICC001


Tgt Ion:	43	Resp:	16779
Ion	Ratio	Lower	Upper
43	100		
86	8.8	6.2	9.2
42	10.4	8.0	12.0
44	5.9	4.0	6.0



#24
 1,1-Dichloroethane
 Concen: 1.071 ppbv
 RT: 2.405 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Tgt Ion:	63	Resp:	12062
Ion	Ratio	Lower	Upper
63	100		
98	2.9	2.8	8.4
100	2.5	1.3	3.9









#25
Ethyl Acetate
Concen: 1.041 ppbv
RT: 2.832 min Scan# 847
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion: 43	Resp: 32108
Ion Ratio	Lower Upper
43 100	
45 11.1	8.1 12.1
61 8.7	7.0 10.4
70 9.9	7.0 10.6



#26
Hexane
Concen: 1.084 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 57	Resp: 14639
Ion Ratio	Lower Upper
57 100	
41 96.8	74.6 111.8
43 222.8	179.0 268.4
56 52.3	42.3 63.5









#27
Carbon Disulfide
Concen: 1.062 ppbv
RT: 2.149 min Scan# 611
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC001


Tgt Ion: 76 Resp: 14637
Ion Ratio Lower Upper
76 100
44 69.6 16.1 24.1#
78 14.8 9.6 14.4#



#28
Methyl tert-Butyl Ether
Concen: 1.092 ppbv
RT: 2.441 min Scan# 727
Delta R.T. 0.006 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 73 Resp: 7236
Ion Ratio Lower Upper
73 100
57 33.1 19.8 29.8#





#29

Chloroform

Concen: 1.016 ppbv

RT: 2.845 min Scan# 81

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC001

Tgt Ion: 83 Resp: 19509

Ion Ratio Lower Upper

83 100

85 56.3 50.5 75.7



#30

Cyclohexane

Concen: 1.084 ppbv

RT: 3.852 min Scan# 1163

Delta R.T. 0.006 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10




Tgt Ion: 84 Resp: 12269


Ion Ratio Lower Upper

84 100

56 127.5 104.7 157.1

41 90.4 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 1.035 ppbv
 RT: 2.716 min Scan# 81
 Delta R.T. -0.000 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC001


Tgt Ion: 61 Resp: 13889
 Ion Ratio Lower Upper
 61 100
 96 60.4 49.3 73.9
 98 43.0 32.1 48.1



#32
 1,1,1-Trichloroethane
 Concen: 1.005 ppbv
 RT: 3.363 min Scan# 1012
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Tgt Ion: 97 Resp: 20745
 Ion Ratio Lower Upper
 97 100
 99 63.6 50.3 75.5
 61 48.3 38.9 58.3









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001


Tgt Ion:114 Resp: 278732
Ion Ratio Lower Upper
114 100
63 26.0 21.0 31.6
88 18.6 14.9 22.3



#34
2-Butanone
Concen: 1.064 ppbv
RT: 2.554 min Scan# 762
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 43 Resp: 20940
Ion Ratio Lower Upper
43 100
72 24.7 18.1 27.1





#35

Carbon Tetrachloride

Concen: 1.016 ppbv

RT: 3.758 min Scan# 1133

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC001





Tgt Ion:117 Resp: 21497

Ion Ratio Lower Upper

117 100

119 89.6 74.6 112.0

121 30.9 23.2 34.8



#36

Benzene

Concen: 1.046 ppbv

RT: 3.648 min Scan# 1100

Delta R.T. -0.000 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10




Tgt Ion: 78 Resp: 27911


Ion Ratio Lower Upper

78 100

77 21.5 17.5 26.3

50 18.6 15.5 23.3





#37

1,2-Dichloroethane

Concen: 1.035 ppbv

RT: 3.214 min Scan# 9

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC001

Tgt Ion: 62 Resp: 16159

Ion Ratio Lower Upper

62 100

98 6.7 5.5 8.3



#38

Trichloroethene

Concen: 0.414 ppbv

RT: 4.544 min Scan# 1377

Delta R.T. 0.009 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Tgt Ion:130 Resp: 4331




Ion Ratio Lower Upper


130 100

95 113.3 85.7 128.5

132 83.5 73.4 110.0

97 62.5 52.7 79.1





#39

1,2-Dichloropropane

Concen: 1.039 ppbv

RT: 4.311 min Scan# 11

Delta R.T. 0.009 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC001





Tgt Ion: 63 Resp: 10147

Ion Ratio Lower Upper

63 100

41 70.5 62.6 93.8

62 56.2 51.0 76.6



#40

1,4-Dioxane

Concen: 0.312 ppbv

RT: 4.577 min Scan# 1387

Delta R.T. 0.009 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Tgt Ion: 88 Resp: 1369




Ion Ratio Lower Upper


88 100

58 234.8 63.0 94.6#

43 0.0 32.5 48.7#

57 0.0 0.0 0.0





#41

Tetrahydrofuran

Concen: 1.029 ppbv

RT: 3.052 min Scan# 914

Delta R.T. 0.006 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC001





Tgt Ion: 42 Resp: 11176

Ion Ratio Lower Upper

42 100

72 40.2 30.6 46.0

71 40.4 30.5 45.7



#42

Bromodichloromethane

Concen: 1.028 ppbv

RT: 4.480 min Scan# 1357

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10




Tgt Ion: 83 Resp: 22150


Ion Ratio Lower Upper

83 100

85 63.8 51.7 77.5

127 5.6 6.3 9.5





#43

Methyl Methacrylate

Concen: 1.046 ppbv

RT: 4.871 min Scan# 1476

Delta R.T. 0.006 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC001





Tgt Ion: 69 Resp: 14821

Ion Ratio Lower Upper

69 100

41 148.1 126.7 190.1

100 35.7 28.6 43.0



#44

2,2,4-Trimethylpentane

Concen: 0.662 ppbv

RT: 4.632 min Scan# 1404

Delta R.T. 0.006 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10




Tgt Ion: 57 Resp: 30372


Ion Ratio Lower Upper

57 100

41 61.1 28.0 42.0#

56 51.9 25.6 38.4#









#45
 t-1,3-Dichloropropene
 Concen: 0.663 ppbv
 RT: 6.402 min Scan# 1951
 Delta R.T. 0.003 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC001


Tgt Ion: 75 Resp: 8750
 Ion Ratio Lower Upper
 75 100
 77 28.5 24.3 36.5



#46
 cis-1,3-Dichloropropene
 Concen: 1.001 ppbv
 RT: 5.590 min Scan# 1700
 Delta R.T. 0.006 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Tgt Ion: 75 Resp: 15527
 Ion Ratio Lower Upper
 75 100
 39 72.5 59.1 88.7
 77 37.9 24.2 36.2#





#47

1,1,2-Trichloroethane

Concen: 0.995 ppbv

RT: 6.577 min Scan# 2005

Delta R.T. 0.013 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC001

Tgt Ion: 97 Resp: 10369





Ion Ratio Lower Upper

97 100

83 97.9 64.6 97.0#

85 56.4 39.4 59.2

99 67.4 47.0 70.6



#48

Dibromochloromethane

Concen: 1.041 ppbv

RT: 7.377 min Scan# 2252

Delta R.T. -0.000 min

Lab File: VL043206.D




Acq: 18 Nov 2025 11:10


Tgt Ion: 129 Resp: 18560

Ion Ratio Lower Upper

129 100

127 72.8 39.3 117.9





#49

Bromoform

Concen: 1.004 ppbv

RT: 9.477 min Scan# 2900

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC001





Tgt Ion:173 Resp: 15858

Ion Ratio Lower Upper

173 100

175 47.9 38.6 58.0

171 53.5 40.6 61.0



#50

4-Methyl-2-Pentanone

Concen: 1.067 ppbv

RT: 5.745 min Scan# 1748

Delta R.T. 0.013 min

Lab File: VL043206.D




Acq: 18 Nov 2025 11:10


Tgt Ion: 43 Resp: 29560

Ion Ratio Lower Upper

43 100

58 36.1 28.4 42.6









#51
2-Hexanone
Concen: 1.043 ppbv
RT: 7.428 min Scan# 2111
Delta R.T. 0.009 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 22438
Ion Ratio Lower Upper
43 100
58 49.4 38.4 57.6
98 0.2 0.0 0.0#



#52
Tetrachloroethene
Concen: 1.015 ppbv
RT: 8.215 min Scan# 2511
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 164 Resp: 10080
Ion Ratio Lower Upper
164 100
166 116.5 99.4 149.2
129 96.3 81.8 122.8
131 113.5 77.4 116.2





#53

Toluene

Concen: 1.036 ppbv

RT: 6.923 min Scan# 2111

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC001

Tgt Ion: 91 Resp: 32951

Ion Ratio Lower Upper

91 100

92 59.0 47.4 71.0



#54

1,2-Dibromoethane

Concen: 1.040 ppbv

RT: 7.645 min Scan# 2335

Delta R.T. 0.006 min

Lab File: VL043206.D




Acq: 18 Nov 2025 11:10


Tgt Ion: 107 Resp: 15865

Ion Ratio Lower Upper

107 100

109 93.3 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 21
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC001


Tgt Ion:117 Resp: 227998
Ion Ratio Lower Upper
117 100
82 66.2 56.1 84.1
119 32.1 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 1.046 ppbv
RT: 8.917 min Scan# 2728
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:131 Resp: 17535
Ion Ratio Lower Upper
131 100
133 97.3 77.4 116.2
119 60.6 49.1 73.7









#57
Chlorobenzene
Concen: 1.060 ppbv
RT: 8.914 min Scan# 21
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion	Ratio	Lower	Upper
112	100		
77	68.6	55.5	83.3
114	29.8	29.1	35.5



#58
Ethyl Benzene
Concen: 1.059 ppbv
RT: 9.344 min Scan# 2860
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion	Ratio	Lower	Upper
91	100		
106	26.6	22.8	34.2





#59

m/p-Xylene

Concen: 0.911 ppbv

RT: 9.545 min Scan# 2921

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC001

Tgt Ion: 91 Resp: 32159

Ion Ratio Lower Upper

91 100

106 0.0 36.1 54.1#



#60

o-Xylene

Concen: 1.054 ppbv

RT: 9.956 min Scan# 3049

Delta R.T. -0.000 min

Lab File: VL043206.D




Acq: 18 Nov 2025 11:10


Tgt Ion: 91 Resp: 37054

Ion Ratio Lower Upper

91 100

106 44.1 20.9 62.8





#61

Styrene

Concen: 1.023 ppbv

RT: 9.866 min Scan# 3020

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_1

ClientSampleId :

VSTDICC001





Tgt Ion:104 Resp: 15730

Ion Ratio Lower Upper

104 100

78 56.9 42.1 63.1

103 48.6 37.2 55.8



#62

Isopropylbenzene

Concen: 1.068 ppbv

RT: 10.532 min Scan# 3227

Delta R.T. -0.000 min

Lab File: VL043206.D




Acq: 18 Nov 2025 11:10


Tgt Ion:105 Resp: 67163

Ion Ratio Lower Upper

105 100

120 25.2 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 1.035 ppbv
RT: 9.956 min Scan# 3048
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion:	83	Resp:	23083
Ion	Ratio	Lower	Upper
83	100		
131	10.0	4.8	14.3
85	66.2	31.9	95.9



#64
n-propylbenzene
Concen: 1.070 ppbv
RT: 10.985 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:	120	Resp:	17289
Ion	Ratio	Lower	Upper
120	100		
91	471.4	377.1	565.7









#65
tert-Butylbenzene
Concen: 1.091 ppbv
RT: 11.526 min Scan# 3533
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC001


Tgt Ion:119 Resp: 61258
Ion Ratio Lower Upper
119 100
91 81.7 65.3 97.9
134 20.0 17.0 25.4



#66
Benzyl Chloride
Concen: 1.005 ppbv
RT: 11.610 min Scan# 3560
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 91 Resp: 5729
Ion Ratio Lower Upper
91 100
126 16.3 14.4 21.6
128 6.5 4.6 6.8





#67

sec-Butylbenzene

Concen: 1.084 ppbv

RT: 11.762 min Scan# 3177

Delta R.T. -0.000 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC001

Tgt Ion:105 Resp: 84989

Ion Ratio Lower Upper

105 100

134 19.3 15.0 22.6



#68

1-Bromo-4-Fluorobenzene

Concen: 9.996 ppbv

RT: 10.370 min Scan# 3177

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10




Tgt Ion: 95 Resp: 167061


Ion Ratio Lower Upper

95 100

174 62.1 50.2 75.2

175 4.6 4.1 6.1









#69
p-Isopropyltoluene
Concen: 1.054 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion:119 Resp: 69993
Ion Ratio Lower Upper
119 100
134 25.0 20.2 30.2
91 28.9 23.8 35.8



#70
n-Butylbenzene
Concen: 1.076 ppbv
RT: 12.277 min Scan# 3766
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion: 91 Resp: 69323
Ion Ratio Lower Upper
91 100
92 55.0 43.5 65.3
134 23.8 19.0 28.6









#71
2-Chlorotoluene
Concen: 1.065 ppbv
RT: 10.895 min Scan# 31
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 51452
Ion Ratio Lower Upper
91 100
126 29.3 11.7 46.7



#72
4-Ethyltoluene
Concen: 1.037 ppbv
RT: 11.118 min Scan# 3408
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:105 Resp: 44403
Ion Ratio Lower Upper
105 100
120 29.8 23.8 35.8
91 13.2 10.3 15.5
77 14.6 11.1 16.7





#73

1,3,5-Trimethylbenzene

Concen: 1.068 ppbv

RT: 11.199 min Scan# 3432

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC001

Tgt Ion:105 Resp: 38350

Ion Ratio Lower Upper

105 100

120 42.3 38.6 57.8



#74

1,2,4-Trimethylbenzene

Concen: 1.073 ppbv

RT: 11.532 min Scan# 3536

Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10

Tgt Ion:105 Resp: 42051




Ion Ratio Lower Upper


105 100

120 46.8 41.7 62.5

91 61.4 70.0 105.0#

77 21.7 23.8 35.8#









#75
1,3-Dichlorobenzene
Concen: 1.050 ppbv
RT: 11.600 min Scan# 3557
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion:146 Resp: 24011
Ion Ratio Lower Upper
146 100
111 45.9 23.4 70.3
148 64.7 31.6 95.0



#76
1,4-Dichlorobenzene
Concen: 1.029 ppbv
RT: 11.665 min Scan# 3577
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:146 Resp: 23543
Ion Ratio Lower Upper
146 100
111 44.5 22.8 68.3
148 66.4 31.8 95.4









#77
1,2-Dichlorobenzene
Concen: 1.103 ppbv
RT: 11.940 min Scan# 3662
Delta R.T. -0.000 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC001


Tgt Ion:146	Resp:	24248
Ion Ratio	Lower	Upper
146	100	
111	46.3	24.3 73.0
148	62.3	31.6 94.7



#78
Hexachloro-1,3-Butadiene
Concen: 1.116 ppbv
RT: 13.879 min Scan# 4261
Delta R.T. 0.003 min
Lab File: VL043206.D
Acq: 18 Nov 2025 11:10

Tgt Ion:225	Resp:	20489
Ion Ratio	Lower	Upper
225	100	
223	63.3	50.2 75.4
227	62.5	51.1 76.7





#79

Naphthalene

Concen: 1.047 ppbv

RT: 13.507 min Scan# 4146

Delta R.T. -0.000 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10





Instrument :

MSVOA_L

ClientSampleId :

VSTDIC001

Tgt Ion:	128	Resp:	36462
Ion Ratio	Lower	Upper	
128	100		
127	14.2	9.9	14.9
129	12.2	8.4	12.6



#80

Naphthalene, 2-methyl-

Concen: 0.807 ppbv




RT: 14.455 min Scan# 4439


Delta R.T. 0.003 min

Lab File: VL043206.D

Acq: 18 Nov 2025 11:10


Tgt Ion:	142	Resp:	7375
Ion Ratio	Lower	Upper	
142	100		
141	77.8	67.1	100.7
115	40.8	31.9	47.9







#81
 1,2,4-Trichlorobenzene
 Concen: 1.014 ppbv
 RT: 13.436 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VL043206.D
 Acq: 18 Nov 2025 11:10

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC001



Tgt Ion:	180	Resp:	15785
Ion Ratio	Lower	Upper	
180	100		
182	98.2	77.4	116.2
184	33.4	24.2	36.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043208.D
 Acq On : 18 Nov 2025 12:16
 Operator : SY/MD
 Sample : VSTDIC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Quant Time: Nov 19 00:51:52 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	101366	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	273996	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	222034	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.370	95	160968	9.890	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.900%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.554	85	1358	0.095	ppbv	99
3) Chlorodifluoromethane	1.476	51	1436	0.102	ppbv #	52
4) Chloromethane	1.534	50	638	0.119	ppbv #	82
5) Vinyl Chloride	1.580	62	664	0.114	ppbv	97
6) Bromomethane	1.667	94	352	0.124	ppbv	88
7) Chloroethane	1.706	64	275	0.121	ppbv	94
8) Dichlorotetrafluoroethane	1.554	85	1358	0.114	ppbv #	78
9) Propene	1.486	41	811	0.124	ppbv #	1
10) Heptane	4.972	43	480	0.027	ppbv #	1
11) Trichlorofluoromethane	1.877	101	1744	0.126	ppbv	87
12) 1,1,2-Trichlorotrifluo...	2.140	101	1249	0.115	ppbv	98
13) Ethanol	1.887	45	1141	0.543	ppbv	75
14) Bromoethene	1.780	108	456	0.110	ppbv #	71
15) Acetone	1.835	43	2574	0.183	ppbv	96
16) 1,3-Butadiene	1.609	39	1049	0.167	ppbv #	57
17) tert-Butyl alcohol	2.039	59	2464	0.133	ppbv #	30
18) 1,1-Dichloroethene	2.033	96	620	0.123	ppbv	81
19) Isopropyl Alcohol	1.887	45	1141	0.143	ppbv #	55
20) Methylene Chloride	2.062	84	3550	0.077	ppbv	91
21) Allyl Chloride	2.097	41	1553	0.164	ppbv #	45
22) trans-1,2-Dichloroethene	2.033	96	620	0.108	ppbv	96
23) Vinyl Acetate	2.460	43	2093	0.123	ppbv #	92
24) 1,1-Dichloroethane	2.408	63	1242	0.112	ppbv #	82
25) Ethyl Acetate	2.832	43	3043	0.100	ppbv #	93
26) Hexane	2.822	57	1613	0.121	ppbv	93
27) Carbon Disulfide	2.152	76	1422	0.105	ppbv #	1
28) Methyl tert-Butyl Ether	2.437	73	585	0.089	ppbv #	50
29) Chloroform	2.845	83	2028	0.107	ppbv	92
30) Cyclohexane	3.845	84	893	0.080	ppbv #	1
31) cis-1,2-Dichloroethene	2.719	61	1331	0.101	ppbv #	89
32) 1,1,1-Trichloroethane	3.369	97	999	0.049	ppbv #	11
34) 2-Butanone	2.560	43	2593	0.134	ppbv	96
35) Carbon Tetrachloride	3.761	117	901	0.043	ppbv #	88
36) Benzene	3.651	78	1569	0.060	ppbv #	79
37) 1,2-Dichloroethane	3.214	62	1537	0.100	ppbv	100
38) Trichloroethene	4.541	130	502	0.049	ppbv #	29
39) 1,2-Dichloropropane	3.949	63	69214	7.207	ppbv #	25
40) 1,4-Dioxane	4.596	88	308	0.071	ppbv #	16
41) Tetrahydrofuran	3.062	42	868	0.081	ppbv #	62
42) Bromodichloromethane	4.493	83	806	0.038	ppbv	95
43) Methyl Methacrylate	4.865	69	396	0.028	ppbv #	1
44) 2,2,4-Trimethylpentane	4.625	57	4980	0.110	ppbv #	81
45) t-1,3-Dichloropropene	6.405	75	645	0.050	ppbv	94


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043208.D
 Acq On : 18 Nov 2025 12:16
 Operator : SY/MD
 Sample : VSTDIC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Quant Time: Nov 19 00:51:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.583	75	359	0.024	ppbv #	72
47) 1,1,2-Trichloroethane	6.574	97	347	0.034	ppbv #	51
48) Dibromochloromethane	7.389	129	237	0.014	ppbv #	1
49) Bromoform	9.474	173	1357	0.087	ppbv #	84
50) 4-Methyl-2-Pentanone	5.732	43	149	0.005	ppbv #	39
51) 2-Hexanone	7.415	43	83	0.004	ppbv #	29
52) Tetrachloroethene	8.218	164	248	0.025	ppbv #	61
53) Toluene	6.927	91	734	0.023	ppbv #	21
54) 1,2-Dibromoethane	7.648	107	719	0.048	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.914	131	760	0.047	ppbv #	9
57) Chlorobenzene	8.917	112	1955	0.086	ppbv #	81
58) Ethyl Benzene	9.348	91	5157	0.121	ppbv #	86
59) m/p-Xylene	9.545	91	3099	0.090	ppbv #	31
60) o-Xylene	9.956	91	3453	0.101	ppbv	91
61) Styrene	9.859	104	1176	0.079	ppbv #	49
62) Isopropylbenzene	10.532	105	6313	0.103	ppbv	96
63) 1,1,2,2-Tetrachloroethane	9.953	83	1380	0.064	ppbv #	58
64) n-propylbenzene	10.989	120	1635	0.104	ppbv	87
65) tert-Butylbenzene	11.529	119	5745	0.105	ppbv	91
66) Benzyl Chloride	11.613	91	125	0.023	ppbv #	65
67) sec-Butylbenzene	11.762	105	7764	0.102	ppbv	94
69) p-Isopropyltoluene	11.921	119	6425	0.099	ppbv	99
70) n-Butylbenzene	12.280	91	5160	0.082	ppbv	93
71) 2-Chlorotoluene	10.898	91	5248	0.111	ppbv	93
72) 4-Ethyltoluene	11.124	105	4061	0.097	ppbv	96
73) 1,3,5-Trimethylbenzene	11.196	105	3445	0.098	ppbv	92
74) 1,2,4-Trimethylbenzene	11.536	105	1485	0.039	ppbv #	73
75) 1,3-Dichlorobenzene	11.604	146	1935	0.087	ppbv	98
76) 1,4-Dichlorobenzene	11.668	146	2131	0.096	ppbv	89
77) 1,2-Dichlorobenzene	11.943	146	2051	0.096	ppbv	96
78) Hexachloro-1,3-Butadiene	13.876	225	1837	0.103	ppbv #	89
79) Naphthalene	13.510	128	2112	0.062	ppbv #	87
80) Naphthalene,2-methyl-	14.507	142	309	0.035	ppbv #	65
81) 1,2,4-Trichlorobenzene	13.439	180	350	0.023	ppbv #	12





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 49 Resp: 101366
Ion Ratio Lower Upper
49 100
128 46.8 23.0 69.0
130 58.3 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.095 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.058 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 85 Resp: 1358
Ion Ratio Lower Upper
85 100
87 31.4 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.102 ppbv
RT: 1.476 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 51 Resp: 1436
Ion Ratio Lower Upper
51 100
67 40.2 14.9 22.3#



#4
Chloromethane
Concen: 0.119 ppbv
RT: 1.534 min Scan# 447
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 50 Resp: 638
Ion Ratio Lower Upper
50 100
52 39.8 24.1 36.1#









#5
 Vinyl Chloride
 Concen: 0.114 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.1


Tgt Ion: 62 Resp: 664
 Ion Ratio Lower Upper
 62 100
 64 32.2 27.3 40.9



#6
 Bromomethane
 Concen: 0.124 ppbv
 RT: 1.667 min Scan# 488
 Delta R.T. 0.003 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16

Tgt Ion: 94 Resp: 352
 Ion Ratio Lower Upper
 94 100
 96 102.8 72.9 109.3









#7
Chloroethane
Concen: 0.121 ppbv
RT: 1.706 min Scan# 50
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 64 Resp: 275
Ion Ratio Lower Upper
64 100
66 30.3 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 0.114 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 85 Resp: 1358
Ion Ratio Lower Upper
85 100
135 48.5 52.6 79.0#









#9
Propene
Concen: 0.124 ppbv
RT: 1.486 min Scan# 41
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 41 Resp: 811
Ion Ratio Lower Upper
41 100
42 79.2 3.7 5.5#



#10
Heptane
Concen: 0.027 ppbv
RT: 4.972 min Scan# 1509
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 43 Resp: 480
Ion Ratio Lower Upper
43 100
57 260.6 39.7 59.5#









#11
 Trichlorofluoromethane
 Concen: 0.126 ppbv
 RT: 1.877 min Scan# 511
 Delta R.T. 0.003 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.1


Tgt Ion:101 Resp: 1744
 Ion Ratio Lower Upper
 101 100
 103 55.2 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.115 ppbv
 RT: 2.140 min Scan# 634
 Delta R.T. 0.003 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16

Tgt Ion:101 Resp: 1249
 Ion Ratio Lower Upper
 101 100
 151 72.4 59.2 88.8









#13
Ethanol
Concen: 0.543 ppbv
RT: 1.887 min Scan# 51
Delta R.T. 0.171 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1


Tgt Ion: 45 Resp: 1141
Ion Ratio Lower Upper
45 100
46 22.5 15.0 59.8



#14
Bromoethene
Concen: 0.110 ppbv
RT: 1.780 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 108 Resp: 456
Ion Ratio Lower Upper
108 100
106 103.9 88.9 133.3
79 90.8 16.5 24.7#









#15
Acetone
Concen: 0.183 ppbv
RT: 1.835 min Scan# 54
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 43 Resp: 2574
Ion Ratio Lower Upper
43 100
58 28.0 20.7 31.1



#16
1,3-Butadiene
Concen: 0.167 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 39 Resp: 1049
Ion Ratio Lower Upper
39 100
54 111.2 59.4 89.2#





#17

tert-Butyl alcohol

Concen: 0.133 ppbv

RT: 2.039 min Scan# 602

Delta R.T. 0.003 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC0.1

Tgt Ion: 59 Resp: 2464

Ion Ratio Lower Upper

59 100

57 37.8 9.0 13.4#



#18

1,1-Dichloroethene

Concen: 0.123 ppbv

RT: 2.033 min Scan# 601

Delta R.T. 0.003 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16




Tgt Ion: 96 Resp: 620


Ion Ratio Lower Upper

96 100

61 178.8 170.6 255.8

98 58.9 52.2 78.4









#19
Isopropyl Alcohol
Concen: 0.143 ppbv
RT: 1.887 min Scan# 511
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 45 Resp: 1141
Ion Ratio Lower Upper
45 100
58 11.7 31.3 46.9#



#20
Methylene Chloride
Concen: 0.077 ppbv
RT: 2.062 min Scan# 610
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 84 Resp: 3550
Ion Ratio Lower Upper
84 100
49 156.9 132.9 199.3
51 38.9 34.8 52.2
86 71.3 48.2 72.2





#21

Allyl Chloride

Concen: 0.164 ppbv

RT: 2.097 min Scan# 61

Delta R.T. 0.006 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC0.1





Tgt Ion: 41 Resp: 1553

Ion Ratio Lower Upper

41 100

76 21.1 22.1 33.1#

39 136.3 61.0 91.4#



#22

trans-1,2-Dichloroethene

Concen: 0.108 ppbv

RT: 2.033 min Scan# 601

Delta R.T. -0.301 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16




Tgt Ion: 96 Resp: 620


Ion Ratio Lower Upper

96 100

61 178.8 139.0 208.6

98 58.9 49.4 74.2









#23
 Vinyl Acetate
 Concen: 0.123 ppbv
 RT: 2.460 min Scan# 717
 Delta R.T. 0.003 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.1


Tgt Ion: 43	Resp: 2093
Ion Ratio	Lower Upper
43 100	
86 8.1	6.2 9.2
42 13.4	8.0 12.0#
44 0.0	4.0 6.0#



#24
 1,1-Dichloroethane
 Concen: 0.112 ppbv
 RT: 2.408 min Scan# 717
 Delta R.T. 0.003 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16

Tgt Ion: 63	Resp: 1242
Ion Ratio	Lower Upper
63 100	
98 13.4	2.8 8.4#
100 0.0	1.3 3.9#









#25
Ethyl Acetate
Concen: 0.100 ppbv
RT: 2.832 min Scan# 84
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:	43	Resp:	3043
Ion	Ratio	Lower	Upper
43	100		
45	10.9	8.1	12.1
61	8.2	7.0	10.4
70	15.1	7.0	10.6



#26
Hexane
Concen: 0.121 ppbv
RT: 2.822 min Scan# 845
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion:	57	Resp:	1613
Ion	Ratio	Lower	Upper
57	100		
41	88.3	74.6	111.8
43	235.4	179.0	268.4
56	61.2	42.3	63.5









#27
Carbon Disulfide
Concen: 0.105 ppbv
RT: 2.152 min Scan# 61
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 76 Resp: 1422
Ion Ratio Lower Upper
76 100
44 325.6 16.1 24.1#
78 45.7 9.6 14.4#



#28
Methyl tert-Butyl Ether
Concen: 0.089 ppbv
RT: 2.437 min Scan# 726
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 73 Resp: 585
Ion Ratio Lower Upper
73 100
57 0.0 19.8 29.8#





#29

Chloroform

Concen: 0.107 ppbv

RT: 2.845 min Scan# 851

Delta R.T. 0.003 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC0.1

Tgt Ion: 83 Resp: 2028

Ion Ratio Lower Upper

83 100

85 69.1 50.5 75.7



#30

Cyclohexane

Concen: 0.080 ppbv

RT: 3.845 min Scan# 1161

Delta R.T. -0.000 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16




Tgt Ion: 84 Resp: 893


Ion Ratio Lower Upper

84 100

56 0.0 104.7 157.1#

41 0.0 69.0 103.6#









#31
 cis-1,2-Dichloroethene
 Concen: 0.101 ppbv
 RT: 2.719 min Scan# 812
 Delta R.T. 0.003 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.1


Tgt Ion: 61 Resp: 1331
 Ion Ratio Lower Upper
 61 100
 96 55.8 49.3 73.9
 98 50.0 32.1 48.1#



#32
 1,1,1-Trichloroethane
 Concen: 0.049 ppbv
 RT: 3.369 min Scan# 1014
 Delta R.T. 0.009 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16

Tgt Ion: 97 Resp: 999
 Ion Ratio Lower Upper
 97 100
 99 130.1 50.3 75.5#
 61 111.0 38.9 58.3#









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:114 Resp: 273996
Ion Ratio Lower Upper
114 100
63 25.3 21.0 31.6
88 18.3 14.9 22.3



#34
2-Butanone
Concen: 0.134 ppbv
RT: 2.560 min Scan# 764
Delta R.T. 0.009 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 43 Resp: 2593
Ion Ratio Lower Upper
43 100
72 24.3 18.1 27.1









#35
Carbon Tetrachloride
Concen: 0.043 ppbv
RT: 3.761 min Scan# 1135
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:117 Resp: 901
Ion Ratio Lower Upper
117 100
119 85.0 74.6 112.0
121 16.7 23.2 34.8#



#36
Benzene
Concen: 0.060 ppbv
RT: 3.651 min Scan# 1101
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 78 Resp: 1569
Ion Ratio Lower Upper
78 100
77 17.4 17.5 26.3#
50 35.3 15.5 23.3#





#37

1,2-Dichloroethane

Concen: 0.100 ppbv

RT: 3.214 min Scan# 9

Delta R.T. 0.003 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC0.1

Tgt Ion: 62 Resp: 1537

Ion Ratio Lower Upper

62 100

98 6.9 5.5 8.3



#38

Trichloroethene

Concen: 0.049 ppbv

RT: 4.541 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Tgt Ion: 130 Resp: 502




Ion Ratio Lower Upper


130 100

95 61.3 85.7 128.5#

132 0.0 73.4 110.0#

97 3.3 52.7 79.1#





#39

1,2-Dichloropropane

Concen: 7.207 ppbv

RT: 3.949 min Scan# 1193

Delta R.T. -0.353 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC0.1





Tgt Ion: 63 Resp: 69214

Ion Ratio Lower Upper

63 100

41 0.0 62.6 93.8#

62 18.5 51.0 76.6#



#40

1,4-Dioxane

Concen: 0.071 ppbv

RT: 4.596 min Scan# 1393

Delta R.T. 0.029 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Tgt Ion: 88 Resp: 308




Ion Ratio Lower Upper


88 100

58 70.8 63.0 94.6

43 182.8 32.5 48.7#

57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 0.081 ppbv
RT: 3.062 min Scan# 919
Delta R.T. 0.016 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 42 Resp: 868
Ion Ratio Lower Upper
42 100
72 46.2 30.6 46.0#
71 0.0 30.5 45.7#



#42
Bromodichloromethane
Concen: 0.038 ppbv
RT: 4.493 min Scan# 1361
Delta R.T. 0.016 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 83 Resp: 806
Ion Ratio Lower Upper
83 100
85 61.0 51.7 77.5
127 9.3 6.3 9.5









#43
Methyl Methacrylate
Concen: 0.028 ppbv
RT: 4.865 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.1


Tgt Ion: 69 Resp: 396
Ion Ratio Lower Upper
69 100
41 0.0 126.7 190.1#
100 0.0 28.6 43.0#



#44
2,2,4-Trimethylpentane
Concen: 0.110 ppbv
RT: 4.625 min Scan# 1402
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 57 Resp: 4980
Ion Ratio Lower Upper
57 100
41 54.9 28.0 42.0#
56 32.9 25.6 38.4





#45

t-1,3-Dichloropropene

Concen: 0.050 ppbv

RT: 6.405 min Scan# 1950

Delta R.T. 0.006 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC0.1

Tgt Ion: 75 Resp: 645

Ion Ratio Lower Upper

75 100

77 33.7 24.3 36.5



#46

cis-1,3-Dichloropropene

Concen: 0.024 ppbv

RT: 5.583 min Scan# 1698

Delta R.T. -0.000 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16




Tgt Ion: 75 Resp: 359


Ion Ratio Lower Upper

75 100

39 48.3 59.1 88.7#

77 17.3 24.2 36.2#





#47

1,1,2-Trichloroethane

Concen: 0.034 ppbv

RT: 6.574 min Scan# 2004

Delta R.T. 0.009 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC0.1

Tgt Ion: 97 Resp: 347





Ion Ratio Lower Upper

97 100

83 37.9 64.6 97.0#

85 39.4 39.4 59.2#

99 0.0 47.0 70.6#



#48

Dibromochloromethane

Concen: 0.014 ppbv

RT: 7.389 min Scan# 2256

Delta R.T. 0.013 min

Lab File: VL043208.D




Acq: 18 Nov 2025 12:16


Tgt Ion: 129 Resp: 237

Ion Ratio Lower Upper

129 100

127 547.7 39.3 117.9#









#49
Bromoform
Concen: 0.087 ppbv
RT: 9.474 min Scan# 2900
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:173 Resp: 1357
Ion Ratio Lower Upper
173 100
175 55.2 38.6 58.0
171 65.3 40.6 61.0#



#50
4-Methyl-2-Pentanone
Concen: 0.005 ppbv
RT: 5.732 min Scan# 1744
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 43 Resp: 149
Ion Ratio Lower Upper
43 100
58 0.0 28.4 42.6#









#51
2-Hexanone
Concen: 0.004 ppbv
RT: 7.415 min Scan# 21
Delta R.T. -0.004 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 43 Resp: 83
Ion Ratio Lower Upper
43 100
58 0.0 38.4 57.6#
98 0.0 0.0 0.0



#52
Tetrachloroethene
Concen: 0.025 ppbv
RT: 8.218 min Scan# 2512
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 164 Resp: 248
Ion Ratio Lower Upper
164 100
166 87.8 99.4 149.2#
129 90.9 81.8 122.8
131 19.7 77.4 116.2#









#53
Toluene
Concen: 0.023 ppbv
RT: 6.927 min Scan# 2113
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 91 Resp: 734
Ion Ratio Lower Upper
91 100
92 0.0 47.4 71.0#



#54
1,2-Dibromoethane
Concen: 0.048 ppbv
RT: 7.648 min Scan# 2336
Delta R.T. 0.009 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 107 Resp: 719
Ion Ratio Lower Upper
107 100
109 85.5 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.875 min Scan# 211
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:117 Resp: 222034
Ion Ratio Lower Upper
117 100
82 66.4 56.1 84.1
119 32.2 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 0.047 ppbv
RT: 8.914 min Scan# 2727
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion:131 Resp: 760
Ion Ratio Lower Upper
131 100
133 0.0 77.4 116.2#
119 0.0 49.1 73.7#





#57

Chlorobenzene

Concen: 0.086 ppbv

RT: 8.917 min Scan# 21

Delta R.T. 0.006 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :

VSTDICC0.1





Tgt Ion:112 Resp: 1955

Ion Ratio Lower Upper

112 100

77 48.2 55.5 83.3#

114 30.3 29.1 35.5



#58

Ethyl Benzene

Concen: 0.121 ppbv

RT: 9.348 min Scan# 2861

Delta R.T. 0.003 min

Lab File: VL043208.D




Acq: 18 Nov 2025 12:16


Tgt Ion: 91 Resp: 5157

Ion Ratio Lower Upper

91 100

106 35.7 22.8 34.2#





#59

m/p-Xylene

Concen: 0.090 ppbv

RT: 9.545 min Scan# 2921

Delta R.T. 0.003 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC0.1

Tgt Ion: 91 Resp: 3099

Ion Ratio Lower Upper

91 100

106 0.0 36.1 54.1#



#60

o-Xylene

Concen: 0.101 ppbv

RT: 9.956 min Scan# 3049

Delta R.T. -0.000 min

Lab File: VL043208.D




Acq: 18 Nov 2025 12:16


Tgt Ion: 91 Resp: 3453

Ion Ratio Lower Upper

91 100

106 47.5 20.9 62.8









#61
Styrene
Concen: 0.079 ppbv
RT: 9.859 min Scan# 3019
Delta R.T. -0.004 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:104 Resp: 1176
Ion Ratio Lower Upper
104 100
78 77.0 42.1 63.1#
103 0.0 37.2 55.8#



#62
Isopropylbenzene
Concen: 0.103 ppbv
RT: 10.532 min Scan# 3227
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion:105 Resp: 6313
Ion Ratio Lower Upper
105 100
120 26.9 19.8 29.8









#63
 1,1,2,2-Tetrachloroethane
 Concen: 0.064 ppbv
 RT: 9.953 min Scan# 3048
 Delta R.T. -0.000 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.1


Tgt Ion: 83 Resp: 1380
 Ion Ratio Lower Upper
 83 100
 131 17.4 4.8 14.3#
 85 99.1 31.9 95.9#



#64
 n-propylbenzene
 Concen: 0.104 ppbv
 RT: 10.989 min Scan# 3368
 Delta R.T. 0.006 min
 Lab File: VL043208.D
 Acq: 18 Nov 2025 12:16

Tgt Ion: 120 Resp: 1635
 Ion Ratio Lower Upper
 120 100
 91 507.0 377.1 565.7





#65

tert-Butylbenzene

Concen: 0.105 ppbv

RT: 11.529 min Scan# 3533

Delta R.T. 0.006 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC0.1





Tgt Ion:119 Resp: 5745

Ion Ratio Lower Upper

119 100

91 90.4 65.3 97.9

134 23.6 17.0 25.4



#66

Benzyl Chloride

Concen: 0.023 ppbv

RT: 11.613 min Scan# 3561

Delta R.T. 0.006 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16




Tgt Ion: 91 Resp: 125


Ion Ratio Lower Upper

91 100

126 0.0 14.4 21.6#

128 0.0 4.6 6.8#









#67
sec-Butylbenzene
Concen: 0.102 ppbv
RT: 11.762 min Scan# 3176
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:105 Resp: 7764
Ion Ratio Lower Upper
105 100
134 21.5 15.0 22.6



#68
1-Bromo-4-Fluorobenzene
Concen: 9.890 ppbv
RT: 10.370 min Scan# 3177
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 95 Resp: 160968
Ion Ratio Lower Upper
95 100
174 62.5 50.2 75.2
175 4.8 4.1 6.1









#69
p-Isopropyltoluene
Concen: 0.099 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:	119	Resp:	6425
Ion	Ratio	Lower	Upper
119	100		
134	24.9	20.2	30.2
91	30.8	23.8	35.8



#70
n-Butylbenzene
Concen: 0.082 ppbv
RT: 12.280 min Scan# 3767
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion:	91	Resp:	5160
Ion	Ratio	Lower	Upper
91	100		
92	60.4	43.5	65.3
134	22.5	19.0	28.6









#71
2-Chlorotoluene
Concen: 0.111 ppbv
RT: 10.898 min Scan# 31
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion: 91 Resp: 5248
Ion Ratio Lower Upper
91 100
126 32.9 11.7 46.7



#72
4-Ethyltoluene
Concen: 0.097 ppbv
RT: 11.124 min Scan# 3410
Delta R.T. 0.006 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion: 105 Resp: 4061
Ion Ratio Lower Upper
105 100
120 27.0 23.8 35.8
91 14.4 10.3 15.5
77 15.3 11.1 16.7





#73

1,3,5-Trimethylbenzene

Concen: 0.098 ppbv

RT: 11.196 min Scan# 3432

Delta R.T. -0.000 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC0.1

Tgt Ion:105 Resp: 3445

Ion Ratio Lower Upper

105 100

120 42.6 38.6 57.8



#74

1,2,4-Trimethylbenzene

Concen: 0.039 ppbv

RT: 11.536 min Scan# 3537

Delta R.T. 0.006 min

Lab File: VL043208.D

Acq: 18 Nov 2025 12:16

Tgt Ion:105 Resp: 1485




Ion Ratio Lower Upper


105 100

120 42.9 41.7 62.5

91 47.0 70.0 105.0#

77 30.7 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 0.087 ppbv
RT: 11.604 min Scan# 3557
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:	146	Resp:	1935
Ion	Ratio	Lower	Upper
146	100		
111	49.3	23.4	70.3
148	63.5	31.6	95.0



#76
1,4-Dichlorobenzene
Concen: 0.096 ppbv
RT: 11.668 min Scan# 3578
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion:	146	Resp:	2131
Ion	Ratio	Lower	Upper
146	100		
111	42.9	22.8	68.3
148	51.4	31.8	95.4









#77
1,2-Dichlorobenzene
Concen: 0.096 ppbv
RT: 11.943 min Scan# 3663
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:146	Resp:	2051
Ion Ratio	Lower	Upper
146	100	
111	50.8	24.3 73.0
148	66.4	31.6 94.7



#78
Hexachloro-1,3-Butadiene
Concen: 0.103 ppbv
RT: 13.876 min Scan# 4260
Delta R.T. -0.000 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion:225	Resp:	1837
Ion Ratio	Lower	Upper
225	100	
223	79.9	50.2 75.4#
227	64.2	51.1 76.7









#79
Naphthalene
Concen: 0.062 ppbv
RT: 13.510 min Scan# 4146
Delta R.T. 0.003 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.1


Tgt Ion:128 Resp: 2112
Ion Ratio Lower Upper
128 100
127 9.7 9.9 14.9#
129 18.1 8.4 12.6#



#80
Naphthalene,2-methyl-
Concen: 0.035 ppbv
RT: 14.507 min Scan# 4455
Delta R.T. 0.055 min
Lab File: VL043208.D
Acq: 18 Nov 2025 12:16

Tgt Ion:142 Resp: 309
Ion Ratio Lower Upper
142 100
141 38.8 67.1 100.7#
115 36.9 31.9 47.9





#81

1,2,4-Trichlorobenzene

Concen: 0.023 ppbv

RT: 13.439 min Scan# 41

Delta R.T. 0.003 min

Lab File: VL043208.D


Acq: 18 Nov 2025 12:16

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.1





Tgt Ion:180 Resp: 350

Ion Ratio Lower Upper

180 100

182 0.0 77.4 116.2#

184 0.0 24.2 36.4#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043209.D
 Acq On : 18 Nov 2025 12:49
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Quant Time: Nov 19 00:52:39 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	98952	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	266012	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	213547	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.371	95	152015	9.712	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.554	85	381	0.027	ppbv	100
3) Chlorodifluoromethane	1.638	51	673	0.049	ppbv	95
4) Chloromethane	1.535	50	260	0.050	ppbv #	71
5) Vinyl Chloride	1.580	62	150	0.026	ppbv #	86
6) Bromomethane	1.667	94	148	0.054	ppbv #	23
7) Chloroethane	1.706	64	109	0.049	ppbv #	76
8) Dichlorotetrafluoroethane	1.554	85	381	0.033	ppbv	99
9) Propene	1.489	41	440	0.069	ppbv #	1
10) Heptane	4.969	43	193	0.011	ppbv #	1
11) Trichlorofluoromethane	1.878	101	277	0.020	ppbv	86
12) 1,1,2-Trichlorotrifluo...	2.137	101	368	0.035	ppbv	96
13) Ethanol	1.719	45	229	0.112	ppbv #	38
14) Bromoethene	1.777	108	102	0.025	ppbv #	55
15) Acetone	1.839	43	1490	0.108	ppbv #	86
16) 1,3-Butadiene	1.606	39	392	0.064	ppbv #	12
17) tert-Butyl alcohol	2.040	59	506	0.028	ppbv #	71
18) 1,1-Dichloroethene	2.033	96	130	0.026	ppbv #	90
19) Isopropyl Alcohol	1.887	45	746	0.096	ppbv #	49
20) Methylene Chloride	2.059	84	3122	Below Cal	#	93
21) Allyl Chloride	2.095	41	774	0.084	ppbv #	23
22) trans-1,2-Dichloroethene	2.337	96	240	0.043	ppbv #	56
23) Vinyl Acetate	2.457	43	627	0.038	ppbv #	78
24) 1,1-Dichloroethane	2.402	63	453	0.042	ppbv #	86
25) Ethyl Acetate	2.836	43	455	0.015	ppbv #	82
26) Hexane	2.820	57	499	0.038	ppbv #	1
27) Carbon Disulfide	2.153	76	477	0.036	ppbv #	1
28) Methyl tert-Butyl Ether	2.438	73	181	0.028	ppbv #	50
29) Chloroform	2.399	83	140	0.008	ppbv	95
30) Cyclohexane	3.846	84	179	0.016	ppbv #	1
31) cis-1,2-Dichloroethene	2.719	61	369	0.029	ppbv #	90
32) 1,1,1-Trichloroethane	3.363	97	153	0.008	ppbv #	1
34) 2-Butanone	2.545	43	88	0.005	ppbv #	53
35) Carbon Tetrachloride	3.758	117	558	0.028	ppbv #	61
36) Benzene	3.651	78	323	0.013	ppbv #	11
37) 1,2-Dichloroethane	3.211	62	298	0.020	ppbv #	80
38) Trichloroethene	4.542	130	123	0.012	ppbv #	49
39) 1,2-Dichloropropane	3.949	63	68235	7.318	ppbv #	24
40) 1,4-Dioxane	4.606	88	75	0.018	ppbv #	1
41) Tetrahydrofuran	3.053	42	118	0.011	ppbv #	37
42) Bromodichloromethane	4.483	83	268	0.013	ppbv #	24
43) Methyl Methacrylate	4.868	69	145	0.011	ppbv #	1
44) 2,2,4-Trimethylpentane	4.629	57	595	0.014	ppbv #	41
45) t-1,3-Dichloropropene	6.403	75	232	0.018	ppbv #	44

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043209.D
 Acq On : 18 Nov 2025 12:49
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03


Quant Time: Nov 19 00:52:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.587	75	123	0.008	ppbv #	37
47) 1,1,2-Trichloroethane	6.574	97	122	0.012	ppbv #	51
48) Dibromochloromethane	7.743	129	34	0.002	ppbv	62
49) Bromoform	9.471	173	102	0.007	ppbv #	1
50) 4-Methyl-2-Pentanone	5.736	43	140	0.005	ppbv #	39
51) 2-Hexanone	7.419	43	61	0.003	ppbv #	29
52) Tetrachloroethene	8.222	164	158	0.017	ppbv #	48
53) Toluene	6.930	91	323	0.011	ppbv #	21
54) 1,2-Dibromoethane	7.645	107	102	0.007	ppbv	87
56) 1,1,1,2-Tetrachloroethane	8.927	131	114	0.007	ppbv #	1
57) Chlorobenzene	8.914	112	429	0.020	ppbv #	55
58) Ethyl Benzene	9.519	91	613	0.015	ppbv	91
59) m/p-Xylene	9.545	91	299	0.009	ppbv #	1
60) o-Xylene	9.956	91	1078	0.033	ppbv	99
61) Styrene	9.869	104	300	0.021	ppbv #	54
62) Isopropylbenzene	10.533	105	1990	0.034	ppbv	90
63) 1,1,2,2-Tetrachloroethane	9.963	83	306	0.015	ppbv #	47
64) n-propylbenzene	10.982	120	316	0.021	ppbv	82
65) tert-Butylbenzene	11.526	119	1740	0.033	ppbv	98
66) Benzyl Chloride	11.607	91	62	0.012	ppbv #	65
67) sec-Butylbenzene	11.762	105	2398	0.033	ppbv #	70
69) p-Isopropyltoluene	11.924	119	1875	0.030	ppbv #	91
70) n-Butylbenzene	12.277	91	1502	0.025	ppbv	94
71) 2-Chlorotoluene	10.895	91	1463	0.032	ppbv	98
72) 4-Ethyltoluene	11.125	105	408	0.010	ppbv #	61
73) 1,3,5-Trimethylbenzene	11.203	105	1041	0.031	ppbv #	68
74) 1,2,4-Trimethylbenzene	11.536	105	789	0.021	ppbv #	49
75) 1,3-Dichlorobenzene	11.610	146	214	0.010	ppbv #	1
76) 1,4-Dichlorobenzene	11.669	146	292	0.014	ppbv #	1
77) 1,2-Dichlorobenzene	11.947	146	561	0.027	ppbv	81
78) Hexachloro-1,3-Butadiene	13.876	225	622	0.036	ppbv #	83
79) Naphthalene	13.520	128	504	0.015	ppbv #	13
80) Naphthalene,2-methyl-	14.475	142	58	0.007	ppbv #	36
81) 1,2,4-Trichlorobenzene	13.442	180	150	0.010	ppbv #	46

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Nov 19 00:52:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 00:48:15 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03


Tgt Ion: 49 Resp: 98952
Ion Ratio Lower Upper
49 100
128 45.9 23.0 69.0
130 58.8 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.027 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.058 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 85 Resp: 381
Ion Ratio Lower Upper
85 100
87 31.8 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.049 ppbv
RT: 1.638 min Scan# 41
Delta R.T. 0.165 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.03


Tgt Ion: 51 Resp: 673
Ion Ratio Lower Upper
51 100
67 20.8 14.9 22.3



#4
Chloromethane
Concen: 0.050 ppbv
RT: 1.535 min Scan# 447
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 50 Resp: 260
Ion Ratio Lower Upper
50 100
52 46.0 24.1 36.1#









#5
 Vinyl Chloride
 Concen: 0.026 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.03


Tgt Ion: 62 Resp: 150
 Ion Ratio Lower Upper
 62 100
 64 26.3 27.3 40.9#



#6
 Bromomethane
 Concen: 0.054 ppbv
 RT: 1.667 min Scan# 488
 Delta R.T. 0.003 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49

Tgt Ion: 94 Resp: 148
 Ion Ratio Lower Upper
 94 100
 96 18.4 72.9 109.3#









#7
Chloroethane
Concen: 0.049 ppbv
RT: 1.706 min Scan# 50
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.03


Tgt Ion: 64 Resp: 109
Ion Ratio Lower Upper
64 100
66 19.6 26.8 40.2#



#8
Dichlorotetrafluoroethane
Concen: 0.033 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 85 Resp: 381
Ion Ratio Lower Upper
85 100
135 64.8 52.6 79.0









#9
Propene
Concen: 0.069 ppbv
RT: 1.489 min Scan# 41
Delta R.T. 0.010 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03


Tgt Ion: 41 Resp: 440
Ion Ratio Lower Upper
41 100
42 147.5 3.7 5.5#



#10
Heptane
Concen: 0.011 ppbv
RT: 4.969 min Scan# 1508
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 43 Resp: 193
Ion Ratio Lower Upper
43 100
57 146.1 39.7 59.5#









#11
Trichlorofluoromethane
Concen: 0.020 ppbv
RT: 1.878 min Scan# 511
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03


Tgt Ion:101 Resp: 277
Ion Ratio Lower Upper
101 100
103 76.6 52.6 79.0



#12
1,1,2-Trichlorotrifluoroethane
Concen: 0.035 ppbv
RT: 2.137 min Scan# 633
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion:101 Resp: 368
Ion Ratio Lower Upper
101 100
151 77.4 59.2 88.8









#13
Ethanol
Concen: 0.112 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.03


Tgt Ion: 45 Resp: 229
Ion Ratio Lower Upper
45 100
46 0.0 15.0 59.8#



#14
Bromoethene
Concen: 0.025 ppbv
RT: 1.777 min Scan# 522
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 108 Resp: 102
Ion Ratio Lower Upper
108 100
106 158.8 88.9 133.3#
79 0.0 16.5 24.7#









#15
Acetone
Concen: 0.108 ppbv
RT: 1.839 min Scan# 541
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03


Tgt Ion: 43 Resp: 1490
Ion Ratio Lower Upper
43 100
58 32.8 20.7 31.1#



#16
1,3-Butadiene
Concen: 0.064 ppbv
RT: 1.606 min Scan# 469
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 39 Resp: 392
Ion Ratio Lower Upper
39 100
54 0.0 59.4 89.2#









#17
tert-Butyl alcohol
Concen: 0.028 ppbv
RT: 2.040 min Scan# 602
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion: 59 Resp: 506
Ion Ratio Lower Upper
59 100
57 0.0 9.0 13.4#



#18
1,1-Dichloroethene
Concen: 0.026 ppbv
RT: 2.033 min Scan# 601
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 96 Resp: 130
Ion Ratio Lower Upper
96 100
61 219.8 170.6 255.8
98 88.5 52.2 78.4#









#19
Isopropyl Alcohol
Concen: 0.096 ppbv
RT: 1.887 min Scan# 51
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion: 45 Resp: 746
Ion Ratio Lower Upper
45 100
58 7.6 31.3 46.9#



#20
Methylene Chloride
Concen: Below Cal
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 84 Resp: 3122
Ion Ratio Lower Upper
84 100
49 161.1 132.9 199.3
51 44.7 34.8 52.2
86 46.3 48.2 72.2#









#21
 Allyl Chloride
 Concen: 0.084 ppbv
 RT: 2.095 min Scan# 619
 Delta R.T. 0.003 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.03


Tgt Ion: 41 Resp: 774
 Ion Ratio Lower Upper
 41 100
 76 19.1 22.1 33.1#
 39 161.5 61.0 91.4#



#22
 trans-1,2-Dichloroethene
 Concen: 0.043 ppbv
 RT: 2.337 min Scan# 695
 Delta R.T. 0.003 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49

Tgt Ion: 96 Resp: 240
 Ion Ratio Lower Upper
 96 100
 61 93.9 139.0 208.6#
 98 66.0 49.4 74.2









#23
 Vinyl Acetate
 Concen: 0.038 ppbv
 RT: 2.457 min Scan# 715
 Delta R.T. 0.000 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDICC0.03


Tgt Ion:	43	Resp:	627
Ion	Ratio	Lower	Upper
43	100		
86	0.0	6.2	9.2#
42	19.3	8.0	12.0#
44	0.0	4.0	6.0#



#24
 1,1-Dichloroethane
 Concen: 0.042 ppbv
 RT: 2.402 min Scan# 715
 Delta R.T. -0.003 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49

Tgt Ion:	63	Resp:	453
Ion	Ratio	Lower	Upper
63	100		
98	0.0	2.8	8.4#
100	0.0	1.3	3.9#









#25
Ethyl Acetate
Concen: 0.015 ppbv
RT: 2.836 min Scan# 847
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.03

Tgt Ion: 43 Resp: 455




Ion	Ratio	Lower	Upper
43	100		
45	0.0	8.1	12.1#
61	8.6	7.0	10.4
70	0.0	7.0	10.6#




#26
Hexane
Concen: 0.038 ppbv
RT: 2.820 min Scan# 844
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 57 Resp: 499

Ion	Ratio	Lower	Upper
57	100		
41	0.0	74.6	111.8#
43	0.0	179.0	268.4#
56	73.3	42.3	63.5#









#27
Carbon Disulfide
Concen: 0.036 ppbv
RT: 2.153 min Scan# 611
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC0.03


Tgt Ion: 76 Resp: 477
Ion Ratio Lower Upper
76 100
44 1027.0 16.1 24.1#
78 43.6 9.6 14.4#



#28
Methyl tert-Butyl Ether
Concen: 0.028 ppbv
RT: 2.438 min Scan# 726
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 73 Resp: 181
Ion Ratio Lower Upper
73 100
57 0.0 19.8 29.8#





#29

Chloroform

Concen: 0.008 ppbv

RT: 2.399 min Scan# 714

Delta R.T. -0.443 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC0.03

Tgt Ion: 83 Resp: 140

Ion Ratio Lower Upper

83 100

85 66.7 50.5 75.7



#30

Cyclohexane

Concen: 0.016 ppbv

RT: 3.846 min Scan# 1161

Delta R.T. 0.000 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49




Tgt Ion: 84 Resp: 179


Ion Ratio Lower Upper

84 100

56 369.8 104.7 157.1#

41 0.0 69.0 103.6#









#31
 cis-1,2-Dichloroethene
 Concen: 0.029 ppbv
 RT: 2.719 min Scan# 81
 Delta R.T. 0.003 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.03


Tgt Ion:	61	Resp:	369
Ion	Ratio	Lower	Upper
61	100		
96	58.5	49.3	73.9
98	28.9	32.1	48.1#



#32
 1,1,1-Trichloroethane
 Concen: 0.008 ppbv
 RT: 3.363 min Scan# 1012
 Delta R.T. 0.003 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49

Tgt Ion:	97	Resp:	153
Ion	Ratio	Lower	Upper
97	100		
99	233.3	50.3	75.5#
61	0.0	38.9	58.3#









#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 3.952 min Scan# 1194
 Delta R.T. 0.003 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.03


Tgt Ion:114 Resp: 266012
 Ion Ratio Lower Upper
 114 100
 63 25.7 21.0 31.6
 88 18.2 14.9 22.3



#34
 2-Butanone
 Concen: 0.005 ppbv
 RT: 2.545 min Scan# 759
 Delta R.T. -0.006 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49

Tgt Ion: 43 Resp: 88
 Ion Ratio Lower Upper
 43 100
 72 0.0 18.1 27.1#





#35

Carbon Tetrachloride

Concen: 0.028 ppbv

RT: 3.758 min Scan# 1133

Delta R.T. 0.003 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.03





Tgt Ion:117 Resp: 558

Ion Ratio Lower Upper

117 100

119 46.9 74.6 112.0#

121 33.5 23.2 34.8



#36

Benzene

Concen: 0.013 ppbv

RT: 3.651 min Scan# 1101

Delta R.T. 0.003 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49




Tgt Ion: 78 Resp: 323


Ion Ratio Lower Upper

78 100

77 0.0 17.5 26.3#

50 81.8 15.5 23.3#





#37

1,2-Dichloroethane

Concen: 0.020 ppbv

RT: 3.211 min Scan# 9

Delta R.T. 0.000 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC0.03

Tgt Ion: 62 Resp: 298

Ion Ratio Lower Upper

62 100

98 0.0 5.5 8.3#



#38

Trichloroethene

Concen: 0.012 ppbv

RT: 4.542 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Tgt Ion: 130 Resp: 123




Ion Ratio Lower Upper


130 100

95 0.0 85.7 128.5#

132 97.5 73.4 110.0

97 44.0 52.7 79.1#





#39

1,2-Dichloropropane

Concen: 7.318 ppbv

RT: 3.949 min Scan# 1193

Delta R.T. -0.353 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC0.03





Tgt Ion: 63 Resp: 68235

Ion Ratio Lower Upper

63 100

41 0.4 62.6 93.8#

62 17.0 51.0 76.6#



#40

1,4-Dioxane

Concen: 0.018 ppbv

RT: 4.606 min Scan# 1396

Delta R.T. 0.039 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Tgt Ion: 88 Resp: 75




Ion Ratio Lower Upper


88 100

58 142.7 63.0 94.6#

43 806.7 32.5 48.7#

57 2290.7 0.0 0.0#









#41
Tetrahydrofuran
Concen: 0.011 ppbv
RT: 3.053 min Scan# 914
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion:	42	Resp:	118
Ion	Ratio	Lower	Upper
42	100		
72	0.0	30.6	46.0#
71	0.0	30.5	45.7#



#42
Bromodichloromethane
Concen: 0.013 ppbv
RT: 4.483 min Scan# 1358
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion:	83	Resp:	268
Ion	Ratio	Lower	Upper
83	100		
85	0.0	51.7	77.5#
127	0.0	6.3	9.5#









#43
Methyl Methacrylate
Concen: 0.011 ppbv
RT: 4.868 min Scan# 1476
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03


Tgt Ion: 69 Resp: 145
Ion Ratio Lower Upper
69 100
41 0.0 126.7 190.1#
100 0.0 28.6 43.0#



#44
2,2,4-Trimethylpentane
Concen: 0.014 ppbv
RT: 4.629 min Scan# 1403
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 57 Resp: 595
Ion Ratio Lower Upper
57 100
41 0.0 28.0 42.0#
56 0.0 25.6 38.4#









#45
t-1,3-Dichloropropene
Concen: 0.018 ppbv
RT: 6.403 min Scan# 1951
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.03


Tgt Ion: 75 Resp: 232
Ion Ratio Lower Upper
75 100
77 0.0 24.3 36.5#



#46
cis-1,3-Dichloropropene
Concen: 0.008 ppbv
RT: 5.587 min Scan# 1699
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 75 Resp: 123
Ion Ratio Lower Upper
75 100
39 17.9 59.1 88.7#
77 0.0 24.2 36.2#









#47
1,1,2-Trichloroethane
Concen: 0.012 ppbv
RT: 6.574 min Scan# 2001
Delta R.T. 0.010 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03


Tgt Ion: 97	Resp:	122
Ion	Ratio	Lower Upper
97	100	
83	65.9	64.6 97.0
85	95.6	39.4 59.2#
99	0.0	47.0 70.6#



#48
Dibromochloromethane
Concen: 0.002 ppbv
RT: 7.743 min Scan# 2365
Delta R.T. 0.366 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 129	Resp:	34
Ion	Ratio	Lower Upper
129	100	
127	111.8	39.3 117.9





#49

Bromoform

Concen: 0.007 ppbv

RT: 9.471 min Scan# 21

Delta R.T. -0.003 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :

VSTDICC0.03





Tgt Ion:173 Resp: 102

Ion Ratio Lower Upper

173 100

175 174.5 38.6 58.0#

171 179.4 40.6 61.0#



#50

4-Methyl-2-Pentanone

Concen: 0.005 ppbv

RT: 5.736 min Scan# 1745

Delta R.T. 0.003 min

Lab File: VL043209.D




Acq: 18 Nov 2025 12:49


Tgt Ion: 43 Resp: 140

Ion Ratio Lower Upper

43 100

58 0.0 28.4 42.6#









#51
2-Hexanone
Concen: 0.003 ppbv
RT: 7.419 min Scan# 21
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03


Tgt Ion: 43	Resp: 61
Ion Ratio	Lower Upper
43 100	
58 0.0	38.4 57.6#
98 0.0	0.0 0.0



#52
Tetrachloroethene
Concen: 0.017 ppbv
RT: 8.222 min Scan# 2513
Delta R.T. 0.010 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt	Ion:164	Resp:	158
Ion	Ratio	Lower	Upper
164	100		
166	25.6	99.4	149.2#
129	104.6	81.8	122.8
131	37.7	77.4	116.2#









#53
Toluene
Concen: 0.011 ppbv
RT: 6.930 min Scan# 2111
Delta R.T. 0.010 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion: 91 Resp: 323
Ion Ratio Lower Upper
91 100
92 0.0 47.4 71.0#



#54
1,2-Dibromoethane
Concen: 0.007 ppbv
RT: 7.645 min Scan# 2335
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 107 Resp: 102
Ion Ratio Lower Upper
107 100
109 102.6 72.1 108.1





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.875 min Scan# 21

Delta R.T. 0.003 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.03





Tgt Ion:117 Resp: 213547

Ion Ratio Lower Upper

117 100

82 66.2 56.1 84.1

119 32.3 25.0 37.6



#56

1,1,1,2-Tetrachloroethane

Concen: 0.007 ppbv

RT: 8.927 min Scan# 2731

Delta R.T. 0.013 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49




Tgt Ion:131 Resp: 114


Ion Ratio Lower Upper

131 100

133 798.2 77.4 116.2#

119 0.0 49.1 73.7#









#57
Chlorobenzene
Concen: 0.020 ppbv
RT: 8.914 min Scan# 2114
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion:112 Resp: 429
Ion Ratio Lower Upper
112 100
77 28.1 55.5 83.3#
114 13.2 29.1 35.5#



#58
Ethyl Benzene
Concen: 0.015 ppbv
RT: 9.519 min Scan# 2914
Delta R.T. 0.175 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 91 Resp: 613
Ion Ratio Lower Upper
91 100
106 33.2 22.8 34.2





#59

m/p-Xylene

Concen: 0.009 ppbv

RT: 9.545 min Scan# 2921

Delta R.T. 0.003 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC0.03

Tgt Ion: 91 Resp: 299

Ion Ratio Lower Upper

91 100

106 341.8 36.1 54.1#



#60

o-Xylene

Concen: 0.033 ppbv

RT: 9.956 min Scan# 3049

Delta R.T. 0.000 min

Lab File: VL043209.D




Acq: 18 Nov 2025 12:49


Tgt Ion: 91 Resp: 1078

Ion Ratio Lower Upper

91 100

106 41.6 20.9 62.8









#61
Styrene
Concen: 0.021 ppbv
RT: 9.869 min Scan# 3020
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03


Tgt Ion:104 Resp: 300
Ion Ratio Lower Upper
104 100
78 49.7 42.1 63.1
103 109.0 37.2 55.8#



#62
Isopropylbenzene
Concen: 0.034 ppbv
RT: 10.533 min Scan# 3227
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion:105 Resp: 1990
Ion Ratio Lower Upper
105 100
120 19.9 19.8 29.8





#63

1,1,2,2-Tetrachloroethane

Concen: 0.015 ppbv

RT: 9.963 min Scan# 3048

Delta R.T. 0.010 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.03





Tgt Ion: 83 Resp: 306

Ion Ratio Lower Upper

83 100

131 12.7 4.8 14.3

85 110.5 31.9 95.9#



#64

n-propylbenzene

Concen: 0.021 ppbv

RT: 10.982 min Scan# 3366

Delta R.T. 0.000 min

Lab File: VL043209.D




Acq: 18 Nov 2025 12:49


Tgt Ion: 120 Resp: 316

Ion Ratio Lower Upper

120 100

91 519.6 377.1 565.7









#65
tert-Butylbenzene
Concen: 0.033 ppbv
RT: 11.526 min Scan# 3534
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion:119 Resp: 1740
Ion Ratio Lower Upper
119 100
91 83.2 65.3 97.9
134 22.4 17.0 25.4



#66
Benzyl Chloride
Concen: 0.012 ppbv
RT: 11.607 min Scan# 3559
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 91 Resp: 62
Ion Ratio Lower Upper
91 100
126 0.0 14.4 21.6#
128 0.0 4.6 6.8#









#67
sec-Butylbenzene
Concen: 0.033 ppbv
RT: 11.762 min Scan# 3176
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion:105 Resp: 2398
Ion Ratio Lower Upper
105 100
134 32.4 15.0 22.6#



#68
1-Bromo-4-Fluorobenzene
Concen: 9.712 ppbv
RT: 10.371 min Scan# 3177
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion: 95 Resp: 152015
Ion Ratio Lower Upper
95 100
174 62.0 50.2 75.2
175 4.6 4.1 6.1





#69

p-Isopropyltoluene

Concen: 0.030 ppbv

RT: 11.924 min Scan# 3

Delta R.T. 0.006 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC0.03





Tgt Ion:119 Resp: 1875

Ion Ratio Lower Upper

119 100

134 26.7 20.2 30.2

91 37.8 23.8 35.8#



#70

n-Butylbenzene

Concen: 0.025 ppbv

RT: 12.277 min Scan# 3766

Delta R.T. 0.003 min

Lab File: VL043209.D

Acq: 18 Nov 2025 12:49




Tgt Ion: 91 Resp: 1502


Ion Ratio Lower Upper

91 100

92 51.6 43.5 65.3

134 19.0 19.0 28.6









#71
2-Chlorotoluene
Concen: 0.032 ppbv
RT: 10.895 min Scan# 311
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.03


Tgt Ion:	91	Resp:	1463
Ion Ratio	Lower	Upper	
91	100		
126	28.2	11.7	46.7



#72
4-Ethyltoluene
Concen: 0.010 ppbv
RT: 11.125 min Scan# 3410
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion:	105	Resp:	408
Ion Ratio	Lower	Upper	
105	100		
120	42.9	23.8	35.8#
91	0.0	10.3	15.5#
77	45.1	11.1	16.7#









#73
 1,3,5-Trimethylbenzene
 Concen: 0.031 ppbv
 RT: 11.203 min Scan# 3434
 Delta R.T. 0.006 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC0.03


Tgt Ion:105 Resp: 1041
 Ion Ratio Lower Upper
 105 100
 120 26.5 38.6 57.8#



#74
 1,2,4-Trimethylbenzene
 Concen: 0.021 ppbv
 RT: 11.536 min Scan# 3537
 Delta R.T. 0.006 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49

Tgt Ion:105 Resp: 789
 Ion Ratio Lower Upper
 105 100
 120 31.1 41.7 62.5#
 91 21.9 70.0 105.0#
 77 14.0 23.8 35.8#









#75
1,3-Dichlorobenzene
Concen: 0.010 ppbv
RT: 11.610 min Scan# 3557
Delta R.T. 0.010 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion:146 Resp: 214
Ion Ratio Lower Upper
146 100
111 126.2 23.4 70.3#
148 144.9 31.6 95.0#



#76
1,4-Dichlorobenzene
Concen: 0.014 ppbv
RT: 11.669 min Scan# 3578
Delta R.T. 0.003 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion:146 Resp: 292
Ion Ratio Lower Upper
146 100
111 0.0 22.8 68.3#
148 157.5 31.8 95.4#









#77
1,2-Dichlorobenzene
Concen: 0.027 ppbv
RT: 11.947 min Scan# 3664
Delta R.T. 0.006 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03


Tgt Ion:	146	Resp:	561
Ion	Ratio	Lower	Upper
146	100		
111	53.7	24.3	73.0
148	85.2	31.6	94.7



#78
Hexachloro-1,3-Butadiene
Concen: 0.036 ppbv
RT: 13.876 min Scan# 4260
Delta R.T. 0.000 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49

Tgt Ion:	225	Resp:	622
Ion	Ratio	Lower	Upper
225	100		
223	78.3	50.2	75.4#
227	75.2	51.1	76.7









#79
Naphthalene
Concen: 0.015 ppbv
RT: 13.520 min Scan# 4146
Delta R.T. 0.013 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC0.03


Tgt Ion:128 Resp: 504
Ion Ratio Lower Upper
128 100
127 65.8 9.9 14.9#
129 0.0 8.4 12.6#



#80
Naphthalene,2-methyl-
Concen: 0.007 ppbv
RT: 14.475 min Scan# 4445
Delta R.T. 0.023 min
Lab File: VL043209.D
Acq: 18 Nov 2025 12:49


Tgt Ion:142 Resp: 58
Ion Ratio Lower Upper
142 100
141 0.0 67.1 100.7#
115 41.4 31.9 47.9







#81
 1,2,4-Trichlorobenzene
 Concen: 0.010 ppbv
 RT: 13.442 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VL043209.D
 Acq: 18 Nov 2025 12:49

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03



Tgt Ion:180 Resp: 150
 Ion Ratio Lower Upper
 180 100
 182 132.0 77.4 116.2#
 184 92.0 24.2 36.4#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043210.D
 Acq On : 18 Nov 2025 13:24
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Quant Time: Nov 19 00:53:26 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 00:48:15 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	98968	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	258461	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	211247	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	159814	10.321	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.499	85	197987	14.184	ppbv	99
3) Chlorodifluoromethane	1.476	51	182813	13.317	ppbv	98
4) Chloromethane	1.531	50	72636	13.918	ppbv	97
5) Vinyl Chloride	1.580	62	71671	12.557	ppbv	96
6) Bromomethane	1.667	94	38265	13.830	ppbv	98
7) Chloroethane	1.703	64	29584	13.309	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	165709	14.271	ppbv	97
9) Propene	1.771	41	77762	12.212	ppbv	99
10) Heptane	4.972	43	225935	12.934	ppbv	100
11) Trichlorofluoromethane	1.874	101	200703	14.840	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.137	101	152772	14.469	ppbv	98
13) Ethanol	1.719	45	25468	12.425	ppbv	99
14) Bromoethene	1.781	108	55813	13.789	ppbv	96
15) Acetone	1.832	43	161036	11.712	ppbv	99
16) 1,3-Butadiene	1.609	39	81077	13.226	ppbv	99
17) tert-Butyl alcohol	2.036	59	257617	14.274	ppbv	99
18) 1,1-Dichloroethene	2.033	96	65731	13.370	ppbv	96
19) Isopropyl Alcohol	1.881	45	105271	13.549	ppbv	99
20) Methylene Chloride	2.059	84	59798	15.164	ppbv	93
21) Allyl Chloride	2.094	41	126959	13.724	ppbv	96
22) trans-1,2-Dichloroethene	2.337	96	77386	13.772	ppbv	94
23) Vinyl Acetate	2.457	43	232315	13.995	ppbv	99
24) 1,1-Dichloroethane	2.405	63	155017	14.292	ppbv	98
25) Ethyl Acetate	2.829	43	394074	13.265	ppbv	99
26) Hexane	2.823	57	172303	13.252	ppbv	97
27) Carbon Disulfide	2.150	76	191612	14.435	ppbv	97
28) Methyl tert-Butyl Ether	2.434	73	91132	14.279	ppbv	97
29) Chloroform	2.845	83	262803	14.221	ppbv	100
30) Cyclohexane	3.852	84	152015	13.946	ppbv	94
31) cis-1,2-Dichloroethene	2.716	61	180647	13.978	ppbv	98
32) 1,1,1-Trichloroethane	3.363	97	276915	13.930	ppbv	98
34) 2-Butanone	2.551	43	253855	13.914	ppbv	100
35) Carbon Tetrachloride	3.755	117	279560	14.250	ppbv	96
36) Benzene	3.651	78	348892	14.096	ppbv	99
37) 1,2-Dichloroethane	3.214	62	214248	14.798	ppbv	98
38) Trichloroethene	4.541	130	137027	14.123	ppbv	93
39) 1,2-Dichloropropane	4.302	63	128700	14.206	ppbv	98
40) 1,4-Dioxane	4.571	88	55905	13.728	ppbv #	95
41) Tetrahydrofuran	3.043	42	139722	13.868	ppbv	99
42) Bromodichloromethane	4.480	83	294008	14.721	ppbv	97
43) Methyl Methacrylate	4.872	69	179110	13.633	ppbv	100
44) 2,2,4-Trimethylpentane	4.632	57	595778	14.014	ppbv	99
45) t-1,3-Dichloropropene	6.406	75	177678	14.508	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043210.D
 Acq On : 18 Nov 2025 13:24
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Quant Time: Nov 19 00:53:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 00:48:15 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.587	75	205226	14.270	ppbv	100
47) 1,1,2-Trichloroethane	6.571	97	132701	13.738	ppbv	92
48) Dibromochloromethane	7.377	129	240736	14.564	ppbv	99
49) Bromoform	9.477	173	211325	14.434	ppbv	99
50) 4-Methyl-2-Pentanone	5.732	43	354421	13.790	ppbv	99
51) 2-Hexanone	7.422	43	280742	14.067	ppbv #	100
52) Tetrachloroethene	8.218	164	125132	13.591	ppbv	93
53) Toluene	6.924	91	414411	14.050	ppbv	99
54) 1,2-Dibromoethane	7.642	107	203913	14.413	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.917	131	214673	13.827	ppbv	99
57) Chlorobenzene	8.914	112	301986	13.923	ppbv	98
58) Ethyl Benzene	9.348	91	579890	14.262	ppbv	99
59) m/p-Xylene	9.545	91	423995	12.959	ppbv #	20
60) o-Xylene	9.956	91	450260	13.826	ppbv	99
61) Styrene	9.862	104	204972	14.382	ppbv	99
62) Isopropylbenzene	10.532	105	803487	13.791	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.956	83	280254	13.557	ppbv	99
64) n-propylbenzene	10.986	120	205637	13.737	ppbv	100
65) tert-Butylbenzene	11.526	119	689352	13.249	ppbv	98
66) Benzyl Chloride	11.610	91	71386	13.520	ppbv	100
67) sec-Butylbenzene	11.762	105	972780	13.394	ppbv	100
69) p-Isopropyltoluene	11.921	119	836888	13.602	ppbv	99
70) n-Butylbenzene	12.277	91	838598	14.050	ppbv	99
71) 2-Chlorotoluene	10.895	91	625884	13.976	ppbv	100
72) 4-Ethyltoluene	11.121	105	553654	13.961	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	460783	13.846	ppbv	97
74) 1,2,4-Trimethylbenzene	11.533	105	486995	13.409	ppbv	97
75) 1,3-Dichlorobenzene	11.604	146	292143	13.784	ppbv	99
76) 1,4-Dichlorobenzene	11.668	146	294723	13.903	ppbv	99
77) 1,2-Dichlorobenzene	11.944	146	273467	13.420	ppbv	98
78) Hexachloro-1,3-Butadiene	13.876	225	211153	12.416	ppbv	99
79) Naphthalene	13.510	128	473229	14.670	ppbv	99
80) Naphthalene,2-methyl-	14.452	142	139019	16.418	ppbv	99
81) 1,2,4-Trichlorobenzene	13.436	180	216260	14.990	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015


Tgt Ion: 49 Resp: 98968
Ion Ratio Lower Upper
49 100
128 46.7 23.0 69.0
130 58.4 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 14.184 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 85 Resp: 197987
Ion Ratio Lower Upper
85 100
87 32.4 25.4 38.2









#3
Chlorodifluoromethane
Concen: 13.317 ppbv
RT: 1.476 min Scan# 428
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion: 51 Resp: 182813
Ion Ratio Lower Upper
51 100
67 19.7 14.9 22.3



#4
Chloromethane
Concen: 13.918 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 50 Resp: 72636
Ion Ratio Lower Upper
50 100
52 31.7 24.1 36.1









#5
Vinyl Chloride
Concen: 12.557 ppbv
RT: 1.580 min Scan# 461
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion: 62 Resp: 71671
Ion Ratio Lower Upper
62 100
64 32.0 27.3 40.9



#6
Bromomethane
Concen: 13.830 ppbv
RT: 1.667 min Scan# 488
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 94 Resp: 38265
Ion Ratio Lower Upper
94 100
96 89.0 72.9 109.3









#7
Chloroethane
Concen: 13.309 ppbv
RT: 1.703 min Scan# 498
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion: 64 Resp: 29584
Ion Ratio Lower Upper
64 100
66 32.6 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 14.271 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 85 Resp: 165709
Ion Ratio Lower Upper
85 100
135 63.7 52.6 79.0





#9

Propene

Concen: 12.212 ppbv

RT: 1.771 min Scan# 51

Delta R.T. 0.291 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC015

Tgt Ion: 41 Resp: 77762

Ion Ratio Lower Upper

41 100

42 5.1 3.7 5.5



#10

Heptane

Concen: 12.934 ppbv

RT: 4.972 min Scan# 1509

Delta R.T. 0.006 min

Lab File: VL043210.D




Acq: 18 Nov 2025 13:24


Tgt Ion: 43 Resp: 225935

Ion Ratio Lower Upper

43 100

57 49.4 39.7 59.5









#11
Trichlorofluoromethane
Concen: 14.840 ppbv
RT: 1.874 min Scan# 511
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion:101 Resp: 200703
Ion Ratio Lower Upper
101 100
103 62.3 52.6 79.0



#12
1,1,2-Trichlorotrifluoroethane
Concen: 14.469 ppbv
RT: 2.137 min Scan# 633
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:101 Resp: 152772
Ion Ratio Lower Upper
101 100
151 72.3 59.2 88.8









#13
Ethanol
Concen: 12.425 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion: 45 Resp: 25468
Ion Ratio Lower Upper
45 100
46 37.9 15.0 59.8



#14
Bromoethene
Concen: 13.789 ppbv
RT: 1.781 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 108 Resp: 55813
Ion Ratio Lower Upper
108 100
106 105.8 88.9 133.3
79 20.6 16.5 24.7









#15
Acetone
Concen: 11.712 ppbv
RT: 1.832 min Scan# 51
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDICC015


Tgt Ion: 43 Resp: 161036
Ion Ratio Lower Upper
43 100
58 25.1 20.7 31.1



#16
1,3-Butadiene
Concen: 13.226 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 39 Resp: 81077
Ion Ratio Lower Upper
39 100
54 73.5 59.4 89.2





#17

tert-Butyl alcohol

Concen: 14.274 ppbv

RT: 2.036 min Scan# 602

Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :





VSTDICC015

Tgt Ion: 59 Resp: 257617

Ion Ratio Lower Upper

59 100

57 10.9 9.0 13.4



#18

1,1-Dichloroethene

Concen: 13.370 ppbv

RT: 2.033 min Scan# 601

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion: 96 Resp: 65731


Ion Ratio Lower Upper

96 100

61 204.3 170.6 255.8

98 64.7 52.2 78.4









#19
Isopropyl Alcohol
Concen: 13.549 ppbv
RT: 1.881 min Scan# 511
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015


Tgt Ion: 45 Resp: 105271
Ion Ratio Lower Upper
45 100
58 38.5 31.3 46.9



#20
Methylene Chloride
Concen: 15.164 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 84 Resp: 59798
Ion Ratio Lower Upper
84 100
49 153.8 132.9 199.3
51 47.0 34.8 52.2
86 58.2 48.2 72.2









#21
 Allyl Chloride
 Concen: 13.724 ppbv
 RT: 2.094 min Scan# 619
 Delta R.T. 0.003 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC015


Tgt Ion: 41	Resp: 126959
Ion Ratio	Lower Upper
41	100
76	27.5 22.1 33.1
39	81.3 61.0 91.4



#22
 trans-1,2-Dichloroethene
 Concen: 13.772 ppbv
 RT: 2.337 min Scan# 695
 Delta R.T. 0.003 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24

Tgt Ion: 96	Resp: 77386
Ion Ratio	Lower Upper
96	100
61	184.6 139.0 208.6
98	63.5 49.4 74.2









#23
 Vinyl Acetate
 Concen: 13.995 ppbv
 RT: 2.457 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24

Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC015

Tgt Ion: 43 Resp: 232315




Ion	Ratio	Lower	Upper
43	100		
86	8.2	6.2	9.2
42	10.1	8.0	12.0
44	5.1	4.0	6.0




#24
 1,1-Dichloroethane
 Concen: 14.292 ppbv
 RT: 2.405 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24

Tgt Ion: 63 Resp: 155017

Ion	Ratio	Lower	Upper
63	100		
98	4.9	2.8	8.4
100	3.0	1.3	3.9









#25
Ethyl Acetate
Concen: 13.265 ppbv
RT: 2.829 min Scan# 847
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion: 43	Resp: 394074
Ion Ratio	Lower Upper
43 100	
45 9.5	8.1 12.1
61 8.4	7.0 10.4
70 8.2	7.0 10.6



#26
Hexane
Concen: 13.252 ppbv
RT: 2.823 min Scan# 845
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 57	Resp: 172303		
Ion	Ratio	Lower	Upper
57	100		
41	95.4	74.6	111.8
43	229.0	179.0	268.4
56	53.5	42.3	63.5









#27
Carbon Disulfide
Concen: 14.435 ppbv
RT: 2.150 min Scan# 611
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion: 76 Resp: 191612
Ion Ratio Lower Upper
76 100
44 22.2 16.1 24.1
78 11.5 9.6 14.4



#28
Methyl tert-Butyl Ether
Concen: 14.279 ppbv
RT: 2.434 min Scan# 725
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 73 Resp: 91132
Ion Ratio Lower Upper
73 100
57 23.5 19.8 29.8









#29
Chloroform
Concen: 14.221 ppbv
RT: 2.845 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion: 83 Resp: 262803
Ion Ratio Lower Upper
83 100
85 63.2 50.5 75.7



#30
Cyclohexane
Concen: 13.946 ppbv
RT: 3.852 min Scan# 1163
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 84 Resp: 152015
Ion Ratio Lower Upper
84 100
56 123.0 104.7 157.1
41 82.7 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 13.978 ppbv
 RT: 2.716 min Scan# 81
 Delta R.T. -0.000 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015


Tgt Ion: 61 Resp: 180647
 Ion Ratio Lower Upper
 61 100
 96 60.9 49.3 73.9
 98 38.3 32.1 48.1



#32
 1,1,1-Trichloroethane
 Concen: 13.930 ppbv
 RT: 3.363 min Scan# 1012
 Delta R.T. 0.003 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24

Tgt Ion: 97 Resp: 276915
 Ion Ratio Lower Upper
 97 100
 99 64.4 50.3 75.5
 61 49.3 38.9 58.3









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion:114 Resp: 258461
Ion Ratio Lower Upper
114 100
63 26.2 21.0 31.6
88 18.8 14.9 22.3



#34
2-Butanone
Concen: 13.914 ppbv
RT: 2.551 min Scan# 761
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 43 Resp: 253855
Ion Ratio Lower Upper
43 100
72 22.4 18.1 27.1









#35
Carbon Tetrachloride
Concen: 14.250 ppbv
RT: 3.755 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion:117 Resp: 279560
Ion Ratio Lower Upper
117 100
119 97.4 74.6 112.0
121 30.4 23.2 34.8



#36
Benzene
Concen: 14.096 ppbv
RT: 3.651 min Scan# 1101
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 78 Resp: 348892
Ion Ratio Lower Upper
78 100
77 22.1 17.5 26.3
50 19.0 15.5 23.3





#37

1,2-Dichloroethane

Concen: 14.798 ppbv

RT: 3.214 min Scan# 9

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC015

Tgt Ion: 62 Resp: 214248

Ion Ratio Lower Upper

62 100

98 6.1 5.5 8.3



#38

Trichloroethene

Concen: 14.123 ppbv

RT: 4.541 min Scan# 1376

Delta R.T. 0.006 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Tgt Ion:130 Resp: 137027




Ion Ratio Lower Upper


130 100

95 109.6 85.7 128.5

132 97.4 73.4 110.0

97 77.0 52.7 79.1





#39

1,2-Dichloropropane

Concen: 14.206 ppbv

RT: 4.302 min Scan# 1

Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC015





Tgt Ion: 63 Resp: 128700

Ion Ratio Lower Upper

63 100

41 75.9 62.6 93.8

62 65.6 51.0 76.6



#40

1,4-Dioxane

Concen: 13.728 ppbv

RT: 4.571 min Scan# 1385

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Tgt Ion: 88 Resp: 55905




Ion Ratio Lower Upper


88 100

58 76.7 63.0 94.6

43 34.0 32.5 48.7

57 0.0 0.0 0.0





#41

Tetrahydrofuran

Concen: 13.868 ppbv

RT: 3.043 min Scan# 914

Delta R.T. -0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC015





Tgt Ion: 42 Resp: 139722

Ion Ratio Lower Upper

42 100

72 39.1 30.6 46.0

71 37.3 30.5 45.7



#42

Bromodichloromethane

Concen: 14.721 ppbv

RT: 4.480 min Scan# 1357

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion: 83 Resp: 294008


Ion Ratio Lower Upper

83 100

85 61.7 51.7 77.5

127 7.6 6.3 9.5





#43

Methyl Methacrylate

Concen: 13.633 ppbv

RT: 4.872 min Scan# 1476

Delta R.T. 0.006 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_1

ClientSampleId :

VSTDIC015





Tgt Ion: 69 Resp: 179110

Ion Ratio Lower Upper

69 100

41 157.9 126.7 190.1

100 35.7 28.6 43.0



#44

2,2,4-Trimethylpentane

Concen: 14.014 ppbv

RT: 4.632 min Scan# 1404

Delta R.T. 0.006 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion: 57 Resp: 595778


Ion Ratio Lower Upper

57 100

41 35.9 28.0 42.0

56 31.6 25.6 38.4





#45

t-1,3-Dichloropropene

Concen: 14.508 ppbv

RT: 6.406 min Scan# 1952

Delta R.T. 0.006 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC015

Tgt Ion: 75 Resp: 177678

Ion Ratio Lower Upper

75 100

77 30.2 24.3 36.5



#46

cis-1,3-Dichloropropene

Concen: 14.270 ppbv

RT: 5.587 min Scan# 1699

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion: 75 Resp: 205226


Ion Ratio Lower Upper

75 100

39 73.9 59.1 88.7

77 29.8 24.2 36.2









#47
1,1,2-Trichloroethane
Concen: 13.738 ppbv
RT: 6.571 min Scan# 2001
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion: 97	Resp: 132701
Ion Ratio	Lower Upper
97 100	
83 90.4	64.6 97.0
85 53.2	39.4 59.2
99 62.2	47.0 70.6



#48
Dibromochloromethane
Concen: 14.564 ppbv
RT: 7.377 min Scan# 2252
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:129	Resp: 240736		
Ion Ratio	Lower	Upper	
129 100			
127 77.5	39.3	117.9	









#49
Bromoform
Concen: 14.434 ppbv
RT: 9.477 min Scan# 2900
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion:173 Resp: 211325
Ion Ratio Lower Upper
173 100
175 48.6 38.6 58.0
171 51.5 40.6 61.0



#50
4-Methyl-2-Pentanone
Concen: 13.790 ppbv
RT: 5.732 min Scan# 1744
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 43 Resp: 354421
Ion Ratio Lower Upper
43 100
58 35.9 28.4 42.6









#51
2-Hexanone
Concen: 14.067 ppbv
RT: 7.422 min Scan# 21
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion: 43 Resp: 280742
Ion Ratio Lower Upper
43 100
58 47.7 38.4 57.6
98 0.1 0.0 0.0#



#52
Tetrachloroethene
Concen: 13.591 ppbv
RT: 8.218 min Scan# 2512
Delta R.T. 0.006 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 164 Resp: 125132
Ion Ratio Lower Upper
164 100
166 128.8 99.4 149.2
129 110.8 81.8 122.8
131 104.7 77.4 116.2









#53
Toluene
Concen: 14.050 ppbv
RT: 6.924 min Scan# 2111
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion: 91 Resp: 414411
Ion Ratio Lower Upper
91 100
92 58.8 47.4 71.0



#54
1,2-Dibromoethane
Concen: 14.413 ppbv
RT: 7.642 min Scan# 2334
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 107 Resp: 203913
Ion Ratio Lower Upper
107 100
109 90.6 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 211247
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015


Tgt Ion:117 Resp: 211247
Ion Ratio Lower Upper
117 100
82 68.4 56.1 84.1
119 32.6 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 13.827 ppbv
RT: 8.917 min Scan# 2728
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:131 Resp: 214673
Ion Ratio Lower Upper
131 100
133 94.9 77.4 116.2
119 61.5 49.1 73.7





#57

Chlorobenzene

Concen: 13.923 ppbv

RT: 8.914 min Scan# 21

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_1

ClientSampleId :

VSTDICC015





Tgt Ion:112 Resp: 301986

Ion Ratio Lower Upper

112 100

77 71.1 55.5 83.3

114 31.5 29.1 35.5



#58

Ethyl Benzene

Concen: 14.262 ppbv

RT: 9.348 min Scan# 2861

Delta R.T. 0.003 min

Lab File: VL043210.D




Acq: 18 Nov 2025 13:24


Tgt Ion: 91 Resp: 579890

Ion Ratio Lower Upper

91 100

106 29.2 22.8 34.2





#59

m/p-Xylene

Concen: 12.959 ppbv

RT: 9.545 min Scan# 2921

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC015

Tgt Ion: 91 Resp: 423995

Ion Ratio Lower Upper

91 100

106 97.4 36.1 54.1#



#60

o-Xylene

Concen: 13.826 ppbv

RT: 9.956 min Scan# 3049

Delta R.T. -0.000 min

Lab File: VL043210.D




Acq: 18 Nov 2025 13:24


Tgt Ion: 91 Resp: 450260

Ion Ratio Lower Upper

91 100

106 42.4 20.9 62.8









#61
Styrene
Concen: 14.382 ppbv
RT: 9.862 min Scan# 3020
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_L
ClientSampleId :
VSTDICC015


Tgt Ion:104 Resp: 204972
Ion Ratio Lower Upper
104 100
78 53.7 42.1 63.1
103 46.9 37.2 55.8



#62
Isopropylbenzene
Concen: 13.791 ppbv
RT: 10.532 min Scan# 3227
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:105 Resp: 803487
Ion Ratio Lower Upper
105 100
120 25.3 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 13.557 ppbv
RT: 9.956 min Scan# 3049
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion: 83 Resp: 280254
Ion Ratio Lower Upper
83 100
131 9.7 4.8 14.3
85 64.9 31.9 95.9



#64
n-propylbenzene
Concen: 13.737 ppbv
RT: 10.986 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 120 Resp: 205637
Ion Ratio Lower Upper
120 100
91 471.6 377.1 565.7









#65
tert-Butylbenzene
Concen: 13.249 ppbv
RT: 11.526 min Scan# 31
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion:119 Resp: 689352
Ion Ratio Lower Upper
119 100
91 83.7 65.3 97.9
134 20.9 17.0 25.4



#66
Benzyl Chloride
Concen: 13.520 ppbv
RT: 11.610 min Scan# 3560
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 91 Resp: 71386
Ion Ratio Lower Upper
91 100
126 17.9 14.4 21.6
128 5.9 4.6 6.8









#67
 sec-Butylbenzene
 Concen: 13.394 ppbv
 RT: 11.762 min Scan# 3176
 Delta R.T. -0.000 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDIC015


Tgt Ion:105 Resp: 972780
 Ion Ratio Lower Upper
 105 100
 134 18.9 15.0 22.6



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.321 ppbv
 RT: 10.367 min Scan# 3176
 Delta R.T. -0.000 min
 Lab File: VL043210.D
 Acq: 18 Nov 2025 13:24

Tgt Ion: 95 Resp: 159814
 Ion Ratio Lower Upper
 95 100
 174 60.3 50.2 75.2
 175 4.7 4.1 6.1









#69
p-Isopropyltoluene
Concen: 13.602 ppbv
RT: 11.921 min Scan# 3656
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion:119 Resp: 836888
Ion Ratio Lower Upper
119 100
134 24.8 20.2 30.2
91 30.3 23.8 35.8



#70
n-Butylbenzene
Concen: 14.050 ppbv
RT: 12.277 min Scan# 3766
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 91 Resp: 838598
Ion Ratio Lower Upper
91 100
92 54.9 43.5 65.3
134 23.4 19.0 28.6









#71
2-Chlorotoluene
Concen: 13.976 ppbv
RT: 10.895 min Scan# 31
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion: 91 Resp: 625884
Ion Ratio Lower Upper
91 100
126 28.9 11.7 46.7



#72
4-Ethyltoluene
Concen: 13.961 ppbv
RT: 11.121 min Scan# 3409
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion: 105 Resp: 553654
Ion Ratio Lower Upper
105 100
120 29.3 23.8 35.8
91 13.2 10.3 15.5
77 14.1 11.1 16.7





#73

1,3,5-Trimethylbenzene

Concen: 13.846 ppbv

RT: 11.199 min Scan# 3432

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_1

ClientSampleId :





VSTDIC015

Tgt Ion:105 Resp: 460783

Ion Ratio Lower Upper

105 100

120 45.9 38.6 57.8



#74

1,2,4-Trimethylbenzene

Concen: 13.409 ppbv

RT: 11.533 min Scan# 3536

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Tgt Ion:105 Resp: 486995




Ion Ratio Lower Upper


105 100

120 50.4 41.7 62.5

91 83.4 70.0 105.0

77 28.5 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 13.784 ppbv
RT: 11.604 min Scan# 3557
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion:146 Resp: 292143
Ion Ratio Lower Upper
146 100
111 48.4 23.4 70.3
148 63.1 31.6 95.0



#76
1,4-Dichlorobenzene
Concen: 13.903 ppbv
RT: 11.668 min Scan# 3578
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:146 Resp: 294723
Ion Ratio Lower Upper
146 100
111 47.2 22.8 68.3
148 63.2 31.8 95.4









#77
1,2-Dichlorobenzene
Concen: 13.420 ppbv
RT: 11.944 min Scan# 3663
Delta R.T. 0.003 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24




Instrument :
MSVOA_1
ClientSampleId :
VSTDIC015


Tgt Ion:146 Resp: 273467
Ion Ratio Lower Upper
146 100
111 50.2 24.3 73.0
148 63.9 31.6 94.7



#78
Hexachloro-1,3-Butadiene
Concen: 12.416 ppbv
RT: 13.876 min Scan# 4260
Delta R.T. -0.000 min
Lab File: VL043210.D
Acq: 18 Nov 2025 13:24

Tgt Ion:225 Resp: 211153
Ion Ratio Lower Upper
225 100
223 64.9 50.2 75.4
227 63.9 51.1 76.7





#79

Naphthalene

Concen: 14.670 ppbv

RT: 13.510 min Scan# 4147

Delta R.T. 0.003 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC015





Tgt Ion:128 Resp: 473229

Ion Ratio Lower Upper

128 100

127 12.4 9.9 14.9

129 11.1 8.4 12.6



#80

Naphthalene,2-methyl-

Concen: 16.418 ppbv

RT: 14.452 min Scan# 4438

Delta R.T. -0.000 min

Lab File: VL043210.D

Acq: 18 Nov 2025 13:24




Tgt Ion:142 Resp: 139019


Ion Ratio Lower Upper

142 100

141 82.6 67.1 100.7

115 40.2 31.9 47.9





#81

1,2,4-Trichlorobenzene

Concen: 14.990 ppbv

RT: 13.436 min Scan# 41

Delta R.T. -0.000 min

Lab File: VL043210.D


Acq: 18 Nov 2025 13:24

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC015





Tgt Ion:180 Resp: 216260

Ion Ratio Lower Upper

180 100

182 95.3 77.4 116.2

184 30.1 24.2 36.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	97854	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.956	114	261914	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.879	117	216212	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.374 95 159242 10.048 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.500%

Target Compounds

					Qvalue	
2) Dichlorodifluoromethane	1.499	85	133374	9.664	ppbv	98
3) Chlorodifluoromethane	1.476	51	124858	9.199	ppbv	98
4) Chloromethane	1.535	50	46906	9.090	ppbv	95
5) Vinyl Chloride	1.580	62	47932	8.493	ppbv	95
6) Bromomethane	1.667	94	22230	8.126	ppbv	99
7) Chloroethane	1.703	64	19854	9.033	ppbv	98
8) Dichlorotetrafluoroethane	1.554	85	110587	9.632	ppbv	97
9) Propene	1.771	41	50816	8.817	ppbv	100
10) Heptane	4.975	43	149202	8.698	ppbv	99
11) Trichlorofluoromethane	1.878	101	134746	10.077	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	100984	9.673	ppbv	99
13) Ethanol	1.719	45	15371	7.584	ppbv	92
14) Bromoethene	1.781	108	36979	9.240	ppbv	98
15) Acetone	1.832	43	104830	7.711	ppbv	99
16) 1,3-Butadiene	1.609	39	54795	9.041	ppbv	96
17) tert-Butyl alcohol	2.036	59	164987	9.245	ppbv	99
18) 1,1-Dichloroethene	2.033	96	43833	9.017	ppbv	99
19) Isopropyl Alcohol	1.884	45	64772	8.431	ppbv	97
20) Methylene Chloride	2.062	84	41254	10.323	ppbv	95
21) Allyl Chloride	2.095	41	75516	8.256	ppbv	97
22) trans-1,2-Dichloroethene	2.341	96	50228	9.041	ppbv	92
23) Vinyl Acetate	2.460	43	140156	8.539	ppbv	98
24) 1,1-Dichloroethane	2.409	63	103187	9.622	ppbv	97
25) Ethyl Acetate	2.829	43	258609	8.820	ppbv	99
26) Hexane	2.826	57	114312	8.892	ppbv	98
27) Carbon Disulfide	2.150	76	125048	9.528	ppbv	97
28) Methyl tert-Butyl Ether	2.438	73	64183	10.171	ppbv	98
29) Chloroform	2.846	83	173977	9.521	ppbv	99
30) Cyclohexane	3.852	84	98598	9.148	ppbv	97
31) cis-1,2-Dichloroethene	2.719	61	119843	9.379	ppbv	98
32) 1,1,1-Trichloroethane	3.363	97	182262	9.358	ppbv	98
34) 2-Butanone	2.551	43	160977	8.707	ppbv	98
35) Carbon Tetrachloride	3.761	117	182902	9.200	ppbv	98
36) Benzene	3.655	78	233094	9.293	ppbv	99
37) 1,2-Dichloroethane	3.214	62	143363	9.772	ppbv	98
38) Trichloroethene	4.542	130	90993	9.272	ppbv	95
39) 1,2-Dichloropropane	4.305	63	83416	9.086	ppbv	95
40) 1,4-Dioxane	4.571	88	32109	7.780	ppbv #	95
41) Tetrahydrofuran	3.046	42	90769	8.891	ppbv	98
42) Bromodichloromethane	4.486	83	195442	9.657	ppbv	96
43) Methyl Methacrylate	4.872	69	116325	8.737	ppbv	98
44) 2,2,4-Trimethylpentane	4.635	57	393438	9.133	ppbv	99
45) t-1,3-Dichloropropene	6.409	75	97587	7.863	ppbv	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111825\
 Data File : VL043211.D
 Acq On : 18 Nov 2025 13:56
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825

Quant Time: Nov 19 01:34:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	115476	7.924	ppbv	97
47) 1,1,2-Trichloroethane	6.577	97	88525	9.086	ppbv	95
48) Dibromochloromethane	7.380	129	159365	9.514	ppbv	99
49) Bromoform	9.477	173	134790	9.085	ppbv	99
50) 4-Methyl-2-Pentanone	5.742	43	224791	8.631	ppbv	99
51) 2-Hexanone	7.429	43	166506	8.233	ppbv #	100
52) Tetrachloroethene	8.222	164	82890	8.884	ppbv	98
53) Toluene	6.924	91	273172	9.197	ppbv	99
54) 1,2-Dibromoethane	7.645	107	124804	8.705	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.921	131	142512	8.969	ppbv	98
57) Chlorobenzene	8.917	112	199373	8.981	ppbv	96
58) Ethyl Benzene	9.351	91	381474	9.167	ppbv	97
59) m/p-Xylene	9.549	91	265475	7.957	ppbv #	13
60) o-Xylene	9.960	91	295836	8.876	ppbv	99
61) Styrene	9.866	104	127117	8.714	ppbv	97
62) Isopropylbenzene	10.536	105	525717	8.816	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.960	83	178591	8.421	ppbv	100
64) n-propylbenzene	10.986	120	131634	8.591	ppbv	96
65) tert-Butylbenzene	11.529	119	450819	8.465	ppbv	99
66) Benzyl Chloride	11.614	91	27878	5.159	ppbv	99
67) sec-Butylbenzene	11.766	105	623520	8.388	ppbv	100
69) p-Isopropyltoluene	11.924	119	520880	8.272	ppbv	99
70) n-Butylbenzene	12.280	91	513027	8.398	ppbv	99
71) 2-Chlorotoluene	10.898	91	396434	8.649	ppbv	99
72) 4-Ethyltoluene	11.122	105	356325	8.779	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	294977	8.660	ppbv	95
74) 1,2,4-Trimethylbenzene	11.533	105	314606	8.463	ppbv	94
75) 1,3-Dichlorobenzene	11.604	146	180491	8.320	ppbv	99
76) 1,4-Dichlorobenzene	11.669	146	179491	8.273	ppbv	99
77) 1,2-Dichlorobenzene	11.944	146	166137	7.966	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	125707	7.222	ppbv	100
79) Naphthalene	13.510	128	273029	8.270	ppbv	99
80) Naphthalene,2-methyl-	14.455	142	74802	8.631	ppbv	97
81) 1,2,4-Trichlorobenzene	13.436	180	125470	8.497	ppbv	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 49 Resp: 97854
Ion Ratio Lower Upper
49 100
128 46.7 23.0 69.0
130 59.0 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 9.664 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 85 Resp: 133374
Ion Ratio Lower Upper
85 100
87 33.0 25.4 38.2









#3
Chlorodifluoromethane
Concen: 9.199 ppbv
RT: 1.476 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 51 Resp: 124858
Ion Ratio Lower Upper
51 100
67 19.6 14.9 22.3



#4
Chloromethane
Concen: 9.090 ppbv
RT: 1.535 min Scan# 447
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 50 Resp: 46906
Ion Ratio Lower Upper
50 100
52 32.8 24.1 36.1









#5
 Vinyl Chloride
 Concen: 8.493 ppbv
 RT: 1.580 min Scan# 461
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825


Tgt Ion: 62 Resp: 47932
 Ion Ratio Lower Upper
 62 100
 64 31.4 27.3 40.9



#6
 Bromomethane
 Concen: 8.126 ppbv
 RT: 1.667 min Scan# 488
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56

Tgt Ion: 94 Resp: 22230
 Ion Ratio Lower Upper
 94 100
 96 92.2 72.9 109.3









#7
Chloroethane
Concen: 9.033 ppbv
RT: 1.703 min Scan# 499
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion: 64 Resp: 19854
Ion Ratio Lower Upper
64 100
66 32.4 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 9.632 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 85 Resp: 110587
Ion Ratio Lower Upper
85 100
135 63.5 52.6 79.0





#9

Propene

Concen: 8.817 ppbv

RT: 1.771 min Scan# 51

Delta R.T. 0.291 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :





ICVVL111825

Tgt Ion: 41 Resp: 50816

Ion Ratio Lower Upper

41 100

42 4.7 3.7 5.5



#10

Heptane

Concen: 8.698 ppbv

RT: 4.975 min Scan# 1510

Delta R.T. 0.010 min

Lab File: VL043211.D




Acq: 18 Nov 2025 13:56


Tgt Ion: 43 Resp: 149202

Ion Ratio Lower Upper

43 100

57 49.1 39.7 59.5









#11
 Trichlorofluoromethane
 Concen: 10.077 ppbv
 RT: 1.878 min Scan# 511
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825


Tgt Ion:101 Resp: 134746
 Ion Ratio Lower Upper
 101 100
 103 63.8 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 9.673 ppbv
 RT: 2.140 min Scan# 634
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56

Tgt Ion:101 Resp: 100984
 Ion Ratio Lower Upper
 101 100
 151 73.1 59.2 88.8









#13
Ethanol
Concen: 7.584 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 45 Resp: 15371
Ion Ratio Lower Upper
45 100
46 42.2 15.0 59.8



#14
Bromoethene
Concen: 9.240 ppbv
RT: 1.781 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 108 Resp: 36979
Ion Ratio Lower Upper
108 100
106 109.3 88.9 133.3
79 21.9 16.5 24.7









#15
Acetone
Concen: 7.711 ppbv
RT: 1.832 min Scan# 51
Delta R.T. 0.000 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 43 Resp: 104830
Ion Ratio Lower Upper
43 100
58 25.2 20.7 31.1



#16
1,3-Butadiene
Concen: 9.041 ppbv
RT: 1.609 min Scan# 470
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 39 Resp: 54795
Ion Ratio Lower Upper
39 100
54 71.3 59.4 89.2









#17
tert-Butyl alcohol
Concen: 9.245 ppbv
RT: 2.036 min Scan# 602
Delta R.T. 0.000 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion: 59 Resp: 164987
Ion Ratio Lower Upper
59 100
57 10.8 9.0 13.4



#18
1,1-Dichloroethene
Concen: 9.017 ppbv
RT: 2.033 min Scan# 601
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 96 Resp: 43833
Ion Ratio Lower Upper
96 100
61 214.0 170.6 255.8
98 62.9 52.2 78.4









#19
Isopropyl Alcohol
Concen: 8.431 ppbv
RT: 1.884 min Scan# 51
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion: 45 Resp: 64772
Ion Ratio Lower Upper
45 100
58 40.8 31.3 46.9



#20
Methylene Chloride
Concen: 10.323 ppbv
RT: 2.062 min Scan# 610
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 84 Resp: 41254
Ion Ratio Lower Upper
84 100
49 155.1 132.9 199.3
51 43.9 34.8 52.2
86 60.8 48.2 72.2





#21

Allyl Chloride

Concen: 8.256 ppbv

RT: 2.095 min Scan# 61

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_1

ClientSampleId :

ICVVL111825





Tgt Ion: 41 Resp: 75516

Ion Ratio Lower Upper

41 100

76 27.9 22.1 33.1

39 79.3 61.0 91.4



#22

trans-1,2-Dichloroethene

Concen: 9.041 ppbv

RT: 2.341 min Scan# 696

Delta R.T. 0.006 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56




Tgt Ion: 96 Resp: 50228


Ion Ratio Lower Upper

96 100

61 185.7 139.0 208.6

98 66.6 49.4 74.2









#23
Vinyl Acetate
Concen: 8.539 ppbv
RT: 2.460 min Scan# 717
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825

Tgt Ion: 43 Resp: 140156




Ion	Ratio	Lower	Upper
43	100		
86	8.4	6.2	9.2
42	9.1	8.0	12.0
44	4.4	4.0	6.0




#24
1,1-Dichloroethane
Concen: 9.622 ppbv
RT: 2.409 min Scan# 717
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 63 Resp: 103187

Ion	Ratio	Lower	Upper
63	100		
98	4.1	2.8	8.4
100	2.6	1.3	3.9









#25
Ethyl Acetate
Concen: 8.820 ppbv
RT: 2.829 min Scan# 847
Delta R.T. 0.000 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion:	43	Resp:	258609
Ion	Ratio	Lower	Upper
43	100		
45	9.6	8.1	12.1
61	8.5	7.0	10.4
70	8.2	7.0	10.6



#26
Hexane
Concen: 8.892 ppbv
RT: 2.826 min Scan# 846
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion:	57	Resp:	114312
Ion	Ratio	Lower	Upper
57	100		
41	96.0	74.6	111.8
43	226.4	179.0	268.4
56	53.6	42.3	63.5









#27
Carbon Disulfide
Concen: 9.528 ppbv
RT: 2.150 min Scan# 61
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion: 76 Resp: 125048
Ion Ratio Lower Upper
76 100
44 21.9 16.1 24.1
78 11.6 9.6 14.4



#28
Methyl tert-Butyl Ether
Concen: 10.171 ppbv
RT: 2.438 min Scan# 726
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 73 Resp: 64183
Ion Ratio Lower Upper
73 100
57 23.6 19.8 29.8





#29

Chloroform

Concen: 9.521 ppbv

RT: 2.846 min Scan# 81

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :





ICVVL111825

Tgt Ion: 83 Resp: 173977

Ion Ratio Lower Upper

83 100

85 64.2 50.5 75.7



#30

Cyclohexane

Concen: 9.148 ppbv

RT: 3.852 min Scan# 1163

Delta R.T. 0.006 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56




Tgt Ion: 84 Resp: 98598


Ion Ratio Lower Upper

84 100

56 127.4 104.7 157.1

41 84.5 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 9.379 ppbv
 RT: 2.719 min Scan# 81
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_1
 ClientSampleId :
 ICVVL111825


Tgt Ion: 61 Resp: 119843
 Ion Ratio Lower Upper
 61 100
 96 62.9 49.3 73.9
 98 42.0 32.1 48.1



#32
 1,1,1-Trichloroethane
 Concen: 9.358 ppbv
 RT: 3.363 min Scan# 1012
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56

Tgt Ion: 97 Resp: 182262
 Ion Ratio Lower Upper
 97 100
 99 64.2 50.3 75.5
 61 49.9 38.9 58.3









#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 3.956 min Scan# 1195
 Delta R.T. 0.006 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_1
 ClientSampleId :
 ICVVL111825


Tgt Ion:114 Resp: 261914
 Ion Ratio Lower Upper
 114 100
 63 26.0 21.0 31.6
 88 18.3 14.9 22.3



#34
 2-Butanone
 Concen: 8.707 ppbv
 RT: 2.551 min Scan# 761
 Delta R.T. 0.000 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56

Tgt Ion: 43 Resp: 160977
 Ion Ratio Lower Upper
 43 100
 72 21.8 18.1 27.1









#35
Carbon Tetrachloride
Concen: 9.200 ppbv
RT: 3.761 min Scan# 1135
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion:117 Resp: 182902
Ion Ratio Lower Upper
117 100
119 94.7 74.6 112.0
121 30.4 23.2 34.8



#36
Benzene
Concen: 9.293 ppbv
RT: 3.655 min Scan# 1102
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 78 Resp: 233094
Ion Ratio Lower Upper
78 100
77 21.8 17.5 26.3
50 20.4 15.5 23.3









#37
1,2-Dichloroethane
Concen: 9.772 ppbv
RT: 3.214 min Scan# 965
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 62 Resp: 143363
Ion Ratio Lower Upper
62 100
98 6.3 5.5 8.3



#38
Trichloroethene
Concen: 9.272 ppbv
RT: 4.542 min Scan# 1376
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 130 Resp: 90993
Ion Ratio Lower Upper
130 100
95 112.7 85.7 128.5
132 94.1 73.4 110.0
97 70.6 52.7 79.1





#39

1,2-Dichloropropane

Concen: 9.086 ppbv

RT: 4.305 min Scan# 1

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :

ICVVL111825





Tgt Ion: 63 Resp: 83416

Ion Ratio Lower Upper

63 100

41 85.0 62.6 93.8

62 64.6 51.0 76.6



#40

1,4-Dioxane

Concen: 7.780 ppbv

RT: 4.571 min Scan# 1385

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Tgt Ion: 88 Resp: 32109




Ion Ratio Lower Upper


88 100

58 74.8 63.0 94.6

43 37.4 32.5 48.7

57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 8.891 ppbv
RT: 3.046 min Scan# 914
Delta R.T. 0.000 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 42 Resp: 90769
Ion Ratio Lower Upper
42 100
72 38.0 30.6 46.0
71 36.1 30.5 45.7



#42
Bromodichloromethane
Concen: 9.657 ppbv
RT: 4.486 min Scan# 1359
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 83 Resp: 195442
Ion Ratio Lower Upper
83 100
85 61.0 51.7 77.5
127 7.6 6.3 9.5









#43
Methyl Methacrylate
Concen: 8.737 ppbv
RT: 4.872 min Scan# 1476
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion: 69 Resp: 116325
Ion Ratio Lower Upper
69 100
41 155.2 126.7 190.1
100 36.5 28.6 43.0



#44
2,2,4-Trimethylpentane
Concen: 9.133 ppbv
RT: 4.635 min Scan# 1405
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 57 Resp: 393438
Ion Ratio Lower Upper
57 100
41 35.8 28.0 42.0
56 31.5 25.6 38.4





#45

t-1,3-Dichloropropene

Concen: 7.863 ppbv

RT: 6.409 min Scan# 1950

Delta R.T. 0.010 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_1

ClientSampleId :





ICVVL111825

Tgt Ion: 75 Resp: 97587

Ion Ratio Lower Upper

75 100

77 30.3 24.3 36.5



#46

cis-1,3-Dichloropropene

Concen: 7.924 ppbv

RT: 5.593 min Scan# 1701

Delta R.T. 0.010 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56




Tgt Ion: 75 Resp: 115476


Ion Ratio Lower Upper

75 100

39 70.4 59.1 88.7

77 30.4 24.2 36.2





#47

1,1,2-Trichloroethane

Concen: 9.086 ppbv

RT: 6.577 min Scan# 2005

Delta R.T. 0.013 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :

ICVVL111825

Tgt Ion: 97 Resp: 88525





Ion Ratio Lower Upper

97 100

83 85.4 64.6 97.0

85 52.6 39.4 59.2

99 62.6 47.0 70.6



#48

Dibromochloromethane

Concen: 9.514 ppbv

RT: 7.380 min Scan# 2253

Delta R.T. 0.003 min

Lab File: VL043211.D




Acq: 18 Nov 2025 13:56


Tgt Ion: 129 Resp: 159365

Ion Ratio Lower Upper

129 100

127 77.3 39.3 117.9





#49

Bromoform

Concen: 9.085 ppbv

RT: 9.477 min Scan# 2900

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :

ICVVL111825





Tgt Ion:173 Resp: 134790

Ion Ratio Lower Upper

173 100

175 47.2 38.6 58.0

171 50.2 40.6 61.0



#50

4-Methyl-2-Pentanone

Concen: 8.631 ppbv

RT: 5.742 min Scan# 1747

Delta R.T. 0.010 min

Lab File: VL043211.D




Acq: 18 Nov 2025 13:56


Tgt Ion: 43 Resp: 224791

Ion Ratio Lower Upper

43 100

58 36.0 28.4 42.6









#51
2-Hexanone
Concen: 8.233 ppbv
RT: 7.429 min Scan# 21
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 43 Resp: 166506
Ion Ratio Lower Upper
43 100
58 47.9 38.4 57.6
98 0.0 0.0 0.0#



#52
Tetrachloroethene
Concen: 8.884 ppbv
RT: 8.222 min Scan# 2513
Delta R.T. 0.010 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 164 Resp: 82890
Ion Ratio Lower Upper
164 100
166 121.3 99.4 149.2
129 100.1 81.8 122.8
131 94.6 77.4 116.2









#53
Toluene
Concen: 9.197 ppbv
RT: 6.924 min Scan# 2111
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 91 Resp: 273172
Ion Ratio Lower Upper
91 100
92 59.8 47.4 71.0



#54
1,2-Dibromoethane
Concen: 8.705 ppbv
RT: 7.645 min Scan# 2335
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 107 Resp: 124804
Ion Ratio Lower Upper
107 100
109 91.1 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.879 min Scan# 21
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion:117 Resp: 216212
Ion Ratio Lower Upper
117 100
82 70.6 56.1 84.1
119 32.3 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 8.969 ppbv
RT: 8.921 min Scan# 2729
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion:131 Resp: 142512
Ion Ratio Lower Upper
131 100
133 94.3 77.4 116.2
119 61.9 49.1 73.7









#57
Chlorobenzene
Concen: 8.981 ppbv
RT: 8.917 min Scan# 21
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion:112 Resp: 199373
Ion Ratio Lower Upper
112 100
77 72.7 55.5 83.3
114 35.1 29.1 35.5



#58
Ethyl Benzene
Concen: 9.167 ppbv
RT: 9.351 min Scan# 2862
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 91 Resp: 381474
Ion Ratio Lower Upper
91 100
106 30.2 22.8 34.2





#59

m/p-Xylene

Concen: 7.957 ppbv

RT: 9.549 min Scan# 2921

Delta R.T. 0.006 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_L

ClientSampleId :





ICVVL111825

Tgt Ion: 91 Resp: 265475

Ion Ratio Lower Upper

91 100

106 102.0 36.1 54.1#



#60

o-Xylene

Concen: 8.876 ppbv

RT: 9.960 min Scan# 3050

Delta R.T. 0.003 min

Lab File: VL043211.D




Acq: 18 Nov 2025 13:56


Tgt Ion: 91 Resp: 295836

Ion Ratio Lower Upper

91 100

106 42.8 20.9 62.8





#61

Styrene

Concen: 8.714 ppbv

RT: 9.866 min Scan# 3020

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_1

ClientSampleId :

ICVVL111825





Tgt Ion:104 Resp: 127117

Ion Ratio Lower Upper

104 100

78 54.7 42.1 63.1

103 48.0 37.2 55.8



#62

Isopropylbenzene

Concen: 8.816 ppbv

RT: 10.536 min Scan# 3228

Delta R.T. 0.003 min

Lab File: VL043211.D




Acq: 18 Nov 2025 13:56


Tgt Ion:105 Resp: 525717

Ion Ratio Lower Upper

105 100

120 25.1 19.8 29.8









#63
1,1,2,2-Tetrachloroethane
Concen: 8.421 ppbv
RT: 9.960 min Scan# 3048
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion: 83 Resp: 178591
Ion Ratio Lower Upper
83 100
131 9.6 4.8 14.3
85 63.9 31.9 95.9



#64
n-propylbenzene
Concen: 8.591 ppbv
RT: 10.986 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 120 Resp: 131634
Ion Ratio Lower Upper
120 100
91 482.2 377.1 565.7









#65
tert-Butylbenzene
Concen: 8.465 ppbv
RT: 11.529 min Scan# 3535
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion:119 Resp: 450819
Ion Ratio Lower Upper
119 100
91 83.0 65.3 97.9
134 20.8 17.0 25.4



#66
Benzyl Chloride
Concen: 5.159 ppbv
RT: 11.614 min Scan# 3561
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 91 Resp: 27878
Ion Ratio Lower Upper
91 100
126 17.3 14.4 21.6
128 5.6 4.6 6.8









#67
 sec-Butylbenzene
 Concen: 8.388 ppbv
 RT: 11.766 min Scan# 30
 Delta R.T. 0.003 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56




Instrument :
 MSVOA_1
 ClientSampleId :
 ICVVL111825


Tgt Ion:105 Resp: 623520
 Ion Ratio Lower Upper
 105 100
 134 19.0 15.0 22.6



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.048 ppbv
 RT: 10.374 min Scan# 3178
 Delta R.T. 0.006 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56

Tgt Ion: 95 Resp: 159242
 Ion Ratio Lower Upper
 95 100
 174 60.4 50.2 75.2
 175 4.7 4.1 6.1









#69
p-Isopropyltoluene
Concen: 8.272 ppbv
RT: 11.924 min Scan# 3657
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion:119 Resp: 520880
Ion Ratio Lower Upper
119 100
134 25.2 20.2 30.2
91 30.6 23.8 35.8



#70
n-Butylbenzene
Concen: 8.398 ppbv
RT: 12.280 min Scan# 3767
Delta R.T. 0.006 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 91 Resp: 513027
Ion Ratio Lower Upper
91 100
92 54.0 43.5 65.3
134 23.1 19.0 28.6









#71
2-Chlorotoluene
Concen: 8.649 ppbv
RT: 10.898 min Scan# 31
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion: 91 Resp: 396434
Ion Ratio Lower Upper
91 100
126 28.9 11.7 46.7



#72
4-Ethyltoluene
Concen: 8.779 ppbv
RT: 11.122 min Scan# 3409
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion: 105 Resp: 356325
Ion Ratio Lower Upper
105 100
120 29.3 23.8 35.8
91 13.5 10.3 15.5
77 13.9 11.1 16.7









#73
1,3,5-Trimethylbenzene
Concen: 8.660 ppbv
RT: 11.199 min Scan# 34
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion:105 Resp: 294977
Ion Ratio Lower Upper
105 100
120 45.0 38.6 57.8



#74
1,2,4-Trimethylbenzene
Concen: 8.463 ppbv
RT: 11.533 min Scan# 3536
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion:105 Resp: 314606
Ion Ratio Lower Upper
105 100
120 51.8 41.7 62.5
91 96.9 70.0 105.0
77 32.1 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 8.320 ppbv
RT: 11.604 min Scan# 3557
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_1
ClientSampleId :
ICVVL111825


Tgt Ion:146 Resp: 180491
Ion Ratio Lower Upper
146 100
111 47.7 23.4 70.3
148 62.7 31.6 95.0



#76
1,4-Dichlorobenzene
Concen: 8.273 ppbv
RT: 11.669 min Scan# 3578
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56

Tgt Ion:146 Resp: 179491
Ion Ratio Lower Upper
146 100
111 46.3 22.8 68.3
148 63.4 31.8 95.4





#77

1,2-Dichlorobenzene

Concen: 7.966 ppbv

RT: 11.944 min Scan# 3

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56

Instrument :

MSVOA_1

ClientSampleId :

ICVVL111825





Tgt Ion:146 Resp: 166137

Ion Ratio Lower Upper

146 100

111 49.2 24.3 73.0

148 63.9 31.6 94.7



#78

Hexachloro-1,3-Butadiene

Concen: 7.222 ppbv

RT: 13.879 min Scan# 4261

Delta R.T. 0.003 min

Lab File: VL043211.D

Acq: 18 Nov 2025 13:56




Tgt Ion:225 Resp: 125707


Ion Ratio Lower Upper

225 100

223 62.7 50.2 75.4

227 63.7 51.1 76.7









#79
Naphthalene
Concen: 8.270 ppbv
RT: 13.510 min Scan# 4146
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56




Instrument :
MSVOA_L
ClientSampleId :
ICVVL111825


Tgt Ion:128 Resp: 273029
Ion Ratio Lower Upper
128 100
127 12.7 9.9 14.9
129 11.0 8.4 12.6



#80
Naphthalene,2-methyl-
Concen: 8.631 ppbv
RT: 14.455 min Scan# 4439
Delta R.T. 0.003 min
Lab File: VL043211.D
Acq: 18 Nov 2025 13:56


Tgt Ion:142 Resp: 74802
Ion Ratio Lower Upper
142 100
141 86.8 67.1 100.7
115 41.7 31.9 47.9







#81
 1,2,4-Trichlorobenzene
 Concen: 8.497 ppbv
 RT: 13.436 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VL043211.D
 Acq: 18 Nov 2025 13:56

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL111825



Tgt Ion:	180	Resp:	125470
Ion Ratio	Lower	Upper	
180	100		
182	96.1	77.4	116.2
184	30.9	24.2	36.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043305.D
 Acq On : 08 Dec 2025 08:44
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Dec 09 00:07:49 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	129333	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	319801	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	246102	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.367 95 185245 10.269 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.700%

Target Compounds

					Qvalue	
2) Dichlorodifluoromethane	1.496	85	194161	10.644	ppbv	98
3) Chlorodifluoromethane	1.473	51	181461	10.115	ppbv	95
4) Chloromethane	1.531	50	69866	10.244	ppbv	96
5) Vinyl Chloride	1.580	62	68051	9.124	ppbv	98
6) Bromomethane	1.667	94	33036	9.137	ppbv	97
7) Chloroethane	1.703	64	28027	9.648	ppbv	93
8) Dichlorotetrafluoroethane	1.554	85	158307	10.433	ppbv	97
9) Propene	1.479	41	71565	9.394	ppbv #	1
10) Heptane	4.969	43	214107	9.444	ppbv	99
11) Trichlorofluoromethane	1.874	101	201211	11.385	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.136	101	154240	11.178	ppbv	97
13) Ethanol	1.719	45	22019	8.220	ppbv	97
14) Bromoethene	1.780	108	55354	10.465	ppbv	96
15) Acetone	1.832	43	154196	8.581	ppbv	99
16) 1,3-Butadiene	1.606	39	76902	9.600	ppbv	97
17) tert-Butyl alcohol	2.036	59	210170	8.911	ppbv	98
18) 1,1-Dichloroethene	2.030	96	67377	10.487	ppbv	94
19) Isopropyl Alcohol	1.884	45	84553	8.327	ppbv #	86
20) Methylene Chloride	2.059	84	60771	11.604	ppbv	96
21) Allyl Chloride	2.091	41	105856	8.756	ppbv	98
22) trans-1,2-Dichloroethene	2.337	96	75565	10.291	ppbv	94
23) Vinyl Acetate	2.457	43	204427	9.424	ppbv #	96
24) 1,1-Dichloroethane	2.405	63	146480	10.334	ppbv	98
25) Ethyl Acetate	2.829	43	373216	9.630	ppbv	99
26) Hexane	2.819	57	168150	9.897	ppbv	98
27) Carbon Disulfide	2.146	76	188702	10.878	ppbv	99
28) Methyl tert-Butyl Ether	2.434	73	84588	10.142	ppbv	96
29) Chloroform	2.842	83	280385	11.610	ppbv	97
30) Cyclohexane	3.845	84	138616	9.731	ppbv	94
31) cis-1,2-Dichloroethene	2.716	61	166393	9.853	ppbv	96
32) 1,1,1-Trichloroethane	3.360	97	280239	10.886	ppbv	99
34) 2-Butanone	2.551	43	234600	10.392	ppbv	100
35) Carbon Tetrachloride	3.752	117	294900	12.149	ppbv	98
36) Benzene	3.648	78	339149	11.074	ppbv	99
37) 1,2-Dichloroethane	3.211	62	209977	11.721	ppbv	98
38) Trichloroethene	4.535	130	135509	11.309	ppbv	98
39) 1,2-Dichloropropane	4.302	63	121360	10.827	ppbv	87
40) 1,4-Dioxane	4.564	88	44522	8.836	ppbv #	98
41) Tetrahydrofuran	3.046	42	120028	9.628	ppbv	98
42) Bromodichloromethane	4.477	83	299904	12.136	ppbv	99
43) Methyl Methacrylate	4.865	69	152174	9.361	ppbv	98
44) 2,2,4-Trimethylpentane	4.629	57	573790	10.909	ppbv	99
45) t-1,3-Dichloropropene	6.402	75	117128	7.729	ppbv	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043305.D
 Acq On : 08 Dec 2025 08:44
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Quant Time: Dec 09 00:07:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.580	75	75161	4.224	ppbv	94
47) 1,1,2-Trichloroethane	6.564	97	135385	11.381	ppbv	95
48) Dibromochloromethane	7.370	129	240386	11.753	ppbv	99
49) Bromoform	9.471	173	217719	12.018	ppbv	98
50) 4-Methyl-2-Pentanone	5.729	43	310023	9.749	ppbv	97
51) 2-Hexanone	7.419	43	219906	8.905	ppbv #	99
52) Tetrachloroethene	8.212	164	123188	10.813	ppbv	98
53) Toluene	6.920	91	368638	10.165	ppbv	100
54) 1,2-Dibromoethane	7.636	107	185272	10.583	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.914	131	231860	12.819	ppbv	97
57) Chlorobenzene	8.911	112	301763	11.943	ppbv	95
58) Ethyl Benzene	9.344	91	527984	11.146	ppbv	95
59) m/p-Xylene	9.542	91	461134	12.143	ppbv #	33
60) o-Xylene	9.953	91	447628	11.799	ppbv	99
61) Styrene	9.859	104	167047	10.061	ppbv	99
62) Isopropylbenzene	10.529	105	797107	11.744	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.950	83	290145	12.020	ppbv	99
64) n-propylbenzene	10.979	120	205431	11.779	ppbv	99
65) tert-Butylbenzene	11.523	119	694372	11.455	ppbv	100
66) Benzyl Chloride	11.607	91	30971	5.035	ppbv	98
67) sec-Butylbenzene	11.759	105	982704	11.615	ppbv	99
69) p-Isopropyltoluene	11.918	119	812059	11.330	ppbv	99
70) n-Butylbenzene	12.274	91	780968	11.231	ppbv	99
71) 2-Chlorotoluene	10.892	91	588943	11.289	ppbv	99
72) 4-Ethyltoluene	11.118	105	539530	11.678	ppbv	99
73) 1,3,5-Trimethylbenzene	11.196	105	445981	11.503	ppbv	97
74) 1,2,4-Trimethylbenzene	11.529	105	499268	11.800	ppbv	91
75) 1,3-Dichlorobenzene	11.597	146	298569	12.092	ppbv	98
76) 1,4-Dichlorobenzene	11.662	146	298168	12.074	ppbv	99
77) 1,2-Dichlorobenzene	11.937	146	286499	12.068	ppbv	99
78) Hexachloro-1,3-Butadiene	13.876	225	220765	11.143	ppbv	99
79) Naphthalene	13.507	128	403341	10.733	ppbv	99
80) Naphthalene,2-methyl-	14.452	142	91763	9.302	ppbv	97
81) 1,2,4-Trichlorobenzene	13.432	180	202078	12.024	ppbv	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 832
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 49 Resp: 129333
Ion Ratio Lower Upper
49 100
128 47.5 23.0 69.0
130 61.0 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 10.644 ppbv
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 85 Resp: 194161
Ion Ratio Lower Upper
85 100
87 30.9 25.4 38.2









#3
Chlorodifluoromethane
Concen: 10.115 ppbv
RT: 1.473 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 51 Resp: 181461
Ion Ratio Lower Upper
51 100
67 20.9 14.9 22.3



#4
Chloromethane
Concen: 10.244 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 50 Resp: 69866
Ion Ratio Lower Upper
50 100
52 32.0 24.1 36.1









#5
Vinyl Chloride
Concen: 9.124 ppbv
RT: 1.580 min Scan# 461
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion: 62 Resp: 68051
Ion Ratio Lower Upper
62 100
64 32.9 27.3 40.9



#6
Bromomethane
Concen: 9.137 ppbv
RT: 1.667 min Scan# 488
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 94 Resp: 33036
Ion Ratio Lower Upper
94 100
96 94.1 72.9 109.3









#7
Chloroethane
Concen: 9.648 ppbv
RT: 1.703 min Scan# 499
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion: 64 Resp: 28027
Ion Ratio Lower Upper
64 100
66 29.5 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 10.433 ppbv
RT: 1.554 min Scan# 453
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 85 Resp: 158307
Ion Ratio Lower Upper
85 100
135 67.8 52.6 79.0









#9
Propene
Concen: 9.394 ppbv
RT: 1.479 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 41 Resp: 71565
Ion Ratio Lower Upper
41 100
42 68.4 3.7 5.5#



#10
Heptane
Concen: 9.444 ppbv
RT: 4.969 min Scan# 1508
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 43 Resp: 214107
Ion Ratio Lower Upper
43 100
57 50.0 39.7 59.5









#11
 Trichlorofluoromethane
 Concen: 11.385 ppbv
 RT: 1.874 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010


Tgt Ion:101 Resp: 201211
 Ion Ratio Lower Upper
 101 100
 103 63.0 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 11.178 ppbv
 RT: 2.136 min Scan# 633
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion:101 Resp: 154240
 Ion Ratio Lower Upper
 101 100
 151 76.6 59.2 88.8









#13
Ethanol
Concen: 8.220 ppbv
RT: 1.719 min Scan# 503
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 45 Resp: 22019
Ion Ratio Lower Upper
45 100
46 39.2 15.0 59.8



#14
Bromoethene
Concen: 10.465 ppbv
RT: 1.780 min Scan# 523
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 108 Resp: 55354
Ion Ratio Lower Upper
108 100
106 106.7 88.9 133.3
79 19.4 16.5 24.7









#15
Acetone
Concen: 8.581 ppbv
RT: 1.832 min Scan# 51
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 43 Resp: 154196
Ion Ratio Lower Upper
43 100
58 26.2 20.7 31.1



#16
1,3-Butadiene
Concen: 9.600 ppbv
RT: 1.606 min Scan# 469
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 39 Resp: 76902
Ion Ratio Lower Upper
39 100
54 71.7 59.4 89.2









#17
tert-Butyl alcohol
Concen: 8.911 ppbv
RT: 2.036 min Scan# 602
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 59 Resp: 210170
Ion Ratio Lower Upper
59 100
57 10.5 9.0 13.4



#18
1,1-Dichloroethene
Concen: 10.487 ppbv
RT: 2.030 min Scan# 600
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 96 Resp: 67377
Ion Ratio Lower Upper
96 100
61 202.7 170.6 255.8
98 62.0 52.2 78.4









#19
Isopropyl Alcohol
Concen: 8.327 ppbv
RT: 1.884 min Scan# 51
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 45 Resp: 84553
Ion Ratio Lower Upper
45 100
58 47.9 31.3 46.9#



#20
Methylene Chloride
Concen: 11.604 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 84 Resp: 60771
Ion Ratio Lower Upper
84 100
49 160.8 132.9 199.3
51 47.0 34.8 52.2
86 64.2 48.2 72.2









#21
 Allyl Chloride
 Concen: 8.756 ppbv
 RT: 2.091 min Scan# 619
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDCCC010


Tgt Ion: 41 Resp: 105856
 Ion Ratio Lower Upper
 41 100
 76 27.6 22.1 33.1
 39 78.1 61.0 91.4



#22
 trans-1,2-Dichloroethene
 Concen: 10.291 ppbv
 RT: 2.337 min Scan# 695
 Delta R.T. 0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion: 96 Resp: 75565
 Ion Ratio Lower Upper
 96 100
 61 163.0 139.0 208.6
 98 61.8 49.4 74.2









#23
 Vinyl Acetate
 Concen: 9.424 ppbv
 RT: 2.457 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDCCC010

Tgt Ion: 43 Resp: 204427




Ion	Ratio	Lower	Upper
43	100		
86	9.5	6.2	9.2#
42	8.6	8.0	12.0
44	6.4	4.0	6.0#




#24
 1,1-Dichloroethane
 Concen: 10.334 ppbv
 RT: 2.405 min Scan# 716
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion: 63 Resp: 146480

Ion	Ratio	Lower	Upper
63	100		
98	5.0	2.8	8.4
100	3.2	1.3	3.9









#25
Ethyl Acetate
Concen: 9.630 ppbv
RT: 2.829 min Scan# 847
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 43 Resp: 373216




Ion	Ratio	Lower	Upper
43	100		
45	9.5	8.1	12.1
61	8.8	7.0	10.4
70	8.2	7.0	10.6




#26
Hexane
Concen: 9.897 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 57 Resp: 168150

Ion	Ratio	Lower	Upper
57	100		
41	95.7	74.6	111.8
43	222.8	179.0	268.4
56	56.4	42.3	63.5









#27
Carbon Disulfide
Concen: 10.878 ppbv
RT: 2.146 min Scan# 611
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion: 76 Resp: 188702
Ion Ratio Lower Upper
76 100
44 20.4 16.1 24.1
78 11.8 9.6 14.4



#28
Methyl tert-Butyl Ether
Concen: 10.142 ppbv
RT: 2.434 min Scan# 725
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 73 Resp: 84588
Ion Ratio Lower Upper
73 100
57 22.7 19.8 29.8









#29
Chloroform
Concen: 11.610 ppbv
RT: 2.842 min Scan# 851
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion: 83 Resp: 280385
Ion Ratio Lower Upper
83 100
85 60.9 50.5 75.7



#30
Cyclohexane
Concen: 9.731 ppbv
RT: 3.845 min Scan# 1161
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 84 Resp: 138616
Ion Ratio Lower Upper
84 100
56 121.7 104.7 157.1
41 84.0 69.0 103.6









#31
 cis-1,2-Dichloroethene
 Concen: 9.853 ppbv
 RT: 2.716 min Scan# 812
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010


Tgt Ion: 61 Resp: 166393
 Ion Ratio Lower Upper
 61 100
 96 64.4 49.3 73.9
 98 42.2 32.1 48.1



#32
 1,1,1-Trichloroethane
 Concen: 10.886 ppbv
 RT: 3.360 min Scan# 1011
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion: 97 Resp: 280239
 Ion Ratio Lower Upper
 97 100
 99 64.0 50.3 75.5
 61 48.7 38.9 58.3





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.949 min Scan# 1193

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Instrument :

MSVOA_L

ClientSampleId :

VSTDCCC010





Tgt Ion:114 Resp: 319801

Ion Ratio Lower Upper

114 100

63 25.6 21.0 31.6

88 18.0 14.9 22.3



#34

2-Butanone

Concen: 10.392 ppbv

RT: 2.551 min Scan# 761

Delta R.T. -0.000 min

Lab File: VL043305.D




Acq: 08 Dec 2025 08:44


Tgt Ion: 43 Resp: 234600

Ion Ratio Lower Upper

43 100

72 22.7 18.1 27.1









#35
Carbon Tetrachloride
Concen: 12.149 ppbv
RT: 3.752 min Scan# 1133
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion:117 Resp: 294900
Ion Ratio Lower Upper
117 100
119 95.5 74.6 112.0
121 30.1 23.2 34.8



#36
Benzene
Concen: 11.074 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 78 Resp: 339149
Ion Ratio Lower Upper
78 100
77 22.7 17.5 26.3
50 19.0 15.5 23.3





#37

1,2-Dichloroethane

Concen: 11.721 ppbv

RT: 3.211 min Scan# 9

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Instrument :

MSVOA_1

ClientSampleId :





VSTDCCC010

Tgt Ion: 62 Resp: 209977

Ion Ratio Lower Upper

62 100

98 6.3 5.5 8.3



#38

Trichloroethene

Concen: 11.309 ppbv

RT: 4.535 min Scan# 1374

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Tgt Ion:130 Resp: 135509




Ion Ratio Lower Upper


130 100

95 106.7 85.7 128.5

132 90.9 73.4 110.0

97 70.3 52.7 79.1





#39

1,2-Dichloropropane

Concen: 10.827 ppbv

RT: 4.302 min Scan# 1

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Instrument :

MSVOA_1

ClientSampleId :

VSTDCCC010





Tgt Ion: 63 Resp: 121360

Ion Ratio Lower Upper

63 100

41 91.0 62.6 93.8

62 72.5 51.0 76.6



#40

1,4-Dioxane

Concen: 8.836 ppbv

RT: 4.564 min Scan# 1383

Delta R.T. -0.003 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Tgt Ion: 88 Resp: 44522




Ion Ratio Lower Upper


88 100

58 80.4 63.0 94.6

43 38.4 32.5 48.7

57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 9.628 ppbv
RT: 3.046 min Scan# 914
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 42 Resp: 120028
Ion Ratio Lower Upper
42 100
72 38.4 30.6 46.0
71 36.1 30.5 45.7



#42
Bromodichloromethane
Concen: 12.136 ppbv
RT: 4.477 min Scan# 1356
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 83 Resp: 299904
Ion Ratio Lower Upper
83 100
85 63.9 51.7 77.5
127 8.0 6.3 9.5









#43
Methyl Methacrylate
Concen: 9.361 ppbv
RT: 4.865 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion: 69 Resp: 152174
Ion Ratio Lower Upper
69 100
41 155.9 126.7 190.1
100 35.6 28.6 43.0



#44
2,2,4-Trimethylpentane
Concen: 10.909 ppbv
RT: 4.629 min Scan# 1403
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 57 Resp: 573790
Ion Ratio Lower Upper
57 100
41 36.1 28.0 42.0
56 32.4 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 7.729 ppbv
RT: 6.402 min Scan# 1951
Delta R.T. 0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion: 75 Resp: 117128
Ion Ratio Lower Upper
75 100
77 30.1 24.3 36.5



#46
cis-1,3-Dichloropropene
Concen: 4.224 ppbv
RT: 5.580 min Scan# 1697
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 75 Resp: 75161
Ion Ratio Lower Upper
75 100
39 67.4 59.1 88.7
77 29.6 24.2 36.2









#47
1,1,2-Trichloroethane
Concen: 11.381 ppbv
RT: 6.564 min Scan# 2001
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010

Tgt Ion: 97 Resp: 135385




Ion	Ratio	Lower	Upper
97	100		
83	84.4	64.6	97.0
85	52.5	39.4	59.2
99	62.6	47.0	70.6




#48
Dibromochloromethane
Concen: 11.753 ppbv
RT: 7.370 min Scan# 2250
Delta R.T. -0.007 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 129 Resp: 240386

Ion	Ratio	Lower	Upper
129	100		
127	77.6	39.3	117.9









#49
Bromoform
Concen: 12.018 ppbv
RT: 9.471 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion:173 Resp: 217719
Ion Ratio Lower Upper
173 100
175 49.1 38.6 58.0
171 52.5 40.6 61.0



#50
4-Methyl-2-Pentanone
Concen: 9.749 ppbv
RT: 5.729 min Scan# 1743
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 43 Resp: 310023
Ion Ratio Lower Upper
43 100
58 34.0 28.4 42.6









#51
2-Hexanone
Concen: 8.905 ppbv
RT: 7.419 min Scan# 21
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 43 Resp: 219906
Ion Ratio Lower Upper
43 100
58 47.6 38.4 57.6
98 1.0 0.0 0.0#



#52
Tetrachloroethene
Concen: 10.813 ppbv
RT: 8.212 min Scan# 2510
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 164 Resp: 123188
Ion Ratio Lower Upper
164 100
166 124.8 99.4 149.2
129 96.7 81.8 122.8
131 97.9 77.4 116.2









#53
Toluene
Concen: 10.165 ppbv
RT: 6.920 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion: 91 Resp: 368638
Ion Ratio Lower Upper
91 100
92 59.4 47.4 71.0



#54
1,2-Dibromoethane
Concen: 10.583 ppbv
RT: 7.636 min Scan# 2332
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 107 Resp: 185272
Ion Ratio Lower Upper
107 100
109 88.4 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 21
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion:117 Resp: 246102
Ion Ratio Lower Upper
117 100
82 71.1 56.1 84.1
119 30.4 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 12.819 ppbv
RT: 8.914 min Scan# 2727
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:131 Resp: 231860
Ion Ratio Lower Upper
131 100
133 94.4 77.4 116.2
119 58.4 49.1 73.7









#57
Chlorobenzene
Concen: 11.943 ppbv
RT: 8.911 min Scan# 211
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion:	112	Resp:	301763
Ion	Ratio	Lower	Upper
112	100		
77	65.2	55.5	83.3
114	30.6	29.1	35.5



#58
Ethyl Benzene
Concen: 11.146 ppbv
RT: 9.344 min Scan# 2860
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:	91	Resp:	527984
Ion	Ratio	Lower	Upper
91	100		
106	31.2	22.8	34.2





#59

m/p-Xylene

Concen: 12.143 ppbv

RT: 9.542 min Scan# 2921

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Instrument :

MSVOA_L

ClientSampleId :





VSTDCCC010

Tgt Ion: 91 Resp: 461134

Ion Ratio Lower Upper

91 100

106 88.8 36.1 54.1#



#60

o-Xylene

Concen: 11.799 ppbv

RT: 9.953 min Scan# 3048

Delta R.T. -0.003 min

Lab File: VL043305.D




Acq: 08 Dec 2025 08:44


Tgt Ion: 91 Resp: 447628

Ion Ratio Lower Upper

91 100

106 42.5 20.9 62.8









#61
Styrene
Concen: 10.061 ppbv
RT: 9.859 min Scan# 3019
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_L
ClientSampleId :
VSTDCCC010


Tgt Ion:104 Resp: 167047
Ion Ratio Lower Upper
104 100
78 51.8 42.1 63.1
103 45.3 37.2 55.8



#62
Isopropylbenzene
Concen: 11.744 ppbv
RT: 10.529 min Scan# 3226
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:105 Resp: 797107
Ion Ratio Lower Upper
105 100
120 25.3 19.8 29.8









#63
1,1,2-Tetrachloroethane
Concen: 12.020 ppbv
RT: 9.950 min Scan# 3048
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion: 83 Resp: 290145
Ion Ratio Lower Upper
83 100
131 9.9 4.8 14.3
85 63.4 31.9 95.9



#64
n-propylbenzene
Concen: 11.779 ppbv
RT: 10.979 min Scan# 3365
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 120 Resp: 205431
Ion Ratio Lower Upper
120 100
91 467.7 377.1 565.7









#65
tert-Butylbenzene
Concen: 11.455 ppbv
RT: 11.523 min Scan# 3533
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion:119 Resp: 694372
Ion Ratio Lower Upper
119 100
91 81.7 65.3 97.9
134 22.0 17.0 25.4



#66
Benzyl Chloride
Concen: 5.035 ppbv
RT: 11.607 min Scan# 3559
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 91 Resp: 30971
Ion Ratio Lower Upper
91 100
126 17.2 14.4 21.6
128 6.0 4.6 6.8









#67
 sec-Butylbenzene
 Concen: 11.615 ppbv
 RT: 11.759 min Scan# 3606
 Delta R.T. -0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDCCC010


Tgt Ion:105 Resp: 982704
 Ion Ratio Lower Upper
 105 100
 134 19.4 15.0 22.6



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.269 ppbv
 RT: 10.367 min Scan# 3176
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion: 95 Resp: 185245
 Ion Ratio Lower Upper
 95 100
 174 68.3 50.2 75.2
 175 5.0 4.1 6.1









#69
p-Isopropyltoluene
Concen: 11.330 ppbv
RT: 11.918 min Scan# 3655
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion:119 Resp: 812059
Ion Ratio Lower Upper
119 100
134 25.9 20.2 30.2
91 29.0 23.8 35.8



#70
n-Butylbenzene
Concen: 11.231 ppbv
RT: 12.274 min Scan# 3765
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion: 91 Resp: 780968
Ion Ratio Lower Upper
91 100
92 55.2 43.5 65.3
134 24.3 19.0 28.6









#71
2-Chlorotoluene
Concen: 11.289 ppbv
RT: 10.892 min Scan# 31
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion: 91 Resp: 588943
Ion Ratio Lower Upper
91 100
126 30.0 11.7 46.7



#72
4-Ethyltoluene
Concen: 11.678 ppbv
RT: 11.118 min Scan# 3408
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:105 Resp: 539530
Ion Ratio Lower Upper
105 100
120 28.8 23.8 35.8
91 12.8 10.3 15.5
77 13.6 11.1 16.7









#73
 1,3,5-Trimethylbenzene
 Concen: 11.503 ppbv
 RT: 11.196 min Scan# 3432
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44




Instrument :
 MSVOA_1
 ClientSampleId :
 VSTDCCC010


Tgt Ion:105 Resp: 445981
 Ion Ratio Lower Upper
 105 100
 120 46.5 38.6 57.8



#74
 1,2,4-Trimethylbenzene
 Concen: 11.800 ppbv
 RT: 11.529 min Scan# 3535
 Delta R.T. -0.000 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Tgt Ion:105 Resp: 499268
 Ion Ratio Lower Upper
 105 100
 120 50.4 41.7 62.5
 91 74.6 70.0 105.0
 77 26.3 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 12.092 ppbv
RT: 11.597 min Scan# 3556
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument : MSVOA_1
ClientSampleId : VSTDCCC010


Tgt Ion:	146	Resp:	298569
Ion Ratio	Lower	Upper	
146	100		
111	44.8	23.4	70.3
148	64.8	31.6	95.0



#76
1,4-Dichlorobenzene
Concen: 12.074 ppbv
RT: 11.662 min Scan# 3576
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:	146	Resp:	298168
Ion Ratio	Lower	Upper	
146	100		
111	44.2	22.8	68.3
148	64.0	31.8	95.4









#77
1,2-Dichlorobenzene
Concen: 12.068 ppbv
RT: 11.937 min Scan# 3661
Delta R.T. -0.003 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44




Instrument :
MSVOA_1
ClientSampleId :
VSTDCCC010


Tgt Ion:146 Resp: 286499
Ion Ratio Lower Upper
146 100
111 47.3 24.3 73.0
148 63.7 31.6 94.7



#78
Hexachloro-1,3-Butadiene
Concen: 11.143 ppbv
RT: 13.876 min Scan# 4260
Delta R.T. -0.000 min
Lab File: VL043305.D
Acq: 08 Dec 2025 08:44

Tgt Ion:225 Resp: 220765
Ion Ratio Lower Upper
225 100
223 63.4 50.2 75.4
227 64.1 51.1 76.7





#79

Naphthalene

Concen: 10.733 ppbv

RT: 13.507 min Scan# 4146

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44

Instrument :

MSVOA_1

ClientSampleId :

VSTDCCC010





Tgt Ion:128 Resp: 403341

Ion Ratio Lower Upper

128 100

127 12.1 9.9 14.9

129 11.2 8.4 12.6



#80

Naphthalene,2-methyl-

Concen: 9.302 ppbv

RT: 14.452 min Scan# 4438

Delta R.T. -0.000 min

Lab File: VL043305.D

Acq: 08 Dec 2025 08:44




Tgt Ion:142 Resp: 91763


Ion Ratio Lower Upper

142 100

141 81.6 67.1 100.7


115 37.9 31.9 47.9







#81
 1,2,4-Trichlorobenzene
 Concen: 12.024 ppbv
 RT: 13.432 min Scan# 41
 Delta R.T. -0.003 min
 Lab File: VL043305.D
 Acq: 08 Dec 2025 08:44

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010



Tgt Ion:180 Resp: 202078
 Ion Ratio Lower Upper
 180 100
 182 97.5 77.4 116.2
 184 30.7 24.2 36.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043307.D
 Acq On : 08 Dec 2025 10:24
 Operator : SY/MD
 Sample : VL1208ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01

Quant Time: Dec 09 00:09:13 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

QLast Update : Wed Nov 19 01:11:25 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.777	49	128373	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	319254	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	250711	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.364 95 185879 10.115 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.100%

Target Compounds

					Qvalue	
2) Dichlorodifluoromethane	1.496	85	195547	10.800	ppbv	100
3) Chlorodifluoromethane	1.473	51	181736	10.206	ppbv	97
4) Chloromethane	1.531	50	70563	10.423	ppbv	100
5) Vinyl Chloride	1.577	62	67921	9.174	ppbv	93
6) Bromomethane	1.664	94	33521	9.340	ppbv	98
7) Chloroethane	1.700	64	28601	9.919	ppbv	91
8) Dichlorotetrafluoroethane	1.551	85	159668	10.601	ppbv	97
9) Propene	1.479	41	71417	9.445	ppbv #	1
10) Heptane	4.962	43	210516	9.355	ppbv	98
11) Trichlorofluoromethane	1.874	101	196801	11.219	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.133	101	153996	11.244	ppbv	97
13) Ethanol	1.716	45	23202	8.726	ppbv	97
14) Bromoethene	1.777	108	55304	10.533	ppbv	94
15) Acetone	1.829	43	152789	8.567	ppbv	99
16) 1,3-Butadiene	1.606	39	76687	9.645	ppbv	97
17) tert-Butyl alcohol	2.033	59	209336	8.942	ppbv	100
18) 1,1-Dichloroethene	2.030	96	65282	10.237	ppbv	93
19) Isopropyl Alcohol	1.881	45	85723	8.506	ppbv #	86
20) Methylene Chloride	2.056	84	60065	11.551	ppbv	95
21) Allyl Chloride	2.091	41	105272	8.773	ppbv	100
22) trans-1,2-Dichloroethene	2.334	96	75422	10.348	ppbv	95
23) Vinyl Acetate	2.454	43	197206	9.159	ppbv #	98
24) 1,1-Dichloroethane	2.402	63	147058	10.453	ppbv	98
25) Ethyl Acetate	2.826	43	372295	9.678	ppbv	99
26) Hexane	2.816	57	166920	9.898	ppbv	98
27) Carbon Disulfide	2.146	76	190369	11.057	ppbv	99
28) Methyl tert-Butyl Ether	2.431	73	77723	9.389	ppbv	98
29) Chloroform	2.839	83	280544	11.703	ppbv	98
30) Cyclohexane	3.842	84	136887	9.681	ppbv	95
31) cis-1,2-Dichloroethene	2.713	61	166197	9.915	ppbv	94
32) 1,1,1-Trichloroethane	3.357	97	278163	10.886	ppbv	99
34) 2-Butanone	2.548	43	232285	10.307	ppbv	99
35) Carbon Tetrachloride	3.752	117	293430	12.109	ppbv	100
36) Benzene	3.645	78	342251	11.194	ppbv	99
37) 1,2-Dichloroethane	3.208	62	212457	11.880	ppbv	98
38) Trichloroethene	4.532	130	137474	11.493	ppbv	98
39) 1,2-Dichloropropane	4.295	63	122348	10.934	ppbv	92
40) 1,4-Dioxane	4.564	88	43982	8.743	ppbv #	97
41) Tetrahydrofuran	3.043	42	121301	9.747	ppbv	98
42) Bromodichloromethane	4.473	83	296233	12.008	ppbv	97
43) Methyl Methacrylate	4.862	69	156667	9.654	ppbv	95
44) 2,2,4-Trimethylpentane	4.625	57	569168	10.839	ppbv	98
45) t-1,3-Dichloropropene	6.393	75	118391	7.826	ppbv	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043307.D
 Acq On : 08 Dec 2025 10:24
 Operator : SY/MD
 Sample : VL1208ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01


Quant Time: Dec 09 00:09:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.580	75	146637	8.255	ppbv	94
47) 1,1,2-Trichloroethane	6.561	97	137247	11.557	ppbv	98
48) Dibromochloromethane	7.370	129	240784	11.793	ppbv	99
49) Bromoform	9.471	173	220417	12.188	ppbv	99
50) 4-Methyl-2-Pentanone	5.723	43	315888	9.951	ppbv	98
51) 2-Hexanone	7.415	43	223708	9.075	ppbv #	98
52) Tetrachloroethene	8.208	164	125681	11.051	ppbv	96
53) Toluene	6.914	91	374608	10.347	ppbv	99
54) 1,2-Dibromoethane	7.636	107	183811	10.518	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.911	131	229368	12.449	ppbv	99
57) Chlorobenzene	8.908	112	305509	11.869	ppbv	99
58) Ethyl Benzene	9.341	91	538550	11.161	ppbv	97
59) m/p-Xylene	9.539	91	420640	10.873	ppbv #	20
60) o-Xylene	9.950	91	446789	11.560	ppbv	98
61) Styrene	9.859	104	167877	9.925	ppbv	99
62) Isopropylbenzene	10.529	105	805674	11.652	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.950	83	285542	11.611	ppbv	99
64) n-propylbenzene	10.979	120	208180	11.718	ppbv	96
65) tert-Butylbenzene	11.523	119	696094	11.273	ppbv	100
66) Benzyl Chloride	11.604	91	31638	5.049	ppbv	98
67) sec-Butylbenzene	11.759	105	973259	11.291	ppbv	99
69) p-Isopropyltoluene	11.914	119	808286	11.070	ppbv	99
70) n-Butylbenzene	12.270	91	780553	11.019	ppbv	100
71) 2-Chlorotoluene	10.888	91	584474	10.997	ppbv	99
72) 4-Ethyltoluene	11.115	105	540638	11.487	ppbv	99
73) 1,3,5-Trimethylbenzene	11.193	105	451214	11.424	ppbv	96
74) 1,2,4-Trimethylbenzene	11.526	105	496984	11.530	ppbv	100
75) 1,3-Dichlorobenzene	11.597	146	297680	11.834	ppbv	98
76) 1,4-Dichlorobenzene	11.662	146	298352	11.859	ppbv	98
77) 1,2-Dichlorobenzene	11.937	146	285835	11.819	ppbv	98
78) Hexachloro-1,3-Butadiene	13.873	225	220056	10.903	ppbv	99
79) Naphthalene	13.504	128	409312	10.692	ppbv	97
80) Naphthalene,2-methyl-	14.449	142	92975	9.252	ppbv	99
81) 1,2,4-Trichlorobenzene	13.432	180	204005	11.915	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01

Quant Time: Dec 09 00:09:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.777 min Scan# 81
Delta R.T. -0.004 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01


Tgt Ion: 49 Resp: 128373
Ion Ratio Lower Upper
49 100
128 48.0 23.0 69.0
130 62.5 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 10.800 ppbv
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 85 Resp: 195547
Ion Ratio Lower Upper
85 100
87 32.0 25.4 38.2









#3
Chlorodifluoromethane
Concen: 10.206 ppbv
RT: 1.473 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 51 Resp: 181736
Ion Ratio Lower Upper
51 100
67 19.9 14.9 22.3



#4
Chloromethane
Concen: 10.423 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 50 Resp: 70563
Ion Ratio Lower Upper
50 100
52 30.2 24.1 36.1









#5
 Vinyl Chloride
 Concen: 9.174 ppbv
 RT: 1.577 min Scan# 460
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24




Instrument :
 MSVOA_1
 ClientSampleId :
 VL1208ABS01


Tgt Ion: 62 Resp: 67921
 Ion Ratio Lower Upper
 62 100
 64 30.2 27.3 40.9



#6
 Bromomethane
 Concen: 9.340 ppbv
 RT: 1.664 min Scan# 487
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion: 94 Resp: 33521
 Ion Ratio Lower Upper
 94 100
 96 93.0 72.9 109.3









#7
Chloroethane
Concen: 9.919 ppbv
RT: 1.700 min Scan# 498
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 64 Resp: 28601
Ion Ratio Lower Upper
64 100
66 28.6 26.8 40.2



#8
Dichlorotetrafluoroethane
Concen: 10.601 ppbv
RT: 1.551 min Scan# 452
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 85 Resp: 159668
Ion Ratio Lower Upper
85 100
135 68.3 52.6 79.0









#9
 Propene
 Concen: 9.445 ppbv
 RT: 1.479 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24




Instrument :
 MSVOA_1
 ClientSampleId :
 VL1208ABS01


Tgt Ion: 41 Resp: 71417
 Ion Ratio Lower Upper
 41 100
 42 67.1 3.7 5.5#



#10
 Heptane
 Concen: 9.355 ppbv
 RT: 4.962 min Scan# 1506
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion: 43 Resp: 210516
 Ion Ratio Lower Upper
 43 100
 57 50.9 39.7 59.5









#11
 Trichlorofluoromethane
 Concen: 11.219 ppbv
 RT: 1.874 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24




Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01


Tgt Ion:101 Resp: 196801
 Ion Ratio Lower Upper
 101 100
 103 65.2 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 11.244 ppbv
 RT: 2.133 min Scan# 632
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion:101 Resp: 153996
 Ion Ratio Lower Upper
 101 100
 151 76.4 59.2 88.8









#13
Ethanol
Concen: 8.726 ppbv
RT: 1.716 min Scan# 503
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 45 Resp: 23202
Ion Ratio Lower Upper
45 100
46 35.5 15.0 59.8



#14
Bromoethene
Concen: 10.533 ppbv
RT: 1.777 min Scan# 522
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 108 Resp: 55304
Ion Ratio Lower Upper
108 100
106 104.9 88.9 133.3
79 17.7 16.5 24.7









#15
Acetone
Concen: 8.567 ppbv
RT: 1.829 min Scan# 51
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 43 Resp: 152789
Ion Ratio Lower Upper
43 100
58 26.6 20.7 31.1



#16
1,3-Butadiene
Concen: 9.645 ppbv
RT: 1.606 min Scan# 469
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 39 Resp: 76687
Ion Ratio Lower Upper
39 100
54 71.8 59.4 89.2









#17
tert-Butyl alcohol
Concen: 8.942 ppbv
RT: 2.033 min Scan# 60
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 59 Resp: 209336
Ion Ratio Lower Upper
59 100
57 11.2 9.0 13.4



#18
1,1-Dichloroethene
Concen: 10.237 ppbv
RT: 2.030 min Scan# 600
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 96 Resp: 65282
Ion Ratio Lower Upper
96 100
61 200.9 170.6 255.8
98 62.7 52.2 78.4









#19
Isopropyl Alcohol
Concen: 8.506 ppbv
RT: 1.881 min Scan# 51
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 45 Resp: 85723
Ion Ratio Lower Upper
45 100
58 47.5 31.3 46.9#



#20
Methylene Chloride
Concen: 11.551 ppbv
RT: 2.056 min Scan# 608
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 84 Resp: 60065
Ion Ratio Lower Upper
84 100
49 158.6 132.9 199.3
51 48.9 34.8 52.2
86 60.8 48.2 72.2









#21
 Allyl Chloride
 Concen: 8.773 ppbv
 RT: 2.091 min Scan# 619
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24




Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01


Tgt Ion: 41 Resp: 105272
 Ion Ratio Lower Upper
 41 100
 76 27.6 22.1 33.1
 39 76.5 61.0 91.4



#22
 trans-1,2-Dichloroethene
 Concen: 10.348 ppbv
 RT: 2.334 min Scan# 694
 Delta R.T. -0.000 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion: 96 Resp: 75422
 Ion Ratio Lower Upper
 96 100
 61 165.0 139.0 208.6
 98 63.3 49.4 74.2









#23
 Vinyl Acetate
 Concen: 9.159 ppbv
 RT: 2.454 min Scan# 715
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Instrument :
 MSVOA_1
 ClientSampleId :
 VL1208ABS01

Tgt Ion: 43 Resp: 197206




Ion	Ratio	Lower	Upper
43	100		
86	8.3	6.2	9.2
42	10.1	8.0	12.0
44	6.8	4.0	6.0




#24
 1,1-Dichloroethane
 Concen: 10.453 ppbv
 RT: 2.402 min Scan# 715
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion: 63 Resp: 147058

Ion	Ratio	Lower	Upper
63	100		
98	4.8	2.8	8.4
100	3.1	1.3	3.9





#25

Ethyl Acetate

Concen: 9.678 ppbv

RT: 2.826 min Scan# 84

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_1

ClientSampleId :

VL1208ABS01

Tgt Ion: 43 Resp: 372295





Ion Ratio Lower Upper

43 100

45 9.8 8.1 12.1

61 8.6 7.0 10.4

70 8.0 7.0 10.6



#26

Hexane

Concen: 9.898 ppbv

RT: 2.816 min Scan# 843

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Tgt Ion: 57 Resp: 166920




Ion Ratio Lower Upper


57 100

41 96.6 74.6 111.8

43 224.2 179.0 268.4

56 56.9 42.3 63.5









#27
Carbon Disulfide
Concen: 11.057 ppbv
RT: 2.146 min Scan# 611
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 76 Resp: 190369
Ion Ratio Lower Upper
76 100
44 19.1 16.1 24.1
78 11.8 9.6 14.4



#28
Methyl tert-Butyl Ether
Concen: 9.389 ppbv
RT: 2.431 min Scan# 724
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 73 Resp: 77723
Ion Ratio Lower Upper
73 100
57 23.8 19.8 29.8





#29

Chloroform

Concen: 11.703 ppbv

RT: 2.839 min Scan# 81

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :





VL1208ABS01

Tgt Ion: 83 Resp: 280544

Ion Ratio Lower Upper

83 100

85 64.6 50.5 75.7



#30

Cyclohexane

Concen: 9.681 ppbv

RT: 3.842 min Scan# 1160

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24




Tgt Ion: 84 Resp: 136887


Ion Ratio Lower Upper

84 100

56 123.4 104.7 157.1

41 84.6 69.0 103.6









#31
cis-1,2-Dichloroethene
Concen: 9.915 ppbv
RT: 2.713 min Scan# 81
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01


Tgt Ion: 61 Resp: 166197
Ion Ratio Lower Upper
61 100
96 67.8 49.3 73.9
98 42.0 32.1 48.1



#32
1,1,1-Trichloroethane
Concen: 10.886 ppbv
RT: 3.357 min Scan# 1010
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 97 Resp: 278163
Ion Ratio Lower Upper
97 100
99 64.3 50.3 75.5
61 49.0 38.9 58.3









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.946 min Scan# 1192
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion:114 Resp: 319254
Ion Ratio Lower Upper
114 100
63 25.8 21.0 31.6
88 17.9 14.9 22.3



#34
2-Butanone
Concen: 10.307 ppbv
RT: 2.548 min Scan# 760
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 43 Resp: 232285
Ion Ratio Lower Upper
43 100
72 22.3 18.1 27.1









#35
Carbon Tetrachloride
Concen: 12.109 ppbv
RT: 3.752 min Scan# 1133
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01


Tgt Ion:117 Resp: 293430
Ion Ratio Lower Upper
117 100
119 92.9 74.6 112.0
121 29.0 23.2 34.8



#36
Benzene
Concen: 11.194 ppbv
RT: 3.645 min Scan# 1099
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 78 Resp: 342251
Ion Ratio Lower Upper
78 100
77 21.3 17.5 26.3
50 19.1 15.5 23.3





#37

1,2-Dichloroethane

Concen: 11.880 ppbv

RT: 3.208 min Scan# 9

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_1

ClientSampleId :





VL1208ABS01

Tgt Ion: 62 Resp: 212457

Ion Ratio Lower Upper

62 100

98 6.2 5.5 8.3



#38

Trichloroethene

Concen: 11.493 ppbv

RT: 4.532 min Scan# 1373

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Tgt Ion:130 Resp: 137474




Ion Ratio Lower Upper


130 100

95 104.4 85.7 128.5

132 88.8 73.4 110.0

97 65.2 52.7 79.1





#39

1,2-Dichloropropane

Concen: 10.934 ppbv

RT: 4.295 min Scan# 11

Delta R.T. -0.007 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :

VL1208ABS01





Tgt Ion: 63 Resp: 122348

Ion Ratio Lower Upper

63 100

41 86.4 62.6 93.8

62 68.2 51.0 76.6



#40

1,4-Dioxane

Concen: 8.743 ppbv

RT: 4.564 min Scan# 1383

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Tgt Ion: 88 Resp: 43982




Ion Ratio Lower Upper


88 100

58 76.2 63.0 94.6

43 37.8 32.5 48.7

57 0.0 0.0 0.0









#41
Tetrahydrofuran
Concen: 9.747 ppbv
RT: 3.043 min Scan# 914
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01


Tgt Ion: 42 Resp: 121301
Ion Ratio Lower Upper
42 100
72 38.9 30.6 46.0
71 36.7 30.5 45.7



#42
Bromodichloromethane
Concen: 12.008 ppbv
RT: 4.473 min Scan# 1355
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 83 Resp: 296233
Ion Ratio Lower Upper
83 100
85 62.5 51.7 77.5
127 8.5 6.3 9.5









#43
Methyl Methacrylate
Concen: 9.654 ppbv
RT: 4.862 min Scan# 1476
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 69 Resp: 156667
Ion Ratio Lower Upper
69 100
41 150.8 126.7 190.1
100 35.4 28.6 43.0



#44
2,2,4-Trimethylpentane
Concen: 10.839 ppbv
RT: 4.625 min Scan# 1402
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 57 Resp: 569168
Ion Ratio Lower Upper
57 100
41 36.5 28.0 42.0
56 32.5 25.6 38.4









#45
t-1,3-Dichloropropene
Concen: 7.826 ppbv
RT: 6.393 min Scan# 1948
Delta R.T. -0.007 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01


Tgt Ion: 75 Resp: 118391
Ion Ratio Lower Upper
75 100
77 29.3 24.3 36.5



#46
cis-1,3-Dichloropropene
Concen: 8.255 ppbv
RT: 5.580 min Scan# 1697
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 75 Resp: 146637
Ion Ratio Lower Upper
75 100
39 69.0 59.1 88.7
77 27.6 24.2 36.2









#47
1,1,2-Trichloroethane
Concen: 11.557 ppbv
RT: 6.561 min Scan# 2000
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01


Tgt Ion:	97	Resp:	137247
Ion	Ratio	Lower	Upper
97	100		
83	81.7	64.6	97.0
85	51.0	39.4	59.2
99	61.1	47.0	70.6



#48
Dibromochloromethane
Concen: 11.793 ppbv
RT: 7.370 min Scan# 2250
Delta R.T. -0.007 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:	129	Resp:	240784
Ion	Ratio	Lower	Upper
129	100		
127	79.2	39.3	117.9





#49

Bromoform

Concen: 12.188 ppbv

RT: 9.471 min Scan# 21

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :

VL1208ABS01





Tgt Ion:173 Resp: 220417

Ion Ratio Lower Upper

173 100

175 48.8 38.6 58.0

171 50.4 40.6 61.0



#50

4-Methyl-2-Pentanone

Concen: 9.951 ppbv

RT: 5.723 min Scan# 1741

Delta R.T. -0.010 min

Lab File: VL043307.D




Acq: 08 Dec 2025 10:24


Tgt Ion: 43 Resp: 315888

Ion Ratio Lower Upper

43 100

58 34.5 28.4 42.6









#51
2-Hexanone
Concen: 9.075 ppbv
RT: 7.415 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01


Tgt Ion: 43 Resp: 223708
Ion Ratio Lower Upper
43 100
58 46.6 38.4 57.6
98 0.9 0.0 0.0#



#52
Tetrachloroethene
Concen: 11.051 ppbv
RT: 8.208 min Scan# 2509
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 164 Resp: 125681
Ion Ratio Lower Upper
164 100
166 125.0 99.4 149.2
129 95.8 81.8 122.8
131 91.5 77.4 116.2





#53

Toluene

Concen: 10.347 ppbv

RT: 6.914 min Scan# 2109

Delta R.T. -0.007 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :





VL1208ABS01

Tgt Ion: 91 Resp: 374608

Ion Ratio Lower Upper

91 100

92 58.7 47.4 71.0



#54

1,2-Dibromoethane

Concen: 10.518 ppbv

RT: 7.636 min Scan# 2332

Delta R.T. -0.003 min

Lab File: VL043307.D




Acq: 08 Dec 2025 10:24


Tgt Ion: 107 Resp: 183811

Ion Ratio Lower Upper

107 100

109 94.4 72.1 108.1









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.869 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion:117 Resp: 250711
Ion Ratio Lower Upper
117 100
82 69.4 56.1 84.1
119 32.6 25.0 37.6



#56
1,1,1,2-Tetrachloroethane
Concen: 12.449 ppbv
RT: 8.911 min Scan# 2726
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:131 Resp: 229368
Ion Ratio Lower Upper
131 100
133 96.8 77.4 116.2
119 60.5 49.1 73.7









#57
Chlorobenzene
Concen: 11.869 ppbv
RT: 8.908 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion	Ratio	Lower	Upper
112	100		
77	69.6	55.5	83.3
114	30.5	29.1	35.5



#58
Ethyl Benzene
Concen: 11.161 ppbv
RT: 9.341 min Scan# 2859
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion	Ratio	Lower	Upper
91	100		
106	30.1	22.8	34.2





#59

m/p-Xylene

Concen: 10.873 ppbv

RT: 9.539 min Scan# 2921

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_1

ClientSampleId :





VL1208ABS01

Tgt Ion: 91 Resp: 420640

Ion Ratio Lower Upper

91 100

106 97.4 36.1 54.1#



#60

o-Xylene

Concen: 11.560 ppbv

RT: 9.950 min Scan# 3047

Delta R.T. -0.007 min

Lab File: VL043307.D




Acq: 08 Dec 2025 10:24


Tgt Ion: 91 Resp: 446789

Ion Ratio Lower Upper

91 100

106 43.1 20.9 62.8









#61
Styrene
Concen: 9.925 ppbv
RT: 9.859 min Scan# 3020
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_L
ClientSampleId :
VL1208ABS01


Tgt Ion:104 Resp: 167877
Ion Ratio Lower Upper
104 100
78 51.5 42.1 63.1
103 47.4 37.2 55.8



#62
Isopropylbenzene
Concen: 11.652 ppbv
RT: 10.529 min Scan# 3226
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:105 Resp: 805674
Ion Ratio Lower Upper
105 100
120 25.6 19.8 29.8





#63

1,1,2,2-Tetrachloroethane

Concen: 11.611 ppbv

RT: 9.950 min Scan# 3048

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :

VL1208ABS01





Tgt Ion: 83 Resp: 285542

Ion Ratio Lower Upper

83 100

131 9.6 4.8 14.3

85 64.5 31.9 95.9



#64

n-propylbenzene

Concen: 11.718 ppbv

RT: 10.979 min Scan# 3365

Delta R.T. -0.003 min

Lab File: VL043307.D




Acq: 08 Dec 2025 10:24


Tgt Ion: 120 Resp: 208180

Ion Ratio Lower Upper

120 100

91 461.2 377.1 565.7









#65
tert-Butylbenzene
Concen: 11.273 ppbv
RT: 11.523 min Scan# 3533
Delta R.T. -0.000 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion:119 Resp: 696094
Ion Ratio Lower Upper
119 100
91 81.5 65.3 97.9
134 21.7 17.0 25.4



#66
Benzyl Chloride
Concen: 5.049 ppbv
RT: 11.604 min Scan# 3558
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 91 Resp: 31638
Ion Ratio Lower Upper
91 100
126 16.9 14.4 21.6
128 6.4 4.6 6.8









#67
sec-Butylbenzene
Concen: 11.291 ppbv
RT: 11.759 min Scan# 3175
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion:105 Resp: 973259
Ion Ratio Lower Upper
105 100
134 19.4 15.0 22.6



#68
1-Bromo-4-Fluorobenzene
Concen: 10.115 ppbv
RT: 10.364 min Scan# 3175
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion: 95 Resp: 185879
Ion Ratio Lower Upper
95 100
174 69.2 50.2 75.2
175 5.1 4.1 6.1









#69
 p-Isopropyltoluene
 Concen: 11.070 ppbv
 RT: 11.914 min Scan# 3654
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24




Instrument :
 MSVOA_1
 ClientSampleId :
 VL1208ABS01


Tgt Ion:119 Resp: 808286
 Ion Ratio Lower Upper
 119 100
 134 25.9 20.2 30.2
 91 29.1 23.8 35.8



#70
 n-Butylbenzene
 Concen: 11.019 ppbv
 RT: 12.270 min Scan# 3764
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Tgt Ion: 91 Resp: 780553
 Ion Ratio Lower Upper
 91 100
 92 54.6 43.5 65.3
 134 24.1 19.0 28.6









#71
2-Chlorotoluene
Concen: 10.997 ppbv
RT: 10.888 min Scan# 31
Delta R.T. -0.007 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion: 91 Resp: 584474
Ion Ratio Lower Upper
91 100
126 30.0 11.7 46.7



#72
4-Ethyltoluene
Concen: 11.487 ppbv
RT: 11.115 min Scan# 3407
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:105 Resp: 540638
Ion Ratio Lower Upper
105 100
120 28.8 23.8 35.8
91 12.9 10.3 15.5
77 13.7 11.1 16.7





#73

1,3,5-Trimethylbenzene

Concen: 11.424 ppbv

RT: 11.193 min Scan# 3432

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_L

ClientSampleId :





VL1208ABS01

Tgt Ion:105 Resp: 451214

Ion Ratio Lower Upper

105 100

120 45.6 38.6 57.8



#74

1,2,4-Trimethylbenzene

Concen: 11.530 ppbv

RT: 11.526 min Scan# 3534

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Tgt Ion:105 Resp: 496984




Ion Ratio Lower Upper


105 100

120 52.1 41.7 62.5

91 88.0 70.0 105.0

77 29.7 23.8 35.8









#75
1,3-Dichlorobenzene
Concen: 11.834 ppbv
RT: 11.597 min Scan# 3556
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion:146 Resp: 297680
Ion Ratio Lower Upper
146 100
111 46.0 23.4 70.3
148 64.9 31.6 95.0



#76
1,4-Dichlorobenzene
Concen: 11.859 ppbv
RT: 11.662 min Scan# 3576
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24

Tgt Ion:146 Resp: 298352
Ion Ratio Lower Upper
146 100
111 43.2 22.8 68.3
148 64.2 31.8 95.4





#77

1,2-Dichlorobenzene

Concen: 11.819 ppbv

RT: 11.937 min Scan# 3

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24

Instrument :

MSVOA_1

ClientSampleId :

VL1208ABS01





Tgt Ion:146 Resp: 285835

Ion Ratio Lower Upper

146 100

111 46.1 24.3 73.0

148 63.9 31.6 94.7



#78

Hexachloro-1,3-Butadiene

Concen: 10.903 ppbv

RT: 13.873 min Scan# 4259

Delta R.T. -0.003 min

Lab File: VL043307.D

Acq: 08 Dec 2025 10:24




Tgt Ion:225 Resp: 220056


Ion Ratio Lower Upper

225 100

223 63.6 50.2 75.4

227 64.7 51.1 76.7









#79
Naphthalene
Concen: 10.692 ppbv
RT: 13.504 min Scan# 4146
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24




Instrument :
MSVOA_1
ClientSampleId :
VL1208ABS01


Tgt Ion:128 Resp: 409312
Ion Ratio Lower Upper
128 100
127 13.5 9.9 14.9
129 11.4 8.4 12.6



#80
Naphthalene,2-methyl-
Concen: 9.252 ppbv
RT: 14.449 min Scan# 4437
Delta R.T. -0.003 min
Lab File: VL043307.D
Acq: 08 Dec 2025 10:24


Tgt Ion:142 Resp: 92975
Ion Ratio Lower Upper
142 100
141 84.6 67.1 100.7
115 39.8 31.9 47.9







#81
 1,2,4-Trichlorobenzene
 Concen: 11.915 ppbv
 RT: 13.432 min Scan# 41
 Delta R.T. -0.003 min
 Lab File: VL043307.D
 Acq: 08 Dec 2025 10:24

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1208ABS01



Tgt Ion:	180	Resp:	204005
Ion	Ratio	Lower	Upper
180	100		
182	94.3	77.4	116.2
184	30.3	24.2	36.4



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043310.D
 Acq On : 08 Dec 2025 12:40
 Operator : SY/MD
 Sample : Q3799-01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1

Quant Time: Dec 09 00:10:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound		R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards							
1) Bromochloromethane		2.781	49	151890	10.000	ppbv	0.00
33) 1,4-Difluorobenzene		3.946	114	389995	10.000	ppbv	0.00
55) Chlorobenzene-d5		8.869	117	311445	10.000	ppbv	0.00
System Monitoring Compounds							
68) 1-Bromo-4-Fluorobenzene		10.364	95	233482	10.227	ppbv	0.00
Spiked Amount		10.000	Range 65 - 135		Recovery	= 102.300%	
Target Compounds							
							Qvalue
3) Chlorodifluoromethane		1.473	51	9263	0.440	ppbv	# 83
4) Chloromethane		1.531	50	17448	2.178	ppbv	95
7) Chloroethane		1.703	64	142	0.042	ppbv	# 41
9) Propene		1.483	41	5806	0.649	ppbv	# 1
11) Trichlorofluoromethane		1.874	101	4433	0.214	ppbv	95
12) 1,1,2-Trichlorotrifluo...		2.136	101	1181	0.073	ppbv	100
13) Ethanol		1.719	45	11664	3.708	ppbv	# 38
15) Acetone		1.832	43	100928	4.783	ppbv	98
16) 1,3-Butadiene		1.599	39	3163	0.336	ppbv	# 40
17) tert-Butyl alcohol		2.039	59	2939	0.106	ppbv	94
19) Isopropyl Alcohol		1.884	45	9295	0.779	ppbv	# 1
20) Methylene Chloride		2.059	84	18675	2.407	ppbv	89
23) Vinyl Acetate		2.447	43	12208	0.479	ppbv	# 87
25) Ethyl Acetate		2.819	43	19515	0.429	ppbv	# 82
26) Hexane		2.819	57	24551	1.230	ppbv	# 44
29) Chloroform		2.845	83	1625	0.057	ppbv	85
34) 2-Butanone		2.551	43	13762	0.500	ppbv	99
35) Carbon Tetrachloride		3.755	117	1879	0.063	ppbv	100
36) Benzene		3.651	78	3478	0.093	ppbv	96
39) 1,2-Dichloropropane		3.946	63	100423	7.346	ppbv	# 24
41) Tetrahydrofuran		3.052	42	902	0.059	ppbv	# 67
53) Toluene		6.910	91	3785	0.086	ppbv	# 66


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043310.D
Acq On : 08 Dec 2025 12:40
Operator : SY/MD
Sample : Q3799-01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AA-1

Quant Time: Dec 09 00:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 832
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1


Tgt Ion: 49 Resp: 151890
Ion Ratio Lower Upper
49 100
128 48.5 23.0 69.0
130 62.9 29.9 89.7



#3
Chlorodifluoromethane
Concen: 0.440 ppbv
RT: 1.473 min Scan# 428
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion: 51 Resp: 9263
Ion Ratio Lower Upper
51 100
67 11.1 14.9 22.3#









#4
Chloromethane
Concen: 2.178 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_1
ClientSampleId :
AA-1


Tgt Ion: 50 Resp: 17448
Ion Ratio Lower Upper
50 100
52 33.1 24.1 36.1



#7
Chloroethane
Concen: 0.042 ppbv
RT: 1.703 min Scan# 499
Delta R.T. 0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion: 64 Resp: 142
Ion Ratio Lower Upper
64 100
66 0.0 26.8 40.2#





#9

Propene

Concen: 0.649 ppbv

RT: 1.483 min Scan# 41

Delta R.T. 0.003 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40

Instrument :

MSVOA_L

ClientSampleId :





AA-1

Tgt Ion: 41 Resp: 5806

Ion Ratio Lower Upper

41 100

42 64.4 3.7 5.5#



#11

Trichlorofluoromethane

Concen: 0.214 ppbv

RT: 1.874 min Scan# 552

Delta R.T. -0.000 min

Lab File: VL043310.D




Acq: 08 Dec 2025 12:40


Tgt Ion: 101 Resp: 4433

Ion Ratio Lower Upper

101 100

103 70.0 52.6 79.0









#12
1,1,2-Trichlorotrifluoroethane
Concen: 0.073 ppbv
RT: 2.136 min Scan# 611
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1


Tgt Ion:101 Resp: 1181
Ion Ratio Lower Upper
101 100
151 73.8 59.2 88.8



#13
Ethanol
Concen: 3.708 ppbv
RT: 1.719 min Scan# 504
Delta R.T. 0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion: 45 Resp: 11664
Ion Ratio Lower Upper
45 100
46 0.0 15.0 59.8#









#15
Acetone
Concen: 4.783 ppbv
RT: 1.832 min Scan# 51
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1


Tgt Ion: 43 Resp: 100928
Ion Ratio Lower Upper
43 100
58 26.8 20.7 31.1



#16
1,3-Butadiene
Concen: 0.336 ppbv
RT: 1.599 min Scan# 467
Delta R.T. -0.007 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion: 39 Resp: 3163
Ion Ratio Lower Upper
39 100
54 23.6 59.4 89.2#





#17

tert-Butyl alcohol

Concen: 0.106 ppbv

RT: 2.039 min Scan# 602

Delta R.T. 0.003 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40

Instrument :

MSVOA_1

ClientSampleId :





AA-1

Tgt Ion: 59 Resp: 2939

Ion Ratio Lower Upper

59 100

57 13.4 9.0 13.4



#19

Isopropyl Alcohol

Concen: 0.779 ppbv

RT: 1.884 min Scan# 555

Delta R.T. 0.003 min

Lab File: VL043310.D




Acq: 08 Dec 2025 12:40


Tgt Ion: 45 Resp: 9295

Ion Ratio Lower Upper

45 100

58 291.0 31.3 46.9#









#20
Methylene Chloride
Concen: 2.407 ppbv
RT: 2.059 min Scan# 608
Delta R.T. 0.003 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Instrument :
MSVOA_L
ClientSampleId :
AA-1

Tgt Ion: 84 Resp: 18675




Ion	Ratio	Lower	Upper
84	100		
49	146.1	132.9	199.3
51	39.7	34.8	52.2
86	58.4	48.2	72.2




#23
Vinyl Acetate
Concen: 0.479 ppbv
RT: 2.447 min Scan# 729
Delta R.T. -0.010 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion: 43 Resp: 12208

Ion	Ratio	Lower	Upper
43	100		
86	11.2	6.2	9.2#
42	5.1	8.0	12.0#
44	10.1	4.0	6.0#









#25
Ethyl Acetate
Concen: 0.429 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.010 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1


Tgt Ion:	43	Resp:	19515
Ion	Ratio	Lower	Upper
43	100		
45	3.6	8.1	12.1#
61	1.5	7.0	10.4#
70	2.4	7.0	10.6#



#26
Hexane
Concen: 1.230 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion:	57	Resp:	24551
Ion	Ratio	Lower	Upper
57	100		
41	101.6	74.6	111.8
43	83.3	179.0	268.4#
56	62.7	42.3	63.5





#29

Chloroform

Concen: 0.057 ppbv

RT: 2.845 min Scan# 81

Delta R.T. 0.003 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40

Instrument :

MSVOA_L

ClientSampleId :





AA-1

Tgt Ion: 83 Resp: 1625

Ion Ratio Lower Upper

83 100

85 51.5 50.5 75.7



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.946 min Scan# 1192

Delta R.T. -0.003 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40




Tgt Ion: 114 Resp: 389995


Ion Ratio Lower Upper

114 100

63 25.9 21.0 31.6

88 18.0 14.9 22.3









#34
2-Butanone
Concen: 0.500 ppbv
RT: 2.551 min Scan# 761
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_L
ClientSampleId :
AA-1


Tgt Ion: 43 Resp: 13762
Ion Ratio Lower Upper
43 100
72 21.9 18.1 27.1



#35
Carbon Tetrachloride
Concen: 0.063 ppbv
RT: 3.755 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion: 117 Resp: 1879
Ion Ratio Lower Upper
117 100
119 93.5 74.6 112.0
121 28.5 23.2 34.8





#36

Benzene

Concen: 0.093 ppbv

RT: 3.651 min Scan# 1101

Delta R.T. 0.003 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40

Instrument :

MSVOA_L

ClientSampleId :

AA-1





Tgt Ion: 78 Resp: 3478

Ion Ratio Lower Upper

78 100

77 21.0 17.5 26.3

50 16.0 15.5 23.3



#39

1,2-Dichloropropane

Concen: 7.346 ppbv

RT: 3.946 min Scan# 1192

Delta R.T. -0.356 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40




Tgt Ion: 63 Resp: 100423


Ion Ratio Lower Upper

63 100

41 0.0 62.6 93.8#

62 16.8 51.0 76.6#









#41
Tetrahydrofuran
Concen: 0.059 ppbv
RT: 3.052 min Scan# 914
Delta R.T. 0.006 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40




Instrument :
MSVOA_1
ClientSampleId :
AA-1


Tgt Ion: 42 Resp: 902
Ion Ratio Lower Upper
42 100
72 0.0 30.6 46.0#
71 39.9 30.5 45.7



#53
Toluene
Concen: 0.086 ppbv
RT: 6.910 min Scan# 2108
Delta R.T. -0.010 min
Lab File: VL043310.D
Acq: 08 Dec 2025 12:40

Tgt Ion: 91 Resp: 3785
Ion Ratio Lower Upper
91 100
92 84.9 47.4 71.0#





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.869 min Scan# 21

Delta R.T. -0.003 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40

Instrument :

MSVOA_L

ClientSampleId :

AA-1





Tgt Ion:117 Resp: 311445

Ion Ratio Lower Upper

117 100

82 66.3 56.1 84.1

119 32.4 25.0 37.6



#68

1-Bromo-4-Fluorobenzene

Concen: 10.227 ppbv

RT: 10.364 min Scan# 3175

Delta R.T. -0.003 min

Lab File: VL043310.D

Acq: 08 Dec 2025 12:40




Tgt Ion: 95 Resp: 233482

Ion Ratio Lower Upper

95 100

174 66.4 50.2 75.2

175 5.1 4.1 6.1



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043311.D
 Acq On : 08 Dec 2025 13:16
 Operator : SY/MD
 Sample : Q3799-01DUP
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DUP

Quant Time: Dec 09 00:11:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.784	49	155357	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	397814	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	318793	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	229577	9.825	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.496	85	8883	0.405	ppbv	96
3) Chlorodifluoromethane	1.473	51	9471	0.440	ppbv #	91
4) Chloromethane	1.531	50	17124	2.090	ppbv #	87
7) Chloroethane	1.703	64	162	0.046	ppbv	99
9) Propene	1.486	41	6430	0.703	ppbv #	1
11) Trichlorofluoromethane	1.875	101	4183	0.197	ppbv	90
12) 1,1,2-Trichlorotrifluo...	2.137	101	1148	0.069	ppbv	88
13) Ethanol	1.719	45	12812	3.982	ppbv	99
15) Acetone	1.832	43	103724	4.806	ppbv	95
16) 1,3-Butadiene	1.599	39	3175	0.330	ppbv #	43
17) tert-Butyl alcohol	2.040	59	2973	0.105	ppbv	99
19) Isopropyl Alcohol	1.884	45	9039	0.741	ppbv #	1
20) Methylene Chloride	2.059	84	24089	3.258	ppbv	95
23) Vinyl Acetate	2.451	43	9886	0.379	ppbv #	97
25) Ethyl Acetate	2.823	43	21691	0.466	ppbv #	80
26) Hexane	2.823	57	29316	1.436	ppbv #	41
29) Chloroform	2.846	83	4392	0.151	ppbv	92
34) 2-Butanone	2.557	43	14218	0.506	ppbv	98
35) Carbon Tetrachloride	3.758	117	1347	0.045	ppbv #	77
36) Benzene	3.648	78	6108	0.160	ppbv	94
39) 1,2-Dichloropropane	3.949	63	101486	7.278	ppbv #	24
53) Toluene	6.924	91	4945	0.110	ppbv #	81
60) o-Xylene	9.519	91	2361	0.048	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 811
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP


Tgt Ion: 49 Resp: 155357
Ion Ratio Lower Upper
49 100
128 49.7 23.0 69.0
130 61.6 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.405 ppbv
RT: 1.496 min Scan# 435
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 85 Resp: 8883
Ion Ratio Lower Upper
85 100
87 29.8 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.440 ppbv
RT: 1.473 min Scan# 41
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP


Tgt Ion: 51 Resp: 9471
Ion Ratio Lower Upper
51 100
67 14.6 14.9 22.3#



#4
Chloromethane
Concen: 2.090 ppbv
RT: 1.531 min Scan# 446
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 50 Resp: 17124
Ion Ratio Lower Upper
50 100
52 37.1 24.1 36.1#









#7
Chloroethane
Concen: 0.046 ppbv
RT: 1.703 min Scan# 499
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP


Tgt Ion: 64 Resp: 162
Ion Ratio Lower Upper
64 100
66 32.8 26.8 40.2



#9
Propene
Concen: 0.703 ppbv
RT: 1.486 min Scan# 432
Delta R.T. 0.006 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 41 Resp: 6430
Ion Ratio Lower Upper
41 100
42 50.0 3.7 5.5#









#11
 Trichlorofluoromethane
 Concen: 0.197 ppbv
 RT: 1.875 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16




Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DUP


Tgt Ion:101 Resp: 4183
 Ion Ratio Lower Upper
 101 100
 103 57.7 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.069 ppbv
 RT: 2.137 min Scan# 633
 Delta R.T. 0.000 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16

Tgt Ion:101 Resp: 1148
 Ion Ratio Lower Upper
 101 100
 151 84.3 59.2 88.8









#13
 Ethanol
 Concen: 3.982 ppbv
 RT: 1.719 min Scan# 503
 Delta R.T. 0.003 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16




Instrument :
 MSVOA_1
 ClientSampleId :
 AA-1DUP


Tgt Ion: 45 Resp: 12812
 Ion Ratio Lower Upper
 45 100
 46 38.2 15.0 59.8



#15
 Acetone
 Concen: 4.806 ppbv
 RT: 1.832 min Scan# 539
 Delta R.T. 0.000 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16

Tgt Ion: 43 Resp: 103724
 Ion Ratio Lower Upper
 43 100
 58 28.4 20.7 31.1









#16
1,3-Butadiene
Concen: 0.330 ppbv
RT: 1.599 min Scan# 410
Delta R.T. -0.006 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP


Tgt Ion: 39 Resp: 3175
Ion Ratio Lower Upper
39 100
54 26.3 59.4 89.2#



#17
tert-Butyl alcohol
Concen: 0.105 ppbv
RT: 2.040 min Scan# 603
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 59 Resp: 2973
Ion Ratio Lower Upper
59 100
57 11.7 9.0 13.4









#19
Isopropyl Alcohol
Concen: 0.741 ppbv
RT: 1.884 min Scan# 51
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP


Tgt Ion: 45 Resp: 9039
Ion Ratio Lower Upper
45 100
58 326.3 31.3 46.9#



#20
Methylene Chloride
Concen: 3.258 ppbv
RT: 2.059 min Scan# 609
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 84 Resp: 24089
Ion Ratio Lower Upper
84 100
49 157.1 132.9 199.3
51 45.8 34.8 52.2
86 61.9 48.2 72.2









#23
 Vinyl Acetate
 Concen: 0.379 ppbv
 RT: 2.451 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DUP

Tgt Ion: 43 Resp: 9886




Ion	Ratio	Lower	Upper
43	100		
86	9.4	6.2	9.2#
42	9.0	8.0	12.0
44	5.3	4.0	6.0




#25
 Ethyl Acetate
 Concen: 0.466 ppbv
 RT: 2.823 min Scan# 845
 Delta R.T. -0.006 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16

Tgt Ion: 43 Resp: 21691

Ion	Ratio	Lower	Upper
43	100		
45	2.1	8.1	12.1#
61	0.9	7.0	10.4#
70	3.3	7.0	10.6#









#26
Hexane
Concen: 1.436 ppbv
RT: 2.823 min Scan# 844
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP


Tgt Ion: 57	Resp: 29316
Ion Ratio	Lower Upper
57 100	
41 101.3	74.6 111.8
43 75.4	179.0 268.4#
56 61.6	42.3 63.5



#29
Chloroform
Concen: 0.151 ppbv
RT: 2.846 min Scan# 852
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 83	Resp: 4392
Ion Ratio	Lower Upper
83 100	
85 57.2	50.5 75.7









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.949 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_1
ClientSampleId :
AA-1DUP


Tgt Ion:114 Resp: 397814
Ion Ratio Lower Upper
114 100
63 25.5 21.0 31.6
88 17.8 14.9 22.3



#34
2-Butanone
Concen: 0.506 ppbv
RT: 2.557 min Scan# 763
Delta R.T. 0.006 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 43 Resp: 14218
Ion Ratio Lower Upper
43 100
72 21.7 18.1 27.1









#35
Carbon Tetrachloride
Concen: 0.045 ppbv
RT: 3.758 min Scan# 1133
Delta R.T. 0.003 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP


Tgt Ion:117 Resp: 1347
Ion Ratio Lower Upper
117 100
119 70.8 74.6 112.0#
121 18.1 23.2 34.8#



#36
Benzene
Concen: 0.160 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 78 Resp: 6108
Ion Ratio Lower Upper
78 100
77 19.3 17.5 26.3
50 16.5 15.5 23.3





#39

1,2-Dichloropropane

Concen: 7.278 ppbv

RT: 3.949 min Scan# 1119

Delta R.T. -0.353 min

Lab File: VL043311.D

Acq: 08 Dec 2025 13:16

Instrument :

MSVOA_1

ClientSampleId :

AA-1DUP





Tgt Ion: 63 Resp: 101486

Ion Ratio Lower Upper

63 100

41 0.0 62.6 93.8#

62 17.4 51.0 76.6#



#53

Toluene

Concen: 0.110 ppbv

RT: 6.924 min Scan# 2112

Delta R.T. 0.003 min

Lab File: VL043311.D




Acq: 08 Dec 2025 13:16


Tgt Ion: 91 Resp: 4945

Ion Ratio Lower Upper

91 100

92 73.7 47.4 71.0#









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 2117
Delta R.T. 0.000 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16




Instrument :
MSVOA_L
ClientSampleId :
AA-1DUP


Tgt Ion:117 Resp: 318793
Ion Ratio Lower Upper
117 100
82 65.7 56.1 84.1
119 32.2 25.0 37.6



#60
o-Xylene
Concen: 0.048 ppbv
RT: 9.519 min Scan# 2914
Delta R.T. -0.437 min
Lab File: VL043311.D
Acq: 08 Dec 2025 13:16

Tgt Ion: 91 Resp: 2361
Ion Ratio Lower Upper
91 100
106 45.6 20.9 62.8








#68
 1-Bromo-4-Fluorobenzene
 Concen: 9.825 ppbv
 RT: 10.367 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: VL043311.D
 Acq: 08 Dec 2025 13:16

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DUP

Tgt Ion: 95 Resp: 229577

Ion	Ratio	Lower	Upper
95	100		
174	68.8	50.2	75.2
175	5.2	4.1	6.1



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043312.D
 Acq On : 08 Dec 2025 13:52
 Operator : SY/MD
 Sample : Q3799-02
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 9B-IA-1

Quant Time: Dec 09 00:11:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Bromochloromethane	2.784	49	137082	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	357699	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	281919	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	213574	10.335	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 103.400%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.499	85	7699	0.398	ppbv	99
3) Chlorodifluoromethane	1.476	51	9652	0.508	ppbv #	78
4) Chloromethane	1.531	50	3969	0.549	ppbv	92
9) Propene	1.486	41	4514	0.559	ppbv #	1
11) Trichlorofluoromethane	1.877	101	3724	0.199	ppbv	89
12) 1,1,2-Trichlorotrifluo...	2.140	101	1125	0.077	ppbv	87
13) Ethanol	1.722	45	50530	17.797	ppbv	99
15) Acetone	1.835	43	166746	8.755	ppbv	99
16) 1,3-Butadiene	1.599	39	943	0.111	ppbv #	38
17) tert-Butyl alcohol	1.884	59	3955	0.158	ppbv	95
19) Isopropyl Alcohol	1.884	45	109957	10.217	ppbv	99
20) Methylene Chloride	2.059	84	6881	0.479	ppbv	83
23) Vinyl Acetate	2.454	43	7277	0.316	ppbv #	92
25) Ethyl Acetate	2.826	43	15972	0.389	ppbv #	91
26) Hexane	2.826	57	15912	0.884	ppbv #	52
29) Chloroform	2.842	83	1770	0.069	ppbv	84
34) 2-Butanone	2.560	43	7477	0.296	ppbv	94
35) Carbon Tetrachloride	3.755	117	1086	0.040	ppbv #	64
36) Benzene	3.648	78	5022	0.147	ppbv #	77
39) 1,2-Dichloropropane	3.952	63	88229	7.037	ppbv #	24
43) Methyl Methacrylate	4.868	69	1326	0.073	ppbv #	1
53) Toluene	6.923	91	2002	0.049	ppbv #	2
60) o-Xylene	9.953	91	1733	0.040	ppbv	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1


Tgt Ion: 49 Resp: 137082
Ion Ratio Lower Upper
49 100
128 49.8 23.0 69.0
130 63.2 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.398 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 85 Resp: 7699
Ion Ratio Lower Upper
85 100
87 31.3 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.508 ppbv
RT: 1.476 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1


Tgt Ion: 51 Resp: 9652
Ion Ratio Lower Upper
51 100
67 8.7 14.9 22.3#



#4
Chloromethane
Concen: 0.549 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 50 Resp: 3969
Ion Ratio Lower Upper
50 100
52 34.3 24.1 36.1





#9

Propene

Concen: 0.559 ppbv

RT: 1.486 min Scan# 41

Delta R.T. 0.006 min

Lab File: VL043312.D

Acq: 08 Dec 2025 13:52

Instrument :

MSVOA_L

ClientSampleId :





9B-IA-1

Tgt Ion: 41 Resp: 4514

Ion Ratio Lower Upper

41 100

42 61.9 3.7 5.5#



#11

Trichlorofluoromethane

Concen: 0.199 ppbv

RT: 1.877 min Scan# 553

Delta R.T. 0.003 min

Lab File: VL043312.D




Acq: 08 Dec 2025 13:52


Tgt Ion:101 Resp: 3724

Ion Ratio Lower Upper

101 100

103 74.7 52.6 79.0









#12
1,1,2-Trichlorotrifluoroethane
Concen: 0.077 ppbv
RT: 2.140 min Scan# 611
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1


Tgt Ion:101 Resp: 1125
Ion Ratio Lower Upper
101 100
151 63.0 59.2 88.8



#13
Ethanol
Concen: 17.797 ppbv
RT: 1.722 min Scan# 505
Delta R.T. 0.006 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 45 Resp: 50530
Ion Ratio Lower Upper
45 100
46 38.1 15.0 59.8









#15
Acetone
Concen: 8.755 ppbv
RT: 1.835 min Scan# 54
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1


Tgt Ion: 43 Resp: 166746
Ion Ratio Lower Upper
43 100
58 26.3 20.7 31.1



#16
1,3-Butadiene
Concen: 0.111 ppbv
RT: 1.599 min Scan# 467
Delta R.T. -0.007 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 39 Resp: 943
Ion Ratio Lower Upper
39 100
54 21.5 59.4 89.2#









#17
tert-Butyl alcohol
Concen: 0.158 ppbv
RT: 1.884 min Scan# 555
Delta R.T. -0.152 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1


Tgt Ion: 59 Resp: 3955
Ion Ratio Lower Upper
59 100
57 13.3 9.0 13.4



#19
Isopropyl Alcohol
Concen: 10.217 ppbv
RT: 1.884 min Scan# 555
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 45 Resp: 109957
Ion Ratio Lower Upper
45 100
58 39.9 31.3 46.9









#20
Methylene Chloride
Concen: 0.479 ppbv
RT: 2.059 min Scan# 608
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1

Tgt Ion: 84 Resp: 6881




Ion	Ratio	Lower	Upper
84	100		
49	136.6	132.9	199.3
51	40.7	34.8	52.2
86	50.3	48.2	72.2




#23
Vinyl Acetate
Concen: 0.316 ppbv
RT: 2.454 min Scan# 731
Delta R.T. -0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 43 Resp: 7277

Ion	Ratio	Lower	Upper
43	100		
86	13.2	6.2	9.2#
42	9.7	8.0	12.0
44	8.7	4.0	6.0#





#25

Ethyl Acetate

Concen: 0.389 ppbv

RT: 2.826 min Scan# 846

Delta R.T. -0.003 min

Lab File: VL043312.D

Acq: 08 Dec 2025 13:52

Instrument :

MSVOA_L

ClientSampleId :

9B-IA-1

Tgt Ion: 43 Resp: 15972





Ion Ratio Lower Upper

43 100

45 6.7 8.1 12.1#

61 4.6 7.0 10.4#

70 6.5 7.0 10.6#



#26

Hexane

Concen: 0.884 ppbv

RT: 2.826 min Scan# 846

Delta R.T. 0.006 min

Lab File: VL043312.D

Acq: 08 Dec 2025 13:52

Tgt Ion: 57 Resp: 15912




Ion Ratio Lower Upper


57 100

41 100.0 74.6 111.8

43 104.4 179.0 268.4#

56 60.9 42.3 63.5





#29

Chloroform

Concen: 0.069 ppbv

RT: 2.842 min Scan# 851

Delta R.T. -0.000 min

Lab File: VL043312.D

Acq: 08 Dec 2025 13:52

Instrument :

MSVOA_L

ClientSampleId :





9B-IA-1

Tgt Ion: 83 Resp: 1770

Ion Ratio Lower Upper

83 100

85 50.8 50.5 75.7



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.952 min Scan# 1194

Delta R.T. 0.003 min

Lab File: VL043312.D

Acq: 08 Dec 2025 13:52




Tgt Ion: 114 Resp: 357699


Ion Ratio Lower Upper

114 100

63 24.7 21.0 31.6

88 18.3 14.9 22.3









#34
2-Butanone
Concen: 0.296 ppbv
RT: 2.560 min Scan# 761
Delta R.T. 0.009 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1


Tgt Ion: 43 Resp: 7477
Ion Ratio Lower Upper
43 100
72 25.3 18.1 27.1



#35
Carbon Tetrachloride
Concen: 0.040 ppbv
RT: 3.755 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 117 Resp: 1086
Ion Ratio Lower Upper
117 100
119 49.7 74.6 112.0#
121 32.1 23.2 34.8









#36
Benzene
Concen: 0.147 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_1
ClientSampleId :
9B-IA-1


Tgt Ion: 78 Resp: 5022
Ion Ratio Lower Upper
78 100
77 32.5 17.5 26.3#
50 8.7 15.5 23.3#



#39
1,2-Dichloropropane
Concen: 7.037 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. -0.350 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 63 Resp: 88229
Ion Ratio Lower Upper
63 100
41 0.4 62.6 93.8#
62 17.3 51.0 76.6#









#43
Methyl Methacrylate
Concen: 0.073 ppbv
RT: 4.868 min Scan# 1476
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_1
ClientSampleId :
9B-IA-1


Tgt Ion: 69 Resp: 1326
Ion Ratio Lower Upper
69 100
41 0.0 126.7 190.1#
100 0.0 28.6 43.0#



#53
Toluene
Concen: 0.049 ppbv
RT: 6.923 min Scan# 2112
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 91 Resp: 2002
Ion Ratio Lower Upper
91 100
92 132.6 47.4 71.0#









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.875 min Scan# 21
Delta R.T. 0.003 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52




Instrument :
MSVOA_L
ClientSampleId :
9B-IA-1


Tgt Ion:117 Resp: 281919
Ion Ratio Lower Upper
117 100
82 64.8 56.1 84.1
119 32.1 25.0 37.6



#60
o-Xylene
Concen: 0.040 ppbv
RT: 9.953 min Scan# 3048
Delta R.T. -0.004 min
Lab File: VL043312.D
Acq: 08 Dec 2025 13:52

Tgt Ion: 91 Resp: 1733
Ion Ratio Lower Upper
91 100
106 47.3 20.9 62.8








#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.335 ppbv
 RT: 10.370 min Scan# 31
 Delta R.T. 0.003 min
 Lab File: VL043312.D
 Acq: 08 Dec 2025 13:52

Instrument :
 MSVOA_L
 ClientSampleId :
 9B-IA-1

Tgt Ion: 95 Resp: 213574
 Ion Ratio Lower Upper
 95 100
 174 68.6 50.2 75.2
 175 4.9 4.1 6.1



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043313.D
 Acq On : 08 Dec 2025 14:29
 Operator : SY/MD
 Sample : Q3799-03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 10-IA-1

Quant Time: Dec 09 00:11:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.784	49	142316	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	361997	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	299956	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	227860	10.363	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 103.600%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.495	85	8224	0.410	ppbv	96
3) Chlorodifluoromethane	1.476	51	8001	0.405	ppbv #	87
4) Chloromethane	1.531	50	3421	0.456	ppbv	95
9) Propene	1.482	41	14885	1.776	ppbv #	86
10) Heptane	4.978	43	813	0.033	ppbv #	23
11) Trichlorofluoromethane	1.874	101	4374	0.225	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.140	101	1155	0.076	ppbv	99
13) Ethanol	1.722	45	357639	121.331	ppbv	98
15) Acetone	1.842	43	5690107	287.779	ppbv #	17
16) 1,3-Butadiene	1.482	39	13670	1.551	ppbv #	17
17) tert-Butyl alcohol	1.897	59	465772	17.946	ppbv #	91
19) Isopropyl Alcohol	1.897	45	4546385	406.914	ppbv #	36
20) Methylene Chloride	2.062	84	89672	15.850	ppbv	92
21) Allyl Chloride	1.897	41	990499	74.458	ppbv #	77
23) Vinyl Acetate	2.557	43	14010	0.587	ppbv #	92
25) Ethyl Acetate	2.832	43	839316	19.682	ppbv #	92
26) Hexane	2.822	57	41166	2.202	ppbv #	1
27) Carbon Disulfide	2.143	76	706	0.037	ppbv #	53
28) Methyl tert-Butyl Ether	2.829	73	34658	3.776	ppbv #	1
29) Chloroform	2.848	83	1442	0.054	ppbv	90
31) cis-1,2-Dichloroethene	2.832	61	101121	5.441	ppbv #	26
34) 2-Butanone	2.557	43	14010	0.548	ppbv	99
35) Carbon Tetrachloride	3.761	117	2513	0.091	ppbv #	84
36) Benzene	3.654	78	5665	0.163	ppbv	98
37) 1,2-Dichloroethane	3.221	62	966	0.048	ppbv	98
39) 1,2-Dichloropropane	3.952	63	92441	7.286	ppbv #	25
41) Tetrahydrofuran	2.829	42	80551	5.708	ppbv #	44
43) Methyl Methacrylate	4.871	69	1199833	65.204	ppbv	96
44) 2,2,4-Trimethylpentane	4.635	57	2206	0.037	ppbv #	41
53) Toluene	6.920	91	18366	0.447	ppbv	100
58) Ethyl Benzene	9.348	91	2583	0.045	ppbv	90
61) Styrene	9.862	104	2173	0.107	ppbv #	89
66) Benzyl Chloride	12.037	91	98244	13.104	ppbv #	66
70) n-Butylbenzene	12.037	91	98244	1.159	ppbv #	66
74) 1,2,4-Trimethylbenzene	11.529	105	2784	0.054	ppbv #	71


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043313.D
Acq On : 08 Dec 2025 14:29
Operator : SY/MD
Sample : Q3799-03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1

Quant Time: Dec 09 00:11:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt Ion: 49 Resp: 142316
Ion Ratio Lower Upper
49 100
128 49.2 23.0 69.0
130 63.6 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.410 ppbv
RT: 1.495 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 85 Resp: 8224
Ion Ratio Lower Upper
85 100
87 34.0 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.405 ppbv
RT: 1.476 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt Ion: 51 Resp: 8001
Ion Ratio Lower Upper
51 100
67 12.6 14.9 22.3#



#4
Chloromethane
Concen: 0.456 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 50 Resp: 3421
Ion Ratio Lower Upper
50 100
52 27.5 24.1 36.1





#9

Propene

Concen: 1.776 ppbv

RT: 1.482 min Scan# 41

Delta R.T. 0.003 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29

Instrument :

MSVOA_L

ClientSampleId :





10-IA-1

Tgt Ion: 41 Resp: 14885

Ion Ratio Lower Upper

41 100

42 0.0 3.7 5.5#



#10

Heptane

Concen: 0.033 ppbv

RT: 4.978 min Scan# 1511

Delta R.T. 0.013 min

Lab File: VL043313.D




Acq: 08 Dec 2025 14:29


Tgt Ion: 43 Resp: 813

Ion Ratio Lower Upper

43 100

57 102.3 39.7 59.5#









#11
 Trichlorofluoromethane
 Concen: 0.225 ppbv
 RT: 1.874 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: VL043313.D
 Acq: 08 Dec 2025 14:29




Instrument :
 MSVOA_L
 ClientSampleId :
 10-IA-1


Tgt Ion:101 Resp: 4374
 Ion Ratio Lower Upper
 101 100
 103 63.7 52.6 79.0



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.076 ppbv
 RT: 2.140 min Scan# 634
 Delta R.T. 0.003 min
 Lab File: VL043313.D
 Acq: 08 Dec 2025 14:29

Tgt Ion:101 Resp: 1155
 Ion Ratio Lower Upper
 101 100
 151 74.5 59.2 88.8









#13
Ethanol
Concen: 121.331 ppbv
RT: 1.722 min Scan# 503
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt Ion: 45 Resp: 357639
Ion Ratio Lower Upper
45 100
46 38.4 15.0 59.8



#15
Acetone
Concen: 287.779 ppbv
RT: 1.842 min Scan# 542
Delta R.T. 0.009 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 43 Resp: 5690107
Ion Ratio Lower Upper
43 100
58 68.0 20.7 31.1#





#16

1,3-Butadiene

Concen: 1.551 ppbv

RT: 1.482 min Scan# 41

Delta R.T. -0.123 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29

Instrument :

MSVOA_L

ClientSampleId :





10-IA-1

Tgt Ion: 39 Resp: 13670

Ion Ratio Lower Upper

39 100

54 4.2 59.4 89.2#



#17

tert-Butyl alcohol

Concen: 17.946 ppbv

RT: 1.897 min Scan# 559

Delta R.T. -0.140 min

Lab File: VL043313.D




Acq: 08 Dec 2025 14:29


Tgt Ion: 59 Resp: 465772

Ion Ratio Lower Upper

59 100

57 7.9 9.0 13.4#









#19
Isopropyl Alcohol
Concen: 406.914 ppbv
RT: 1.897 min Scan# 51
Delta R.T. 0.016 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt Ion: 45 Resp: 4546385
Ion Ratio Lower Upper
45 100
58 0.0 31.3 46.9#



#20
Methylene Chloride
Concen: 15.850 ppbv
RT: 2.062 min Scan# 610
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 84 Resp: 89672
Ion Ratio Lower Upper
84 100
49 154.0 132.9 199.3
51 47.3 34.8 52.2
86 66.4 48.2 72.2





#21

Allyl Chloride

Concen: 74.458 ppbv

RT: 1.897 min Scan# 51

Delta R.T. -0.195 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29

Instrument :

MSVOA_L

ClientSampleId :

10-IA-1





Tgt Ion: 41 Resp: 990499

Ion Ratio Lower Upper

41 100

76 0.0 22.1 33.1#

39 87.1 61.0 91.4



#23

Vinyl Acetate

Concen: 0.587 ppbv

RT: 2.557 min Scan# 763

Delta R.T. 0.100 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29

Tgt Ion: 43 Resp: 14010




Ion Ratio Lower Upper


43 100

86 0.0 6.2 9.2#

42 9.8 8.0 12.0

44 4.9 4.0 6.0









#25
Ethyl Acetate
Concen: 19.682 ppbv
RT: 2.832 min Scan# 84
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt	Ion: 43	Resp:	839316
Ion	Ratio	Lower	Upper
43	100		
45	14.0	8.1	12.1#
61	12.1	7.0	10.4#
70	10.3	7.0	10.6



#26
Hexane
Concen: 2.202 ppbv
RT: 2.822 min Scan# 845
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 57	Resp: 41166		
Ion	Ratio	Lower	Upper
57	100		
41	120.7	74.6	111.8#
43	2046.6	179.0	268.4#
56	61.0	42.3	63.5









#27
Carbon Disulfide
Concen: 0.037 ppbv
RT: 2.143 min Scan# 61
Delta R.T. -0.004 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt Ion: 76 Resp: 706
Ion Ratio Lower Upper
76 100
44 0.0 16.1 24.1#
78 32.9 9.6 14.4#



#28
Methyl tert-Butyl Ether
Concen: 3.776 ppbv
RT: 2.829 min Scan# 847
Delta R.T. 0.394 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 73 Resp: 34658
Ion Ratio Lower Upper
73 100
57 119.3 19.8 29.8#





#29

Chloroform

Concen: 0.054 ppbv

RT: 2.848 min Scan# 8

Delta R.T. 0.006 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29

Instrument :

MSVOA_L

ClientSampleId :





10-IA-1

Tgt Ion: 83 Resp: 1442

Ion Ratio Lower Upper

83 100

85 55.1 50.5 75.7



#31

cis-1,2-Dichloroethene

Concen: 5.441 ppbv

RT: 2.832 min Scan# 848

Delta R.T. 0.116 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29




Tgt Ion: 61 Resp: 101121


Ion Ratio Lower Upper

61 100

96 0.0 49.3 73.9#

98 0.0 32.1 48.1#









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.955 min Scan# 1193
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt Ion:114 Resp: 361997
Ion Ratio Lower Upper
114 100
63 25.6 21.0 31.6
88 18.1 14.9 22.3



#34
2-Butanone
Concen: 0.548 ppbv
RT: 2.557 min Scan# 763
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 43 Resp: 14010
Ion Ratio Lower Upper
43 100
72 22.1 18.1 27.1









#35
Carbon Tetrachloride
Concen: 0.091 ppbv
RT: 3.761 min Scan# 1135
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument : MSVOA_1
ClientSampleId : 10-IA-1


Tgt Ion:117 Resp: 2513
Ion Ratio Lower Upper
117 100
119 74.3 74.6 112.0#
121 31.5 23.2 34.8



#36
Benzene
Concen: 0.163 ppbv
RT: 3.654 min Scan# 1102
Delta R.T. 0.006 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 78 Resp: 5665
Ion Ratio Lower Upper
78 100
77 22.2 17.5 26.3
50 18.3 15.5 23.3









#37
1,2-Dichloroethane
Concen: 0.048 ppbv
RT: 3.221 min Scan# 90
Delta R.T. 0.009 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_1
ClientSampleId :
10-IA-1


Tgt Ion: 62 Resp: 966
Ion Ratio Lower Upper
62 100
98 6.2 5.5 8.3



#39
1,2-Dichloropropane
Concen: 7.286 ppbv
RT: 3.952 min Scan# 1194
Delta R.T. -0.350 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 63 Resp: 92441
Ion Ratio Lower Upper
63 100
41 0.2 62.6 93.8#
62 18.7 51.0 76.6#





#41

Tetrahydrofuran

Concen: 5.708 ppbv

RT: 2.829 min Scan# 84

Delta R.T. -0.217 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29

Instrument :

MSVOA_L

ClientSampleId :

10-IA-1





Tgt Ion: 42 Resp: 80551

Ion Ratio Lower Upper

42 100

72 0.7 30.6 46.0#

71 8.1 30.5 45.7#



#43

Methyl Methacrylate

Concen: 65.204 ppbv

RT: 4.871 min Scan# 1478

Delta R.T. 0.006 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29




Tgt Ion: 69 Resp: 1199833


Ion Ratio Lower Upper

69 100

41 151.4 126.7 190.1

100 36.1 28.6 43.0





#44

2,2,4-Trimethylpentane

Concen: 0.037 ppbv

RT: 4.635 min Scan# 1405

Delta R.T. 0.009 min

Lab File: VL043313.D

Acq: 08 Dec 2025 14:29

Instrument :

MSVOA_L

ClientSampleId :

10-IA-1





Tgt Ion: 57 Resp: 2206

Ion Ratio Lower Upper

57 100

41 0.0 28.0 42.0#

56 0.0 25.6 38.4#



#53

Toluene

Concen: 0.447 ppbv

RT: 6.920 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VL043313.D




Acq: 08 Dec 2025 14:29


Tgt Ion: 91 Resp: 18366

Ion Ratio Lower Upper

91 100

92 59.4 47.4 71.0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 21
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt Ion:117 Resp: 299956
Ion Ratio Lower Upper
117 100
82 63.9 56.1 84.1
119 32.3 25.0 37.6



#58
Ethyl Benzene
Concen: 0.045 ppbv
RT: 9.348 min Scan# 2861
Delta R.T. 0.003 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion: 91 Resp: 2583
Ion Ratio Lower Upper
91 100
106 23.5 22.8 34.2









#61
Styrene
Concen: 0.107 ppbv
RT: 9.862 min Scan# 3020
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_1
ClientSampleId :
10-IA-1


Tgt Ion:	104	Resp:	2173
Ion	Ratio	Lower	Upper
104	100		
78	64.7	42.1	63.1#
103	48.6	37.2	55.8



#66
Benzyl Chloride
Concen: 13.104 ppbv
RT: 12.037 min Scan# 3692
Delta R.T. 0.430 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29

Tgt Ion:	91	Resp:	98244
Ion	Ratio	Lower	Upper
91	100		
126	0.2	14.4	21.6#
128	1.1	4.6	6.8#









#68
1-Bromo-4-Fluorobenzene
Concen: 10.363 ppbv
RT: 10.367 min Scan# 3176
Delta R.T. -0.000 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29




Instrument :
MSVOA_L
ClientSampleId :
10-IA-1


Tgt Ion: 95 Resp: 227860
Ion Ratio Lower Upper
95 100
174 69.3 50.2 75.2
175 4.9 4.1 6.1



#70
n-Butylbenzene
Concen: 1.159 ppbv
RT: 12.037 min Scan# 3692
Delta R.T. -0.237 min
Lab File: VL043313.D
Acq: 08 Dec 2025 14:29


Tgt Ion: 91 Resp: 98244
Ion Ratio Lower Upper
91 100
92 79.3 43.5 65.3#
134 8.5 19.0 28.6#







#74
 1,2,4-Trimethylbenzene
 Concen: 0.054 ppbv
 RT: 11.529 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: VL043313.D
 Acq: 08 Dec 2025 14:29

Instrument :
 MSVOA_L
 ClientSampleId :
 10-IA-1



Tgt Ion:	105	Resp:	2784
Ion	Ratio	Lower	Upper
105	100		
120	38.2	41.7	62.5#
91	51.1	70.0	105.0#
77	21.4	23.8	35.8#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043314.D
 Acq On : 08 Dec 2025 15:37
 Operator : SY/MD
 Sample : Q3799-04
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 11-IA-1

Quant Time: Dec 09 00:12:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.781	49	138499	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.946	114	354324	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	279503	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.364	95	214484	10.469	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.700%


Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.496	85	7849	0.402	ppbv	96
3) Chlorodifluoromethane	1.473	51	9090	0.473	ppbv #	83
4) Chloromethane	1.531	50	3636	0.498	ppbv	94
9) Propene	1.483	41	6390	0.783	ppbv #	86
11) Trichlorofluoromethane	1.878	101	4305	0.227	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.140	101	1091	0.074	ppbv	100
13) Ethanol	1.722	45	1527543	532.508	ppbv	91
15) Acetone	1.832	43	153023	7.952	ppbv	100
16) 1,3-Butadiene	1.599	39	2201	0.257	ppbv #	29
17) tert-Butyl alcohol	2.040	59	3031	0.120	ppbv	100
19) Isopropyl Alcohol	1.884	45	507615	46.685	ppbv #	37
20) Methylene Chloride	2.059	84	6519	0.396	ppbv	88
21) Allyl Chloride	2.040	41	1053	0.081	ppbv #	21
23) Vinyl Acetate	2.447	43	4178	0.180	ppbv #	84
25) Ethyl Acetate	2.829	43	27132	0.654	ppbv #	95
26) Hexane	2.823	57	22119	1.216	ppbv #	58
29) Chloroform	2.845	83	1255	0.049	ppbv	93
34) 2-Butanone	2.554	43	8673	0.347	ppbv #	90
35) Carbon Tetrachloride	3.755	117	2524	0.094	ppbv #	86
36) Benzene	3.648	78	5240	0.154	ppbv	94
37) 1,2-Dichloroethane	3.205	62	720	0.036	ppbv #	80
39) 1,2-Dichloropropane	3.946	63	88097	7.094	ppbv #	23
52) Tetrachloroethene	8.205	164	443	0.035	ppbv #	53
53) Toluene	6.917	91	7807	0.194	ppbv #	70
58) Ethyl Benzene	9.345	91	2046	0.038	ppbv	93
60) o-Xylene	9.953	91	1391	0.032	ppbv	78
74) 1,2,4-Trimethylbenzene	11.523	105	2617	0.054	ppbv #	46
79) Naphthalene	13.507	128	5897	0.138	ppbv #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.781 min Scan# 811
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_1
ClientSampleId :
11-IA-1


Tgt Ion: 49 Resp: 138499
Ion Ratio Lower Upper
49 100
128 49.7 23.0 69.0
130 63.4 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.402 ppbv
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 85 Resp: 7849
Ion Ratio Lower Upper
85 100
87 29.4 25.4 38.2









#3
Chlorodifluoromethane
Concen: 0.473 ppbv
RT: 1.473 min Scan# 411
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_1
ClientSampleId :
11-IA-1


Tgt Ion: 51 Resp: 9090
Ion Ratio Lower Upper
51 100
67 11.0 14.9 22.3#



#4
Chloromethane
Concen: 0.498 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 50 Resp: 3636
Ion Ratio Lower Upper
50 100
52 33.1 24.1 36.1





#9

Propene

Concen: 0.783 ppbv

RT: 1.483 min Scan# 41

Delta R.T. 0.003 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Instrument :

MSVOA_1

ClientSampleId :





11-IA-1

Tgt Ion: 41 Resp: 6390

Ion Ratio Lower Upper

41 100

42 0.0 3.7 5.5#



#11

Trichlorofluoromethane

Concen: 0.227 ppbv

RT: 1.878 min Scan# 553

Delta R.T. 0.003 min

Lab File: VL043314.D




Acq: 08 Dec 2025 15:37


Tgt Ion: 101 Resp: 4305

Ion Ratio Lower Upper

101 100

103 67.4 52.6 79.0





#12

1,1,2-Trichlorotrifluoroethane

Concen: 0.074 ppbv

RT: 2.140 min Scan# 61

Delta R.T. 0.003 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Instrument :

MSVOA_L

ClientSampleId :





11-IA-1

Tgt Ion:101 Resp: 1091

Ion Ratio Lower Upper

101 100

151 74.1 59.2 88.8



#13

Ethanol

Concen: 532.508 ppbv

RT: 1.722 min Scan# 505

Delta R.T. 0.006 min

Lab File: VL043314.D




Acq: 08 Dec 2025 15:37


Tgt Ion: 45 Resp: 1527543

Ion Ratio Lower Upper

45 100

46 42.8 15.0 59.8









#15
Acetone
Concen: 7.952 ppbv
RT: 1.832 min Scan# 511
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_L
ClientSampleId :
11-IA-1


Tgt Ion: 43 Resp: 153023
Ion Ratio Lower Upper
43 100
58 25.9 20.7 31.1



#16
1,3-Butadiene
Concen: 0.257 ppbv
RT: 1.599 min Scan# 467
Delta R.T. -0.006 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 39 Resp: 2201
Ion Ratio Lower Upper
39 100
54 14.1 59.4 89.2#









#17
tert-Butyl alcohol
Concen: 0.120 ppbv
RT: 2.040 min Scan# 602
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_1
ClientSampleId :
11-IA-1


Tgt Ion: 59 Resp: 3031
Ion Ratio Lower Upper
59 100
57 11.1 9.0 13.4



#19
Isopropyl Alcohol
Concen: 46.685 ppbv
RT: 1.884 min Scan# 555
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 45 Resp: 507615
Ion Ratio Lower Upper
45 100
58 0.9 31.3 46.9#









#20
Methylene Chloride
Concen: 0.396 ppbv
RT: 2.059 min Scan# 608
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_1
ClientSampleId :
11-IA-1


Tgt Ion:	84	Resp:	6519
Ion	Ratio	Lower	Upper
84	100		
49	148.0	132.9	199.3
51	51.3	34.8	52.2
86	66.4	48.2	72.2



#21
Allyl Chloride
Concen: 0.081 ppbv
RT: 2.040 min Scan# 603
Delta R.T. -0.052 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion:	41	Resp:	1053
Ion	Ratio	Lower	Upper
41	100		
76	0.0	22.1	33.1#
39	0.0	61.0	91.4#









#23
Vinyl Acetate
Concen: 0.180 ppbv
RT: 2.447 min Scan# 732
Delta R.T. -0.010 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Instrument :
MSVOA_1
ClientSampleId :
11-IA-1

Tgt Ion: 43 Resp: 4178




Ion	Ratio	Lower	Upper
43	100		
86	6.2	6.2	9.2
42	0.0	8.0	12.0#
44	9.1	4.0	6.0#




#25
Ethyl Acetate
Concen: 0.654 ppbv
RT: 2.829 min Scan# 847
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 43 Resp: 27132

Ion	Ratio	Lower	Upper
43	100		
45	10.0	8.1	12.1
61	5.8	7.0	10.4#
70	5.6	7.0	10.6#









#26
Hexane
Concen: 1.216 ppbv
RT: 2.823 min Scan# 845
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_L
ClientSampleId :
11-IA-1


Tgt Ion:	57	Resp:	22119
Ion	Ratio	Lower	Upper
57	100		
41	104.5	74.6	111.8
43	125.3	179.0	268.4#
56	66.2	42.3	63.5#



#29
Chloroform
Concen: 0.049 ppbv
RT: 2.845 min Scan# 852
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion:	83	Resp:	1255
Ion	Ratio	Lower	Upper
83	100		
85	57.6	50.5	75.7









#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.946 min Scan# 1193
Delta R.T. -0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_1
ClientSampleId :
11-IA-1


Tgt Ion:114 Resp: 354324
Ion Ratio Lower Upper
114 100
63 24.9 21.0 31.6
88 18.1 14.9 22.3



#34
2-Butanone
Concen: 0.347 ppbv
RT: 2.554 min Scan# 762
Delta R.T. 0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 43 Resp: 8673
Ion Ratio Lower Upper
43 100
72 27.2 18.1 27.1#









#35
Carbon Tetrachloride
Concen: 0.094 ppbv
RT: 3.755 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_L
ClientSampleId :
11-IA-1


Tgt Ion:117 Resp: 2524
Ion Ratio Lower Upper
117 100
119 92.0 74.6 112.0
121 0.0 23.2 34.8#



#36
Benzene
Concen: 0.154 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 78 Resp: 5240
Ion Ratio Lower Upper
78 100
77 25.4 17.5 26.3
50 21.4 15.5 23.3





#37

1,2-Dichloroethane

Concen: 0.036 ppbv

RT: 3.205 min Scan# 91

Delta R.T. -0.006 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Instrument :

MSVOA_1

ClientSampleId :





11-IA-1

Tgt Ion: 62 Resp: 720

Ion Ratio Lower Upper

62 100

98 0.0 5.5 8.3#



#39

1,2-Dichloropropane

Concen: 7.094 ppbv

RT: 3.946 min Scan# 1192

Delta R.T. -0.356 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37




Tgt Ion: 63 Resp: 88097


Ion Ratio Lower Upper

63 100

41 0.0 62.6 93.8#

62 16.4 51.0 76.6#









#52
Tetrachloroethene
Concen: 0.035 ppbv
RT: 8.205 min Scan# 2110
Delta R.T. -0.007 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_1
ClientSampleId :
11-IA-1


Tgt Ion:	164	Resp:	443
Ion	Ratio	Lower	Upper
164	100		
166	119.6	99.4	149.2
129	49.9	81.8	122.8#
131	1.0	77.4	116.2#



#53
Toluene
Concen: 0.194 ppbv
RT: 6.917 min Scan# 2110
Delta R.T. -0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion:	91	Resp:	7807
Ion	Ratio	Lower	Upper
91	100		
92	82.0	47.4	71.0#









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.869 min Scan# 211
Delta R.T. -0.003 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_1
ClientSampleId :
11-IA-1


Tgt Ion:117 Resp: 279503
Ion Ratio Lower Upper
117 100
82 65.9 56.1 84.1
119 32.6 25.0 37.6



#58
Ethyl Benzene
Concen: 0.038 ppbv
RT: 9.345 min Scan# 2860
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion: 91 Resp: 2046
Ion Ratio Lower Upper
91 100
106 32.1 22.8 34.2





#60

o-Xylene

Concen: 0.032 ppbv

RT: 9.953 min Scan# 3049

Delta R.T. -0.003 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37

Instrument :

MSVOA_1

ClientSampleId :





11-IA-1

Tgt Ion: 91 Resp: 1391

Ion Ratio Lower Upper

91 100

106 55.5 20.9 62.8



#68

1-Bromo-4-Fluorobenzene

Concen: 10.469 ppbv

RT: 10.364 min Scan# 3175

Delta R.T. -0.003 min

Lab File: VL043314.D

Acq: 08 Dec 2025 15:37




Tgt Ion: 95 Resp: 214484


Ion Ratio Lower Upper

95 100

174 68.8 50.2 75.2

175 4.8 4.1 6.1









#74
1,2,4-Trimethylbenzene
Concen: 0.054 ppbv
RT: 11.523 min Scan# 3533
Delta R.T. -0.006 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37




Instrument :
MSVOA_1
ClientSampleId :
11-IA-1

Tgt Ion:	105	Resp:	2617
Ion	Ratio	Lower	Upper
105	100		
120	30.7	41.7	62.5#
91	13.4	70.0	105.0#
77	19.9	23.8	35.8#



#79
Naphthalene
Concen: 0.138 ppbv
RT: 13.507 min Scan# 4146
Delta R.T. -0.000 min
Lab File: VL043314.D
Acq: 08 Dec 2025 15:37

Tgt Ion:	128	Resp:	5897
Ion	Ratio	Lower	Upper
128	100		
127	85.8	9.9	14.9#
129	6.9	8.4	12.6#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043315.D
 Acq On : 08 Dec 2025 16:25
 Operator : SY/MD
 Sample : Q3799-05
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 13-IA-1

Quant Time: Dec 09 00:12:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards							
1)	Bromochloromethane	2.777	49	133616	10.000	ppbv	0.00
33)	1,4-Difluorobenzene	3.943	114	344933	10.000	ppbv	0.00
55)	Chlorobenzene-d5	8.865	117	261904	10.000	ppbv	0.00
System Monitoring Compounds							
68)	1-Bromo-4-Fluorobenzene	10.361	95	195928	10.206	ppbv	0.00
Spiked Amount		10.000	Range	65 - 135	Recovery	=	102.100%
Target Compounds							Qvalue
2)	Dichlorodifluoromethane	1.496	85	7550	0.401	ppbv	94
3)	Chlorodifluoromethane	1.473	51	24896	1.343	ppbv #	81
4)	Chloromethane	1.528	50	3225	0.458	ppbv	93
9)	Propene	1.486	41	6937	0.881	ppbv #	1
11)	Trichlorofluoromethane	1.874	101	4607	0.252	ppbv	94
12)	1,1,2-Trichlorotrifluo...	2.136	101	1083	0.076	ppbv	96
13)	Ethanol	1.716	45	117047	42.294	ppbv	99
15)	Acetone	1.829	43	65404	3.523	ppbv	100
16)	1,3-Butadiene	1.599	39	2213	0.267	ppbv #	40
17)	tert-Butyl alcohol	2.036	59	2347	0.096	ppbv	99
19)	Isopropyl Alcohol	1.881	45	152976	14.583	ppbv #	38
20)	Methylene Chloride	2.056	84	8141	0.763	ppbv #	96
23)	Vinyl Acetate	2.447	43	2075	0.093	ppbv #	62
25)	Ethyl Acetate	2.819	43	32153	0.803	ppbv #	82
26)	Hexane	2.819	57	39916	2.274	ppbv #	44
29)	Chloroform	2.836	83	836	0.034	ppbv	89
34)	2-Butanone	2.551	43	6283	0.258	ppbv	99
36)	Benzene	3.645	78	6800	0.206	ppbv	97
39)	1,2-Dichloropropane	3.943	63	88762	7.342	ppbv #	24
53)	Toluene	6.914	91	7294	0.186	ppbv	93
74)	1,2,4-Trimethylbenzene	11.526	105	2441	0.054	ppbv #	51
75)	1,3-Dichlorobenzene	11.659	146	947	0.036	ppbv #	79
76)	1,4-Dichlorobenzene	11.659	146	947	0.036	ppbv #	79
79)	Naphthalene	13.500	128	59509	1.488	ppbv #	1
80)	Naphthalene,2-methyl-	14.452	142	1880	0.179	ppbv #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043315.D
Acq On : 08 Dec 2025 16:25
Operator : SY/MD
Sample : Q3799-05
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
13-IA-1

Quant Time: Dec 09 00:12:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.777 min Scan# 831
Delta R.T. -0.004 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion: 49 Resp: 133616
Ion Ratio Lower Upper
49 100
128 49.4 23.0 69.0
130 62.4 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.401 ppbv
RT: 1.496 min Scan# 435
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion: 85 Resp: 7550
Ion Ratio Lower Upper
85 100
87 35.4 25.4 38.2









#3
Chlorodifluoromethane
Concen: 1.343 ppbv
RT: 1.473 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_1
ClientSampleId :
13-IA-1


Tgt Ion: 51 Resp: 24896
Ion Ratio Lower Upper
51 100
67 9.9 14.9 22.3#



#4
Chloromethane
Concen: 0.458 ppbv
RT: 1.528 min Scan# 445
Delta R.T. -0.003 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion: 50 Resp: 3225
Ion Ratio Lower Upper
50 100
52 26.3 24.1 36.1





#9

Propene

Concen: 0.881 ppbv

RT: 1.486 min Scan# 41

Delta R.T. 0.006 min

Lab File: VL043315.D

Acq: 08 Dec 2025 16:25

Instrument :

MSVOA_L

ClientSampleId :





13-IA-1

Tgt Ion: 41 Resp: 6937

Ion Ratio Lower Upper

41 100

42 45.2 3.7 5.5#



#11

Trichlorofluoromethane

Concen: 0.252 ppbv

RT: 1.874 min Scan# 552

Delta R.T. -0.000 min

Lab File: VL043315.D




Acq: 08 Dec 2025 16:25


Tgt Ion:101 Resp: 4607

Ion Ratio Lower Upper

101 100

103 60.7 52.6 79.0









#12
1,1,2-Trichlorotrifluoroethane
Concen: 0.076 ppbv
RT: 2.136 min Scan# 61
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion:101 Resp: 1083
Ion Ratio Lower Upper
101 100
151 70.9 59.2 88.8



#13
Ethanol
Concen: 42.294 ppbv
RT: 1.716 min Scan# 503
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion: 45 Resp: 117047
Ion Ratio Lower Upper
45 100
46 38.1 15.0 59.8









#15
Acetone
Concen: 3.523 ppbv
RT: 1.829 min Scan# 51
Delta R.T. -0.003 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion: 43 Resp: 65404
Ion Ratio Lower Upper
43 100
58 25.8 20.7 31.1



#16
1,3-Butadiene
Concen: 0.267 ppbv
RT: 1.599 min Scan# 467
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion: 39 Resp: 2213
Ion Ratio Lower Upper
39 100
54 23.1 59.4 89.2#





#17

tert-Butyl alcohol

Concen: 0.096 ppbv

RT: 2.036 min Scan# 602

Delta R.T. -0.000 min

Lab File: VL043315.D

Acq: 08 Dec 2025 16:25

Instrument :

MSVOA_L

ClientSampleId :





13-IA-1

Tgt Ion: 59 Resp: 2347

Ion Ratio Lower Upper

59 100

57 11.6 9.0 13.4



#19

Isopropyl Alcohol

Concen: 14.583 ppbv

RT: 1.881 min Scan# 554

Delta R.T. -0.000 min

Lab File: VL043315.D




Acq: 08 Dec 2025 16:25


Tgt Ion: 45 Resp: 152976

Ion Ratio Lower Upper

45 100

58 1.2 31.3 46.9#









#20
Methylene Chloride
Concen: 0.763 ppbv
RT: 2.056 min Scan# 608
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion:	84	Resp:	8141
Ion	Ratio	Lower	Upper
84	100		
49	166.8	132.9	199.3
51	52.5	34.8	52.2#
86	65.1	48.2	72.2



#23
Vinyl Acetate
Concen: 0.093 ppbv
RT: 2.447 min Scan# 729
Delta R.T. -0.010 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion:	43	Resp:	2075
Ion	Ratio	Lower	Upper
43	100		
86	23.0	6.2	9.2#
42	0.0	8.0	12.0#
44	22.1	4.0	6.0#









#25
Ethyl Acetate
Concen: 0.803 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.010 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion: 43	Resp: 32153
Ion Ratio	Lower Upper
43 100	
45 3.3	8.1 12.1#
61 1.9	7.0 10.4#
70 2.6	7.0 10.6#



#26
Hexane
Concen: 2.274 ppbv
RT: 2.819 min Scan# 844
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion: 57	Resp: 39916
Ion Ratio	Lower Upper
57 100	
41 100.1	74.6 111.8
43 81.2	179.0 268.4#
56 58.6	42.3 63.5









#29
Chloroform
Concen: 0.034 ppbv
RT: 2.836 min Scan# 84
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion: 83 Resp: 836
Ion Ratio Lower Upper
83 100
85 71.5 50.5 75.7



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.943 min Scan# 1191
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion: 114 Resp: 344933
Ion Ratio Lower Upper
114 100
63 25.8 21.0 31.6
88 17.8 14.9 22.3









#34
2-Butanone
Concen: 0.258 ppbv
RT: 2.551 min Scan# 71
Delta R.T. -0.000 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion: 43 Resp: 6283
Ion Ratio Lower Upper
43 100
72 23.0 18.1 27.1



#36
Benzene
Concen: 0.206 ppbv
RT: 3.645 min Scan# 1099
Delta R.T. -0.003 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion: 78 Resp: 6800
Ion Ratio Lower Upper
78 100
77 22.1 17.5 26.3
50 21.8 15.5 23.3





#39

1,2-Dichloropropane

Concen: 7.342 ppbv

RT: 3.943 min Scan# 1191

Delta R.T. -0.359 min

Lab File: VL043315.D

Acq: 08 Dec 2025 16:25

Instrument :

MSVOA_L

ClientSampleId :

13-IA-1





Tgt Ion: 63 Resp: 88762

Ion Ratio Lower Upper

63 100

41 0.0 62.6 93.8#

62 16.9 51.0 76.6#



#53

Toluene

Concen: 0.186 ppbv

RT: 6.914 min Scan# 2109

Delta R.T. -0.007 min

Lab File: VL043315.D




Acq: 08 Dec 2025 16:25


Tgt Ion: 91 Resp: 7294

Ion Ratio Lower Upper

91 100

92 64.2 47.4 71.0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.865 min Scan# 21
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion:117 Resp: 261904
Ion Ratio Lower Upper
117 100
82 64.6 56.1 84.1
119 32.6 25.0 37.6



#68
1-Bromo-4-Fluorobenzene
Concen: 10.206 ppbv
RT: 10.361 min Scan# 3174
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion: 95 Resp: 195928
Ion Ratio Lower Upper
95 100
174 67.0 50.2 75.2
175 5.0 4.1 6.1









#74
1,2,4-Trimethylbenzene
Concen: 0.054 ppbv
RT: 11.526 min Scan# 3535
Delta R.T. -0.003 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion:	105	Resp:	2441
Ion	Ratio	Lower	Upper
105	100		
120	34.5	41.7	62.5#
91	18.6	70.0	105.0#
77	21.0	23.8	35.8#



#75
1,3-Dichlorobenzene
Concen: 0.036 ppbv
RT: 11.659 min Scan# 3575
Delta R.T. 0.058 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25

Tgt Ion:	146	Resp:	947
Ion	Ratio	Lower	Upper
146	100		
111	72.0	23.4	70.3#
148	69.5	31.6	95.0









#76
1,4-Dichlorobenzene
Concen: 0.036 ppbv
RT: 11.659 min Scan# 311
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25




Instrument :
MSVOA_L
ClientSampleId :
13-IA-1


Tgt Ion:146 Resp: 947
Ion Ratio Lower Upper
146 100
111 72.0 22.8 68.3#
148 69.5 31.8 95.4



#79
Naphthalene
Concen: 1.488 ppbv
RT: 13.500 min Scan# 4144
Delta R.T. -0.007 min
Lab File: VL043315.D
Acq: 08 Dec 2025 16:25


Tgt Ion:128 Resp: 59509
Ion Ratio Lower Upper
128 100
127 975.9 9.9 14.9#
129 3.9 8.4 12.6#







#80
 Naphthalene, 2-methyl-
 Concen: 0.179 ppbv
 RT: 14.452 min Scan# 44
 Delta R.T. -0.000 min
 Lab File: VL043315.D
 Acq: 08 Dec 2025 16:25

Instrument :
 MSVOA_L
 ClientSampleId :
 13-IA-1



Tgt Ion:	142	Resp:	1880
Ion	Ratio	Lower	Upper
142	100		
141	382.4	67.1	100.7#
115	45.4	31.9	47.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
 Data File : VL043316.D
 Acq On : 08 Dec 2025 17:00
 Operator : SY/MD
 Sample : Q3799-06
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 16-IA-1

Quant Time: Dec 09 00:13:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Nov 19 01:11:25 2025
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards							
1)	Bromochloromethane	2.784	49	146093	10.000	ppbv	0.00
33)	1,4-Difluorobenzene	3.949	114	393021	10.000	ppbv	0.00
55)	Chlorobenzene-d5	8.872	117	306652	10.000	ppbv	0.00
System Monitoring Compounds							
68)	1-Bromo-4-Fluorobenzene	10.367	95	232936	10.363	ppbv	0.00
Spiked Amount		10.000	Range	65 - 135	Recovery	=	103.600%
Target Compounds							
							Qvalue
2)	Dichlorodifluoromethane	1.499	85	8618	0.418	ppbv	92
3)	Chlorodifluoromethane	1.466	51	58484	2.886	ppbv #	62
4)	Chloromethane	1.531	50	5664	0.735	ppbv	98
6)	Bromomethane	1.667	94	156	0.038	ppbv #	4
9)	Propene	1.560	41	27719	3.221	ppbv #	1
10)	Heptane	4.972	43	1069	0.042	ppbv #	28
11)	Trichlorofluoromethane	1.874	101	4165	0.209	ppbv	86
12)	1,1,2-Trichlorotrifluo...	2.136	101	1349	0.087	ppbv	86
13)	Ethanol	1.722	45	462202	152.750	ppbv	99
15)	Acetone	1.832	43	604849	29.800	ppbv	97
16)	1,3-Butadiene	1.832	39	29138	3.220	ppbv #	14
17)	tert-Butyl alcohol	1.910	59	1075072	40.352	ppbv #	92
19)	Isopropyl Alcohol	1.910	45	7888776	687.812	ppbv #	38
20)	Methylene Chloride	2.062	84	44935	7.301	ppbv	89
21)	Allyl Chloride	1.910	41	2138169	156.575	ppbv #	74
23)	Vinyl Acetate	2.470	43	3183	0.130	ppbv #	33
25)	Ethyl Acetate	2.826	43	77111	1.761	ppbv #	93
26)	Hexane	2.826	57	52020	2.710	ppbv #	72
29)	Chloroform	2.839	83	1029	0.038	ppbv	85
31)	cis-1,2-Dichloroethene	2.832	61	4180	0.219	ppbv #	26
34)	2-Butanone	2.554	43	12307	0.444	ppbv	99
35)	Carbon Tetrachloride	3.755	117	1777	0.060	ppbv #	83
36)	Benzene	3.648	78	7554	0.201	ppbv	95
39)	1,2-Dichloropropane	3.949	63	96525	7.007	ppbv #	25
53)	Toluene	6.920	91	11377	0.255	ppbv	100
58)	Ethyl Benzene	9.344	91	13162	0.223	ppbv	93
59)	m/p-Xylene	9.523	91	42461	0.897	ppbv	97
60)	o-Xylene	9.953	91	11276	0.239	ppbv	90
61)	Styrene	9.862	104	2334	0.113	ppbv	97
66)	Benzyl Chloride	11.542	91	558	0.073	ppbv #	78
70)	n-Butylbenzene	12.034	91	48515	0.560	ppbv #	65
74)	1,2,4-Trimethylbenzene	11.529	105	2546	0.048	ppbv #	48
79)	Naphthalene	13.510	128	2251	0.048	ppbv #	85
80)	Naphthalene,2-methyl-	14.458	142	1788	0.145	ppbv #	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043316.D
Acq On : 08 Dec 2025 17:00
Operator : SY/MD
Sample : Q3799-06
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
16-IA-1

Quant Time: Dec 09 00:13:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.784 min Scan# 81
Delta R.T. 0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 49 Resp: 146093
Ion Ratio Lower Upper
49 100
128 50.3 23.0 69.0
130 64.6 29.9 89.7



#2
Dichlorodifluoromethane
Concen: 0.418 ppbv
RT: 1.499 min Scan# 436
Delta R.T. 0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 85 Resp: 8618
Ion Ratio Lower Upper
85 100
87 36.4 25.4 38.2









#3
Chlorodifluoromethane
Concen: 2.886 ppbv
RT: 1.466 min Scan# 41
Delta R.T. -0.007 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 51 Resp: 58484
Ion Ratio Lower Upper
51 100
67 1.7 14.9 22.3#



#4
Chloromethane
Concen: 0.735 ppbv
RT: 1.531 min Scan# 446
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 50 Resp: 5664
Ion Ratio Lower Upper
50 100
52 28.7 24.1 36.1









#6
Bromomethane
Concen: 0.038 ppbv
RT: 1.667 min Scan# 41
Delta R.T. 0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 94 Resp: 156
Ion Ratio Lower Upper
94 100
96 0.0 72.9 109.3#



#9
Propene
Concen: 3.221 ppbv
RT: 1.560 min Scan# 455
Delta R.T. 0.081 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 41 Resp: 27719
Ion Ratio Lower Upper
41 100
42 154.5 3.7 5.5#









#10
Heptane
Concen: 0.042 ppbv
RT: 4.972 min Scan# 11
Delta R.T. 0.006 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 43 Resp: 1069
Ion Ratio Lower Upper
43 100
57 0.0 39.7 59.5#



#11
Trichlorofluoromethane
Concen: 0.209 ppbv
RT: 1.874 min Scan# 552
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 101 Resp: 4165
Ion Ratio Lower Upper
101 100
103 54.8 52.6 79.0









#12
1,1,2-Trichlorotrifluoroethane
Concen: 0.087 ppbv
RT: 2.136 min Scan# 611
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion:101 Resp: 1349
Ion Ratio Lower Upper
101 100
151 62.1 59.2 88.8



#13
Ethanol
Concen: 152.750 ppbv
RT: 1.722 min Scan# 505
Delta R.T. 0.006 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 45 Resp: 462202
Ion Ratio Lower Upper
45 100
46 37.9 15.0 59.8









#15
Acetone
Concen: 29.800 ppbv
RT: 1.832 min Scan# 51
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 43 Resp: 604849
Ion Ratio Lower Upper
43 100
58 27.2 20.7 31.1



#16
1,3-Butadiene
Concen: 3.220 ppbv
RT: 1.832 min Scan# 539
Delta R.T. 0.226 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 39 Resp: 29138
Ion Ratio Lower Upper
39 100
54 1.7 59.4 89.2#









#17
tert-Butyl alcohol
Concen: 40.352 ppbv
RT: 1.910 min Scan# 50
Delta R.T. -0.126 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 59 Resp: 1075072
Ion Ratio Lower Upper
59 100
57 8.1 9.0 13.4#



#19
Isopropyl Alcohol
Concen: 687.812 ppbv
RT: 1.910 min Scan# 563
Delta R.T. 0.029 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 45 Resp: 7888776
Ion Ratio Lower Upper
45 100
58 1.0 31.3 46.9#









#20
Methylene Chloride
Concen: 7.301 ppbv
RT: 2.062 min Scan# 61
Delta R.T. 0.006 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Instrument :
MSVOA_L
ClientSampleId :
16-IA-1

Tgt Ion: 84 Resp: 44935




Ion	Ratio	Lower	Upper
84	100		
49	144.4	132.9	199.3
51	47.5	34.8	52.2
86	58.7	48.2	72.2




#21
Allyl Chloride
Concen: 156.575 ppbv
RT: 1.910 min Scan# 563
Delta R.T. -0.181 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 41 Resp: 2138169

Ion	Ratio	Lower	Upper
41	100		
76	0.0	22.1	33.1#
39	89.9	61.0	91.4









#23
Vinyl Acetate
Concen: 0.130 ppbv
RT: 2.470 min Scan# 71
Delta R.T. 0.013 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_1
ClientSampleId :
16-IA-1


Tgt Ion:	43	Resp:	3183
Ion	Ratio	Lower	Upper
43	100		
86	2.0	6.2	9.2#
42	49.9	8.0	12.0#
44	25.5	4.0	6.0#



#25
Ethyl Acetate
Concen: 1.761 ppbv
RT: 2.826 min Scan# 846
Delta R.T. -0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion:	43	Resp:	77111
Ion	Ratio	Lower	Upper
43	100		
45	13.6	8.1	12.1#
61	6.9	7.0	10.4#
70	6.2	7.0	10.6#









#26
Hexane
Concen: 2.710 ppbv
RT: 2.826 min Scan# 84
Delta R.T. 0.006 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 57	Resp: 52020
Ion Ratio	Lower Upper
57 100	
41 97.1	74.6 111.8
43 154.4	179.0 268.4#
56 58.0	42.3 63.5



#29
Chloroform
Concen: 0.038 ppbv
RT: 2.839 min Scan# 850
Delta R.T. -0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 83	Resp: 1029
Ion Ratio	Lower Upper
83 100	
85 74.9	50.5 75.7









#31
 cis-1,2-Dichloroethene
 Concen: 0.219 ppbv
 RT: 2.832 min Scan# 848
 Delta R.T. 0.116 min
 Lab File: VL043316.D
 Acq: 08 Dec 2025 17:00




Instrument :
 MSVOA_L
 ClientSampleId :
 16-IA-1


Tgt Ion: 61 Resp: 4180
 Ion Ratio Lower Upper
 61 100
 96 0.0 49.3 73.9#
 98 0.0 32.1 48.1#



#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 3.949 min Scan# 1193
 Delta R.T. -0.000 min
 Lab File: VL043316.D
 Acq: 08 Dec 2025 17:00

Tgt Ion: 114 Resp: 393021
 Ion Ratio Lower Upper
 114 100
 63 24.6 21.0 31.6
 88 17.7 14.9 22.3









#34
2-Butanone
Concen: 0.444 ppbv
RT: 2.554 min Scan# 761
Delta R.T. 0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 43 Resp: 12307
Ion Ratio Lower Upper
43 100
72 22.9 18.1 27.1



#35
Carbon Tetrachloride
Concen: 0.060 ppbv
RT: 3.755 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 117 Resp: 1777
Ion Ratio Lower Upper
117 100
119 73.4 74.6 112.0#
121 32.6 23.2 34.8









#36
Benzene
Concen: 0.201 ppbv
RT: 3.648 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_1
ClientSampleId :
16-IA-1


Tgt Ion:	78	Resp:	7554
Ion	Ratio	Lower	Upper
78	100		
77	19.3	17.5	26.3
50	21.4	15.5	23.3



#39
1,2-Dichloropropane
Concen: 7.007 ppbv
RT: 3.949 min Scan# 1193
Delta R.T. -0.353 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion:	63	Resp:	96525
Ion	Ratio	Lower	Upper
63	100		
41	0.2	62.6	93.8#
62	18.5	51.0	76.6#









#53
Toluene
Concen: 0.255 ppbv
RT: 6.920 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1


Tgt Ion: 91 Resp: 11377
Ion Ratio Lower Upper
91 100
92 59.0 47.4 71.0



#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.872 min Scan# 2714
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 117 Resp: 306652
Ion Ratio Lower Upper
117 100
82 64.3 56.1 84.1
119 32.1 25.0 37.6









#58
Ethyl Benzene
Concen: 0.223 ppbv
RT: 9.344 min Scan# 21
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_1
ClientSampleId :
16-IA-1


Tgt Ion: 91 Resp: 13162
Ion Ratio Lower Upper
91 100
106 32.4 22.8 34.2



#59
m/p-Xylene
Concen: 0.897 ppbv
RT: 9.523 min Scan# 2915
Delta R.T. -0.020 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 91 Resp: 42461
Ion Ratio Lower Upper
91 100
106 47.0 36.1 54.1









#60
o-Xylene
Concen: 0.239 ppbv
RT: 9.953 min Scan# 3049
Delta R.T. -0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_1
ClientSampleId :
16-IA-1


Tgt Ion: 91 Resp: 11276
Ion Ratio Lower Upper
91 100
106 48.2 20.9 62.8



#61
Styrene
Concen: 0.113 ppbv
RT: 9.862 min Scan# 3020
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion:104 Resp: 2334
Ion Ratio Lower Upper
104 100
78 55.6 42.1 63.1
103 47.8 37.2 55.8









#66
Benzyl Chloride
Concen: 0.073 ppbv
RT: 11.542 min Scan# 3153
Delta R.T. -0.065 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_1
ClientSampleId :
16-IA-1


Tgt Ion: 91 Resp: 558
Ion Ratio Lower Upper
91 100
126 6.1 14.4 21.6#
128 7.3 4.6 6.8#



#68
1-Bromo-4-Fluorobenzene
Concen: 10.363 ppbv
RT: 10.367 min Scan# 3176
Delta R.T. -0.000 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion: 95 Resp: 232936
Ion Ratio Lower Upper
95 100
174 68.8 50.2 75.2
175 5.0 4.1 6.1





#70

n-Butylbenzene

Concen: 0.560 ppbv

RT: 12.034 min Scan# 3

Delta R.T. -0.240 min

Lab File: VL043316.D

Acq: 08 Dec 2025 17:00





Instrument :

MSVOA_L

ClientSampleId :

16-IA-1

Tgt Ion:	91	Resp:	48515
Ion Ratio	Lower	Upper	
91	100		
92	81.0	43.5	65.3#
134	8.9	19.0	28.6#



#74

1,2,4-Trimethylbenzene

Concen: 0.048 ppbv




RT: 11.529 min Scan# 3535


Delta R.T. -0.000 min

Lab File: VL043316.D

Acq: 08 Dec 2025 17:00

Tgt Ion:	105	Resp:	2546
Ion Ratio	Lower	Upper	
105	100		
120	69.1	41.7	62.5#
91	21.6	70.0	105.0#
77	6.3	23.8	35.8#









#79
Naphthalene
Concen: 0.048 ppbv
RT: 13.510 min Scan# 4146
Delta R.T. 0.003 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00




Instrument :
MSVOA_L
ClientSampleId :
16-IA-1

Tgt Ion:128 Resp: 2251
Ion Ratio Lower Upper
128 100
127 14.0 9.9 14.9
129 20.9 8.4 12.6#



#80
Naphthalene,2-methyl-
Concen: 0.145 ppbv
RT: 14.458 min Scan# 4440
Delta R.T. 0.006 min
Lab File: VL043316.D
Acq: 08 Dec 2025 17:00

Tgt Ion:142 Resp: 1788
Ion Ratio Lower Upper
142 100
141 86.1 67.1 100.7
115 64.3 31.9 47.9#



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043331.D
Acq On : 09 Dec 2025 07:30
Operator : SY/MD
Sample : Q3799-03DL2 400X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL2

Quant Time: Dec 10 02:12:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	2.780	49	122404	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	307022	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.869	117	245820	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.367	95	171652	9.526	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.300%
Target Compounds						
					Qvalue	
13) Ethanol	1.719	45	710	0.280	ppbv #	38
15) Acetone	1.832	43	127495	7.497	ppbv	100
19) Isopropyl Alcohol	1.884	45	32181	3.349	ppbv #	1
20) Methylene Chloride	2.055	84	3914	Below Cal	#	83
25) Ethyl Acetate	2.829	43	1495	0.041	ppbv #	86
39) 1,2-Dichloropropane	3.946	63	78942	7.336	ppbv #	24


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120825\
Data File : VL043331.D
Acq On : 09 Dec 2025 07:30
Operator : SY/MD
Sample : Q3799-03DL2 400X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
10-IA-1DL2

Quant Time: Dec 10 02:12:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111825AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Nov 19 01:11:25 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.780 min Scan# 81
Delta R.T. -0.001 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30




Instrument :
MSVOA_1
ClientSampleId :
10-IA-1DL2


Tgt Ion: 49 Resp: 122404
Ion Ratio Lower Upper
49 100
128 49.5 23.0 69.0
130 64.0 29.9 89.7



#13
Ethanol
Concen: 0.280 ppbv
RT: 1.719 min Scan# 504
Delta R.T. 0.003 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30

Tgt Ion: 45 Resp: 710
Ion Ratio Lower Upper
45 100
46 0.0 15.0 59.8#









#15
Acetone
Concen: 7.497 ppbv
RT: 1.832 min Scan# 51
Delta R.T. -0.000 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30




Instrument :
MSVOA_1
ClientSampleId :
10-IA-1DL2


Tgt Ion: 43 Resp: 127495
Ion Ratio Lower Upper
43 100
58 25.7 20.7 31.1



#19
Isopropyl Alcohol
Concen: 3.349 ppbv
RT: 1.884 min Scan# 555
Delta R.T. 0.003 min
Lab File: VL043331.D
Acq: 09 Dec 2025 07:30

Tgt Ion: 45 Resp: 32181
Ion Ratio Lower Upper
45 100
58 101.7 31.3 46.9#





#25

Ethyl Acetate

Concen: 0.041 ppbv

RT: 2.829 min Scan# 847

Delta R.T. -0.000 min

Lab File: VL043331.D

Acq: 09 Dec 2025 07:30

Instrument :

MSVOA_1

ClientSampleId :

10-IA-1DL2

Tgt Ion: 43 Resp: 1495





Ion Ratio Lower Upper

43 100

45 16.9 8.1 12.1#

61 14.4 7.0 10.4#

70 6.3 7.0 10.6#



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.949 min Scan# 1193

Delta R.T. -0.000 min

Lab File: VL043331.D

Acq: 09 Dec 2025 07:30




Tgt Ion: 114 Resp: 307022


Ion Ratio Lower Upper

114 100

63 25.7 21.0 31.6

88 17.7 14.9 22.3





#39

1,2-Dichloropropane

Concen: 7.336 ppbv

RT: 3.946 min Scan# 1192

Delta R.T. -0.356 min

Lab File: VL043331.D

Acq: 09 Dec 2025 07:30

Instrument :

MSVOA_1

ClientSampleId :

10-IA-1DL2





Tgt Ion: 63 Resp: 78942

Ion Ratio Lower Upper

63 100

41 0.3 62.6 93.8#

62 17.0 51.0 76.6#



#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.869 min Scan# 2713

Delta R.T. -0.004 min

Lab File: VL043331.D

Acq: 09 Dec 2025 07:30




Tgt Ion: 117 Resp: 245820


Ion Ratio Lower Upper

117 100

82 63.9 56.1 84.1

119 32.4 25.0 37.6





#68
 1-Bromo-4-Fluorobenzene
 Concen: 9.526 ppbv
 RT: 10.367 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: VL043331.D
 Acq: 09 Dec 2025 07:30

Instrument :
 MSVOA_L
 ClientSampleId :
 10-IA-1DL2

Tgt Ion: 95 Resp: 171652
 Ion Ratio Lower Upper
 95 100
 174 69.1 50.2 75.2
 175 5.0 4.1 6.1

