



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3810

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB149	Mercury	0.076	+/-0.2	U	0.20	CV	12/10/2025	09:16	LB138176

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3810

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB90	Mercury	0.076	+/-0.2	U	0.20	CV	12/10/2025	09:20	LB138176
CCB91	Mercury	0.076	+/-0.2	U	0.20	CV	12/10/2025	09:50	LB138176
CCB92	Mercury	0.076	+/-0.2	U	0.20	CV	12/10/2025	10:17	LB138176
CCB93	Mercury	0.076	+/-0.2	U	0.20	CV	12/10/2025	10:47	LB138176

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3810

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	13:21	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	13:21	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	13:21	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	13:21	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	13:21	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	13:21	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	13:21	LB138187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3810

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	13:59	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	13:59	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	13:59	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	13:59	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	13:59	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	13:59	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	13:59	LB138187
CCB02	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	14:49	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	14:49	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	14:49	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	14:49	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	14:49	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	14:49	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	14:49	LB138187
CCB03	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	15:50	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	15:50	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	15:50	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	15:50	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	15:50	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	15:50	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	15:50	LB138187
CCB04	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	16:46	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	16:46	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	16:46	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	16:46	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	16:46	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	16:46	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	16:46	LB138187
CCB05	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	17:53	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	17:53	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	17:53	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	17:53	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	17:53	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	17:53	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	17:53	LB138187
CCB06	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	18:45	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	18:45	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	18:45	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	18:45	LB138187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile

SDG No.: Q3810

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	12/10/2025	18:45	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	18:45	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	18:45	LB138187
CCB07	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	19:37	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	19:37	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	19:37	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	19:37	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	19:37	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	19:37	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	19:37	LB138187
CCB08	Arsenic	5.12	+/-10	U	20.0	P	12/10/2025	20:25	LB138187
	Barium	14.6	+/-50	U	100	P	12/10/2025	20:25	LB138187
	Cadmium	0.50	+/-3	U	6.00	P	12/10/2025	20:25	LB138187
	Chromium	2.12	+/-5	U	10.0	P	12/10/2025	20:25	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	20:25	LB138187
	Selenium	9.64	+/-10	U	20.0	P	12/10/2025	20:25	LB138187
	Silver	1.62	+/-5	U	10.0	P	12/10/2025	20:25	LB138187