

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Dal-Tile</u>	SDG No.: <u>Q3810</u>
Contract: <u>DALT01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	7.31			-20	20	12/10/2025	13:38	LB138187
	Barium	3.06	6.0	51	-94	106	12/10/2025	13:38	LB138187
	Cadmium	2.15	1.0	215	-5	7	12/10/2025	13:38	LB138187
	Chromium	42.5	52.0	82	42	62	12/10/2025	13:38	LB138187
	Lead	8.91			-12	12	12/10/2025	13:38	LB138187
	Selenium	-1.91			-20	20	12/10/2025	13:38	LB138187
	Silver	5.22			-10	10	12/10/2025	13:38	LB138187
ICSAB01	Arsenic	99.4	104	96	88.4	120	12/10/2025	13:42	LB138187
	Barium	496	537	92	437	637	12/10/2025	13:42	LB138187
	Cadmium	941	972	97	826	1120	12/10/2025	13:42	LB138187
	Chromium	502	542	93	460	624	12/10/2025	13:42	LB138187
	Lead	59.8	49.0	122	37	61	12/10/2025	13:42	LB138187
	Selenium	42.4	46.0	92	26	66	12/10/2025	13:42	LB138187
	Silver	178	201	89	170	232	12/10/2025	13:42	LB138187