

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Dal-Tile</u>	SDG No.: <u>Q3810</u>
Contract: <u>DALT01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	7.31			-20	20	12/10/2025	13:38	LB138187
	Barium	3.06	6.0	51	-94	106	12/10/2025	13:38	LB138187
	Cadmium	2.15	1.0	215	-5	7	12/10/2025	13:38	LB138187
	Chromium	42.5	52.0	82	42	62	12/10/2025	13:38	LB138187
	Lead	8.91			-12	12	12/10/2025	13:38	LB138187
	Selenium	-1.91			-20	20	12/10/2025	13:38	LB138187
	Silver	5.22			-10	10	12/10/2025	13:38	LB138187
ICSAB01	Arsenic	99.4	104	96	88.4	120	12/10/2025	13:42	LB138187
	Barium	496	537	92	437	637	12/10/2025	13:42	LB138187
	Cadmium	941	972	97	826	1120	12/10/2025	13:42	LB138187
	Chromium	502	542	93	460	624	12/10/2025	13:42	LB138187
	Lead	59.8	49.0	122	37	61	12/10/2025	13:42	LB138187
	Selenium	42.4	46.0	92	26	66	12/10/2025	13:42	LB138187
	Silver	178	201	89	170	232	12/10/2025	13:42	LB138187
ICSA	Arsenic	-51.1			-20	20	12/10/2025	13:46	LB138187
	Barium	-25.7	6.0	428	-94	106	12/10/2025	13:46	LB138187
	Cadmium	0.43	1.0	43	-5	7	12/10/2025	13:46	LB138187
	Chromium	33.2	52.0	64	42	62	12/10/2025	13:46	LB138187
	Lead	46.9			-12	12	12/10/2025	13:46	LB138187
	Selenium	-4.52			-20	20	12/10/2025	13:46	LB138187
	Silver	-6.72			-10	10	12/10/2025	13:46	LB138187
ICSAB	Arsenic	89.2	104	86	88.4	120	12/10/2025	13:50	LB138187
	Barium	476	537	89	437	637	12/10/2025	13:50	LB138187
	Cadmium	932	972	96	826	1120	12/10/2025	13:50	LB138187
	Chromium	517	542	95	460	624	12/10/2025	13:50	LB138187
	Lead	72.6	49.0	148	37	61	12/10/2025	13:50	LB138187
	Selenium	3.16	46.0	7	26	66	12/10/2025	13:50	LB138187
	Silver	166	201	83	170	232	12/10/2025	13:50	LB138187
ICSA01	Arsenic	2.67			-20	20	12/11/2025	11:54	LB138198
	Barium	4.34	6.0	72	-94	106	12/11/2025	11:54	LB138198
	Cadmium	3.39	1.0	339	-5	7	12/11/2025	11:54	LB138198
	Chromium	52.4	52.0	101	42	62	12/11/2025	11:54	LB138198
	Lead	7.82			-12	12	12/11/2025	11:54	LB138198
	Selenium	1.32			-20	20	12/11/2025	11:54	LB138198
	Silver	7.73			-10	10	12/11/2025	11:54	LB138198
ICSAB01	Arsenic	105	104	101	88.4	120	12/11/2025	11:59	LB138198
	Barium	502	537	94	437	637	12/11/2025	11:59	LB138198
	Cadmium	946	972	97	826	1120	12/11/2025	11:59	LB138198
	Chromium	535	542	99	460	624	12/11/2025	11:59	LB138198
	Lead	58.0	49.0	118	37	61	12/11/2025	11:59	LB138198
	Selenium	46.2	46.0	100	26	66	12/11/2025	11:59	LB138198

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Dal-Tile

SDG No.: Q3810

Contract: DALT01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Silver	180	201	90	170	232	12/11/2025	11:59	LB138198
ICSA	Arsenic	-8.88			-20	20	12/11/2025	12:05	LB138198
	Barium	8.79	6.0	146	-94	106	12/11/2025	12:05	LB138198
	Cadmium	3.18	1.0	318	-5	7	12/11/2025	12:05	LB138198
	Chromium	50.7	52.0	98	42	62	12/11/2025	12:05	LB138198
	Lead	19.3			-12	12	12/11/2025	12:05	LB138198
	Selenium	-8.60			-20	20	12/11/2025	12:05	LB138198
	Silver	7.22			-10	10	12/11/2025	12:05	LB138198
ICSAB	Arsenic	60.5	104	58	88.4	120	12/11/2025	12:09	LB138198
	Barium	446	537	83	437	637	12/11/2025	12:09	LB138198
	Cadmium	945	972	97	826	1120	12/11/2025	12:09	LB138198
	Chromium	518	542	96	460	624	12/11/2025	12:09	LB138198
	Lead	107	49.0	218	37	61	12/11/2025	12:09	LB138198
	Selenium	82.5	46.0	179	26	66	12/11/2025	12:09	LB138198
	Silver	182	201	90	170	232	12/11/2025	12:09	LB138198