

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Dal-Tile

SDG No.: Q3810

Contract: DALT01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170871BS Mercury	ug/L	4.0	3.88		97	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Dal-Tile

SDG No.: Q3810

Contract: DALT01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170887BS							
Arsenic	ug/L	4000	3800		95	80 - 120	P
Barium	ug/L	1000	918		92	80 - 120	P
Cadmium	ug/L	1000	924		92	80 - 120	P
Chromium	ug/L	2000	1940		97	80 - 120	P
Lead	ug/L	5000	4600		92	80 - 120	P
Selenium	ug/L	10000	9720		97	80 - 120	P
Silver	ug/L	380	355		93	80 - 120	P