

Laboratory Control Sample Summary

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Project: Semi Annual Sampling

Run No.: LB138190

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138190BS							
TSS	mg/L	550	577		105	1	90-110	12/11/2025