

Report of Analysis

Client: Dal Tile - Sunnyvale Plant
Project: Semi Annual Sampling
Client Sample ID: Outfall 002
Lab Sample ID: Q3811-02
Level (low/med): low

Date Collected: 11/20/25
Date Received: 12/08/25
SDG No.: Q3811
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	110		1	5.67	50.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-36-0	Antimony	3.38	U	1	3.38	25.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-38-2	Arsenic	2.56	U	1	2.56	10.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-39-3	Barium	95.9		1	7.28	50.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-41-7	Beryllium	0.28	U	1	0.28	3.00	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-42-8	Boron	85.2		1	7.85	50.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-43-9	Cadmium	0.25	U	1	0.25	3.00	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-70-2	Calcium	52600		1	117	1000	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-47-3	Chromium	2.94	J	1	1.06	5.00	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-48-4	Cobalt	1.13	U	1	1.13	15.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-50-8	Copper	2.30	U	1	2.30	10.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7439-89-6	Iron	63.5		1	11.7	50.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7439-92-1	Lead	1.15	U	1	1.15	6.00	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7439-95-4	Magnesium	3350		1	122	1000	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7439-96-5	Manganese	2.97	U	1	2.97	10.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7439-97-6	Mercury	0.076	UN	1	0.076	0.20	ug/L	12/11/25 09:25	12/11/25 16:33	7470A	M7470A
7439-98-7	Molybdenum	9.37	U	1	9.37	100	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-02-0	Nickel	2.44	J	1	1.53	20.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-09-7	Potassium	2410		1	459	1000	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7782-49-2	Selenium	4.82	U	1	4.82	10.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-22-4	Silver	0.81	U	1	0.81	5.00	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-23-5	Sodium	14900		1	434	1000	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-28-0	Thallium	2.19	U	1	2.19	20.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-31-5	Tin	2.57	U	1	2.57	20.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-32-6	Titanium	3.21	U	1	3.21	20.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-62-2	Vanadium	5.28	J	1	3.13	20.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A
7440-66-6	Zinc	8.33	U	1	8.33	20.0	ug/L	12/10/25 11:30	12/11/25 17:32	6010D	M3010A

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits