



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB151	Mercury	0.076	+/-0.2	U	0.20	CV	12/11/2025	14:51	LB138196

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Mercury	0.076	+/-0.2	U	0.20	CV	12/11/2025	14:58	LB138196
CCB06	Mercury	0.076	+/-0.2	U	0.20	CV	12/11/2025	15:31	LB138196
CCB07	Mercury	0.076	+/-0.2	U	0.20	CV	12/11/2025	15:59	LB138196
CCB08	Mercury	0.076	+/-0.2	U	0.20	CV	12/11/2025	16:26	LB138196
CCB09	Mercury	0.076	+/-0.2	U	0.20	CV	12/11/2025	16:53	LB138196
CCB10	Mercury	0.076	+/-0.2	U	0.20	CV	12/11/2025	17:02	LB138196

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	12/11/2025	11:46	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	11:46	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	11:46	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	11:46	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	11:46	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	11:46	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	11:46	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	11:46	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	11:46	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	11:46	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	11:46	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	11:46	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	11:46	LB138198
	Magnesium	244	+/-1000	U	2000	P	12/11/2025	11:46	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	11:46	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	11:46	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	11:46	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	11:46	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	11:46	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	11:46	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	11:46	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	11:46	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	11:46	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	11:46	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	11:46	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	11:46	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	12/11/2025	12:18	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	12:18	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	12:18	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	12:18	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	12:18	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	12:18	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	12:18	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	12:18	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	12:18	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	12:18	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	12:18	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	12:18	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	12:18	LB138198
	Magnesium	244	+/-1000	U	2000	P	12/11/2025	12:18	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	12:18	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	12:18	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	12:18	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	12:18	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	12:18	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	12:18	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	12:18	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	12:18	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	12:18	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	12:18	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	12:18	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	12:18	LB138198
CCB02	Aluminum	11.3	+/-50	U	100	P	12/11/2025	13:08	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	13:08	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	13:08	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	13:08	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	13:08	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	13:08	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	13:08	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	13:08	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	13:08	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	13:08	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	13:08	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	13:08	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	13:08	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Magnesium	244	+/-1000	U	2000	P	12/11/2025	13:08	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	13:08	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	13:08	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	13:08	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	13:08	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	13:08	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	13:08	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	13:08	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	13:08	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	13:08	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	13:08	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	13:08	LB138198
CCB03	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	13:08	LB138198
	Aluminum	11.3	+/-50	U	100	P	12/11/2025	13:59	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	13:59	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	13:59	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	13:59	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	13:59	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	13:59	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	13:59	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	13:59	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	13:59	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	13:59	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	13:59	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	13:59	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	13:59	LB138198
	Magnesium	244	+/-1000	U	2000	P	12/11/2025	13:59	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	13:59	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	13:59	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	13:59	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	13:59	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	13:59	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	13:59	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	13:59	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	13:59	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	13:59	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	13:59	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	13:59	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	13:59	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Aluminum	20.3	+/-50	J	100	P	12/11/2025	15:04	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	15:04	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	15:04	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	15:04	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	15:04	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	15:04	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	15:04	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	15:04	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	15:04	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	15:04	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	15:04	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	15:04	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	15:04	LB138198
	Magnesium	244	+/-1000	U	2000	P	12/11/2025	15:04	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	15:04	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	15:04	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	15:04	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	15:04	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	15:04	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	15:04	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	15:04	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	15:04	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	15:04	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	15:04	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	15:04	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	15:04	LB138198
CCB05	Aluminum	13.8	+/-50	J	100	P	12/11/2025	15:54	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	15:54	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	15:54	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	15:54	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	15:54	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	15:54	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	15:54	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	15:54	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	15:54	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	15:54	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	15:54	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	15:54	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	15:54	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Magnesium	244	+/-1000	U	2000	P	12/11/2025	15:54	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	15:54	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	15:54	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	15:54	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	15:54	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	15:54	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	15:54	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	15:54	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	15:54	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	15:54	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	15:54	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	15:54	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	15:54	LB138198
CCB06	Aluminum	11.3	+/-50	U	100	P	12/11/2025	16:44	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	16:44	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	16:44	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	16:44	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	16:44	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	16:44	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	16:44	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	16:44	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	16:44	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	16:44	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	16:44	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	16:44	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	16:44	LB138198
	Magnesium	244	+/-1000	U	2000	P	12/11/2025	16:44	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	16:44	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	16:44	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	16:44	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	16:44	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	16:44	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	16:44	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	16:44	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	16:44	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	16:44	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	16:44	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	16:44	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	16:44	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Aluminum	11.3	+/-50	U	100	P	12/11/2025	17:57	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	17:57	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	17:57	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	17:57	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	17:57	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	17:57	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	17:57	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	17:57	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	17:57	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	17:57	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	17:57	LB138198
	Iron	23.9	+/-50	J	100	P	12/11/2025	17:57	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	17:57	LB138198
	Magnesium	244	+/-1000	U	2000	P	12/11/2025	17:57	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	17:57	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	17:57	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	17:57	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	17:57	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	17:57	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	17:57	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	17:57	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	17:57	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	17:57	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	17:57	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	17:57	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	17:57	LB138198
CCB08	Aluminum	11.3	+/-50	U	100	P	12/11/2025	18:52	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	18:52	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	18:52	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	18:52	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	18:52	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	18:52	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	18:52	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	18:52	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	18:52	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	18:52	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	18:52	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	18:52	LB138198
	Lead	2.30	+/-6	U	12.0	P	12/11/2025	18:52	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Magnesium	244	+/-1000	U	2000	P	12/11/2025	18:52	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	18:52	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	18:52	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	18:52	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	18:52	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	18:52	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	18:52	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	18:52	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	18:52	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	18:52	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	18:52	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	18:52	LB138198
CCB09	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	18:52	LB138198
	Aluminum	11.3	+/-50	U	100	P	12/11/2025	19:45	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	19:45	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	19:45	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	19:45	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	19:45	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	19:45	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	19:45	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	19:45	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	19:45	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	19:45	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	19:45	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	19:45	LB138198
	Lead	3.43	+/-6	J	12.0	P	12/11/2025	19:45	LB138198
	Magnesium	244	+/-1000	U	2000	P	12/11/2025	19:45	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	19:45	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	19:45	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	19:45	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	19:45	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	19:45	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	19:45	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	19:45	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	19:45	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	19:45	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	19:45	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	19:45	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	19:45	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB10	Aluminum	11.3	+/-50	U	100	P	12/11/2025	20:37	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	20:37	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	20:37	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	20:37	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	20:37	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	20:37	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	20:37	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	20:37	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	20:37	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	20:37	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	20:37	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	20:37	LB138198
	Lead	2.34	+/-6	J	12.0	P	12/11/2025	20:37	LB138198
	Magnesium	244	+/-1000	U	2000	P	12/11/2025	20:37	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	20:37	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	20:37	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	20:37	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	20:37	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	20:37	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	20:37	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	20:37	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	20:37	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	20:37	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	20:37	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	20:37	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	20:37	LB138198
CCB11	Aluminum	17.9	+/-50	J	100	P	12/11/2025	20:59	LB138198
	Antimony	6.76	+/-25	U	50.0	P	12/11/2025	20:59	LB138198
	Arsenic	5.12	+/-10	U	20.0	P	12/11/2025	20:59	LB138198
	Barium	14.6	+/-50	U	100	P	12/11/2025	20:59	LB138198
	Beryllium	0.56	+/-3	U	6.00	P	12/11/2025	20:59	LB138198
	Boron	15.7	+/-50	U	100	P	12/11/2025	20:59	LB138198
	Cadmium	0.50	+/-3	U	6.00	P	12/11/2025	20:59	LB138198
	Calcium	234	+/-1000	U	2000	P	12/11/2025	20:59	LB138198
	Chromium	2.12	+/-5	U	10.0	P	12/11/2025	20:59	LB138198
	Cobalt	2.26	+/-15	U	30.0	P	12/11/2025	20:59	LB138198
	Copper	4.60	+/-10	U	20.0	P	12/11/2025	20:59	LB138198
	Iron	23.4	+/-50	U	100	P	12/11/2025	20:59	LB138198
	Lead	2.32	+/-6	J	12.0	P	12/11/2025	20:59	LB138198

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal Tile - Sunnyvale Plant

SDG No.: Q3811

Contract: DALT03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Magnesium	244	+/-1000	U	2000	P	12/11/2025	20:59	LB138198
	Manganese	5.94	+/-10	U	20.0	P	12/11/2025	20:59	LB138198
	Molybdenum	18.7	+/-100	U	200	P	12/11/2025	20:59	LB138198
	Nickel	3.06	+/-20	U	40.0	P	12/11/2025	20:59	LB138198
	Potassium	918	+/-1000	U	2000	P	12/11/2025	20:59	LB138198
	Selenium	9.64	+/-10	U	20.0	P	12/11/2025	20:59	LB138198
	Silver	1.62	+/-5	U	10.0	P	12/11/2025	20:59	LB138198
	Sodium	868	+/-1000	U	2000	P	12/11/2025	20:59	LB138198
	Thallium	4.38	+/-20	U	40.0	P	12/11/2025	20:59	LB138198
	Tin	5.14	+/-20	U	40.0	P	12/11/2025	20:59	LB138198
	Titanium	6.42	+/-20	U	40.0	P	12/11/2025	20:59	LB138198
	Vanadium	6.26	+/-20	U	40.0	P	12/11/2025	20:59	LB138198
	Zinc	16.7	+/-20	U	40.0	P	12/11/2025	20:59	LB138198