

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Dal Tile - Sunnyvale Plant</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3811</u>
Contract:	<u>DALT03</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3811-01</u>	Client ID:	<u>Outfall 001DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3811-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Dal Tile - Sunnyvale Plant</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3811</u>
Contract:	<u>DALT03</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3811-01MS</u>	Client ID:	<u>Outfall 001MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3811-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	5.08		4.25		18		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Dal Tile - Sunnyvale Plant **Level:** LOW **SDG No.:** Q3811
Contract: DALT03 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3811-03 **Client ID:** Outfall 003DUP
Percent Solids for Sample: NA **Duplicate ID** Q3811-03DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	878		875		0		P
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	188		182		3		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Boron	ug/L	20	522		522		0		P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Calcium	ug/L	20	56900		56300		1		P
Chromium	ug/L	20	3.81	J	4.03	J	6		P
Cobalt	ug/L	20	15.0	U	15.0	U			P
Copper	ug/L	20	3.45	J	3.39	J	2		P
Iron	ug/L	20	411		409		0		P
Lead	ug/L	20	1.64	J	1.32	J	22		P
Magnesium	ug/L	20	15200		15200		0		P
Manganese	ug/L	20	19.7		19.2		3		P
Molybdenum	ug/L	20	16.0	J	16.0	J	0		P
Nickel	ug/L	20	3.24	J	3.18	J	2		P
Potassium	ug/L	20	4100		4060		1		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	24800		24300		2		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Tin	ug/L	20	20.0	U	20.0	U			P
Titanium	ug/L	20	32.0		33.3		4		P
Vanadium	ug/L	20	10.2	J	9.89	J	3		P
Zinc	ug/L	20	18.0	J	18.3	J	2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Dal Tile - Sunnyvale Plant</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3811</u>
Contract:	<u>DALT03</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3811-03MS</u>	Client ID:	<u>Outfall 003MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3811-03MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1700		1720		1		P
Antimony	ug/L	20	357		360		1		P
Arsenic	ug/L	20	362		361		0		P
Barium	ug/L	20	274		289		5		P
Beryllium	ug/L	20	88.3		88.0		0		P
Boron	ug/L	20	659		668		1		P
Cadmium	ug/L	20	84.1		83.7		0		P
Calcium	ug/L	20	57800		58800		2		P
Chromium	ug/L	20	180		185		3		P
Cobalt	ug/L	20	85.7		85.9		0		P
Copper	ug/L	20	132		133		1		P
Iron	ug/L	20	1770		1790		1		P
Lead	ug/L	20	442		443		0		P
Magnesium	ug/L	20	16200		16400		1		P
Manganese	ug/L	20	106		111		5		P
Molybdenum	ug/L	20	207		212		2		P
Nickel	ug/L	20	238		237		0		P
Potassium	ug/L	20	8570		8580		0		P
Selenium	ug/L	20	857		855		0		P
Silver	ug/L	20	30.6		29.7		3		P
Sodium	ug/L	20	26900		27800		3		P
Thallium	ug/L	20	792		828		4		P
Tin	ug/L	20	297		299		1		P
Titanium	ug/L	20	114		117		3		P
Vanadium	ug/L	20	143		144		1		P
Zinc	ug/L	20	109		109		0		P