

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3813

Client ID : SSTDCCC0.4

Date Analyzed: 12/12/2025

Lab File ID: BN038343.D

Time Analyzed: 09:24

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5196	7.724	14288	10.52	7596	14.39
UPPER LIMIT	10392	8.224	28576	11.019	15192	14.887
LOWER LIMIT	2598	7.224	7144	10.019	3798	13.887
EPA SAMPLE NO.						
01 PB170912BL	6576	7.72	16790	10.52	8053	14.39
02 BR-04DR-451-120825	5531	7.72	14873	10.52	8212	14.39
03 EB01-120825	6003	7.72	15645	10.52	8460	14.39
04 PB170912BS	6159	7.72	16009	10.52	7792	14.39
05 PB170912BSD	5504	7.72	14243	10.52	6818	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.