

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3813

Client ID: SSTDCCC0.4

Date Analyzed: 12/12/2025

Lab File ID: BN038343.D

Time Analyzed: 09:24

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	14177	17.148	11670	21.339	12444	23.633
UPPER LIMIT	28354	17.648	23340	21.839	24888	24.133
LOWER LIMIT	7088.5	16.648	5835	20.839	6222	23.133
EPA SAMPLE NO.						
01 PB170912BL	12820	17.15	8612	21.34	9324	23.63
02 BR-04DR-451-120825	16033	17.15	12704	21.34	12519	23.63
03 EB01-120825	16049	17.15	12341	21.34	12255	23.63
04 PB170912BS	13155	17.15	8480	21.34	8947	23.63
05 PB170912BSD	11665	17.15	8305	21.34	8801	23.62

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.