

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q3813
 Lab File ID: VV039517.D Date Analyzed: 12/10/2025
 Instrument ID: MSVOA_V Time Analyzed: 09:10
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 (DFB) AREA #	RT #	IS2 (CBZ) AREA #	RT #	IS3 (DCB) AREA #	RT #
12 HOUR STD	6796	5.57	6633	8.79	3778	11.18
UPPER LIMIT	13592	5.736	13266	8.955	7556	11.345
LOWER LIMIT	3398	5.396	3316.5	8.615	1889	11.005
EPA SAMPLE NO.						
BR-04DR-451-120825-SIM	6208	5.57	6043	8.80	3061	11.20
EB01-120825	7218	5.57	6792	8.79	3677	11.18
TB01-120825	6637	5.57	6660	8.80	3485	11.20
VHBLK001	6680	5.57	6693	8.80	3533	11.20
VBLK239	5966	5.57	5775	8.80	2945	11.20

IS1 (DFB) = 1,4-Difluorobenzene
 IS2 (CBZ) = Chlorobenzene-d5
 IS3 (DCB) = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.