

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN112425

Review By	John Carlone	Review On	11/25/2025 9:36:43 AM
Supervise By	Maresh Dadoda	Supervise On	11/26/2025 1:52:19 PM
SubDirectory	VN112425	HP Acquire Method	MSVOA_N HP Processing Method 82N112425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP136378
Initial Calibration Stds	VP136444,VP136445,VP136446,VP136447,VP136448,VP136449
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP136450
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088343.D	24 Nov 2025 08:17		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN088344.D	24 Nov 2025 12:38		JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN088345.D	24 Nov 2025 13:11	COM.#74 Peak not detected in 1 and 5 PPB	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN088346.D	24 Nov 2025 13:32		JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN088347.D	24 Nov 2025 13:53	Method failed for Comp. #74	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN088348.D	24 Nov 2025 14:14		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN088349.D	24 Nov 2025 14:35		JC\MD	Ok,M
8	IBLK	IBLK	VN088350.D	24 Nov 2025 14:56		JC\MD	Ok
9	VSTDICV050	ICVVN112425	VN088351.D	24 Nov 2025 15:18		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN121025

Review By	Mahesh Dadoda	Review On	12/11/2025 11:46:28 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/11/2025 12:33:13 PM
SubDirectory	VN121025	HP Acquire Method	HP Processing Method 82N112425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136601
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136602,VP136603

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088486.D	10 Dec 2025 09:29		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088487.D	10 Dec 2025 10:01	CCC failed High For Com.#42	JC\MD	Ok,M
3	VN1210MBL01	VN1210MBL01	VN088488.D	10 Dec 2025 10:36		JC\MD	Ok
4	VN1210WBL01	VN1210WBL01	VN088489.D	10 Dec 2025 10:57		JC\MD	Ok
5	VN1210MBS01	VN1210MBS01	VN088490.D	10 Dec 2025 11:30		JC\MD	Ok,M
6	VN1210WBS01	VN1210WBS01	VN088491.D	10 Dec 2025 12:03		JC\MD	Ok,M
7	Q3813-01	BR-04DR-451-120825	VN088492.D	10 Dec 2025 12:23	vial A pH<2 CCC failed high	JC\MD	Ok
8	Q3813-03	EB01-120825	VN088493.D	10 Dec 2025 12:44	vial A pH<2 EB	JC\MD	Ok
9	Q3813-04	TB01-120825	VN088494.D	10 Dec 2025 13:04	vial A pH<2 TB	JC\MD	Ok
10	Q3770-01DL	4207DL	VN088495.D	10 Dec 2025 13:25		JC\MD	Ok
11	VN1210WBSD01	VN1210WBSD01	VN088496.D	10 Dec 2025 13:45	BSD failed Low for com.#73	JC\MD	Ok,M
12	Q3809-15	TRIP	VN088497.D	10 Dec 2025 14:06	vial A pH<2 TB;BSD Failed Low	JC\MD	ReRun
13	Q3814-02	TP-5	VN088498.D	10 Dec 2025 14:27	vial A pH#5.0	JC\MD	Ok
14	Q3814-04	TP-6	VN088499.D	10 Dec 2025 14:47	vial A pH#5.0	JC\MD	Ok
15	Q3814-06	TP-7	VN088500.D	10 Dec 2025 15:08	vial A pH#5.0	JC\MD	Ok
16	Q3814-08	TP-8	VN088501.D	10 Dec 2025 15:29	vial A pH#5.0	JC\MD	Ok
17	Q3814-10	TP-9	VN088502.D	10 Dec 2025 15:50	vial A pH#5.0	JC\MD	Ok
18	Q3814-12	TP-12	VN088503.D	10 Dec 2025 16:10	vial A pH#5.0	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN121025

Review By	Maresh Dadoda	Review On	12/11/2025 11:46:28 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/11/2025 12:33:13 PM
SubDirectory	VN121025	HP Acquire Method	HP Processing Method 82N112425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136601		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136602,VP136603		

19	Q3814-14	TP-10	VN088504.D	10 Dec 2025 16:31	vial A pH#5.0	JC\MD	Ok
20	Q3814-16	TP-4	VN088505.D	10 Dec 2025 16:52	vial A pH#5.0	JC\MD	Ok
21	Q3815-02	RE	VN088506.D	10 Dec 2025 17:13	vial A pH#5.0	JC\MD	Ok
22	Q3826-02	TP-13	VN088507.D	10 Dec 2025 17:34	vial A pH#5.0	JC\MD	Ok
23	Q3826-04	TP-14	VN088508.D	10 Dec 2025 17:54	vial A pH#5.0	JC\MD	Ok
24	Q3826-06	TP-15	VN088509.D	10 Dec 2025 18:15	vial A pH#5.0	JC\MD	Ok
25	Q3826-08	TP-11	VN088510.D	10 Dec 2025 18:36	vial A pH#5.0	JC\MD	Ok
26	Q3826-10	TP-1	VN088511.D	10 Dec 2025 18:57	vial A pH#5.0	JC\MD	Ok
27	Q3826-12	TP-2	VN088512.D	10 Dec 2025 19:18	vial A pH#5.0 Surrogate fail	JC\MD	Ok
28	Q3826-14	TP-3	VN088513.D	10 Dec 2025 19:38	vial A pH#5.0	JC\MD	Ok
29	VSTDCCC050	VSTDCCC050EC	VN088514.D	10 Dec 2025 19:59		JC\MD	Ok,M

M : Manual Integration