

Surrogate Summary

SDG No.: **Q3813**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3813-01	BR-04DR-451-120825	1,2-Dichloroethane-d4	50	56.7	113		70 (74)	130 (125)
		Dibromofluoromethane	50	47.0	94		70 (75)	130 (124)
		Toluene-d8	50	44.9	90		70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.3	93		70 (77)	130 (121)
Q3813-03	EB01-120825	1,2-Dichloroethane-d4	50	58.2	116		70 (74)	130 (125)
		Dibromofluoromethane	50	49.4	99		70 (75)	130 (124)
		Toluene-d8	50	46.5	93		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101		70 (77)	130 (121)
Q3813-04	TB01-120825	1,2-Dichloroethane-d4	50	60.9	122		70 (74)	130 (125)
		Dibromofluoromethane	50	49.0	98		70 (75)	130 (124)
		Toluene-d8	50	45.1	90		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.7	97		70 (77)	130 (121)
VN1210WBL01	VN1210WBL01	1,2-Dichloroethane-d4	50	57.8	116		70 (74)	130 (125)
		Dibromofluoromethane	50	50.8	102		70 (75)	130 (124)
		Toluene-d8	50	47.0	94		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.0	96		70 (77)	130 (121)
VN1210WBS01	VN1210WBS01	1,2-Dichloroethane-d4	50	57.3	115		70 (74)	130 (125)
		Dibromofluoromethane	50	55.6	111		70 (75)	130 (124)
		Toluene-d8	50	53.9	108		70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.1	110		70 (77)	130 (121)
VN1210WBSD01	VN1210WBSD01	1,2-Dichloroethane-d4	50	60.8	122		70 (74)	130 (125)
		Dibromofluoromethane	50	55.3	111		70 (75)	130 (124)
		Toluene-d8	50	52.5	105		70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.1	108		70 (77)	130 (121)