

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
Lab Code: ACE SDG NO.: Q3813
Lab File ID: VN088487.D Date Analyzed: 12/10/2025
Instrument ID: MSVOA_N Time Analyzed: 10:01
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	249947	8.21	468726	9.08	435160	11.84
UPPER LIMIT	499894	8.706	937452	9.583	870320	12.341
LOWER LIMIT	124974	7.706	234363	8.583	217580	11.341
EPA SAMPLE NO.						
BR-04DR-451-120825	212180	8.21	494367	9.08	485409	11.84
EB01-120825	207235	8.21	478662	9.08	473650	11.84
TB01-120825	225847	8.21	531810	9.08	534272	11.84
VN1210WBL01	199142	8.20	459524	9.08	467279	11.84
VN1210WBS01	279896	8.21	512638	9.08	464243	11.84
VN1210WBSD01	226609	8.21	440776	9.08	408174	11.84

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.