

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3813
 Lab File ID: VN088487.D Date Analyzed: 12/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:01
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	215555	13.77				
UPPER LIMIT	431110	14.27				
LOWER LIMIT	107778	13.27				
EPA SAMPLE NO.						
BR-04DR-451-120825	222072	13.77				
EB01-120825	215051	13.77				
TB01-120825	252826	13.77				
VN1210WBL01	211178	13.77				
VN1210WBS01	225326	13.77				
VN1210WBSD01	199717	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.