

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: Case No.: 12345 MA No.: SDG No.:
Instrument ID: KONELAB Start Date : 12/10/2025
Analytical Method: Spectrophotometry Run Batch : LB138186

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Cyanide	0.999876	0.000907	0.002491	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: Case No.: 12345 MA No.: SDG No.:
Instrument ID: KONELAB Start Date : 12/10/2025
Analytical Method: Spectrophotometry Run Batch : LB138186

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------