

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121225\
 Data File : BN038380.D
 Acq On : 13 Dec 2025 11:03
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 14 22:22:13 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.459	0.461	-0.4	89	0.00
3	n-Nitrosodimethylamine	0.575	0.586	-1.9	92	0.00
4 S	2-Fluorophenol	0.996	0.925	7.1	84	0.00
5 S	Phenol-d6	1.184	1.111	6.2	87	0.00
6	bis(2-Chloroethyl)ether	1.158	1.152	0.5	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	-0.01
8 S	Nitrobenzene-d5	0.337	0.328	2.7	90	-0.01
9	Naphthalene	1.182	1.187	-0.4	91	-0.01
10	Hexachlorobutadiene	0.210	0.210	0.0	90	-0.01
11 SURR	2-Methylnaphthalene-d10	0.564	0.556	1.4	91	0.00
12	2-Methylnaphthalene	0.751	0.738	1.7	91	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.01
14 S	2,4,6-Tribromophenol	0.177	0.149	15.8	83	-0.01
15 S	2-Fluorobiphenyl	1.927	2.161	-12.1	102	0.00
16	Acenaphthylene	1.964	1.881	4.2	92	-0.01
17	Acenaphthene	1.369	1.310	4.3	91	-0.01
18	Fluorene	1.761	1.679	4.7	92	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	-0.01
20	4,6-Dinitro-2-methylphenol	0.066	0.037	43.9#	59	0.00
21	4-Bromophenyl-phenylether	0.297	0.285	4.0	91	0.00
22	Hexachlorobenzene	0.357	0.360	-0.8	95	-0.01
23	Atrazine	0.202	0.185	8.4	90	0.00
24	Pentachlorophenol	0.139	0.113	18.7	86	0.00
25	Phenanthrene	1.392	1.365	1.9	92	0.00
26	Anthracene	1.200	1.147	4.4	92	-0.01
27 SURR	Fluoranthene-d10	0.989	0.972	1.7	91	0.00
28	Fluoranthene	1.480	1.449	2.1	91	-0.01
29 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
30	Pyrene	1.969	1.872	4.9	90	0.00
31 S	Terphenyl-d14	0.893	0.865	3.1	90	0.00
32	Benzo(a)anthracene	1.545	1.529	1.0	95	0.00
33	Chrysene	1.701	1.698	0.2	91	0.00
34	Bis(2-ethylhexyl)phthalate	0.848	0.777	8.4	85	0.00
35 I	Perylene-d12	1.000	1.000	0.0	96	-0.01
36	Indeno(1,2,3-cd)pyrene	2.123	2.105	0.8	95	-0.02
37	Benzo(b)fluoranthene	1.837	1.739	5.3	93	-0.02
38	Benzo(k)fluoranthene	1.832	1.746	4.7	93	-0.01
39 C	Benzo(a)pyrene	1.503	1.438	4.3	94	-0.01
40	Dibenzo(a,h)anthracene	1.637	1.622	0.9	96	-0.02
41	Benzo(g,h,i)perylene	1.829	1.750	4.3	92	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4