

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121225\
 Data File : BN038380.D
 Acq On : 13 Dec 2025 11:03
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 14 22:22:13 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	90	0.00
2	1,4-Dioxane	0.400	0.402	-0.5	89	0.00
3	n-Nitrosodimethylamine	0.400	0.407	-1.7	92	0.00
4 S	2-Fluorophenol	0.400	0.371	7.3	84	0.00
5 S	Phenol-d6	0.400	0.375	6.3	87	0.00
6	bis(2-Chloroethyl)ether	0.400	0.398	0.5	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	90	-0.01
8 S	Nitrobenzene-d5	0.400	0.389	2.8	90	-0.01
9	Naphthalene	0.400	0.401	-0.3	91	-0.01
10	Hexachlorobutadiene	0.400	0.401	-0.3	90	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.394	1.5	91	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	91	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.336	16.0	83	-0.01
15 S	2-Fluorobiphenyl	0.400	0.449	-12.2	102	0.00
16	Acenaphthylene	0.400	0.383	4.3	92	-0.01
17	Acenaphthene	0.400	0.383	4.3	91	-0.01
18	Fluorene	0.400	0.381	4.8	92	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.222	44.5#	59	0.00
21	4-Bromophenyl-phenylether	0.400	0.385	3.8	91	0.00
22	Hexachlorobenzene	0.400	0.403	-0.8	95	-0.01
23	Atrazine	0.400	0.366	8.5	90	0.00
24	Pentachlorophenol	0.400	0.324	19.0	86	0.00
25	Phenanthrene	0.400	0.392	2.0	92	0.00
26	Anthracene	0.400	0.382	4.5	92	-0.01
27 SURR	Fluoranthene-d10	0.400	0.393	1.8	91	0.00
28	Fluoranthene	0.400	0.392	2.0	91	-0.01
29 I	Chrysene-d12	0.400	0.400	0.0	92	0.00
30	Pyrene	0.400	0.380	5.0	90	0.00
31 S	Terphenyl-d14	0.400	0.388	3.0	90	0.00
32	Benzo(a)anthracene	0.400	0.396	1.0	95	0.00
33	Chrysene	0.400	0.399	0.3	91	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.367	8.3	85	0.00
35 I	Perylene-d12	0.400	0.400	0.0	96	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.397	0.8	95	-0.02
37	Benzo(b)fluoranthene	0.400	0.379	5.3	93	-0.02
38	Benzo(k)fluoranthene	0.400	0.381	4.8	93	-0.01
39 C	Benzo(a)pyrene	0.400	0.383	4.3	94	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.396	1.0	96	-0.02
41	Benzo(g,h,i)perylene	0.400	0.383	4.3	92	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4