

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN121025

Review By	rahul	Review On	12/11/2025 4:30:54 PM
Supervise By	mohammad	Supervise On	12/13/2025 12:56:07 AM
SubDirectory	BN121025	HP Acquire Method	BNA_N, 8270_HP Processing Method BN121025

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6947,SP6946,SP6945,SP6944,SP6943,SP6942,SP6940
CCC	SP6945
Internal Standard/PEM	SP6907,10ul/1000ul sample
ICV/I.BLK	SP6949
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038297.D	10 Dec 2025 09:12		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN038298.D	10 Dec 2025 09:52		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN038299.D	10 Dec 2025 10:28		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN038300.D	10 Dec 2025 11:05		RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN038301.D	10 Dec 2025 11:41		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN038302.D	10 Dec 2025 12:17		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN038303.D	10 Dec 2025 12:53		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN038304.D	10 Dec 2025 13:30	Comp# 20 (4,6-Dinitro-2-methylphenol) removed from 5.0 ng level	RC/JU	Ok
9	SSTDICV0.4	ICVBN121025	BN038305.D	10 Dec 2025 14:06		RC/JU	Ok
10	PB170590BL	PB170590BL	BN038306.D	10 Dec 2025 14:42		RC/JU	Ok
11	Q3483-11	HW1025-PT-BNA-SOIL	BN038307.D	10 Dec 2025 15:18	Need 100X Dilution	RC/JU	Dilution
12	Q3483-11DL	HW1025-PT-BNA-SOIL	BN038308.D	10 Dec 2025 15:54		RC/JU	Ok,M
13	SSTDCCC0.4	SSTDCCC0.4EC	BN038309.D	10 Dec 2025 17:15		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN121225

Review By	rahul	Review On	12/12/2025 3:41:10 PM
Supervise By	Jagrut	Supervise On	12/15/2025 2:17:39 PM
SubDirectory	BN121225	HP Acquire Method	BNA_N, 8270_HP Processing Method BN121025

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6875 SP6947,SP6946,SP6945,SP6944,SP6943,SP6942,SP6940
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6945 SP6907,10ul/1000ul sample SP6949

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038342.D	12 Dec 2025 08:45		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038343.D	12 Dec 2025 09:24	Compound #20, 24 fail low.	RC/JU	Ok
3	PB170912BL	PB170912BL	BN038344.D	12 Dec 2025 10:00		RC/JU	Ok
4	Q3788-01DL	OWBR-05-135-120525	BN038345.D	12 Dec 2025 10:36		RC/JU	Ok
5	Q3813-01	BR-04DR-451-120825	BN038346.D	12 Dec 2025 11:12		RC/JU	Ok
6	Q3813-03	EB01-120825	BN038347.D	12 Dec 2025 11:49		RC/JU	Ok
7	Q3828-01	BR-02-133-120925	BN038348.D	12 Dec 2025 12:25		RC/JU	Ok
8	Q3828-03	OWBR-06-219-120925	BN038349.D	12 Dec 2025 13:01		RC/JU	Ok,M
9	Q3828-05	RMW-03B-90-120925	BN038350.D	12 Dec 2025 13:38	Need 5X Dilution	RC/JU	Dilution
10	Q3828-06	OWBR-03-138-120925	BN038351.D	12 Dec 2025 14:14		RC/JU	Ok
11	Q3828-08	FB01-120925	BN038352.D	12 Dec 2025 15:26		RC/JU	Ok
12	Q3835-01	MW-11A-13.5-121025	BN038353.D	12 Dec 2025 16:02		RC/JU	Ok
13	Q3835-02	OWBR-01-170-121025	BN038354.D	12 Dec 2025 16:38		RC/JU	Ok
14	Q3835-04	MW-118-37.5-121025	BN038355.D	12 Dec 2025 17:14	Need 20X Dilution	RC/JU	Dilution
15	Q3835-05	MW-01-7-121025	BN038356.D	12 Dec 2025 17:50		RC/JU	Ok
16	PB170912BS	PB170912BS	BN038357.D	12 Dec 2025 18:27		RC/JU	Ok,M
17	PB170912BSD	PB170912BSD	BN038358.D	12 Dec 2025 19:03		RC/JU	Ok,M
18	SSTDCCC0.4	SSTDCCC0.4EC	BN038359.D	12 Dec 2025 20:16	Compound #20 fail low.	RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN121225

Review By	rahul	Review On	12/12/2025 3:41:10 PM
Supervise By	Jagrut	Supervise On	12/15/2025 2:17:39 PM
SubDirectory	BN121225	HP Acquire Method	BNA_N, 8270_HP Processing Method BN121025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	SP6875 SP6947,SP6946,SP6945,SP6944,SP6943,SP6942,SP6940		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6945 SP6907,10ul/1000ul sample SP6949		

19	DFTPP	DFTPP	BN038360.D	12 Dec 2025 21:31		RC/JU	Ok
20	SSTDCCC0.4	SSTDCCC0.4	BN038361.D	12 Dec 2025 22:10	Compound #20, 24 fail low.	RC/JU	Ok
21	PB170917BL	PB170917BL	BN038362.D	12 Dec 2025 22:47	CCC fail low for compound #24.	RC/JU	ReRun
22	Q3839-01	C0AM5	BN038363.D	12 Dec 2025 23:24	CCC fail low for compound #24.	RC/JU	ReRun
23	Q3839-02	C0AM6	BN038364.D	13 Dec 2025 00:00	CCC fail low for compound #24.	RC/JU	ReRun
24	Q3839-03MS	C0AM6MS	BN038365.D	13 Dec 2025 00:37	CCC fail low for compound #24.	RC/JU	ReRun
25	Q3839-04MSD	C0AM6MSD	BN038366.D	13 Dec 2025 01:13	CCC fail low for compound #24.	RC/JU	ReRun
26	Q3839-05	C0AN1	BN038367.D	13 Dec 2025 01:50	CCC fail low for compound #24.	RC/JU	ReRun
27	Q3839-06	C0AN2	BN038368.D	13 Dec 2025 02:27	CCC fail low for compound #24.	RC/JU	ReRun
28	Q3839-07	C0AN3	BN038369.D	13 Dec 2025 03:03	CCC fail low for compound #24.	RC/JU	ReRun
29	Q3839-08	C0AN4	BN038370.D	13 Dec 2025 03:40	CCC fail low for compound #24.	RC/JU	ReRun
30	Q3839-11	C0AN7	BN038371.D	13 Dec 2025 04:16	CCC fail low for compound #24.	RC/JU	ReRun
31	Q3839-12	C0AS1	BN038372.D	13 Dec 2025 04:53	CCC fail low for compound #24.	RC/JU	ReRun
32	Q3839-13	C0AS2	BN038373.D	13 Dec 2025 05:29	CCC fail low for compound #24.	RC/JU	ReRun
33	Q3839-14	C0AS3	BN038374.D	13 Dec 2025 06:06	CCC fail low for compound #24.	RC/JU	ReRun
34	Q3839-15	C0AS4	BN038375.D	13 Dec 2025 06:42	CCC fail low for compound #24.	RC/JU	ReRun
35	Q3839-16	C0B02	BN038376.D	13 Dec 2025 07:18	CCC fail low for compound #24.	RC/JU	ReRun
36	PB170917BS	PB170917BS	BN038377.D	13 Dec 2025 07:55		RC/JU	Ok,M
37	SSTDCCC0.4	SSTDCCC0.4EC	BN038378.D	13 Dec 2025 09:08		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN121225

Review By	rahul	Review On	12/12/2025 3:41:10 PM
Supervise By	Jagrut	Supervise On	12/15/2025 2:17:39 PM
SubDirectory	BN121225	HP Acquire Method	BNA_N, 8270_HP Processing Method BN121025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	SP6875 SP6947,SP6946,SP6945,SP6944,SP6943,SP6942,SP6940		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6945 SP6907,10ul/1000ul sample SP6949		

38	DFTPP	DFTPP	BN038379.D	13 Dec 2025 10:24		RC/JU	Ok
39	SSTDCCC0.4	SSTDCCC0.4	BN038380.D	13 Dec 2025 11:03	Compound #20 fail low.	RC/JU	Ok
40	PB170919BL	PB170919BL	BN038381.D	13 Dec 2025 11:39	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
41	Q3828-05DL	RMW-03B-90-120925D	BN038382.D	13 Dec 2025 12:16		RC/JU	Ok
42	Q3839-17	C0B11	BN038383.D	13 Dec 2025 12:52		RC/JU	Ok
43	Q3839-18	C0AM7	BN038384.D	13 Dec 2025 13:28	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
44	Q3839-19MS	C0AM7MS	BN038385.D	13 Dec 2025 14:04	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
45	Q3839-19MSD	C0AM7MSD	BN038386.D	13 Dec 2025 14:40	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
46	Q3839-09	C0AN5	BN038387.D	13 Dec 2025 15:16		RC/JU	Ok,M
47	Q3839-10	C0AN6	BN038388.D	13 Dec 2025 15:52	Need 5X.	RC/JU	Dilution
48	Q3839-21	C0B09	BN038389.D	13 Dec 2025 16:28	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
49	Q3839-22	C0AT4	BN038390.D	13 Dec 2025 17:04	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
50	Q3839-23	C0AT5	BN038391.D	13 Dec 2025 17:41	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
51	Q3839-24	C0AT6	BN038392.D	13 Dec 2025 18:17	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
52	Q3839-25	C0AT7	BN038393.D	13 Dec 2025 18:54	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
53	Q3839-26	C0AT8	BN038394.D	13 Dec 2025 19:30	Wrong Sample (Total Scan sample)	RC/JU	Not Ok
54	PB170919BS	PB170919BS	BN038395.D	13 Dec 2025 20:06	Wrong Sample (Total Scan sample)	RC/JU	Not Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN121225

Review By	rahul	Review On	12/12/2025 3:41:10 PM
Supervise By	Jagrut	Supervise On	12/15/2025 2:17:39 PM
SubDirectory	BN121225	HP Acquire Method	BNA_N, 8270_HP Processing Method BN121025

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6947,SP6946,SP6945,SP6944,SP6943,SP6942,SP6940
CCC	SP6945
Internal Standard/PEM	SP6907,10ul/1000ul sample
ICV/I.BLK	SP6949
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

55	SSTDCCC0.4	SSTDCCC0.4EC	BN038396.D	13 Dec 2025 21:18		RC/JU	Ok
----	------------	--------------	------------	-------------------	--	-------	----

M : Manual Integration