

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3828

Client ID : SSTDCCC0.4

Date Analyzed: 12/12/2025

Lab File ID: BN038343.D

Time Analyzed: 09:24

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5196	7.724	14288	10.52	7596	14.39
UPPER LIMIT	10392	8.224	28576	11.019	15192	14.887
LOWER LIMIT	2598	7.224	7144	10.019	3798	13.887
EPA SAMPLE NO.						
01 PB170912BL	6576	7.72	16790	10.52	8053	14.39
02 PB170912BS	6159	7.72	16009	10.52	7792	14.39
03 BR-02-133-120925	7468	7.72	19798	10.52	10393	14.39
04 OWBR-06-219-120925	5145	7.72	13627	10.52	6975	14.39
05 RMW-03B-90-120925	5116	7.72	13853	10.52	6995	14.39
06 PB170912BSD	5504	7.72	14243	10.52	6818	14.39
07 OWBR-03-138-120925	5333	7.72	14542	10.52	7924	14.39
08 FB01-120925	6161	7.72	15794	10.52	8352	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.