

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3828

Client ID : SSTDCCC0.4

Date Analyzed: 12/13/2025

Lab File ID: BN038380.D

Time Analyzed: 11:03

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5610	7.717	14718	10.52	7660	14.39
UPPER LIMIT	11220	8.217	29436	11.019	15320	14.887
LOWER LIMIT	2805	7.217	7359	10.019	3830	13.887
EPA SAMPLE NO.						
01 RMW-03B-90-120925DL	6069	7.72	16077	10.52	8437	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.