

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3828

Client ID: SSTDCCC0.4

Date Analyzed: 12/13/2025

Lab File ID: BN038380.D

Time Analyzed: 11:03

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	13635	17.136	10585	21.331	11030	23.619
UPPER LIMIT	27270	17.636	21170	21.831	22060	24.119
LOWER LIMIT	6817.5	16.636	5292.5	20.831	5515	23.119
EPA SAMPLE NO.						
01 RMW-03B-90-120925DL	15660	17.14	11931	21.33	12174	23.62

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.