



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121125\
Data File : VX048821.D
Acq On : 11 Dec 2025 12:08
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Title : SW846 8260
Last Update : Fri Dec 05 03:27:34 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	8900	PASS
75	95	30	60	50.6	25115	PASS
95	95	100	100	100.0	49597	PASS
96	95	5	9	6.4	3166	PASS
173	174	0.00	2	1.3	505	PASS
174	95	50	100	78.2	38795	PASS
175	174	5	9	8.0	3096	PASS
176	174	95	101	96.1	37275	PASS
177	176	5	9	6.7	2501	PASS