



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048839.D
Acq On : 12 Dec 2025 08:44
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Title : SW846 8260
Last Update : Fri Dec 05 03:27:34 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	10337	PASS
75	95	30	60	49.8	27880	PASS
95	95	100	100	100.0	55947	PASS
96	95	5	9	6.7	3751	PASS
173	174	0.00	2	1.4	584	PASS
174	95	50	100	76.7	42912	PASS
175	174	5	9	7.6	3272	PASS
176	174	95	101	97.0	41611	PASS
177	176	5	9	6.0	2478	PASS