

Surrogate Summary

SDG No.: **Q3828**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3828-01	BR-02-133-120925	1,2-Dichloroethane-d4	50	61.1	122		70 (74)	130 (125)
		Dibromofluoromethane	50	48.7	97		70 (75)	130 (124)
		Toluene-d8	50	47.9	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	59.9	120		70 (77)	130 (121)
Q3828-03	OWBR-06-219-120925	1,2-Dichloroethane-d4	50	56.3	113		70 (74)	130 (125)
		Dibromofluoromethane	50	48.2	96		70 (75)	130 (124)
		Toluene-d8	50	47.1	94		70 (86)	130 (113)
		4-Bromofluorobenzene	50	58.8	118		70 (77)	130 (121)
Q3828-05	RMW-03B-90-120925	1,2-Dichloroethane-d4	50	53.4	107		70 (74)	130 (125)
		Dibromofluoromethane	50	47.1	94		70 (75)	130 (124)
		Toluene-d8	50	48.3	97		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.8	98		70 (77)	130 (121)
Q3828-05DL	RMW-03B-90-120925DL	1,2-Dichloroethane-d4	50	58.2	116		70 (74)	130 (125)
		Dibromofluoromethane	50	47.5	95		70 (75)	130 (124)
		Toluene-d8	50	45.9	92		70 (86)	130 (113)
		4-Bromofluorobenzene	50	58.1	116		70 (77)	130 (121)
Q3828-06	OWBR-03-138-120925	1,2-Dichloroethane-d4	50	61.1	122		70 (74)	130 (125)
		Dibromofluoromethane	50	49.7	99		70 (75)	130 (124)
		Toluene-d8	50	48.2	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	59.8	120		70 (77)	130 (121)
Q3828-08	FB01-120925	1,2-Dichloroethane-d4	50	52.5	105		70 (74)	130 (125)
		Dibromofluoromethane	50	49.4	99		70 (75)	130 (124)
		Toluene-d8	50	49.7	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.0	92		70 (77)	130 (121)
Q3828-09	TB01-120925	1,2-Dichloroethane-d4	50	52.3	105		70 (74)	130 (125)
		Dibromofluoromethane	50	48.8	98		70 (75)	130 (124)
		Toluene-d8	50	49.3	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.5	95		70 (77)	130 (121)
VX1211WBL01	VX1211WBL01	1,2-Dichloroethane-d4	50	47.5	95		70 (74)	130 (125)
		Dibromofluoromethane	50	47.7	95		70 (75)	130 (124)
		Toluene-d8	50	49.4	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.4	87		70 (77)	130 (121)
VX1211WBS01	VX1211WBS01	1,2-Dichloroethane-d4	50	56.0	112		70 (74)	130 (125)
		Dibromofluoromethane	50	54.6	109		70 (75)	130 (124)
		Toluene-d8	50	53.0	106		70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.2	110		70 (77)	130 (121)
VX1211WBSD0	VX1211WBSD01	1,2-Dichloroethane-d4	50	53.0	106		70 (74)	130 (125)
		Dibromofluoromethane	50	53.2	106		70 (75)	130 (124)
		Toluene-d8	50	53.1	106		70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.6	111		70 (77)	130 (121)
VX1212WBL01	VX1212WBL01	1,2-Dichloroethane-d4	50	57.7	115		70 (74)	130 (125)
		Dibromofluoromethane	50	47.6	95		70 (75)	130 (124)
		Toluene-d8	50	47.8	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	60.2	120		70 (77)	130 (121)
VX1212WBS01	VX1212WBS01	1,2-Dichloroethane-d4	50	60.2	120		70 (74)	130 (125)
		Dibromofluoromethane	50	56.5	113		70 (75)	130 (124)
		Toluene-d8	50	55.6	111		70 (86)	130 (113)
		4-Bromofluorobenzene	50	60.4	121		70 (77)	130 (121)

