

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3830

Project: Monthly 2025

Run No.: LB138190

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138190BS							
TSS	mg/L	550	577		105	1	90-110	12/11/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3830

Project: Monthly 2025

Run No.: LB138192

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138192BS							
TPH	mg/L	20.0	18.2		91	1	78-114	12/11/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3830

Project: Monthly 2025

Run No.: LB138192

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138192BSD							
TPH	mg/L	20.0	18.7		94	1	78-114	12/11/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3830

Project: Monthly 2025

Run No.: LB138203

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138203BS							
BOD5	mg/L	198	183		92	1	84.6-115.4	12/11/2025