

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WOOD06

Lab Code: ACE

SDG NO.: Q3834

Continuing Calib Date: 12/11/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 15:07

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.77	5.77	5.67	5.87	0.00
Aroclor-1016-2	(2)	5.79	5.79	5.69	5.89	0.00
Aroclor-1016-3	(3)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-4	(4)	5.95	5.95	5.85	6.05	0.00
Aroclor-1016-5	(5)	6.24	6.24	6.14	6.34	0.00
Aroclor-1260-1	(1)	7.36	7.36	7.26	7.46	0.00
Aroclor-1260-2	(2)	7.61	7.61	7.51	7.71	0.00
Aroclor-1260-3	(3)	7.97	7.97	7.87	8.07	0.00
Aroclor-1260-4	(4)	8.20	8.19	8.09	8.29	-0.01
Aroclor-1260-5	(5)	8.52	8.52	8.42	8.62	0.00
Tetrachloro-m-xylene		4.62	4.61	4.51	4.71	-0.01
Decachlorobiphenyl		10.37	10.37	10.27	10.47	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WOOD06

Lab Code: ACE

SDG NO.: Q3834

Continuing Calib Date: 12/11/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 15:07

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.85	4.85	4.75	4.95	0.00
Aroclor-1016-2	(2)	4.87	4.87	4.77	4.97	0.00
Aroclor-1016-3	(3)	5.04	5.04	4.94	5.14	0.00
Aroclor-1016-4	(4)	5.08	5.08	4.98	5.18	0.00
Aroclor-1016-5	(5)	5.30	5.30	5.20	5.40	0.00
Aroclor-1260-1	(1)	6.32	6.32	6.22	6.42	0.00
Aroclor-1260-2	(2)	6.51	6.51	6.41	6.61	0.00
Aroclor-1260-3	(3)	6.66	6.66	6.56	6.76	0.00
Aroclor-1260-4	(4)	7.13	7.13	7.03	7.23	0.00
Aroclor-1260-5	(5)	7.37	7.37	7.27	7.47	0.00
Tetrachloro-m-xylene		3.77	3.77	3.67	3.87	0.00
Decachlorobiphenyl		8.77	8.77	8.67	8.87	0.00