

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WOOD06
Lab Code: ACE **SDG NO.:** Q3834
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076943.D **Time Analyzed:** 08:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.769	5.667	5.867	521.880	500.000	4.4
Aroclor-1016-2	5.791	5.687	5.887	537.110	500.000	7.4
Aroclor-1016-3	5.854	5.750	5.950	530.290	500.000	6.1
Aroclor-1016-4	5.951	5.847	6.047	537.470	500.000	7.5
Aroclor-1016-5	6.243	6.140	6.340	532.180	500.000	6.4
Aroclor-1260-1	7.360	7.256	7.456	513.510	500.000	2.7
Aroclor-1260-2	7.614	7.509	7.709	504.280	500.000	0.9
Aroclor-1260-3	7.971	7.867	8.067	513.580	500.000	2.7
Aroclor-1260-4	8.198	8.093	8.293	506.040	500.000	1.2
Aroclor-1260-5	8.522	8.416	8.616	507.940	500.000	1.6
Decachlorobiphenyl	10.373	10.265	10.465	51.060	50.000	2.1
Tetrachloro-m-xylene	4.617	4.514	4.714	53.790	50.000	7.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WOOD06
Lab Code: ACE **SDG NO.:** Q3834
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076943.D **Time Analyzed:** 08:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.849	4.750	4.950	485.090	500.000	-3.0
Aroclor-1016-2	4.866	4.767	4.967	504.950	500.000	1.0
Aroclor-1016-3	5.041	4.942	5.142	499.810	500.000	0.0
Aroclor-1016-4	5.083	4.984	5.184	504.840	500.000	1.0
Aroclor-1016-5	5.295	5.197	5.397	495.920	500.000	-0.8
Aroclor-1260-1	6.323	6.224	6.424	487.500	500.000	-2.5
Aroclor-1260-2	6.510	6.412	6.612	479.570	500.000	-4.1
Aroclor-1260-3	6.662	6.564	6.764	482.860	500.000	-3.4
Aroclor-1260-4	7.131	7.033	7.233	487.900	500.000	-2.4
Aroclor-1260-5	7.372	7.273	7.473	481.060	500.000	-3.8
Decachlorobiphenyl	8.764	8.666	8.866	46.740	50.000	-6.5
Tetrachloro-m-xylene	3.771	3.671	3.871	51.980	50.000	4.0