

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WOOD06
Lab Code: ACE **SDG NO.:** Q3834
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076958.D **Time Analyzed:** 15:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.770	5.667	5.867	532.650	500.000	6.5
Aroclor-1016-2	5.791	5.687	5.887	545.770	500.000	9.2
Aroclor-1016-3	5.854	5.750	5.950	532.740	500.000	6.5
Aroclor-1016-4	5.951	5.847	6.047	531.430	500.000	6.3
Aroclor-1016-5	6.244	6.140	6.340	539.630	500.000	7.9
Aroclor-1260-1	7.361	7.256	7.456	514.800	500.000	3.0
Aroclor-1260-2	7.614	7.509	7.709	508.240	500.000	1.6
Aroclor-1260-3	7.972	7.867	8.067	520.180	500.000	4.0
Aroclor-1260-4	8.199	8.093	8.293	508.900	500.000	1.8
Aroclor-1260-5	8.523	8.416	8.616	512.100	500.000	2.4
Decachlorobiphenyl	10.373	10.265	10.465	50.980	50.000	2.0
Tetrachloro-m-xylene	4.617	4.514	4.714	54.350	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WOOD06
Lab Code: ACE **SDG NO.:** Q3834
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076958.D **Time Analyzed:** 15:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.849	4.750	4.950	496.210	500.000	-0.8
Aroclor-1016-2	4.866	4.767	4.967	507.400	500.000	1.5
Aroclor-1016-3	5.042	4.942	5.142	505.930	500.000	1.2
Aroclor-1016-4	5.083	4.984	5.184	503.980	500.000	0.8
Aroclor-1016-5	5.296	5.197	5.397	501.550	500.000	0.3
Aroclor-1260-1	6.324	6.224	6.424	481.200	500.000	-3.8
Aroclor-1260-2	6.511	6.412	6.612	473.820	500.000	-5.2
Aroclor-1260-3	6.664	6.564	6.764	472.910	500.000	-5.4
Aroclor-1260-4	7.133	7.033	7.233	476.100	500.000	-4.8
Aroclor-1260-5	7.374	7.273	7.473	467.630	500.000	-6.5
Decachlorobiphenyl	8.766	8.666	8.866	44.760	50.000	-10.5
Tetrachloro-m-xylene	3.771	3.671	3.871	51.760	50.000	3.5