

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121125\
 Data File : BP026290.D
 Acq On : 11 Dec 2025 13:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 11 13:24:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.506	0.541	-6.9	85	0.00
3	Pyridine	1.453	1.436	1.2	80	0.00
4	n-Nitrosodimethylamine	0.542	0.534	1.5	80	0.00
5 S	2-Fluorophenol	1.244	1.256	-1.0	79	0.00
6	Aniline	2.061	1.828	11.3	75	0.00
7 S	Phenol-d6	1.570	1.449	7.7	74	0.00
8	2-Chlorophenol	1.414	1.388	1.8	77	0.00
9	Benzaldehyde	0.847	0.977	-15.3	74	0.00
10 C	Phenol	1.689	1.553	8.1	74	0.00
11	bis(2-Chloroethyl)ether	1.324	1.227	7.3	75	0.00
12	1,3-Dichlorobenzene	1.530	1.597	-4.4	80	0.00
13 C	1,4-Dichlorobenzene	1.541	1.595	-3.5	79	0.00
14	1,2-Dichlorobenzene	1.480	1.514	-2.3	79	0.00
15	Benzyl Alcohol	1.142	1.030	9.8	75	0.00
16	2,2'-oxybis(1-Chloropropane	1.848	1.677	9.3	76	-0.01
17	2-Methylphenol	1.146	1.047	8.6	75	0.00
18	Hexachloroethane	0.560	0.595	-6.2	82	-0.01
19 P	n-Nitroso-di-n-propylamine	0.968	0.848	12.4	74	-0.01
20	3+4-Methylphenols	1.583	1.399	11.6	74	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	71	0.00
22	Acetophenone	0.507	0.522	-3.0	75	0.00
23 S	Nitrobenzene-d5	0.351	0.372	-6.0	76	0.00
24	Nitrobenzene	0.355	0.373	-5.1	76	0.00
25	Isophorone	0.689	0.684	0.7	75	-0.01
26 C	2-Nitrophenol	0.180	0.194	-7.8	75	0.00
27	2,4-Dimethylphenol	0.318	0.300	5.7	77	0.00
28	bis(2-Chloroethoxy)methane	0.433	0.427	1.4	74	0.00
29 C	2,4-Dichlorophenol	0.324	0.342	-5.6	75	0.00
30	1,2,4-Trichlorobenzene	0.333	0.365	-9.6	75	-0.01
31	Naphthalene	1.081	1.117	-3.3	74	0.00
32	Benzoic acid	0.255	0.306	-20.0	135	0.03
33	4-Chloroaniline	0.453	0.432	4.6	73	0.00
34 C	Hexachlorobutadiene	0.217	0.255	-17.5	79	-0.01
35	Caprolactam	0.114	0.106	7.0	71	-0.02
36 C	4-Chloro-3-methylphenol	0.341	0.335	1.8	74	0.00
37	2-Methylnaphthalene	0.762	0.776	-1.8	75	-0.01
38	1-Methylnaphthalene	0.744	0.750	-0.8	73	-0.02
39 I	Acenaphthene-d10	1.000	1.000	0.0	71	-0.02
40	1,2,4,5-Tetrachlorobenzene	0.611	0.680	-11.3	75	0.00
41 P	Hexachlorocyclopentadiene	0.381	0.366	3.9	84	0.00
42 S	2,4,6-Tribromophenol	0.260	0.292	-12.3	77	-0.01
43 C	2,4,6-Trichlorophenol	0.415	0.449	-8.2	75	0.00
44	2,4,5-Trichlorophenol	0.462	0.495	-7.1	76	0.00
45 S	2-Fluorobiphenyl	1.376	1.492	-8.4	75	-0.02
46	1,1'-Biphenyl	1.510	1.570	-4.0	73	-0.01
47	2-Chloronaphthalene	1.190	1.251	-5.1	73	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121125\
 Data File : BP026290.D
 Acq On : 11 Dec 2025 13:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 11 13:24:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.326	0.334	-2.5	75	0.00
49	Acenaphthylene	1.953	2.022	-3.5	74	-0.01
50	Dimethylphthalate	1.490	1.551	-4.1	75	-0.02
51	2,6-Dinitrotoluene	0.295	0.322	-9.2	75	-0.01
52 C	Acenaphthene	1.142	1.151	-0.8	73	-0.02
53	3-Nitroaniline	0.344	0.341	0.9	72	-0.02
54 P	2,4-Dinitrophenol	0.157	0.252	-60.5#	318#	-0.01
55	Dibenzofuran	1.776	1.836	-3.4	73	-0.02
56 P	4-Nitrophenol	0.307	0.274	10.7	72	0.00
57	2,4-Dinitrotoluene	0.438	0.476	-8.7	77	-0.01
58	Fluorene	1.405	1.474	-4.9	74	-0.02
59	2,3,4,6-Tetrachlorophenol	0.391	0.426	-9.0	77	-0.02
60	Diethylphthalate	1.495	1.583	-5.9	77	-0.01
61	4-Chlorophenyl-phenylether	0.700	0.753	-7.6	74	-0.02
62	4-Nitroaniline	0.357	0.345	3.4	73	-0.02
63	Azobenzene	1.257	1.263	-0.5	74	-0.02
64 I	Phenanthrene-d10	1.000	1.000	0.0	71	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.148	-25.4#	131	-0.01
66 c	n-Nitrosodiphenylamine	0.618	0.641	-3.7	73	-0.02
67	4-Bromophenyl-phenylether	0.222	0.243	-9.5	75	-0.02
68	Hexachlorobenzene	0.258	0.288	-11.6	76	-0.01
69	Atrazine	0.233	0.243	-4.3	74	-0.02
70 C	Pentachlorophenol	0.179	0.196	-9.5	79	-0.01
71	Phenanthrene	1.127	1.193	-5.9	75	0.00
72	Anthracene	1.148	1.204	-4.9	73	0.00
73	Carbazole	1.087	1.131	-4.0	75	0.00
74	Di-n-butylphthalate	1.284	1.405	-9.4	77	-0.01
75 C	Fluoranthene	1.372	1.484	-8.2	76	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
77	Benzydine	0.628	0.520	17.2	66	0.00
78	Pyrene	1.303	1.318	-1.2	75	0.00
79 S	Terphenyl-d14	0.979	1.046	-6.8	78	0.00
80	Butylbenzylphthalate	0.572	0.593	-3.7	76	0.00
81	Benzo(a)anthracene	1.372	1.422	-3.6	76	0.00
82	3,3'-Dichlorobenzidine	0.499	0.524	-5.0	76	0.00
83	Chrysene	1.299	1.339	-3.1	74	0.00
84	Bis(2-ethylhexyl)phthalate	0.860	0.932	-8.4	81	0.00
85 c	Di-n-octyl phthalate	1.517	1.646	-8.5	79	0.00
86 I	Perylene-d12	1.000	1.000	0.0	72	0.01
87	Indeno(1,2,3-cd)pyrene	1.512	1.659	-9.7	75	0.00
88	Benzo(b)fluoranthene	1.215	1.240	-2.1	73	0.00
89	Benzo(k)fluoranthene	1.216	1.285	-5.7	77	0.00
90 C	Benzo(a)pyrene	1.153	1.208	-4.8	75	0.00
91	Dibenzo(a,h)anthracene	1.236	1.353	-9.5	75	0.02
92	Benzo(g,h,i)perylene	1.232	1.361	-10.5	75	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121125\
Data File : BP026290.D
Acq On : 11 Dec 2025 13:04
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 11 13:24:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 05 14:52:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0