

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: WOOD06
 Lab Code: ACE SDG No.: Q3834
 Instrument ID: MSVOA_Y Calibration Date/Time: 12/11/2025 07:51
 Lab File ID: VY023939.D Init. Calib. Date(s): 12/10/2025 12/10/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:44 13:11
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Acetone	250.000	279.10	-11.6392	20