

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121025

Review By	Semsettin Yesilyurt	Review On	12/11/2025 8:23:18 AM
Supervise By	Maresh Dadoda	Supervise On	12/11/2025 12:00:11 PM
SubDirectory	VY121025	HP Acquire Method	MSVOA_W HP Processing Method 82Y121025S.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136604 VP136606,VP136608,VP136610,VP136612,VP136614,VP136616
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136146 VP136618

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023923.D	10 Dec 2025 07:23		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023924.D	10 Dec 2025 08:01	Need ICAL	SY/MD	Not Ok
3	VIBLK	VIBLK	VY023925.D	10 Dec 2025 08:34		SY/MD	Ok
4	VSTDICC005	VSTDICC005	VY023926.D	10 Dec 2025 08:58	Not used	SY/MD	Not Ok
5	VSTDICC010	VSTDICC010	VY023927.D	10 Dec 2025 09:21	Not used	SY/MD	Not Ok
6	VSTDICC020	VSTDICC020	VY023928.D	10 Dec 2025 09:44		SY/MD	Ok,M
7	VSTDICCC050	VSTDICCC050	VY023929.D	10 Dec 2025 10:06	Comp. #10,16 on Linear Regression	SY/MD	Ok,M
8	VSTDICC100	VSTDICC100	VY023930.D	10 Dec 2025 10:29		SY/MD	Ok,M
9	VSTDICC150	VSTDICC150	VY023931.D	10 Dec 2025 10:52	Not used	SY/MD	Not Ok
10	VIBLK	VIBLK	VY023932.D	10 Dec 2025 11:16		SY/MD	Ok
11	VSTDICC010	VSTDICC010	VY023933.D	10 Dec 2025 11:39		SY/MD	Ok,M
12	VSTDICC005	VSTDICC005	VY023934.D	10 Dec 2025 12:02		SY/MD	Ok,M
13	VSTDICC150	VSTDICC150	VY023935.D	10 Dec 2025 13:11		SY/MD	Ok,M
14	VIBLK	VIBLK	VY023936.D	10 Dec 2025 13:36		SY/MD	Ok
15	VSTDICV050	ICVVY121025	VY023937.D	10 Dec 2025 13:58	ICV Failed for comp. #13	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121125

Review By	Semsettin Yesilyurt	Review On	12/12/2025 11:26:56 AM
Supervise By	Maresh Dadoda	Supervise On	12/12/2025 1:22:49 PM
SubDirectory	VY121125	HP Acquire Method	MSVOA_Y HP Processing Method 82Y121025S.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136626
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136629,VP136630

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023938.D	11 Dec 2025 06:48		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023939.D	11 Dec 2025 07:51		SY/MD	Ok,M
3	VY1211SBL01	VY1211SBL01	VY023940.D	11 Dec 2025 08:57		SY/MD	Ok
4	VY1211SBS01	VY1211SBS01	VY023941.D	11 Dec 2025 09:22		SY/MD	Ok,M
5	VY1211SBSD01	VY1211SBSD01	VY023942.D	11 Dec 2025 09:44		SY/MD	Ok,M
6	Q3831-01	PL-01-12102025	VY023943.D	11 Dec 2025 10:21	vial A	SY/MD	Ok
7	Q3821-01	TP-A-B	VY023944.D	11 Dec 2025 10:45	vial A	SY/MD	Ok
8	Q3827-01	RT4262	VY023945.D	11 Dec 2025 11:09	vial A	SY/MD	Ok
9	VIBLK	VIBLK	VY023946.D	11 Dec 2025 11:32		SY/MD	Ok
10	VY1211SBSD02	VY1211SBSD02	VY023947.D	11 Dec 2025 12:03	Not Required	SY/MD	Not Ok
11	Q3815-01	RE	VY023948.D	11 Dec 2025 12:28	Internal Standard Fail,see non conformance,vial A	SY/MD	Ok,M
12	Q3809-06	B-2	VY023949.D	11 Dec 2025 12:52	vial B	SY/MD	Ok
13	Q3809-12	SW-9	VY023950.D	11 Dec 2025 13:15	vial B	SY/MD	Ok
14	Q3809-13	SW-10	VY023951.D	11 Dec 2025 13:44	vial B	SY/MD	Ok
15	Q3809-14	B-4	VY023952.D	11 Dec 2025 14:08	vial B	SY/MD	Ok
16	Q3809-07RE	B-3RE	VY023953.D	11 Dec 2025 14:31	Internal Standard Fail,vial B	SY/MD	Confirms
17	Q3809-08	SW-5	VY023954.D	11 Dec 2025 14:55	vial B	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY121125

Review By	Semsettin Yesilyurt	Review On	12/12/2025 11:26:56 AM		
Supervise By	Mahesh Dadoda	Supervise On	12/12/2025 1:22:49 PM		
SubDirectory	VY121125	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y121025S.M
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP136626			
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP136629,VP136630			

18	Q3809-11	SW-8	VY023955.D	11 Dec 2025 15:18	vial B	SY/MD	Ok
19	Q3809-09	SW-6	VY023956.D	11 Dec 2025 15:41	vial B	SY/MD	Ok,M
20	VIBLK	VIBLK	VY023957.D	11 Dec 2025 16:05	vial B	SY/MD	Ok
21	Q3834-29	B127(5.5-8)120925	VY023958.D	11 Dec 2025 16:28	Surrogate fail,see non conformance.vial A	SY/MD	Ok,M
22	Q3834-30	B127(8-10)120925	VY023959.D	11 Dec 2025 16:52	Internal Standard Fail,vial A,see non coformance	SY/MD	Ok,M
23	VIBLK	VIBLK	VY023960.D	11 Dec 2025 17:15		SY/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VY023961.D	11 Dec 2025 17:38		SY/MD	Ok,M

M : Manual Integration