

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: WOOD06
Lab Code: ACE SDG NO.: Q3834
Lab File ID: VY023939.D Date Analyzed: 12/11/2025
Instrument ID: MSVOA_Y Time Analyzed: 07:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	547245	7.71	774590	8.62	666641	11.42
UPPER LIMIT	1094490	8.213	1549180	9.122	1333280	11.92
LOWER LIMIT	273623	7.213	387295	8.122	333321	10.92
EPA SAMPLE NO.						
B127(5.5-8)120925	354418	7.71	527728	8.62	482887	11.42
B127(8-10)120925	190433 *	7.71	261771 *	8.62	194490 *	11.42
VY1211SBL01	771211	7.72	1297187	8.62	1132884	11.42
VY1211SBS01	538181	7.71	764222	8.62	660539	11.42
VY1211SBSD01	500795	7.71	721889	8.62	623185	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.