

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: WOOD06
 Lab Code: ACE SDG NO.: Q3834
 Lab File ID: VY023939.D Date Analyzed: 12/11/2025
 Instrument ID: MSVOA_Y Time Analyzed: 07:51
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	349523	13.353				
UPPER LIMIT	699046	13.853				
LOWER LIMIT	174762	12.853				
EPA SAMPLE NO.						
B127(5.5-8)120925	546103	13.35				
B127(8-10)120925	81268 *	13.35				
VY1211SBL01	487980	13.35				
VY1211SBS01	353090	13.35				
VY1211SBSD01	329812	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.