

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>JAC005</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3835</u>
Instrument ID: <u>MSVOA_V</u>	Calibration Date(s): <u>12/03/2025</u> <u>12/03/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>09:55</u> <u>12:18</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRFAL1 = VV039445.D		RRFAL2 = VV039446.D		RRFAL3 = VV039447.D			
		RRFAL4 = VV039448.D		RRFAL5 = VV039449.D		RRFAL6 = VV039450.D			
COMPOUND	RRFAL1	RRFAL2	RRFAL3	RRFAL4	RRFAL5	RRFAL6	RRF	% RSD	
Vinyl chloride	1.464	0.963	1.023	1.100	0.991	0.950	1.082	18	
Vinyl Chloride-d3		0.473	0.499	0.533	0.491	0.458	0.491	5.8	
1,2-Dichloroethane-d4		0.383	0.341	0.334	0.318	0.288	0.333	10.3	
1,2-Dichloropropane-d6		0.397	0.402	0.463	0.425	0.422	0.422	6.2	
Toluene-d8		0.674	0.616	0.726	0.738	0.805	0.712	10	
1,1,2,2-Tetrachloroethane-d2		0.215	0.249	0.262	0.243	0.220	0.238	8.4	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.