

Surrogate Summary

SDG No.: **Q3835**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SWSFAM_VOCSIM**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3835-03	OWBR-01-170-121025-SIM	Vinyl chloride-d3	0.5	0.46	92		40	130
		1,2-Dichloroethane-d4	0.5	0.38	75		70	130
		1,2-Dichloropropane-d6	0.5	0.37	73		60	140
		Toluene-d8	0.5	0.42	85		70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.48	95		65	120
Q3835-06	TB01-121025	Vinyl chloride-d3	0.5	0.52	104		40	130
		1,2-Dichloroethane-d4	0.5	0.46	91		70	130
		1,2-Dichloropropane-d6	0.5	0.42	83		60	140
		Toluene-d8	0.5	0.46	91		70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.54	108		65	120
Q3835-07	VHBLK001	Vinyl chloride-d3	0.5	0.53	106		40	130
		1,2-Dichloroethane-d4	0.5	0.43	86		70	130
		1,2-Dichloropropane-d6	0.5	0.41	81		60	140
		Toluene-d8	0.5	0.45	90		70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.52	103		65	120
VV1211WBL01	VBLK240	Vinyl chloride-d3	0.5	0.61	122		40	130
		1,2-Dichloroethane-d4	0.5	0.50	100		70	130
		1,2-Dichloropropane-d6	0.5	0.47	94		60	140
		Toluene-d8	0.5	0.52	104		70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.59	118		65	120