

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120425

Review By	John Carlone	Review On	12/5/2025 8:05:59 AM
Supervise By	Semsettin Yesilyurt	Supervise On	12/5/2025 2:14:46 PM
SubDirectory	VX120425	HP Acquire Method	HP Processing Method 82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP136497
Initial Calibration Stds	VP136526,VP136527,VP136528,VP136529,VP136530,VP136531
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP136532
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048713.D	04 Dec 2025 08:15		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX048714.D	04 Dec 2025 08:47		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX048715.D	04 Dec 2025 09:16		JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX048716.D	04 Dec 2025 09:36	Comp. #13 on Linear Regression	JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX048717.D	04 Dec 2025 09:57	Comp. #95 on Quadratic Regression	JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX048718.D	04 Dec 2025 10:17		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX048719.D	04 Dec 2025 10:38		JC/MD	Ok,M
8	IBLK	IBLK	VX048720.D	04 Dec 2025 10:58		JC/MD	Ok
9	VSTDICV050	ICVVX120425	VX048721.D	04 Dec 2025 14:04		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121125

Review By	John Carlone	Review On	12/12/2025 8:45:14 AM
Supervise By	Maresh Dadoda	Supervise On	12/12/2025 11:46:33 AM
SubDirectory	VX121125	HP Acquire Method	MSVOA_X HP Processing Method 82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136631
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136632,VP136633 VP136145

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048821.D	11 Dec 2025 12:08		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048822.D	11 Dec 2025 13:02	pH#Lot#V12668	JC/MD	Ok,M
3	VX1211MBL01	VX1211MBL01	VX048823.D	11 Dec 2025 13:31		JC/MD	Ok
4	VX1211WBL01	VX1211WBL01	VX048824.D	11 Dec 2025 14:06		JC/MD	Ok
5	VX1211WBS01	VX1211WBS01	VX048825.D	11 Dec 2025 14:31		JC/MD	Ok,M
6	VX1211WBSD01	VX1211WBSD01	VX048826.D	11 Dec 2025 14:52		JC/MD	Ok,M
7	Q3828-08	FB01-120925	VX048827.D	11 Dec 2025 15:37	vial A pH<2 FB	JC/MD	Ok
8	Q3828-09	TB01-120925	VX048828.D	11 Dec 2025 15:58		JC/MD	Ok
9	Q3835-06	TB01-121025	VX048829.D	11 Dec 2025 16:19	vial A pH<2 TB	JC/MD	Ok
10	Q3835-01	MW-11A-13.5-121025	VX048830.D	11 Dec 2025 16:39	vial A pH<2 Need 50X	JC/MD	Dilution
11	Q3835-02	OWBR-01-170-121025	VX048831.D	11 Dec 2025 17:00	vial A pH<2 E flag of previous sample	JC/MD	ReRun
12	Q3835-04	MW-11B-37.5-121025	VX048832.D	11 Dec 2025 17:21	vial A pH<2 Need 100X	JC/MD	Dilution
13	Q3835-05	MW-01-7-121025	VX048833.D	11 Dec 2025 17:42	vial A pH<2 E flag of previous sample	JC/MD	ReRun
14	Q3828-01	BR-02-133-120925	VX048834.D	11 Dec 2025 18:02	Confirm Concentration	JC/MD	Not Ok
15	Q3828-03	OWBR-06-219-120925	VX048835.D	11 Dec 2025 18:23	Confirm Concentration	JC/MD	Not Ok
16	Q3828-05	RMW-03B-90-120925	VX048836.D	11 Dec 2025 18:44	vial A pH<2 Need 20X	JC/MD	Dilution
17	Q3828-06	OWBR-03-138-120925	VX048837.D	11 Dec 2025 19:05	vial A pH<2 E flag of previous sample	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121125

Review By	John Carlone	Review On	12/12/2025 8:45:14 AM
Supervise By	Maresh Dadoda	Supervise On	12/12/2025 11:46:33 AM
SubDirectory	VX121125	HP Acquire Method	MSVOA_X HP Processing Method 82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136631
CCC	VP136632,VP136633
Internal Standard/PEM	VP136145
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	VSTDCCC050	VSTDCCC050EC	VX048838.D	11 Dec 2025 19:25		JC/MD	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121225

Review By	Mahesh Dadoda	Review On	12/16/2025 1:08:47 PM
Supervise By	Semsettin Yesilyurt	Supervise On	12/16/2025 3:06:43 PM
SubDirectory	VX121225	HP Acquire Method	HP Processing Method 82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136634
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136635,VP136636

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048839.D	12 Dec 2025 08:44		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048840.D	12 Dec 2025 09:12	pH#lot#v14222	JC/MD	Ok,M
3	VX1212MBL01	VX1212MBL01	VX048841.D	12 Dec 2025 10:09		JC/MD	Ok
4	VX1212WBL01	VX1212WBL01	VX048842.D	12 Dec 2025 10:31		JC/MD	Ok
5	VX1212WBS01	VX1212WBS01	VX048843.D	12 Dec 2025 10:52	Comp. #24 Recovery failed high	JC/MD	Ok,M
6	VX1212WBSD01	VX1212WBSD01	VX048844.D	12 Dec 2025 11:20	Comp. #24 Recovery failed high	JC/MD	Ok,M
7	Q3848-01	MW-18B-57.5-121125	VX048845.D	12 Dec 2025 11:42	vial A,pH<2 Need 10X	JC/MD	Dilution
8	IBLK	IBLK	VX048846.D	12 Dec 2025 12:04		JC/MD	Ok
9	Q3828-05DL	RMW-03B-90-120925D	VX048847.D	12 Dec 2025 12:26		JC/MD	Ok
10	Q3835-01DL	MW-11A-13.5-121025D	VX048848.D	12 Dec 2025 12:48		JC/MD	Ok
11	Q3835-04DL	MW-11B-37.5-121025D	VX048849.D	12 Dec 2025 13:09		JC/MD	Ok
12	Q3828-01	BR-02-133-120925	VX048850.D	12 Dec 2025 13:31	vial B,pH<2	JC/MD	Ok
13	Q3828-03	OWBR-06-219-120925	VX048851.D	12 Dec 2025 13:53	vial B,pH<2	JC/MD	Ok
14	Q3828-06	OWBR-03-138-120925	VX048852.D	12 Dec 2025 14:15	vial B,pH<2	JC/MD	Ok
15	Q3835-02	OWBR-01-170-121025	VX048853.D	12 Dec 2025 14:36	vial B,pH<2 Surrogate Fail	JC/MD	ReRun
16	Q3835-05	MW-01-7-121025	VX048854.D	12 Dec 2025 14:58	vial B,pH<2	JC/MD	Ok
17	Q3848-03	MW-18B-57.5-121125-F	VX048855.D	12 Dec 2025 15:20	vial A,pH<2,Surrogate Fail;Need 10X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121225

Review By	Maresh Dadoda	Review On	12/16/2025 1:08:47 PM
Supervise By	Semsettin Yesilyurt	Supervise On	12/16/2025 3:06:43 PM
SubDirectory	VX121225	HP Acquire Method	HP Processing Method 82X120425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP136634		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136635,VP136636		

18	Q3849-01	OW-07B-65-121125	VX048856.D	12 Dec 2025 15:42	vial A,pH<2 E,surrogate fail	JC/MD	ReRun
19	Q3849-07	RMW-02B-67-121125	VX048857.D	12 Dec 2025 16:03	vial A,pH<2 ,Need 50X	JC/MD	Dilution
20	Q3849-08	MW-06-7-121125	VX048858.D	12 Dec 2025 16:25	vial A,pH<2 Surrogate Fail; Need 200X	JC/MD	Dilution
21	Q3849-09	MW-19B-71.5-121125	VX048859.D	12 Dec 2025 16:47	vial A,pH<2 ,Need 100X	JC/MD	Dilution
22	Q3849-11	TB01-121125	VX048860.D	12 Dec 2025 17:09	vial A,pH<2 ,E flag of previous sample	JC/MD	ReRun
23	Q3848-01DL	MW-18B-57.5-121125	VX048861.D	12 Dec 2025 17:31		JC/MD	Ok
24	Q3813-01	BR-04DR-451-120825	VX048862.D	12 Dec 2025 17:52	carryover	JC/MD	Not Ok
25	Q3826-12	TP-2	VX048863.D	12 Dec 2025 18:14	Surrogate Fail;carryover	JC/MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VX048864.D	12 Dec 2025 18:36		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121525

Review By	Mahesh Dadoda	Review On	12/16/2025 12:57:31 PM
Supervise By	Semsettin Yesilyurt	Supervise On	12/16/2025 3:07:47 PM
SubDirectory	VX121525	HP Acquire Method	MSVOA_X HP Processing Method 82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136641
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136642,VP136645 VP136145

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048865.D	15 Dec 2025 08:18		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048866.D	15 Dec 2025 08:54	pH#lot#v14222	JC/MD	Ok,M
3	VX1215MBL01	VX1215MBL01	VX048867.D	15 Dec 2025 09:42		JC/MD	Ok
4	VX1215WBL01	VX1215WBL01	VX048868.D	15 Dec 2025 10:03		JC/MD	Ok
5	VX1215WBS01	VX1215WBS01	VX048869.D	15 Dec 2025 10:45		JC/MD	Ok,M
6	VX1215WBSD01	VX1215WBSD01	VX048870.D	15 Dec 2025 11:13		JC/MD	Ok,M
7	Q3848-03DL	MW-18B-57.5-121125-F	VX048871.D	15 Dec 2025 11:50	Surrogate fail	JC/MD	Not Ok
8	Q3835-02	OWBR-01-170-121025	VX048872.D	15 Dec 2025 12:10	vial C pH<2	JC/MD	Ok
9	Q3848-03DL	MW-18B-57.5-121125-F	VX048873.D	15 Dec 2025 12:31	Not match exactly	JC/MD	Not Ok
10	Q3863-01	MW-17B-55-121225	VX048874.D	15 Dec 2025 12:52	vial A pH<2 ,Need 100X	JC/MD	Dilution
11	VIBLK	VIBLK	VX048875.D	15 Dec 2025 13:13		JC/MD	Ok
12	VIBLK	VIBLK	VX048876.D	15 Dec 2025 14:20		JC/MD	Ok
13	Q3848-03DL	MW-18B-57.5-121125-F	VX048877.D	15 Dec 2025 14:44		JC/MD	Ok
14	Q3863-01DL	MW-17B-55-121225DL	VX048878.D	15 Dec 2025 15:08		JC/MD	Ok
15	Q3863-10	EB01-121225	VX048879.D	15 Dec 2025 15:43	vial A pH<2 ,EB	JC/MD	Ok
16	Q3863-12	TB01-121225	VX048880.D	15 Dec 2025 16:03	vial A pH<2 ,TB	JC/MD	Ok
17	Q3863-02MS	MW-17B-55-121225MS	VX048881.D	15 Dec 2025 16:24	vial A pH<2	JC/MD	Ok,M
18	Q3863-03MSD	MW-17B-55-121225MS	VX048882.D	15 Dec 2025 16:45	vial A pH<2	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX121525

Review By	Mahesh Dadoda	Review On	12/16/2025 12:57:31 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	12/16/2025 3:07:47 PM		
SubDirectory	VX121525	HP Acquire Method	MSVOA_X	HP Processing Method	82X120425W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136641
CCC	VP136642,VP136645
Internal Standard/PEM	VP136145
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VIBLK	VIBLK	VX048883.D	15 Dec 2025 17:05		JC/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VX048884.D	15 Dec 2025 17:26		JC/MD	Ok,M

M : Manual Integration