

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3835</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>JACOBS Engineering Group, Inc.</u>	Datafile :	<u>VX048826.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1211WBSD01	Vinyl chloride	20	18.6	ug/L	93	10		70 (65)	130 (117)	20 (19)
	1,1-Dichloroethene	20	19.2	ug/L	96	13		70 (74)	130 (110)	20 (20)
	1,1-Dichloroethane	20	22.3	ug/L	112	13		70 (78)	130 (112)	20 (20)
	cis-1,2-Dichloroethene	20	21.4	ug/L	107	12		70 (77)	130 (110)	20 (20)
	1,1,1-Trichloroethane	20	21.1	ug/L	106	14		70 (80)	130 (108)	20 (20)
	Benzene	20	21.4	ug/L	107	15		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	21.7	ug/L	109	16		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	20.9	ug/L	104	19		70 (77)	130 (113)	20 (15)
	1,1,2-Trichloroethane	20	21.8	ug/L	109	19		70 (83)	130 (112)	20 (20)
	Tetrachloroethene	20	19.9	ug/L	100	23	*	70 (67)	130 (123)	20 (15)