

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3835
 Lab File ID: VX048822.D Date Analyzed: 12/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 13:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	165373	5.53	271652	6.74	243221	10.04
UPPER LIMIT	330746	6.031	543304	7.239	486442	10.537
LOWER LIMIT	82686.5	5.031	135826	6.239	121611	9.537
EPA SAMPLE NO.						
MW-11A-13.5-121025	195889	5.54	321296	6.75	267936	10.04
MW-11B-37.5-121025	184411	5.54	307688	6.75	247986	10.04
TB01-121025	218560	5.54	357885	6.75	316450	10.04
VX1211WBL01	251542	5.53	411094	6.74	347963	10.04
VX1211WBS01	197059	5.53	327688	6.74	298607	10.04
VX1211WBSD01	160103	5.54	261734	6.75	237436	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.