

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
Lab Code: ACE SDG NO.: Q3835
Lab File ID: VX048840.D Date Analyzed: 12/12/2025
Instrument ID: MSVOA_X Time Analyzed: 09:12
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	152770	5.53	255110	6.74	228178	10.04
UPPER LIMIT	305540	6.031	510220	7.238	456356	10.537
LOWER LIMIT	76385	5.031	127555	6.238	114089	9.537
EPA SAMPLE NO.						
MW-11A-13.5-121025DL	143724	5.54	302254	6.75	330363	10.04
MW-11B-37.5-121025DL	120522	5.54	248427	6.74	265146	10.04
MW-01-7-121025	121695	5.54	251822	6.75	279208	10.04
VX1212WBL01	157967	5.54	331453	6.75	357991	10.04
VX1212WBS01	146797	5.54	247822	6.74	232443	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.